



DPC: 18/30384670 DC

BSI Group Headquarters

389 Chiswick High Road London W4 4AL

Tel: +44 (0)20 8996 9000

Fax: +44 (0)20 8996 7400

www.bsigroup.com

Date: 21 November 2018

Origin: International

Latest date for receipt of comments: 09 January 2019

Project No. 2018/03085

Responsible committee: EPL/29 Electroacoustics

Interested committees:

Title: Draft BS IEC 63219 Definition and verification of hearing aid features.

Please notify the secretary if you are aware of any keywords that might assist in classifying or identifying the standard or if the content of this standard

- i) has any issues related to 3rd party IPR, patent or copyright
- ii) affects other national standard(s)
- iii) requires additional national guidance or information

WARNING: THIS IS A DRAFT AND MUST NOT BE REGARDED OR USED AS A BRITISH STANDARD. THIS DRAFT IS NOT CURRENT BEYOND 09 January 2019

This draft is issued to allow comments from interested parties; all comments will be given consideration prior to publication. No acknowledgement will normally be sent. **See overleaf for information on the submission of comments.**

No copying is allowed, in any form, without prior written permission from BSI except as permitted under the Copyright, Designs and Patent Act 1988 or for circulation within a nominating organization for briefing purposes. Electronic circulation is limited to dissemination by e-mail within such an organization by committee members.

Further copies of this draft may be purchased from BSI Shop <http://shop.bsigroup.com> or from BSI Customer Services, Tel: +44(0) 20 8996 9001 or email cservices@bsigroup.com. British, International and foreign standards are also available from BSI Customer Services.

Information on the co-operating organizations represented on the committees referenced above may be obtained from <https://standardsdevelopment.bsigroup.com/>.

Introduction

This draft standard is based on national and international discussions. Your comments on this draft are invited and will assist in the preparation of the consequent standard.

For international standards, comments will be reviewed by the relevant UK national committee before sending the consensus UK vote and comments to the international committee, which will then decide appropriate action. If the international standard is approved, it is usual for the text to be published as a British Standard.

For national standards, comments will be reviewed by the relevant UK national committee and the resulting standards published as a British Standard.

UK Vote

Please indicate whether you consider the UK should submit a negative (with supporting technical reasons) or positive vote on this draft. Please indicate if you are aware of any reason why this draft standard should not be published as a British Standard.

Submission of Comments

- **Annotated drafts are not acceptable and will be rejected.**
- All comments should be submitted online at <https://standardsdevelopment.bsigroup.com/>. You will need to register in order to comment.

Template for comments and secretariat observations

Date: xx/xx/20xx	Document: ISO/DIS xxxx
------------------	------------------------

1	2	(3)	4	5	(6)	7
MB	Clause No./ Subclause No./Annex (e.g. 3.1)	Paragraph/Figure/ Table/Note	Type of comment	Commend (justification for change) by the MB	Proposed change by the MB	Secretariat observations on each comment submitted
	3.1	Definition 1	ed	Definition is ambiguous and needs clarifying.	Amend to read '...so that the mains connector to which no connection...'	
	6.4	Paragraph 2	te	The use of the UV photometer as an alternative cannot be supported as serious problems have been encountered in its use in the UK.	Delete reference to UV photometer.	



29/1012/CD

COMMITTEE DRAFT (CD)

PROJECT NUMBER:

IEC 63219 ED1

DATE OF CIRCULATION:

2018-11-16

CLOSING DATE FOR COMMENTS:

2019-02-08

SUPERSEDES DOCUMENTS:

29/997/NP, 29/1006A/RVN

IEC TC 29 : ELECTROACOUSTICS

SECRETARIAT:

Denmark

SECRETARY:

Ms Lise Aagesen

OF INTEREST TO THE FOLLOWING COMMITTEES:

PROPOSED HORIZONTAL STANDARD:



Other TC/SCs are requested to indicate their interest, if any, in this CD to the secretary.

FUNCTIONS CONCERNED:

☐ EMC

☐ ENVIRONMENT

☐ QUALITY ASSURANCE

☐ SAFETY

This document is still under study and subject to change. It should not be used for reference purposes.

Recipients of this document are invited to submit, with their comments, notification of any relevant patent rights of which they are aware and to provide supporting documentation.


TITLE:

Definition and verification of hearing aid features

NOTE FROM TC/SC OFFICERS:

A NWIP for definition and verification of hearing aid features, doc. 29/997/NP, was circulated for voting by 2018-08-31 and approved as given in doc. 29/1006/RVN. The item was allocated to WG 13 "Hearing aids" with Mr Hendrik Hussted, Germany, as project leader. The comments were discussed at the last WG 13 meeting, September 2018 in Ottawa, and consequently, IEC/TC 29 took a decision to proceed with the NP as a first CD, doc. 29/1005/DL, Decision 11 refers.

Copyright © 2018 International Electrotechnical Commission, IEC. All rights reserved. It is permitted to download this electronic file, to make a copy and to print out the content for the sole purpose of preparing National Committee positions. You may not copy or "mirror" the file or printed version of the Document, or any part of it, for any other purpose without permission in writing from IEC.



Documents

Document Sharing Hub

CONTENTS

FOREWORD.....	3
1 Scope.....	5
2 Normative references	5
3 Terms and definitions	5
4 Application to non air conduction hearing aids	7
5 Examples of test signals for common hearing situations	8
6 Verification of noise reduction	8
7 Strategies of hearing aid programs and their verification	9
7.1 Manual selection of hearing aid programs	9
7.1.1 Description	9
7.1.2 Verification	9
7.2 Automatic selection of hearing aid programs	10
7.2.1 Description	10
7.2.2 Verification for adjustable hearing aid programs	10
7.2.3 Plausibility check for not-configurable hearing aid programs	11
8 Verification of feedback reduction	11
8.1 Introduction.....	11
8.2 Attachment of the hearing aid	11
8.2.1 General remarks.....	11
8.2.2 Attachment for air conduction hearing aids	11
8.3 Measurement procedure	12
8.4 Post Processing.....	13
9 Verification of a multichannel signal processing.....	14
9.1 Visualization of the effect of multichannel signal processing	14
9.2 Evaluation of the number of independent channels	14
10 Verification of an output limiter	15
Bibliography.....	16
Figure 1 – Visualization of the manual selection of hearing aid programs (HAP).	9
Figure 2 – Visualization of the automatic selection of hearing aid programs (HAP).	10
Figure 3 – Attachment for air conduction hearing aids for the verification of feedback reduction.	12
Figure 4 – Example of the plot of the results of the feedback reduction measurement.....	13
Table 1 – Examples of test signals for different hearing situations	8
Table 2 – Symbols used for the results of the feedback reduction measurement.	13
Table 3 – Example results of the procedure for the verification of a multichannel signal processing	15

INTERNATIONAL ELECTROTECHNICAL COMMISSION

DEFINITION AND VERIFICATION OF HEARING AID FEATURES

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 63219 has been prepared by IEC technical committee 29: Electroacoustics.

The text of this International Standard is based on the following documents:

FDIS	Report on voting
XX/XX/FDIS	XX/XX/RVD

Full information on the voting for the approval of this International Standard can be found in the report on voting indicated in the above table.

This document has been drafted in accordance with the ISO/IEC Directives, Part 2.

94 The committee has decided that the contents of this document will remain unchanged until the
95 stability date indicated on the IEC website under "http://webstore.iec.ch" in the data related to
96 the specific document. At this date, the document will be

- 97 • reconfirmed,
- 98 • withdrawn,
- 99 • replaced by a revised edition, or
- 100 • amended.

101

102 The National Committees are requested to note that for this document the stability date
103 is 20XX..

104 THIS TEXT IS INCLUDED FOR THE INFORMATION OF THE NATIONAL COMMITTEES AND WILL BE
105 DELETED AT THE PUBLICATION STAGE.

106

107