



DPC: 21/30390836 DC

BSI Group Headquarters

389 Chiswick High Road London W4 4AL

Tel: +44 (0)20 8996 9000

Fax: +44 (0)20 8996 7400

www.bsigroup.com

Date: 22 July 2021

Origin: European

Latest date for receipt of comments: 14 September 2021

Project No. 2019/00308

Responsible committee: CH/150/4 Surgical Implants - Bone and Joint Replacements

Interested committees:

Title: Draft BS EN ISO 21536 Non-active surgical implants — Joint replacement implants — Specific requirements for knee-joint replacement implants.

Please notify the secretary if you are aware of any keywords that might assist in classifying or identifying the standard or if the content of this standard

- i) has any issues related to 3rd party IPR, patent or copyright
- ii) affects other national standard(s)
- iii) requires additional national guidance or information

WARNING: THIS IS A DRAFT AND MUST NOT BE REGARDED OR USED AS A BRITISH STANDARD. THIS DRAFT IS NOT CURRENT BEYOND 14 September 2021

This draft is issued to allow comments from interested parties; all comments will be given consideration prior to publication. No acknowledgement will normally be sent. **See overleaf for information on the submission of comments.**

No copying is allowed, in any form, without prior written permission from BSI except as permitted under the Copyright, Designs and Patent Act 1988 or for circulation within a nominating organization for briefing purposes. Electronic circulation is limited to dissemination by e-mail within such an organization by committee members.

Further copies of this draft may be purchased from BSI Shop <http://shop.bsigroup.com> or from BSI Customer Services, Tel: +44(0) 20 8996 9001 or email cservices@bsigroup.com. British, International and foreign standards are also available from BSI Customer Services.

Information on the co-operating organizations represented on the committees referenced above may be obtained from <https://standardsdevelopment.bsigroup.com/>.

Responsible Committee Secretary: Ms Helena Barrell (BSI)

Direct tel:

E-mail: helena.barrell@bsigroup.com

Full document details: 21/30390836 DC

Introduction

This draft standard is based on national and international discussions. Your comments on this draft are invited and will assist in the preparation of the consequent standard.

For international standards, comments will be reviewed by the relevant UK national committee before sending the consensus UK vote and comments to the international committee, which will then decide appropriate action. If the international standard is approved, it is usual for the text to be published as a British Standard.

For national standards, comments will be reviewed by the relevant UK national committee and the resulting standards published as a British Standard.

UK Vote

Please indicate whether you consider the UK should submit a negative (with supporting technical reasons) or positive vote on this draft. Please indicate if you are aware of any reason why this draft standard should not be published as a British Standard.

Submission of Comments

- **Annotated drafts are not acceptable and will be rejected.**
- All comments should be submitted online at <https://standardsdevelopment.bsigroup.com/>. You will need to register in order to comment.

Template for comments and secretariat observations

Date: xx/xx/20xx	Document: ISO/DIS xxxx
------------------	------------------------

1	2	(3)	4	5	(6)	7
MB	Clause No./ Subclause No./Annex (e.g. 3.1)	Paragraph/Figure/ Table/Note	Type of comment	Commend (justification for change) by the MB	Proposed change by the MB	Secretariat observations on each comment submitted
	3.1	Definition 1	ed	Definition is ambiguous and needs clarifying.	Amend to read '...so that the mains connector to which no connection...'	
	6.4	Paragraph 2	te	The use of the UV photometer as an alternative cannot be supported as serious problems have been encountered in its use in the UK.	Delete reference to UV photometer.	

DRAFT INTERNATIONAL STANDARD

ISO/DIS 21536

ISO/TC 150/SC 4

Secretariat: **BSI**

Voting begins on:
2021-07-22

Voting terminates on:
2021-10-14

Non-active surgical implants — Joint replacement implants — Specific requirements for knee-joint replacement implants

ICS: 11.040.40

THIS DOCUMENT IS A DRAFT CIRCULATED FOR COMMENT AND APPROVAL. IT IS THEREFORE SUBJECT TO CHANGE AND MAY NOT BE REFERRED TO AS AN INTERNATIONAL STANDARD UNTIL PUBLISHED AS SUCH.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNOLOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STANDARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION.

This document is circulated as received from the committee secretariat.

ISO/CEN PARALLEL PROCESSING



Reference number
ISO/DIS 21536:2021(E)

AVA Documents
Document Sharing Hub
© ISO 2021



COPYRIGHT PROTECTED DOCUMENT

© ISO 2021

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	2
4 Intended performance	5
5 Design attributes	6
5.1 General	6
5.2 Thickness of knee joint components	6
5.2.1 General	6
5.2.2 Thickness of UHMWPE in tibial inserts, monobloc tibial components, mobile-bearing components, patella inserts and monobloc patellar components	7
5.2.3 Thickness of metal and ceramic in femoral components, tibial trays, mobile-bearing components and patellar trays	7
5.3 Surface finish of non-articulating regions of knee joint components	7
5.4 Surface finish of articulating surfaces of knee joint components	7
6 Materials	7
7 Design evaluation	8
7.1 General	8
7.2 Pre-clinical evaluation	8
7.2.1 General	8
7.2.2 Test methods and performance requirements	10
7.3 Clinical investigation	15
8 Manufacture	16
9 Sterilization	16
10 Packaging	16
11 Information to be supplied by the manufacturer	16
11.1 General	16
11.2 Product type and dimensions	16
11.3 Constructional and functional compatibility of components	17
11.4 Marking	17
11.5 Information for the patient	17
11.6 Electronic instructions for use	17
Annex A (informative) Evaluation of range of relative angular motion of components of fully constrained total knee joint replacement implants	18
Annex ZA (informative) Relationship between this European standard and the General Safety and Performance Requirements of Regulation (EU) 2017/745 aimed to be covered	19
Bibliography	22

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 21536 was prepared by Technical Committee ISO/TC 150, *Implants for surgery*, Subcommittee SC 4, *Bone and joint replacements*.

This third edition cancels and replaces the second edition (ISO 21536:2007), which has been technically revised.

Introduction

There are three levels of standards dealing with non-active surgical implants. These are as follows, with level 1 being the highest:

- level 1: general requirements for non-active surgical implants and instrumentation used in association with implants;
- level 2: particular requirements for families of non-active surgical implants;
- level 3: specific requirements for types of non-active surgical implant.

This standard is a level 3 standard and contains requirements applying specifically to knee joint replacements.

The level 1 standard, ISO 14630, contains requirements that apply to all non-active surgical implants. It also indicates that there are additional requirements in the level 2 and level 3 standards.

The level 2 standards apply to more restricted sets or families of implants such as those designed for use in osteosynthesis, cardiovascular surgery or joint replacement. For joint replacement implants the level 2 standard is ISO 21534.

To address all requirements, it is recommended that a standard of the lowest available level be consulted first.

Non-active surgical implants — Joint replacement implants — Specific requirements for knee-joint replacement implants

1 Scope

This document provides specific requirements for knee joint replacement implants. With regard to safety, this document specifies requirements for intended performance, design attributes, materials, design evaluation, manufacture, sterilization, packaging, information supplied by the manufacturer and methods of test.

This document applies to both total and partial knee joint replacement implants. It applies to these replacements both with and without the replacement of the patella-femoral joint. It applies to components made of metallic and non-metallic materials.

This document applies to a wide variety of knee replacement implants, but for some specific knee replacement implant types, some considerations, not specifically covered in this document, may be applicable. Further details are given in [Clause 7.2.1.1](#).

The requirements which are specified in this document are not intended to require the re-design or re-testing of devices which have been legally marketed and for which there is a history of sufficient and safe clinical use. For such devices compliance with this document shall be demonstrated by providing evidence of the sufficient and safe clinical use.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 7207-1:2007, *Implants for surgery — Components for partial and total knee joint prostheses — Part 1: Classification, definitions and designation of dimensions*

ISO 7207-2, *Implants for surgery — Components for partial and total knee joint prostheses — Part 2: Articulating surfaces made of metal, ceramic and plastics materials*

ISO 14243-1, *Implants for surgery — Wear of total knee-joint prostheses — Part 1: Loading and displacement parameters for wear-testing machines with load control and corresponding environmental conditions for test*

ISO 14243-2, *Implants for surgery — Wear of total knee-joint prostheses — Part 2: Methods of measurement*

ISO 14243-3, *Implants for surgery — Wear of total knee-joint prostheses — Part 3: Loading and displacement parameters for wear-testing machines with displacement control and corresponding environmental conditions for test*

ISO 14243-5, *Implants for surgery — Wear of total knee prostheses — Part 5: Durability performance of the patellofemoral joint*

ISO 14630, *Non-active surgical implants — General requirements*

ISO 14879-1, *Implants for surgery — Total knee-joint prostheses — Part 1: Determination of endurance properties of knee tibial trays*

ISO 17853, *Wear of implant materials — Polymer and metal wear particles — Isolation and characterization*