



DPC: 21/30433083 DC

BSI Group Headquarters

389 Chiswick High Road London W4 4AL

Tel: +44 (0)20 8996 9000

Fax: +44 (0)20 8996 7400

www.bsigroup.com

Date: 21 May 2021

Origin: International

Latest date for receipt of comments: 16 June 2021

Project No. 2021/00228

Responsible committee: CH/62/4 Electromedical equipment

Interested committees:

Title: Draft BS EN IEC 60601-2-2 Amd.1 Amendment 1 - Medical electrical equipment. Part 2-2: Particular requirements for the basic safety and essential performance of high frequency surgical equipment and high frequency surgical accessories.

Please notify the secretary if you are aware of any keywords that might assist in classifying or identifying the standard or if the content of this standard

- i) has any issues related to 3rd party IPR, patent or copyright
- ii) affects other national standard(s)
- iii) requires additional national guidance or information

WARNING: THIS IS A DRAFT AND MUST NOT BE REGARDED OR USED AS A BRITISH STANDARD. THIS DRAFT IS NOT CURRENT BEYOND 16 June 2021

This draft is issued to allow comments from interested parties; all comments will be given consideration prior to publication. No acknowledgement will normally be sent. **See overleaf for information on the submission of comments.**

No copying is allowed, in any form, without prior written permission from BSI except as permitted under the Copyright, Designs and Patent Act 1988 or for circulation within a nominating organization for briefing purposes. Electronic circulation is limited to dissemination by e-mail within such an organization by committee members.

Further copies of this draft may be purchased from BSI Shop <http://shop.bsigroup.com> or from BSI Customer Services, Tel: +44(0) 20 8996 9001 or email cservices@bsigroup.com. British, International and foreign standards are also available from BSI Customer Services.

Information on the co-operating organizations represented on the committees referenced above may be obtained from <https://standardsdevelopment.bsigroup.com/>.

Responsible Committee Secretary: Miss Jacinta Hampton (BSI)

Direct tel:

E-mail: jacinta.hampton@bsigroup.com

Full document details: 21/30433083 DC

Introduction

This draft standard is based on national and international discussions. Your comments on this draft are invited and will assist in the preparation of the consequent standard.

For international standards, comments will be reviewed by the relevant UK national committee before sending the consensus UK vote and comments to the international committee, which will then decide appropriate action. If the international standard is approved, it is usual for the text to be published as a British Standard.

For national standards, comments will be reviewed by the relevant UK national committee and the resulting standards published as a British Standard.

UK Vote

Please indicate whether you consider the UK should submit a negative (with supporting technical reasons) or positive vote on this draft. Please indicate if you are aware of any reason why this draft standard should not be published as a British Standard.

Submission of Comments

- **Annotated drafts are not acceptable and will be rejected.**
- All comments should be submitted online at <https://standardsdevelopment.bsigroup.com/>. You will need to register in order to comment.

Template for comments and secretariat observations

Date: xx/xx/20xx	Document: ISO/DIS xxxx
------------------	------------------------

1	2	(3)	4	5	(6)	7
MB	Clause No./ Subclause No./Annex (e.g. 3.1)	Paragraph/Figure/ Table/Note	Type of comment	Commend (justification for change) by the MB	Proposed change by the MB	Secretariat observations on each comment submitted
	3.1	Definition 1	ed	Definition is ambiguous and needs clarifying.	Amend to read '...so that the mains connector to which no connection...'	
	6.4	Paragraph 2	te	The use of the UV photometer as an alternative cannot be supported as serious problems have been encountered in its use in the UK.	Delete reference to UV photometer.	



62D/1861/CD

COMMITTEE DRAFT (CD)

PROJECT NUMBER:

IEC 60601-2-2/AMD1 ED6

DATE OF CIRCULATION:

2021-05-21

CLOSING DATE FOR COMMENTS:

2021-07-16

SUPERSEDES DOCUMENTS:

62D/1822/RR

IEC SC 62D : ELECTROMEDICAL EQUIPMENT

SECRETARIAT:

United States of America

SECRETARY:

Ms Ladan Bulookbashi

OF INTEREST TO THE FOLLOWING COMMITTEES:

PROPOSED HORIZONTAL STANDARD:



Other TC/SCs are requested to indicate their interest, if any, in this CD to the secretary.

FUNCTIONS CONCERNED:

☐ EMC

☐ ENVIRONMENT

☐ QUALITY ASSURANCE

☒ SAFETY

This document is still under study and subject to change. It should not be used for reference purposes.

Recipients of this document are invited to submit, with their comments, notification of any relevant patent rights of which they are aware and to provide supporting documentation.

TITLE:

Amendment 1 - Medical electrical equipment - Part 2-2: Particular requirements for the basic safety and essential performance of high frequency surgical equipment and high frequency surgical accessories

NOTE FROM TC/SC OFFICERS:

The 6th Edition of IEC 60601-2-2 was published in 2017. This amendment is intended to align the standard to IEC 60601-1:2005/AMD2:2020. Additional issues reported from the National Committees and other interested parties since the publication in 2017 are addressed. National Committees are invited to submit comments in case they identified any such need.

Copyright © 2021 International Electrotechnical Commission, IEC. All rights reserved. It is permitted to download this electronic file, to make a copy and to print out the content for the sole purpose of preparing National Committee positions. You may not copy or "mirror" the file or printed version of the document, or any part of it, for any other purpose without permission in writing from IEC.

CONTENTS

FOREWORD.....	3
INTRODUCTION TO THE AMENDMENT	5
201.1 Scope, object and related standards	6
201.1.3 Collateral standards.....	6
201.2 Normative references.....	6
201.3 Terms and definitions.....	6
201.4 General requirements	7
201.4.11 Power input	7
201.5 Me equipment identification, marking and documents	7
202.2 Normative references.....	9
202.3 Terms and definitions.....	9
202.8 Electromagnetic IMMUNITY requirements for ME EQUIPMENT and ME SYSTEMS.....	10
202.8.1 General.....	10
202.201 Index of defined terms	10
208 General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems	10
Annex A Particular guidance and rationale.....	11
AA.2 Rationale for particular clauses and subclauses.....	11
Figure 201.107 – Measurement of HF LEAKAGE CURRENT from a BIPOLAR ELECTRODE	8
Table 201.101 – Colours of indicator lights and their meaning for HF SURGICAL EQUIPMENT	7