



DPC: 24 / 30475052 DC

BSI Group Headquarters

389 Chiswick High Road London W4 4AL

Tel: +44 (0)20 8996 9000

Fax: +44 (0)20 8996 7400

www.bsigroup.com

Date: 12 March 2024

Origin: International

Latest date for receipt of comments: 3 April 2024

Project No. 2023/01238

Responsible committee: CH/121/9 Lung Ventilators & Related Equipment

Interested committees:

Title: BS EN ISO 80601-2-67 Medical electrical equipment

Part 2-67: Particular requirements for basic safety and essential performance of oxygen-conserving equipment

Please notify the secretary if you are aware of any keywords that might assist in classifying or identifying the standard or if the content of this standard

- i) has any issues related to 3rd party IPR, patent or copyright
- ii) affects other national standard(s)
- iii) requires additional national guidance or information

**WARNING: THIS IS A DRAFT AND MUST NOT BE REGARDED OR USED AS A BRITISH STANDARD.
THIS DRAFT IS NOT CURRENT BEYOND 3 April 2024**

This draft is issued to allow comments from interested parties; all comments will be given consideration prior to publication. No acknowledgement will normally be sent. **See overleaf for information on the submission of comments.**

No copying is allowed, in any form, without prior written permission from BSI except as permitted under the Copyright, Designs and Patent Act 1988 or for circulation within a nominating organization for briefing purposes. Electronic circulation is limited to dissemination by e-mail within such an organization by committee members.

Further copies of this draft may be purchased from BSI Shop <http://shop.bsigroup.com> or from BSI Customer Services, Tel: +44(0) 20 8996 9001 or email cservices@bsigroup.com. British, International and foreign standards are also available from BSI Customer Services.

Information on the co-operating organizations represented on the committees referenced above may be obtained from <http://standardsdevelopment.bsigroup.com>

Responsible Committee Secretary: Ms Allie Ives (BSI)

Direct tel:

E-mail: allie.ives@bsigroup.com

Introduction

This draft standard is based on national and international discussions. Your comments on this draft are invited and will assist in the preparation of the consequent standard.

For international standards, comments will be reviewed by the relevant UK national committee before sending the consensus UK vote and comments to the international committee, which will then decide appropriate action. If the international standard is approved, it is usual for the text to be published as a British Standard.

For national standards, comments will be reviewed by the relevant UK national committee and the resulting standards published as a British Standard.

UK Vote

Please indicate whether you consider the UK should submit a negative (with supporting technical reasons) or positive vote on this draft. Please indicate if you are aware of any reason why this draft standard should not be published as a British Standard.

Submission of Comments

- **Annotated drafts are not acceptable and will be rejected.**
- All comments should be submitted online at <https://standardsdevelopment.bsigroup.com/>. You will need to register in order to comment.

Template for comments and secretariat observations

Date: xx/xx/20xx	Document: ISO/DIS xxxx
------------------	------------------------

1	2	(3)	4	5	(6)	7
MB	Clause No./ Subclause No./Annex (e.g. 3.1)	Paragraph/Figure/ Table/Note	Type of comment	Commend (justification for change) by the MB	Proposed change by the MB	Secretariat observations on each comment submitted
	3.1	Definition 1	ed	Definition is ambiguous and needs clarifying.	Amend to read '...so that the mains connector to which no connection...'	
	6.4	Paragraph 2	te	The use of the UV photometer as an alternative cannot be supported as serious problems have been encountered in its use in the UK.	Delete reference to UV photometer.	



62D/2123/CD

COMMITTEE DRAFT (CD)

PROJECT NUMBER:

ISO 80601-2-67 ED3

DATE OF CIRCULATION:

2024-03-08

CLOSING DATE FOR COMMENTS:

2024-05-03

SUPERSEDES DOCUMENTS:

62D/2055/RR

IEC SC 62D : PARTICULAR MEDICAL EQUIPMENT, SOFTWARE, AND SYSTEMS

SECRETARIAT:

United States of America

SECRETARY:

Ms Ladan Bulookbashi

OF INTEREST TO THE FOLLOWING COMMITTEES:

PROPOSED HORIZONTAL STANDARD:

☐

Other TC/SCs are requested to indicate their interest, if any, in this CD to the secretary.

FUNCTIONS CONCERNED:

☐ EMC

☐ ENVIRONMENT

☐ QUALITY ASSURANCE

☒ SAFETY

This document is still under study and subject to change. It should not be used for reference purposes.

Recipients of this document are invited to submit, with their comments, notification of any relevant patent rights of which they are aware and to provide supporting documentation.

Recipients of this document are invited to submit, with their comments, notification of any local regulations or technical reasons that may exist and should be considered should this proposal proceed, recognizing that failure to address such requirements could result in the need for new or existing "In Some Countries" clauses. (SEE [AC/22/2007](#) OR [NEW GUIDANCE DOC](#)).


TITLE:

Medical electrical equipment - Part 2-67: Particular requirements for basic safety and essential performance of oxygen-conserving equipment

NOTE FROM TC/SC OFFICERS:

This is a joint project with ISO/TC 121/SC 3.

Copyright © 2024 International Electrotechnical Commission, IEC. All rights reserved. It is permitted to download this electronic file, to make a copy and to print out the content for the sole purpose of preparing National Committee positions. You may not copy or "mirror" the file or printed version of the document, or any part of it, for any other purpose without permission in writing from IEC.



Documents

Document Sharing Hub

Medical electrical equipment — Part 2-67: Particular requirements for basic safety and essential performance of oxygen conserving equipment

Appareils électromédicaux — Partie 2-67: Exigences particulières pour la sécurité de base et les performances essentielles des économiseurs d'oxygène

CD stage

Warning for WDs and CDs

This document is not an ISO International Standard. It is distributed for review and comment. It is subject to change without notice and may not be referred to as an International Standard.

Recipients of this draft are invited to submit, with their comments, notification of any relevant patent rights of which they are aware and to provide supporting documentation.

© ISO 2024, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Ch. de Blandonnet 8 • CP 401
CH-1214 Vernier, Geneva, Switzerland
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
copyright@iso.org
www.iso.org

	Page
Contents	
Foreword	2
Introduction	4
201.1 Scope, object and related standards	1
201.2 Normative references	3
201.3 Terms and definitions	4
201.4 General requirements	13
201.5 General requirements for testing of <i>ME equipment</i>	16
201.6 Classification of <i>ME equipment</i> and <i>ME systems</i>	17
201.7 <i>ME equipment</i> identification, marking and documents	17
201.8 Protection against electrical hazards from <i>ME equipment</i>	24
201.9 Protection against mechanical hazards of <i>ME equipment</i> and <i>ME systems</i>	24
201.10 Protection against unwanted and excessive radiation hazards	24
201.11 Protection against excessive temperatures and other hazards	24
201.12 Accuracy of controls and instruments and protection against hazardous outputs	26
201.13 Hazardous situations and fault conditions	29
201.14 Programmable electrical medical systems (PEMS)	29
201.15 Construction of <i>ME equipment</i>	30
201.16 <i>ME systems</i>	30
201.17 Electromagnetic compatibility of <i>ME equipment</i> and <i>ME systems</i>	31
201.101 Gas connections	31
201.102 Requirements for parts and accessories	32
201.103 Oxygen pressure regulators	34
202 Electromagnetic disturbances – Requirements and tests	34
202.4.3.1 Configurations	34
206 Usability	35
Annex C (informative) Guide to marking and labelling requirements for <i>ME equipment</i> and <i>ME systems</i>	36
Annex D (informative) Symbols on marking	41
Annex AA (informative) Particular guidance and rationale	42
Annex BB (informative) Reference to the IMDRF essential principles and labelling guidances	51
Annex CC (informative) Terminology — Alphabetized index of defined terms	55
Bibliography	58

Foreword

ISO (the International Organization for Standardization) and IEC (the International Electrotechnical Commission) form the specialized system for worldwide standardization. National bodies that are members of ISO or IEC participate in the development of International Standards through technical committees established by the respective organization to deal with particular fields of technical activity. ISO and IEC technical committees collaborate in fields of mutual interest. Other international organizations, governmental and non-governmental, in liaison with ISO and IEC, also take part in the work.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives or www.iec.ch/members_experts/refdocs).

ISO and IEC draw attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO and IEC take no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO and IEC had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at www.iso.org/patents and <https://patents.iec.ch>. ISO and IEC shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see www.iso.org/iso/foreword.html. In the IEC, see www.iec.ch/understanding-standards.

This document was prepared jointly by Technical Committee ISO/TC 121, *Anaesthetic and respiratory equipment*, Subcommittee SC 3, *Respiratory devices and related equipment used for patient care*, and Technical Committee IEC/TC 62, *Medical equipment, software, and systems*, Subcommittee SC D, *Particular medical equipment, software, and systems*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 215, *Respiratory and anaesthetic equipment*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This **third** edition cancels and replaces the **second** edition (ISO 80601-2-67:2020), which has been technically revised.

The main changes are as follows:

- updated references, where appropriate;
- harmonization with ISO 20417, where appropriate;
- updated uncertainty of measurement requirements;
- added *marking* requirements for *gas intake port*, external gas sources and MR compatibility;

78 — requirements for *processing* of the *enclosure*;

79 — added *cybersecurity* recommendations; and

80 — updated *connector* requirements.

81 A list of all parts in the ISO 80601 and IEC 80601 series can be found on the ISO and IEC websites.

82 Any feedback or questions on this document should be directed to the user's national standards body. A
 83 complete listing of these bodies can be found at www.iso.org/members.html and [www.iec.ch/national-](http://www.iec.ch/national-committees)
 84 [committees](http://www.iec.ch/national-committees).

Introduction

Long-term oxygen therapy has been demonstrated in randomized, controlled clinical trials to prolong survival in *patients* with chronic respiratory disease and documented hypoxemia. Typical sources of therapeutic long-term oxygen therapy include gaseous oxygen from cylinders or from liquid oxygen and oxygen from an oxygen concentrator.

Most clinicians prescribe low flow oxygen therapy as continuous flow oxygen (CFO) delivery in l/min. CFO systems deliver the flow of oxygen without regard for the *patient's* breathing rate or pattern. Outside of the institutional care setting, the provision of CFO therapy is often a significant expense and can limit the mobility of a *patient* to the immediate vicinity of a stationary or fixed oxygen delivery system. To support mobility, *patients* use CFO from portable liquid or compressed oxygen systems with a limited storage capacity that can limit a *patient's* time and activities while away from a stationary oxygen supply.

Conserving equipment that delivers supplemental oxygen as a bolus conserves usage while allowing satisfactory *patient* arterial oxygen saturation (SaO_2) to be maintained during daily activities. *Conserving equipment* delivers supplemental oxygen unlike CFO in that the therapy gas flow is delivered only during the inspiratory phase of the breathing cycle, when it is most likely to reach the alveoli. During both the expiratory and pause phase of the breathing cycle, the flow of supplemental oxygen is stopped, minimizing waste. Because flow over time produces a volume, the bolus delivered by the *conserving equipment* is typically represented as a volume of gas. Therapy using *conserving equipment* versus CFO results in lower operating costs and longer ambulatory times for *patients* using the same CFO storage capacity.

Operation of *conserving equipment* from various *manufacturers* might differ in the dose delivery mechanism resulting in variations in oxygen therapy to the *patient*. The use of CFO numerical *markings* for dose settings on *conserving equipment* might not directly correlate with CFO settings and might lead to misinterpretation of gas delivery rates and volumes for a particular *patient*. This might result in incorrect *patient* setup and therapy delivery over all breathing rates and patterns versus CFO. Because of the differences in delivery, settings, and *markings* versus CFO therapy, *conserving equipment* use has requirements for *patient* titration to determine the proper setting(s) needed to provide adequate SaO_2 levels for the *patient* breathing patterns.

In this document, the following print types are used:

- requirements and definitions: roman type;
- *terms defined in Clause 3 of the general standard, in this particular document or as noted: italic type;*
and
- informative material appearing outside of tables, such as notes, examples and references: in smaller type.
Normative text of tables is also in a smaller type.

In referring to the structure of this document, the term.

- “clause” means one of the three numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 201 includes subclauses 201.7, 201.8, etc.); and
- “subclause” means a numbered subdivision of a clause (e.g. 201.7, 201.8 and 201.9 are all subclauses of Clause 201).

References to clauses within this document are preceded by the term “Clause” followed by the clause number. References to subclauses within this particular document are by number only.

126 In this document, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination
127 of the conditions is true.

128 For the purposes of this document, the auxiliary verb:

129 — “shall” **indicates** a requirement;

130 — “should” **indicates** a requirement or a test is **recommendation**;

131 — “may” **indicates** a permission;

132 — “can” is used to describe a possibility or capability; and

133 — “must” is used to express an external constraint.

134 Annex C contains a guide to the *marking* and labelling requirements in this document.

135 Annex D contains a summary of the *symbols* referenced in this document.

136 **Requirements in this document have been decomposed so that each requirement is uniquely delineated.**
137 **This is done to support automated requirements tracking.**

138

Medical electrical equipment —

Part 2-67:

Particular requirements for basic safety and essential performance of oxygen conserving equipment

201.1 Scope, object and related standards

NOTE There is guidance or rationale for this clause contained in Clause AA.2.

IEC 60601-1:2005+AMD1:2012+AMD2:2020, Clause 1 applies, except as follows:

NOTE The general standard is IEC 60601-1:2005+AMD1:2012+AMD2:2020.

201.1.1 Scope

IEC 60601-1:2005+AMD1:2012+AMD2:2020, 1.1 is replaced by:

This document is applicable to the *basic safety* and *essential performance* of oxygen conserving equipment, hereafter referred to as *ME equipment*, in combination with its *accessories* intended to conserve supplemental oxygen by delivering gas intermittently and synchronized with the *patient's* inspiratory cycle, when used in the *home healthcare environment*. Oxygen conserving equipment is typically used by a *lay operator*.

NOTE 1 *Conserving equipment* can also be used in professional health care facilities.

This document is also applicable to *conserving equipment* that is incorporated with other equipment.

EXAMPLE *Conserving equipment* combined with a pressure regulator^[4], an oxygen concentrator^[11] or liquid oxygen equipment^[7].

This document is also applicable to those *accessories* intended by their *manufacturer* to be connected to *conserving equipment*, where the characteristics of those *accessories* can affect the *basic safety* or *essential performance* of the *conserving equipment*.

This document is intended to clarify the difference in operation of various *conserving equipment* models, as well as between the operation of *conserving equipment* and continuous flow oxygen equipment, by requiring standardized performance testing and labelling.

This document is only applicable to active devices (e.g. pneumatically or electrically powered) and is not applicable to non-active devices (e.g. reservoir cannulas).

If a clause or subclause is specifically intended to be applicable to *ME equipment* only, or to *ME systems* only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to *ME equipment* and to *ME systems*, as relevant.

Hazards inherent in the intended physiological function of *ME equipment* or *ME systems* within the scope of this document are not covered by specific requirements in this document except in IEC 60601-1:2005+AMD1:2012+AMD2:2020, 7.2.13 and 8.4.1.

NOTE 2 Additional information can be found in IEC 60601-1:2005+AMD1:2012+AMD2:2020, 4.2.

201.1.2 Object

IEC 60601-1:2005+AMD1:2012+AMD2:2020, 1.2 is replaced by:

The object of this document is to establish particular *basic safety* and *essential performance* requirements for *conserving equipment* [as defined in 201.3.207] and its *accessories*.

NOTE 1 *Accessories* are included because *accessories* can have a significant impact on the *basic safety* or *essential performance* of *conserving equipment*.

NOTE 2 This document has been prepared to address the relevant *essential principles*^[16] and labelling^[17] guidances of the International Medical Devices Regulators Forum (IMDRF) as indicated in Annex BB.

NOTE 3 This document has been prepared to address the relevant general safety and performance requirements of European regulation (EU) 2017/745^[15].

201.1.3 Collateral standards

IEC 60601-1:2005+AMD1:2012+AMD2:2020, 1.3 applies with the following addition:

IEC 60601-1-2+AMD1:2020 and IEC 60601-1-6+AMD1:2013+AMD2:2020 apply as modified in Clauses 202 and 206 respectively. IEC 60601-1-3 and IEC 60601-1-9 do not apply. All other published collateral standards in the IEC 60601-1 series apply as published.

201.1.4 Particular standards

Replacement:

In the IEC 60601 series, particular standards define *basic safety* and *essential performance* requirements, and may modify, replace or delete requirements contained in the general standard and collateral standards as appropriate for the particular *ME equipment* under consideration.

A requirement of a particular standard takes priority over the general standard.

For brevity, IEC 60601-1+AMD1:2012+AMD2:2020 is referred to in this document as the general standard. Collateral standards are referred to by their document number.

The numbering of clauses and subclauses of this document corresponds to that of the general standard with the prefix "201" (e.g. 201.1 in this document addresses the content of Clause 1 of the general standard) or applicable collateral standard with the prefix "20x", where x is the final digit(s) of the collateral standard document number (e.g. 202.4 in this document addresses the content of Clause 4 of the IEC 60601-1-2 collateral standard, 206.4 in this document addresses the content of Clause 4 of the IEC 60601-1-6 collateral standard, etc.). The changes to the text of the general standard are specified by the use of the following words:

"Replacement" means that the clause or subclause of the general standard or applicable collateral standard is replaced completely by the text of this document.

"Addition" means that the text of this document is additional to the requirements of the general standard or applicable collateral standard.

"Amendment" means that the clause or subclause of the general standard or applicable collateral standard is amended as indicated by the text of this document.

Subclauses, figures or tables which are additional to those of the general standard are numbered starting from 201.101. However, due to the fact that definitions in the general standard are numbered 3.1 through 3.154, additional definitions in this document are numbered beginning from 201.3.201. Additional annexes are lettered AA, BB, etc., and additional items aa), bb), etc.

Subclauses, figures or tables which are additional to those of a collateral standard are numbered starting from 20x, where "x" is the number of the collateral standard, e.g. 202 for IEC 60601-1-2, 206 for IEC 60601-1-6, etc.

The term "this document" is used to make reference to the general standard, any applicable collateral standards and this particular document taken together.

Where there is no corresponding clause or subclause in this document, the clause or subclause of the general standard or applicable collateral standard, although possibly not relevant, applies without modification; where it is intended that any part of the general standard or applicable collateral standard, although possibly relevant, is not to be applied, a statement to that effect is given in this document.

201.2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60601-1:2005+AMD1:2012+AMD2:2020, Clause 2 applies, except as follows:

Addition:

ISO 32:1977, *Gas cylinders for medical use — Marking for identification of content*

ISO 5359:2014+AMD1:2017, *Low-pressure hose assemblies for use with medical gases*

ISO 7396-1:2016+AMD1:2017, *Medical gas pipeline systems — Part 1: Pipeline systems for compressed medical gases and vacuum*

ISO 10524-1:2018, *Pressure regulators for use with medical gases — Part 1: Pressure regulators and pressure regulators with flow-metering devices*

ISO 10524-3:2019, *Pressure regulators for use with medical gases — Part 3: Pressure regulators integrated with cylinder valves (VIPRs)*

ISO 14937:2009, *Sterilization of health care products — General requirements for characterization of a sterilizing agent and the development, validation and routine control of a sterilization process for medical devices*

ISO 17664-1:2021, *Processing of health care products — Information to be provided by the medical device manufacturer for the processing of medical devices — Part 1: Critical and semi-critical medical devices*

ISO 17664-2:2021, *Processing of health care products — Information to be provided by the medical device manufacturer for the processing of medical devices — Part 2: Non-critical medical devices*

ISO 18562-1:2024, *Biocompatibility evaluation of breathing gas pathways in healthcare applications — Part 1: Evaluation and testing within a risk management process*

ISO 20417:2021, *Medical devices — Information to be supplied by the manufacturer*

ISO 80369-1:2021, *Small-bore connectors for liquids and gases in healthcare applications — Part 1: General requirements*

IEC 60601-1:2005+AMD1:2012+AMD2:2020, *Medical electrical equipment — Part 1: General requirements for basic safety and essential performance*

IEC Guide 115:2023, *Application of uncertainty of measurement to conformity assessment activities in the electrotechnical sector*

EN 13544-2:2002+AMD1:2009, *Respiratory therapy equipment — Part 2: Tubing and connectors*

201.3 Terms and definitions

For the purposes of this document, the terms and definitions given in IEC 60601-1:2005+AMD1:2012+AMD2:2020 and the following apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

— ISO Online browsing platform: available at <https://www.iso.org/obp>

— IEC Electropedia: available at <http://www.electropedia.org/>

NOTE An alphabetized index of defined terms is found in Annex CC.

Addition:

201.3.201 accompanying information

information accompanying or *marked* on a medical device or *accessory* for the user or those accountable for the installation, use, *processing*, maintenance, decommissioning and disposal of the medical device or *accessory*, particularly regarding safe use

Note 1 to entry: The *accompanying information* shall be regarded as part of the medical device or *accessory*.

Note 2 to entry: The *accompanying information* can consist of the label, *marking*, *instructions for use*, *technical description*, installation manual, quick reference guide, etc.

Note 3 to entry: *Accompanying information* is not necessarily a written or printed document but could involve auditory, visual, or tactile materials and multiple media types (e.g., CD/DVD-ROM, USB stick, website).

[SOURCE: ISO 20417:2021, 3.2, modified — deleted note 4.]

¹ Under preparation. Stage at the time of publication: ISO/FDIS 80369-1:2024.

273 **201.3.202**
 274 **alarm limit**
 275 threshold used by an *alarm system* to determine an *alarm condition*

276 [SOURCE: IEC 60601-1-8:2006+AMD1:2012+AMD2:2020, 3.3]

277 **201.3.203**
 278 **attack**
 279 attempt to destroy, expose, alter, disable, steal or gain unauthorized access to or make unauthorized use
 280 of an asset

281 [SOURCE: IEC 81001-5-1:2021, 3.5]

282 **201.3.204**
 283 **biocompatibility**
 284 ability of a medical device, *accessory* or material to perform with an appropriate host response in a
 285 specific application

286 Note 1 to entry: A medical device or *accessory* may produce some level of adverse effect, but that level may be
 287 determined to be acceptable when considering the benefits provided by the medical device or *accessory*.

288 [SOURCE: ISO 18562-1:2024, 3.6]

289 **201.3.205**
 290 **cleaning**
 291 removal of contaminants to the extent necessary for further *processing* or for *intended use*

292 Note 1 to entry: *Cleaning* consists of the removal of adherent soil (e.g. blood, protein substances, and other debris)
 293 from the surfaces, crevices, serrations, joints, and lumens of a medical device by a manual or automated *process* that
 294 prepares the items for safe handling or further *processing*.

295 [SOURCE: ISO 17664-2:2021, 3.1, modified — replaced 'and/or' with 'or'.]

296 **201.3.206**
 297 **connector**
 298 fitting to join two or more components

299 EXAMPLE *Connectors for low-pressure hose assembly* are any of a range of mating components intended to
 300 maintain gas specificity by the allocation of a set of different diameters to the mating *connectors* for each particular
 301 gas.

302 [SOURCE: ISO 4135:2022, 3.1.4.5]

303 **201.3.207**
 304 **conserving equipment**
 305 *ME equipment* intended to conserve supplemental oxygen by delivering gas intermittently and
 306 synchronized with the *patient's* inspiratory cycle

307 Note 1 to entry: *Conserving equipment* can be electrically or pneumatically powered.

201.3.208

conserving equipment with monitoring function

conserving equipment suitable for use with *patients* where monitoring of oxygen delivery via the *conserving equipment* is indicated

201.3.209

cybersecurity

state where information and systems are protected from unauthorized activities, such as access, use, disclosure, disruption, modification, or destruction to a degree that the related *risks* to violation of confidentiality, integrity, and availability are maintained at an acceptable level throughout the life cycle

[SOURCE: IEC 81001-5-1:2021, 3.30]

201.3.210

disinfection

process to inactivate viable microorganisms to a level previously specified as being appropriate for a defined purpose

[SOURCE: ISO 17664-2:2021, 3.5]

201.3.211

essential function

function or capability that is required to maintain *basic safety*, *essential performance*, a minimum of clinical functionality as specified by the *manufacturer*, and operational availability for the medical device

Note 1 to entry: *Essential functions* include, but are not limited to, the *safety* instrumented function (*basic safety* and *essential performance*), the control function and the availability of urgently needed functions and such allowing the *operator* to view and manipulate the *medical device* safely with the most urgently needed performance (operational availability). The loss of *essential function* is commonly termed loss of protection, loss of control and loss of view respectively.

Note 2 to entry: The term is derived from IEC 62443-4-2:2019, 3.1.20, and has been refined for the purpose and scope of this document.

[SOURCE: IEC/TR 60601-4-5:2021, 3.10]

201.3.212

essential principles

essential principles of safety and performance

fundamental high-level requirements that when complied with ensure a medical device is safe and performs as intended

[SOURCE: ISO 16142-1:2016, 3.3]

201.3.213

exhaust port

port of the medical equipment or device from which gas is discharged to the atmosphere during *normal use*, either directly or via an anaesthetic gas scavenging system

345 [SOURCE: ISO 19223:2019, 3.14.2]

346 **201.3.214**

347 **firecall**

348 method established to provide emergency access to a secure medical device

349 Note 1 to entry: In an emergency situation, unprivileged users can gain access to key systems to correct the
350 problem. When a *firecall* is used, there is usually a review *process* to ensure that the access was used properly to
351 correct a problem. These methods generally either provide a one-time use user identifier (ID) or one-time password
352 or other suitable measures.

353 Note 2 to entry: Also referred to as "break glass" feature.

354 [SOURCE: IEC/TR 60601-4-5:2021, 3.11]

355 **201.3.215**

356 **flow-direction-sensitive component**

357 component or *accessory* through which gas flow is in one direction only for proper functioning or *patient*
358 safety

359 [SOURCE: ISO 4135:2022, 3.1.4.15]

360 **201.3.216**

361 **gas intake port**

362 port through which gas is drawn for use by the *patient*

363 Note 1 to entry: Gas is drawn at a sub-ambient pressure at a *gas intake port*, in opposition to an *inlet*, at which gas
364 is provided by a medical gas supply system.

365 [SOURCE: ISO 4135:2022, 3.1.4.21, modified — replaced "apposition" with "opposition".]

366 **201.3.217**

367 **gas pathway**

368 interior surfaces, over which gases or liquids that can be inspired pass

369 Note 1 to entry: The *gas pathway* is bounded by the ports through which gases or liquids enter the medical device.
370 This can include the *patient* interface or the interior surfaces of *enclosures* that are in contact with gases or liquids
371 that can be inspired.

372 Note 2 to entry: The *gas pathway* can include some surfaces in the expiratory pathway.

373 Note 3 to entry: *Patient* contact surfaces such as the outer surfaces of a tracheal tube or the cushion of a *mask* are
374 evaluated according to the ISO 10993 series.

375 **EXAMPLE 1** The ventilator breathing system, *inlet* filter, gas mixer, blower and internal piping.

376 **EXAMPLE 2** Enclosed chamber of an incubator including the mattress or the inner surface of an oxygen hood.

377 **EXAMPLE 3** The inner surfaces of breathing tubes, tracheal tubes or *masks* and mouthpieces.

378 [SOURCE: ISO 18562-1:2024, 3.11]

201.3.218

high-pressure inlet

inlet to which gas is supplied at a pressure exceeding 100 kPa above ambient

Note 1 to entry: The phrases 'low-pressure' and 'high-pressure' are used differently in various contexts, including breathing system pressures (typically less than 10 kPa), terminal *outlet* pressures (less than 600 kPa), manifold pressures (typically up to 3 000 kPa) and cylinder pressures (typically less than 30 000 kPa).

[SOURCE: ISO 4135:2022, 3.1.4.24]

201.3.219

home healthcare environment

dwelling place in which a *patient* lives or other places where *patients* are present, excluding professional healthcare facility environments where *operators* with medical training are continually available when *patients* are present

EXAMPLE In a car, bus, train, boat or plane, in a wheelchair or walking outdoors.

Note 1 to entry: Professional healthcare facilities include hospitals, physician offices, freestanding surgical centres, dental offices, freestanding birthing centres, limited care facilities, first aid rooms or rescue rooms, multiple treatment facilities and emergency medical services.

Note 2 to entry: Nursing homes are considered *home healthcare environments*.

Note 3 to entry: Other places where a *patient* is present include the outdoor environment, while working and in vehicles.

[SOURCE: IEC 60601-1-11:2015+AMD1:2020, 3.1, modified — deleted "For the purpose of this collateral standard,".]

201.3.220

I:E ratio

ratio of the *inspiratory time* to the expiratory time in a respiratory cycle

[SOURCE: ISO 19223:2019, 3.4.19, modified — deleted notes.]

201.3.221

immunity

the ability of *ME equipment* or an *ME system* to perform without degradation in the presence of an electromagnetic disturbance

[SOURCE: IEC 60601-1-2:2014+AMD1:2020, 3.8]

201.3.222

information supplied by the manufacturer

information related to the identification and use of a medical device or *accessory*, in whatever form provided, intended to ensure the safe and effective use of the medical device or *accessory*

Note 1 to entry: For the purposes of this document, *e-documentation* is included in *information supplied by the manufacturer*.

Note 2 to entry: For the purposes of this document, shipping documents and promotional material are excluded from *information supplied by the manufacturer*. However, some *authorities having jurisdiction* can consider such supplemental information as *information supplied by the manufacturer*.

Note 3 to entry: The primary purpose of *information supplied by the manufacturer* is to identify the medical device and its *manufacturer*, and provide essential information about its safety, performance, and appropriate use to the user or other relevant persons.

[SOURCE: ISO 20417:2021, 3.10, modified — deleted note 4.]

201.3.223

inlet

opening through which gas or other material is pushed by an elevated upstream pressure

[SOURCE: ISO 4135:2022, 3.1.4.26, modified — deleted note.]

201.3.224

inspiratory time

t_i

duration of an inflation phase or inspiratory phase

[SOURCE: ISO 19223:2019, 3.4.8, modified — deleted notes.]

201.3.225

instructions for use

IFU

portion of the *accompanying information* that is essential for the safe and effective use of a medical device or *accessory* directed to the user of the medical device

Note 1 to entry: For the purposes of this document, a user can be either a *lay* user or professional user with relevant specialized training.

Note 2 to entry: For the purposes of this document, instructions for the professional *processing* between uses of a medical device or *accessory* can be included in the *instructions for use*.

Note 3 to entry: For the purposes of this document, information indicated on a graphical user interface (GUI) is considered as appearing on the item.

Note 4 to entry: The *instructions for use*, or portions thereof, can be located on the display of a medical device or *accessory*.

Note 5 to entry: Medical devices or *accessories* that can be used safely and effectively without *instructions for use* are exempted from having *instructions for use* by some *authorities having jurisdiction*.

[SOURCE: ISO 20417:2021, 3.11, modified — deleted note 6.]

201.3.226

lay

lay person

<adj> term referring to non-professional or professional without relevant specialized training

EXAMPLE *Lay operator, lay responsible organization.*

[SOURCE: IEC 60601-1-11:2015+AMD1:2020, 3.2]

201.3.227

low-pressure hose assembly

assembly consisting of a flexible hose with permanently attached gas-specific *inlet* and *outlet connectors* and designed to conduct a *medical gas* at pressures less than 1 400 kPa

Note 1 to entry: The phrases 'low-pressure' and 'high-pressure' are used differently in various contexts, including breathing system pressures (typically less than 10 kPa), terminal *outlet* pressures (less than 1 400 kPa), manifold pressures (typically up to 3 000 kPa) and cylinder pressures (typically less than 30 000 kPa).

[SOURCE: ISO 4135:2022, 3.2.3.1]

201.3.228

lung

each of the pair of compliant organs within the ribcage (thorax), bounded by the terminal bronchiole and the visceral pleura, which during *ventilation* provide gas/blood interfaces that enable oxygen from the gas to pass into the blood and carbon dioxide to be removed

[SOURCE: ISO 19223:2019, 3.1.16, modified — deleted notes.]

201.3.229

marking

information, in text or graphical format, durably affixed, printed, etched (or equivalent) to a medical device or *accessory*

Note 1 to entry: For the purposes of this document, the term *marked* is used to designate the corresponding act.

Note 2 to entry: For the purposes of this document, *marking* is different from 'direct *marking*' as commonly described in unique device identification (UDI) standards and regulations. A UDI 'direct *marking*' is a type of *marking*.

[SOURCE: ISO 20417:2021, 3.16, modified — deleted note 3.]

201.3.230

mask

device which provides a non-invasive interface between the *patient's airway* and a *patient-connection* port or other connection to a source of respirable gas

[SOURCE: ISO 4135:2022, 3.8.6.4]

201.3.231

maximum limited pressure

$P_{\text{Lim,max}}$

highest airway pressure that can occur during *normal use* or under *single fault condition*

[SOURCE: ISO 19223:2019, 3.13.3, modified — deleted notes.]

- 486 **201.3.232**
 487 **medical gas pipeline system**
 488 combination of a supply system, a monitoring and *alarm system* and a pipeline distribution system with
 489 terminal units for provision of medical gases or vacuum
 490 [SOURCE: ISO 4135:2022, 3.2.1.1]
- 491 **201.3.233**
 492 **outlet**
 493 opening through which gas leaves a device or component
 494 [SOURCE: ISO 4135:2022, 3.1.4.40]
- 495 **201.3.234**
 496 **processing**
 497 <preparation of medical device, *accessory*> activity to prepare a new or used medical device and *accessory*
 498 for its *intended use*
 499 [SOURCE: ISO 20417:2021, 3.20]
- 500 **201.3.235**
 501 **security level**
 502 level corresponding to the required set of countermeasures and inherent *cybersecurity* properties of
 503 devices and systems for a zone or conduit based on assessment of *risk* for the zone or conduit
 504 [SOURCE: IEC/TR 60601-4-5:2021, 3.23]
- 505 **201.3.236**
 506 **set rate**
 507 number of assured inflations that are set to occur in a specified period of time, expressed as breaths per
 508 minute
 509 [SOURCE: ISO 19223:2019, 3.5.1.1, modified — deleted notes and examples.]
- 510 **201.3.237**
 511 **single use**
 512 <medical device, *accessory*> intended by the *manufacturer* to be used on an individual *patient* or
 513 specimen during a single *procedure* and then disposed of
 514 Note 1 to entry: A *single use* medical device or *accessory* is not intended by its *manufacturer* to be further *processed*
 515 and used again.
 516 [SOURCE: ISO 20417:2021, 3.26]
- 517 **201.3.238**
 518 **standard temperature and pressure dry**
 519 **STPD**
 520 pressure of 101,325 kPa at a temperature of 20 °C, dry