



DPC: 24 / 30485247 DC

BSI Group Headquarters

389 Chiswick High Road London W4 4AL

Tel: +44 (0)20 8996 9000

Fax: +44 (0)20 8996 7400

www.bsigroup.com

Date: 05 January 2024

Origin: International

Latest date for receipt of comments: 17 January 2024

Project No. 2023/01542

Responsible committee: EXL/31 Equipment for explosive atmospheres

Interested committees:

Title: BS EN IEC 60079-29-0 Explosive atmospheres

Part 29-0: Gas detectors - General requirements and test methods, and possible supplementary parts

Please notify the secretary if you are aware of any keywords that might assist in classifying or identifying the standard or if the content of this standard

- i) has any issues related to 3rd party IPR, patent or copyright
- ii) affects other national standard(s)
- iii) requires additional national guidance or information

**WARNING: THIS IS A DRAFT AND MUST NOT BE REGARDED OR USED AS A BRITISH STANDARD.
THIS DRAFT IS NOT CURRENT BEYOND 17 January 2024**

This draft is issued to allow comments from interested parties; all comments will be given consideration prior to publication. No acknowledgement will normally be sent. **See overleaf for information on the submission of comments.**

No copying is allowed, in any form, without prior written permission from BSI except as permitted under the Copyright, Designs and Patent Act 1988 or for circulation within a nominating organization for briefing purposes. Electronic circulation is limited to dissemination by e-mail within such an organization by committee members.

Further copies of this draft may be purchased from BSI Shop <http://shop.bsigroup.com> or from BSI Customer Services, Tel: +44(0) 20 8996 9001 or email cservices@bsigroup.com. British, International and foreign standards are also available from BSI Customer Services.

Information on the co-operating organizations represented on the committees referenced above may be obtained from <http://standardsdevelopment.bsigroup.com>

Responsible Committee Secretary: **Mr Christopher Brown (BSI)**Direct tel: **02089967441**E-mail: **christopher.brown@bsigroup.com**

Introduction

This draft standard is based on national and international discussions. Your comments on this draft are invited and will assist in the preparation of the consequent standard.

For international standards, comments will be reviewed by the relevant UK national committee before sending the consensus UK vote and comments to the international committee, which will then decide appropriate action. If the international standard is approved, it is usual for the text to be published as a British Standard.

For national standards, comments will be reviewed by the relevant UK national committee and the resulting standards published as a British Standard.

UK Vote

Please indicate whether you consider the UK should submit a negative (with supporting technical reasons) or positive vote on this draft. Please indicate if you are aware of any reason why this draft standard should not be published as a British Standard.

Submission of Comments

- **Annotated drafts are not acceptable and will be rejected.**
- All comments should be submitted online at <https://standardsdevelopment.bsigroup.com/>. You will need to register in order to comment.

Template for comments and secretariat observations

Date: xx/xx/20xx	Document: ISO/DIS xxxx
------------------	------------------------

1	2	(3)	4	5	(6)	7
MB	Clause No./ Subclause No./Annex (e.g. 3.1)	Paragraph/Figure/ Table/Note	Type of comment	Commend (justification for change) by the MB	Proposed change by the MB	Secretariat observations on each comment submitted
	3.1	Definition 1	ed	Definition is ambiguous and needs clarifying.	Amend to read '...so that the mains connector to which no connection...'	
	6.4	Paragraph 2	te	The use of the UV photometer as an alternative cannot be supported as serious problems have been encountered in its use in the UK.	Delete reference to UV photometer.	



31/1757/CD

COMMITTEE DRAFT (CD)

PROJECT NUMBER: IEC 60079-29-0 ED1	
DATE OF CIRCULATION: 2023-12-22	CLOSING DATE FOR COMMENTS: 2024-02-16
SUPERSEDES DOCUMENTS: 31/1711/CD, 31/1746A/CC	

IEC TC 31 : EQUIPMENT FOR EXPLOSIVE ATMOSPHERES	
SECRETARIAT: United Kingdom	SECRETARY: Mr Tom Stack
OF INTEREST TO THE FOLLOWING COMMITTEES:	PROPOSED HORIZONTAL STANDARD: ☐ Other TC/SCs are requested to indicate their interest, if any, in this CD to the secretary.
FUNCTIONS CONCERNED: ☐ EMC ☐ ENVIRONMENT ☐ QUALITY ASSURANCE ☒ SAFETY	

This document is still under study and subject to change. It should not be used for reference purposes.

Recipients of this document are invited to submit, with their comments, notification of any relevant patent rights of which they are aware and to provide supporting documentation.

Recipients of this document are invited to submit, with their comments, notification of any local regulations or technical reasons that may exist and should be considered should this proposal proceed, recognizing that failure to address such requirements could result in the need for new or existing "In Some Countries" clauses. (SEE [AC/22/2007](#) OR [NEW GUIDANCE DOC](#)).

TITLE: Explosive atmospheres - Part 29-0: Gas detectors – General requirements and test methods, and possible supplementary parts.
--

NOTE FROM TC/SC OFFICERS: As per 5.4 of the IEC/TC 31 GWP this CD is being circulated as a complete text including editing marks to the existing text. This is for ease of understanding of what has been changed, added or removed thus saving time as in the past experts did not realize critical text had been deleted.
--

Copyright © 2023 International Electrotechnical Commission, IEC. All rights reserved. It is permitted to download this electronic file, to make a copy and to print out the content for the sole purpose of preparing National Committee positions. You may not copy or "mirror" the file or printed version of the document, or any part of it, for any other purpose without permission in writing from IEC.

CONTENTS

1	CONTENTS	2
2	FOREWORD	5
3	INTRODUCTION	7
4	1 Scope	8
5	2 Normative references	10
6	3 Terms and definitions	11
7	4 General requirements	18
8	4.1 Introduction	18
9	4.2 Design	19
10	4.2.1 General	19
11	4.2.2 Indicators and means of indication of measured values	20
12	4.2.3 Alarm signal(s)	22
13	4.2.4 Fault signals	23
14	4.2.5 Adjustments	24
15	4.2.6 Battery-powered equipment	25
16	4.2.7 Gas detection transmitter for use with separate gas detection control units	25
17	4.2.8 Separate gas detection control units for use with gas detection transmitter(s)	25
18	4.2.9 Open path gas detection equipment (Type OF & Type OT)	25
19	4.2.10 Software-controlled equipment	25
20	4.3 Labelling and marking	27
21	4.4 Instruction manual Instructions	27
22	5 Test methods	31
23	5.1 General	31
24	5.2 General requirements for tests	31
25	5.2.1 Samples and sequence of tests	31
26	5.2.2 Preparation of equipment before testing	33
27	5.2.3 Mask for calibration and test (excludes Type OF & Type OT)	34
28	5.2.4 Open path gas detection equipment (Type OF & Type OT)	35
29	5.3 Normal conditions for test	37
30	5.3.1 General	37
31	5.3.2 Test gas(es)	37
32	5.3.3 Flow rate for test gases	39
33	5.3.4 Power supply	39
34	5.3.5 Temperature	40
35	5.3.6 Pressure	40
36	5.3.7 Humidity	40
37	5.3.8 Acclimation time	40
38	5.3.9 Orientation	40
39	5.3.10 Communications options	40
40	5.4 Tests	40
41	5.4.1 General	40
42	5.4.2 Unpowered storage	41
43	5.4.3 Vibration	42
44	5.4.4 Drop test	43
45	5.4.5 Field calibration kit	43



49	5.4.6	Linearity	44
50	5.4.7	Alarm set-point(s)	44
51	5.4.8	Time to alarm or alarm reading (equipment with electrochemical	
52		sensors only)	45
53	5.4.9	Stability	46
54	5.4.10	Gas concentrations above the upper limit of indication	47
55	5.4.11	Poisons and other gases	48
56	5.4.12	Temperature	49
57	5.4.13	Pressure	50
58	5.4.14	Humidity of test gas	50
59	5.4.15	Air velocity	51
60	5.4.16	Flow rate (aspirated equipment only)	51
61	5.4.17	Warm-up time	52
62	5.4.18	Time of response	52
63	5.4.19	Time of recovery	53
64	5.4.20	Sampling probe	53
65	5.4.21	Operation at or below the lower limit of measurement (type IN oxygen	
66		detection equipment only)	53
67	5.4.22	Extended operation in test gas (Toxic gas detection equipment only)	53
68	5.4.23	Orientation	54
69	5.4.24	Battery capacity for battery-powered equipment	54
70	5.4.25	Power supply variations for externally powered equipment	55
71	5.4.26	Electromagnetic immunity	55
72	5.4.27	Fault signal tests	58
73	5.4.28	Software controlled equipment	59
74	5.4.29	Protection against water	60
75	5.4.30	Water vapour interference (Type OF & Type OT)	61
76	5.4.31	Alignment (Type OF & Type OT)	62
77	5.4.32	Beam block fault (Type OF & Type OT)	62
78	5.4.33	Partial obscuration (Type OF & Type OT)	63
79	5.4.34	Long range operation (Type OF & Type OT)	63
80	5.4.35	Direct solar radiation (Type OF & Type OT)	63
81	Annex A (normative)	Acceptance Criterion for Flammable Gas Detection Equipment	65
82	A.1	Flammable gas detection equipment (Type Group I)	65
83	A.2	Flammable gas detection equipment (Type Group II)	67
84	A.3	Hydrogen gas detection equipment (Type Group II)	69
85	Annex B (normative)	Acceptance Criterion for Oxygen Gas Detection Equipment	73
86	B.1	Oxygen gas detection equipment (Type DE & Type IN)	73
87	Annex C (normative)	Acceptance Criterion for Toxic Gas Detection Equipment	75
88	C.1	Toxic gas detection equipment (Type SM & Type HM)	75
89	Annex D (normative)	Acceptance Criterion for Open Path Gas Detection Equipment	78
90	D.1	Open path gas detection equipment (Type OF and OT)	78
91	Annex E (informative)	Determination of time of response and time of recovery	80
92	E.1	Aspirated equipment	80
93	E.1.1	Test rig	80
94	E.1.2	Equipment without internal pump	80
95	E.1.3	Equipment with internal pump	80
96	E.2	Diffusion equipment	81
97	E.2.1	Calibration mask method	81

98	E.2.2	Diffusion or flow methods	81
99	Annex F (informative)	Open Path Gas Detection Equipment Water Vapour Test	
100	Apparatus.....		83
101	F.1	Open path gas detection equipment water vapour test apparatus	83
102	■	(informative) Equipment Types.....	84
103	F.2	Equipment types covered by this standard	84
104	Bibliography		86
105			
106	Figure 1 – Relationship between indication range and measuring range		15
107	Figure 2 – Warm-up time in clean air (typical).....		17
108	Figure 3 – Warm-up time in standard test gas (typical)		17
109	Table 1 – Indicators		21
110	Figure 1 – Equipment for gas calibration and speed of response test		36
111	Figure E.1 – Schematic example of test rig for use with aspirated equipment		81
112	Figure E.2 – Schematic example of test chamber for diffusion method		82
113	Figure E.3 – Schematic example of test chamber for flow method		82
114	Figure F.1 – Water vapour test apparatus.....		83
115			
116			

INTERNATIONAL ELECTROTECHNICAL COMMISSION

**EXPLOSIVE ATMOSPHERES - Part 29-0: Gas detection oners equipment –
General requirements and test methods**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 60079-29-0 has been prepared by the IEC technical committee 31: Equipment for explosive atmospheres.

The text of this International Standard is based on the following documents:

FDIS	Report on voting

Full information on the voting for the approval of this International Standard can be found in the report on voting indicated in the above table.

This document has been drafted in accordance with the ISO/IEC Directives, Part 2.

A list of all parts in the IEC 60079-29 series can be found on the IEC website.

166 The committee has decided that the contents of this document will remain unchanged until the
167 stability date indicated on the IEC website under "<http://webstore.iec.ch>" in the data related to
168 the specific document. At this date, the document will be

- 169 • reconfirmed,
- 170 • withdrawn,
- 171 • replaced by a revised edition, or
- 172 • amended.

173

IMPORTANT – The 'colour inside' logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.

174

175

INTRODUCTION

~~This part of IEC 60079-29 specifies general requirements for the design and describes the test methods that apply to portable, transportable and fixed equipment intended to measure the concentration of gases and vapours in industrial and commercial applications including workplace atmospheres.~~

This part of IEC 60079-29 specifies general requirements that apply to flammable, oxygen and toxic gas detection equipment intended to detect gases and vapours for personal or property protection from an immediate hazard in industrial and commercial applications. This document also includes test methods and acceptance criteria for performance of gas detection equipment whose primary purpose is to provide an indication, alarm or other output function.

~~Performance requirements are defined in supplementing parts of IEC 60079-29. Because a wide range of environmental conditions can be They are intended to apply under environmental conditions present at the site of operation. However, because a wide range of environmental conditions are encountered in practice, this these supplementing parts specify requirements that have to be fulfilled by equipment when tested under prescribed laboratory conditions.~~

~~Electrical equipment used for the direct detection and direct concentration measurement of toxic gases and vapours generate readings in clean air (nominally zero), which vary with environmental conditions and time. This document therefore includes test methods and requirements for acceptable variations in measured values at application of zero gas and of defined test gases.~~

Terms and definitions, general requirements, test methods and performance requirements for Type HM toxic gas detection equipment that complement this document are set out in IEC 62990-1Ed.2.

Consideration should also be given to the following relevant standards:

~~General requirements for construction, testing and performance of open path detectors for flammable gases are set out in IEC 60079-29-4, Explosive atmospheres — Part 29-4: Gas detectors — Performance requirements of open path detectors for flammable gases.~~

Guidance for the selection, installation, use and maintenance of gas detecting equipment is set out in IEC 60079-29-2: *Explosive atmospheres – Part 29-2, Gas detectors – Selection, installation, use and maintenance of detectors for flammable gases and oxygen.*

Guidance for the selection, installation, use and maintenance of toxic gas detecting equipment is set out in IEC 62990-2: Workplace atmospheres — Part 2: Gas detectors — Selection, installation, use and maintenance of detectors for toxic gases and vapours.

Guidance for functional safety of fixed gas detection systems is set out in IEC 60079-29-3: *Explosive atmospheres – Part 29-3, Gas detectors – Guidance on functional safety of fixed gas detection systems.*

Part 29-0: Gas detection ~~ioners~~ equipment – General requirements and test methods

1 Scope

~~This part of IEC 60079-29 specifies general requirements for the design and describes the test methods as well as performance requirements that apply to portable, transportable and fixed equipment intended to measure the concentration of gases and vapours in industrial and commercial applications including explosive atmospheres, mines susceptible to firedamp, and workplace atmospheres.~~

~~This part of IEC 60079-29 specifies general requirements that apply to flammable, oxygen and toxic gas detection equipment intended to detect gases and vapours for personal or property protection from an immediate hazard in industrial and commercial applications. This document also includes test methods and acceptance criteria for performance of This document is applicable to gas detection equipment whose primary purpose is to provide an indication, alarm or other output function. It is applicable to equipment in which the sensor generates an electrical/electronic signal correlating with the target gas concentration.~~

~~NOTE: The term Gas Detection Equipment is often referred to as the term Gas Detector.~~

This document applies to the following gas detection equipment:

- Gas Type “FL”:
Gas detection equipment intended for the detection of flammable gas.
- Gas Type “O2”:
Gas detection equipment intended for the detection of Oxygen.
- Gas Type “TO”:
Gas detection equipment intended for the detection of toxic gas.

NOTE 1: Occupational exposure measurement in the region of Occupational Exposure Limit Values utilizes Type HM gas detection equipment with performance requirements in accordance with IEC 62990-1.

- ~~• Type Group I Flammable gas detection equipment:
Equipment intended for the detection of flammable gases in mines susceptible to firedamp.~~
- ~~• Type Group II Flammable gas detection equipment:
Equipment intended for the detection of flammable gases in explosive atmosphere other than mines susceptible to fire damp.~~
- ~~• Type DE oxygen detection equipment:
Equipment intended for monitoring oxygen deficiency or oxygen enrichment.~~
- ~~• Type IN oxygen detection equipment:
Equipment for monitoring inertisation as measuring function for explosion protection.~~
- ~~NOTE: Inertisation is an explosion protection technique where an explosive atmosphere is purged with inert gas.~~
- ~~• Type SM (Safety Monitoring) toxic gas detection equipment:
Equipment for general gas detection applications (e.g. safety warning, leak detection), and typically using alarm signalling.~~

~~NOTE: Toxic gas detection equipment for Health Monitoring (Type HM) performance requirements reside in IEC 62990-1.~~

- ~~• Type OF (Open Path Flammable) and Type OT (Open Path Toxic) gas detection equipment: Equipment for the detection and measuring of gases or vapours in ambient air by measuring the spectral absorption by the gases or vapours over an open area or cross-duct optical path.~~

~~For Type OF and Type OT, this document specifies performance requirements for portable, transportable and fixed equipment for the detection and measuring of gases or vapours in ambient air by measuring the spectral absorption by the gases or vapours over optical paths typically ranging from less than one metre up to several hundred metres.~~

~~Such equipment measures the integral concentration of the absorbing gas over the optical path in units such as LFL metre for flammable gases or ppm metre for toxic gases.~~

~~NOTE 1 Actual values of concentration in units such as %LFL or ppm can be deduced only where it can be assumed that the concentration is uniform over the optical path. See Appendix X for additional guidance.~~

~~NOTE 2 The 'integral concentration' of the absorbing gas is sometimes termed the 'column density' of the absorbing gas.~~

~~NOTE 3 This standard is based upon absorption techniques using infrared radiation. Other techniques and applications may require additional test considerations to address the particular characteristics of those approaches.~~

~~Equipment falling within the scope of this standard is classified by the following configurations:~~

- ~~a) separate optical transmitter and receiver, located at respective ends of the path through the atmosphere to be monitored.~~
- ~~b) an optical transceiver (i.e. combined transmitter and receiver) and a dedicated reflector located at respective ends of the path through the atmosphere to be monitored.~~
- ~~c) an optical transceiver that utilizes topographical features in the environment instead of a dedicated reflector at the end of a path through the atmosphere to be monitored.~~

~~Additionally, open path gas detectors are characterised by the following features:~~

- ~~a) The optical transmitter and receiver, or transceiver and reflector, are mechanically decoupled.~~
- ~~b) The optical path length may vary according to particular applications and installations (i.e. it is not fixed).~~
- ~~c) There is no single mechanical enclosure of any type across the optical path (e.g. environmental cover).~~
- ~~d) It is assumed the gas concentration across the optical path is not uniform.~~

~~The particular technical challenges experienced by open path gas detectors that are not experienced by single point detectors include:~~

- ~~a) The maintenance of proper alignment between transmitter and receiver or transceiver and reflector.~~
- ~~b) The detection and resolution of obstructions in the optical path.~~
- ~~c) Ensuring immunity from interfering light sources.~~

~~This standard is also applicable when an equipment manufacturer makes any claims regarding any special features of construction or performance that exceed the minimum requirements of this standard or when the equipment is used in applications where a uniform gas concentration is assumed to be present. All such claims shall be verified and the test procedures should be extended or supplemented, where necessary, to verify the claimed performance. The additional tests shall be agreed between the manufacturer and the test laboratory and identified and described in the test report.~~

~~This standard does not apply to any of following:~~

~~a) equipment consisting of a passive optical receiver without a dedicated optical source;~~

~~b) equipment intended for the detection of dusts or mists in air;~~

~~c) equipment intended for the detection of explosives; and~~

~~This standard is applicable to equipment which is intended for use in hazardous or non-hazardous areas, or both. Equipment for use in hazardous areas is also required to have explosion protection in accordance with the 60079 series of standards.~~

NOTE 2: This standard is intended to provide for the supply of equipment giving a level of performance suitable for general purpose applications. However, for specific applications a prospective purchaser or an appropriate authority ~~might~~^{may} additionally require equipment to be submitted for particular tests or approval. Such tests or approval are regarded as additional to and separate from the compliance with this document~~provisions of the standards referred to above.~~

This document is not applicable to equipment:

– used for medical applications;

– used only in laboratories for analysis or measurement;

– used only for process monitoring or control purposes (such as a gas analyser);

– used in the domestic environment;

– used in environmental air pollution monitoring;

– used for flue gas analysis;

– used for sampling systems external from the gas detection equipment;

– with samplers and concentrators such as sorbents or paper tape having an irreversible indication;

– consisting of a passive optical receiver without a dedicated optical source.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60068-2-6, *Environmental testing – Part 2-6: Tests – Test Fc: Vibration (sinusoidal)*

IEC 60079-0, *Explosive atmospheres – Part 0: Equipment – General requirements*

IEC 60529, *Degrees of protection provided by enclosures (IP Code)*

IEC 61000-4-29, *Electromagnetic compatibility (EMC) – Part 4-29: Testing and measurement techniques – Voltage dips, short interruptions and voltage variations on d.c. input power port immunity tests*

IEC 61000-6-3, *Electromagnetic compatibility (EMC) – Part 6-3: Generic standards – Emission standard for residential, commercial and light-industrial environments*

IEC 61000-6-4, *Electromagnetic compatibility (EMC) – Part 6-4: Generic standards – Emission standard for industrial environments*

347 IEC 61326-1, *Electrical equipment for measurement, control and laboratory use – EMC*
348 *requirements – Part 1: General requirements*

349 [IEC 80079-20-1, Explosive atmospheres – Part 20: Material characteristics for gas and vapour](#)
350 [classification, Section 1: Test methods and data](#)

351

352 **3 Terms and definitions**

353 For the purposes of this document, the following terms and definitions apply.

354 ISO and IEC maintain terminological databases for use in standardization at the following
355 addresses:

356 • IEC Electropedia: available at <http://www.electropedia.org/>

357 • ISO Online browsing platform: available at <http://www.iso.org/obp>

358 **3.1** 359 **gas properties**

360 **3.1.1**
361 **ambient air, <gas detection>**
362 normal atmosphere surrounding the equipment

363 **3.1.2**
364 **clean air, <gas detection>**
365 air that is free of gases or vapours to which the sensor is sensitive or which influence the
366 performance of the sensor

367 **3.1.3**
368 **reference air, <gas detection>**
369 air with an oxygen concentration of $(21 \pm 0,4) \%$

370 **3.1.4**
371 **zero test gas, <gas detection>**
372 gas, which is free of the gas(es) to be measured and interfering and contaminating substances,
373 the purpose of which is calibration/adjustment of the equipment zero

374 **3.1.5**
375 **standard test gas, <gas detection>**
376 test gas to be used for all tests unless otherwise stated

377 **3.1.6**
378 **flammable gas, <gas detection>**
379 ADMITTED: flammable gases
380 DEPRECATED: combustible gas
381 gas or vapour which, when mixed with air in a certain proportion, will form an explosive
382 atmosphere

383 Note 1 to entry: For the purposes of this part of IEC 60079-29, the term "flammable gas" includes flammable
384 vapours.

385 Note 2 to entry: For the purposes of this part of IEC 60079-29, the terms "combustible gas" and "flammable gas"
386 are equivalent.

387 [SOURCE: IEC 60070-10-1:2020 with "or vapour" dropped from term and two notes to entry added. Admitted and
388 deprecated terms also added]

3.1.7**toxic gas, <gas detection>**

gas or vapour that can be harmful to human health and/or the performance of persons due to its physical or physico-chemical properties

3.1.8**poisons, <gas detection>**

<for sensing elements> substances that lead to temporary or permanent change of performance, particularly loss of sensitivity of the sensing element

3.1.9**volume fraction (v/v)**

quotient of the volume of a specified component and the sum of the volumes of all components of a gas mixture before mixing, all volumes referring to the pressure and the temperature of the gas mixture

Note 1 to entry: The volume fraction and volume concentration take the same value if, at the same state conditions, the sum of the component volumes before mixing and the volume of the mixture are equal. However, because the mixing of two or more gases at the same state conditions is usually accompanied by a slight contraction or, less frequently, a slight expansion, this is not generally the case.

3.2**gas detection equipment****3.2.1****alarm-only equipment, <gas detection>**

equipment with an alarm but not having an indication of measured value

3.2.2**aspirated equipment, <gas detection>**

equipment that samples the gas by drawing it to the gas sensor

Note 1 to entry: A hand operated or electric pump is often used to draw gas to the sensor.

3.2.3**automatically aspirated equipment, <gas detection>**

aspirated equipment with an integral pump or separate pump, which is connected directly to the equipment

3.2.4**diffusion equipment, <gas detection>**

equipment in which the transfer of gas from the atmosphere to the sensor takes place without aspirated flow

3.2.5**fixed equipment**

equipment fastened to a support, or otherwise secured in a specific location when energized

[SOURCE: IEC 60070-0:2017]

3.2.6**portable equipment, <gas detection>**

equipment intended to be carried by a person during its operation

Note 1 to entry: Portable equipment carried by a person during its operation is sometimes referred to as hand-held equipment.

Note 2 to entry: Hand-held gas detection equipment, typically less than 1 kg, requires use of only one hand to operate. Larger equipment that can be operated by the user while it is carried either by hand, by a shoulder strap or carrying harness, can be equipped with or without a hand directed probe.

[SOURCE: IEC 60070-0:2017 with addition of Note 2 to entry]

3.2.7**transportable equipment**

equipment not intended to be carried by a person during operation, nor intended for fixed installation

[SOURCE: IEC 60070-0:2017]

3.2.8**gas detection transmitter**

fixed gas detection equipment that provide a conditioned electronic signal or output indication to a generally accepted industry standard (such as 4-20 mA), intended to be utilized with separate gas detection control units or signal processing data acquisition, central monitoring and similar systems, which typically process information from various locations and sources including, but not limited to gas detection equipment

3.2.9**gas detection control unit**

equipment intended to provide display indication, alarm functions, output contacts or alarm signal outputs or any combination when operated with remote sensor(s)

3.2.10**separate gas detection control unit**

equipment intended to provide display indication, alarm functions, output contacts or alarm signal outputs or any combination when operated with gas detection transmitter(s)

3.2.11**equipment with integral sensor(s), <gas detection>**

equipment that provides display indication, alarm functions, output contacts or alarm signal outputs using a sensor which is within or directly assembled to the equipment housing

3.2.12**accessory, <gas detection>**

component which can be fitted to the equipment for special purpose

EXAMPLE: External gas pump, sampling probe, hoses, collecting cone, weather protection device.

3.3**sensors****3.3.1****sensing element, <gas detection>**

part of the sensor that is sensitive to the gas or vapour to be measured

3.3.2**measuring principle, <gas detection>**

principle that makes the sensing element or the sensor sensitive to the gas or vapour to be measured

3.3.3**sensor, <gas detection>**

assembly in which the sensing element is housed and that may also contain associated circuit components

3.3.4**integral sensor, <gas detection>**

sensor that is within or directly assembled to the equipment housing

3.3.5**remote sensor, <gas detection>**

sensor which is installed separately, but is connected to a gas detection control unit, gas detection transmitter, or to transportable or portable equipment

3.4**supply of gas to equipment****3.4.1****sample line, <gas detection>**

means by which the gas being sampled is conveyed to the sensor

Note 1 to entry: Accessories such as filters or water traps are often included in the sample line.

3.4.2**sampling probe, <gas detection>**

separate accessory sample line which is optionally attached to the equipment

Note 1 to entry: It is usually short

3.4.3**Adjustment, <gas detection>**

procedure carried out to minimize the deviation of the measured value from the test gas concentration

Note 1 to entry: When the equipment is adjusted to give an indication of zero in zero test gas, the procedure is called 'zero adjustment'.

3.4.4**calibration, <gas detection>**

procedure which establishes the relationship between an indication and the concentration of a test gas

3.4.5**calibration kit, <gas detection>**

means of presenting test gas to the equipment for the purpose of calibrating, adjusting or verifying the operation of the equipment

Note 1 to entry: The calibration kit can be used for verifying the operation of the alarms if the concentration of the test gas is beyond the alarm set-point.

Note 2 to entry: A mask for calibration and test is an example of a field calibration kit.

3.4.6**mask for calibration and test, <gas detection>**

device that can be attached to the equipment to present a test gas to the sensor in a reproducible manner

3.5**signals and alarms****3.5.1****measured value, <gas detection>**

calculated concentration of gas or vapour that results from processing the sensor signal

Note 1 to entry: The measured value may be further processed before indication on output or display (e.g. filtering, averaging, suppression of indication).

3.5.2**indication, <gas detection>**

representation of the measured value on an output or display

3.5.3**indication range, <gas detection>**

range of indications of gas concentration over which the equipment is capable of indicating

SEE Figure 1.

3.5.4**lower limit of indication, <gas detection>**

smallest indication within the indication range

SEE Figure 1.

3.5.5**upper limit of indication, <gas detection>**

largest indication within the indication range

SEE Figure 1.

3.5.6**measuring range, <gas detection>**

range of measured values of gas concentration over which the accuracy of the equipment lies within specified limits

SEE Figure 1.

3.5.7**lower limit of measurement, <gas detection>**

smallest measured value within the measuring range


SEE Figure 1.

3.5.8**upper limit of measurement, <gas detection>**

largest measured value within the measuring range

SEE Figure 1.

Note 1 to entry: Indications below the lower limit of measurement or above the upper limit of measurement will not necessarily meet the requirements of this document.



IEC

Figure 1 – Relationship between indication range and measuring range

3.5.9**alarm set point, <gas detection>**

setting of the equipment at which the measured concentration will cause the equipment to initiate an indication, alarm or other output function

3.5.10**latching alarm, <gas detection>**

alarm that, once activated, requires deliberate action to be deactivated

3.5.11**fault signal, <gas detection>**

audible, visible or other type of output, different from the alarm signal, permitting, directly or indirectly, a warning or indication that the equipment is not working satisfactorily

3.5.12**special state, <gas detection>**

state of the equipment other than those in which monitoring of gas concentration or alarming is the intent

Note 1 to entry: Special state includes warm-up, calibration mode and fault condition.

3.6**times****3.6.1****drift, <gas detection>**

variation in the equipment indication over time at any fixed gas volume fraction (including clean air) under constant ambient conditions

3.6.2**final indication, <gas detection>**

indication given by the equipment after stabilization

3.6.3**time of response, <gas detection>**

$t(x)$

time interval, with the equipment in a warmed-up condition, between the time when an instantaneous rising concentration between two test gases is produced at the equipment inlet, and the time when the indication reaches a stated percentage (x) of the differences of the two indications

Note 1 to entry: For alarm-only equipment the stated indication can be represented by the activation of the alarm set at a stated value.

3.6.4**time of recovery, <gas detection>**

$t(x)$

time interval, with the equipment in a warmed-up condition, between the time when an instantaneous falling concentration between two test gases is produced at the equipment inlet, and the time when the indication reaches a stated percentage (x) of the differences of the two indications

Note 1 to entry: For alarm-only equipment the stated indication can be represented by the de-activation of the alarm set at a stated value.

3.6.5**warm-up time, <gas detection>**

time interval, with the equipment in a stated atmosphere, between the time when the equipment is switched on and the time when the indication reaches and remains within the stated tolerances

Note 1 to entry: See equipment warm-up time in Figures 2 and 3.




Figure 2 – Warm-up time in clean air (typical)




Figure 3 – Warm-up time in standard test gas (typical)

3.7 open path

3.7.1 beam block signal, <gas detection>

3.7.2 audible, visual or other type of output which provides, directly or indirectly, a warning or indication that the optical path is obscured or that the signal detected is too weak to enable the equipment to function normally

inhibition signal, <gas detection>

audible, visual, or other type of output which provides, directly or indirectly, a warning or indication that normal operation has been suspended

3.7.3 integral concentration, <gas detection>

mathematical integral of the gas concentration along the optical path.

Note 1 to entry: It is expressed in units of concentration multiplied by distance, e.g. LFL.metre for flammable gases or ppm.metre for toxic gases

Note 2 to entry: 100 % LFL x 1 metre = 1 LFL.metre; 10 % LFL x 10 metre = 1 LFL.metre

3.7.4**open path, <gas detection>**

distance in space which traverses the area (or part of the area) in which the atmosphere is being monitored and through which gases in the atmosphere are free to move

3.7.5**optical axis, <gas detection>**

median line of the optical path

3.7.6**optical path, <gas detection>**

path traversed by optical radiation from an optical transmitter to an optical receiver.

Note 1 to entry: The radiation can traverse the open path once, twice or many times depending on the form taken by the instrument

3.7.7**open path transmitter, <gas detection>**

assembly in which the optical transmitting element(s) are housed and which may contain associated optical and electrical components

3.7.8**open path transceiver, <gas detection>**

assembly in which the optical detecting element(s) and optical transmitting element(s) are housed and which may contain associated optical and electrical components

3.7.9**open path receiver, <gas detection>**

assembly in which the optical detecting element(s) are housed and which may contain associated optical and electrical components

4 General requirements**4.1 Introduction**

In addition to the requirements of this document, gas detection equipment intended for use in explosive atmospheres shall comply with the relevant requirements of IEC 60079-0 and the Type(s) of Protection standards. The ambient temperature and pressure ranges of such equipment conforming to this document shall not exceed the ambient temperature and pressure ranges of the Type(s) of Protection.

The following figure [identifies components](#) ~~provides the general concept~~ of the gas detection equipment ~~and gas detection system~~:




Figure 4: Functional Components of Gas Detection Equipment

4.2 Design

4.2.1 General

~~The equipment shall be constructed in such a manner that regular functional checks can be undertaken by the user and that it can be equipped with a field calibration kit for application of test gas.~~

~~Gas detection equipment shall be constructed of materials known to be resistant to degradation by the environment in which they are intended to be used. Equipment specifically intended for use in the presence of corrosive vapours or gases, or which can produce corrosive by products as a result of the detection process (e.g. catalytic oxidation or other chemical process) shall be constructed of materials known to be resistant to corrosion by such substances.~~

~~Materials and components used in the construction of the equipment which are relevant to performance of the equipment shall be used within the material and component manufacturer's ratings or limitations unless otherwise specified by appropriate safety standards. If such material and component ratings are not available, demonstrated functionality of the equipment may be used. All materials along the gas path shall not cause an inaccurate reading due to adsorption and absorption effects of materials in contact with the gas(es) to be detected.~~

The equipment shall be constructed in such a manner that:

- a) regular functional checks can be undertaken by the user; and
- b) that it can be equipped with a field calibration kit for application of test gas;
- c) materials and components used in the construction of the equipment which are relevant to performance of the equipment shall be used within the material and component manufacturer's ratings or limitations unless otherwise specified by appropriate safety standards. If such material and component ratings are not available, demonstrated functionality of the equipment may be used. All materials along the gas path shall not cause an inaccurate reading due to adsorption and absorption effects of materials in contact with the gas(es) to be detected;
- d) any function of the gas detection equipment that is defined by the manufacturer as safety related shall not be adversely affected by any failure of non-safety related function of the equipment; and

EXAMPLE: When the equipment has an analogue output safety function and a digital output non-safety function, any failure of the digital output shall not interfere with the function of the analogue output.

~~For aspirated equipment and automatically aspirated equipment, inlet and outlet ports shall be unambiguously marked to ensure the correct connection of any sample and exhaust lines.~~

~~Any failure of the gas detection equipment that is not relevant to safety or health shall not adversely affect those functions of the equipment that are related to safety.~~

~~EXAMPLE: Equipment with 4 to 20 mA safety related output and with Highway Addressable Remote Transducer (HART) non safety related communication. Any failure of HART shall not affect the 4 to 20 mA output.~~

4.2.2 Indicators and means of indication ~~of measured values~~

4.2.2.1 Indicators

Distinguishable indicators shall be provided as shown in Table 1:

State	Equipment Type		
	Portable	Transportable	Fixed
~~Power~~	~~Visual~~	~~Visual~~	~~Visual~~
Alarm	Visual and Audible	Visual	Visual
Fault	Visual and Audible	Visual	Visual
<u>Power</u>	<u>Visual</u>	<u>Visual</u>	<u>Visual</u>
<u>Low battery</u>	<u>Visual and Audible</u>	<u>Visual and Audible</u>	<u>Visual or Audible</u>
Warm-up	Visual or Audible	Visual	Visual
Calibration	Visual	Visual	Visual

Other special state	Visual	Visual	Visual
---------------------	--------	--------	--------

Table 1 – Indicators

A display may be utilized as a means of a visual indicator.

For fixed equipment, the indicators shall be provided at the gas detection control unit.

~~It is permitted that the indications related to gas detection transmitters and remote sensors are shown only at the separate gas detection control unit or gas detection control unit.~~

If audible indicators are provided for transportable or fixed equipment, they shall at a minimum indicate alarms, and may additionally indicate special states.

4.2.2.2 Colours of visual indicators

If only one visual indicator is provided for signalling alarms, special states and other indications, it shall be coloured red.

If a display is the only visual indicator, indicator colour is not required but shall provide distinguishable text or symbol for intended function.

If separate visual indicators are used or if a multi-coloured visual indicator is provided, the colours shall be used in the following order of priority ((a) being highest priority):

- a) alarm indicators shall be coloured RED;
- b) special state indicators shall be coloured YELLOW;
- c) power ~~supply~~ indicators shall be coloured GREEN;
- d) normal operation indicators, if fitted, shall be coloured GREEN.

If a multi-coloured visual indicator is provided, the equipment shall provide an additional indication by means other than colour.

If there is more than one visual indicator of the same colour with different functions, the function of each indicator shall be provided adjacent to each indicator.

NOTE This may be achieved by printed labels, text, marks or icons on a display.

4.2.2.3 Display and other outputs for indication ~~of measured values~~

For equipment with a display for indication shall indicate of measured values, the gas ~~concentration name~~, ~~measured value and the~~ unit of measurement and the gas name for Type O₂ and Type TO shall be indicated. It is permissible to mark the ~~gas name and~~ unit of measurement and the gas name adjacent to the display. It is permissible to provide abbreviations.

Within the measuring range, the resolution of the ~~display and all other outputs for indication of the measured value~~ shall be less than or equal to 1 % of the upper limit of measurement. For oxygen detection equipment, the resolution of the indication shall be less than or equal to 0,1 % (v/v) of oxygen.

Within the measuring range, all outputs shall present the same value within the resolution of each output.

NOTE: These outputs can include analogue signals such as 4-20 mA, digital signals such as ModBus and RS485, or displayed value.

Any measurements above the upper limit of indication shall be clearly indicated.

If the equipment has more than one indication range, the range selected shall be clearly identified.

For alarm-only equipment, the manufacturer shall identify suitable methods for connecting, indicating or recording devices for the purpose of testing the compliance of the equipment with this document. The indication on the display and all other outputs for indication of the measured value shall not contradict the results obtained by the indicating or recording devices.

NOTE Such suitable methods will be necessary for alarm-only equipment.

4.2.2.4 Suppression of indications

Suppression of indication (i.e. deadband) is only permitted according to the following rules:

1. For Type DE equipment: Measured values within an interval of $\pm 0,5$ % (v/v) oxygen around the nominal indication for reference in ambient air may be indicated as the nominal indication for reference in ambient air.

NOTE 1 These requirements are for the equipment operating on site under varying air conditions such as temperature, humidity, velocity, and pressure.

2. For all other types of equipment:

- a) Measured values within the measuring range above the suppression threshold value (i.e. between the lower limit of measurement and the upper limit of measurement) shall be indicated;
- b) Measured values below the suppression threshold value (values below zero included) specified in the applicable performance standard shall be indicated as:
 - zero; or
 - the measured value; or
 - another indication that the measured value is below the lower limit of measurement;
- ~~c) Below minus 2 times the suppression threshold value the equipment shall indicate the measured value or provide a fault signal.~~

NOTE 2 Under-range faults are defined in 4.2.4.

3. It shall be possible to configure the equipment such that in calibration mode the any kind of suppression of indication the measured value shall be automatically disabled.

NOTE 3 The deviation of the zero indication may be beneficial for maintenance purpose.

4. It shall be possible to configure the equipment such that in measuring mode any kind of suppression of indication is disabled.

4.

4.2.3 Alarm signal(s)

Alarm and time-weighted average (TWA) alarm set points shall not be adjustable below zero and the upper level of the indication outside the measuring range.

For equipment with multiple configurable alarm setpoints, there shall be a documented order for the alarms from most sensitive to the least sensitive. ~~A~~The more sensitive setpoint shall always produce an alarm response prior to or equal to the less sensitive setpoint.

It is permitted that alarms are configurable to be latching or non-latching.

~~If alarms are not configurable to be latching or non-latching, a~~ At least one alarm shall be configured latching as a factory default setting. The factory setting at delivery may deviate from the factory default setting on user request.

If alarms are not configurable to be latching or non-latching, at least one alarm ~~shall be configured latching as a factory default setting. The factory setting at delivery may deviate from the factory default setting on user request.~~

For Type ~~O2 DE~~ oxygen detection equipment, the latching properties of alarms for oxygen deficiency at decreasing oxygen concentration or for oxygen enrichment at increasing ~~or decreasing~~ oxygen concentration shall be considered separately.

Alarms shall remain in operation while the alarm condition is present. Audible alarm signals may be silenced by the user if other alarm signal(s) remain active.

If it is possible to ~~disable/de-activate~~ alarm indicators or alarm signal outputs, ~~this deactivation shall be indicated by~~ a signal shall be provided to indicate that alarms are disabled. For fixed equipment, this deactivation shall include a transmittable output signal if the alarm indicators or alarm signal outputs are not automatically re-enabled within 15 min.

EXAMPLE It might be necessary to de-activate alarm indicators or alarm signal outputs during calibration/adjustment and maintenance procedures.

4.2.3.1 ~~4.2.3.1~~ Group I portable equipment indicating up to 5 % (v/v)

Alarm devices shall not be adjustable above 3 % (v/v). An additional ~~over-range~~ alarm, which indicates when upper limit of measurement~~full-scale~~ has been exceeded, may be provided.

4.2.3.2 ~~4.2.3.2~~ Group II portable equipment indicating up to 100 % LFL

Alarm devices shall not be adjustable above 60 % LFL. An additional ~~over-range~~ alarm, which indicates when upper limit of measurement~~full-scale~~ has been exceeded, may be provided.

4.2.3.3 ~~4.2.3.3~~ Open path gas detection equipment

When non-latching alarm devices, output contacts, or signal outputs are provided to indicate detection of an integral gas concentration or path-averaged %LFL in excess of a pre-set alarm level, the fact shall be indicated clearly and prominently in the instructions.

The operation of any other output functions shall be clearly stated in the manual.

4.2.4 Fault signals

Externally powered equipment shall provide a fault signal in the event of ~~failure of power~~ failure to the equipment or if the supply voltage falls below the lower supply voltage fault limit as specified by the manufacturer.

~~Externally powered equipment shall provide a fault signal when the supply voltage falls below the lower supply voltage fault limit specified by the manufacturer.~~

NOTE The absence of a persistent signal can be interpreted as a fault signal (e.g. 0 mA or -20 mA signal).

A short-circuit or open-circuit in any field wiring connections shall be indicated by a fault signal. It is permitted for the equipment to enter the "calibration" special state under the short-circuit or open-circuit condition if this calibration activation procedure is specified in the instruction manual/instructions.

~~Under the above conditions~~ The equipment may also indicate an alarm under field wiring fault conditions.

Measured values below zero (~~for example, e.g.~~ caused by drift) shall be indicated by ~~an under-range~~ fault signal at no less than -10 % of the upper limit of measurement ~~at the latest~~. Equipment with a measuring range up to 20 % LFL shall provide a fault signal at no greater than -20 % of the upper limit of the measuring range. This requirement is not applicable for Type DE oxygen detection equipment.

~~Equipment where the sensor or sensing element can be replaced without opening the housing shall provide a fault signal if the sensor or sensing element is not re-connected within 15 min.~~

Automatically aspirated equipment shall be provided with an integral flow-indicating device that produces a fault signal in the event of low flow.

Equipment with an optical sensor shall provide a fault signal in the event the optical beam is blocked.

Equipment where the sensor or sensing element can be replaced without entering a special state, the equipment shall provide a fault signal if the sensor or sensing element is disconnected and not re-connected within 15 min.

4.2.5 Special state indication

If a special state is entered by equipment this shall be indicated by a signal. For fixed equipment, this shall include a contact or other transmittable output signal.

4.2.5.2.6 Adjustments

All adjustment functions shall be designed so as to discourage unauthorized or inadvertent interference with the equipment.

EXAMPLES: Procedural devices, in the case of a keyboard instrument, or mechanical devices such as a cover requiring the use of a special tool.

~~Ex Equipment Fixed explosion-protected equipment housed in explosion-protected enclosures~~ shall be designed so that, if any facilities for adjustment are necessary for routine ~~re~~calibration and for resetting or like functions, these facilities shall be externally accessible and shall not degrade the ~~Ex Type of explosion-p~~Protection of the equipment.

The adjustments to the reference air for Type DE equipment and to of the zero test gas for all other Types of equipment and of the sensitivity shall be designed so that adjustments of one will not affect the other.:-

~~a) adjustment of one will not affect the other;~~

~~or~~

~~b) it shall not be possible to adjust only one and the sequence of adjustments shall ensure that the affected one is adjusted second. A zero adjustment may be allowed without a mandatory span adjustment provided the new zero is no more than 5 % of the measuring range when referenced to the last adjustment of zero and sensitivity.~~

The equipment shall not perform an automatic adjustment in ambient air during warmstart-up without acknowledgement of an trained/authorised user. If equipment prompts the user for

adjustment in ambient air during warmstart-up and the user makes no selection, equipment shall continue to warmstart-up without adjustment, after a delay of not more than 15 s.

4.2.64.2.7 Battery-powered equipment

Equipment powered with integral batteries shall be provided with indication of low battery condition. ~~For portable and transportable equipment, t~~The indication of low battery condition shall be in accordance with Table 1~~visual and audible~~. It is permitted that the audible signal can be silenced only after activation.

The equipment shall be able to operate and signal alarms for at least 15 min after indication that the low-battery condition has been reached.

~~4.2.7 Gas detection transmitter for use with separate gas detection control units~~

~~A specification shall be included in the instruction manual~~instructions~~that describes the relationship the gas concentration (measured by the equipment) has with the corresponding indication (transfer function). As a minimum, the relationship between the gas concentrations at 0 %, 10 %, 30 %, 50 %, 70 %, 90 % and 100 % of the upper limit of indication and the corresponding indications shall be defined. The upper limit of indication and status signals (e.g. fault, calibration mode) and indications of over and under range condition shall also be specified.~~

~~4.2.8 Separate gas detection control units for use with gas detection transmitter(s)~~

~~A specification shall be included in the instruction manual~~instructions~~that describes the relationship the input signal has with the calculated gas concentration (transfer function). As a minimum, the relationship between the input signals and the measured values at 0 %, 10 %, 30 %, 50 %, 70 %, 90 % and 100 % of the upper limit of indication shall be defined. Required inputs for upper limit of indication, status signals (e.g. fault, calibration mode) and indication of over and under range condition shall also be specified.~~

~~4.2.9 Open path gas detection equipment (Type OF & Type OT)~~

~~4.2.9.1 Alarm or output functions~~

~~When non-latching alarm devices, output contacts, or signal outputs are provided to indicate detection of an integral gas concentration or path-averaged %LFL in excess of a pre-set alarm level, the fact shall be indicated clearly and prominently in the instruction manual~~instructions~~.~~

~~The operation of any other output functions shall be clearly stated in the manual.~~

~~4.2.9.2 Fault signals~~

~~Equipment shall provide a fault signal if any of the following conditions (as a minimum) occur:~~

- ~~a) under range indication (below the zero point) between zero and 10 % full scale equivalent;~~
- ~~b) beam blockage;~~
- ~~c) low battery indication, if applicable;~~
- ~~d) a short-circuit or open-circuit in connections to any remote sensor, if applicable.~~

~~Such signals shall be differentiated from any alarms.~~

4.2.104.2.8 Software-controlled equipment

4.2.10.14.2.8.1 General

The risks to safety related functions arising from faults in software-controlled equipment shall be identified and mitigated.~~In the design of software-controlled equipment, the risks to safety shall~~