



DPC: 24/30506234 DC

BSI Group Headquarters

389 Chiswick High Road London W4 4AL

Tel: +44 (0)20 8996 9000

Fax: +44 (0)20 8996 7400

www.bsigroup.com

Date: 03 December 2024

Origin: International

Latest date for receipt of comments: 26 January 2025

Project No. 2022/02261

Responsible committee: CH/121/9 Lung Ventilators & Related Equipment

Interested committees:

Title: Draft BS EN ISO 12487 Medical electrical equipment — Clinical performance evaluation of clinical thermometers.

Please notify admin.start@bsigroup.com if you are aware of any keywords that might assist in clarifying or identifying the standard or of the content of this standard:

- i) has any issues related to 3rd party IPR, patent or copyright
- ii) affects other national standard(s)
- iii) requires additional national guidance or information

WARNING: THIS IS A DRAFT AND MUST NOT BE REGARDED OR USED AS A BRITISH STANDARD. THIS DRAFT IS NOT CURRENT BEYOND 26 January 2025

This draft is issued to allow comments from interested parties; all comments will be given consideration prior to publication. No acknowledgement will normally be sent. **See overleaf for information on the submission of comments.**

No copying is allowed, in any form, without prior written permission from BSI except as permitted under the Copyright, Designs and Patent Act 1988 or for circulation within a nominating organization for briefing purposes. Electronic circulation is limited to dissemination by e-mail within such an organization by committee members.

Further copies of this draft may be purchased from BSI Shop <http://shop.bsigroup.com> or from BSI Customer Services, Tel: +44(0) 20 8996 9001 or email cservices@bsigroup.com. British, International and foreign standards are also available from BSI Customer Services.

Information on the co-operating organizations represented on the committees referenced above may be obtained from <https://standardsdevelopment.bsigroup.com/>.

E-mail: admin.start@bsigroup.com

Introduction

This draft standard is based on national and international discussions. Your comments on this draft are invited and will assist in the preparation of the consequent standard.

For international standards, comments will be reviewed by the relevant UK national committee before sending the consensus UK vote and comments to the international committee, which will then decide appropriate action. If the international standard is approved, it is usual for the text to be published as a British Standard.

For national standards, comments will be reviewed by the relevant UK national committee and the resulting standards published as a British Standard.

UK Vote

Please indicate whether you consider the UK should submit a negative (with supporting technical reasons) or positive vote on this draft. Please indicate if you are aware of any reason why this draft standard should not be published as a British Standard.

Submission of Comments

- **Annotated drafts are not acceptable and will be rejected.**
- All comments should be submitted online at <https://standardsdevelopment.bsigroup.com/>. You will need to register in order to comment.

Template for comments and secretariat observations

Date: xx/xx/20xx	Document: ISO/DIS xxxx
------------------	------------------------

1	2	(3)	4	5	(6)	7
MB	Clause No./ Subclause No./Annex (e.g. 3.1)	Paragraph/ Figure/Table/Note	Type of comment	Comment (justification for change) by the MB	Proposed change by the MB	Secretariat observations on each comment submitted
	3.1	Definition 1	ed	Definition is ambiguous and needs clarifying.	Amend to read '...so that the mains connector to which no connection...'	
	6.4	Paragraph 2	te	The use of the UV photometer as an alternative cannot be supported as serious problems have been encountered in its use in the UK.	Delete reference to UV photometer.	



DRAFT International Standard

ISO/DIS 12487

Medical electrical equipment — Clinical performance evaluation of clinical thermometers

ICS: 11.040.55

ISO/TC 121/SC 3

Secretariat: **ANSI**

Voting begins on:
2024-12-03

Voting terminates on:
2025-02-25

This document is circulated as received from the committee secretariat.

This draft is submitted to a parallel vote in ISO and in IEC.

ISO/CEN PARALLEL PROCESSING

THIS DOCUMENT IS A DRAFT CIRCULATED FOR COMMENTS AND APPROVAL. IT IS THEREFORE SUBJECT TO CHANGE AND MAY NOT BE REFERRED TO AS AN INTERNATIONAL STANDARD UNTIL PUBLISHED AS SUCH.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNOLOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STANDARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION.

Reference number
ISO/DIS 12487:2024(en)



COPYRIGHT PROTECTED DOCUMENT

© ISO 2024

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

	Contents	Page
1		
2	1 Scope 1	1
3	2 Normative references 1	1
4	3 Terms and definitions.....1	1
5	4 General requirements.....7	7
6	4.1 Clinical investigation methods 7	7
7	4.2 Good clinical practice.....7	7
8	4.3 Status of previous <i>clinical investigations</i> 8	8
9	4.4 Disclosure of summary of <i>clinical investigation</i> 8	8
10	5 Clinical investigation.....8	8
11	5.1 Subject requirements.....8	8
12	5.1.1 Number.....8	8
13	5.1.2 Age distribution 8	8
14	5.1.3 Body temperature distribution 9	9
15	5.1.4 Indicated reference site 9	9
16	5.1.5 Exception criteria 10	10
17	5.2 Clinical investigation method with a <i>reference site thermometer</i> 10	10
18	5.2.1 Environmental condition of <i>clinical investigation</i> 10	10
19	5.2.2 Subject preparation for body temperature stabilization.....10	10
20	5.2.3 Reference site thermometer measurement.....11	11
21	5.2.4 Clinical thermometer-under-test measurement.....11	11
22	5.3 Procedure.....12	12
23	5.3.1 Clinical thermometers with <i>time prediction</i> 12	12
24	5.3.2 Clinical thermometers with <i>site conversion</i> 12	12
25	6 Data analysis 13	13
26	6.1 Correction of the readings of the <i>reference site thermometer</i> 13	13
27	6.2 Clinical bias calculation 13	13
28	6.3 Standard deviation of the <i>clinical bias</i> calculation.....14	14
29	6.4 Clinical repeatability calculation.....14	14
30	7 Acceptance criteria 15	15
31	7.1 Clinical bias 15	15
32	7.2 Standard deviation of the <i>clinical bias</i> 15	15
33	7.3 Clinical repeatability.....15	15
34	Annex A (informative) Particular guidance and rationale 16	16
35	A.1 General guidance 16	16
36	A.2 Rationale for particular clauses and subclauses.....16	16
37	Annex B (informative) Terminology — alphabetized index of defined terms 22	22
38	Bibliography 24	24
39		
40		

Foreword

ISO (the International Organization for Standardization) and IEC (the International Electrotechnical Commission) form the specialized system for worldwide standardization. National bodies that are members of ISO or IEC participate in the development of International Standards through technical committees established by the respective organization to deal with particular fields of technical activity. ISO and IEC technical committees collaborate in fields of mutual interest. Other international organizations, governmental and non-governmental, in liaison with ISO and IEC, also take part in the work. In the field of information technology, ISO and IEC have established a joint technical committee, ISO/IEC JTC 1.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO and IEC shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 121, *Anaesthetic and respiratory equipment*, Subcommittee SC 3, *Respiratory devices and related equipment used for patient care*, and Technical Committee IEC/TC 62, *Electrical equipment in medical practice*, Subcommittee SC 62D, *Electromedical equipment* in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 205, *Non-active medical devices*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This first edition of ISO 12487 together with the third edition of ISO 80601-2-56, cancels and replaces the second edition of ISO 80601-2-56 published in 2017 and its Amendment 1 (2019).

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

73 Determining body temperature is an important *procedure* that is clinically used to assess the status of
74 *patients* as well as blood pressure, SpO₂ and pulse rate. This document is intended to provide the
75 necessary requirements for the *clinical investigation* to ensure that the *essential performance* of these
76 *clinical thermometers* is at an adequate level.

77 This document has been drafted in accordance with the ISO/IEC Directives, Part 2.

78 In this document, the following print types are used.

79 — Requirements and definitions: roman type.

80 — *Terms defined in clause 3 of this document or as noted: italic type.*

81 — Informative material appearing outside of tables, such as notes, examples and references: in smaller
82 type. Normative text of tables is also in a smaller type.

83 In referring to the structure of this document, the term

84 — “clause” means one of the numbered divisions within the table of contents, inclusive of all
85 subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.), and

86 — “subclause” means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses
87 of Clause 7).

88 References to clauses within this document are preceded by the term “Clause” followed by the clause
89 number. References to subclauses within this particular document are by number only.

90 In this document, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination
91 of the conditions is true.

92 In this document, the following verbal forms are used:

93 — “shall” indicates a requirement;

94 — “should” indicates a recommendation;

95 — “may” indicates a permission;

96 — “can” is used to describe a possibility or capability; and

97 — “must” is used to indicate an external constraint.

98
99

Medical electrical equipment — Clinical performance evaluation of clinical thermometers

1 Scope

This document specifies the requirements and methods for the *clinical investigation* of *ME equipment* used to measure the body temperature in *indirect measurement mode*.

This document covers both intermittently and continuously measuring *clinical thermometers*.

This document is not intended for *clinical thermometers* measuring the body temperature in *direct measurement mode*.

NOTE 1 For *clinical thermometers* in *direct measurement mode* determining the technical accuracy in accordance with IEC 80601-2-56 is considered sufficient.

NOTE 2 For *clinical thermometers* with claimed measurement time shorter than 60 seconds (for methods such as oral or rectal measurement), or shorter than 5 minutes (for methods such as axillary measurement) are treated as *predictive type thermometers* and fall under the scope of this document.

This document specifies additional disclosure requirements.

This document is not applicable to the *clinical investigation* of a screening thermographs for human febrile temperature screening whose *laboratory accuracy* requirements are described in IEC 80601-2-59.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 80601-2-56:—¹, *Medical electrical equipment — Part 2-56: Particular requirements for basic safety and essential performance of clinical thermometers for body temperature measurement*

ISO 14155:—², *Clinical investigation of medical devices for human subjects — Good clinical practice*

IEC 60601-1:2005+AMD1:2012+AMD2:2020, *Medical electrical equipment — Part 1: General requirements for basic safety and essential performance +Amendment 1:2012 +Amendment 2:2020*

JCGM 100:2008, *Evaluation of measurement data — Guide to the expression of uncertainty in measurement* (GUM 1995 with minor corrections)

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

¹ Under preparation. Stage at the time of publication: ISO/DIS 80601-2-56:2024.

² Under preparation. Stage at the time of publication: ISO/DIS 14155:2024.

129 ISO and IEC maintain terminological databases for use in standardization at the following addresses:

130 — ISO Online browsing platform: available at <http://www.iso.org/obp>

131 — IEC Electropedia: available at <http://www.electropedia.org/>

132 3.1

133 accompanying information

134 information accompanying or *marked* on a medical device or accessory for the user or those accountable
135 for the installation, use, processing, maintenance, decommissioning and disposal of the medical device or
136 accessory, particularly regarding safe use

137 Note 1 to entry: The *accompanying information* shall be regarded as part of the medical device or accessory.

138 Note 2 to entry: The *accompanying information* can consist of the label, *marking*, *instructions for use*, *technical*
139 *description*, installation manual, quick reference guide, etc.

140 Note 3 to entry: *Accompanying information* is not necessarily a written or printed document but could involve
141 auditory, visual, or tactile materials and multiple media types (e.g., CD/DVD-ROM, USB stick, website).

142 [SOURCE: ISO 20417:2021, 3.2, modified — deleted note 4.]

143 3.2

144 clinical accuracy

145 closeness of agreement between the *output temperature* of a *clinical thermometer* and the reference value
146 of the temperature of the reference site that the *clinical thermometer* purports to represent

147 [SOURCE: ISO 80601-2-56:—, 201.3.202]

148 3.3

149 clinical bias

150 Δ_{cb}

151 mean difference between *output temperatures* of a *clinical thermometer* and a *reference thermometer* for
152 the intended *reference body site*

153 Note 1 to entry: The *measuring site* can be the same as or different from the *reference body site*.

154 3.4

155 clinical investigation

156 systematic investigation in one or more human subjects, undertaken to assess the clinical performance,
157 effectiveness or safety of a medical device

158 Note 1 to entry: For the purpose of this document, “clinical trial” or “clinical study” are synonymous with “clinical
159 investigation”.

160 [SOURCE: ISO 14155:—, 3.9]

161 3.5

162 clinical investigation report

163 document describing the design, conduct, statistical analysis and results of a *clinical investigation*

164 [SOURCE: ISO 14155:—, 3.11]