



DPC: 25/30512266 DC

BSI Group Headquarters

389 Chiswick High Road London W4 4AL

Tel: +44 (0)20 8996 9000

Fax: +44 (0)20 8996 7400

www.bsigroup.com

Date: 14 March 2025

Origin: International

Latest date for receipt of comments: 02 April 2025

Project No. 2025/01045

Responsible committee: EPL/66 Safety of measuring, control and laboratory equipment

Interested committees:

Title: Draft BS EN 61010-2-030 Ed.4.0 BS EN IEC 61010-2-030 Safety requirements for electrical equipment for measurement, control, and laboratory use. Part 2-030: Particular requirements for equipment having testing or measuring circuits.

Please notify admin.start@bsigroup.com if you are aware of any keywords that might assist in clarifying or identifying the standard or of the content of this standard:

- i) has any issues related to 3rd party IPR, patent or copyright
- ii) affects other national standard(s)
- iii) requires additional national guidance or information

WARNING: THIS IS A DRAFT AND MUST NOT BE REGARDED OR USED AS A BRITISH STANDARD. THIS DRAFT IS NOT CURRENT BEYOND 02 April 2025

This draft is issued to allow comments from interested parties; all comments will be given consideration prior to publication. No acknowledgement will normally be sent. **See overleaf for information on the submission of comments.**

No copying is allowed, in any form, without prior written permission from BSI except as permitted under the Copyright, Designs and Patent Act 1988 or for circulation within a nominating organization for briefing purposes. Electronic circulation is limited to dissemination by e-mail within such an organization by committee members.

Further copies of this draft may be purchased from BSI Shop <http://shop.bsigroup.com> or from BSI Customer Services, Tel: +44(0) 20 8996 9001 or email cservices@bsigroup.com. British, International and foreign standards are also available from BSI Customer Services.

Information on the co-operating organizations represented on the committees referenced above may be obtained from <https://standardsdevelopment.bsigroup.com/>.

Introduction

This draft standard is based on national and international discussions. Your comments on this draft are invited and will assist in the preparation of the consequent standard.

For international standards, comments will be reviewed by the relevant UK national committee before sending the consensus UK vote and comments to the international committee, which will then decide appropriate action. If the international standard is approved, it is usual for the text to be published as a British Standard.

For national standards, comments will be reviewed by the relevant UK national committee and the resulting standards published as a British Standard.

UK Vote

Please indicate whether you consider the UK should submit a negative (with supporting technical reasons) or positive vote on this draft. Please indicate if you are aware of any reason why this draft standard should not be published as a British Standard.

Submission of Comments

- **Annotated drafts are not acceptable and will be rejected.**
- All comments should be submitted online at <https://standardsdevelopment.bsigroup.com/>. You will need to register in order to comment.

Template for comments and secretariat observations

Date: xx/xx/20xx	Document: ISO/DIS xxxx
------------------	------------------------

1	2	(3)	4	5	(6)	7
MB	Clause No./ Subclause No./Annex (e.g. 3.1)	Paragraph/ Figure/Table/Note	Type of comment	Comment (justification for change) by the MB	Proposed change by the MB	Secretariat observations on each comment submitted
	3.1	Definition 1	ed	Definition is ambiguous and needs clarifying.	Amend to read '...so that the mains connector to which no connection...'	
	6.4	Paragraph 2	te	The use of the UV photometer as an alternative cannot be supported as serious problems have been encountered in its use in the UK.	Delete reference to UV photometer.	



66/841/CD

COMMITTEE DRAFT (CD)

PROJECT NUMBER: IEC 61010-2-030 ED4	
DATE OF CIRCULATION: 2025-03-07	CLOSING DATE FOR COMMENTS: 2025-05-02
SUPERSEDES DOCUMENTS: 66/837/RR	

IEC TC 66 : SAFETY OF MEASURING, CONTROL AND LABORATORY EQUIPMENT	
SECRETARIAT: United Kingdom	SECRETARY: Ms Stephanie Lavy
OF INTEREST TO THE FOLLOWING COMMITTEES: TC 65, TC 85	HORIZONTAL FUNCTION(S): TC 66 Horizontal Group Safety
ASPECTS CONCERNED: Safety	

This document is still under study and subject to change. It should not be used for reference purposes.

Recipients of this document are invited to submit, with their comments, notification of any relevant patent rights of which they are aware and to provide supporting documentation.

Recipients of this document are invited to submit, with their comments, notification of any local regulations or technical reasons that may exist and should be considered should this proposal proceed, recognizing that failure to address such requirements could result in the need for new or existing "In Some Countries" clauses. (SEE [AC/22/2007](#) OR [NEW GUIDANCE DOC](#)).

TITLE:

Safety requirements for electrical equipment for measurement, control, and laboratory use - Part 2-030: Particular requirements for equipment having testing or measuring circuits

NOTE FROM TC/SC OFFICERS:

Copyright © 2025 International Electrotechnical Commission, IEC. All rights reserved. It is permitted to download this electronic file, to make a copy and to print out the content for the sole purpose of preparing National Committee positions. You may not copy or "mirror" the file or printed version of the document, or any part of it, for any other purpose without permission in writing from IEC.

CONTENTS

1		
2	FOREWORD.....	4
3	INTRODUCTION.....	7
4	1 Scope and object.....	8
5	2 Normative references	9
6	3 Terms and definitions	10
7	4 Assessment, tests, and environmental conditions	10
8	5 Marking and documentation	11
9	6 Protection against electric shock	13
10	7 Protection against mechanical HAZARDS.....	20
11	8 Protection against the spread of fire	20
12	9 Protection against the spread of fire and arc flash	20
13	10 Equipment temperature limits and resistance to temperature	25
14	11 Protection against HAZARDS from fluids and solid foreign objects	25
15	12 Protection against injury from radiation, sonic and ultrasonic pressure	25
16	13 Protection against hazardous substances, explosion and implosion.....	25
17	14 Components and subassemblies	25
18	15 Interlocks and safety-related controls	26
19	16 Protection against HARM related to the use of the equipment	26
20	17 Assessment of RISKS and situations not fully covered by Clauses 6 to 16	28
21	101 MEASURING CIRCUITS	28
22	Annex K (normative) Insulation requirements not covered by 6.7	32
23	Annex L (informative) Index of defined terms	44
24	Annex AA (normative) MEASUREMENT CATEGORIES	45
25	Annex BB (informative) HAZARDS pertaining to measurements performed in certain	
26	environments.....	48
27	Annex CC (informative) 4 mm "banana" TERMINALS	51
28	Annex DD (informative) Flowchart for insulation according to the type of circuit.....	53
29	Annex EE (informative) Determination of CLEARANCES for Table 101	56
30	Bibliography.....	57
31		
32	Figure 101 – Duration of current flow versus body current for a.c. and d.c. currents.....	14
33	Figure K.101 – Test circuit for evaluation of TRANSIENT OVERVOLTAGE attenuation.....	38
34	Figure AA.1 – Example to identify the locations of MEASUREMENT CATEGORIES	46
35	Figure CC.1 – Recommended dimensions of 4 mm TERMINALS	52
36	Figure DD.1 – Requirements for CLEARANCE, CREEPAGE DISTANCE and solid insulation.....	55
37		
38	Table 101 – CLEARANCES for unmated measuring circuit TERMINALS.....	16
39	Table K.15 – CLEARANCE values for the calculation of K.3.2	34
40	Table K.16 – Test voltages based on CLEARANCES	35
41	Table K.101 – Impulse voltages for circuits connected to MAINS	38
42	Table K.102 – CLEARANCES for measuring circuits RATED for MEASUREMENT CATEGORIES.....	40

43	Table K.103 – Impulse test voltages for testing electric strength of solid insulation for	
44	measuring circuits RATED for MEASUREMENT CATEGORIES	41
45	Table K.104 – a.c. test voltages for testing electric strength of solid insulation for	
46	measuring circuits RATED for MEASUREMENT CATEGORIES	41
47	Table K.105 – Minimum values for distance or thickness of solid insulation for	
48	measuring circuits RATED for MEASUREMENT CATEGORIES	43
49	Table AA.1 – Characteristics of MEASUREMENT CATEGORIES	47
50	Table EE.1 – CLEARANCE values for Table 101	56
51		
52		

INTERNATIONAL ELECTROTECHNICAL COMMISSION

**SAFETY REQUIREMENTS FOR ELECTRICAL EQUIPMENT
FOR MEASUREMENT, CONTROL, AND LABORATORY USE –****Part 2-030: Particular requirements for equipment
having testing or measuring circuits**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) IEC draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). IEC takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, IEC had not received notice of (a) patent(s), which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at <https://patents.iec.ch>. IEC shall not be held responsible for identifying any or all such patent rights.

IEC 61010-2-030 has been prepared by IEC technical committee 66: Safety of measuring, control and laboratory equipment. It is an International Standard.

It has the status of a group safety publication in accordance with IEC Guide 104.

This fourth edition cancels and replaces the third edition published in 2023. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- a) IEC 61010-2-130 has been considered relevant to this document;

- b) added a definition for MEASURING CIRCUITS;
- c) impact of amendment 2 to IEC 61010-1:2010/AMD1/2016 has been taken into consideration:
- 1) Clause 2 and Bibliography have been updated;
 - 2) MAINS definition has been modified
 - 3) in 4.3.2.5, specific requirements for power supply have been deleted since Amendment 2 of IEC 61010 1 take them now into account;
 - 4) Figure 101 is replaced by Figure 3;
 - 5) modifications of 6.8.3.1 and 6.8.3.2 have been deleted since Amendment 2 of IEC 61010 1 take them now into account;
 - 6) conformity statements of K.3.2 and K.101.2 have been modified;
 - 7) a few conformity statements for constructional requirements of solid insulation have been modified;
- d) in 4.4.2.101, requirements for surge protective devices have been modified;
- e) requirements for markings of HAZARDOUS LIVE output TERMINALS have been added to a new 5.1.5.101.4 and the previous 5.1.5.101.4 has been renumbered 5.1.5.101.5; RATINGS for the output voltage have been added in 5.4.2;
- f) new requirements for equipment operation in 5.4.4;
- g) exception of 6.1.2 has been better framed;
- h) in 6.3.2 c), level for inductive stored energy has been added;
- i) 6.5.2.101 has been deleted;
- j) 6.5.5 has been modified to take into account the automatic disconnection of HAZARDOUS LIVE voltage source
- k) 6.5.6 has been modified to take into account voltage-limiting devices used in MEASURING CIRCUITS;
- l) in 6.6.101.1, a new requirement for TERMINALS is added;
- m) in 6.6.101.2, the conformity statement has been modified;
- n) in 6.8.1, a precision is given when testing solid insulation;
- o) a new 6.101 gives requirement for HAZARD indicator in case of hazardous voltage or energy;
- p) 9.101.1 and 9.101.2.1 have been redrafted for clarification, in particular to take into account the recommendations of new Annex N about REASONABLY FORESEEABLE MISUSE; an exception has been added for MEASURING CIRCUIT not RATED for MEASUREMENT CATEGORIES III or IV
- q) a new 14.9 with requirements for capacitors has been added;
- r) Clause 101 has been rephrased and completed with requirements for equipment injecting voltage and current;
- 1) 101.2 has been moved to 9.101.4 as this subclause is relevant for prevention of fire;
 - 2) 101.3 has been moved to 16.4 and 16.101.

The text of this International Standard is based on the following documents:

Draft	Report on voting
66/xxx/FDIS	66/xxx/RVD

Full information on the voting for its approval can be found in the report on voting indicated in the above table.

The language used for the development of this International Standard is English.

This document was drafted in accordance with ISO/IEC Directives, Part 2, and developed in accordance with ISO/IEC Directives, Part 1 and ISO/IEC Directives, IEC Supplement, available at www.iec.ch/members_experts/refdocs. The main document types developed by IEC are described in greater detail at www.iec.ch/publications.

A list of all parts of the IEC 61010 series, under the general title *Safety requirements for electrical equipment for measurement, control, and laboratory use*, can be found on the IEC website.

This document is to be used in conjunction with IEC 61010-1:2010, IEC 61010-1:2010/AMD1:2016 and IEC 61010-1:2010/AMD2:—¹.

This document supplements or modifies the corresponding clauses in IEC 61010-1 so as to convert that publication into the IEC standard: *Particular requirements for equipment having testing or measuring circuits*.

Where a particular subclause of IEC 61010-1 is not mentioned in this document, that subclause applies as far as is reasonable. Where this document states "addition", "modification", "replacement", or "deletion", the relevant requirement, test specification or note in IEC 61010-1 should be adapted accordingly.

In this standard:

- the following print types are used:
 - requirements: in roman type;
 - NOTES: in small roman type;
 - *conformity and tests: in italic type*;
 - terms used throughout this standard which have been defined in Clause 3: SMALL ROMAN CAPITALS;
- subclauses, figures, tables and notes which are additional to those in IEC 61010-1 are numbered starting from 101. Additional annexes are lettered starting from AA and additional list items are lettered from aa).

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under webstore.iec.ch in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

IMPORTANT – The "colour inside" logo on the cover page of this document indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.

¹ Second amendment under preparation. Stage at the time of publication: IEC CDV 61010-1/12:2025.

INTRODUCTION

IEC 61010-1 specifies the safety requirements that are generally applicable to all equipment within its scope. For certain types of equipment, the requirements of IEC 61010-1 and its amendments will be supplemented or modified by the special requirements of one or more standard from the IEC 61010-2 series which is/are read in conjunction with the requirements of IEC 61010-1.

1) This document specifies the safety requirements for equipment with testing or measuring circuits which are connected for test or measurement purposes to devices or circuits outside the measurement equipment itself.

2) IEC 61010-2-032 specifies the safety requirements for hand-held and hand-manipulated current sensors for measuring, detecting, injecting current, or indicating current waveforms on circuits without physically opening the current path of the circuit being measured, included clamp multimeters.

Most of the requirements of this document have been included in IEC 61010-2-032. Equipment within the scopes of both this document and IEC 61010-2-032 is considered to be covered by the requirements of IEC 61010-2-032.

However, for current sensors in combined equipment with PROTECTIVE BONDING and automatic disconnection of the supply, this document and IEC 61010-2-032 are read in conjunction.

3) IEC 61010-2-033 specifies the safety requirements for hand-held multimeters **and other meters** for domestic and professional use, capable of measuring mains voltage, intended to measure voltage and other electrical quantities such as resistance or current.

All relevant requirements of this document have been included in IEC 61010-2-033.

4) IEC 61010-2-034 specifies the safety requirements for measurement equipment for insulation resistance and test equipment for electric strength which are connected to units, lines or circuits for test or measurement purposes.

All relevant requirements of this document have been included in IEC 61010-2-034. However, for equipment within the scope of IEC 61010-2-032 and IEC 61010-2-034, these standards are read in conjunction.

IEC 61010-031 specifies the safety requirements for hand-held and hand-manipulated probe assemblies and their related accessories intended to be used in particular with equipment in the scope of this document, IEC 61010-2-032, IEC 61010-2-033 and IEC 61010-2-034. These probe assemblies are for non-contact or direct electrical connection between a part and electrical test and measurement equipment. They may be fixed to the equipment or be detachable accessories for the equipment.

IEC 61010-2-130 specifies particular safety requirements for equipment and accessories intended to be used in educational establishment by pupil OPERATORS. Pupil OPERATORS are for the sake of IEC 61010-2-130, children between the ages of 3 years and 16 years operating electrical equipment under supervision of a RESPONSIBLE BODY. For equipment used in educational establishments, the requirements of IEC 61010-2-130 will supplement or modify the requirements of this document.

SAFETY REQUIREMENTS FOR ELECTRICAL EQUIPMENT FOR MEASUREMENT, CONTROL, AND LABORATORY USE –

Part 2-030: Particular requirements for equipment having testing or measuring circuits

All Clauses and their subclauses, all Annexes, their Clauses and subclauses, and Bibliography of IEC 61010-1:2010+AMD1:2016+AMD2:—² apply except as follows.

1.1.1 Equipment included in scope

Replace the existing text with the following new text:

This document specifies safety requirements for equipment having testing or measuring circuits which are connected for test or measurement purposes to devices or circuits outside the measurement equipment itself.

These include testing or measuring circuits which are part of electrical test and measurement equipment, laboratory equipment, or process control equipment. These circuits in equipment have additional protective means between the circuit and an OPERATOR.

NOTE These testing and measuring circuits can, for example:

- measure voltages in circuits of other equipment,
- measure temperature of a separate device via a thermocouple,
- measure force on a separate device via a strain gauge,
- inject a voltage or current onto a circuit to analyse or test a new design.

This group safety publication focusing on safety essential requirements is primarily intended to be used as a product safety standard for the products mentioned in the scope, but is also intended to be used by technical committees in the preparation of publications for products similar to those mentioned in the scope of this document, in accordance with the principles laid down in IEC Guide 104 and ISO/IEC Guide 51.

One of the responsibilities of a technical committee is, wherever applicable, to make use of basic safety publications and/or group safety publications in the preparation of its publications.

1.2.1 Aspects included in scope

Replace item c) and item h) from the list of the second paragraph with the following new item c) and item h):

² Second amendment under preparation. Stage at the time of publication: IEC CDV 61010-1/12:2025.

- c) spread of fire or arc flash from the equipment (see Clause 9);
- h) HAZARDS related to the use and REASONABLY FORESEEABLE MISUSE of the equipment (see Clause 16 and Clause 101).

Insert the following new paragraph between the third paragraph and the Note:

Annex BB provides guidance to equipment manufacturers on HAZARDS that should be considered for equipment intended for performing tests and measurements on hazardous conductors, including MAINS conductors and telecommunication network conductors.

2 Normative references

Replace the following two existing normative references:

IEC 61672-1, *Electroacoustics – Sound level meters – Part 1: Specifications*

IEC 61672-2, *Electroacoustics – Sound level meters – Part 2: Pattern evaluation tests*

with the following new normative references:

IEC 61672-1:2013, *Electroacoustics – Sound level meters – Part 1: Specifications*

IEC 61672-2:2013, *Electroacoustics – Sound level meters – Part 2: Pattern evaluation tests*
IEC 61672-2:2013/AMD1:2017

Insert the following five new normative references in alphanumeric order:

IEC 60364-4-44:2007, *Low-voltage electrical installations – Part 4-44: Protection for safety – Protection against voltage disturbances and electromagnetic disturbances*

IEC 60364-4-44:2007/AMD1:2015

IEC 60364-4-44:2007/AMD2:2018

IEC 61000-4-5:2014, *Electromagnetic compatibility (EMC) – Part 4-5: Testing and measurement techniques – Surge immunity test*

IEC 61000-4-5:2014/AMD1:2017

IEC 61010-1:2010, *Safety requirements for electrical equipment for measurement, control, and laboratory use – Part 1: General requirements*

IEC 61010-1:2010/AMD1:2016

IEC 61010-1:2010/AMD2:—³

IEC 61010-2-032:—⁴, *Safety requirements for electrical equipment for measurement, control, and laboratory use – Part 2-032: Particular requirements for hand-held and hand-manipulated current sensors for electrical test and measurement*

³ Second amendment under preparation. Stage at the time of publication: IEC CDV 61010-1/A2:2025.

⁴ Sixth edition under preparation. Stage at the time of publication: IEC CD 61010-2-032:2025

297 IEC 61010-031:2022, *Safety requirements for electrical equipment for measurement, control and*
298 *laboratory use – Part 031: Safety requirements for hand-held probe assemblies for electrical*
299 *measurement and test*

300

301 **3.5 Safety terms**

302 *Replace the definition of 3.5.6 with the following new definition:*

303 **3.5.6**

304 **MAINS**

305 a.c or d.c. electrical power distribution system (external to the equipment)

306 NOTE 1 to entry: MAINS include public or private electrical utilities and, unless otherwise specified in this document,
307 equivalent sources such as motor-driven generators and uninterruptible power supplies.

308 NOTE 2 to entry: This definition differs from the IEC 61010-1/A2 definition by allowing voltages greater than 1 000 V
309 a.c. and 1 500 V d.c. and can be used to power the equipment or be measured.

310 [IEC 60664-1:2020, 3.1.2 (mains supply) modified]

311

312 *Insert the following two new terms and definitions between the existing terms 3.5.7 and 3.5.8:*

313 **3.5.101**

314 **MEASURING CIRCUIT**

315 testing or measuring circuit (internal to the equipment) which is connected for test or
316 measurement purposes to devices or circuits outside the equipment itself

317 **3.5.102**

318 **MEASUREMENT CATEGORY**

319 classification of MEASURING CIRCUITS according to the type of MAINS to which they are intended
320 to be connected

321

322 **4.3.2.6 Input and output voltages**

323 *Replace the existing title and text of 4.3.2.6 with the following new title and text:*

324 **4.3.2.6 Input and output voltages or currents**

325 Input and output voltages or currents, including floating voltages but excluding the electrical
326 supply voltage (see 4.3.2.5), shall be set to any voltage or current within their RATED range, in
327 normal and reverse polarity if possible.

328

329 **4.4.2.1 General**

330 *Replace the item 2) of the first paragraph with the following new item 2):*

331 2) Applicable fault conditions specified in 4.4.2.2 to 4.4.2.15 and 4.4.2.101.

332

Add the following new subclause after 4.4.2.15:

4.4.2.101 Surge protective devices

Surge protective devices used in MAINS CIRCUITS or MEASURING CIRCUITS connected to MAINS shall be short-circuited and open-circuited, only one at a time in turn in any convenient order.

Add the following new subclause after 5.1.5.2:

5.1.5.101 MEASURING CIRCUIT TERMINALS

5.1.5.101.1 General

Some MEASURING CIRCUIT TERMINALS for the equipment within the scope of this document also serve as output TERMINALS.

Except as permitted in 5.1.5.105:

- a) the value of the nominal a.c. r.m.s. line-to-neutral or d.c. voltage of MAINS being measured shall be marked for MEASURING CIRCUIT TERMINALS RATED for MEASUREMENT CATEGORIES, or the value of the RATED voltage to earth for other MEASURING CIRCUIT TERMINALS, and

NOTE CLEARANCES and solid insulation for MEASUREMENT CATEGORIES are specified for a nominal a.c. r.m.s. line-to-neutral or d.c. voltage of MAINS being measured. Neutral is considered to be earthed (see Annex I).

- a) the value of the RATED voltage or the RATED current, as applicable, of each pair or set of MEASURING CIRCUIT TERMINALS that are intended to be used together shall be marked, and
- b) the pertinent MEASUREMENT CATEGORY for each individual, pair, or set of MEASURING CIRCUIT TERMINALS, or symbol 14 of Table 1 shall be marked as specified in 5.1.5.101.2 and 5.1.5.101.3, if applicable.

MEASURING CIRCUIT TERMINALS are usually arranged in pairs or sets. Each pair or set of TERMINALS may have a RATED voltage or a RATED current, or both, within that set, and each individual TERMINAL may have a RATED voltage to earth. For some equipment, the RATED voltage between TERMINALS may be different from the RATED voltage to earth. Markings shall be clear to avoid misunderstanding.

Symbol 14 of Table 1 shall be marked if current MEASURING CIRCUIT TERMINALS are not intended for connection to the secondary of current transformers without internal protection (see 9.101.4.2).

Markings shall be placed adjacent to the TERMINALS. However, if there is insufficient space (as in multi-input equipment), the marking may be on the RATING plate or scale plate, or the TERMINAL may be marked with symbol 14 of Table 1.

For any set of MEASURING CIRCUIT TERMINALS, symbol 14 of Table 1 does not need to be marked more than once, if it is close to the TERMINALS.

Conformity is checked by inspection and, if applicable, as specified in 5.1.5.101.2, 5.1.5.101.3 and 5.1.5.101.4, taking the exceptions in 5.1.5.101.5 into account.

5.1.5.101.2 MEASURING CIRCUIT TERMINALS RATED for MEASUREMENT CATEGORIES

The relevant MEASUREMENT CATEGORY shall be marked for TERMINALS of MEASURING CIRCUITS RATED for MEASUREMENT CATEGORIES. The MEASUREMENT CATEGORY markings shall be "CAT II", "CAT III" or "CAT IV" as applicable.

373 The RATED voltage of the TERMINALS of a MEASURING CIRCUIT intended for MAINS voltage
 374 measurements shall be equal to or higher than their RATED a.c. r.m.s. line-to-neutral or d.c.
 375 voltage.

376 Marking those TERMINALS with more than one type of MEASUREMENT CATEGORY and its RATED
 377 voltage is permissible.

378 *Conformity is checked by inspection.*

379 **5.1.5.101.3 MEASURING CIRCUIT TERMINALS RATED for connection to voltages above the**
 380 **levels of 6.3.1**

381 Symbol 14 of Table 1 shall be marked adjacent to the TERMINALS for MEASURING CIRCUIT
 382 TERMINALS RATED for connection to voltages above the levels of 6.3.1, but that are not RATED for
 383 MEASUREMENT CATEGORIES (see also 5.4.1 b b)).

384 *Conformity is checked by inspection.*

385 **5.1.5.101.4 HAZARDOUS LIVE output TERMINALS**

386 Output TERMINALS which can be HAZARDOUS LIVE shall be marked with symbol 12 of Table 1 in
 387 close proximity to those TERMINALS.

388 *Conformity is checked by inspection.*

389 **5.1.5.101.5 MEASURING CIRCUIT TERMINALS which are permanently connected, dedicated,**
 390 **or for non-HAZARDOUS LIVE voltages**

391 MEASURING CIRCUIT TERMINALS do not need to be marked if:

- 392 a) they are intended to be permanently connected and not ACCESSIBLE (see 5.4.3 aa) and bb)),
 393 or
- 394 b) they are dedicated only for connection to specific TERMINALS of other equipment, or
- 395 c) it is obvious from other indications that the RATED voltage does not exceed the levels of
 396 6.3.1.

397 NOTE Examples of acceptable indications that the RATED voltage of the inputs are intended to not exceed the
 398 levels of 6.3.1 include:

- 399 – the full scale deflection marking of a single-range indicating voltmeter or ammeter or maximum marking of a
 400 multi-range multimeter;
- 401 – the maximum range marking of a voltage selector switch;
- 402 – a marked voltage or power RATING expressed in dB, mW or W, where the equivalent value, as explained in
 403 the documentation, does not exceed 30 V a.c.

404 *Conformity is checked by inspection.*

405

406 **5.4.1 GENERAL**

407 *Add the following two new items to the list of the first paragraph after item k):*

- 408 aa) information about each relevant MEASUREMENT CATEGORY if the MEASURING CIRCUIT is
 409 RATED for MEASUREMENT CATEGORIES (see 5.1.5.101.2);
- 410 bb) for MEASURING CIRCUITS that are not RATED for MEASUREMENT CATEGORIES, but that could
 411 be misused by connection to such circuits, a warning not to use the equipment for
 412 measurements on MAINS, and a detailed RATING including TRANSIENT OVERVOLTAGES (see
 413 AA.2.4 for more information).

414

415 *Add a new paragraph between the first and the second paragraph:*

416 Some equipment may have multiple MEASUREMENT CATEGORY RATINGS for the same MEASURING
417 CIRCUIT. For such equipment, the documentation shall clearly identify the MEASUREMENT
418 CATEGORIES where the equipment is intended to be used and where it shall not be used.

419

420 **5.4.2 Equipment RATINGS**

421 *Add the following new item to the list of the first paragraph after item f):*

422 aa) when relevant, the output voltage or voltage range, frequency and current RATING.

423

424 **5.4.3 Equipment installation**

425 *Add the following two new items to the list of the first paragraph after item i):*

426 aa) for MEASURING CIRCUIT TERMINALS intended for permanent connection and that are RATED
427 for MEASUREMENT CATEGORIES, information regarding the MEASUREMENT CATEGORY,
428 RATED voltages or RATED currents as applicable (see 5.1.5.101.2);

429 bb) for MEASURING CIRCUIT TERMINALS intended for permanent connection and that are not
430 RATED for MEASUREMENT CATEGORIES, information regarding the RATED voltages, RATED
431 currents, and RATED TRANSIENT OVERVOLTAGES as applicable (see 5.1.5.101.5).

432

433 **5.4.4 Equipment operation**

434 *Add the following two new item to the list of the first paragraph after item k):*

435 aa) instructions to warn the OPERATOR that a HAZARDOUS LIVE voltage can be present on
436 locking or screw-held type TERMINALS during NORMAL USE if it is not feasible for operating
437 reasons to prevent it (see 6.1.2 aa));

438 bb) when performing an d.c. current measurement in series with an inductive circuit,
439 instructions to warn the OPERATOR that a HAZARDOUS LIVE voltage can be present if the
440 circuit is suddenly opened (a probe falls off for example) (see 101 d));

441 cc) instructions for a daily or routine check to ensure the correct functionality of the equipment
442 before use when one HAZARD indicator has been considered to be sufficient (see 6.101).

443

444 **6.1.2 Exceptions**

445 *Add the following new item to the list after item b):*

446 aa) locking or screw-held type MEASURING CIRCUITS TERMINALS, including TERMINALS which
447 do not require the use of a TOOL (see 5.4.4 aa)). Symbol 12 of Table 1 shall be marked
448 adjacent to the TERMINALS. Clause 17 applies to this situation.

449

450 **6.2.1 General**

451 *Insert the following NOTE 101 after the third paragraph:*

NOTE 101 For determination of ACCESSIBLE parts of equipment intended to be used in educational establishment by pupil OPERATORS, see IEC 61010-2-130.

6.3.2 Levels in SINGLE FAULT CONDITION

Replace the existing item c) with the following new item c):

c) The levels of:

- 1) 1,6 J for inductive or capacitive stored energy, except
- 2) the capacitance from line B of Figure 3 for voltages up to 40 kV peak or DC.

6.5.1 General

Replace the existing item c) with the following new item c):

c) automatic disconnection of HAZARDOUS LIVE voltage source (see 6.5.5);

Replace in Figure 4, the text in the cell referenced 6.5.1 c) by the text "Automatic disconnection of HAZARDOUS LIVE voltage source"

6.5.2.3 PROTECTIVE CONDUCTOR TERMINAL

Replace the existing item h) 2) with the following new item h) 2):

h) 2) the PROTECTIVE BONDING shall not be interrupted by any switching or interrupting device. Devices used for indirect bonding in MEASURING CIRCUITS (see 6.5.5 and 6.5.6) are permitted to be part of the PROTECTIVE BONDING.

6.5.5 Automatic disconnection of the supply

Replace the existing title of 6.5.5 with the following new title:

6.5.5 Automatic disconnection of HAZARDOUS LIVE voltage source

Add the following five paragraphs after the last paragraph:

For MEASURING CIRCUITS, voltage-sensitive tripping devices can be used to establish indirect bonding between the PROTECTIVE CONDUCTOR TERMINAL and ACCESSIBLE conductive parts if these become HAZARDOUS LIVE as a result of a fault.

Voltage-sensitive tripping devices interrupt the HAZARDOUS LIVE voltage source, or connect the ACCESSIBLE conductive parts to the PROTECTIVE CONDUCTOR TERMINAL whenever the voltage across them reaches the relevant levels of 6.3.2 a). The tripping duration versus the current shall not exceed the levels of Figure 2.

Conformity is checked by applying successively the relevant voltage level of 6.3.2 a) and the maximum RATED voltage of the MEASURING CIRCUIT between the ACCESSIBLE conductive parts

and the PROTECTIVE CONDUCTOR TERMINAL. The current between the ACCESSIBLE conductive parts and the PROTECTIVE CONDUCTOR TERMINAL is measured with the circuit of Figure A.1. During and after the test, no HAZARD shall arise, nor shall there be any evidence of fire, arcing, explosion, or damage to voltage-sensitive tripping devices, including traces on the printed wiring board.

Voltage-sensitive tripping devices used to establish indirect bonding shall have at least the voltage and current RATINGS of the MEASURING CIRCUIT TERMINALS.

Conformity is checked by inspection.

6.5.6 Current or voltage-limiting device

Add the following four new paragraphs after the last paragraph:

For MEASURING CIRCUITS, voltage-limiting devices can be used to establish indirect bonding between the PROTECTIVE CONDUCTOR TERMINAL and ACCESSIBLE conductive parts if these become HAZARDOUS LIVE as a result of a fault.

Voltage-limiting devices which become conductive when the voltage across them exceeds the relevant levels of 6.3.2 a) shall have overcurrent protection to prevent damage of the device.

Voltage-limiting devices are not suitable in the case of voltage applied from external sources to the accessible conductive parts.

Conformity is checked by introducing a single fault within the equipment or from external connection while the equipment is operated in NORMAL USE such that the maximum hazardous voltage presents itself to the ACCESSIBLE conductive parts protected by indirect bonding. The current between the ACCESSIBLE conductive parts and the PROTECTIVE CONDUCTOR TERMINAL is measured with the circuit of Figure A.1. The duration of current flow versus the body current shall not exceed the levels of Figure 2. During and after the test, no HAZARD shall arise, nor shall there be any evidence of fire, arcing, explosion, or damage to overcurrent protection devices, except for fuses which can open.

Add the following new subclause and new Table 101 after 6.6.4:

6.6.101 MEASURING CIRCUIT TERMINALS

6.6.101.1 General

HAZARDOUS LIVE TERMINALS shall be anchored, fitted or designed so that internal conductors and TERMINAL mechanical mountings will not work loose when the TERMINALS are tightened or loosened, or when connections are made.

When determining the values of CREEPAGE DISTANCES for MEASURING CIRCUIT TERMINALS of HAND-HELD EQUIPMENT intended to be connected only to a hand-held probe assembly complying with IEC 61010-031 or to a current sensor complying with IEC 61010-2-032, the applicable values of CREEPAGE DISTANCES from material group I are allowed to be applied to all material groups.

Requirements for MEASURING CIRCUIT TERMINALS in unmated position, partially mated or mated position are defined respectively in 6.6.101.2, 6.6.101.3 and 6.6.101.4. Requirements for specialized MEASURING CIRCUIT TERMINALS are defined in 6.6.101.5.

Annex CC provides information regarding the recommended dimensions of 4 mm "banana" TERMINALS.

6.6.101.2 MEASURING CIRCUIT TERMINALS in unmated position

The following requirements apply to MEASURING CIRCUITS TERMINALS in unmated position when RATED voltages are applied to any other TERMINALS of the equipment.

- 1) ACCESSIBLE parts of locking-type or screw-held-type TERMINALS in unmated position including TERMINALS which do not require the use of a TOOL for unlocking or unscrewing shall not be HAZARDOUS LIVE.

NOTE See also the exception of 6.1.2 aa) which permits these parts to be ACCESSIBLE to an OPERATOR during NORMAL USE while they are HAZARDOUS LIVE only if it is not feasible for operating reasons to prevent it.

Conformity is checked by inspection.

- 2) ACCESSIBLE parts of other unmated TERMINALS shall be insulated from HAZARDOUS LIVE parts by PROTECTIVE IMPEDANCE or CLEARANCES and CREEPAGE DISTANCES meeting the requirements of 2 a) and 2 b) of this subclause 6.6.101.2.

- a) For TERMINALS with a voltage RATING up to 3 000 V, the CLEARANCES shall be at least the applicable values of Table 101.

Table 101 – CLEARANCES for unmated MEASURING CIRCUIT TERMINALS

Maximum voltage applied to the conductive parts of the TERMINAL V	CLEARANCE	
	a.c. r.m.s. mm	d.c. mm
600	0,8	0,8
1 000	1,0	0,8
1 500	2,0	1,1
2 000	3,2	1,8
3 000	6,4	3,5
For maximum voltages above 30 V a.c. r.m.s. or 60 V d.c. up to 600 V, CLEARANCES are 0,8 mm. Linear interpolation is allowed above 600 V. NOTE See Annex EE for determination of the CLEARANCES.		

For TERMINALS with a voltage RATING above 3 000 V, the CLEARANCE shall be at least the D_2 value of Table K.15 with U_m equal to 1,25 times the peak value of the voltage (see K.3.2).

If the equipment is RATED to operate at an altitude greater than 2 000 m, the value of the CLEARANCE shall be multiplied by the applicable factor of Table 3.

Conformity is checked by inspection and measurement of CLEARANCE from the closest approach of the test finger touching the external parts of the TERMINAL in the least favourable position (see Figure 1).

If the CLEARANCES are less than the minimum required, they shall be verified by the a.c. voltage test.

Conformity is checked by the a.c. voltage test of 6.8.3.1 with a duration of at least 3 cycles using the applicable test voltage for impulse of Table K.16 for the required CLEARANCE.

Correction factors of Table 10 are applicable to the values of test voltages for CLEARANCES given in Table K.16.

- b) The CREEPAGE DISTANCE values shall be at least the applicable CLEARANCE values defined in 2 a) of this subclause 6.6.101.2.