

Chemical indicators— Guidance for the selection, use, and interpretation of results

Chemical indicators—Guidance for the selection, use, and interpretation of results

Approved 18 October 1999

Abstract: This document provides guidance for the selection and use of chemical indicators that are intended for use with sterilization processes employing steam, ethylene oxide, irradiation, or dry heat. This technical information report is based on the ANSI/AAMI ST60:1996 standard for chemical indicators.

Keywords: chemical indicator, CIER vessel, resistometer, steam, ethylene oxide, irradiation, dry heat

Published by the

Association for the Advancement of Medical Instrumentation
1110 N. Glebe Road, Suite 220
Arlington, VA 22201-5762

© 2000 by the Association for the Advancement of Medical Instrumentation

All Rights Reserved

Publication, reproduction, photocopying, storage, or transmission, electronically or otherwise, of all or any part of this document without the prior written permission of the Association for the Advancement of Medical Instrumentation is strictly prohibited by law. It is illegal under federal law (17 U.S.C. § 101, *et seq.*) to make copies of all or any part of this document (whether internally or externally) without the prior written permission of the Association for the Advancement of Medical Instrumentation. Violators risk legal action, including civil and criminal penalties, and damages of \$100,000 per offense. For permission regarding the use of all or any part of this document, contact AAMI, 1110 N. Glebe Road, Suite 220, Arlington, VA 22201-5762. Phone: (703) 525-4890; Fax: (703) 525-1067.

Printed in the United States of America

ISBN 1-57020-127-7

AAMI TECHNICAL INFORMATION REPORT

A Technical Information Report (TIR) is a publication of the Association for the Advancement of Medical Instrumentation (AAMI) Standards Board that addresses a particular aspect of medical technology.

Although the material presented in a TIR may need further evaluation by experts, releasing the information is valuable because the industry and the professions have an immediate need for it.

A TIR differs markedly from a standard or recommended practice, and readers should understand the differences among these documents.

Standards and recommended practices are subject to a formal process of committee approval, public review, and resolution of all comments. This process of consensus is supervised by the AAMI Standards Board and, in the case of American National Standards, the American National Standards Institute.

A TIR is not subject to the same formal approval process as a standard. However, a technical committee and the AAMI Standards Board do approve a TIR for distribution.

Another difference is that although both standards and TIRs are periodically reviewed, a standard must be acted upon—either reaffirmed, revised, or withdrawn—and the action must be formally approved usually every 5 years but at least every 10 years. For a TIR, AAMI consults with a technical committee about every 5 years after the publication date (and periodically thereafter) for guidance on whether the document is still useful—that is, to check that the information is relevant or of historical value. If the information is not useful, the TIR is removed from circulation.

A TIR may be developed because it is more responsive to underlying safety or performance issues than a standard or recommended practice or because achieving consensus is extremely difficult or unlikely. Unlike a standard, a TIR permits the inclusion of differing viewpoints on technical issues.

CAUTION NOTICE: This AAMI technical information report may be revised or withdrawn at any time. Because it addresses a rapidly evolving field or technology, readers are cautioned to ensure that they have also considered information that may be more recent than this document.

Contents

	Page
1 Scope.....	1
2 Cited references	1
3 Definitions	2
4 General considerations.....	2
5 Classes of chemical indicators	3
5.1 General	3
5.2 Class 1—Process indicators.....	4
5.3 Class 2—Indicators for use in specific tests	4
5.4 Class 3—Single-parameter indicators	5
5.5 Class 4—Multiparameter indicators.....	6
5.6 Class 5—Integrating indicators.....	8
6 Role of chemical indicators.....	8
6.1 General	8
6.2 Design considerations (by class)	9
6.2.1 Process indicators	9
6.2.2 Single-parameter indicators.....	10
6.2.3 Multiparameter indicators	10
6.2.4 Integrating indicators	10
6.3 Chemical indicators performance evaluation	10
7 Selection and use of chemical indicators.....	11
7.1 Selecting chemical indicators.....	11
7.1.1 General	11
7.1.2 External chemical indicators	12
7.1.3 Internal chemical indicators	12
7.2 Retrieval and interpretation	12
7.3 Chemical indicators showing a “Fail” response.....	12
7.4 Questions to ask the manufacturer about chemical indicator performance characteristics	12
7.5 Interpretation of chemical indicator responses	13
7.6 Use of chemical indicators in sterility assurance procedures	13
8 Labeling of chemical indicators.....	14
9 Personnel training.....	14
10 Storage and handling.....	14
11 Recordkeeping.....	14
Annexes	
A Labeling requirements—ANSI/AAMI ST60, subclause 5.6	15
B Chemical indicators for use in new sterilization processes.....	16

Tables

1	Critical parameters for each sterilization method.....	3
2	Tolerances and limiting values for the response to critical parameters for Class 3 indicators	5
3	Tolerances and limiting values for the response to critical parameters for Class 4 indicators	6
4	Tolerances and limiting values for the response to critical parameters for Class 5 indicators	8

Figure

1	Graph of a standard sterilization cycle compared to that of a resistometer vessel.....	10
---	--	----

Committee representation

Association for the Advancement of Medical Instrumentation Sterilization Standards Committee

This Technical Information Report was developed and balloted by the AAMI Chemical Indicators Working Group under the auspices of the AAMI Sterilization Standards Committee. Approval of this Technical Information Report does not necessarily imply that all working group members voted affirmative.

At the time this document was published, the **AAMI Sterilization Standards Committee** had the following members:

<i>Cochairs:</i>	Virginia C. Chamberlain, PhD William E. Young
<i>Members:</i>	Zoe Z. Aler, RN, Timonium, MD Trabue D. Bryans, Viromed Biosafety Labs Virginia C. Chamberlain, PhD, Hendersonville, NC Anne M. Cofield, CRCST, International Association of Healthcare Central Service Materiel Management Neal E. Danielson, Wichita, KS Dorothy M. Fogg, RN, BSN, MA, Association of Perioperative Registered Nurses Lisa Foster, Ion Beam Applications James M. Gibson, Jr., JM Gibson Associates Barbara J. Goodman, RN, BS, CNOR, Rising Sun, MD Joel R. Gorski, PhD, NAMSA Susan Hadfield, Standards Council of Canada Victoria Hitchins, PhD, U.S. Food and Drug Administration Gretchen Keenan, 3M Health Care Sue Kuhnert, STS duoTEK Byron J. Lambert, PhD, Guidant Corporation Paul S. Malchesky, STERIS Corporation Patrick McCormick, PhD, Bausch & Lomb, Inc. Robert F. Morrissey, PhD, Johnson and Johnson S. Richard Nusbaum, Pennsylvania Engineering Co. David Orton, CR Bard Barry F.J. Page, Garner, NC Michael H. Scholla, MS, PhD, Dupont Medical Packaging Systems Janet K. Schultz, MSN, RN, Roswell, GA Harry L. Shaffer, Titan Corporation Robert J. Sharbaugh, PhD, CIC, Association for Professionals in Infection Control and Epidemiology Frank Sizemore, American Society for Healthcare Central Service Professionals James L. Whitby, MA, MB, FRCP, London, ON Thelma Wilcott, Becton Dickinson & Company Steve C. Yeadon, BS, Alcon Labs William E. Young, Baxter Healthcare Corporation
<i>Alternates:</i>	Bettye Beebe, Alcon Labs Carl W. Bruch, PhD, Pennsylvania Engineering Co. Louis M. Glasgow, Bausch & Lomb, Inc. Joyce M. Hansen, Baxter Healthcare Corporation Lois A. Jones, Becton Dickinson & Company Susan G. Klacik, AS, BS, International Association of Healthcare Central Service Materiel Managers Sandra A. Lee, RN, STERIS Corporation Chiu Lin, PhD, U.S. Food and Drug Administration Bruce Schullo, Ion Beam Applications Phil M. Schneider, 3M Health Care James Whitbourne, STS duoTEK

At the time this document was published, the **AAMI Chemical Indicators Working Group** had the following members:

Cochairs: Joel R. Gorski, PhD
Marvin L. Hart

Members: Richard B. Barrett, PhD, Tempil, Inc.
Susanna F. Barrett, U.S. Food and Drug Administration, Center for Devices and Radiological Health
Carl W. Bruch, PhD, Consultant, Hudson, WI
Martin S. Favero, PhD, Advanced Sterilization Products
Christopher Fricker, PhD, STERIS Corporation
Joel R. Gorski, PhD, NAMSA
Charles Hughes, SPS Medical
Jim Kaiser, Getinge/Castle, Inc.
Marvin L. Hart, Marvin L. Hart Associates, Inc.
Gretchen Keenan, 3M Health Care
Colleen P. Landers, RN, Timmins and District Hospital, Timmins, Ontario, Canada
Patrick McCormick, PhD, Bausch & Lomb, Inc.
Frank E. Platko, PhD, Propper Manufacturing Company, Inc.
Rose Marie Proietti, RN, MBA, Medical Action Industries, Inc.
Marimargaret Reichert, RN, MA, Retired, Olmsted Falls, OH
Jonathan Wilder, H&W Technology, LLC

Alternates: Karla Byrne, Getinge/Castle, Inc.
Catherine J. Finocchario, Bausch & Lomb, Inc.
Steve Kirckof, D.Eng., 3M Health Care
Paul S. Malchesky, D.Eng., STERIS Corporation
Candace McManus, PhD, U.S. Food and Drug Administration, Center for Devices and Radiological Health, Office of Compliance
Shaundrea L. Rechsteiner, NAMSA
Gary Socola, SPS Medical

NOTE—Participation by federal agency representatives in the development of this Technical Information Report does not constitute endorsement by the federal government or any of its agencies.
