American National Standard

ANSI/AAMI/ISO 15223-2:2010

Medical devices — Symbols to be used with medical device labels, labeling, and information to be supplied — Part 2: Symbol development, selection and validation

Association for the Advancement of Medical Instrumentation

Documents
Document Sharing Hub

AAMI
Single user license only. Copying, networking, and distribution prohibited.
Full document details: ANSI/AAMI/ISO 15223-2:2010, Medical devices — Symbols to be used with medical device labels, labeling, and information to be supp...

Objectives and uses of AAMI standards and recommended practices

It is most important that the objectives and potential uses of an AAMI product standard or recommended practice are clearly understood. The objectives of AAMI's technical development program derive from AAMI's overall mission: the advancement of medical instrumentation. Essential to such advancement are (1) a continued increase in the safe and effective application of current technologies to patient care, and (2) the encouragement of new technologies. It is AAMI's view that standards and recommended practices can contribute significantly to the advancement of medical instrumentation, provided that they are drafted with attention to these objectives and provided that arbitrary and restrictive uses are avoided.

A voluntary *standard* for a *medical device* recommends to the manufacturer the information that should be provided with or on the product, basic safety and performance criteria that should be considered in qualifying the device for clinical use, and the measurement techniques that can be used to determine whether the device conforms with the safety and performance criteria and/or to compare the performance characteristics of different products. Some standards emphasize the information that should be provided with the device, including performance characteristics, instructions for use, warnings and precautions, and other data considered important in ensuring the safe and effective use of the device in the clinical environment. Recommending the disclosure of performance characteristics often necessitates the development of specialized test methods to facilitate uniformity in reporting; reaching consensus on these tests can represent a considerable part of committee work. When a drafting committee determines that clinical concerns warrant the establishment of *minimum* safety and performance criteria, referee tests must be provided and the reasons for establishing the criteria must be documented in the rationale.

A *recommended practice* provides guidelines for the use, care, and/or processing of a medical device or system. A recommended practice does not address device performance *per se*, but rather procedures and practices that will help ensure that a device is used safely and effectively and that its performance will be maintained.

Although a device standard is primarily directed to the manufacturer, it may also be of value to the potential purchaser or user of the device as a frame of reference for device evaluation. Similarly, even though a recommended practice is usually oriented towards healthcare professionals, it may be useful to the manufacturer in better understanding the environment in which a medical device will be used. Also, some recommended practices, while not addressing device performance criteria, provide guidelines to industrial personnel on such subjects as sterilization processing, methods of collecting data to establish safety and efficacy, human engineering, and other processing or evaluation techniques; such guidelines may be useful to health care professionals in understanding industrial practices.

In determining whether an AAMI standard or recommended practice is relevant to the specific needs of a potential user of the document, several important concepts must be recognized:

All AAMI standards and recommended practices are *voluntary* (unless, of course, they are adopted by government regulatory or procurement authorities). The application of a standard or recommended practice is solely within the discretion and professional judgment of the user of the document.

Each AAMI standard or recommended practice reflects the collective expertise of a committee of health care professionals and industrial representatives, whose work has been reviewed nationally (and sometimes internationally). As such, the consensus recommendations embodied in a standard or recommended practice are intended to respond to clinical needs and, ultimately, to help ensure patient safety. A standard or recommended practice is limited, however, in the sense that it responds generally to perceived risks and conditions that may not always be relevant to specific situations. A standard or recommended practice is an important *reference* in responsible decision-making, but it should never *replace* responsible decision-making.

Despite periodic review and revision (at least once every five years), a standard or recommended practice is necessarily a static document applied to a dynamic technology. Therefore, a standards user must carefully review the reasons why the document was initially developed and the specific rationale for each of its provisions. This review will reveal whether the document remains relevant to the specific needs of the user.

Particular care should be taken in applying a product standard to existing devices and equipment, and in applying a recommended practice to current procedures and practices. While observed or potential risks with existing equipment typically form the basis for the safety and performance criteria defined in a standard, professional judgment must be used in applying these criteria to existing equipment. No single source of information will serve to identify a particular product as "unsafe". A voluntary standard can be used as one resource, but the ultimate decision as to product safety and efficacy must take into account the specifics of its utilization and, of course, cost-benefit considerations. Similarly, a recommended practice should be analyzed in the context of the specific needs and resources of the individual institution or firm. Again, the rationale accompanying each AAMI standard and recommended practice is an excellent guide to the reasoning and data underlying its provision.

In summary, a standard or recommended practice is truly useful only when it is used in conjunction with other sources of information and policy guidance and in the context of professional experience and judgment.

INTERPRETATIONS OF AAMI STANDARDS AND RECOMMENDED PRACTICES

Requests for interpretations of AAMI standards and recommended practices must be made in writing, to the AAMI Vice President, Standards Policy and Programs. An official interpretation must be approved by letter ballot of the originating committee and subsequently reviewed and approved by the AAMI Standards Board. The interpretation will become official and representation of the Association only upon exhaustion of any appeals and upon publication of notice of interpretation in the "Standards Monitor" section of the *AAMI News*. The Association for the Advancement of Medical Instrumentation disclaims responsibility for any characterization or explanation of a standard or recommended practice which has not been developed and communicated in accordance with this procedure and which is not published, by appropriate notice, as an *official interpretation* in the *AAMI News*.

AVA Documents
Document Sharing Hub
AAMI
Single user license only. Copying, networking, and distribution prohibited.
Full document details: ANSI/AAMI/ISO 15223-2:2010, Medical devices — Symbols to be used with medical device labels, labeling, and information to be supp...

Medical devices — Symbols to be used with medical device labels, labeling, and information to be supplied — Part 2: Symbol development, selection and validation

Approved 8 March 2010 by
Association for the Advancement of Medical Instrumentation

Approved 20 April 2010 by
American National Standards Institute, Inc.

Abstract: Specifies a process for developing, registering, and validating symbols for use in the labeling of medical devices.

Keywords: medical device symbols, validation

Documents
Document Sharing Hub

AAMI
Single user license only. Copying, networking, and distribution prohibited.
Full document details: ANSI/AAMI/ISO 15223-2:2010, Medical devices — Symbols to be used with medical device labels, labeling, and information to be supp...

AAMI Standard

This Association for the Advancement of Medical Instrumentation (AAMI) standard implies a consensus of those substantially concerned with its scope and provisions. The existence of an AAMI standard does not in any respect preclude anyone, whether they have approved the standard or not, from manufacturing, marketing, purchasing, or using products, processes, or procedures not conforming to the standard. AAMI standards are subject to periodic review, and users are cautioned to obtain the latest editions.

CAUTION NOTICE: This AAMI standard may be revised or withdrawn at any time. AAMI procedures require that action be taken to reaffirm, revise, or withdraw this standard no later than five years from the date of publication. Interested parties may obtain current information on all AAMI standards by calling or writing AAMI, or by visiting the AAMI website at www.aami.org.

All AAMI standards, recommended practices, technical information reports, and other types of technical documents developed by AAMI are *voluntary*, and their application is solely within the discretion and professional judgment of the user of the document. Occasionally, voluntary technical documents are adopted by government regulatory agencies or procurement authorities, in which case the adopting agency is responsible for enforcement of its rules and regulations.

Published by

Association for the Advancement of Medical Instrumentation
4301 N. Fairfax Drive, Suite 301
Arlington, VA 22203-1633
www.aami.org

© 2010 by the Association for the Advancement of Medical Instrumentation

All Rights Reserved

This publication is subject to copyright claims of ISO, ANSI, and AAMI. No part of this publication may be reproduced or distributed in any form, including an electronic retrieval system, without the prior written permission of AAMI. All requests pertaining to this document should be submitted to AAMI. It is illegal under federal law (17 U.S.C. § 101, *et seq*.) to make copies of all or any part of this document (whether internally or externally) without the prior written permission of the Association for the Advancement of Medical Instrumentation. Violators risk legal action, including civil and criminal penalties, and damages of $100,000 per offense. For permission regarding the use of all or any part of this document, complete the reprint request form at www.aami.org or contact AAMI, 4301 N. Fairfax Drive, Suite 301, Arlington, VA 22203-1633. Phone: (703) 525-4890; Fax: (703) 525-1067.

Printed in the United States of America

ISBN 1–57020–387-3

AAMI
Single user license only. Copying, networking, and distribution prohibited.
Full document details: ANSI/AAMI/ISO 15223-2:2010, Medical devices — Symbols to be used with medical device labels, labeling, and information to be supp...

Contents

Page

Glossary of equivalent standards iv

Committee representation vi

Background of ANSI/AAMI adoption of ISO 15223-2:2010 viii

Foreword iv

Introduction x

1 Scope 1

2 Normative references 1

3 Terms and definitions 1

4 Principles for identification and development of new symbols 3
4.1 Identifying the need for a symbol 3
4.2 Symbols with horizontal applications 3
4.3 Symbols for use within a restricted range of device types 3

5 Process for selecting and validating symbols for inclusion in ISO 15223-1 3
5.1 General 3
5.2 Initial evaluation 4
5.3 Second evaluation 4

6 Classification of risk 8

7 Concept development 8
7.1 Existence of other symbols 8
7.2 Symbol design 9

8 Evaluation 9
8.1 Testing early symbol concepts 9
8.2 Comprehension testing 10
8.3 Memory testing 10
8.4 Usability testing 10

9 Acceptance criteria 11
9.1 General 11
9.2 Symbols with no to low safety relevance 11
9.3 Symbols with moderate to high safety relevance 11

Annex A (normative) Information to be provided during the symbol development process for adoption of a symbol into ISO 15223-1 12

Annex B (normative) ISO/TC 145/SC 3 proposal for graphical symbols 13

Annex C (normative) IEC/SC 3C proposal for graphical symbol form 15

Bibliography 17

Documents
Document Sharing Hub

AAMI
Single user license only. Copying, networking, and distribution prohibited.
Full document details: ANSI/AAMI/ISO 15223-2:2010, Medical devices — Symbols to be used with medical device labels, labeling, and information to be supp...

Glossary of equivalent standards

International Standards adopted in the United States may include normative references to other International Standards. For each International Standard that has been adopted by AAMI (and ANSI), the table below gives the corresponding U.S. designation and level of equivalency to the International Standard. NOTE: Documents are sorted by international designation. The code in the US column, "(R)20xx" indicates the year the document was officially reaffirmed by AAMI. E.g., ANSI/AAMI/ISO 10993-4:2002/(R)2009 indicates that 10993-4, originally approved and published in 2002, was reaffirmed without change in 2009.

Other normatively referenced International Standards may be under consideration for U.S. adoption by AAMI; therefore, this list should not be considered exhaustive.

International designation	U.S. designation	Equivalency
IEC 60601-1:2005 Technical Corrigendum 1 and 2	ANSI/AAMI ES60601-1:2005 and ANSI/AAMI ES60601-1:2005/A2:2010 ANSI/AAMI ES60601-1:2005/C1:2009 (amdt)	Major technical variations C1 Identical to Corrigendum 1 & 2
IEC 60601-1-2:2007	ANSI/AAMI/IEC 60601-1-2:2007	Identical
IEC 60601-2-2:2009	ANSI/AAMI/IEC 60601-2-2:2009	Identical
IEC 60601-2-4:2002	ANSI/AAMI DF80:2003/(R)2010	Major technical variations
IEC 60601-2-19:2009	ANSI/AAMI/IEC 60601-2-19:2009	Identical
IEC 60601-2-20:2009	ANSI/AAMI/IEC 60601-2-20:2009	Identical
IEC 60601-2-21:2009	ANSI/AAMI/IEC 60601-2-21:2009	Identical
IEC 60601-2-24:1998	ANSI/AAMI ID26:2004/(R)2009	Major technical variations
IEC 60601-2-47:2001	ANSI/AAMI EC38:2007	Major technical variations
IEC 60601-2-50:2009	ANSI/AAMI/IEC 60601-2-50:2009	Identical
IEC 80601-2-30:2009 and Technical Corrigendum 1	ANSI/AAMI/IEC 80601-2-30:2009 and ANSI/AAMI/IEC 80601-2-30:2009/ C1:2009 (amdt) – consolidated text	Identical (with inclusion) C1 Identical to Corrigendum 1
IEC 80601-2-58:2008	ANSI/AAMI/IEC 80601-2-58:2008	Identical
IEC/TR 60878:2009	ANSI/AAMI/IEC TIR60878:2003	Identical
IEC/TR 62296:2009	ANSI/AAMI/IEC TIR62296:2009	Identical
IEC 62304:2006	ANSI/AAMI/IEC 62304:2006	Identical
IEC/TR 62348:2006	ANSI/AAMI/IEC TIR62348:2006	Identical
IEC/TR 62354:2009	ANSI/AAMI/IEC TIR62354:2009	Identical
IEC/TR 80002-1:2009	ANSI/IEC/TR 80002-1:2009	Identical
ISO 5840:2005	ANSI/AAMI/ISO 5840:2005/(R)2010	Identical
ISO 7198:1998	ANSI/AAMI/ISO 7198:1998/2001/(R)2010	Identical
ISO 7199:2009	ANSI/AAMI/ISO 7199:2009	Identical
ISO 8637:2004	ANSI/AAMI RD16:2007	Major technical variations
ISO 8638:2004	ANSI/AAMI RD17:2007	Major technical variations
ISO 10993-1:2009	ANSI/AAMI/ISO 10993-1:2009	Identical
ISO 10993-2:2006	ANSI/AAMI/ISO 10993-2:2006	Identical
ISO 10993-3:2003	ANSI/AAMI/ISO 10993-3:2003/(R)2009	Identical
ISO 10993-4:2002 and Amendment 1:2006	ANSI/AAMI/ISO 10993-4:2002/(R)2009 and Amendment 1:2006/(R)2009	Identical
ISO 10993-5:2009	ANSI/AAMI/ISO 10993-5:2009	Identical
ISO 10993-6:2007	ANSI/AAMI/ISO 10993-6:2007	Identical
ISO 10993-7:2008	ANSI/AAMI/ISO 10993-7:2008	Identical
ISO 10993-9:2009	ANSI/AAMI/ISO 10993-9:2009	Identical
ISO 10993-10:2002 and Amendment 1:2006	ANSI/AAMI BE78:2002/(R)2008 ANSI/AAMI BE78:2002/A1:2006/(R)2008	Minor technical variations Identical
ISO 10993-11:2006	ANSI/AAMI/ISO 10993-11:2006	Identical
ISO 10993-12:2007	ANSI/AAMI/ISO 10993-12:2007	Identical
ISO 10993-13:1998	ANSI/AAMI/ISO 10993-13:1999/(R)2004	Identical
ISO 10993-14:2001	ANSI/AAMI/ISO 10993-14:2001/(R)2006	Identical
ISO 10993-15:2000	ANSI/AAMI/ISO 10993-15:2000/(R)2006	Identical
ISO 10993-16:2010	ANSI/AAMI/ISO 10993-16:2010	Identical
ISO 10993-17:2002	ANSI/AAMI/ISO 10993-17:2002/(R)2008	Identical
ISO 10993-18:2005	ANSI/AAMI BE83:2006	Major technical variations
ISO/TS 10993-19:2006	ANSI/AAMI/ISO TIR10993-19:2006	Identical

Documents Document Sharing Hub
AAMI
Single user license only. Copying, networking, and distribution prohibited.
Full document details: ANSI/AAMI/ISO 15223-2:2010, Medical devices — Symbols to be used with medical device labels, labeling, and information to be supp...

International designation	U.S. designation	Equivalency
ISO/TS 10993-20:2006	ANSI/AAMI/ISO TIR10993-20:2006	Identical
ISO 11135-1:2007	ANSI/AAMI/ISO 11135-1:2007	Identical
ISO/TS 11135-2:2008	ANSI/AAMI/ISO TIR11135-2:2008	Identical
ISO 11137-1:2006	ANSI/AAMI/ISO 11137-1:2006/(R)2010	Identical
ISO 11137-2:2006 (2006-08-01 corrected version)	ANSI/AAMI/ISO 11137-2:2006	Identical
ISO 11137-3:2006	ANSI/AAMI/ISO 11137-3:2006/(R)2010	Identical
ISO 11138-1: 2006	ANSI/AAMI/ISO 11138-1:2006/(R)2010	Identical
ISO 11138-2: 2006	ANSI/AAMI/ISO 11138-2:2006/(R)2010	Identical
ISO 11138-3: 2006	ANSI/AAMI/ISO 11138-3:2006/(R)2010	Identical
ISO 11138-4: 2006	ANSI/AAMI/ISO 11138-4:2006/(R)2010	Identical
ISO 11138-5: 2006	ANSI/AAMI/ISO 11138-5:2006/(R)2010	Identical
ISO/TS 11139:2006	ANSI/AAMI/ISO 11139:2006	Identical
ISO 11140-1:2005	ANSI/AAMI/ISO 11140-1:2005/(R)2010	Identical
ISO 11140-3:2007	ANSI/AAMI/ISO 11140-3:2007	Identical
ISO 11140-4:2007	ANSI/AAMI/ISO 11140-4:2007	Identical
ISO 11140-5:2007	ANSI/AAMI/ISO 11140-5:2007	Identical
ISO 11607-1:2006	ANSI/AAMI/ISO 11607-1:2006	Identical
ISO 11607-2:2006	ANSI/AAMI/ISO 11607-2:2006	Identical
ISO 11737-1: 2006	ANSI/AAMI/ISO 11737-1:2006	Identical
ISO 11737-2:2009	ANSI/AAMI/ISO 11737-2:2009	Identical
ISO 13408-1:2008	ANSI/AAMI/ISO 13408-1:2008	Identical
ISO 13408-2:2003	ANSI/AAMI/ISO 13408-2:2003	Identical
ISO 13408-3:2006	ANSI/AAMI/ISO 13408-3:2006	Identical
ISO 13408-4:2005	ANSI/AAMI/ISO 13408-4:2005	Identical
ISO 13408-5:2006	ANSI/AAMI/ISO 13408-5:2006	Identical
ISO 13408-6:2006	ANSI/AAMI/ISO 13408-6:2006	Identical
ISO 13485:2003	ANSI/AAMI/ISO 13485:2003/(R)2009	Identical
ISO 14155-1:2003	ANSI/AAMI/ISO 14155-1:2003/(R)2008	Identical
ISO 14155-2:2003	ANSI/AAMI/ISO 14155-2:2003/(R)2008	Identical
ISO 14160:1998	ANSI/AAMI/ISO 14160:1998/(R)2008	Identical
ISO 14161:2009	ANSI/AAMI/ISO 14161:2009	Identical
ISO 14708-3:2008	ANSI/AAMI/ISO 14708-3:2008	Identical
ISO 14708-4:2008	ANSI/AAMI/ISO 14708-4:2008	Identical
ISO 14708-5:2010	ANSI/AAMI /ISO 14708-5:2010	Identical
ISO 14937:2009	ANSI/AAMI/ISO 14937:2009	Identical
ISO/TR 14969:2004	ANSI/AAMI/ISO TIR14969:2004	Identical
ISO 14971:2007	ANSI/AAMI/ISO 14971:2007	Identical
ISO 15223-1:2007 and A1:2008	ANSI/AAMI/ISO 15223-1:2007 and A1:2008	Identical
ISO 15223-2:2010	ANSI/AAMI/ISO 15223-2:2010	Identical
ISO 15225:2010	ANSI/AAMI/ISO 15225:2010	Identical
ISO 15674:2009	ANSI/AAMI/ISO 15674:2009	Identical
ISO 15675:2009	ANSI/AAMI/ISO 15675:2009	Identical
ISO 15882:2008	ANSI/AAMI/ISO 15882:2008	Identical
ISO 15883-1:2006	ANSI/AAMI ST15883-1:2009	Major technical variations
ISO/TR 16142:2006	ANSI/AAMI/ISO TIR16142:2005	Identical
ISO 17664:2004	ANSI/AAMI ST81:2004	Major technical variations
ISO 17665-1:2006	ANSI/AAMI/ISO 17665-1:2006	Identical (with inclusions)
ISO/TS 17665-2:2009	ANSI/AAMI/ISO TIR17665-2:2009	Identical
ISO 18472:2006	ANSI/AAMI/ISO 18472:2006	Identical
ISO/TS 19218:2005	ANSI/AAMI/ISO 19218:2005	Identical
ISO 22442-1:2007	ANSI/AAMI/ISO 22442-1:2007	Identical
ISO 22442-2:2007	ANSI/AAMI/ISO 22442-2:2007	Identical
ISO 22442-3:2007	ANSI/AAMI/ISO 22442-3:2007	Identical
ISO 25539-1:2003 and A1:2005	ANSI/AAMI/ISO 25539-1:2003/(R)2009 and A1:2005/(R)2009	Identical
ISO 25539-2:2008	ANSI/AAMI/ISO 25539-2:2008	Identical
ISO 81060-1:2007	ANSI/AAMI/ISO 81060-1:2007	Identical
ISO 81060-2:2009	ANSI/AAMI/ISO 81060-2:2009	Identical

Documents
Document Sharing Hub
AAMI
Single user license only. Copying, networking, and distribution prohibited.
Full document details: ANSI/AAMI/ISO 15223-2:2010, Medical devices — Symbols to be used with medical device labels, labeling, and information to be supp...

Committee representation

Association for the Advancement of Medical Instrumentation

Quality Management and Corresponding General Aspects for Medical Devices Committee

The adoption of ISO 15223-2 was initiated by the AAMI Quality Management and Corresponding General Aspects for Medical Devices Committee, which also functions as a U.S. Technical Advisory Group to the relevant work in the International Organization for Standardization (ISO). U.S. representatives from the AAMI Symbols and Nomenclature Working Group (U.S. Sub-TAG for ISO/TC 210/WG 3), chaired by Leighton W. Hansel of Abbott Laboratories and Charles B. Sidebottom of Medtronic, Inc. played an active part in developing the ISO standard.

At the time this document was published, the **AAMI Quality Management and Corresponding General Aspects for Medical Devices Committee** had the following members:

Cochairs
Carol L. Herman, FDA/CDRH
Charles B. Sidebottom, PE, Medtronic Inc.

Members
Leighton W. Hansel, Abbott Laboratories
Carol L. Herman, FDA/CDRH
Ed R. Kimmelman, BME, JD, (Independent Expert)
David Osborn, Philips Healthcare
Harvey Rudolph, PhD, Underwriters Laboratories Inc.
Charles B. Sidebottom, PE, Medtronic Inc.
Al Van Houdt, Spacelabs Medical Inc.

Alternates
David J. Geraghty, Spacelabs Medical Inc
Sherry Leichtweis, Abbott Laboratories
Luann M. Pendy, Medtronic Inc
Kimberly A. Trautman, FDA/CDRH

At the time this document was published, the **AAMI Symbols and Nomenclature Working Group** had the following members:

Cochairs
Leighton W. Hansel, Abbott Laboratories
Charles B. Sidebottom, PE, Medtronic Inc.

Members
Krisann M. Anderson, St Jude Medical Inc.
Candice Betz, Becton Dickinson & Company
Charles Cogdill, Boston Scientific Corporation
Rich Eaton, Medical Imaging & Technology Alliance a Division of NEMA
Christine M. Flahive, Chris Flahive Associates
Nancy George, CSQE CQA, Software Quality Management Inc.
David J. Geraghty, Spacelabs Medical Inc.
Leighton W. Hansel, Abbott Laboratories
Christopher Heckert, Zimmer Inc.
Steve Hellstrom, Hospira Worldwide Inc.
Carol L. Herman, FDA/CDRH
Joshua Kim, Welch Allyn Inc.
David Osborn, Philips Healthcare
Mandy Savino, Covidien
Charles B. Sidebottom, PE, Medtronic Inc.
John G. Smith, CareFusion
Nancy J. Stark, PhD, Clinical Device Group Inc.
Richard C. Thorne, Eli Lilly & Company

Alternates
Richard H. Bean, Zimmer Inc.
Tom C. Gorgol, Eli Lilly & Company
Michael Jaffe, PhD, Philips Healthcare

Documents Document Sharing Hub
AAMI
Single user license only. Copying, networking, and distribution prohibited.
Full document details: ANSI/AAMI/ISO 15223-2:2010, Medical devices — Symbols to be used with medical device labels, labeling, and information to be supp...

Patricia A. Melerski, Hospira Worldwide Inc.
Susan Qualey, Spacelabs Medical Inc.
Kay Sachs-Campbell, Boston Scientific Corporation
Bradford Marshall Spring, Becton Dickinson & Company
Victoria Wagman, FDA/CDRH

NOTE—Participation by federal agency representatives in the development of this standard does not constitute endorsement by the federal government or any of its agencies.

Documents
Document Sharing Hub
AAMI
Single user license only. Copying, networking, and distribution prohibited.
Full document details: ANSI/AAMI/ISO 15223-2:2010, Medical devices — Symbols to be used with medical device labels, labeling, and information to be supp...