American National Standard

ANSI/AAMI/ISO 26722:2009

Water treatment equipment for hemodialysis applications and related therapies

Association for the Advancement of Medical Instrumentation

AAMI
Single user license only. Copying, networking, and distribution prohibited.

Documents
Document Sharing Hub

Full document details: ANSI/AAMI/ISO 26722:2009, Water treatment equipment for hemodialysis applications and related therapies

Objectives and uses of AAMI standards and recommended practices

It is most important that the objectives and potential uses of an AAMI product standard or recommended practice are clearly understood. The objectives of AAMI's technical development program derive from AAMI's overall mission: the advancement of medical instrumentation. Essential to such advancement are (1) a continued increase in the safe and effective application of current technologies to patient care, and (2) the encouragement of new technologies. It is AAMI's view that standards and recommended practices can contribute significantly to the advancement of medical instrumentation, provided that they are drafted with attention to these objectives and provided that arbitrary and restrictive uses are avoided.

A voluntary *standard* for a *medical device* recommends to the manufacturer the information that should be provided with or on the product, basic safety and performance criteria that should be considered in qualifying the device for clinical use, and the measurement techniques that can be used to determine whether the device conforms with the safety and performance criteria and/or to compare the performance characteristics of different products. Some standards emphasize the information that should be provided with the device, including performance characteristics, instructions for use, warnings and precautions, and other data considered important in ensuring the safe and effective use of the device in the clinical environment. Recommending the disclosure of performance characteristics often necessitates the development of specialized test methods to facilitate uniformity in reporting; reaching consensus on these tests can represent a considerable part of committee work. When a drafting committee determines that clinical concerns warrant the establishment of *minimum* safety and performance criteria, referee tests must be provided and the reasons for establishing the criteria must be documented in the rationale.

A *recommended practice* provides guidelines for the use, care, and/or processing of a medical device or system. A recommended practice does not address device performance *per se*, but rather procedures and practices that will help ensure that a device is used safely and effectively and that its performance will be maintained.

Although a device standard is primarily directed to the manufacturer, it may also be of value to the potential purchaser or user of the device as a frame of reference for device evaluation. Similarly, even though a recommended practice is usually oriented towards healthcare professionals, it may be useful to the manufacturer in better understanding the environment in which a medical device will be used. Also, some recommended practices, while not addressing device performance criteria, provide guidelines to industrial personnel on such subjects as sterilization processing, methods of collecting data to establish safety and efficacy, human engineering, and other processing or evaluation techniques; such guidelines may be useful to health care professionals in understanding industrial practices.

In determining whether an AAMI standard or recommended practice is relevant to the specific needs of a potential user of the document, several important concepts must be recognized:

All AAMI standards and recommended practices are *voluntary* (unless, of course, they are adopted by government regulatory or procurement authorities). The application of a standard or recommended practice is solely within the discretion and professional judgment of the user of the document.

Each AAMI standard or recommended practice reflects the collective expertise of a committee of health care professionals and industrial representatives, whose work has been reviewed nationally (and sometimes internationally). As such, the consensus recommendations embodied in a standard or recommended practice are intended to respond to clinical needs and, ultimately, to help ensure patient safety. A standard or recommended practice is limited, however, in the sense that it responds generally to perceived risks and conditions that may not always be relevant to specific situations. A standard or recommended practice is an important *reference* in responsible decision-making, but it should never *replace* responsible decision-making.

Despite periodic review and revision (at least once every five years), a standard or recommended practice is necessarily a static document applied to a dynamic technology. Therefore, a standards user must carefully review the reasons why the document was initially developed and the specific rationale for each of its provisions. This review will reveal whether the document remains relevant to the specific needs of the user.

Particular care should be taken in applying a product standard to existing devices and equipment, and in applying a recommended practice to current procedures and practices. While observed or potential risks with existing equipment typically form the basis for the safety and performance criteria defined in a standard, professional judgment must be used in applying these criteria to existing equipment. No single source of information will serve to identify a particular product as "unsafe". A voluntary standard can be used as one resource, but the ultimate decision as to product safety and efficacy must take into account the specifics of its utilization and, of course, cost-benefit considerations. Similarly, a recommended practice should be analyzed in the context of the specific needs and resources of the individual institution or firm. Again, the rationale accompanying each AAMI standard and recommended practice is an excellent guide to the reasoning and data underlying its provision.

In summary, a standard or recommended practice is truly useful only when it is used in conjunction with other sources of information and policy guidance and in the context of professional experience and judgment.

INTERPRETATIONS OF AAMI STANDARDS AND RECOMMENDED PRACTICES

Requests for interpretations of AAMI standards and recommended practices must be made in writing, to the AAMI Vice President, Standards Policy and Programs. An official interpretation must be approved by letter ballot of the originating committee and subsequently reviewed and approved by the AAMI Standards Board. The interpretation will become official and representation of the Association only upon exhaustion of any appeals and upon publication of notice of interpretation in the "Standards Monitor" section of the *AAMI News*. The Association for the Advancement of Medical Instrumentation disclaims responsibility for any characterization or explanation of a standard or recommended practice which has not been developed and communicated in accordance with this procedure and which is not published, by appropriate notice, as an *official interpretation* in the *AAMI News*.

AAMI
Single user license only. Copying, networking, and distribution prohibited.

Full document details: ANSI/AAMI/ISO 26722:2009, Water treatment equipment for hemodialysis applications and related therapies

American National Standard

ANSI/AAMI/ISO 26722:2009
(Revision of AAMI RD62:2006
and ANSI/AAMI RD62:2006/A1:2009)

Water treatment equipment for hemodialysis applications and related therapies

Approved 22 March 2011 by
Association for the Advancement of Medical Instrumentation

Approved 22 April 2011
American National Standards Institute, Inc.

Abstract: Addressed to the manufacturer and/or provider of water treatment systems and/or devices used for the express purpose of providing water for hemodialysis or related therapies. Covers devices used to treat water intended for use in the delivery of hemodialysis and related therapies, including water used for: (1) the preparation of concentrates from powder or other highly concentrated media at a dialysis facility; (2) the preparation of dialysis fluid, including dialysis fluid that may be used for the preparation of substitution fluid; (3) the reprocessing of dialyzers for multiple uses. Scope includes all devices, piping and fittings between the point at which potable water is delivered to the water treatment system and the point of use of the dialysis water, e.g., water purification devices, online water quality monitors (such as conductivity monitors), and piping systems for the distribution of dialysis water.

Keywords: device, labeling, monitoring, quality, safety

AAMI
Single user license only. Copying, networking, and distribution prohibited.

Full document details: ANSI/AAMI/ISO 26722:2009, Water treatment equipment for hemodialysis applications and related therapies

AAMI Standard

This Association for the Advancement of Medical Instrumentation (AAMI) standard implies a consensus of those substantially concerned with its scope and provisions. The existence of an AAMI standard does not in any respect preclude anyone, whether they have approved the standard or not, from manufacturing, marketing, purchasing, or using products, processes, or procedures not conforming to the standard. AAMI standards are subject to periodic review, and users are cautioned to obtain the latest editions.

CAUTION NOTICE: This AAMI standard may be revised or withdrawn at any time. AAMI procedures require that action be taken to reaffirm, revise, or withdraw this standard no later than 5 years from the date of publication. Interested parties may obtain current information on all AAMI standards by calling or writing AAMI.

All AAMI standards, recommended practices, technical information reports, and other types of technical documents developed by AAMI are *voluntary*, and their application is solely within the discretion and professional judgment of the user of the document. Occasionally, voluntary technical documents are adopted by government regulatory agencies or procurement authorities, in which case the adopting agency is responsible for enforcement of its rules and regulations.

Published by

Association for the Advancement of Medical Instrumentation
4301 N. Fairfax Dr., Suite 301
Arlington, VA 22203-1633
www.aami.org

© 2011 by the Association for the Advancement of Medical Instrumentation

All Rights Reserved

Publication, reproduction, photocopying, storage, or transmission, electronically or otherwise, of all or any part of this document without the prior written permission of the Association for the Advancement of Medical Instrumentation is strictly prohibited by law. It is illegal under federal law (17 U.S.C. § 101, et seq.) to make copies of all or any part of this document (whether internally or externally) without the prior written permission of the Association for the Advancement of Medical Instrumentation. Violators risk legal action, including civil and criminal penalties, and damages of $100,000 per offense. For permission regarding the use of all or any part of this document, complete the reprint request form at www.aami.org or contact AAMI, 4301 N. Fairfax Drive Suite 301, Arlington, VA 22203-1633. Phone: +1-703- 525-4890; Fax: +1-703-525-1067.

Printed in the United States of America

ISBN 1–57020–414–4

AAMI
Single user license only. Copying, networking, and distribution prohibited.

Full document details: ANSI/AAMI/ISO 26722:2009, Water treatment equipment for hemodialysis applications and related therapies

Contents

Page

Glossary of equivalent standards ..... **iv**

Committee representation ..... **vi**

Background of US adoption of ISO 26722:2009 ..... **vii**

Foreword ..... **viii**

Introduction ..... **ix**

1 Scope **1**

2 Normative references **1**

3 Terms and definitions **2**

4 Requirements **5**
4.1 Dialysis water quality requirements **5**
4.2 Water treatment equipment requirements **7**

5 Tests **12**
5.1 Compliance with dialysis water quality requirements **12**
5.2 Compliance with water treatment equipment requirements **13**

6 Labeling **16**
6.1 General **16**
6.2 Device markings **17**
6.3 Product literature **17**

Annex A (informative) **Rationale for the development and provisions of this International Standard** **20**

Annex B (informative) **Reference tables from ISO 13959** **30**

Bibliography **32**

Documents
Document Sharing Hub

AAMI
Single user license only. Copying, networking, and distribution prohibited.

Full document details: ANSI/AAMI/ISO 26722:2009, Water treatment equipment for hemodialysis applications and related therapies

Glossary of equivalent standards

International Standards adopted in the United States may include normative references to other International Standards. For each International Standard that has been adopted by AAMI (and ANSI), the table below gives the corresponding U.S. designation and level of equivalency to the International Standard. NOTE: Documents are sorted by international designation. The code in the US column, "(R)20xx" indicates the year the document was officially reaffirmed by AAMI. E.g., ANSI/AAMI/ISO 10993-4:2002/(R)2009 indicates that 10993-4, originally approved and published in 2002, was reaffirmed without change in 2009.

Other normatively referenced International Standards may be under consideration for U.S. adoption by AAMI; therefore, this list should not be considered exhaustive.

International designation	U.S. designation	Equivalency
IEC 60601-1:2005 Technical Corrigendum 1 and 2	ANSI/AAMI ES60601-1:2005 and ANSI/AAMI ES60601-1:2005/A2:2010 ANSI/AAMI ES60601-1:2005/C1:2009 (amdt)	Major technical variations C1 Identical to Corrigendum 1 & 2
IEC 60601-1-2:2007	ANSI/AAMI/IEC 60601-1-2:2007	Identical
IEC 60601-2-2:2009	ANSI/AAMI/IEC 60601-2-2:2009	Identical
IEC 60601-2-4:2010	ANSI/AAMI/IEC 60601-2-4:2010	Identical
IEC 60601-2-16:2008	ANSI/AAMI/IEC 60601-2-16:2008	Identical
IEC 60601-2-19:2009	ANSI/AAMI/IEC 60601-2-19:2009	Identical
IEC 60601-2-20:2009	ANSI/AAMI/IEC 60601-2-20:2009	Identical
IEC 60601-2-21:2009	ANSI/AAMI/IEC 60601-2-21:2009	Identical
IEC 60601-2-24:1998	ANSI/AAMI ID26:2004/(R)2009	Major technical variations
IEC 60601-2-47:2001	ANSI/AAMI EC38:2007	Major technical variations
IEC 60601-2-50:2009	ANSI/AAMI/IEC 60601-2-50:2009	Identical
IEC 80001-1:2010	ANSI/AAMI/IEC 80001-1:2010	Identical
IEC 80601-2-30:2009 and Technical Corrigendum 1	ANSI/AAMI/IEC 80601-2-30:2009 and ANSI/AAMI/IEC 80601-2-30:2009/ C1:2009 (amdt) – consolidated text	Identical (with inclusion) C1 Identical to Corrigendum 1
IEC 80601-2-58:2008	ANSI/AAMI/IEC 80601-2-58:2008	Identical
IEC/TR 60878:2009	ANSI/AAMI/IEC TIR60878:2003	Identical
IEC/TR 62296:2009	ANSI/AAMI/IEC TIR62296:2009	Identical
IEC 62304:2006	ANSI/AAMI/IEC 62304:2006	Identical
IEC/TR 62348:2006	ANSI/AAMI/IEC TIR62348:2006	Identical
IEC/TR 62354:2009	ANSI/AAMI/IEC TIR62354:2009	Identical
IEC 62366:2007	ANSI/AAMI/IEC 62377:2007	Identical
IEC/TR 80002-1:2009	ANSI/IEC/TR 80002-1:2009	Identical
ISO 5840:2005	ANSI/AAMI/ISO 5840:2005/(R)2010	Identical
ISO 7198:1998	ANSI/AAMI/ISO 7198:1998/2001/(R)2010	Identical
ISO 7199:2009	ANSI/AAMI/ISO 7199:2009	Identical
ISO 8637:2010	ANSI/AAMI/ISO 8637:2010	Identical
ISO 8638:2010	ANSI/AAMI/ISO 8638:2010	Identical
ISO 10993-1:2009	ANSI/AAMI/ISO 10993-1:2009	Identical
ISO 10993-2:2006	ANSI/AAMI/ISO 10993-2:2006/(R)2010	Identical
ISO 10993-3:2003	ANSI/AAMI/ISO 10993-3:2003/(R)2009	Identical
ISO 10993-4:2002 and Amendment 1:2006	ANSI/AAMI/ISO 10993-4:2002/(R)2009 and Amendment 1:2006/(R)2009	Identical
ISO 10993-5:2009	ANSI/AAMI/ISO 10993-5:2009	Identical
ISO 10993-6:2007	ANSI/AAMI/ISO 10993-6:2007/(R)2010	Identical
ISO 10993-7:2008	ANSI/AAMI/ISO 10993-7:2008	Identical
ISO 10993-9:2009	ANSI/AAMI/ISO 10993-9:2009	Identical
ISO 10993-10:2010	ANSI/AAMI/ISO 10993-10:2010	Identical
ISO 10993-11:2006	ANSI/AAMI/ISO 10993-11:2006/(R)2010	Identical
ISO 10993-12:2007	ANSI/AAMI/ISO 10993-12:2007	Identical
ISO 10993-13:2010	ANSI/AAMI/ISO 10993-13:2010	Identical
ISO 10993-14:2001	ANSI/AAMI/ISO 10993-14:2001/(R)2006	Identical
ISO 10993-15:2000	ANSI/AAMI/ISO 10993-15:2000/(R)2006	Identical
ISO 10993-16:2010	ANSI/AAMI/ISO 10993-16:2010	Identical
ISO 10993-17:2002	ANSI/AAMI/ISO 10993-17:2002/(R)2008	Identical
ISO 10993-18:2005	ANSI/AAMI BE83:2006	Major technical variations
ISO/TS 10993-19:2006	ANSI/AAMI/ISO TIR10993-19:2006	Identical
ISO/TS 10993-20:2006	ANSI/AAMI/ISO TIR10993-20:2006	Identical
ISO 11135-1:2007	ANSI/AAMI/ISO 11135-1:2007	Identical
ISO/TS 11135-2:2008	ANSI/AAMI/ISO TIR11135-2:2008	Identical
ISO 11137-1:2006	ANSI/AAMI/ISO 11137-1:2006/(R)2010	Identical

© 2011 Association for the Advancement of Medical Instrumentation ■ ANSI/AAMI/ISO 26722:2009

Documents
Document Sharing Hub

AAMI
Single user license only. Copying, networking, and distribution prohibited.

Full document details: ANSI/AAMI/ISO 26722:2009, Water treatment equipment for hemodialysis applications and related therapies

International designation	U.S. designation	Equivalency
ISO 11137-2:2006 (2006-08-01 corrected version)	ANSI/AAMI/ISO 11137-2:2006	Identical
ISO 11137-3:2006	ANSI/AAMI/ISO 11137-3:2006/(R)2010	Identical
ISO 11138-1: 2006	ANSI/AAMI/ISO 11138-1:2006/(R)2010	Identical
ISO 11138-2: 2006	ANSI/AAMI/ISO 11138-2:2006/(R)2010	Identical
ISO 11138-3: 2006	ANSI/AAMI/ISO 11138-3:2006/(R)2010	Identical
ISO 11138-4: 2006	ANSI/AAMI/ISO 11138-4:2006/(R)2010	Identical
ISO 11138-5: 2006	ANSI/AAMI/ISO 11138-5:2006/(R)2010	Identical
ISO/TS 11139:2006	ANSI/AAMI/ISO 11139:2006	Identical
ISO 11140-1:2005	ANSI/AAMI/ISO 11140-1:2005/(R)2010	Identical
ISO 11140-3:2007	ANSI/AAMI/ISO 11140-3:2007	Identical
ISO 11140-4:2007	ANSI/AAMI/ISO 11140-4:2007	Identical
ISO 11140-5:2007	ANSI/AAMI/ISO 11140-5:2007	Identical
ISO 11607-1:2006	ANSI/AAMI/ISO 11607-1:2006/(R)2010	Identical
ISO 11607-2:2006	ANSI/AAMI/ISO 11607-2:2006/(R)2010	Identical
ISO 11663:2009	ANSI/AAMI/ISO 11633:2009	Identical
ISO 11737-1: 2006	ANSI/AAMI/ISO 11737-1:2006	Identical
ISO 11737-2:2009	ANSI/AAMI/ISO 11737-2:2009	Identical
ISO/TS 12417:2011	ANSI/AAMI/ISO TIR12417:2011	Identical
ISO 13408-1:2008	ANSI/AAMI/ISO 13408-1:2008	Identical
ISO 13408-2:2003	ANSI/AAMI/ISO 13408-2:2003	Identical
ISO 13408-3:2006	ANSI/AAMI/ISO 13408-3:2006	Identical
ISO 13408-4:2005	ANSI/AAMI/ISO 13408-4:2005	Identical
ISO 13408-5:2006	ANSI/AAMI/ISO 13408-5:2006	Identical
ISO 13408-6:2006	ANSI/AAMI/ISO 13408-6:2006	Identical
ISO 13485:2003	ANSI/AAMI/ISO 13485:2003/(R)2009	Identical
ISO 13958:2009	ANSI/AAMI/ISO 13958:2009	Identical
ISO 13959:2009	ANSI/AAMI/ISO 13959:2009	Identical
ISO 14155:2011	ANSI/AAMI/ISO 14155:2011	Identical
ISO 14160:1998	ANSI/AAMI/ISO 14160:1998/(R)2008	Identical
ISO 14161:2009	ANSI/AAMI/ISO 14161:2009	Identical
ISO 14708-3:2008	ANSI/AAMI/ISO 14708-3:2008	Identical
ISO 14708-4:2008	ANSI/AAMI/ISO 14708-4:2008	Identical
ISO 14708-5:2010	ANSI/AAMI /ISO 14708-5:2010	Identical
ISO 14937:2009	ANSI/AAMI/ISO 14937:2009	Identical
ISO/TR 14969:2004	ANSI/AAMI/ISO TIR14969:2004	Identical
ISO 14971:2007	ANSI/AAMI/ISO 14971:2007/(R)2010	Identical
ISO 15223-1:2007 and A1:2008	ANSI/AAMI/ISO 15223-1:2007 and A1:2008	Identical
ISO 15223-2:2010	ANSI/AAMI/ISO 15223-2:2010	Identical
ISO 15225:2010	ANSI/AAMI/ISO 15225:2010	Identical
ISO 15674:2009	ANSI/AAMI/ISO 15674:2009	Identical
ISO 15675:2009	ANSI/AAMI/ISO 15675:2009	Identical
ISO 15882:2008	ANSI/AAMI/ISO 15882:2008	Identical
ISO 15883-1:2006	ANSI/AAMI ST15883-1:2009	Major technical variations
ISO/TR 16142:2006	ANSI/AAMI/ISO TIR16142:2005	Identical
ISO 17664:2004	ANSI/AAMI ST81:2004	Major technical variations
ISO 17665-1:2006	ANSI/AAMI/ISO 17665-1:2006	Identical (with inclusions)
ISO/TS 17665-2:2009	ANSI/AAMI/ISO TIR17665-2:2009	Identical
ISO 18472:2006	ANSI/AAMI/ISO 18472:2006/(R)2010	Identical
ISO/TS 19218:2005	ANSI/AAMI/ISO 19218:2005	Identical
ISO 20857:2010	ANSI/AAMI/ISO 20857:2010	Identical
ISO 22442-1:2007	ANSI/AAMI/ISO 22442-1:2007	Identical
ISO 22442-2:2007	ANSI/AAMI/ISO 22442-2:2007	Identical
ISO 22442-3:2007	ANSI/AAMI/ISO 22442-3:2007	Identical
ISO 23500:2011	ANSI/AAMI/ISO 23500:2011	Identical
ISO 25539-1:2003 and A1:2005	ANSI/AAMI/ISO 25539-1:2003/(R)2009 and A1:2005/(R)2009	Identical
ISO 25539-2:2008	ANSI/AAMI/ISO 25539-2:2008	Identical
ISO 26722:2009	ANSI/AAMI/ISO 26722:2009	Identical
ISO 27186:2010	ANSI/AAMI/ISO 27186:2010	Identical
ISO 80369-1:2010	ANSI/AAMI/ISO 80369-1:2010	Identical
ISO 81060-1:2007	ANSI/AAMI/ISO 81060-1:2007	Identical
ISO 81060-2:2009	ANSI/AAMI/ISO 81060-2:2009	Identical

Documents
Document Sharing Hub
AAMI
Single user license only. Copying, networking, and distribution prohibited.
Full document details: ANSI/AAMI/ISO 26722:2009, Water treatment equipment for hemodialysis applications and related therapies

Committee representation

Association for the Advancement of Medical Instrumentation

Renal Disease and Detoxification Committee

The adoption of ISO 26722:2009 as an American National Standard was initiated by the AAMI Renal Disease and Detoxification Committee.

Committee approval of the standard does not necessarily imply that all committee members voted for its approval. At the time this document was published, the **AAMI Renal Disease and Detoxification Committee** had the following members:

Cochairs:
Conor Curtin
Richard A. Ward, PhD

Members:
G Steven Acres, MD, Carolina Regional Nephrology Associates
Larry Alexander, Florian Services
James Weldon Baker, AmeriWater
Robert Berube, Church & Dwight Co Inc.
Arnold Boehnlein, Baxter Healthcare Corporation
Karla S. Byrne, Rockwell Medical Inc.
Wayne Carlson, Minntech Corporation
Danilo B. Concepcion, CHT CCHT CBNT, St Joseph Hospital Renal Center
Conor Curtin, Fresenius Medical Care North America
Jim Curtis, Home Dialysis Plus LTD
R. Barry Deeter, RN MSN, University of Utah Dialysis Program
Robert Dudek, Siemens Water Technologies Corp
Gema Gonzalez, FDA/CDRH
Elizabeth Howard, Davita
Byron L. Jacobs, CBET, Sanford USD Medical Center
David C. Katz, El Camino Hospital
Thomas Kelly, Medcatalyst Consulting LLC
Fei M. Law, Gambro Renal Products Inc.
Nathan W. Levin, MD, Renal Research Institute LLC
Jo Ann Maltais, PhD, Maltais Consulting
Duane Martz, AAEE/MBA, B. Braun of America
Lane McCarthy, CCHT, Hortense & Louis Rubin Dialysis Center
Bruce H. Merriman, Central Florida Kidney Centers
Judith Noble-Wang, Centers for Disease Control and Prevention
Glenda Payne, RN, MS,CNN, Centers for Medicare & Medicaid Services
James D. Stewardson, Brighton, CO
Vern S. Taaffe, Reprocessing Products Corp
Denny Treu, BSME, NxStage Medical Inc.
David J. Vanella, RAC, Renal Solutions Inc.
Richard A. Ward, PhD, University of Louisville

Alternates
John Bacon, Rockwell Medical Inc.
Ted A. Kasparek, Davita
Mike Lorenson, Reprocessing Products Corp
Gregory Montgomery, Siemens Water Technologies Corp
John A. Rickert, Minntech Corporation
Brooks E Rogers, Fresenius Medical Care North America
Steve Rowles, Church & Dwight Co Inc.
Ty Shockley, Baxter Healthcare Corporation
David Updyke, Renal Solutions Inc.
Gary Warns, Gambro Renal Products Inc.
Michael Webb, NxStage Medical Inc.

NOTE—Participation by federal agency representatives in the development of this technical information report does not constitute endorsement by the federal government or any of its agencies.

Documents Document Sharing Hub
AAMI
Single user license only. Copying, networking, and distribution prohibited.
Full document details: ANSI/AAMI/ISO 26722:2009, Water treatment equipment for hemodialysis applications and related therapies

Background on AAMI adoption of ISO 26722:2009

As indicated in the foreword to the main body of this document (page viii), the International Organization for Standardization (ISO) is a worldwide federation of national standards bodies. The United States is one of the ISO members that took an active role in the development of this standard, which was developed by ISO Technical Committee 150, Subcommittee 2, *Cardiovascular implants and extracorporeal systems*, to fill a need for standards that ensure equipment used to treat water for hemodialysis adequately protects patients from adverse effects arising from known chemical and microbial contaminants found in drinking water supplies

U.S. participation in this ISO TC is organized through the U.S. Technical Advisory Group for ISO/TC 150/SC 2, administered by the Association for the Advancement of Medical Instrumentation (AAMI). The U.S. TAG for ISO/TC 150/SC 2 supports the requirements provided in this document.

The U.S. adoption of ANSI/AAMI/ISO 26722:2009 was approved by the American National Standards Institute (ANSI) as a partial revision of ANSI/AAMI RD62:2006 and ANSI/AAMI RD62:2006/A1:2009, *Water treatment equipment for hemodialysis applications*, on 22 April 2011 (ANSI/AAMI/ISO 13959, *Water for hemodialysis and related therapies,* completes the revision*)*. The AAMI Renal Disease and Detoxification Committee (U.S. sub-TAG for ISO/TC 150/SC 2/WG 5, Renal replacement, detoxification and apheresis) contributed to the development and initiated the U.S. adoption of ISO 26722:2009.

The major difference between ANSI/AAMI RD62:2006 and ANSI/AAMI/ISO 26722:2009 is that the water quality standard found in ANSI/AAMI RD62:2006 is not included in ANSI/AAMI/ISO 26722:2009 but is now published in a separate standard, ANSI/AAMI/ISO 13959:2009, *Water for hemodialysis and related therapies*.

AAMI and ANSI procedures require that standards be reviewed and, if necessary, revised every five years to reflect technological advances that may have occurred since publication.

AAMI (and ANSI) have adopted other ISO standards. See the Glossary of Equivalent Standards for a list of ISO standards adopted by AAMI which gives the corresponding U.S. designation and the level of equivalency with the ISO standard.

The concepts incorporated in this standard should not be considered inflexible or static. This standard, like any other, must be reviewed and updated periodically to assimilate progressive technological developments. To remain relevant, it must be modified as technological advances are made and as new data come to light.

Suggestions for improving this standard are invited. Comments and suggested revisions should be sent to Standards Department, AAMI, 4301 N. Fairfax Drive, Suite 301, Arlington, VA 22203-1633.

NOTE—Beginning with the foreword on page viii, this American National Standard is identical to ISO 26722:2009.

Documents
Document Sharing Hub
AAMI
Single user license only. Copying, networking, and distribution prohibited.
Full document details: ANSI/AAMI/ISO 26722:2009, Water treatment equipment for hemodialysis applications and related therapies

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 26722 was prepared by Technical Committee ISO/TC 150, *Implants for surgery*, Subcommittee SC 2, *Cardiovascular implants and extracorporeal systems*.

Documents Document Sharing Hub
AAMI
Single user license only. Copying, networking, and distribution prohibited.
Full document details: ANSI/AAMI/ISO 26722:2009, Water treatment equipment for hemodialysis applications and related therapies

Introduction

This International Standard reflects the conscientious efforts of concerned physicians, clinical engineers, nurses, dialysis technicians and dialysis patients, in consultation with device manufacturers and government representatives, to develop an International Standard for performance levels that could be reasonably achieved at the time of publication. The term “consensus,” as applied to the development of voluntary medical device International Standards, does not imply unanimity of opinion, but rather reflects the compromise necessary in some instances when a variety of interests must be merged.

The provisions of this International Standard apply to individual water treatment devices and to water treatment systems assembled from one or more of these devices. In the first instance, this International Standard is directed at the individual or company that specifies the complete water treatment system and, second, at the supplier who assembles and installs the system. Since systems may be assembled from a number of individual water treatment devices, the provisions of this International Standard are also directed at the manufacturers of these devices, provided that the manufacturer indicates that the device is intended for use in hemodialysis applications. This International Standard is written principally to address water treatment systems for dialysis facilities treating multiple patients. However, many of its provisions equally apply to water treatment systems used in applications where a single patient is treated, such as in a home dialysis or acute hospital dialysis setting. Specifically, requirements for the chemical and microbiological quality of water are considered to apply in all settings, regardless of whether a single patient or many patients are being treated.

The verbal forms used in this International Standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this International Standard, the auxiliary verb:

— “shall” means that compliance with a requirement or a test is mandatory for compliance with this International Standard;

— “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this International Standard;

— “may” is used to describe a permissible way to achieve compliance with a requirement or test.

The requirements established by this International Standard will help protect hemodialysis patients from adverse effects arising from known chemical and microbial contaminants found in water supplies. However, proper dialysis and patient safety is ultimately dependent on the quality of the dialysis fluid. Since the manufacturer or supplier of water treatment equipment does not have control over the dialysis fluid, any reference to dialysis fluid in this International Standard is for clarification only and not a requirement of the manufacturer. The responsibility for assuring that the dialysis fluid is not contaminated, mismatched, or otherwise damaging to the patient rests with the clinical professionals caring for the patient under the supervision of the medical director. Recommendations on the preparation and handling of water and dialysis fluid in a dialysis facility are provided in ISO 23500.

Documents
Document Sharing Hub
AAMI
Single user license only. Copying, networking, and distribution prohibited.
Full document details: ANSI/AAMI/ISO 26722:2009, Water treatment equipment for hemodialysis applications and related therapies

© 2011 Association for the Advancement of Medical Instrumentation ■ ANSI/AAMI/ISO 26722:2009

Documents
Document Sharing Hub
AAMI
Single user license only. Copying, networking, and distribution prohibited.
Full document details: ANSI/AAMI/ISO 26722:2009, Water treatment equipment for hemodialysis applications and related therapies

Water treatment equipment for hemodialysis applications and related therapies

1 Scope

This International Standard is addressed to the manufacturer and/or supplier of water treatment systems and/or devices used for the express purpose of providing water for hemodialysis or related therapies.

This International Standard covers devices used to treat water intended for use in the delivery of hemodialysis and related therapies. Included in the scope of this International Standard is water used for: (1) the preparation of concentrates from powder or other highly concentrated media at a dialysis facility; (2) the preparation of dialysis fluid that may be used for the preparation of substitution fluid; (3) the reprocessing of dialyzers for multiple uses.

Included within the scope of this International Standard are all devices, piping and fittings between the point at which potable water is delivered to the water treatment system and the point of use of the dialysis water. Examples of devices included within the scope of this International Standard are water purification devices, online water quality monitors (such as conductivity monitors), and piping systems for the distribution of dialysis water.

Excluded from the scope of this International Standard are dialysis fluid supply systems that proportion water and concentrates to produce dialysis fluid, sorbent dialysis fluid regeneration systems that regenerate and recirculate small volumes of the dialysis fluid, dialysis concentrates, hemodiafiltration systems, hemofiltration systems, systems that process dialyzers for multiple uses, and peritoneal dialysis systems. Some of these devices, such as dialysis fluid delivery systems and concentrates, are addressed in other International Standards. Also excluded from the scope of this International Standard are requirements for the ongoing monitoring of the purity of water used for dialysis fluid, concentrate preparation or dialyzer reprocessing.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 13959:2009, *Water for hemodialysis and related therapies*

© 2011 Association for the Advancement of Medical Instrumentation ■ ANSI/AAMI/ISO 26722:2009

Documents
Document Sharing Hub
AAMI
Single user license only. Copying, networking, and distribution prohibited.
Full document details: ANSI/AAMI/ISO 26722:2009, Water treatment equipment for hemodialysis applications and related therapies

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

3.1
action level
concentration of a contaminant at which steps should be taken to interrupt the trend toward higher, unacceptable levels

3.2
chlorine, combined
chlorine that is chemically combined, such as in chloramine compounds

NOTE No direct test exists for measuring combined chlorine, but it can be measured indirectly by measuring both total and free chlorine and calculating the difference.

3.3
chlorine, free
dissolved molecular chlorine

3.4
chlorine, total
sum of **combined chlorine** (3.2) and **free chlorine** (3.3)

NOTE Chlorine can exist in water as dissolved molecular chlorine (free chlorine) or in chemically combined forms (combined chlorine). Where chloramine is used to disinfect water supplies, chloramine is usually the principal component of combined chlorine.

3.5
device
individual water purification unit, such as a softener, carbon adsorption bed, reverse osmosis unit or deionizer

NOTE This term is synonymous with the term “component” as used by the U.S. Food and Drug Administration (see Reference [26]).

3.6
dialysis fluid
aqueous fluid containing electrolytes and usually buffer and glucose, which is intended to exchange solutes with blood during hemodialysis

NOTE 1 The word “dialysis fluid” is used throughout this document to mean the fluid made from dialysis water and concentrates which is delivered to the dialyzer by the dialysis fluid delivery system. Such phrases as “dialysate,” “dialysis solution,” or “dialyzing fluid” may be used in place of dialysis fluid.

NOTE 2 The dialysis fluid entering the dialyzer is referred to as “fresh dialysis fluid,” while the fluid leaving the dialyzer is referred to as “spent dialysis fluid.”

NOTE 3 Dialysis fluid does not include prepackaged parenteral fluids used in some renal replacement therapies, such as hemodiafiltration and hemofiltration.

3.7
dialysis fluid delivery system
device that: (1) prepares dialysis fluid online from dialysis water and concentrates or that stores and distributes premixed dialysis fluid; (2) circulates the dialysis fluid through the dialyzer; (3) monitors the

Documents Document Sharing Hub
AAMI
Single user license only. Copying, networking, and distribution prohibited.
Full document details: ANSI/AAMI/ISO 26722:2009, Water treatment equipment for hemodialysis applications and related therapies