

American National Standard



ANSI/AAMI/ ISO 80369-3: 2016/A1:2019

Small-bore connectors
for liquids and gases in
healthcare applications—
Part 3: Connectors for
enteral applications—
Amendment 1

Small-bore connectors for liquids and gases in healthcare applications—Part 3: Connectors for enteral applications—Amendment 1

Approved 11 July 2019 by
AAMI

Approved 9 December 2019 by
American National Standards Institute

Abstract: This amendment (1) modifies the scope by removing the following exclusion: "Medical devices for rectal drainage, rectal administration of medicines or fluid, and any other rectal access medical device" and (2) makes dimensional changes to revise the through bore of the male connector to return to the dimension that was used for all of the misconnection analysis which was performed during the development of the connector. The through bore will also be further specified denoting a maximum draft angle for a minimum distance.

AAMI Standard

This Association for the Advancement of Medical Instrumentation (AAMI) standard implies a consensus of those substantially concerned with its scope and provisions. The existence of an AAMI standard does not in any respect preclude anyone, whether they have approved the standard or not, from manufacturing, marketing, purchasing, or using products, processes, or procedures not conforming to the standard. AAMI standards are subject to periodic review, and users are cautioned to obtain the latest editions.

CAUTION NOTICE: This AAMI standard may be revised or withdrawn at any time. AAMI procedures require that action be taken to reaffirm, revise, or withdraw this standard no later than 5 years from the date of publication. Interested parties may obtain current information on all AAMI standards by calling or writing AAMI.

All AAMI standards, recommended practices, technical information reports, and other types of technical documents developed by AAMI are *voluntary*, and their application is solely within the discretion and professional judgment of the user of the document. Occasionally, voluntary technical documents are adopted by government regulatory agencies or procurement authorities, in which case the adopting agency is responsible for enforcement of its rules and regulations.

Published by

AAMI
901 N. Glebe Road, Suite 300
Arlington, VA 22203

© 2019 by the Association for the Advancement of Medical Instrumentation

All Rights Reserved

This publication is subject to copyright claims of ISO and AAMI. No part of this publication may be reproduced or distributed in any form, including an electronic retrieval system, without the prior written permission of AAMI. All requests pertaining to this draft should be submitted to AAMI. It is illegal under federal law (17 U.S.C. § 101, *et seq.*) to make copies of all or any part of this document (whether internally or externally) without the prior written permission of the Association for the Advancement of Medical Instrumentation. Violators risk legal action, including civil and criminal penalties, and damages of \$100,000 per offense. For permission regarding the use of all or any part of this document, contact AAMI, 901 N. Glebe Road, Suite 300, Arlington, VA 22203 Phone: (703) 525-4890; Fax: (703) 525-1067.

Printed in the United States of America

ISBN 978-1-57020-736-5

Contents

Page

Committee representation.....	iv
Background of ANSI/AAMI adoption of ISO 80369-3:2016/Amd1:2019	vi
Foreword.....	vii
ANSI/AAMI/ISO 80369-3:2016/Amd1:2019	1