



ASA/ANSI S1.4-2014/Part 1 / IEC
61672-1:2013
(Revision of ANSI S1.4-1983 [R2006],
ANSI S1.4a-1985 [R2006], and ANSI
S1.43-1997 [R2007])

Reaffirmed by ANSI on January 19, 2024

AMERICAN NATIONAL STANDARD

Electroacoustics – Sound Level Meters – Part 1: Specifications

(a nationally adopted international standard)

Secretariat:

Acoustical Society of America

Approved on July 21, 2014:

American National Standards Institute, Inc.

Abstract

This part of ANSI/ASA S1.4 / IEC 61672 gives electroacoustical performance specifications for three kinds of sound-measuring instruments: a time-weighting sound level meter that measures exponential-time-weighted, frequency-weighted sound levels; an integrating-averaging sound level meter that measures time-averaged, frequency-weighted sound levels; and an integrating sound level meter that measures frequency-weighted sound exposure levels.

ANSI/ASA S1.4-2014/Part 1 / IEC 61672-1:2013
(Revision of ANSI S1.4-1983 [R2006], ANSI S1.4a-1985 [R2006], and ANSI S1.43-1997 [R2007])

Reaffirmed by ANSI
January 19, 2024

AMERICAN NATIONAL STANDARD

**Electroacoustics – Sound Level Meters –
Part 1: Specifications**

(a nationally adopted international standard)

Secretariat:

Acoustical Society of America

Approved on July 21, 2014 by:

American National Standards Institute, Inc.

Abstract

This part of ANSI/ASA S1.4 / IEC 61672 gives electroacoustical performance specifications for three kinds of sound-measuring instruments: a time-weighting sound level meter that measures exponential-time-weighted, frequency-weighted sound levels; an integrating-averaging sound level meter that measures time-averaged, frequency-weighted sound levels; and an integrating sound level meter that measures frequency-weighted sound exposure levels.

AMERICAN NATIONAL STANDARD

**Electroacoustics – Sound Level Meters –
Part 1: Specifications
(a nationally adopted international standard)**

ANSI/ASA S1.4-2014/Part 1 /
IEC 61672-1:2013

Accredited Standards Committee S1, Acoustics

Standards Secretariat
Acoustical Society of America
1305 Walt Whitman Road, Suite 300
Melville, NY 11747

The American National Standards Institute, Inc. (ANSI) is the national coordinator of voluntary standards development and the clearinghouse in the U.S.A. for information on national and international standards.

The Acoustical Society of America (ASA) is an organization of scientists and engineers formed in 1929 to increase and diffuse the knowledge of acoustics and to promote its practical applications.



AMERICAN NATIONAL STANDARDS ON ACOUSTICS

The Acoustical Society of America (ASA) provides the Secretariat for Accredited Standards Committees S1 on Acoustics, S2 on Mechanical Vibration and Shock, S3 on Bioacoustics, S3/SC 1 on Animal Bioacoustics, and S12 on Noise. These committees have wide representation from the technical community (manufacturers, consumers, trade associations, organizations with a general interest, and government representatives). The standards are published by the Acoustical Society of America as American National Standards after approval by their respective Standards Committees and the American National Standards Institute (ANSI).


These standards are developed and published as a public service to provide standards useful to the public, industry, and consumers, and to Federal, State, and local governments.

Each of the Accredited Standards Committees (operating in accordance with procedures approved by ANSI) is responsible for developing, voting upon, and maintaining or revising its own Standards. The ASA Standards Secretariat administers Committee organization and activity and provides liaison between the Accredited Standards Committees and ANSI. After the Standards have been produced and adopted by the Accredited Standards Committees, and approved as American National Standards by ANSI, the ASA Standards Secretariat arranges for their publication and distribution.

An American National Standard implies a consensus of those substantially concerned with its scope and provisions. Consensus is established when, in the judgment of the ANSI Board of Standards Review, substantial agreement has been reached by directly and materially affected interests. Substantial agreement means much more than a simple majority, but not necessarily unanimity. Consensus requires that all views and objections be considered and that a concerted effort be made towards their resolution.

The use of an American National Standard is completely voluntary. Their existence does not in any respect preclude anyone, whether he or she has approved the Standards or not, from manufacturing, marketing, purchasing, or using products, processes, or procedures not conforming to the Standards.

NOTICE: This American National Standard may be revised or withdrawn at any time. The procedures of the American National Standards Institute require that action be taken periodically to reaffirm, revise, or withdraw this Standard.



Acoustical Society of America
ASA Secretariat
1305 Walt Whitman Road, Suite 300
Melville, New York 11747
Telephone: 1 (631) 390-0215
Fax: 1 (631) 923-2875
E-mail: asastds@acousticalsociety.org

© 2014 by Acoustical Society of America. This standard may not be reproduced in whole or in part in any form for sale, promotion, or any commercial purpose, or any purpose not falling within the provisions of the U.S. Copyright Act of 1976, without prior written permission of the publisher. For permission, address a request to the Standards Secretariat of the Acoustical Society of America.

These materials are subject to copyright claims of IEC and ASA. No part of this publication may be reproduced in any form, including an electronic retrieval system, without the prior written permission of the Acoustical Society of America (ASA). All requests pertaining to this standard should be submitted to the ASA.

Contents

1	Scope	1
2	Normative references	2
3	Terms and definitions	2
4	Reference environmental conditions	9
5	Performance specifications	9
5.1	General	9
5.2	Adjustments at the calibration check frequency	12
5.3	Corrections to indicated levels	13
5.4	Directional response	15
5.5	Frequency weightings	16
5.6	Level linearity	18
5.7	Self-generated noise	19
5.8	Time-weightings F and S	19
5.9	Toneburst response	20
5.10	Response to repeated tonebursts	20
5.11	Overload indication	22
5.12	Under-range indication	23
5.13	C-weighted peak sound level	23
5.14	Stability during continuous operation	24
5.15	High-level stability	24
5.16	Reset	25
5.17	Thresholds	25
5.18	Display	25
5.19	Analogue or digital output	25
5.20	Timing facilities	26
5.21	Radio frequency emissions and disturbances to a public power supply	26
5.22	Crosstalk	27
5.23	Power supply	27
6	Environmental, electrostatic, and radio-frequency requirements	28
6.1	General	28
6.2	Static pressure	28
6.3	Air temperature	29
6.4	Humidity	29
6.5	Electrostatic discharge	29
6.6	A.C. power-frequency and radio-frequency fields	29
6.7	Mechanical vibration	31
7	Provision for use with auxiliary devices	31
8	Marking	32

9	Instruction Manual.....	32
9.1	General	32
9.2	Information for operation	32
9.3	Information for testing	36
	Annex A (informative) Relationship between tolerance interval, corresponding acceptance interval and the maximum-permitted uncertainty of measurement.....	39
	Annex B (normative) Maximum-permitted uncertainties of measurement.....	40
	Annex C (informative) Example assessments of conformance to specifications of this standard.....	42
	Annex D (normative) Frequencies at fractional-octave intervals	45
	Annex E (normative) Analytical expressions for frequency-weightings C, A, and Z.....	47
	Annex F (informative) Additional information for U.S. users regarding low-frequency tolerances in this standard vs. previous versions of ANSI S1.4	49

Tables

Table 1 – Acceptance limits for the difference between a measured windscreen correction and the corresponding correction given in the Instruction Manual	14
Table 2 – Acceptance limits for deviations of directional response from the design goal	16
Table 3 – Frequency weightings and acceptance limits	17
Table 4 – Reference 4 kHz toneburst responses and acceptance limits.....	21
Table 5 – Reference differences for C-weighted peak sound levels and acceptance limits	24
Table 6 – Limits for conducted disturbance to the voltage of a public supply of electric power	27
Table B.1 – Maximum-permitted uncertainties of measurement for a coverage probability of 95 %	40
Table C.1 – Examples of assessment of conformance	43
Table D.1 – Frequencies at one-third-octave intervals	45
Table D.2 – Frequencies at one-sixth-octave intervals.....	46
Table D.3 – Frequencies at one-twelfth-octave intervals.....	46
Table F.1 — Frequency response tolerance limits in ANSI S1.4-1983	50

Figures

Figure 1 – Principal steps involved in forming a time-weighted sound level	4
--	---

Figure A.1 – Relationship between tolerance interval, corresponding acceptance interval and the maximum-permitted uncertainty of measurement	39
Figure C.1 – Examples of assessment of conformance	44

Foreword

[This Foreword is for information only and is not a part of the American National Standard ANSI/ASA S1.4-2014/Part 1 / IEC 61672-1:2013 American National Standard Electroacoustics – Sound level meters – Part 1: Specifications. As such, this Foreword may contain material that has not been subjected to public review or a consensus process. In addition, it does not contain requirements necessary for conformance to the standard.]

This standard comprises a part of a group of definitions, standards, and specifications for use in acoustics. It was developed and approved by Accredited Standards Committee S1 Acoustics, under its approved operating procedures. Those procedures have been accredited by the American National Standards Institute (ANSI). The Scope of Accredited Standards Committee S1 is as follows:

Standards, specifications, methods of measurement and test, and terminology in the field of physical acoustics, including architectural acoustics, electroacoustics, sonics and ultrasonics, and underwater sound, but excluding those aspects which pertain to biological safety, tolerances, and comfort.

This nationally adopted international standard is a revision of ANSI S1.4-1983 (R 2006), ANSI S1.4A-1985 (R 2006), and ANSI S1.43-1997 (R 2007), which have been technically revised.

This standard is identical to IEC 61672-1:2013, which was prepared by IEC Technical Committee 29, in cooperation with the International Organization of Legal Metrology (OIML). However, in conformance with ANSI and IEC rules, the words “this part of ANSI/ASA S1.4 / IEC 61672” replace the words “this part of IEC 61672” where they appear in the IEC document, decimal points were substituted in place of the decimal commas used in IEC documents, and American English spelling is used in place of British English spelling. An informational annex is also included that is not found in IEC 61672/Part 1. In accordance with clause A.1 b) of *ANSI Procedures for the National Adoption of ISO and IEC Standards as American National Standards*, a standard will still be considered identical if it includes “any regional or national informative material (e.g. informative annexes that do not alter, add to or delete from the provision of the ISO or IEC standard); examples of informative material are advice to users, training guidance or suggested forms or reports.”

This standard includes six Annexes. Annexes B, D and E are normative and are considered to be a part of this standard. Annexes A and C are informative and are not considered part of this standard. Annex F is additional, totally informative material for U.S. users and is shown in **bold, blue text within a blue border**.

The ANSI/ASA equivalents for the IEC standards in the IEC 61672 series are given below:

- ANSI/ASA S1.4-2014/Part 1 / IEC 61672-1:2013 is an identical national adoption of IEC 61672-1:2013.
- ANSI/ASA S1.4-2014/Part 2 / IEC 61672-2:2013 is an identical national adoption of IEC 61672-2:2013.
- ANSI/ASA S1.4-2014/Part 3 / IEC 61672-3:2013 is an identical national adoption of IEC 61672-3:2013.

At the time this Standard was submitted to Accredited Standards Committee S1, Acoustics for approval, the membership was as follows:

R.J. Peppin, *Chair*
A. Scharine, *Vice-Chair*

S.B. Blaeser, *Secretary*

3M Personal Safety Division, Detection Solutions	P.J. Battenberg
.....	M. Wurm (Alt.)
Acoustical Society of America	R.J. Peppin
Air-Conditioning, Heating and Refrigeration Institute	S.J. Lind
.....	D. Abbate (Alt.)
American Industrial Hygiene Association	D. Driscoll
.....	D. Sandfort (Alt.)
Campanella Associates	A.J. Campanella
Casella USA	R. Brauch
ETS-Lindgren Acoustic Systems	D. Winker
.....	M. Black (Alt.)
G.R.A.S. Sound & Vibration	J. Soendergaard
.....	B. Schustrich (Alt.)
Information Technology Industry Council	W.M. Beltman
.....	J. Rosenberg (Alt.)
National Council of Acoustical Consultants	E. Logsdon
.....	G.E. Winzer (Alt.)
National Institute of Standards & Technology (NIST)	D.J. Evans
.....	R.P. Wagner (Alt.)
PCB Group	K. Cox
.....	L. Harbaugh (Alt.)
Scantek, Inc.	S. Marshall
.....	M. Golden (Alt.)
U.S. Air Force (USAF)	R.L. McKinley
.....	F. Mobley (Alt.)
U.S. Army Construction Engineering Research Laboratory	M. Swearingen
.....	M.J. White (Alt.)
U.S. Army Research Laboratory, Human Research and Engineering Directorate	A. Scharine
.....	M. Ericson (Alt.)
U.S. Department of Labor – Mine Safety and Health Administration	J.P. Seiler
.....	J. Homer (Alt.)

Individual Experts of Accredited Standards Committee S1, Acoustics, were:

V. Buzduga
P. Hanes

W.W. Lang
V. Nedzelnitsky
P.D. Schomer

C. Walber
L. Wu

Working Group S1/WG 17, Sound Level Meters and Integrating Sound Level Meters, which assisted Accredited Standards Committee S1, Acoustics, in the development of this standard, had the following membership.

G.R. Stephany, Chair

M. Alexander
R.J. Anderson
P.J. Battenberg

B.M. Brooks
K. Cox
J.J. Earshen

P. Hanes
R.J. Peppin
J.P. Seiler

Suggestions for improvements to this standard will be welcomed. They should be sent to Accredited Standards Committee S1, Acoustics, in care of the Standards Secretariat of the Acoustical Society of America, 1305 Walt Whitman Road, Suite 300, Melville, New York 11747. Telephone: 631-390-0215; FAX: 631-923-2875; E-mail: asastds@acousticalsociety.org.

Introduction

U.S. NOTE: The following introductory text describes the differences between the current version of IEC 61672-1 and the 2002 version; it does NOT present the differences between this version and the previous version of ANSI S1.4-1983, ANSI S1.4A-1985, or ANSI S1.43-1997.

For assessments of conformance to performance specifications, this second edition of IEC 61672-1 uses different criteria than were used for the 2002 first edition.

In the period from 1961 to 1985, International Standards for sound level meters did not provide any requirements or recommendations to account for the uncertainty of measurement in assessments of conformance to specifications.

This absence of requirements or recommendations to account for uncertainty of measurement created ambiguity in determinations of conformance to specifications for situations where a measured deviation from a design goal was close to a limit of the allowed deviation. If conformance was determined based on whether a measured deviation did or did not exceed the limits, the end-user of the sound level meter incurred the risk that the true deviation from a design goal exceeded the limits.

To remove this ambiguity, IEC Technical Committee 29, at its meeting in 1996, adopted a policy to account for measurement uncertainty in assessments of conformance in International Standards that it prepares.

The first edition (2002) of IEC 61672-1 accounted for measurement uncertainty by giving two explicit criteria for determining conformance to the specifications. The two criteria were (a) that measured deviations from design goals, extended by the expanded uncertainty of measurement, do not exceed the applicable tolerance limits and (b) that the expanded uncertainty of measurement does not exceed agreed-upon maximum values. For most performance specifications, the tolerance limits were calculated essentially by extending the allowances for design and manufacturing from the 1979 and 1985 International Standards for sound level meters by the applicable maximum-permitted expanded uncertainties of measurement. Tolerance limits were intended to represent the limits for true deviations from design goals with a coverage probability of 95 %.

This second edition of IEC 61672-1 uses an amended criterion for assessing conformance to a specification. Conformance is demonstrated when (a) measured deviations from design goals do not exceed the applicable *acceptance limits* and (b) the uncertainty of measurement does not exceed the corresponding maximum-permitted uncertainty. Acceptance limits are analogous to the allowances for design and manufacturing implied in the first edition (2002) of IEC 61672-1. Actual and maximum-permitted uncertainties are determined for a coverage probability of 95 %. The amended criterion for assessing conformance does not necessitate any change to the design of a sound level meter in order to conform to the specifications of this International Standard.

The maximum-permitted uncertainties of measurement are not equivalent to the uncertainties associated with the measurement of a sound level. The uncertainty of a measured sound level is evaluated from the anticipated deviations of the electroacoustical performance of the sound level meter from the relevant design goals as well as estimates of the uncertainties associated with the specific measurement situation. Unless more-specific information is available, the evaluation of the contribution of a specific sound level meter to a total measurement uncertainty can be based on the acceptance limits and maximum-permitted uncertainties specified in this standard.

American National Standard

Electroacoustics – Sound level meters – Part 1: Specifications

1 Scope

This part of ANSI/ASA S1.4 / IEC 61672 gives electroacoustical performance specifications for three kinds of sound measuring instruments:

- a time-weighting sound level meter that measures exponential-time-weighted, frequency-weighted sound levels;
- an integrating-averaging sound level meter that measures time-averaged, frequency-weighted sound levels; and
- an integrating sound level meter that measures frequency-weighted sound exposure levels.

Sound level meters conforming to the requirements of this standard have a specified frequency response for sound incident on the microphone from one principal direction in an acoustic free field or successively from random directions.

Sound level meters specified in this standard are intended to measure sounds generally in the range of human hearing.

NOTE The AU frequency weighting specified in IEC 61012 can be applied for measurements of A-weighted sound levels of audible sound in the presence of a source that contains spectral components at frequencies greater than 20 kHz.¹

Two performance categories, class 1 and class 2, are specified in this standard. In general, specifications for class 1 and class 2 sound level meters have the same design goals and differ mainly in the acceptance limits and the range of operational temperature. Acceptance limits for class 2 are greater than, or equal to, those for class 1.

This standard is applicable to a range of designs for sound level meters. A sound level meter may be a self-contained hand-held instrument with an attached microphone and a built-in display device. A sound level meter may be comprised of separate components in one or more enclosures and may be capable of displaying a variety of acoustical signal levels. Sound level meters may include extensive analogue or digital signal processing, separately or in combination, with multiple analogue and digital outputs. Sound level meters may include general-purpose computers, recorders, printers, and other devices that form a necessary part of the complete instrument.

¹ IEC 61012, *Filters for the measurement of audible sound in the presence of ultrasound*.

Sound level meters may be designed for use with an operator present or for automatic and continuous measurements of sound level without an operator present. Specifications in this standard for the response to sound waves apply without an operator present in the sound field.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60942, *Electroacoustics – Sound calibrators*

IEC 61000-4-2:2008, *Electromagnetic compatibility (EMC) – Part 4-2: Testing and measurement techniques – Electrostatic discharge immunity test*

IEC 61000-6-2:2005, *Electromagnetic compatibility (EMC) – Part 6-2: Generic standards – Immunity for industrial environments*

IEC 61094-6, *Measurement microphones – Part 6: Electrostatic actuators for determination of frequency response*

IEC 61183, *Electroacoustics – Random-incidence and diffuse-field calibration of sound level meters*

IEC 62585, *Electroacoustics – Methods to determine corrections to obtain the free-field response of a sound level meter*

ISO/IEC Guide 98-4:2012, *Evaluation of measurement data – The role of measurement uncertainty in conformance assessment*

ISO/IEC Guide 99, *International vocabulary of metrology – Basic and general concepts and associated terms* (VIM)

CISPR 16-1-1:2010, *Specification for radio disturbance and immunity measuring apparatus and methods – Part 1-1: Radio disturbance and immunity measuring apparatus – Measuring apparatus*²

Amendment 1:2010

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO/IEC Guide 98-4, ISO/IEC Guide 99, and IEC 61000-6-2, as well as the following apply.

NOTE All quantities are expressed in SI units.

² CISPR = *International Special Committee on Radio Interference*.

3.1**sound pressure**

difference between an instantaneous total pressure and the corresponding static pressure

Note 1 to entry: Sound pressure is expressed in pascals (Pa).

3.2**sound pressure level**

ten times the logarithm to the base 10 of the ratio of the time-mean-square of a sound-pressure signal to the square of the reference value

Note 1 to entry: Sound pressure level is expressed in decibels (dB).

Note 2 to entry: The reference value is 20 μ Pa.

3.3**frequency weighting**

difference, as a specified function of frequency, between the level of the frequency-weighted signal indicated on the display device and the corresponding level of a constant-amplitude sinusoidal input signal

Note 1 to entry: Level difference is expressed in decibels (dB).

3.4**time weighting**

exponential function of time, of a specified time constant, that weights the square of a sound-pressure signal

3.5**sound level****frequency-weighted sound pressure level**

level with time weighting or time averaging of the square of a frequency-weighted sound-pressure signal

Note 1 to entry: Sound level is expressed in decibels (dB).

3.6**time-weighted sound level**

ten times the logarithm to the base 10 of the ratio of the running time average of the time-weighted square of a frequency-weighted sound-pressure signal to the square of the reference value

Note 1 to entry: Time-weighted sound level is expressed in decibels (dB).

Note 2 to entry: For time-weighted sound level, example letter symbols are L_{AF} , L_{AS} , L_{CF} , and L_{CS} for frequency weightings A and C and time weightings F and S.


Note 3 to entry: In symbols and as an example, A-weighted and F-time-weighted sound level $L_{AF}(t)$ at observation time t can be represented by

$$L_{AF}(t) = 10 \lg \left[\frac{(1/\tau_F) \int_{-\infty}^t p_A^2(\xi) e^{-(t-\xi)/\tau_F} d\xi}{p_0^2} \right] \text{ dB} \quad (1)$$

where

- τ_F is the exponential time constant in seconds for the F time weighting;
- ξ is a dummy variable of time integration from some time in the past, as indicated by $-\infty$ for the lower limit of the integral, to the time of observation t ;
- $p_A(\xi)$ is the A-weighted instantaneous sound-pressure signal; and
- p_0 is the reference value of 20 μPa .

Note 4 to entry: The sketch in Figure 1 illustrates the process indicated by Equation (1).



IEC 2243/13

Key

- a Start with a frequency-weighted electrical input signal
- b Square the input signal
- c Apply a low-pass filter with one real pole at $-1/\tau$ (exponential time weighting)
- d Take the base-10 logarithm
- e Display the result in decibels with the square of a reference value of 20 μPa

Figure 1 – Principal steps involved in forming a time-weighted sound level

3.7

maximum time-weighted sound level

greatest time-weighted sound level within a stated time interval

Note 1 to entry: Maximum time-weighted sound level is expressed in decibels (dB).

Note 2 to entry: Example letter symbols for maximum time-weighted sound level are L_{AFmax} , L_{ASmax} , L_{CFmax} , and L_{CSmax} for frequency weightings A and C and time weightings F and S.

3.8

peak sound pressure

greatest sound pressure (positive or negative) during a stated time interval

Note 1 to entry: Peak sound pressure is expressed in pascals (Pa).

Note 2 to entry: A peak sound pressure can arise from a positive or negative instantaneous sound pressure.

3.9

peak sound level

ten times the logarithm to the base 10 of the ratio of the square of a frequency-weighted peak sound-pressure signal to the square of the reference value

Note 1 to entry: Peak sound level is expressed in decibels (dB).

Note 2 to entry: The reference value is 20 μPa .

3.10

time-averaged sound level

equivalent continuous sound level

ten times the logarithm to the base 10 of the ratio of the time average of the square of a frequency-weighted sound-pressure signal during a stated time interval to the square of the reference value

Note 1 to entry: Time-averaged or equivalent continuous sound level is expressed in decibels (dB).

Note 2 to entry: In symbols and as an example, time-averaged, A-weighted sound level $L_{\text{Aeq},T}$ is given by

$$L_{\text{Aeq},T} = 10 \lg \left[\frac{(1/T) \int_{t-T}^t p_A^2(\xi) d\xi}{p_0^2} \right] \text{ dB} \quad (2)$$

where

- ξ is a dummy variable of time integration over the averaging time interval ending at the time of observation t ;
- T is the averaging time interval;
- $p_A(\xi)$ is the A-weighted sound-pressure signal; and
- p_0 is the reference value of 20 μPa .

Note 3 to entry: In principle, time weighting is not involved in a determination of time-averaged sound level.

3.11

sound exposure

time integral of the square of a frequency-weighted sound-pressure signal over a stated time interval or event of stated duration

Note 1 to entry: Duration of integration is included implicitly in the time integral and is not always reported explicitly, although it is useful to state the nature of the event. For measurements of sound exposure over a specified time interval, duration of integration is usually reported and indicated by a suitable subscript to the letter symbol, for example as $E_{\text{A},1\text{h}}$.

Note 2 to entry: In symbols and as an example, A-weighted sound exposure $E_{\text{A},T}$ is represented by

$$E_{\text{A},T} = \int_{t_1}^{t_2} p_A^2(t) dt \quad (3)$$

where $p_A^2(t)$ is the square of the A-weighted sound-pressure signal during integration time T starting at t_1 and ending at t_2 .

Note 3 to entry: The unit of sound exposure is pascal-squared seconds (Pa^2s) if sound pressure is in pascals and running time is in seconds.

Note 4 to entry: For applications such as measurement of exposure to noise in the workplace, sound exposure in pascal-squared hours is more convenient than pascal-squared seconds.

3.12**sound exposure level**

ten times the logarithm to the base 10 of the ratio of a sound exposure to the reference value

Note 1 to entry: Sound exposure level is expressed in decibels (dB).

Note 2 to entry: In symbols and as an example, A-weighted sound exposure level $L_{AE,T}$ is related to the corresponding time-averaged, A-weighted sound level $L_{Aeq,T}$ by

$$L_{AE,T} = 10 \lg \left[\frac{\int_{t_1}^{t_2} p_A^2(t) dt}{p_0^2 T_0} \right] \text{ dB} = 10 \lg \left(\frac{E_{A,T}}{E_0} \right) \text{ dB} = L_{Aeq,T} + 10 \lg \left(\frac{T}{T_0} \right) \text{ dB} \quad (4)$$

where

- $E_{A,T}$ is the A-weighted sound exposure in pascal-squared seconds over time interval T (see Equation (3));
- E_0 is the reference value given by $p_0^2 T_0 = (20 \mu\text{Pa})^2 \times (1 \text{ s}) = 400 \times 10^{-12} \text{ Pa}^2\text{s}$;
- T is the measurement time interval, in seconds, starting at t_1 and ending at t_2 , and
- T_0 is the reference value of 1 s for sound exposure level.

Note 3 to entry: Time-averaged, A-weighted sound level $L_{Aeq,T}$ during averaging time interval T is related to the corresponding A-weighted sound exposure $E_{A,T}$ or the A-weighted sound exposure level $L_{AE,T}$ occurring within that interval by

$$E_{A,T} = p_0^2 T \left(10^{0.1 L_{Aeq,T}} \right) \quad (5)$$

or

$$L_{Aeq,T} = 10 \lg \left(\frac{E_{A,T}}{p_0^2 T} \right) \text{ dB} = L_{AE,T} - 10 \lg \left(\frac{T}{T_0} \right) \text{ dB} \quad (6)$$

3.13**microphone**

electroacoustic transducer by which electrical signals are obtained from acoustic oscillations

[SOURCE: IEC 60050-801:1994, definition 801-26-01]

3.14**microphone reference point**

point specified on, or close to, the microphone to describe the position of the microphone

Note 1 to entry: The microphone reference point can be at the center of the diaphragm of the microphone.

3.15**reference direction**

inward direction toward the microphone reference point and specified for determining the directional response and frequency weighting of a sound level meter

Note 1 to entry: The reference direction can be specified with respect to an axis of symmetry.

3.16**sound-incidence angle**

angle between the reference direction and a line between the acoustic center of a sound source and the microphone reference point

Note 1 to entry: Sound-incidence angle is expressed in degrees.

3.17**relative directional response**

for any frequency weighting and at any frequency of incident sinusoidal sounds, and in a specified plane containing the principal axis of the microphone, sound level indicated at a given sound-incidence angle minus the sound level indicated for sound at the same frequency from the same source and incident from the reference direction

Note 1 to entry: Relative directional response is expressed in decibels.

3.18**directivity factor**

for a sound level meter, a measure of the deviation from an ideal directional response with equal sensitivity at all possible angles of sound incidence on the microphone

Note 1 to entry: Directivity factor is non-dimensional.

3.19**directivity index**

ten times the base-ten logarithm of the directivity factor

Note 1 to entry: Directivity index is expressed in decibels.

3.20**relative frequency-weighted free-field response**

for a given frequency, time-weighted or time-averaged, frequency-weighted sound level indicated by a sound level meter in response to plane progressive sinusoidal sound incident on the microphone from the reference direction minus the corresponding time-weighted or time-averaged sound level present at the position of the microphone reference point for the sound level meter and from the same sound source but in the absence of the sound level meter

Note 1 to entry: Relative frequency-weighted free-field response is expressed in decibels (dB).

Note 2 to entry: Relative frequency-weighted free-field response is called free-field sensitivity level in IEC 61183.

3.21**relative frequency-weighted random-incidence response**

for a given frequency, time-averaged, frequency-weighted sound level indicated by a sound level meter in response to random-incidence sound minus the time-averaged sound pressure level present at the position of the microphone reference point for the sound level meter and from the same sound source but in the absence of the sound level meter

Note 1 to entry: Relative frequency-weighted random-incidence response is expressed in decibels (dB).

Note 2 to entry: Relative frequency-weighted random-incidence response is called random-incidence sensitivity level in IEC 61183.