
AMERICAN NATIONAL STANDARD

**Electroacoustics – Sound Level Meters –
Part 2: Pattern Evaluation Tests
(a nationally adopted international standard)**

ANSI/ASA S1.4-2014/Part 2 /
IEC 61672-2:2013

Accredited Standards Committee S1, Acoustics

Standards Secretariat
Acoustical Society of America
1305 Walt Whitman Road, Suite 300
Melville, NY 11747

The American National Standards Institute, Inc. (ANSI) is the national coordinator of voluntary standards development and the clearinghouse in the U.S.A. for information on national and international standards.

The Acoustical Society of America (ASA) is an organization of scientists and engineers formed in 1929 to increase and diffuse the knowledge of acoustics and to promote its practical applications.



Electroacoustics – Sound Level Meters – Part 2: Pattern Evaluation Tests (a nationally adopted international standard)

Secretariat:

Acoustical Society of America

Approved on July 21, 2014 by:

American National Standards Institute, Inc.

Abstract

ANSI/ASA S1.4-2014/Part 2 / IEC 61672-2:2013 provides details of the tests necessary to verify conformance to all mandatory specifications given in ANSI/ASA S1.4-2014/Part 1 / IEC 61672-1:2013 for time-weighting sound level meters, integrating-averaging sound level meters, and integrating sound level meters. Pattern-evaluation tests apply for each channel of a multi-channel sound level meter, as necessary. Tests and test methods are applicable to class 1 and class 2 sound level meters. The aim is to ensure that all laboratories use consistent methods to perform pattern-evaluation tests. This edition constitutes a technical revision; the main technical changes with regard to the previous edition concern conformance to specifications which is now demonstrated when measured deviations from design goals do not exceed the applicable acceptance limits, and when the uncertainty of measurement does not exceed the corresponding maximum permitted uncertainty, with both uncertainties determined for a coverage probability of 95 %. In this document, references to IEC 61672-1, IEC 61672-2, and IEC 61672-3 refer to the second editions of the IEC standards unless stated otherwise. They refer equally to the ANSI/ASA S1.4-2014 / IEC 61672:2013 editions.

Procedures for the pattern-evaluation testing of sound level meters designed to conform to the specifications of IEC 61672-1:2002 were given in IEC 61672-2:2003.

9.15	Overload indication	32
9.16	C-weighted peak sound level.....	33
9.17	Reset.....	34
9.18	Electrical output	34
9.19	Timing facilities	34
9.20	Crosstalk in multi-channel sound level meter systems.....	34
9.21	Power supply	34
10	Pattern evaluation report.....	35

Foreword

[This Foreword is for information only, and is not a part of the American National Standard ANSI/ASA S1.4-2014/Part 2 / IEC 61672-2:2013 American National Standard Electroacoustics – Sound Level Meters – Part 2: Pattern Evaluation Tests. As such, this Foreword may contain material that has not been subjected to public review or a consensus process. In addition, it does not contain requirements necessary for conformance to the standard.]

This standard comprises a part of a group of definitions, standards, and specifications for use in acoustics. It was developed and approved by Accredited Standards Committee S1 Acoustics, under its approved operating procedures. Those procedures have been accredited by the American National Standards Institute (ANSI). The Scope of Accredited Standards Committee S1 is as follows:

Standards, specifications, methods of measurement and test, and terminology in the field of physical acoustics, including architectural acoustics, electroacoustics, sonics and ultrasonics, and underwater sound, but excluding those aspects which pertain to biological safety, tolerances, and comfort.

This standard is identical to IEC 61672-2:2013, which was prepared by IEC Technical Committee 29, in cooperation with the International Organization of Legal Metrology (OIML). However, in conformance with ANSI and IEC rules, the words “this part of ANSI/ASA S1.4 / IEC 61672” replace the words “this part of IEC 61672” where they appear in the IEC document, decimal points were substituted in place of the decimal commas used in IEC documents, and American English spelling is used in place of British English spelling.

The ANSI/ASA equivalents for the ISO standards in the IEC 61672 series and other referenced nationally adopted standards are given below:

- ANSI/ASA S1.4-2014/Part 1 / IEC 61672-1:2013 is an identical national adoption of IEC 61672-1:2013.
- ANSI/ASA S1.4-2014/Part 2 / IEC 61672-2:2013 is an identical national adoption of IEC 61672-2:2013.
- ANSI/ASA S1.4-2014/Part 3 / IEC 61672-3:2013 is an identical national adoption of IEC 61672-3:2013.

This nationally adopted international standard is one of three standards that, taken together, revise and replace ANSI/ASA S1.4-1983 (R2006), ANSI S1.4A-1985 (R2006) and ANSI S1.43-1997 (R2007). The subject matter in this document (Part 2) was not part of the superseded American National Standards.

U.S. NOTE: The following introductory text describes the differences between the current version of IEC 61672-2 and the 2003 version; it does NOT present the differences between this version and the previous version of ANSI S1.4-1983, ANSI S1.4A-1985, or ANSI S1.43-1997.

This second edition cancels and replaces the first edition published in 2003. This second edition constitutes a technical revision.

The main technical changes with regard to the previous edition are as follows:

In this second edition, conformance to specifications is demonstrated when:

- a) measured deviations from design goals do not exceed the applicable acceptance limits, and
- b) the uncertainty of measurement does not exceed the corresponding maximum permitted uncertainty, with both uncertainties determined for a coverage probability of 95 %.

At the time this Standard was submitted to Accredited Standards Committee S1, Acoustics for approval, the membership was as follows:

R.J. Peppin, *Chair*
A. Scharine, *Vice-Chair*

S.B. Blaeser, *Secretary*

3M Personal Safety Division, Detection Solutions	P.J. Battenberg
.....	M. Wurm (Alt.)
Acoustical Society of America	R.J. Peppin
Air-Conditioning, Heating and Refrigeration Institute	S.J. Lind
.....	D. Abbate (Alt.)
American Industrial Hygiene Association	D. Driscoll
.....	D. Sandfort (Alt.)
Campanella Associates	A.J. Campanella
Casella USA	R. Brauch
ETS-Lindgren Acoustic Systems	D. Winker
.....	M. Black (Alt.)
G.R.A.S. Sound & Vibration	J. Soendergaard
.....	B. Schustrich (Alt.)
Information Technology Industry Council	W.M. Beltman
.....	J. Rosenberg (Alt.)
National Council of Acoustical Consultants	E. Logsdon
.....	G.E. Winzer (Alt.)
National Institute of Standards & Technology (NIST)	D.J. Evans
.....	R.P. Wagner (Alt.)
PCB Group	K. Cox
.....	L. Harbaugh (Alt.)
Scantek, Inc.	S. Marshall
.....	M. Golden (Alt.)
U.S. Air Force (USAF)	R.L. McKinley
.....	F. Mobley (Alt.)
U.S. Army Construction Engineering Research Laboratory	M. Swearingen
.....	M.J. White (Alt.)
U.S. Army Research Laboratory, Human Research and Engineering Directorate	A. Scharine
.....	M. Ericson (Alt.)
U.S. Department of Labor – Mine Safety and Health Administration	J.P. Seiler
.....	J. Homer (Alt.)

Individual Experts of Accredited Standards Committee S1, Acoustics, were:

V. Buzduga
P. Hanes

W.W. Lang
V. Nedzelnitsky
P.D. Schomer

C. Walber
L. Wu

Working Group S1/WG 17, Sound Level Meters and Integrating Sound Level Meters, which assisted Accredited Standards Committee S1, Acoustics, in the development of this standard, had the following membership:

G.R. Stephany, Chair

M. Alexander
R.J. Anderson
P.J. Battenberg

B.M. Brooks
K. Cox
J.J. Earshen

P. Hanes
R.J. Peppin
J.P. Seiler

Suggestions for improvements to this standard will be welcomed. They should be sent to Accredited Standards Committee S1, Acoustics, in care of the Standards Secretariat of the Acoustical Society of America, 1305 Walt Whitman Road, Suite 300, Melville, New York 11747. Telephone: 631-390-0215; FAX: 631-923-2875; E-mail: asastds@acousticalsociety.org.

American National Standard

Electroacoustics – Sound Level Meters – Part 2: Pattern Evaluation Tests

1 Scope

This part of ANSI/ASA S1.4/IEC 61672 provides details of the tests necessary to verify conformance to all mandatory specifications given in IEC 61672-1 for time-weighting sound level meters, integrating-averaging sound level meters, and integrating sound level meters. Pattern-evaluation tests apply for each channel of a multi-channel sound level meter, as necessary. Tests and test methods are applicable to class 1 and class 2 sound level meters. The aim is to ensure that all laboratories use consistent methods to perform pattern-evaluation tests.

NOTE 1 In this document, references to IEC 61672-1, IEC 61672-2, and IEC 61672-3 refer to the second editions unless stated otherwise.

U.S. INFORMATIVE NOTE: In this document, references to IEC 61672-1, IEC 61672-2, and IEC 61672-3 refer to the three parts of ANSI/ASA S1.4-2014 / IEC 61672:2013. See the list provided in the Foreword.

NOTE 2 Procedures for the pattern-evaluation testing of sound level meters designed to conform to the specifications of IEC 61672-1:2002 were given in IEC 61672-2:2003.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60942, *Electroacoustics – Sound calibrators*

IEC 61000-4-2:2008, *Electromagnetic compatibility (EMC) – Part 4-2: Testing and measurement techniques – Electrostatic discharge immunity test*

IEC 61000-4-3:2010, *Electromagnetic compatibility (EMC) – Part 4-3: Testing and measurement techniques – Radiated, radio-frequency, electromagnetic-field immunity test*

IEC 61000-4-6:2008, *Electromagnetic compatibility (EMC) – Part 4-6: Testing and measurement techniques – Immunity to conducted disturbances, induced by radio-frequency fields*

IEC 61000-6-2:2005, *Electromagnetic compatibility (EMC) – Part 6-2: Generic standards – Immunity for industrial environments*

IEC 61094-1, *Measurement microphones – Part 1: Specifications for laboratory standard microphones*

IEC 61094-5, *Measurement microphones – Part 5: Methods for pressure calibration of working standard microphones by comparison*

IEC 61183, *Electroacoustics – Random-incidence and diffuse-field calibration of sound level meters*

IEC 61672-1, *Electroacoustics – Sound level meters – Part 1: Specifications*

IEC 62585, *Electroacoustics – Methods to determine corrections to obtain the free-field response of a sound level meter*

CISPR 16-1-1, *Specification for radio disturbance and immunity measuring apparatus and methods – Part 1-1: Radio disturbance and immunity measuring apparatus – Measuring apparatus*¹

CISPR 16-1-2:2006, *Specification for radio disturbance and immunity measuring apparatus and methods – Part 1-2: Radio disturbance and immunity measuring apparatus – Ancillary equipment – Conducted disturbances*

CISPR 16-2-1:2010 (Ed. 2.1), *Specification for radio disturbance and immunity measuring apparatus and methods – Part 2-1: Methods of measurement of disturbances and immunity – Conducted disturbance measurements*

CISPR 16-2-3:2010 (Ed. 3.1), *Specification for radio disturbance and immunity measuring apparatus and methods – Part 2-3: Methods of measurement of disturbances and immunity – Radiated disturbance measurements*

CISPR 22:2008, *Information technology equipment – Radio disturbance characteristics – Limits and methods of measurement*

ISO/IEC Guide 98-3, *Uncertainty of measurement – Part 3: Guide to the expression of uncertainty in measurement* (GUM:1995)

ISO/IEC Guide 99, *International vocabulary of metrology – Basic and general concepts and associated terms* (VIM)

ISO 26101:2012, *Acoustics – Test methods for the qualification of free-field environments*

3 Terms and definitions

For the purposes of this document, in addition to the terms and definitions given in IEC 61672-1 and IEC 62585, the terms and definitions given in IEC 61000-4-2, IEC 61000-4-3, IEC 61000-4-6, ISO/IEC Guide 98-3, and ISO/IEC Guide 99 also apply.

4 Submission for testing

4.1 At least three specimens of the same pattern of sound level meter shall be submitted for pattern-evaluation testing. As a minimum, the laboratory shall select two of the specimens for testing. At least one of the two specimens shall then be tested fully according to the procedures of this Standard. The laboratory shall decide whether the full tests shall also be performed on the second specimen or whether additional limited testing is adequate to approve the pattern.

¹ In English, CISPR stands for International Special Committee on Radio Interference.

4.2 An Instruction Manual and all items or accessories that are identified in the Instruction Manual as integral components for the normal mode of operation shall be submitted along with the three sound level meters. Examples of additional items or accessories include a microphone extension device or cable and peripheral equipment.

4.3 If the manufacturer of the sound level meter supplies devices that are to be connected to the sound level meter by cables, then the devices and cables shall be submitted with the sound level meter.

4.4 A calibrated sound calibrator of a model specified in the Instruction Manual for the sound level meter shall be supplied with the sound level meter. An Instruction Manual for the sound calibrator shall also be provided. As required by IEC 61672-1, the model of the calibrator shall conform to the relevant specifications of IEC 60942 for the class of sound calibrator.

5 Marking of the sound level meter and information in the Instruction Manual

5.1 It shall be verified that the sound level meter is marked according to the requirements of IEC 61672-1.

5.2 It shall be verified that the Instruction Manual contains all the information that is required by IEC 61672-1, as relevant to the facilities provided by the sound level meter.

5.3 If the sound level meter does not conform to the requirements of 5.1 and 5.2, no pattern-evaluation tests shall be performed.

5.4 After completion of all tests, the information shall be reviewed to ensure that it is correct and that no applicable acceptance limits are exceeded.

6 Mandatory facilities and general requirements

6.1 No test specified in this part of ANSI/ASA S1.4 / IEC 61672 shall be omitted unless the sound level meter does not possess the facility described for the test. When the design of a sound level meter which has been pattern approved is changed and a new pattern approval is requested, then, at the discretion of the laboratory, it is not necessary to repeat those tests for electroacoustical performance characteristics that are not affected by the design change.

6.2 A time-weighting sound level meter shall be verified to be able to display A-frequency-weighted, F-time-weighted sound level and to be able to indicate overload and under-range conditions.

6.3 An integrating-averaging sound level meter shall be verified to be able to display A-frequency-weighted, time-averaged sound level and to be able to indicate overload and under-range conditions.

6.4 An integrating sound level meter shall be verified to be able to display A-frequency-weighted sound exposure level and to be able to indicate overload and under-range conditions.

6.5 All display devices for the sound level meter shall be verified to be able to display sound levels or sound exposure levels with the resolution required by IEC 61672-1. The range of the display shall be at least the minimum specified in IEC 61672-1.

6.6 If a sound level meter is capable of measuring maximum or peak sound levels, or both, it shall be verified that a "hold" feature is provided.

6.7 A class 1 sound level meter shall be verified to have frequency-weighting C.

6.8 If the sound level meter is capable of indicating C-weighted peak sound levels, it shall be verified that the capability to display C-weighted, time-weighted sound level or C-weighted, time-averaged sound level is also provided.

6.9 For sound level meters with multiple level ranges, it shall be verified that overlap of the level ranges conforms to the specifications of IEC 61672-1.

6.10 For sound level meters that can display more than one measurement quantity, it shall be verified that there is a means to ascertain the quantity that is being displayed.

6.11 If the sound level meter does not possess the mandatory facilities listed in 6.2 through 6.10, as applicable, the sound level meter does not conform to the specifications of IEC 61672-1 and no pattern-evaluation tests shall be performed.

6.12 For all pattern-evaluation tests, the configuration of the sound level meter, or the multi-channel sound level meter system, shall be as specified in the Instruction Manual for one of the normal modes of operation, including all required accessories. The configuration shall include a windscreen if a windscreen is an integral component for the normal mode of operation, or if the Instruction Manual states that the sound level meter conforms to the specifications of IEC 61672-1 with a windscreen installed around the microphone. The model of the windscreen shall be as stated in the Instruction Manual for use with the sound level meter. All configurations of the sound level meter that are stated in the Instruction Manual as conforming to the requirements of IEC 61672-1 shall be tested.

6.13 If the Instruction Manual states that the sound level meter conforms to the specifications of IEC 61672-1 with optional facilities installed, the configuration with the optional facilities installed shall also be tested to verify conformance to the relevant specifications.

6.14 If an electrical output is provided on the sound level meter and the laboratory intends to utilize the electrical output instead of the display device, the laboratory shall verify that changes in the levels of applied acoustical or electrical input signals produce changes in the signal levels indicated on the display device and at the electrical output that are in accordance with the specifications of IEC 61672-1. This requirement applies to each channel of a multi-channel system. Where multiple outputs are present, if an output is specified in the Instruction Manual for testing, that output should be used for pattern-evaluation tests.

6.15 For all tests, the sound level meter shall be powered from its preferred supply.

6.16 The sound level meter shall be allowed to reach equilibrium with the prevailing environmental conditions before switching on the power to perform a test.

6.17 Tests for conformance to the specifications for the effects of changes in environmental conditions preferably should be conducted before tests for conformance to the specifications for electroacoustical performance.

6.18 If the sound level meter has more than one signal-processing channel, pattern-evaluation tests shall be performed for each channel that utilizes unique signal processing techniques. For multi-channel systems with the same functional equivalence in all channels, the number of

channels to be tested may be less than the number of available channels, at the discretion of the laboratory. For a multi-channel system, the number of channels to be tested should be determined from consideration of a scenario for which there is an array of microphones supplying signals to each input with each channel processing the signals in an identical manner. Selection of how many and which channels to test should consider differences, as described in the Instruction Manual, in the implementation of signal-processing techniques in the various channels. If any special procedure for testing the channels for identical functionality is described in the Instruction Manual, that procedure should be followed.

NOTE If the sound level meter is a multi-channel device (for example, a sound level meter with two or more separate signal inputs with non-parallel processing of digitized data by time-sharing, but quasi-parallel display for the displayed signals), it is usually possible to test the channels for identical functionality either by setting the functions of the channels for identical processing and reading the display(s) or by allowing the channel functionality to rotate by a special test setting procedure thereby allowing comparison of display(s).

6.19 Conformance to a performance specification is demonstrated when the following criteria are both satisfied: (a) a measured deviation from a design goal does not exceed the applicable acceptance limit and (b) the corresponding uncertainty of measurement does not exceed the corresponding maximum-permitted uncertainty of measurement given in IEC 61672-1 for the same coverage probability of 95%. IEC 61672-1 gives example assessments of conformance using these criteria.

6.20 The laboratory shall use instruments with current calibrations for the appropriate quantities. The calibrations shall be traceable to national standards, as required.

6.21 Laboratories performing pattern evaluation tests shall calculate all uncertainties of measurements in accordance with the guidelines given in the ISO/IEC Guide 98-3. Actual measurement uncertainties shall be calculated for a coverage probability of 95 %. Calculation of the actual measurement uncertainty for a particular test should consider at least the following components, as applicable.

- The uncertainty attributed to calibration of the individual instruments and equipment used to perform the test, including the sound calibrator, where applicable;
- The uncertainty resulting from environmental effects or corrections;
- The uncertainty resulting from small errors that may be present in the applied signals;
- The uncertainty attributed to effects associated with the repeatability of the results of the measurements. When a laboratory is only required to perform a single measurement, it is necessary for the laboratory to make an estimate of the random contribution to the total measurement uncertainty. The estimate should be determined from an earlier evaluation of several measurements of the performance of similar sound level meters;
- The uncertainty associated with the resolution of the display device of the sound level meter under test. For digital display devices that indicate signal levels with a resolution of 0.1 dB, the uncertainty component should be taken as a rectangular distribution with semi-range of 0.05 dB;
- The uncertainty associated with the device used to mount the sound level meter in the free-field test facility;

- The uncertainty resulting from the deviation of the sound field in the free-field test facility from an ideal free sound field; and
- The uncertainty associated with each correction applied to the measurement data.

6.22 If the uncertainty of measurement exceeds the maximum-permitted uncertainty of measurement, the result of the test shall not be used to demonstrate conformance to a specification and pattern approval shall not be granted.

6.23 As appropriate, the laboratory shall utilize the recommendations given in the Instruction Manual for performing the pattern-evaluation tests.

7 Environmental, electrostatic, and radio-frequency tests

7.1 General

7.1.1 Before conducting, but not during, the tests described in the Clause 7, the indication of A-weighted sound level at the calibration check frequency shall be checked by application of the sound calibrator specified in 4.4. If necessary, the sound level meter shall be adjusted to indicate the required sound level under reference environmental conditions. For multi-channel devices, the corresponding indications shall be checked for all channels selected for testing.

7.1.2 Environmental conditions at the time of checking the indication shall be recorded.

7.1.3 The effect of environmental conditions on the sound pressure level produced by the sound calibrator shall be accounted for in accordance with the procedure in the Instruction Manual for the sound calibrator and data from its calibration. The effects shall be evaluated relative to the sound pressure level produced under reference conditions.

7.1.4 For environmental tests, a sound calibrator shall be used to provide a known sound pressure level at the microphone of the sound level meter. For class 1 sound level meters, the calibrator shall conform to either the class LS or class 1 specifications of IEC 60942. For class 2 sound level meters, the calibrator shall conform to either class LS, class 1, or class 2 specifications of IEC 60942. If the sound calibrator conforms to the requirements of the applicable performance class for a nominal frequency of 1 kHz, the environmental tests shall be performed at the nominal frequency of 1 kHz. The effects of static pressure, air temperature, and relative humidity on the sound pressure level produced in the coupler of the sound calibrator, over the range of environmental conditions specified for the tests, shall be known.

NOTE The range of environmental conditions specified for pattern-evaluation tests exceeds the range specified in IEC 60942 for class LS sound calibrators.

7.1.5 The sound level meter shall be set to perform a typical measurement of time-weighted sound level, time-averaged sound level, or sound exposure level on the reference level range. The frequency weighting shall be set to A weighting.

7.1.6 Time-weighted sound levels, time-averaged sound levels, or sound exposure levels indicated by the sound level meter in response to the signal from the sound calibrator shall be recorded for each test condition. When necessary, time-averaged sound levels shall be calculated from the indications of sound exposure level and elapsed time in accordance with IEC 61672-1. Averaging times for time-averaged sound levels or integration times for sound exposure levels shall be recorded.