
AMERICAN NATIONAL STANDARD

Acoustics – Determination of sound power levels and sound energy levels of noise sources using sound pressure – Precision methods for anechoic rooms and hemi-anechoic rooms

(a Nationally Adopted International Standard)

ANSI/ASA S12.55-2012 / ISO 3745:2012

Accredited Standards Committee S12, Noise

Standards Secretariat
Acoustical Society of America
35 Pinelawn Road, Suite 114 E
Melville, NY 11747-3177

The American National Standards Institute, Inc. (ANSI) is the national coordinator of voluntary standards development and the clearinghouse in the U.S.A. for information on national and international standards.

The Acoustical Society of America (ASA) is an organization of scientists and engineers formed in 1929 to increase and diffuse the knowledge of acoustics and to promote its practical applications.



AMERICAN NATIONAL STANDARD

**Acoustics – Determination of sound power levels and
sound energy levels of noise sources using sound
pressure – Precision methods for anechoic rooms and
hemi-anechoic rooms**

(a nationally adopted international standard)

Secretariat:

Acoustical Society of America

Approved on August 23, 2012 by:

American National Standards Institute, Inc.

Abstract

This American National Standard specifies methods for measuring the sound pressure levels on a measurement surface enveloping a noise source (machinery or equipment) in an anechoic room or a hemi-anechoic room. The sound power level (or, in the case of impulsive or transient noise emission, the sound energy level) produced by the noise source, in frequency bands of width one-third octave or with frequency weighting A applied, is calculated using those measurements, including corrections to allow for any differences between the meteorological conditions at the time and place of the test and those corresponding to a reference characteristic acoustic impedance.

AMERICAN NATIONAL STANDARDS ON ACOUSTICS

The Acoustical Society of America (ASA) provides the Secretariat for Accredited Standards Committees S1 on Acoustics, S2 on Mechanical Vibration and Shock, S3 on Bioacoustics, S3/SC 1 on Animal Bioacoustics, and S12 on Noise. These committees have wide representation from the technical community (manufacturers, consumers, trade associations, organizations with a general interest, and government representatives). The standards are published by the Acoustical Society of America as American National Standards after approval by their respective Standards Committees and the American National Standards Institute (ANSI).

These standards are developed and published as a public service to provide standards useful to the public, industry, and consumers, and to Federal, State, and local governments.

Each of the Accredited Standards Committees (operating in accordance with procedures approved by ANSI) is responsible for developing, voting upon, and maintaining or revising its own Standards. The ASA Standards Secretariat administers Committee organization and activity and provides liaison between the Accredited Standards Committees and ANSI. After the Standards have been produced and adopted by the Accredited Standards Committees, and approved as American National Standards by ANSI, the ASA Standards Secretariat arranges for their publication and distribution.

An American National Standard implies a consensus of those substantially concerned with its scope and provisions. Consensus is established when, in the judgment of the ANSI Board of Standards Review, substantial agreement has been reached by directly and materially affected interests. Substantial agreement means much more than a simple majority, but not necessarily unanimity. Consensus requires that all views and objections be considered and that a concerted effort be made towards their resolution.

The use of an American National Standard is completely voluntary. Their existence does not in any respect preclude anyone, whether he or she has approved the Standards or not, from manufacturing, marketing, purchasing, or using products, processes, or procedures not conforming to the Standards.

NOTICE: This American National Standard may be revised or withdrawn at any time. The procedures of the American National Standards Institute require that action be taken periodically to reaffirm, revise, or withdraw this Standard.



Acoustical Society of America
ASA Secretariat
35 Pinelawn Road, Suite 114E
Melville, New York 11747-3177
Telephone: 1 (631) 390-0215
Fax: 1 (631) 390-0217
E-mail: asastds@aip.org

© 2012 by Acoustical Society of America. This standard may not be reproduced in whole or in part in any form for sale, promotion, or any commercial purpose, or any purpose not falling within the provisions of the U.S. Copyright Act of 1976, without prior written permission of the publisher. For permission, address a request to the Standards Secretariat of the Acoustical Society of America.

These materials are subject to copyright claims of ISO, IEC, ANSI, and ASA. No part of this publication may be reproduced in any form, including an electronic retrieval system, without the prior written permission of the Acoustical Society of America (ASA). All requests pertaining to this standard should be submitted to the Acoustical Society of America (ASA).

Contents

1	Scope	1
1.1	General	1
1.2	Types of noise and noise sources	1
1.3	Test room	1
1.4	Measurement uncertainty	1
2	Normative references	1
3	Terms and definitions	2
4	Reference meteorological conditions	8
5	Test rooms	8
5.1	Acoustic criterion for adequacy of the test room	8
5.2	Criteria for background noise	9
5.3	Criterion for air temperature	10
6	Instrumentation	12
6.1	Instruments for acoustical measurements	12
6.2	Instruments for meteorological measurements	12
7	Definition, location, installation and operation of noise source under test	13
7.1	General	13
7.2	Auxiliary equipment	13
7.3	Noise source location	13
7.4	Mounting of the noise source	13
7.5	Operation of source during test	14
8	Measurement surface	15
8.1	Spherical measurement surface for use in an anechoic room	15
8.2	Hemispherical measurement surface for use in a hemi-anechoic room	15
9	Determination of sound power levels and sound energy levels	17
9.1	Measurements in the test room	17
9.2	Measurement of meteorological conditions	17
9.3	Microphone positions	17
9.4	Determination of sound power levels of a noise source which emits steady or non-steady noise	19
9.5	Determination of sound energy levels for a noise source which emits impulsive noise	23
9.6	Calculation of directivity indices	26
9.7	Calculation of surface sound pressure level non-uniformity index	26
9.8	Frequency-weighted sound power level and sound energy level	27
10	Measurement uncertainty	27
10.1	Methodology	27
10.2	Determination of σ_{omc}	28
10.3	Determination of σ_{R0}	28
10.4	Typical upper bound values of σ_{R0}	29
10.5	Total standard deviation, σ_{tot} , and expanded measurement uncertainty, U	30
11	Information to be recorded	31
11.1	General	31

11.2	Noise source under test.....	31
11.3	Test room.....	31
11.4	Instrumentation	31
11.5	Acoustical data	32
12	Test report	32
Annex A	(normative) General procedures for qualification of anechoic and hemi-anechoic rooms.....	33
A.1	General	33
A.2	Instrumentation and measuring equipment	33
A.3	Installation of test sources and microphones	36
A.4	Test procedures.....	37
A.5	Qualification procedure.....	39
Annex B	(normative) Qualification procedure for spaces within test rooms used in the determination of sound power levels and sound energy levels of specific noise sources.....	41
B.1	General	41
B.2	Properties of reflecting plane	41
B.3	Procedure using two measurement spheres or hemispheres with different radii (two-surface method)	42
Annex C	(normative) Calculation of A-weighted sound power levels and A-weighted sound energy levels from one-third-octave band levels.....	43
C.1	A-weighted sound power levels.....	43
C.2	A-weighted sound energy levels.....	43
C.3	Values of j and C_j for use in calculations	43
Annex D	(normative) Array of microphone positions on a spherical measurement surface in a free field.....	45
Annex E	(normative) Arrays of microphone positions on a hemispherical measurement surface in a hemi-free field	48
Annex F	(normative) Coaxial circular paths of microphones on a hemispherical measurement surface in a hemi-free field	53
Annex G	(normative) Meridional paths of microphones on a hemispherical measurement surface in a hemi-free field	54
Annex H	(normative) Spiral paths of microphones on a hemispherical measurement surface in a hemi-free field.....	56
Annex I	(informative) Guidance on the development of information on measurement uncertainty	57
I.1	General	57
I.2	Considerations on the total standard deviation, σ_{tot}	57
I.3	Considerations on σ_{omc}	57
I.4	Considerations on σ_{R0}	59
I.5	Combined standard uncertainty.....	66
I.6	Measurement uncertainty based on reproducibility data.....	66
I.7	Example of an uncertainty budget for a reference sound source in a national laboratory	67

Figures

Figure 1 — Reference box, center for the measurement surface (when using the geometric center of the source) and characteristic source dimensions for application in an anechoic and a hemi-anechoic room	16
Figure A.1 — Microphone positions on the spherical surface for test sound source directivity measurements	35
Figure D.1 — Microphone positions on the spherical measurement surface	47
Figure E.1 — Microphone positions on the hemispherical measurement surface (general case)	51
Figure E.2 — Microphone positions on the hemispherical measurement surface for broadband omnidirectional sources	52
Figure F.1 — Coaxial circular paths for a moving microphone	53
Figure G.1 — Meridional paths for a moving microphone	54
Figure G.2 — Example of a mechanical system to realize a meridional path	55
Figure G.3 — Example of an electronic control circuit.....	55
Figure H.1 — Spiral path for a moving microphone.....	56

Tables

Table 1 — Maximum background noise levels in test room for absolute criteria	11
Table 2 — Typical upper bound values of the standard deviation of reproducibility of the method, σ_{R0} , for sound power levels and sound energy levels determined in a hemi-anechoic room in accordance with this American National Standard	30
Table 3 — Typical upper bound values of the standard deviation of reproducibility of the method, σ_{R0} , for the sound power levels and sound energy levels determined in an anechoic room, in accordance with this American National Standard	30
Table A.1 — Allowable deviation in directivity of the test source.....	34
Table A.2 — Maximum allowable deviation of measured sound pressure levels from theoretical levels using the inverse square law	40
Table C.1 — Values of j and C_j for mid-band frequencies of one-third-octave bands.....	44
Table D.1 — Microphone positions.....	46
Table E.1 — Microphone positions (general case).....	49
Table E.2 — Microphone positions for a broadband omnidirectional source	50

Table I.1 — Examples of calculated total standard deviations, σ_{tot} , for three different cases 59

Table I.2 — Example of an uncertainty budget for determinations of σ_{R0} for sound power level and sound energy level, valid for frequencies from 500 Hz to 4 kHz, or for A-weighted measurements of a source with a relatively flat frequency spectrum..... 62

Foreword

[This Foreword is for information only, and is not a part of the American National Standard ANSI/ASA S12.55-2012 / ISO 3745:2012 American National Standard Acoustics — Determination of sound power levels and sound energy levels of noise sources using sound pressure — Precision methods for anechoic rooms and hemi-anechoic rooms.]

This standard comprises a part of a group of definitions, standards, and specifications for use in noise. It was developed and approved by Accredited Standards Committee S12, Noise, under its approved operating procedures. Those procedures have been accredited by the American National Standards Institute (ANSI). The Scope of Accredited Standards Committee S12 is as follows:

Standards, specifications, and terminology in the field of acoustical noise pertaining to methods of measurement, evaluation, and control, including biological safety, tolerance, and comfort, and physical acoustics as related to environmental and occupational noise.

This standard is a revision of ANSI S12.55-2006 / ISO 3745:2003, which has been technically revised.

This Standard is identical to International Standard ISO 3745, Acoustics — Determination of sound power levels and sound energy levels of noise sources using sound pressure — Precision methods for anechoic rooms and hemi-anechoic rooms, which was prepared by Technical Committee ISO/TC 43 Subcommittee SC 1, Noise. However, in conformance with ANSI and ISO rules, the words "American National Standard" replace the words "International Standard" where they appear in the ISO document, decimal points were substituted in place of the decimal commas used in ISO documents, and American English spelling is used in place of British English spelling. In addition, editorial correction was made to the caption of Figure E.2. In ISO 3745:2012 this figure and the previous one were numbered E.1.

The ANSI or ANSI/ASA equivalents for the ISO standards in the ISO 3740 series and other referenced nationally adopted standards are given below:

- ANSI/ASA S12.5 / ISO 6926 is an identical national adoption of ISO 6926;
- ANSI/ASA S12.50/ISO 3740 is an identical national adoption of ISO 3740;
- ANSI/ASA S12.51/ISO 3741 is an identical national adoption of ISO 3741;
- ANSI/ASA S12.53/Part 1/ISO 3743-1 is an identical national adoption of ISO 3743-1;
- ANSI/ASA S12.53/Part 2/ISO 3743-2 is an identical national adoption of ISO 3743-2;
- ANSI/ASA S12.54/ISO 3744 is an identical national adoption of ISO 3744;
- ANSI/ASA S12.55/ISO 3745 is an identical national adoption of ISO 3745;
- ANSI/ASA S12.56/ISO 3746 is an identical national adoption of ISO 3746; and
- ANSI/ASA S12.57/ISO 3747 is an identical national adoption of ISO 3747.

At the time this Standard was submitted to Accredited Standards Committee S12, Noise for approval, the membership was as follows:

W.J. Murphy, *Chair*
S.J. Lind, *Vice-Chair*

S.B. Blaeser, *Secretary*

3M Occupational Health & Environmental Safety Division E.H. Berger

Acoustical Society of America R.D. Hellweg
..... D. Lubman (Alt.)

Air-Conditioning, Heating and Refrigeration Institute	S. Lind D. Abbate (Alt.)
Air Movement & Control Association, Inc.	J.A. Brooks M. Stevens (Alt.)
American Academy of Audiology	D. Ostergren S. Gordon-Salant (Alt.)
American Academy of Otolaryngology - Head and Neck Surgery	R.A. Dobie L.A. Michael (Alt.)
American Industrial Hygiene Association	D. Driscoll S.N. Hacker (Alt.)
American Speech-Language-Hearing Association	L.A. Wilber V. Gladstone (Alt.)
Caterpillar, Inc.	K.G. Meitl
Compressed Air and Gas Institute	R.C. Johnson D.R. Bookshar (Alt.)
Council for Accreditation in Occupational Hearing Conservation	VACANT L.D. Hager (Alt.)
Emerson Electric – Copeland Corporation	A.T. Herfat G. Williamson (Alt.)
ETS – Lindgren Acoustic Systems	D. Winker M. Black (Alt.)
ExxonMobil	B. Moulton
G.R.A.S. Sound & Vibration	J. Soendergaard B. Schustrich (Alt.)
GE Aviation	R. Majjigi R. Babbitt (Alt.)
General Motors	D.B. Moore
Harley-Davidson Motor Company	D. Jaeger
Information Technology Industry Council	W.M. Beltman J. Rosenberg (Alt.)
Institute of Noise Control Engineering	R.J. Peppin B. Tinianov (Alt.)
International Safety Equipment Association	J. Birkner C.Z. Fargo (Alt.)
John Deere	K.B. Washburn K. Cone (Alt.)
National Council of Acoustical Consultants	J. Erdreich G.E. Winzer (Alt.)
National Hearing Conservation Association	P.A. Brogan

National Institute for Occupational Safety and Health	W.J. Murphy E. Zechmann (Alt.)
National Institute of Standards and Technology – National Voluntary Laboratory Accreditation Program	B.A. Sandoval
National Park Service	G.R. Stanley K. Fristrup (Alt.)
Noise Control Engineering, Inc.	M.A. Bahtiarian R. Fischer (Alt.)
Noise Pollution Clearinghouse	L. Blomberg
PCB Group	K. Cox L. Harbaugh (Alt.)
Power Tool Institute, Inc.	W.D. Spencer M. Hickok (Alt.)
Schomer and Associates, Inc.	P.D. Schomer
Sperian Hearing Protection, LLC	R. Ghent B. Witt (Alt.)
U.S. Air Force	R.L. McKinley H. Gallagher (Alt.)
U.S. Army Aeromedical Research Lab	W.A. Ahroon
U.S. Army Construction Engineering Research Laboratory	M.J. White M. Swearingen (Alt.)
U.S. Army Public Health Command	C. Stewart M. Grantham (Alt.)
U.S. Army Research Laboratory, Human Research and Engineering Directorate	J. Gaston M.S. Binseel (Alt.)
U.S. Department of Transportation	A. Konheim
U.S. Naval Surface Warfare Center - Carderock	M. Craun
Université du Québec ETS	J. Voix F. Laville (Alt.)

Individual Experts of Accredited Standards Committee S12, Noise, were:

P.K. Baade	W.W. Lang	J. Schmitt
E.H. Berger	D. Lubman	P.D. Schomer
B.M. Brooks	D. Michaud	L.C. Sutherland
A.J. Campanella	N.P. Miller	W.R. Thornton
L.S. Finegold	W.J. Murphy	L.A. Wilber
R.D. Godfrey	M.A. Nobile	G.E. Winzer
R.D. Hellweg	R.J. Peppin	

Working Group S12/WG 23, *Determination of Sound Power*, which assisted Accredited Standards Committee S12, Noise, in the development of this standard, had the following membership:

B.M. Brooks, Co-Chair
J. Schmitt, Co-Chair

M.C. Black
S. Bly
A.J. Campanella
D.A. Collings
K. Cunefare

R.D. Hellweg
A.T. Herfat
S. Keith
G.C. Maling

M.A. Nobile
M. O'Connell
P.D. Schomer
J. Weinstein

Suggestions for improvements of this standard will be welcomed. They should be sent to Accredited Standards Committee S12, Noise, in care of the Standards Secretariat of the Acoustical Society of America, 35 Pinelawn Road, Suite 114E, Melville, New York 11747-3177. Telephone: 631-390-0215; FAX: 631-390-0217; E-mail: asastds@aip.org.

Introduction

This American National Standard is a national adoption of one of the series ISO 3741^[3] to ISO 3747^[8], which specify various methods for determining the sound power levels and sound energy levels of noise sources including machinery, equipment and their sub-assemblies. The selection of one of the methods from the series for use in a particular application depends on the purpose of the test to determine the sound power level or sound energy level and on the facilities available. General guidelines to assist in the selection are provided in ISO 3740^[2]. ISO 3741^[3] to ISO 3747^[8] give only general principles regarding the operating and mounting conditions of the machinery or equipment for the purposes of the test. It is important that test codes be established for individual kinds of noise source, in order to give detailed requirements on mounting, loading and operating conditions under which the sound power levels or sound energy levels are to be obtained and to select the appropriate measurement surface and microphone array from among those specified in this American National Standard.

The methods given in this American National Standard require the source to be mounted in either an anechoic room or a hemi-anechoic room having specified acoustical characteristics. The methods are then based on the premise that the sound power or sound energy of the source is directly proportional to the mean-square sound pressure over a hypothetical measurement surface enclosing the source and otherwise depends on the physical constants of air.

The methods specified in this American National Standard permit the determination of the sound power level and the sound energy level in frequency bands and/or with frequency A-weighting applied.

The methods give a precision grade of accuracy (grade 1) as defined in ISO 12001. The resulting sound power levels and sound energy levels include corrections to allow for any differences that might exist between the meteorological conditions under which the tests are conducted and reference meteorological conditions. For applications where there are large uncertainties due to operating conditions or where reduced accuracy is acceptable, reference can be made to the more practical methods of ISO 3744^[6] or ISO 3746^[7]. Guidance on evaluation of measurement uncertainty is given in Annex I.

AMERICAN NATIONAL STANDARD

Acoustics – Determination of sound power levels and sound energy levels of noise sources using sound pressure – Precision methods for anechoic rooms and hemi-anechoic rooms (a nationally adopted international standard)

1 Scope

1.1 General

This American National Standard specifies methods for measuring the sound pressure levels on a measurement surface enveloping a noise source (machinery or equipment) in an anechoic room or a hemi-anechoic room. The sound power level (or, in the case of impulsive or transient noise emission, the sound energy level) produced by the noise source, in frequency bands of width one-third octave or with frequency weighting A applied, is calculated using those measurements, including corrections to allow for any differences between the meteorological conditions at the time and place of the test and those corresponding to a reference characteristic acoustic impedance.

In general, the frequency range of interest includes the one-third-octave bands with mid-band frequencies from 100 Hz to 10,000 Hz. In practice, the range is extended or restricted to frequencies beyond or within these limits, to those between which the test room is qualified for the purposes of the measurements.

1.2 Types of noise and noise sources

The methods specified in this American National Standard are suitable for all types of noise (steady, non-steady, fluctuating, isolated bursts of sound energy, etc.) defined in ISO 12001.

The noise source under test can be a device, machine, component or sub-assembly. The maximum size of the noise source depends on specified requirements regarding the radius of the hypothetical sphere or hemisphere used as the enveloping measurement surface.

1.3 Test room

The test rooms that are applicable for measurements made in accordance with this American National Standard are an anechoic room or hemi-anechoic room, also called, respectively, a free-field test room or hemi-free-field test room.

1.4 Measurement uncertainty

Information is given on the uncertainty of the sound power levels and sound energy levels determined in accordance with this American National Standard, for measurements made in limited bands of frequency and with frequency weighting A applied. The uncertainty conforms to ISO 12001:1996, accuracy grade 1 (precision grade).

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 5725 (all parts), *Accuracy (trueness and precision) of measurement methods and results*

ISO 9613-1:1993, *Acoustics — Attenuation of sound during propagation outdoors — Part 1: Calculation of the absorption of sound by the atmosphere*

ISO 12001:1996, *Acoustics — Noise emitted by machinery and equipment — Rules for the drafting and presentation of a noise test code*

ISO/IEC Guide 98-3, *Uncertainty in measurement — Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)*

IEC 60942:2003, *Electroacoustics — Sound calibrators*

IEC 61183, *Electroacoustics — Random-incidence and diffuse-field calibration of sound level meters*

IEC 61260:1995 + AM1:2001, *Electroacoustics — Octave-band and fractional-octave-band filters*

IEC 61672-1:2002, *Electroacoustics — Sound level meters — Part 1: Specifications.*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

3.1

sound pressure

p

difference between instantaneous pressure and static pressure

NOTE 1 Adapted from ISO 80000-8:2007^[22], 8-9.2.

NOTE 2 Sound pressure is expressed in pascals.

3.2

sound pressure level

L_p

ten times the logarithm to the base 10 of the ratio of the square of the sound pressure, p , to the square of a reference value, p_0 , expressed in decibels

$$L_p = 10 \lg \frac{p^2}{p_0^2} \text{ dB} \quad (1)$$

where the reference value, p_0 , is 20 μPa

[ISO/TR 25417:2007^[20], 2.2]

NOTE 1 If specific frequency and time weightings as specified in IEC 61672-1 and/or specific frequency bands are applied, this should be indicated by appropriate subscripts; e.g. L_{pA} denotes the A-weighted sound pressure level.

NOTE 2 This definition is technically in accordance with ISO 80000-8:2007^[22], 8-22.

3.3 time-averaged sound pressure level

$L_{p,T}$

ten times the logarithm to the base 10 of the ratio of the time average of the square of the sound pressure, p , during a stated time interval of duration, T (starting at t_1 and ending at t_2), to the square of a reference value, p_0 , expressed in decibels

$$L_{p,T} = 10 \lg \left[\frac{\frac{1}{T} \int_{t_1}^{t_2} p^2(t) dt}{p_0^2} \right] \text{ dB} \quad (2)$$

where the reference value, p_0 , is 20 μPa

NOTE 1 In general, the subscript “ T ” is omitted since time-averaged sound pressure levels are necessarily determined over a certain measurement time interval.

NOTE 2 Time-averaged sound pressure levels are often A-weighted, in which case they are denoted by $L_{pA,T}$, which is usually abbreviated to L_{pA} .

NOTE 3 Adapted from ISO/TR 25417:2007^[20], 2.3.

3.4 single event time-integrated sound pressure level

L_E

ten times the logarithm to the base 10 of the ratio of the integral of the square of the sound pressure, p , of an isolated single sound event (burst of sound or transient sound) over a stated time interval T (starting at t_1 and ending at t_2) to a reference value, E_0 , expressed in decibels

$$L_E = 10 \lg \left[\frac{\int_{t_1}^{t_2} p^2(t) dt}{E_0} \right] \text{ dB} \quad (3)$$

where the reference value, E_0 , is $(20 \mu\text{Pa})^2 \text{ s} = 4 \times 10^{-10} \text{ Pa}^2 \text{ s}$

[ISO 3741:2010^[3], 3.4]

NOTE 1 This quantity can be obtained by $L_{p,T} + 10 \lg(T/T_0)$ dB, where $T_0 = 1 \text{ s}$.

NOTE 2 When used to measure sound immission (see ISO 11690-1^[19]), this quantity is usually called “sound exposure level” (see ISO/TR 25417:2007^[20], 2.7).

3.5 measurement time interval

T

portion or a multiple of an operational period or operational cycle of the noise source under test for which the time-averaged sound pressure level is determined

NOTE Measurement time interval is expressed in seconds.

[ISO 3741:2010^[3], 3.5]

3.6

free sound field

sound field in a homogeneous, isotropic medium free of boundaries

NOTE In practice, a free sound field is a field in which the influence of reflections at the boundaries or other disturbing objects is negligible over the frequency range of interest.

[ISO/TR 25417:2007^[20], 2.17]

3.7

anechoic room

anechoic test room

free-field test room

test room in which a free sound field is obtained

3.8

free sound field over a reflecting plane

free sound field in the half-space above an infinite reflecting plane in the absence of any other obstacles

3.9

reflecting plane

sound reflecting planar surface on which the noise source under test is located

3.10

hemi-anechoic room

hemi-anechoic test room

hemi-free-field test room

test room in which a free sound field over a reflecting plane is obtained

3.11

frequency range of interest

for general purposes, the frequency range of one-third-octave bands with nominal mid-band frequencies from 100 Hz to 10,000 Hz

NOTE For special purposes, the frequency range may be extended or reduced, provided that the test environment and instrument specifications are satisfactory for use over the modified frequency range. Changes to the frequency range of interest should be made clear in the test report. For sources in which the A-weighted sound power levels are determined by sound at predominantly high or low frequencies, the frequency range of interest should be extended to include these frequencies.

3.12

measurement radius

r

radius of a spherical or hemispherical measurement surface

NOTE Measurement radius is expressed in meters.

3.13**measurement surface**

hypothetical surface of area, S , on which the microphone positions are located at which the sound pressure levels are measured, enveloping the noise source under test and, in the case of a hemi-anechoic room, terminating on the reflecting plane on which the source is located

NOTE The measurement surface area is expressed in meters squared.

3.14**characteristic source dimension**

d_0

distance from the origin of the co-ordinate system to the farthest corner of the reference box, where the reference box is defined as a hypothetical rectangular parallelepiped that just encloses the source including all the significant sound radiating components and any test table on which the source may be mounted; in the case of a hemi-anechoic room the reference box terminates on the reflecting plane

NOTE 1 Characteristic source dimension is expressed in meters.

NOTE 2 For illustration see Figure 1.

3.15**background noise**

noise from all sources other than the noise source under test

NOTE Background noise includes contributions from airborne sound, noise from structure-borne vibration, and electrical noise in the instrumentation.

3.16**background noise correction**

K_1

correction applied to each of the measured sound pressure levels on the measurement surface to account for the influence of background noise

NOTE 1 Background noise correction is expressed in decibels.

NOTE 2 The background noise correction is frequency dependent; the correction in the case of a frequency band is denoted K_{1f} , where f denotes the relevant mid-band frequency; in the case of A-weighting, the quantity is denoted K_{1A} .

3.17**surface time-averaged sound pressure level**

$\overline{L_p}$

mean (energy average) of the time-averaged sound pressure levels at all the microphone positions, or traverses, on the measurement surface, with the background noise corrections, K_1 , applied at each microphone position or traverse

$$\overline{L_p} = 10 \lg \left[\frac{\sum_{i=1}^{N_M} 10^{0.1 L_{pi}(\text{ST})}}{N_M} \right] \text{dB} \quad (4)$$

where

$L_{pi(ST)}$ is the background noise corrected time-averaged sound pressure level for the i th microphone position or traverse on the measurement surface, with the noise source under test in operation, in decibels;

N_M is the number of microphone positions or traverses.

NOTE Surface time-averaged sound pressure level is expressed in decibels.

3.18

surface single event time-integrated sound pressure level

$\overline{L_E}$

mean (energy average) of the single event time-integrated sound pressure levels at all the microphone positions on the measurement surface, with the background noise correction, K_1 , applied at each microphone position

$$\overline{L_E} = 10 \lg \left[\frac{\sum_{i=1}^{N_M} 10^{0.1 L_{Ei(ST)}}}{N_M} \right] \text{ dB} \quad (5)$$

where

$L_{Ei(ST)}$ is the background noise corrected single event time-integrated sound pressure level for the i th microphone position on the measurement surface, with the noise source under test in operation, in decibels;

N_M is the number of microphone positions.

NOTE Surface single event time-integrated sound pressure level is expressed in decibels.

3.19

sound power

P

through a surface, product of the sound pressure, p , and the component of the particle velocity, u_n , at a point on the surface in the direction normal to the surface, integrated over that surface

[ISO 80000-8:2007^[22], 8-16 reproduced in ISO/TR 25417:2007^[20], 2.8]

NOTE 1 Sound power is expressed in watts.

NOTE 2 The quantity relates to the rate per time at which airborne sound energy is radiated by a source.

3.20

sound power level

L_W

ten times the logarithm to the base 10 of the ratio of the sound power of a source, P , to a reference value, P_0 , expressed in decibels

$$L_W = 10 \lg \frac{P}{P_0} \text{ dB} \quad (6)$$

where the reference value, P_0 , is 1 pW

NOTE 1 If a specific frequency weighting as specified in IEC 61672-1 and/or specific frequency bands are applied, this should be indicated by appropriate subscripts; e.g. $L_{w,A}$ denotes the A-weighted sound power level.

NOTE 2 This definition is technically in accordance with ISO 80000-8:2007^[22], 8-23.

[ISO/TR 25417:2007^[20], 2.9]

3.21 sound energy

J

integral of the sound power, P , over a stated time interval of duration T (starting at t_1 and ending at t_2)

$$J = \int_{t_1}^{t_2} P(t) dt \quad (7)$$

NOTE 1 Sound energy is expressed in joules.

NOTE 2 The quantity is particularly relevant for non-stationary, intermittent sound events.

[ISO/TR 25417:2007^[20], 2.10]

3.22 sound energy level

L_J

ten times the logarithm to the base 10 of the ratio of the sound energy, J , to a reference value, J_0 , expressed in decibels

$$L_J = 10 \lg \frac{J}{J_0} \text{ dB} \quad (8)$$

where the reference value, J_0 , is 1 pJ

NOTE If a specific frequency weighting as specified in IEC 61672-1 and/or specific frequency bands are applied, this should be indicated by appropriate subscripts; e.g. $L_{J,A}$ denotes the A-weighted sound energy level.

[ISO/TR 25417:2007^[20], 2.11]

3.23 directivity index

D_{li}

measure of the extent to which a noise source under test radiates sound in the direction of the i th microphone position on a measurement surface, relative to the mean sound radiation over the measurement surface

$$D_{li} = L_{pi} - \overline{L_p} \quad (9)$$

where

L_{pi} is the background noise corrected sound pressure level (either time-averaged or single event time-integrated) for the i th microphone position on the measurement surface, with the noise source under test in operation, in decibels;

$\overline{L_p}$ is the surface sound pressure level (either time-averaged or single event time-integrated), in decibels.

3.24

surface sound pressure level non-uniformity index

V_I

measure of the variability of measured sound pressure levels over the measurement surface

$$V_I = \sqrt{\frac{1}{(N_M - 1)} \sum_{i=1}^{N_M} (L_{pi} - L_{pav})^2} \quad (10)$$

where

L_{pi} is the background noise corrected sound pressure level (either time-averaged or single event time-integrated) for the i th microphone position on the measurement surface with the noise source under test in operation, in decibels;

L_{pav} is the arithmetic average of the background noise corrected sound pressure levels (either time-averaged or single event time-integrated) over all the microphone positions on the measurement surface for the noise source under test, in decibels;

N_M is the number of microphone positions.

NOTE When V_I is determined at a particular measurement radius, the quantity is denoted $V_{I,r}$.

4 Reference meteorological conditions

Reference meteorological conditions for the purpose of calculating the sound power level and sound energy level, corresponding to a reference characteristic acoustic impedance of air $\rho c = 411.5 \text{ N s/m}^3$ (where ρ is the density of air and c is the speed of sound) are:

- a) air temperature: 23.0 °C;
- b) static pressure: 101.325 kPa;
- c) relative humidity: 50 %.

5 Test rooms

5.1 Acoustic criterion for adequacy of the test room

Anechoic or hemi-anechoic rooms that are applicable for measurements in accordance with this American National Standard either satisfy:

- a) Annex A over the frequency range of interest, for use in general purpose measurements; or

- b) Annex B over the frequency range of interest, for determination of sound power levels of a specific noise source.

Annex A and Annex B specify procedures for determining the extent of deviations of the test room from the ideal free-field condition or the ideal hemi-free-field condition, and criteria are given to assess the adequacy of the test room. Qualification procedures for the test room shall be in accordance with Annex A or Annex B.

For sources in which the A-weighted sound levels are determined by sound at predominantly high or low frequencies, outside the nominal frequency range of interest (see 3.11), the frequency range of interest shall be extended to include these frequencies, and this shall be clearly stated in the test report.

NOTE If it is necessary to make measurements in test rooms or spaces within test rooms where the requirements of Annex A or Annex B are exceeded, see ISO 3744^[6], ISO 3746^[7], ISO 9614-1^[15] or ISO 9614-2^[16].

5.2 Criteria for background noise

5.2.1 Relative criteria

5.2.1.1 General

The difference between the level of the background noise and that of the noise source under test (when measured in the presence of this background noise) averaged (see 9.4.3) over all microphone positions or traverses, shall be at least 6 dB for all frequency bands, and for one-third-octave bands of mid-band frequency from 250 Hz to 5,000 Hz shall be at least 10 dB. If this requirement is met, the background noise criteria of this American National Standard are satisfied.

NOTE 1 The same criteria are applied to single event levels, where the measurement time interval used to measure the background noise is the same as the measurement time interval associated with the measurement of the single event.

NOTE 2 The noise associated with the microphone traversing mechanism, if one is used for the measurements, is considered to be part of the background noise. In such cases, the background noise should be measured with the traversing mechanism operating.

5.2.1.2 Frequency band measurements

The requirements of 5.2.1.1 may not be achievable in all frequency bands, even when the background noise levels in the test room are extremely low and well controlled. Therefore, for the purposes of determining compliance with the background noise criteria given in 5.2.1.1, any band may be excluded from the frequency range of interest if the A-weighted sound power level (see Annex C) of that band (after correcting for background noise) is at least 15 dB below the highest A-weighted sound power level in any frequency band.

5.2.1.3 A-weighted measurements

If the A-weighted sound power level or sound energy level is to be determined from frequency band levels and reported, the following steps shall be followed to determine whether this quantity meets the background noise criteria of this American National Standard:

- a) the A-weighted sound power level or sound energy level is computed in accordance with the procedures in this American National Standard using the data from every frequency band within the frequency range of interest;

- b) the computation is repeated, but excluding those bands for which $\Delta L_p < 6$ dB for one-third-octave bands of mid-band frequency 200 Hz and below and 6,300 Hz and above, and for which $\Delta L_p < 10$ dB for one-third-octave bands of mid-band frequency from 250 Hz to 5,000 Hz.

If the difference between these two levels is less than 0.5 dB, the A-weighted sound power level or sound energy level determined from the data for all bands shall be considered as conforming to the background noise criteria of this American National Standard.

5.2.2 Absolute criteria

If it can be demonstrated that the background noise levels in the test room at the time of the measurements are less than or equal to those given in Table 1 for all bands within the frequency range of interest, the measurements can be taken as having met the background noise requirements of this American National Standard, even if the 6 dB and 10 dB requirements (see 5.2.1.1) are not met for all bands. It can be assumed that the source emits little or no measurable noise in these frequency bands, and that the data reported represent an upper bound to the sound power level or sound energy level in these bands.

If levels of noise measured at the smallest possible distance from the source (see Clause 8) are less than or equal to those given in Table 1, the frequency range of interest may be restricted to a contiguous range of frequencies that includes both the lowest and highest frequencies at which the sound pressure level from the noise source exceeds the corresponding value in Table 1. In such cases, the applicable frequency range of interest shall be reported.

NOTE The absolute background criteria are intended for general sound power measurements; they exceed the threshold of audibility and may not be suitable for all uses.

5.2.3 Statement of non-conformity with criteria

If the above background noise criteria are not satisfied, the report shall clearly state that the background noise requirements of this American National Standard have not been met, and, in the case of frequency band measurements, shall identify the particular frequency bands that do not meet the criteria. Furthermore, the report shall not state or imply that the measurements have been made "in full conformity" with ISO 3745.

5.3 Criterion for air temperature

The air temperature in the test room shall be within the range 15 °C to 30 °C.

NOTE The equations in Clause 9 used for the calculation of the sound power level and the sound energy level include a multiplication factor which is an approximation to cover a variety of possible sound-generating mechanisms (monopole, dipole, quadrupole, etc.) which pertain to different kinds of noise source; the range of air temperature is limited in order to guarantee a deviation smaller than 0.2 dB in the result.

Table 1 — Maximum background noise levels in test room for absolute criteria

One-third-octave mid-band frequency Hz	Maximum band sound pressure level dB
50	44
63	38
80	32
100	27
125	22
160	16
200	13
250	11
315	9
400	8
500	7
630	7
800	7
1,000	7
1,250	7
1,600	7
2,000	7
2,500	8
3,150	8
4,000	8
5,000	8
6,300	8
8,000	12
10,000	14
12,500	11
16,000	46
20,000	46

6 Instrumentation

6.1 Instruments for acoustical measurements

6.1.1 General

The instruments for measuring sound pressure levels, including microphone(s) as well as cable(s), windscreen(s), recording devices and other accessories, if used, shall meet the requirements of IEC 61672-1:2002, class 1. Filters shall meet the requirements of IEC 61260:1995, class 1.

The microphone shall be oriented so that the reference direction of the microphone (as specified in IEC 61672-1) is normal to the measurement surface.

6.1.2 Calibration

At the beginning and the end of every measurement session and at least at the beginning and the end of each measurement day, the entire sound pressure level measuring system shall be checked at one or more frequencies by means of a sound calibrator meeting the requirements of IEC 60942:2003, class 1. Without any further adjustment, the difference between the readings of two consecutive checks shall be less or equal to 0.3 dB. If this value is exceeded, the results of measurements obtained after the previous satisfactory check shall be discarded.

6.1.3 Verification

Compliance of the sound pressure level measuring instrument, the filters and the sound calibrator with the relevant requirements shall be verified by the existence of a valid certificate of compliance. If applicable, random incidence response of the microphone shall be verified by a procedure from IEC 61183. All compliance testing shall be conducted by a laboratory being accredited or otherwise nationally authorized to perform the relevant tests and calibrations and ensuring metrological traceability to the appropriate measurement standards.

Unless national regulations dictate otherwise, it is recommended that the sound calibrator should be calibrated at intervals not exceeding 1 year, the compliance of the instrumentation system with the requirements of IEC 61672-1 should be verified at intervals not exceeding 2 years, and the compliance of the filter set with the requirements of IEC 61260 should be verified at intervals not exceeding 2 years.

6.2 Instruments for meteorological measurements

6.2.1 General

The maximum permissible error for instruments used for meteorological measurements shall be

- a) ± 1 K for temperature measuring devices;
- b) ± 10 % for relative humidity measuring devices;
- c) ± 2 kPa for static pressure measuring devices.

NOTE It is not necessary to specify values for these uncertainties if their influence on the final measured value is determined to be within acceptable limits using ISO/IEC Guide 98-3.

6.2.2 Verification

For instruments used for monitoring meteorological conditions, e.g. to ensure that any of these parameters is within a specified interval, manufacturer's specifications are deemed sufficient to verify these requirements.

If any of the meteorological parameters has a direct influence on the measurement result, the compliance of the measuring instrument with the relevant requirement has to be verified by a valid calibration certificate from a laboratory being accredited or otherwise nationally authorized to perform the relevant calibration and ensuring metrological traceability to the appropriate measurement standards.

7 Definition, location, installation and operation of noise source under test

7.1 General

The manner in which the noise source under test is installed and operated may have a significant influence on the sound power or sound energy emitted by a noise source. This clause specifies conditions that are intended to minimize variations in the noise emission due to the installation and operating conditions of the noise source under test. Relevant instructions of a noise test code, if any exists for the family of machinery or equipment to which the noise source under test belongs, shall be followed. The same installation, mounting and operating conditions of the noise source under test shall be used for the determination of emission sound pressure levels and sound power levels.

Particularly for large machines, it is necessary to decide which components, sub-assemblies, auxiliary equipment, power sources; etc., belong to the noise source under test.

7.2 Auxiliary equipment

Care shall be taken to ensure that any electrical conduits, piping or air ducts connected to the noise source under test do not radiate significant amounts of sound energy into the test environment.

If possible, all auxiliary equipment necessary for the operation of the noise source under test that is not a part of it shall be located outside the test room. If this is impractical, care shall be taken to minimize any sound radiated into the test environment from such equipment. The noise source under test shall be taken to include all significant sources of sound emission, including auxiliary equipment which cannot either be removed or adequately quieted.

7.3 Noise source location

In locating the source within the test room, it is important to allow sufficient space around the source so that the measurement surface can envelop it in accordance with the requirements of Clause 8.

Detailed information on installation conditions shall be based upon the general requirements of this American National Standard and particular requirements of the relevant noise test code, if one exists.

7.4 Mounting of the noise source

7.4.1 General

In many cases, the sound power or sound energy emitted by a noise source under test depends upon the support or mounting conditions of the noise source. Whenever a typical mounting condition exists for a noise source, that condition shall be used or simulated, if practicable.