
AMERICAN NATIONAL STANDARD

Procedures for Testing Basic Vestibular Function

ANSI/ASA S3.45-2009

Accredited Standards Committee S3, Bioacoustics

Standards Secretariat
Acoustical Society of America
35 Pinelawn Road, Suite 114 E
Melville, NY 11747-3177

The American National Standards Institute, Inc. (ANSI) is the national coordinator of voluntary standards development and the clearinghouse in the U.S.A. for information on national and international standards.

The Acoustical Society of America (ASA) is an organization of scientists and engineers formed in 1929 to increase and diffuse the knowledge of acoustics and to promote its practical applications.



AMERICAN NATIONAL STANDARD

Procedures for Testing Basic Vestibular Function

Secretariat:

Acoustical Society of America

Approved January 8, 2009:

American National Standards Institute, Inc.

Abstract

This Standard defines procedures for performing and reporting a battery of tests for the evaluation of human vestibular function. Six different tests are specified. Stimuli are presented to evoke eye movement by a subject whose response is determined either by measurement of electrical signals generated by the eye movements or by image processing methods applied to video eye movements. The Standard specifies test procedures, measurements, data analysis, and data reporting requirements. These tests, including the data analysis and reporting procedures, are called the Basic Vestibular Function Test Battery. Test interpretation is not a part of this Standard.

AMERICAN NATIONAL STANDARDS ON ACOUSTICS

The Acoustical Society of America (ASA) provides the Secretariat for Accredited Standards Committees S1 on Acoustics, S2 on Mechanical Vibration and Shock, S3 on Bioacoustics, S3/SC 1 on Animal Bioacoustics, and S12 on Noise. These committees have wide representation from the technical community (manufacturers, consumers, trade associations, organizations with a general interest, and government representatives). The standards are published by the Acoustical Society of America as American National Standards after approval by their respective Standards Committees and the American National Standards Institute (ANSI).


These standards are developed and published as a public service to provide standards useful to the public, industry, and consumers, and to Federal, State, and local governments.

Each of the Accredited Standards Committees (operating in accordance with procedures approved by ANSI) is responsible for developing, voting upon, and maintaining or revising its own Standards. The ASA Standards Secretariat administers Committee organization and activity and provides liaison between the Accredited Standards Committees and ANSI. After the Standards have been produced and adopted by the Accredited Standards Committees, and approved as American National Standards by ANSI, the ASA Standards Secretariat arranges for their publication and distribution.

An American National Standard implies a consensus of those substantially concerned with its scope and provisions. Consensus is established when, in the judgment of the ANSI Board of Standards Review, substantial agreement has been reached by directly and materially affected interests. Substantial agreement means much more than a simple majority, but not necessarily unanimity. Consensus requires that all views and objections be considered and that a concerted effort be made towards their resolution.

The use of an American National Standard is completely voluntary. Their existence does not in any respect preclude anyone, whether he or she has approved the Standards or not, from manufacturing, marketing, purchasing, or using products, processes, or procedures not conforming to the Standards.

NOTICE: This American National Standard may be revised or withdrawn at any time. The procedures of the American National Standards Institute require that action be taken periodically to reaffirm, revise, or withdraw this Standard.



Acoustical Society of America
ASA Secretariat
35 Pinelawn Road, Suite 114E
Melville, New York 11747-3177
Telephone: 1 (631) 390-0215
Fax: 1 (631) 390-0217
E-mail: asastds@aip.org

© 2009 by Acoustical Society of America. This standard may not be reproduced in whole or in part in any form for sale, promotion, or any commercial purpose, or any purpose not falling within the provisions of the U.S. Copyright Act of 1976, without prior written permission of the publisher. For permission, address a request to the Standards Secretariat of the Acoustical Society of America.

Contents

1	Scope, purpose, and applications	1
1.1	Scope	1
1.2	Purpose.....	2
1.3	Applications.....	2
2	Normative references	2
3	Terms and definitions	2
4	Instrumentation.....	4
4.1	Overview of instrumentation	4
4.2	Component and setup features common to and needed for both methods of eye movement measurement.....	4
4.3	Components and setup features specific to electro-oculography.....	6
4.4	Videonystagmography and display.....	8
5	Spontaneous and gaze-evoked nystagmus.....	9
5.1	Test purpose	9
5.2	Spontaneous test.....	9
5.3	Gaze test.....	9
5.4	Data reporting	9
5.5	Measurement of slow component velocity.....	10
6	Saccade test.....	11
6.1	Test purpose	11
6.2	Visual target.....	11
6.3	Procedure	11
6.4	Evaluation	11
7	Pursuit test	11
7.1	Test purpose	11
7.2	Visual target.....	11
7.3	Procedure	12
7.4	Evaluation	12
8	Positioning and positional nystagmus	12
8.1	Test purpose	12
8.2	Dynamic (positioning) tests.....	12
8.3	Static (positional) tests.....	13
8.4	Results to be reported	14
9	Caloric test	14
9.1	Test purpose	14
9.2	Test environment	15
9.3	Data acquisition	15

9.4	General test method	15
9.5	Fixation suppression	16
9.6	Test details	16
9.7	Role of the tester	17
9.8	Analysis of test data	17
9.9	Data to be reported	18
10	Report of test results	20
10.1	Form of report	20
10.2	Standardized worksheet	20
	Annex A (informative) Test interpretation	22
	Annex B (informative) Air caloric stimulation	24
	Bibliography	25

Tables

Table 1	Positioning and positional maneuvers	13
Table 2	Standard caloric stimulus values	16

Figures

Figure 1	Example of electro-oculogram (EOG) calibration	6
Figure 2	Normal electrode placement	6
Figure 3	Estimate of slope of slow component of nystagmus	10
Figure 4	Start and end positions for example positioning (dynamic) and positional (static) maneuvers: A) Positioning, head straight while sitting, and head extended while supine; B) Positioning, head left; C) Positioning, supine, head left	14
Figure 5	Example of calculation of slow component velocity of nystagmus	18
Figure 6	Worksheet for the basic vestibular function test battery	21
Figure A.1	Sample test report – normal subject	22
Figure A.2	Sample test report – abnormal subject	23

Foreword

[This Foreword is for information only, and is not a part of the American National Standard ANSI/ASA S3.45-2009 American National Standard Procedures for Testing Basic Vestibular Function.]

This standard comprises a part of a group of definitions, standards, and specifications for use in bioacoustics. It was developed and approved by Accredited Standards Committee S3, Bioacoustics, under its approved operating procedures. Those procedures have been accredited by the American National Standards Institute (ANSI). The Scope of Accredited Standards Committee S3 is as follows:

Standards, specifications, methods of measurement and test, and terminology in the fields of psychological and physiological acoustics, including aspects of general acoustics which pertain to biological safety, tolerance and comfort.

This standard is a revision of ANSI S3.45-1999, which has been technically revised. An alternate means for the measurement of eye movements using video-oculography (VOG) has been added, since this method has advantages of being less invasive, and can be used to display and measure torsional eye movements that are not possible with electro-oculography.

A 5-year exception to the standard, which permitted use of older computerized equipment which sampled ENG signals at 50 samples per second, was eliminated since current technology is capable of supporting the 100-sample-per-second rate specified in the standard.

The issues that needed to be addressed before the use of air as a medium for caloric irrigations could be incorporated into the standard have been added in an informative annex.

One normative reference was added in Clause 2, and additional informative resources were added to the bibliography, primarily regarding the use of video-oculography.

This standard is not comparable to any existing ISO Standard.

At the time this Standard was submitted to Accredited Standards Committee S3, Bioacoustics, for approval, the membership was as follows:

C.A. Champlin, *Chair*

R.F. Burkard, *Vice-Chair*

S.B. Blaeser, *Secretary*

Acoustical Society of America C.A. Champlin
..... R.F. Burkard (Alt.)

American Academy of Audiology Y. Szymko-Bennett
..... D.A. Fabry (Alt.)

American Academy of Otolaryngology, Head and Neck Surgery, Inc. R.A. Dobie
..... L.A. Michael (Alt.)

American Industrial Hygiene Association	T.K. Madison
.....	D. Driscoll (Alt.)
American Speech-Hearing-Language Association (ASHA)	L.A. Wilber
.....	V. Gladstone (Alt.)
Beltone/GN Resound	S. Petrovic
Council for Accreditation in Occupational Hearing Conservation (CAOHC)	E.H. Berger
.....	J. Banach (Alt.)
Etymotic Research, Inc.	M.C. Killion
.....	R. Scicluna (Alt.)
Frye Electronics, Inc.	G.J. Frye
.....	K.E. Frye (Alt.)
Hearing Industries Association	T.A. Victorian
.....	C.M. Rogin (Alt.)
National Hearing Conservation Association	T. Schulz
National Institute for Occupational Safety and Health	M. Stephenson
.....	W.J. Murphy (Alt.)
National Institute of Standards and Technology	V. Nedzelnitsky
.....	R. Wagner (Alt.)
Quest Technologies, Inc.	M. Wurm
.....	P. Battenberg (Alt.)
Starkey Laboratories	D.A. Preves
.....	T. Burns (Alt.)
U.S. Air Force	R. McKinley
U.S. Army Aeromedical Research Lab	W. Ahroon
.....	Vacant (Alt.)
U.S. Army CERL	L. Pater
.....	D. Delaney (Alt.)
U.S. Army Research Laboratory, Human Research and Engineering Directorate	T.R. Letowski
.....	P. Henry (Alt.)

Individual Experts of the Accredited Standards Committee S3, Bioacoustics, were:

J.R. Bareham	T. Frank	P.D. Schomer
A.J. Brammer	K.D. Kryter	L.A. Wilber
A.J. Campanella	R. McKinley	W.A. Yost

Working Group S3/WG 82, Basic Vestibular Function Test Battery, which assisted Accredited Standards Committee S3, Bioacoustics, in the development of this standard, had the following membership.

C. Wall, III, Chair

R. Burkard
J.M. Furman
R. Miles

E.M. Monsell
N.T. Shepard

Suggestions for improvements of this standard will be welcomed. They should be sent to Accredited Standards Committee S3, Bioacoustics, in care of the Standards Secretariat of the Acoustical Society of America, 35 Pinelawn Road, Suite 114E, Melville, New York 11747-3177. Telephone: 631-390-0215; FAX: 631-390-0217; E-mail: asastds@aip.org.

Introduction

The vestibular system provides information related to a person's orientation in space using information from five organs in each inner ear. Three of these organs sense angular acceleration in approximately orthogonal planes, while two sense linear acceleration and gravity. This peripheral information is integrated with that of other senses, including vision and proprioception, to give information about the orientation of the individual in space and to allow for appropriate compensatory reflexive movements. These reflexes can adapt in ways that enhance a person's ability to cope with changing conditions.

Two major reflexes involved are the vestibulo-ocular reflex and the vestibulo-spinal reflex. The former allows one to keep visual objects of interest stable on the retina while the person is in motion. The latter allows one to maintain postural stability when subjected to motion. These reflexive movements can be used to evaluate the response to vestibular stimuli. This Standard utilizes the vestibulo-ocular reflex as its basis for test procedures.

A variety of test methods has been used in clinical and research evaluations of vestibular function. Some require relatively modest test equipment; others require relatively complex and expensive mechanical equipment. Lack of standardization in test protocols and data reporting often makes comparison of data between facilities and test conductors difficult. The goal of this Standard is to specify a battery of tests that use defined stimuli, data collection and recording methods, data analysis procedures, and data reporting requirements.

Depending on the system used to measure the movement in the eyes, the test battery is commonly known as "Electronystagmography" or "Videonystagmography." Electronystagmography (ENG) measures eye movements using the method of electro-oculography (EOG). In this method, skin-mounted electrodes measure changes in potentials across the eyeball due to movement of the eye in its socket. The corneal retinal potential is the source of the measured voltage. Videonystagmography (VNG) measures eye movements using video-oculography (VOG). In this method images of the eyes from video cameras are processed to give two dimensional (2D) or three dimensional (3D) eye position estimates. For VNG, 2D (horizontal and vertical) eye movements are typically provided, although some systems can also provide torsional eye movement estimates.

Use of the Standard will greatly facilitate comparison of data among different clinical settings as well as between these data and the data from research laboratories.

American National Standard

Procedures for Testing Basic Vestibular Function

1 Scope, purpose, and applications

1.1 Scope

1.1.1 General

This Standard specifies procedures for conducting six separate tests, which, together with the data analysis and reporting requirements specified in this Standard, constitute the Basic Vestibular Function Test Battery. The six tests cover:

- (a) spontaneous nystagmus (see 5.2);
- (b) gaze-evoked nystagmus (see 5.3);
- (c) saccade test (see 6);
- (d) pursuit testing (see 7);
- (e) positioning and positional nystagmus (see 8); and
- (f) caloric testing (see 9).

Each of these tests evaluates a response of the vestibular system by the vestibulo-ocular reflex through the measurement of electrical signals generated by eye movements recorded by electro-oculography.

1.1.2 Subjects

This Standard may be used to evaluate basic vestibular function in any human subject without restriction on age or sex.

NOTE Children under the age of about 4 may not be able to participate in all tests for various reasons, e.g., inability to follow verbal instructions.

Test subjects, however, must have functional retinas in both eyes and be able to see light through each eye separately to determine conjugacy of eye movements (i.e., ability to move both eyes synchronously or together).

Applicability of this Standard to some subjects may be limited by two variables. The test utilizing caloric stimulation depends upon heat transfer to the vestibular end organ, and is highly variable among individuals. The electro-oculogram, one method of measuring eye movement used in this Standard, depends on the existence of a corneoretinal potential which may be absent in certain cases of visual pathology. The use of video methods to measure eye movements requires that the eye can be visualized by video cameras well enough so that image processing methodology (e.g., tracking the iris-pupil limbus) can be effectively applied.

1.2 Purpose

The purpose of this Standard is to define a basic battery of tests to evaluate the vestibular function in humans. These tests define a minimum set of procedures for the purpose of establishing a common basis for comparison of data on subjects with both normal and dysfunctional vestibular function. These tests are not necessarily intended to replace more extensive tests used for detailed clinical or research purposes in specific applications but are intended to form a minimal base to be included in such studies.

1.3 Applications

Evaluation of vestibular function may be used in assessing disease, aging, and ability to function in certain occupations. Susceptibility to falling down, vertigo, motion sickness, and the ability to adapt to a weightless environment are related in greater or lesser degree to vestibular function.

This Standard does not include interpretation of the tests of vestibular function. For informational purposes, examples of test interpretations for subjects with normal and abnormal test results are provided in Annex A.

2 Normative references

The following referenced documents are indispensable for the application of this standard. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ANSI Z136.1-2007, *American National Standard for Safe Use of Lasers*

IEC 60825-1 *Safety of laser products - Part 1: Equipment classification and requirements*

UL 544, *Standard for Safety for Medical and Dental Equipment*

3 Terms and definitions

For the purposes of this Standard, the following definitions apply. More detailed discussion of the terminology employed is found in the references listed in the Bibliography.

3.1

electro-oculogram

record, usually made on a strip-chart recorder, of the differences in electrical potentials, at the skin at the sides of the eyes, that result from changes of the eyeball's gaze position

NOTE Abbreviation: EOG.

3.2

fast component

portion of the asymmetrical, sawtooth-shaped wave form of vestibular nystagmus having the greater slope magnitude