

STANDARDS AUSTRALIA

Amendment No. 1
to
AS 2252.2—2009
Controlled environments
Part 2: Biological safety cabinets Class II—Design

CORRECTION

The 2009 edition of AS 2252.2 is amended as follows; the amendment(s) should be inserted in the appropriate place(s).

SUMMARY: This Amendment applies to Clause 5.2.4.

Published on 21 June 2010.

Clause 5.2.4

Delete the text and *replace* with the following:

When determined in accordance with AS 1807.1, the average velocity of the laminar flow air shall be as per the manufacturer's specification. Where a specification is not available the average velocity shall be between 0.40 and 0.45 m/s. All readings shall fall within 20% of the average.

AMDT
No. 1
JUN
2010

This page has been left intentionally blank.