AS 2928—1987

Australian Standard®

OCCLUDED-EAR SIMULATOR FOR THE MEASUREMENT OF EARPHONES COUPLED TO THE EAR BY EAR INSERTS

Documents
Document Sharing Hub

Full document details: AS 2928-1987

This Australian standard was prepared by Committee AK/11, Audiology. It was approved on behalf of the Council of the Standards Association of Australia on 25 November 1986 and published on 2 February 1987.

The following interests are represented on Committee AK/11:

Audiological Society of Australia
Better Hearing Australia
Department of Defence
Hearing Aid Council of Australia
National Acoustical Laboratories
Otolaryngological Society of Australia
State Pollution Control Commission of New South Wales
University of Melbourne
University of Queensland

Review of Australian Standards. *To keep abreast of progress in industry, Australian Standards are subject to periodic review and are kept up to date by the issue of amendments or new editions as necessary. It is important therefore that Standards users ensure that they are in possession of the latest edition, and any amendments thereto.*

Full details of all Australian Standards and related publications will be found in the Standards Australia Catalogue of Publications; this information is supplemented each month by the magazine 'The Australian Standard', which subscribing members receive, and which gives details of new publications, new editions and amendments, and of withdrawn Standards.

Suggestions for improvements to Australian Standards, addressed to the head office of Standards Australia, are welcomed. Notification of any inaccuracy or ambiguity found in an Australian Standard should be made without delay in order that the matter may be investigated and appropriate action taken.

This standard was issued in draft form for comment as DR 85264.

Documents
Document Sharing Hub

Full document details: AS 2928-1987

AS 2928—1987

Australian Standard®

OCCLUDED-EAR SIMULATOR FOR THE MEASUREMENT OF EARPHONES COUPLED TO THE EAR BY EAR INSERTS

First published 1987

PUBLISHED BY STANDARDS AUSTRALIA
(STANDARDS ASSOCIATION OF AUSTRALIA)
1 THE CRESCENT, HOMEBUSH, NSW 2140

ISBN 0 7262 4508 9

Documents
Document Sharing Hub

Full document details: AS 2928-1987

PREFACE

This standard was prepared by the Association's Committee on Audiology. It is identical with and has been reproduced from IEC Publication 711, Occluded-ear Simulator for the Measurement of Earphones Coupled to the Ear by Ear Inserts.

The purpose of this standard is to specify an occluded-ear simulator intended for the calibration of insert earphones in the frequency range 100 Hz to 10 000 Hz in terms of the sound pressure at the eardrum.

It should be noted that this standard does not supersede AS 1089—1971, Reference Coupler for the Measurement of the Electro-acoustic Characteristics of Hearing Aid Earphones, but is an addition to it.

For the purpose of this Australian standard, the text of the IEC standard used herein should be modified as follows:

Cross-reference: The reference to the IEC standard should be replaced by a reference to the Australian standard as follows:

Reference to IEC standard	*Appropriate Australian Standard*
IEC 126, IEC Reference Coupler for the Measurement of Hearing Aids Using Earphones Coupled to the Ear by Means of Ear Inserts	AS 1089, Reference Coupler for the Measurement of the Electro-acoustic Characteristics of Hearing Aid Earphones

© Copyright — STANDARDS AUSTRALIA

Users of Standards are reminded that copyright subsists in all Standards Australia publications and software. Except where the Copyright Act allows and except where provided for below no publications or software produced by Standards Australia may be reproduced, stored in a retrieval system in any form or transmitted by any means without prior permission in writing from Standards Australia. Permission may be conditional on an appropriate royalty payment. Requests for permission and information on commercial software royalties should be directed to the head office of Standards Australia.

Standards Australia will permit up to 10 percent of the technical content pages of a Standard to be copied for use exclusively in-house by purchasers of the Standard without payment of a royalty or advice to Standards Australia.

Standards Australia will also permit the inclusion of its copyright material in computer software programs for no royalty payment provided such programs are used exclusively in-house by the creators of the programs.

Care should be taken to ensure that material used is from the current edition of the Standard and that it is updated whenever the Standard is amended or revised. The number and date of the Standard should therefore be clearly identified.

The use of material in print form or in computer software programs to be used commercially, with or without payment, or in commercial contracts is subject to the payment of a royalty. This policy may be varied by Standards Australia at any time.

AVA Documents
Document Sharing Hub

Full document details: AS 2928-1987

CONTENTS

Page

Clause

1. Scope 4
2. Object 4
3. Definitions 4
4. General requirements 5
5. Performance specifications 6
6. Performance tests 8
7. Coupling arrangements 10
8. Calibration 10
9. Atmospheric reference conditions 10

FIGURES 11

APPENDIX A — Figure A1. — Example of one specific design of occluded-ear simulator 15

Documents
Document Sharing Hub

Full document details: AS 2928-1987

STANDARDS ASSOCIATION OF AUSTRALIA

Australian Standard

for

OCCLUDED-EAR SIMULATOR FOR THE MEASUREMENT OF EARPHONES COUPLED TO THE EAR BY EAR INSERTS

1. Scope

This standard specifies an occluded-ear simulator intended for the calibration of insert earphones in the frequency range 100 Hz to 10 000 Hz in terms of the sound pressure at the eardrum.

The occluded-ear simulator is also designed as the basis for a later extension intended to simulate the complete ear canal for the calibration of earphones coupled to the ear by means of open mould fittings or similar devices. A document for such a device is under consideration.

2. Object

The occluded-ear simulator simulates the median values of relevant acoustical characteristics of adult normal human ears.

The occluded-ear simulator does not simulate the leakage between an earmould and a human ear canal; therefore, the results obtained with the simulator may deviate from the performance of an insert earphone on a real ear, especially at low frequencies.

Moreover, large performance variations among individual ears will occur and this should be borne in mind when employing results obtained with the simulator.

3. Definitions

For the purpose of this standard, the following definitions apply:

3.1 *Ear insert*

An ear insert is a device used to provide the acoustic coupling between an earphone and the ear canal (e.g. an earmould or a similar device with or without a connecting tube).

3.2 *Insert earphone*

An insert earphone is a small earphone coupled to the ear canal by means of an ear insert or attached to a connecting element which is inserted into the ear canal. This ear insert may be a part of the insert earphone.

3.3 *Ear insert simulator*

An ear insert simulator (e.g. earmould simulator) is an insert which terminates the entrance of the ear simulator and provides for passage of sound into the occluded-ear simulator through an opening on its axis.

COPYRIGHT

Documents
Document Sharing Hub

Full document details: AS 2928-1987