

Australian Standard™

Electroacoustics—Audiological equipment

Part 3: Auditory test signals of short duration for audiometric and neuro- otological purposes

[IEC title: Audiometers, Part 3: Auditory test signals of short
duration for audiometric and neuro-otological purposes]



Standards Australia



Documents

Document Sharing Hub

This Australian Standard was prepared by Committee AV-003, Acoustics Human Effects. It was approved on behalf of the Council of Standards Australia on 31 July 2002 and published on 16 September 2002.

The following are represented on Committee AV-003:

- Association of Australian Acoustical Consultants
- Association of Consulting Engineers Australia
- Australian Acoustical Society
- Australian Chamber of Commerce and Industry
- Australian Hearing
- Institution of Engineers Australia
- N.S.W. Rural Fire Service
- New South Wales Nurses Association
- Royal Institution of Naval Architects
- Safety Institute of Australia (Incorporated)
- Victorian WorkCover Authority
- WorkCover New South Wales

Keeping Standards up-to-date

Standards are living documents which reflect progress in science, technology and systems. To maintain their currency, all Standards are periodically reviewed, and new editions are published. Between editions, amendments may be issued. Standards may also be withdrawn. It is important that readers assure themselves they are using a current Standard, which should include any amendments which may have been published since the Standard was purchased.

Detailed information about Standards can be found by visiting the Standards Australia web site at www.standards.com.au and looking up the relevant Standard in the on-line catalogue.

Alternatively, the printed Catalogue provides information current at 1 January each year, and the monthly magazine, *The Australian Standard*, has a full listing of revisions and amendments published each month.

We also welcome suggestions for improvement in our Standards, and especially encourage readers to notify us immediately of any apparent inaccuracies or ambiguities. Contact us via email at mail@standards.com.au, or write to the Chief Executive, Standards Australia International Ltd, GPO Box 5420, Sydney, NSW 2001.

This Standard was issued in draft form for comment as DR 02237.

Australian Standard™

Electroacoustics—Audiological equipment

Part 3: Auditory test signals of short duration for audiometric and neuro- otological purposes

Originated as part of AS 2586—1985.
Revised and redesignated in part as AS IEC 60645.3—2002.

COPYRIGHT

© Standards Australia International

All rights are reserved. No part of this work may be reproduced or copied in any form or by any means, electronic or mechanical, including photocopying, without the written permission of the publisher.

Published by Standards Australia International Ltd
GPO Box 5420, Sydney, NSW 2001, Australia

ISBN 0 7337 4756 6

PREFACE

This Standard was prepared by the Australian members of the Joint Standards Australia/Standards New Zealand Committee AV-003, Acoustics Human Effects to be used as a means to demonstrate compliance with the relevant essential principles in new medical device legislation.

After consultation with stakeholders in both countries, Standards Australia and Standards New Zealand decided to develop this Standard as an Australian, rather than an Australian/New Zealand Standard.

This Standard is identical with and has been reproduced from IEC 60645-3:1994, *Audiometers, Part 3: Auditory test signals of short duration for audiometric and neuro-otological purposes*.

The objective of this Standard is to specify a means of describing the physical characteristics of audiometric test and reference signals of short duration and methods for their measurement. The Standard also specifies a psychoacoustic method for determining the level of the test signal in terms of hearing level.

This Standard provides for the use of the following Australian and Australian/New Zealand Standards as equivalents to particular International Standards referenced herein:

<i>Reference to International Standard</i>		<i>Australian/New Zealand Standard</i>	
IEC		AS/NZS	
60373	Mechanical coupler for measurements of bone vibrators	1591	Acoustics—Instrumentation for audiometry
		1591.4	Part 4: A mechanical coupler for calibration of bone vibrators
		AS	
60651	Sound level meters	1259	Acoustics—Sound level meters
		1259.1	Part 1: Non-integrating
60711	Occluded-ear simulator for the measurement of earphones coupled to the ear by ear inserts	2928	Occluded-ear simulator for the measurement of earphones coupled to the ear by ear inserts

As this Standard is reproduced from an International Standard, the following applies:

- (a) Its number does not appear on each page of text and its identity is shown only on the cover and title page.
- (b) In the source text ‘this part of IEC 645’ should read ‘this Australian Standard’.
- (c) A full point substitutes for a comma when referring to a decimal marker.

CONTENTS

	Page
INTRODUCTION	iv
Clause	
1 Scope	1
2 Normative references	1
3 Explanation of terms	2
4 Reference signals	3
4.1 Reference click	3
4.2 Reference brief tone	3
5 Calibration and measurement of short-duration signals	4
6 Determination of the signal level in terms of hearing level	4
7 Instruction manual	5
Figures	3
Annex A (Informative) Bibliography	6

INTRODUCTION

This International Standard provides a means of specifying acoustic and vibratory test signals of short duration used for audiometric and neuro-otological purposes. Examples of test methods, where such signals are commonly used, are the recording of brainstem evoked potentials and evoked oto-acoustic emissions. Reference broadband clicks and brief tones are described in order to provide a basis for calibration and as a recommendation for use when there is no specific reason to have an alternative signal. The method of measurement of acoustic and vibratory signals is described as well as guidance given regarding psychoacoustical calibration for air and bone conduction stimulation.