

AS/NZS 3200.2.1:1999
IEC 60601-2-1:1998

Australian/New Zealand Standard™

Medical electrical equipment

Part 2.1: Particular requirements for safety—Electron accelerators in the range 1 MeV to 50 MeV

[IEC title: Medical electrical equipment Part 2.1: Particular requirements for the safety of electron accelerators in the range 1 MeV to 50 MeV]

AS/NZS 3200.2.1:1999

This Joint Australian/New Zealand Standard was prepared by Joint Technical Committee HE/3, Medical Electrical Equipment. It was approved on behalf of the Council of Standards Australia on 12 March 1999 and on behalf of the Council of Standards New Zealand on 7 April 1999. It was published on 5 June 1999.

The following interests are represented on Committee HE/3:

Australasian College of Physical Scientists and Engineers in Medicine
 Australasian Society for Ultrasound in Medicine
 Australian & New Zealand College of Anaesthetists
 Australian Dental Association
 Australian Institute of Radiography
 Australian Radiation Laboratory
 Australian Society of Anaesthetists
 College of Biomedical Engineering Institution of Engineers Australia
 Commonwealth Department of Health and Family Services
 Department of Defence (Australia)
 Medical Industry Association of Australia
 Ministry of Commerce, New Zealand
 Royal Australasian College of Surgeons
 Royal Australasian College of Radiologists

Review of Standards. To keep abreast of progress in industry, Joint Australian/New Zealand Standards are subject to periodic review and are kept up to date by the issue of amendments or new editions as necessary. It is important therefore that Standards users ensure that they are in possession of the latest edition, and any amendments thereto.

Full details of all Joint Standards and related publications will be found in the Standards Australia and Standards New Zealand Catalogue of Publications; this information is supplemented each month by the magazines 'The Australian Standard' and 'Standards New Zealand', which subscribing members receive, and which give details of new publications, new editions and amendments, and of withdrawn Standards.

Suggestions for improvements to Joint Standards, addressed to the head office of either Standards Australia or Standards New Zealand, are welcomed. Notification of any inaccuracy or ambiguity found in a Joint Australian/New Zealand Standard should be made without delay in order that the matter may be investigated and appropriate action taken.

AS/NZS 3200.2.1:1999

Australian/New Zealand Standard™

Medical electrical equipment

Part 2.1: Particular requirements for safety—Electron accelerators in the range 1 MeV to 50 MeV

First published as AS/NZS 3200.2.1:1999.

Published jointly by:

Standards Australia
1 The Crescent,
Homebush NSW 2140 Australia

Standards New Zealand
Level 10, Radio New Zealand House,
155 The Terrace,
Wellington 6001 New Zealand

PREFACE

This Standard was prepared by the Joint Standards Australia/Standards New Zealand Committee HE/3, Medical Electrical Equipment as a Joint Standard.

This Particular Standard is identical with and has been reproduced from IEC 60601-2-1:1998 which modifies and supplements the corresponding Clauses of IEC 60601-1:1988, *Medical electrical equipment, Part 1: General requirements for safety* which has been adopted as AS/NZS 3200.1.0 hereinafter referred to as the General Standard.

The General Standard details electrical safety requirements for all types of medical electrical equipment. A Particular Standard details additional safety requirements for a related group of medical electrical devices. A Collateral Standard details additional safety requirements for a range of devices within the scope of the General Standard which may not be related but share common problems.

In the text of this Standard, the following print types are used:

- (a) Requirements, compliance with which can be tested and definitions
..... in large roman type
- (b) Explanations, advice, introductions, general statements, exceptions and references
..... in smaller roman type
- (c) Headings of sub-clauses and text specifications in italic type
- (d) Terms used throughout the Standard, which have been defined in Clause 2 and which are also in the index IN SMALL CAPITALS

Some pages of the original, which relate to IEC administrative matters, are also omitted from this edition.

As this publication has been reproduced from an international Standard, the following modifications apply:

- (i) its number does not appear on each page of text and its identity is shown on the cover and title page.
- (ii) the words this 'Australian/New Zealand Standard' should replace the words 'this International Standard' wherever they appear.
- (iii) substitute a full point for a comma when it appears as a decimal marker.

The references to international Standards should be replaced by references to the following Australian or Joint Australian/New Zealand Standards:

<i>Reference to International Standard or other publication</i>		<i>Australian/New Zealand Standard</i>	
IEC		AS/NZS	
60601-1	Medical electrical equipment— Part 1: General requirements for safety. Amendment 1, Amendment 2	3200 3200.1.0*	Medical electrical equipment— Part 1.0 General requirements for safety—Parent Standard
60601-1-4	Medical electrical equipment— Part 1: General requirements for safety, 4. Collateral Standard: Programmable electrical medical systems	3200.1.4	Part 1.4: General requirements for safety—Collateral Standard: Programmable electrical medical systems
60788	Medical radiology—Terminology	—	

* This version includes both IEC amendments.

IEC		AS/NZS	
61217	Radiotherapy equipment— Co-ordinates, movements and scales	4495	Radiotherapy equipment— Coordinates, movements and scales

The terms 'normative' and 'informative' have been used in this Standard to define the application of the annex to which they apply. A 'normative' annex is an integral part of a Standard, whereas an 'informative' annex is only for information and guidance.

© Copyright – STANDARDS AUSTRALIA/STANDARDS NEW ZEALAND

Users of Standards are reminded that copyright subsists in all Standards Australia and Standards New Zealand publications and software. Except where the Copyright Act allows and except where provided for below no publications or software produced by Standards Australia or Standards New Zealand may be reproduced, stored in a retrieval system in any form or transmitted by any means without prior permission in writing from Standards Australia or Standards New Zealand. Permission may be conditional on an appropriate royalty payment. Australian requests for permission and information on commercial software royalties should be directed to the head office of Standards Australia. New Zealand requests should be directed to Standards New Zealand.

Up to 10 percent of the technical content pages of a Standard may be copied for use exclusively in-house by purchasers of the Standard without payment of a royalty or advice to Standards Australia or Standards New Zealand.

Inclusion of copyright material in computer software programs is also permitted without royalty payment provided such programs are used exclusively in-house by the creators of the programs.

Care should be taken to ensure that material used is from the current edition of the Standard and that it is updated whenever the Standard is amended or revised. The number and date of the Standard should therefore be clearly identified.

The use of material in print form or in computer software programs to be used commercially, with or without payment, or in commercial contracts is subject to the payment of a royalty. This policy may be varied by Standards Australia or Standards New Zealand at any time.

CONTENTS

	Page
INTRODUCTION	viii
Clause	
SECTION ONE – GENERAL	
1 Scope and object	1
2 Terminology and definitions	3
4 General requirements for tests.....	4
5 Classification	4
6 Identification, marking and documents	8
SECTION TWO – ENVIRONMENTAL CONDITIONS	
10 Environmental conditions	11
SECTION THREE – PROTECTION AGAINST ELECTRIC SHOCK HAZARDS	
16 ENCLOSURES and PROTECTIVE COVERS.....	11
18 Protective earthing, functional earthing and potential equalization.....	13
19 Continuous LEAKAGE CURRENTS and PATIENT AUXILIARY CURRENTS.....	13
SECTION FOUR – PROTECTION AGAINST MECHANICAL HAZARDS	
22 Moving parts.....	14
27 Pneumatic and hydraulic power.....	17
28 Suspended masses	17
SECTION FIVE – PROTECTION AGAINST HAZARDS FROM UNWANTED OR EXCESSIVE RADIATION	
29 X-RADIATION	17
29 IONIZING RADIATION safety requirements	17
29.1 Protection against incorrect ABSORBED DOSE in the TREATMENT VOLUME.....	18
29.1.1 Monitoring and control of ABSORBED DOSE.....	18
29.1.1.1 DOSE MONITORING SYSTEMS	18
29.1.1.2 RADIATION DETECTORS	19
29.1.1.3 Selection and DISPLAY of DOSE MONITOR UNITS	20
29.1.1.4 TERMINATION OF IRRADIATION by the DOSE MONITORING SYSTEM.....	20
29.1.1.5 Monitoring of distribution of ABSORBED DOSE.....	21
29.1.2 CONTROLLING TIMER.....	22
29.1.3 ABSORBED DOSE RATE	23
29.1.4 Selection and DISPLAY of RADIATION TYPE	24
29.1.5 Selection and DISPLAY of ENERGY	25

Clause		Page
29.1.6	Selection and DISPLAY of STATIONARY RADIOTHERAPY and MOVING BEAM RADIOTHERAPY	25
29.1.7	RADIATION BEAM production and distribution systems	27
29.1.7.1	Selection and DISPLAY of TARGETS or other movable RADIATION BEAM production devices	27
29.1.7.2	Selection and DISPLAY of FIELD FLATTENING and BEAM SCATTERING FILTERS	27
29.1.7.3	RADIATION BEAM distribution systems not using FIELD FLATTENING or BEAM SCATTERING FILTERS	28
29.1.8	Selection and DISPLAY of WEDGE FILTERS	29
29.1.9	ELECTRON BEAM APPLICATORS and trays for RADIATION BEAM modifying devices	30
29.1.10	Control of EQUIPMENT use.....	31
29.1.11	Starting conditions	32
29.1.12	INTERRUPTION OF IRRADIATION	32
29.1.13	TERMINATION OF IRRADIATION.....	33
29.1.14	Abnormal TERMINATION OF IRRADIATION	33
29.1.15	PROGRAMMABLE ELECTRONIC SUBSYSTEMS	34
29.2	Protection against STRAY RADIATION in the RADIATION FIELD	35
29.2.1	STRAY X-RADIATION during ELECTRON IRRADIATION	35
29.2.2	RELATIVE SURFACE DOSE during X-IRRADIATION	35
29.2.3	STRAY NEUTRON RADIATION	36
29.3	Protection against RADIATION in the PATIENT plane outside the RADIATION FIELD .	36
29.3.1	LEAKAGE RADIATION through BEAM LIMITING DEVICES	36
29.3.1.1	X-RADIATION	37
29.3.1.2	ELECTRON RADIATION.....	38
29.3.2	LEAKAGE RADIATION (excluding NEUTRONS) outside the area <i>M</i>	39
29.3.3	LEAKAGE NEUTRON RADIATION outside the area <i>M</i>	40
29.3.4	LEAKAGE RADIATION under fault conditions.....	41
29.4	RADIATION safety for PATIENTS and others.....	41
29.4.1	LEAKAGE X-RADIATION outside the PATIENT plane	41
29.4.2	LEAKAGE NEUTRON RADIATION outside the PATIENT plane	42
29.4.3	Emission of IONIZING RADIATION after TERMINATION OF IRRADIATION due to INDUCED RADIOACTIVITY	42
29.4.4	Retractable RADIATION BEAM shield	43
29.4.5	Adventitious IONIZING RADIATION	43
36	ELECTROMAGNETIC COMPATIBILITY.....	44

SECTION SIX – PROTECTION AGAINST HAZARDS OF IGNITION OF FLAMMABLE ANAESTHETIC MIXTURES

SECTION NINE – ABNORMAL OPERATION AND FAULT CONDITIONS; ENVIRONMENTAL TESTS

Clause	<i>Page</i>
52 Abnormal operation and fault conditions	45

SECTION TEN – CONSTRUCTIONAL REQUIREMENTS

57 MAINS PARTS, components and layout	45
Appendix L (normative) References – Publications mentioned in this standard	55
Annex AA (informative) Terminology – Index of defined terms	56
Annex BB (informative) Bibliography	59
Figure 101 – Limits of STRAY X-RADIATION during ELECTRON IRRADIATION (29.2.1)	47
Figure 102 – Limits of RELATIVE SURFACE DOSE during X-IRRADIATION (29.2.2)	48
Figure 103 – Elevation view – Application of LEAKAGE RADIATION requirements (29.3 and 29.4)	49
Figure 104 – 24 measurement points for averaging LEAKAGE RADIATION during X-IRRADIATION (29.3.1.1)	50
Figure 105 – Limits of LEAKAGE RADIATION through the BEAM LIMITING DEVICES during ELECTRON IRRADIATION (29.3.1.2)	51
Figure 106 – Measurement points for averaging LEAKAGE RADIATION during ELECTRON IRRADIATION (29.3.1.2)	52
Figure 107 – 24 measurement points for averaging LEAKAGE RADIATION outside area <i>M</i> (29.3.2)	53
Figure 108 – EQUIPMENT movements and scales	54
Table 101 – Data required in the technical description to support clause 29 SITE TEST compliance	7
Table 102 – Clauses and subclauses in this Particular Standard that require the provision of information in the ACCOMPANYING DOCUMENTS, INSTRUCTIONS FOR USE and the technical description	12
Table 103 – Limits of STRAY X-RADIATION during ELECTRON IRRADIATION	35
Table 104 – Limits of RELATIVE SURFACE DOSE during X-IRRADIATION	36

INTRODUCTION

The use of ELECTRON ACCELERATORS for RADIOTHERAPY purposes may expose PATIENTS to danger if the EQUIPMENT fails to deliver the required dose to the PATIENT, or if the EQUIPMENT design does not satisfy standards of electrical and mechanical safety. The EQUIPMENT may also cause danger to persons in the vicinity if the EQUIPMENT itself fails to contain the RADIATION adequately and/or if there are inadequacies in the design of the TREATMENT ROOM.

This Particular Standard establishes requirements to be complied with by MANUFACTURERS in the design and construction of ELECTRON ACCELERATORS for use in RADIOTHERAPY; it does not attempt to define their optimum performance requirements. Its purpose is to identify those features of design that are regarded, at the present time, as essential for the safe operation of such EQUIPMENT; it places limits on the degradation of EQUIPMENT performance beyond which it can be presumed that a fault condition exists and where an INTERLOCK then operates to prevent continued operation of the EQUIPMENT.

Clause 29 contains limits beyond which INTERLOCKS prevent, INTERRUPT or TERMINATE IRRADIATION in order to avoid an unsafe condition. TYPE TESTS that are performed by the MANUFACTURER, and/or SITE TESTS, which are not necessarily performed by the MANUFACTURER, are SPECIFIED for each requirement.

It should be understood that, before installation, a MANUFACTURER can provide a compliance certificate relating only to TYPE TESTS. Data available from SITE TESTS should be incorporated in the ACCOMPANYING DOCUMENTS, in the form of a SITE TEST report, by those who test the EQUIPMENT at installation.

AUSTRALIAN/NEW ZEALAND STANDARD

Medical electrical equipment**Part 2.1:****Particular requirements for safety—Electron accelerators in the range 1 MeV to 50 MeV**

SECTION ONE – GENERAL

The clauses and subclauses of this section of the General Standard apply, except as follows:

1 Scope and object**1.1 Scope***Addition:*

This Particular Standard, with the inclusion of TYPE TESTS and SITE TESTS, applies respectively to the manufacture and some installation¹⁾ aspects of ELECTRON ACCELERATORS²⁾

- intended for RADIOTHERAPY in human medical practice, including those in which the selection and DISPLAY of operating parameters can be controlled automatically by PROGRAMMABLE ELECTRONIC SUBSYSTEMS (PESS),
- that, under NORMAL CONDITIONS (NC) and in NORMAL USE, deliver a RADIATION BEAM of X-RADIATION and/or ELECTRON RADIATION having
 - NOMINAL ENERGY in the range 1 MeV to 50 MeV,
 - maximum ABSORBED DOSE³⁾ RATES between $0,001 \text{ Gy} \times \text{s}^{-1}$ and $1 \text{ Gy} \times \text{s}^{-1}$ at 1 m from the RADIATION SOURCE,
 - NORMAL TREATMENT DISTANCES (NTDs) between 0,5 m and 2 m from the RADIATION SOURCE,

and

- intended to be
 - for NORMAL USE, operated under the authority of appropriately licensed or QUALIFIED PERSONS by OPERATORS having the required skills for a particular medical application, for particular specified clinical purposes, e.g. STATIONARY RADIOTHERAPY or MOVING BEAM RADIOTHERAPY,
 - maintained in accordance with the recommendations given in the INSTRUCTIONS FOR USE,
 - subject to regular quality assurance performance and calibration checks by a QUALIFIED PERSON,
 - used within the environmental and electrical supply conditions SPECIFIED in the technical description.

¹⁾ In this Particular Standard, all references to installation refer to installation in the USER's premises.

²⁾ See ICRP 33 (128) – (134) and (144) – (156).

³⁾ In this Particular Standard, all references to ABSORBED DOSE refer to ABSORBED DOSE in water.

1.2 Object

Addition:

This Particular Standard establishes requirements to ensure the IONIZING RADIATION safety and enhanced electrical and mechanical safety of ELECTRON ACCELERATORS, and specifies tests to check compliance with those requirements.

NOTE – The adoption of this standard helps to ensure that the EQUIPMENT

- maintains PATIENT safety during EQUIPMENT movements and failure of the SUPPLY MAINS,
- delivers the pre-selected RADIATION TYPE, NOMINAL ENERGY, and ABSORBED DOSE,
- delivers the RADIATION in accordance with the pre-selected relationship of the RADIATION BEAM to the PATIENT, by utilizing STATIONARY RADIOTHERAPY, MOVING BEAM RADIOTHERAPY, RADIATION BEAM modifying devices, etc., without causing unnecessary risk to the PATIENT, the OPERATOR, other persons or the environment.

1.3 Particular Standards

Additional subclauses:

1.3.101 Relationship to other standards and documents

NOTE – See Appendix L for normative references.

1.3.102 IEC 60601-1

The requirements of this Particular Standard take priority over those of all other standards; it is to be read in conjunction with IEC 60601-1:1988, *Medical electrical equipment – Part 1: General requirements for safety* with its amendments 1 (1991) and 2 (1995) – hereinafter referred to as the General Standard – which it amends and supplements. As in the General Standard, the requirements are followed by compliance tests.

Where there is no corresponding section, clause or subclause in this Particular Standard, the section, clause or subclause of the General Standard, although possibly not relevant, applies without modification. Where it is intended that any part of the General Standard, although possibly relevant, is not to be applied, a statement to that effect is given in this Particular Standard. Unless otherwise stated, all clauses of the General Standard shall apply. The term "this Standard" is used throughout to refer to the General Standard and this Particular Standard taken together.

The numbering of sections, clauses and subclauses of this Particular Standard corresponds to that of the General Standard. The changes to the text of the General and Collateral Standards are SPECIFIED by the use of the following words:

- "Replacement" means that the clause or subclause of the General Standard is replaced completely by the text of this Particular Standard;
- "Addition" means that the text of this Particular Standard is additional to the requirements of the General Standard;
- "Amendment" means that the clause or subclause of the General Standard is amended as indicated by the text of this Particular Standard.

Subclauses, figures or tables that are additional to those of the General Standard are numbered starting from 101: additional annexes are lettered AA, BB, etc., and additional items *aa*), *bb*), etc.

1.3.103 IEC 61217

This standard gives guidance on the designation of EQUIPMENT movements, the marking of scales, their zero positions and the direction of movement with increasing value (see 6.3.101).

1.5 Collateral Standards

Additional subclauses:

1.5.101 IEC 60601-1-1

This Collateral Standard does not apply.

1.5.102 IEC 60601-1-2

ELECTROMAGNETIC COMPATIBILITY – Requirements and tests; see clause 36.

NOTE – IEC 60601-1-2 applies to MEDICAL ELECTRICAL EQUIPMENT and ITE (INFORMATION TECHNOLOGY EQUIPMENT) used in medical applications. ELECTRON ACCELERATORS and their integral ITE are not excluded from compliance with IEC 60601-1-2; it has not been possible at this date to determine fully whether further amendments to its requirements or tests will need to be made in addition to those given in clause 36 of this Particular Standard.

1.5.103 IEC 60601-1-3

This Collateral Standard does not apply.

1.5.104 IEC 60601-1-4

PROGRAMMABLE ELECTRONIC SUBSYSTEMS (PESS) – Requirements and tests: see 29.1.15 and 52.1b).

2 Terminology and definitions**2.1 EQUIPMENT parts, auxiliaries and ACCESSORIES**

Addition:

NOTE – Annex AA lists defined terms alphabetically with their source reference.

Additional definitions:

2.1.101**CONTROLLING TIMER**

device to measure the time during which IRRADIATION occurs and, if a predetermined time is reached, TO TERMINATE IRRADIATION

2.1.102**ELECTRON BEAM APPLICATOR**

BEAM LIMITING DEVICE for ELECTRON RADIATION BEAMS

2.1.103**GANTRY**

that part of the EQUIPMENT supporting the RADIATION HEAD

2.1.104**GEOMETRICAL RADIATION FIELD**

geometrical projection of the distal end of the BEAM LIMITING DEVICE on a plane orthogonal to the REFERENCE AXIS, as seen from the centre of the front surface of the TARGET/ELECTRON RADIATION window. The GEOMETRICAL RADIATION FIELD may be defined at any distance from the front surface of the TARGET for X-RADIATION, or from the ELECTRON RADIATION window for ELECTRON RADIATION

2.1.105**HARD-WIRED**

term used where the features of a system can be modified only by physically removing and re-routing wires

2.1.106**INTERRUPTION OF IRRADIATION / TO INTERRUPT IRRADIATION**

stopping of/to stop IRRADIATION and movements with the possibility of continuing without reselecting operating conditions

2.1.107**MOVING BEAM RADIOTHERAPY**

RADIOTHERAPY with any planned displacement of the RADIATION FIELD or PATIENT relative to each other or with any planned change of ABSORBED DOSE distribution

2.1.108**NOMINAL ENERGY (hereinafter referred to as ENERGY)**

- a) for ELECTRON RADIATION, ENERGY stated by the MANUFACTURER to characterize the RADIATION BEAM. This ENERGY is approximately equal to the most probable energy at the surface of the measuring PHANTOM, $E_{p,0}$ (See ICRU Report 35: Section 3.3; Energy, $E_{p,0}$)
- b) for X-RADIATION, the ENERGY stated by the MANUFACTURER to characterize the RADIATION BEAM

2.1.109**NORMAL TREATMENT DISTANCE (abbreviation: NTD)**

- a) for ELECTRON IRRADIATION, SPECIFIED distance measured along the REFERENCE AXIS from the ELECTRON RADIATION window to the distal end of the ELECTRON BEAM APPLICATOR or to a SPECIFIED plane
- b) for X-IRRADIATION, SPECIFIED distance measured along the REFERENCE AXIS from the front surface of the TARGET to the ISOCENTRE or, for EQUIPMENT without an ISOCENTRE, to a SPECIFIED plane

2.1.110**PASSWORD**

for EQUIPMENT under the control of a PROGRAMMABLE ELECTRONIC SUBSYSTEM (PESS), sequence of keystrokes that permits OPERATOR access for NORMAL USE or to reset INTERLOCKS and, with a different sequence of keystrokes, permits access for adjustment and maintenance

2.1.111**PATIENT SUPPORT**

the assembly of EQUIPMENT that supports the PATIENT

2.1.112**PRIMARY/SECONDARY DOSE MONITORING COMBINATION**

utilization of two DOSE MONITORING SYSTEMS where one is arranged to be the PRIMARY and the other the SECONDARY DOSE MONITORING SYSTEM

2.1.113**QUALIFIED PERSON**

person recognised by a competent authority as having the requisite knowledge and training to perform particular duties

2.1.114**RADIATION TYPE**

nature of the waves or corpuscles comprising the RADIATION, e.g. whether the RADIATION is X-RADIATION or ELECTRON RADIATION

2.1.115**REDUNDANT DOSE MONITORING COMBINATION**

utilization of two DOSE MONITORING SYSTEMS where both systems are arranged to TERMINATE IRRADIATION according to the pre-selected number of DOSE MONITOR UNITS

2.1.116**RELATIVE SURFACE DOSE**

ratio of the ABSORBED DOSE on the REFERENCE AXIS, at the depth of 0,5 mm, to the maximum ABSORBED DOSE on the REFERENCE AXIS, both measured in a PHANTOM with its surface at a SPECIFIED distance

2.1.117**SITE TEST**

after installation, test of an individual device or EQUIPMENT to establish compliance with SPECIFIED criteria

2.1.118**TERMINATION OF IRRADIATION / TO TERMINATE IRRADIATION**

stopping off/to stop IRRADIATION and movements, with no possibility of restarting without the reselection of all operating conditions (i.e. returning / to return to the PREPARATORY STATE)

NOTE – Events that TERMINATE IRRADIATION and stop movements include the following events

- when the pre-selected value of DOSE MONITOR UNITS is reached,
- when the pre-selected value of elapsed time is reached,
- a deliberate manual act,
- the operation of an INTERLOCK,
- in MOVING BEAM RADIOTHERAPY, when the pre-selected value of an angular or linear dimension is exceeded.

2.1.119**TRANSMISSION DETECTOR**

RADIATION DETECTOR through which the RADIATION BEAM passes

2.1.120**TYPE TEST**

for a particular design of device or EQUIPMENT, test by the MANUFACTURER to establish compliance with SPECIFIED criteria

4 General requirements for tests

4.1 Tests

Additional subclause:

4.1.101 Test grades

Three grades of TYPE TEST and two of SITE TEST procedures are SPECIFIED in clause 29 of this Particular Standard. Their requirements are as follows:

- *TYPE TEST grade A: An analysis of EQUIPMENT design, as related to the SPECIFIED RADIATION safety provisions, which shall result in a statement included in the technical description, regarding the working principles or constructional means by which the requirement is fulfilled.*
- *TYPE TEST / SITE TEST grade B: Visual inspection or functional test or measurement of the EQUIPMENT. The test shall be in accordance with the procedure SPECIFIED in this Particular Standard and shall be based on operating states, including fault condition states, which are achievable only without interference with the circuitry or construction of the EQUIPMENT.*
- *TYPE TEST / SITE TEST grade C: Functional test or measurement of the EQUIPMENT. The test shall be in accordance with the principle SPECIFIED in this Particular Standard. The SITE TEST procedure shall be included in the technical description. When the procedure involves operating states that require interference with circuitry or the construction of the EQUIPMENT, the test should be performed by, or under the direct supervision of, the MANUFACTURER or his agent.*

5 Classification

Replacement:

EQUIPMENT and its APPLIED PARTS shall be classified by marking and/or identification as described in clause 6. This includes:

5.1 According to the type of protection against electric shock:

- CLASS I EQUIPMENT.

5.2 According to the degree of protection against electric shock:

- TYPE B APPLIED PARTS.

5.3 According to the degree of protection against ingress of water detailed in IEC 60529 (see 6.1 I))

- IPX0, unless otherwise SPECIFIED.

5.4 According to the method(s) of sterilization or disinfection recommended in the INSTRUCTIONS FOR USE.

5.5 According to the degree of safety of application in the presence of a FLAMMABLE ANAESTHETIC MIXTURE WITH AIR, or WITH OXYGEN OR NITROUS OXIDE:

- EQUIPMENT not suitable for use in the presence of a FLAMMABLE ANAESTHETIC MIXTURE WITH AIR or WITH OXYGEN OR NITROUS OXIDE.

5.6 According to the mode of operation:

- CONTINUOUS OPERATION WITH INTERMITTENT LOADING, unless otherwise SPECIFIED.

Addition:

**Table 101 – Data required in the technical description to support
clause 29 SITE TEST compliance**

Compliance subclause	Statement regarding data from TYPE TESTS grade A	Details of, and results from, TYPE TESTS grade B	Details of, and results from, TYPE TESTS grade C	SPECIFIC procedures and test conditions for SITE TESTS grade C
29.1.1.1	b) d) e)		d)	a) b) c)
29.1.1.2	a) b)		b) c)	b)
29.1.1.3	d) e) f)			
29.1.1.4	a) b) c)			b) c)
29.1.1.5	a)			b)
29.1.2	a) b) c)			b) c)
29.1.3	b) c) d) e)			b) c) d) e)
29.1.4	a) b) c) d) e) f) g)			
29.1.5	e)		e)	
29.1.6	e) h)		f) g)	e) f) g) h)
29.1.7.1				d)
29.1.7.2				d)
29.1.7.3.1	†			†
29.1.7.3.2				d)
29.1.9	b)			
29.1.10	e)			
29.1.11	†			
29.1.12	a) c)			
29.1.13	a)			
29.1.14	a) b)			a) b)
29.1.15	†			†
29.2.1	†			
29.2.2		†		
29.2.3			†	
29.3.1.1		a) b) c) d)		
29.3.1.2	c)	a) 1), a) 2), b)		c)
29.3.2		†		†
29.3.3		†		
29.3.4	†		† ← or →	†
29.4.1		a) b)		
29.4.2			†	
29.4.3	b)	a)		
29.4.4	†			
29.4.5		†		
NOTE – † denotes requirement of subclause having no other specific identification.				

6 Identification, marking and documents

6.1 Marking on the outside of EQUIPMENT or EQUIPMENT parts

d) *Minimum requirements for marking on EQUIPMENT and on interchangeable parts*

Addition:

The dimensions of the GEOMETRICAL RADIATION FIELD at NTD and the distance from the distal end to NTD shall be clearly legible on the outside of all interchangeable and non-adjustable BLDs and ELECTRON BEAM APPLICATORS.

Each manually interchangeable WEDGE FILTER shall be clearly marked to establish its identity.

z) *Removable protective means*

Addition:

Where the requirements of this item are wholly or partly met by the nature of the installation, compliance at installation should be checked by inspection; the results should be included in the SITE TEST report.

6.2 Marking on the inside of EQUIPMENT or EQUIPMENT parts

Addition:

aa) Removal of the covers of the RADIATION HEAD shall expose symbol 14 of table D.1 of the General Standard, indicating "Attention, consult ACCOMPANYING DOCUMENTS".

6.3 Marking of controls and instruments

Additional subclause:

6.3.101 Provision of scales and indications for moving parts

The following shall be provided:

- a) a mechanical scale or a numerical indication for each available movement;
- b) a LIGHT FIELD, with an indication of the position of the REFERENCE AXIS;
- c) a scale or numerical indication of the distance along the REFERENCE AXIS from the front surface of the RADIATION SOURCE to the surface of the PATIENT (RADIATION SOURCE TO SKIN DISTANCE).

The designation, direction of increasing value and zero position of all movements shall comply with IEC 61217 (see figure 108).

Compliance is checked by inspection.

6.7 Indicator lights and push-buttons

a) Colours of indicator lights

Addition:

Where indicator lights are used on the TREATMENT CONTROL PANEL (TCP) or other control panels, the colours of the lights shall accord with the following:

- RADIATION BEAM "on" yellow⁴⁾
- READY STATE green⁴⁾
- urgent action required in response to an unintended state of operation.. red
- PREPARATORY STATE other colour

Light emitting diodes (LEDs) are not considered to be indicator lights when

- on any one TCP, all indications for which no particular colour is required are given by LEDs of the same colour, and
- the indications for which particular colours are required are clearly distinguishable.

6.8 ACCOMPANYING DOCUMENTS

6.8.1 General

Addition:

NOTE – The INSTRUCTIONS FOR USE and technical description form part of the ACCOMPANYING DOCUMENTS. With the exception of data required in the technical description to support clause 29 SITE TEST compliance, which are set out in table 101, clauses and subclauses in this Particular Standard that require the provision of information in the ACCOMPANYING DOCUMENTS, INSTRUCTIONS FOR USE and technical description are given in table 102.

6.8.2 INSTRUCTIONS FOR USE

a) General information

Addition:

- INSTRUCTIONS FOR USE shall contain
 - an explanation of the function of all INTERLOCKS and other RADIATION safety devices,
 - instructions for checking their correct operation,
 - a recommendation of the frequency with which such checks should be made,

⁴⁾ In the TREATMENT ROOM or at other locations, these states may require urgent action or caution; different colours, in accordance with table 3 of the General Standard, may therefore be used in such locations.

- the recommended inspection or replacement intervals for any parts having a safety function that are subject to impairment caused, during the NORMAL USE of the EQUIPMENT, by the effects of IONIZING RADIATION on the dielectric and/or mechanical strength of those parts.

j) Environmental protection

Addition:

NOTE – The USER's RADIOLOGICAL PROTECTION adviser is, generally, the person responsible for the identification and disposal of material that may exhibit RADIOACTIVITY.

To assist the USER's RADIOLOGICAL PROTECTION adviser, the following data shall be provided:

- ENERGIES and corresponding maximum ABSORBED DOSE RATES at NTD under conditions of NORMAL USE for
 - X-RADIATION (with and without any ADDED FILTER if NORMAL USE is possible in both these states), and
 - ELECTRON RADIATION;
- dimensioned shape of the maximum GEOMETRICAL RADIATION FIELDS at NTD for X-RADIATION and ELECTRON RADIATION;
- location(s), referenced to accessible points on the RADIATION HEAD, of:
 - the front surface of the TARGET, and
 - the ELECTRON BEAM RADIATION window;
- available directions of the RADIATION BEAM;
- if a RADIATION BEAM shield is incorporated, its transmission factor at each X-RADIATION ENERGY;
- guidance and precautions regarding the identification, handling and disposal of EQUIPMENT or EQUIPMENT parts that may exhibit RADIOACTIVITY.

6.8.3 Technical description

a) General

Addition:

- full details of the environmental conditions and power supply required for NORMAL USE.

Additional items:

aa) SITE TEST information

The technical description shall contain the following:

- statements resulting from TYPE TESTS grade A;
- details of, and results from, TYPE TESTS grades B and C;
- SPECIFIC procedures and test conditions for SITE TESTS grade C;

NOTE 1 – See table 101 for clause 29 SITE TEST compliance data requirements.

- instructions on how to generate a described fault condition or, if not practicable, how to generate a test signal as close as practicable to the source of the signal that would have generated it, with a statement confirming that the test signal simulates the one that would be produced in a particular fault condition;

NOTE – In some cases, one test signal may simulate more than one fault condition.

- instructions on how to reset the EQUIPMENT to the NORMAL USE condition after the completion of the SITE TESTS and how to verify this condition.

NOTE 2 – The person responsible for the SITE TESTS should record the results in a report which will then form part of the ACCOMPANYING DOCUMENTS; in addition, the SITE TEST report should contain at least the following:

- name and address of USER site;
- MODEL OR TYPE REFERENCE (type number) and SERIAL NUMBER of the EQUIPMENT;
- name, status and employment address of all personnel taking part in the tests, and dates of their participation;
- environmental and power supply conditions;
- the actual conditions, when test conditions, procedures or devices differ from those given by the MANUFACTURER, or where the information cannot be derived from this Particular Standard.

Compliance is checked by inspection of the technical description.

SECTION TWO – ENVIRONMENTAL CONDITIONS

The clauses and subclauses of this section of the General Standard apply, except as follows:

10 Environmental conditions

Addition:

NOTE – See 1.1, third dash, fourth bullet, and 6.8.3 a): *Addition*.

10.2.2 Power supply

- a) EQUIPMENT shall be suitable for a power supply having:

Amendment:

Amend second dash to read as follows:

- a sufficiently low internal impedance to prevent voltage fluctuations between the on-load and off-load steady states exceeding $\pm 5\%$.

SECTION THREE – PROTECTION AGAINST ELECTRIC SHOCK HAZARDS

The clauses and subclauses of this section of the General Standard apply, except as follows:

16 ENCLOSURES and PROTECTIVE COVERS

Additional item:

- aa) Where the requirements of this clause are wholly or partly met by the nature of the installation, compliance test methods shall be SPECIFIED in the ACCOMPANYING DOCUMENTS.

Compliance at installation should be checked by inspection and tests; the results should be included in the SITE TEST report (see also 57.1 a)).

Addition:

Table 102 – Clauses and subclauses in this Particular Standard that require the provision of information in the ACCOMPANYING DOCUMENTS, INSTRUCTIONS FOR USE and the technical description

NOTE – The check reference is given as an aid for checking the availability of compliance documentation.

Check reference	ACCOMPANYING DOCUMENTS	INSTRUCTIONS FOR USE	Technical description
1		1.1	
2			1.1
3			4.1.101
4		5.4	
5	6.1 d) via IEC 60601-1: 6.8.1/ 6.1 d)		
6	6.2 aa)		
7	6.8 / 6.8.1 (Note)		
8		6.8.2 a), j)	
9			6.8.3 a), aa)
10	16 aa)		
11 : 1, 2, 4		22.4.1 a) 1), a) 2)	22.4.1 a) 4)
12		22.4.2 d)	
13		22.4.3 e)	
14	22.7.101		
15	28.101 b)		
16			29 (see table 101)
17	29.1.1.1 c) (Note)		
18			29.1.1.1 d)
19		29.1.2 b)	
20: b, c, e			29.1.3 b), c), e).
21			29.1.4 f) (Note)
22		29.1.5 d)	
23			29.1.7.3.1
24		29.1.10 e)	
25: a, c		29.1.12 a), c)	
26		29.1.13 a)	
27		29.1.14 a)	
28		29.1.15 f)	
29			29.4.1 b)
30			29.4.3 b)
31	36		
32		52.1 b)	
33			57.1 a)