

STANDARDS AUSTRALIA/STANDARDS NEW ZEALAND

Pr BB

Amendment No. 1

to

AS/NZS 3200.2.31:1996

Approval and test specification—Medical electrical equipment

Part 2.31: Particular requirements for safety—External cardiac pacemakers
with internal power source

REVISED TEXT

The 1996 edition of AS/NZS 3200.2.31 is amended as follows; the amendment(s) should be inserted in the appropriate place.

SUMMARY: This Amendment is in two parts. Part A applies to the title. Part B is identical with and reproduced from IEC 60601-2-31 Amendment 1:1998.

Published on 5 March 1999.

Approved for publication in New Zealand on behalf of the Standards Council of New Zealand on 26 November 1998.

AMDT
No. 1
MAR.
1999

PART A**Title**

Delete the existing title of the Standard wherever it occurs and *replace* with the following:

Medical electrical equipment, Part 2.31: Particular requirements for safety—External cardiac pacemakers with internal power source.

AMDT
No. 1
MAR.
1999

PART B

Replace existing text as specified below.

Page 9

INTRODUCTION

Replace the text of the first paragraph by the following:

This Particular Standard concerns the safety of PACEMAKERS. The relationship of this Particular Standard with IEC 60601-1 (including its amendments) and the Collateral Standards is explained in 1.3.

Replace the text of the fourth paragraph by the following:

PACEMAKERS differ in the various ways in which they maintain and monitor cardiac activity in different circumstances. The simplest model stimulates the atrium or ventricle independently of the cardiac activity; others detect atrial or ventricular activity and stimulate the atrium or ventricle as and when this is necessary; others, more complex, detect the spontaneous heart activity and stimulate appropriately the atrium and/or the ventricle. Certain PACEMAKERS work on preset frequency values, amplitudes and impulse duration. Others can have several values for parameters.

Delete the sixth paragraph.

ISBN 0 7337 2494 9

Addition:

This Particular Standard amends and supplements a set of IEC publications, hereinafter referred to as "General Standard", consisting of IEC 60601-1:1988, *Medical electrical equipment – Part 1: General requirements for safety*, amendment 1, amendment 2; IEC 60601-1-1:1992, *Medical electrical equipment – Part 1: General requirements for safety*, 1. Collateral Standard: *Safety requirements for medical electrical systems*, amendment 1; IEC 60601-1-2:1993, *Medical electrical equipment – Part 1: General requirements for safety*, 2. Collateral Standard: *Electromagnetic compatibility – Requirements and tests*, and IEC 60601-1-4:1996, *Medical electrical equipment – Part 1: General requirements for safety*, 4. Collateral Standard: *Programmable electronic medical systems*.

For brevity, IEC 60601-1 is referred to in this Particular Standards either as the "General Standard" or as the "General Requirement(s)", and IEC 60601-1-1, IEC 60601-1-2 and IEC 60601-1-4 as the "Collateral Standards".

The term "this Standard" covers this Particular Standard, used together with the General Standard and Collateral Standards.

2 Terminology and definitions

Replace the text of 2.1.102 by the following:

2.1.102**MAXIMUM TRACKING RATE**

maximum ventricular pacing rate in response to sensed atrial activity

Replace the text of 2.1.105 by the following:

2.1.105**POST-VENTRICULAR ATRIAL REFRACTOR PERIOD (PVARP)**

period after a ventricular event (whether sensed or paced), during which synchronous ventricular pacing is disabled, regardless of any atrial event

6 Identification, marking and documents**6.8 Accompanying documents****6.8.2 Instructions for use**

Add, on page 17, the following:

a)* Replacement:

Replace the text of the third dash by the following:

- Instructions for use shall include warnings regarding potential changes in the behaviour of the PULSE GENERATOR caused by electromagnetic or other interference sources (e.g. communication transmitters in hospitals, emergency transport vehicles, cellular telephones, etc.) and the effects of therapeutic and diagnostic energy sources (e.g. external cardioversion, diathermy, TENS devices, high-frequency surgical equipment, magnetic resonance imaging or similar sources) on the PULSE GENERATOR. This shall include advice on recognizing when the behaviour of the PULSE GENERATOR is being influenced by external interference sources and steps to be taken to avoid such interference.

AMDT
No. 1
MAR.
1999

Page 17

aa) Supplementary instructions for use

3)* *Replace the text of the fifth indent by the following:*

- sensing amplifier blanking period(s) (if a sensing function is provided);

6)* *Replace the existing text by the following:*

6) Not used.

Page 21

12)* *Replace the text of the fourth indent by the following:*

- inspection of the NON-IMPLANTABLE PULSE GENERATOR and PATIENT CABLE for signs of physical damage or contamination, in particular damage or contamination that may have a detrimental effect on the electrical isolation properties of the EQUIPMENT;

Add new items 13) and 14) as follows:

13)* A warning that, before handling the EXTERNAL PULSE GENERATOR, the PATIENT CABLE or indwelling LEADS, steps should be taken to equalize the electrostatic potential between the USER and the PATIENT, for example by touching the PATIENT at a site remote from the pacing LEAD.

14)* A caution that, when clinically indicated, supplemental monitoring of the PATIENT should be considered.

Page 25

36* Electromagnetic compatibility

Replace the text of the first two paragraphs by the following:

This clause of the General Standard applies, except as follows:

36.202.1* ELECTROSTATIC DISCHARGE

Replacement:

Construction of the EQUIPMENT shall ensure a sufficient degree of protection against SAFETY HAZARDS caused by repeated exposure to ELECTROSTATIC DISCHARGE.

Replace the last sentence in the third paragraph of the compliance test by the following:

No inappropriate delivery of energy to the APPLIED PART shall occur at any severity level specified in table 102.

AMDT
No. 1
MAR.
1999

Page 33

51 Protection against hazardous output

Add the following:

51.1* Intentional exceeding of safety limits

Replacement:

If the EQUIPMENT incorporates features which require PULSE RATES above the rate limit (see 51.104), the runaway rate protection may be disarmed when the feature is in use. The means for disarming the runaway rate protection shall require the USER to engage continuously the activating mechanism.

Compliance is checked by inspection and by a functional test.

Page 35

51.104 Rate limit (runaway protection)

Add an asterisk to the subclause number; delete "a)" and delete paragraph b).

51.106* MAXIMUM TRACKING RATE

Replace the text of the first paragraph by the following:

In DUAL CHAMBER modes incorporating atrial-synchronous ventricular pacing, a means shall be provided to set a limit at which the ventricle is paced in response to sensed atrial activity. The EQUIPMENT shall respond to sensed atrial activity above the MAXIMUM TRACKING RATE in a manner stated by the manufacturer.

Page 37

56 Components and general assembly

56.3 Connections – General

Delete "Addition" and the two paragraphs that follow.

Annex AA
(informative)**General guidance and rationale**

Page 43

Hazard analysis*Replace the text of the first paragraph by the following:*

EXTERNAL PACEMAKERS are used to treat PATIENTS who have symptomatic or acute bradycardia as well as for temporary pacing related to other medical procedures. PATIENT safety is affected by the medical procedure involved, by the understanding of EQUIPMENT function by the clinician and by EQUIPMENT function. The requirements as specified in this Standard are considered to provide for an acceptable level of safety.

Page 45

Table AA.1, change the fourth entry under "Unwanted stimulation" to read:

Unwanted stimulation	Noise	Noise reversion Warnings	6.8.2 aa) 3) and 51.105 6.8.2 a) Third dash
----------------------	-------	-----------------------------	--

Table AA.1, change the second entry under "Micro/macro shock" to read:

Micro/macro shock	Injection current	Warnings	6.8.2 a) Third dash, 6.8.2 aa) 5), 6.8.2 aa) 7) and 6.8.2 aa) 8)
-------------------	-------------------	----------	--

Page 49

Add a new subclause as follows:

- 6.8.2 a) Sources of electrical interference may affect the operation of the EQUIPMENT. In the presence of excessive levels of interference, the EQUIPMENT may:
- fail to pace,
 - revert to asynchronous pacing, or
 - inappropriately track the interference as cardiac activity.

Page 51

Delete 6.8.2 aa) 6).

Add new items 6.8.2 aa) 13) and 6.8.2 aa) 14) as follows:

- 6.8.2 aa) 13) Although believed to be, at best, a rare complication of pacing, there is a theoretical possibility that a static discharge to the EXTERNAL PULSE GENERATOR or a PATIENT CABLE connected to it could transfer minimally sufficient energy to the PATIENT to produce cardiac depolarization. If this were to occur in an electrically unstable PATIENT during the vulnerable portion of the cardiac cycle, a potentially lethal arrhythmia might be induced. No documented cases or anecdotal reports of such an event are known. It should be noted that there are ways that one or more asynchronous pulses may be delivered to the PATIENT (e.g. noise reversion, loss of sensing) all of which are much more likely and which typically are cautioned about in the labelling. While only rarely have these common occurrences precipitated an arrhythmia, medical literature leaves no doubt as to the potential for serious consequences. Therefore, a warning that care should be taken to discharge any static electricity that has accumulated on the attending health care professional or the PATIENT before touching the EQUIPMENT is appropriate.
- 6.8.2 aa) 14) The pulse energy delivered to the PATIENT is a consequence of the setting of the EXTERNAL PULSE GENERATOR and interaction of that output with a dynamic PATIENT/LEAD environment. The acute load presented by the temporary PATIENT/LEAD system can vary over a range of several hundred ohms. While much of this variation may be clinically inconsequential, "significant" departures from the pre-set level of energy output can occur. Since what constituted a "significant" departure from the pre-set level of energy output will vary widely from PATIENT to PATIENT depending on many factors, including the pre-set margin of safety for capture, selecting a limit that could be monitored by the EQUIPMENT and that would apply to all PATIENTS would necessarily leave other PATIENTS largely unprotected. The output circuitry cannot readily determine if the output resulted in capture of the heart.

Subclause 19.4 j) 4)

Replace the text of the second paragraph by the following:

NON-IMPLANTABLE PULSE GENERATORS may provide a "recharge" pulse whose area (integral of amplitude over time) is equal to that of the pacing pulse and of opposite polarity. The purpose of a recharge pulse is to make the net current through the tissue and the LEAD zero. Since the "recharge" pulse would immediately follow the pace pulse, the measurement for PATIENT AUXILIARY CURRENT (d.c. offset) is performed just before the start of the pace pulse so that the "recharge" pulse is not included in the measurement.

AMDT
No. 1
MAR.
1999

Page 55

36 Electromagnetic compatibility

Add, after the first paragraph on page 57, the following:

This test is intended to demonstrate that the insulation resistance is high enough and the capacitance is low enough that the energy is not conducted through the ENCLOSURE to the circuitry causing inappropriate energy to be sent to the APPLIED PART.

Page 59

Add the following new subclause:

51.1 Intentional exceeding of safety limits

If high pacing rates are used in specific circumstances, extra precautions should be taken to prevent accidental high rate stimulation and to prevent the EQUIPMENT from being inadvertently left with the runaway rate protection feature disabled.

Page 61

Replace 51.104 a) by 51.104.

Delete 51.104 b).

Replace 51.106 by the following:

51.106 If DUAL CHAMBER modes incorporating atrial-synchronous ventricular pacing are available in the EQUIPMENT, a means should be provided to limit the ventricular pacing rate in response to sensed atrial activity to prevent deterioration of the haemodynamic state of the PATIENT. This value is independent of the runaway limit which is intended to prevent an excessively high pacing rate in the event of a SINGLE FAULT CONDITION.

Delete 56.3 aa).

AS/NZS 3200.2.31:1996

IEC 601-2-31:1994

Australian/New Zealand Standard[®]

**Approval and test specification—
Medical electrical equipment**

**Part 2.31: Particular requirements
for safety—External cardiac
pacemakers with internal power
source**

[IEC title: Medical electrical equipment, Part 2: Particular requirements
for the safety of external cardiac pacemakers with internal power
source]

STANDARDS AUSTRALIA ✓

STANDARDS NEW ZEALAND



Documents

Document Sharing Hub

AS/NZS 3200.2.31:1996

This Joint Australian/New Zealand Standard was prepared by Joint Technical Committee HT/15, Medical Electrical Equipment General Safety Aspects. It was approved on behalf of the Council of Standards Australia on 21 October 1995 and on behalf of the Council of Standards New Zealand on 13 November 1995. It was published on 5 March 1996.

The following interests are represented on Committee HT/15:

Australasian College of Physical Scientists and Engineers in Medicine
 Australian and New Zealand College of Anaesthetists
 Australian Federation for Medical and Biological Engineering
 Australian Private Hospitals Association
 Canterbury Health, New Zealand
 Department of Defence, Australia
 Department of Health, N.S.W.
 Department of Health, W.A.
 Department of Human Services and Health, Australia
 Department of Veterans Affairs, Australia
 Institute of Biomedical Engineering, Australia
 Medical Industry Association of Australia
 Ministry of Commerce, New Zealand
 Monash Medical Centre, Vic.
 Public Works Department, N.S.W.
 Queensland Health
 Society for Medical and Biological Engineering
 Wellington Area Health Board, New Zealand

Review of Standards. To keep abreast of progress in industry, Joint Australian/New Zealand Standards are subject to periodic review and are kept up to date by the issue of amendments or new editions as necessary. It is important therefore that Standards users ensure that they are in possession of the latest edition, and any amendments thereto.

Full details of all Joint Standards and related publications will be found in the Standards Australia and Standards New Zealand Catalogue of Publications; this information is supplemented each month by the magazines 'The Australian Standard' and 'Standards New Zealand', which subscribing members receive, and which give details of new publications, new editions and amendments, and of withdrawn Standards.

Suggestions for improvements to Joint Standards, addressed to the head office of either Standards Australia or Standards New Zealand, are welcomed. Notification of any inaccuracy or ambiguity found in a Joint Australian/New Zealand Standard should be made without delay in order that the matter may be investigated and appropriate action taken.

AS/NZS 3200.2.31:1996

Australian/New Zealand Standard[®]

**Approval and test specification—
Medical electrical equipment**

**Part 2.31: Particular requirements
for safety—External cardiac
pacemakers with internal power
source**

PUBLISHED JOINTLY BY:

STANDARDS AUSTRALIA
1 The Crescent,
Homebush NSW 2140 Australia

STANDARDS NEW ZEALAND
Level 10, Standards House,
155 The Terrace,
Wellington 6001 New Zealand

ISBN 0 7337 0178 7



Documents

Document Sharing Hub

PREFACE

This Standard was prepared by the Joint Standards Australia/Standards New Zealand Committee HT/15 on Medical Electrical Equipment General Safety Aspects, as a Joint Standard.

This Standard is identical with and has been reproduced from IEC 601-2-31:1994 *Medical electrical equipment Part 2: Particular requirements for the safety of external cardiac pacemakers with internal power source*.

This Standard is one of a series of Approval and Test Specifications issued by Standards Australia and Standards New Zealand for various categories of medical equipment. It is supplementary to AS 3200.1/NZS 6150:1990, *Approval and test specification—Medical electrical equipment, Part 1: General requirements for safety*.

The international Standard IEC 601-2-31 modifies and supplements the corresponding Clauses of IEC 601-1:1988, *Medical electrical equipment, Part 1: General requirements for safety* which has been adopted as AS 3200.1 (NZS 6150) hereinafter referred to as the General Standard. The requirements of a Particular Standard take priority, where appropriate, over those of the General Standard.

In the text of this Standard, the following print types are used:

- (i) Requirements, compliance with which can be tested and definitions in large roman type
- (ii) Explanations, advice, introductions, general statements, exceptions and references in smaller roman type
- (iii) Headings of sub-clauses and text specifications in italic
- (iv) Terms used throughout the Standard, which have been defined in Clause 2 and which are also in the index IN SMALL CAPITALS

* An asterisk is placed before each Clause for which rationale is included in Annex AA.

Under arrangements made between Standards Australia/Standards New Zealand and the IEC, as well as certain other Standards organizations, users of this Standard are advised that the number of this Standard is not reproduced on each page; its identity is shown only on the cover.

For the purpose of this Standard, the IEC text should be modified as follows:

- (a) *Terminology* The words ‘this Australian/New Zealand Standard’ should replace the words ‘this International Standard’ wherever they appear.
- (b) *Decimal marker* Substitute a full point for a comma where it appears as a decimal marker.
- (c) *References* The references to international Standards should be replaced by references to the following Australian or Joint Australian/New Zealand Standards:

Reference to International Standard or other publication		Australian/New Zealand Standard	
IEC		AS	
86	Primary batteries	2176	Primary batteries
86-1	Part 1: General		
86-2	Part 2: Specification sheets		
601	Medical electrical equipment	3200	Approval and test specification—Medical electrical equipment
601-1	Part 1: General requirements for safety	3200.1/NZS 6150 Part 1:	General requirements for safety
601-1/Amdt 1	Amendment No. 1 (1991)	3200.1/Amdt 1	Amendment No. 1 (1992)

iii

IEC		AS/NZS	
601-1-2	Part 1: General requirements for safety. 2. Collateral standard: Electromagnetic compatibility—Requirements and tests	3200.1.2	General requirements for safety—Collateral Standard: Electromagnetic compatibility—Requirements and tests
801	Electromagnetic compatibility for industrial measurement and control equipment process	—	
801-1-2	Part 2:Electrostatic discharge requirements	—	
ISO			
5841	Cardiac pacemakers	—	
5841-1	Part 1:Implantable pacemakers	—	

The term 'informative' has been used in this Standard to define the application of the Annex to which it applies. An 'informative' Annex is for information and guidance only.

© Copyright – STANDARDS AUSTRALIA/STANDARDS NEW ZEALAND

Users of Standards are reminded that copyright subsists in all Standards Australia and Standards New Zealand publications and software. Except where the Copyright Act allows and except where provided for below no publications or software produced by Standards Australia or Standards New Zealand may be reproduced, stored in a retrieval system in any form or transmitted by any means without prior permission in writing from Standards Australia or Standards New Zealand. Permission may be conditional on an appropriate royalty payment. Australian requests for permission and information on commercial software royalties should be directed to the head office of Standards Australia. New Zealand requests should be directed to Standards New Zealand.

Up to 10 percent of the technical content pages of a Standard may be copied for use exclusively in-house by purchasers of the Standard without payment of a royalty or advice to Standards Australia or Standards New Zealand.

Inclusion of copyright material in computer software programs is also permitted without royalty payment provided such programs are used exclusively in-house by the creators of the programs.

Care should be taken to ensure that material used is from the current edition of the Standard and that it is updated whenever the Standard is amended or revised. The number and date of the Standard should therefore be clearly identified.

The use of material in print form or in computer software programs to be used commercially, with or without payment, or in commercial contracts is subject to the payment of a royalty. This policy may be varied by Standards Australia or Standards New Zealand at any time.

CONTENTS

Clause	SECTION ONE: GENERAL	Page
1	Scope and object	1
2	Terminology and definitions	2
5	Classification	3
6	Identification, marking and documents	3
7	Power input	6
SECTION TWO: ENVIRONMENTAL CONDITIONS		
SECTION THREE: PROTECTION AGAINST ELECTRICAL SHOCK HAZARDS		
14	Requirements related to classification	7
19	Continuous LEAKAGE CURRENTS and PATIENT AUXILIARY CURRENTS	7
SECTION FOUR: PROTECTION AGAINST MECHANICAL HAZARDS		
SECTION FIVE: PROTECTION AGAINST HAZARDS FROM UNWANTED OR EXCESSIVE RADIATION		
36	Electromagnetic compatibility	8
SECTION SIX: PROTECTION AGAINST HAZARDS OF IGNITION OF FLAMMABLE ANAESTHETIC MIXTURES		
SECTION SEVEN: PROTECTION AGAINST EXCESSIVE TEMPERATURES AND OTHER SAFETY HAZARDS		
44	Overflow, spillage, leakage, humidity, ingress of liquids, cleaning, sterilization and disinfection	9
49	Interruption of the power supply	10
SECTION EIGHT: ACCURACY OF OPERATING DATA AND PROTECTION AGAINST HAZARDOUS OUTPUT		
50	Accuracy of operating data	11
51	Protection against hazardous output	12