

AS/NZS 3824:1998
IEC 61859:1997

AS/NZS 3824

Australian/New Zealand Standard™

Guidelines for radiotherapy
treatment rooms design

STANDARDS AUSTRALIA

STANDARDS NEW ZEALAND

Documents
AVA Document Sharing Hub

AS/NZS 3824:1998

This Joint Australian/New Zealand Standard was prepared by Joint Technical Committee HE/3, Medical Electrical Equipment. It was approved on behalf of the Council of Standards Australia on 1 May 1998 and on behalf of the Council of Standards New Zealand on 7 May 1998. It was published on 5 July 1998.

The following interests are represented on Committee HE/3:

Australasian College of Physical Scientists and Engineers in Medicine
 Australasian Society for Ultrasound in Medicine
 Australian Dental Association
 Australian Institute of Radiography
 Australian Radiation Laboratory
 Australian Society of Anaesthetists
 College of Biomedical Engineering Institution of Engineers Australia
 Commonwealth Department of Health and Family Services
 Department of Defence (Australia)
 Medical Industry Association of Australia
 Ministry of Commerce, New Zealand
 Royal Australasian College of Surgeons
 Royal Australasian College of Radiologists

Review of Standards. To keep abreast of progress in industry, Joint Australian/New Zealand Standards are subject to periodic review and are kept up to date by the issue of amendments or new editions as necessary. It is important therefore that Standards users ensure that they are in possession of the latest edition, and any amendments thereto.

Full details of all Joint Standards and related publications will be found in the Standards Australia and Standards New Zealand Catalogue of Publications; this information is supplemented each month by the magazines 'The Australian Standard' and 'Standards New Zealand', which subscribing members receive, and which give details of new publications, new editions and amendments, and of withdrawn Standards.

Suggestions for improvements to Joint Standards, addressed to the head office of either Standards Australia or Standards New Zealand, are welcomed. Notification of any inaccuracy or ambiguity found in a Joint Australian/New Zealand Standard should be made without delay in order that the matter may be investigated and appropriate action taken.

AS/NZS 3824:1998

Australian/New Zealand Standard™

**Guidelines for radiotherapy
treatment rooms design**

First published as AS/NZS 3824:1998.

Published jointly by:

Standards Australia
1 The Crescent,
Homebush NSW 2140 Australia

Standards New Zealand
Level 10, Radio New Zealand House,
155 The Terrace,
Wellington 6001 New Zealand

PREFACE

This Standard was prepared by the Joint Standards Australia/Standards New Zealand Committee HE/3, Medical Electrical Equipment.

This Standard is identical with and has been reproduced from IEC 61859:1997, *Guidelines for radiotherapy treatment rooms design*.

The even numbered pages of the original publication are in French and are omitted from this version. Some other pages of the original, which relate to IEC administrative matters, are also omitted from this edition.

As this publication has been reproduced from an International Standard, the following modifications apply:

- (a) Its number does not appear on each page of text and its identity is shown on the cover and title page.
- (b) The words ‘this Australian/New Zealand Standard’ should replace the words ‘this International Standard’ wherever they appear.
- (c) Substitute a full point for a comma where it appears as a decimal marker.

The references to International Standards should be replaced by references to the following Australian or Joint Australian/New Zealand Standards:

<i>Reference to International Standard or other Publication</i>		<i>Australian/New Zealand Standard</i>	
IEC		AS/NZS	
60601-1-2	Medical electrical equipment Part 1: General requirements for safety— 2. Collateral standard: Electromagnetic compatibility— Requirements and tests	3200	Approval and test specification—Medical electrical equipment
		3200.1.2	Part 1.2: General requirements for safety—Collateral Standard: Electromagnetic compatibility— Requirements and tests
60601-2-1	Medical electrical equipment Part 2: Particular requirements for medical electron accelerators in the range 1 MeV to 50 MeV— Section One: General Section Two: Radiation safety for equipment	—	
60601-2-8	Medical electrical equipment Part 2: Particular requirements for the safety of therapeutic X-ray generators	3200	Approval and test specification—Medical electrical equipment
		3200.2.8	Part 2.8: Particular requirements for safety—Therapeutic X-ray generators
60601-2-11	Medical electrical equipment Part 2: Particular requirements for the safety of gamma beam radiotherapy equipment	—	

iii

IEC	AS/NZS
60601-2-17 Medical electrical equipment Part 2: Particular requirements for the safety of remote-controlled automatically driven gamma-ray afterloading equipment	3200 Approval and test specifications—Medical electrical equipment 3200.2.17 Part 2.17: Particular requirements for safety—Remote-controlled automatically-driven gamma-ray afterloading equipment

Clause 10 of the IEC document is headed 'Reference documents'. The documents listed are not International Standards and do not appear in the text of the document. More appropriately, the clause should be considered to be a bibliography.

© Copyright – STANDARDS AUSTRALIA/STANDARDS NEW ZEALAND

Users of Standards are reminded that copyright subsists in all Standards Australia and Standards New Zealand publications and software. Except where the Copyright Act allows and except where provided for below no publications or software produced by Standards Australia or Standards New Zealand may be reproduced, stored in a retrieval system in any form or transmitted by any means without prior permission in writing from Standards Australia or Standards New Zealand. Permission may be conditional on an appropriate royalty payment. Australian requests for permission and information on commercial software royalties should be directed to the head office of Standards Australia. New Zealand requests should be directed to Standards New Zealand.

Up to 10 percent of the technical content pages of a Standard may be copied for use exclusively in-house by purchasers of the Standard without payment of a royalty or advice to Standards Australia or Standards New Zealand.

Inclusion of copyright material in computer software programs is also permitted without royalty payment provided such programs are used exclusively in-house by the creators of the programs.

Care should be taken to ensure that material used is from the current edition of the Standard and that it is updated whenever the Standard is amended or revised. The number and date of the Standard should therefore be clearly identified.

The use of material in print form or in computer software programs to be used commercially, with or without payment, or in commercial contracts is subject to the payment of a royalty. This policy may be varied by Standards Australia or Standards New Zealand at any time.

CONTENTS

	Page
INTRODUCTION	v
Clause	
1 Scope and object	1
1.1 Scope	1
1.2 Object	1
1.3 Collateral and Particular Standards	1
2 Electrical requirements	2
2.1 SUPPLY MAINS	2
2.2 Isolation from the SUPPLY MAINS	2
2.3 Electromagnetic compatibility	3
2.4 TREATMENT ROOM lighting	3
2.5 Electrical FIXED MAINS SOCKET-OUTLETS	3
3 Environmental requirements	3
3.1 Ventilation	3
3.2 Temperature, pressure and humidity requirements	4
3.3 Air filtration	4
3.4 Water supply	4
3.5 Cleaning requirements	4
3.6 Fire protection	5
3.7 Magnetic field	5
3.8 Floor covering and wall coating	5
4 TREATMENT ROOM entrance system	5
4.1 Entrance barrier system	5
4.2 TREATMENT ROOM entrance INTERLOCKS	5
4.3 DISPLAY of entrance INTERLOCK status	6
4.4 Motorized TREATMENT ROOM door	6
5 Movable floor	6
6 ACCESSORIES handling and storage	6
7 Miscellaneous	7
7.1 Video and audio surveillance	7
7.2 PATIENT alignment system	7
7.3 Indicator lights and warning signs	7
7.4 Access ducts to the TREATMENT ROOM	7
8 Protection against IONIZING RADIATION outside the TREATMENT ROOM	8
9 Information to be given by the RADIOTHERAPY EQUIPMENT MANUFACTURER	8
10 Reference documents	9