

AS/NZS 4543.2:1999

IEC 101331-2:1994

Australian/New Zealand Standard™

**Protective devices against
diagnostic medical X-radiation**

Part 2: Protective glass plates

STANDARDS AUSTRALIA

STANDARDS NEW ZEALAND



Documents

Document Sharing Hub

AS/NZS 4543.2:1999

This Joint Australian/New Zealand Standard was prepared by Joint Technical Committee HE/3, Medical Electrical Equipment. It was approved on behalf of the Council of Standards Australia on 12 February 1999 and on behalf of the Council of Standards New Zealand on 18 January 1999. It was published on 5 March 1999.

The following interests are represented on Committee HT/23:

Australasian College of Physical Scientists and Engineers in Medicine
 Australasian Society for Ultrasound in Medicine
 Australian Dental Association
 Australian Institute of Radiography
 Australian Radiation Laboratory
 Australian Society of Anaesthetists
 College of Biomedical Engineering Institution of Engineers Australia
 Commonwealth Department of Health and Family Services
 Department of Defence (Australia)
 Medical Industry Association of Australia
 Ministry of Commerce, New Zealand
 Royal Australasian College of Surgeons
 Royal Australasian College of Radiologists

Review of Standards. To keep abreast of progress in industry, Joint Australian/New Zealand Standards are subject to periodic review and are kept up to date by the issue of amendments or new editions as necessary. It is important therefore that Standards users ensure that they are in possession of the latest edition, and any amendments thereto.

Full details of all Joint Standards and related publications will be found in the Standards Australia and Standards New Zealand Catalogue of Publications; this information is supplemented each month by the magazines 'The Australian Standard' and 'Standards New Zealand', which subscribing members receive, and which give details of new publications, new editions and amendments, and of withdrawn Standards.

Suggestions for improvements to Joint Standards, addressed to the head office of either Standards Australia or Standards New Zealand, are welcomed. Notification of any inaccuracy or ambiguity found in a Joint Australian/New Zealand Standard should be made without delay in order that the matter may be investigated and appropriate action taken.

This Standard was issued in draft form for comment as DR 95391.



Documents

Document Sharing Hub

AS/NZS 4543.2:1999

Australian/New Zealand Standard™

**Protective devices against
diagnostic medical X-radiation**

Part 2: Protective glass plates

First published as AS/NZS 4543.2:1999.

Published jointly by:

Standards Australia
1 The Crescent,
Homebush NSW 2140 Australia

Standards New Zealand
Level 10, Radio New Zealand House,
155 The Terrace,
Wellington 6001 New Zealand

ISBN 0 7337 2558 9

 **Documents**
Document Sharing Hub

PREFACE

This Standard was prepared by the Joint Standards Australia/Standards New Zealand Committee, HE/3, Medical Electrical Equipment, as a Joint Standard.

This Standard is identical with and has been reproduced from IEC 61331-2:1994, *Protective devices against diagnostic medical X-radiation Part 2: Protective glass plates*.

All IEC Standards have been renumbered as the IEC 60XXX series, therefore references in the source text should be modified by adding 60000 to the number.

Under arrangements made between Standards Australia/Standards New Zealand and the IEC, as well as certain other Standards organizations, users of this Standard are advised that the number of this Standard is not reproduced on each page; its identity is shown only on the cover.

For the purpose of this Standard, the IEC text should be modified as follows:

- (a) *Terminology* The words 'this Australian/New Zealand Standard' should replace the words 'this part of International Standard IEC 1331' or 'this part 2' wherever they appear.
- (b) *Decimal marker* Substitute a full point for a comma where it appears as a decimal marker.
- (c) *References* The references to international Standards should be replaced by references to the following Australian or Joint Australian/New Zealand Standards:

<i>Reference to International Standard or other publication</i>		<i>Australian/New Zealand Standard</i>	
IEC		AS/NZS	
6788	Medical radiology—Terminology	—	
61331	Protective devices against diagnostic medical X-radiation	4543	Protective devices against diagnostic medical X-radiation
61331.1	Part 1: Determination of attenuation properties of materials	4543.1	Part 1: Determination of attenuation properties of materials

The term 'normative' has been used in this Standard to define the application of the annex to which it applies. A 'normative' annex is an integral part of the Standard.

© Copyright – STANDARDS AUSTRALIA/STANDARDS NEW ZEALAND

Users of Standards are reminded that copyright subsists in all Standards Australia and Standards New Zealand publications and software. Except where the Copyright Act allows and except where provided for below no publications or software produced by Standards Australia or Standards New Zealand may be reproduced, stored in a retrieval system in any form or transmitted by any means without prior permission in writing from Standards Australia or Standards New Zealand. Permission may be conditional on an appropriate royalty payment. Australian requests for permission and information on commercial software royalties should be directed to the head office of Standards Australia. New Zealand requests should be directed to Standards New Zealand.

Up to 10 percent of the technical content pages of a Standard may be copied for use exclusively in-house by purchasers of the Standard without payment of a royalty or advice to Standards Australia or Standards New Zealand.

Inclusion of copyright material in computer software programs is also permitted without royalty payment provided such programs are used exclusively in-house by the creators of the programs.

Care should be taken to ensure that material used is from the current edition of the Standard and that it is updated whenever the Standard is amended or revised. The number and date of the Standard should therefore be clearly identified.

The use of material in print form or in computer software programs to be used commercially, with or without payment, or in commercial contracts is subject to the payment of a royalty. This policy may be varied by Standards Australia or Standards New Zealand at any time.

CONTENTS

Clause		Page
1	Scope and object	1
	1.1 Scope	1
	1.2 Object	1
2	Normative references	1
3	Terminology	2
	3.1 Degree of requirements	2
	3.2 Use of terms	2
	3.3 Defined terms	2
4	Dimensions	2
	4.1 Thickness of PROTECTIVE GLASS PLATES	2
	4.2 Plane dimensions of PROTECTIVE GLASS PLATES	3
	4.3 Designation of nominal sizes	4
5	Geometrical accuracy of PROTECTIVE GLASS PLATES	4
	5.1 Squareness	4
	5.2 Flatness	4
	5.3 Parallelism	4
	5.4 Narrow sides	4
	5.5 Edges	4
6	Optical quality of material	5
	6.1 Definition of zones for determining homogeneity	5
	6.2 Bubbles	5
	6.3 Streaks and other inhomogeneities	5
7	Transmittance	6
8	ATTENUATION properties	6
	8.1 Minimum value of ATTENUATION EQUIVALENT	6
	8.2 Determination	6
	8.3 Information	6
	8.4 Verification	7
9	Marking	7
10	ACCOMPANYING DOCUMENTS	7
11	Statement of compliance	7