

AS/NZS IEC 60601.2.35:2022
IEC 60601-2-35:2020



Australian/New Zealand Standard™

Medical electrical equipment

Part 2.35: Particular requirements for the basic safety and essential performance of heating devices using blankets, pads and mattresses and intended for heating in medical use



AS/NZS IEC 60601.2.35:2022

This Joint Australian/New Zealand Standard™ was prepared by Joint Technical Committee HE-003, Medical Electrical Equipment. It was approved on behalf of the Council of Standards Australia on 6 April 2022 and by the New Zealand Standards Approval Board on 4 May 2022.

This Standard was published on 20 May 2022.

The following are represented on Committee HE-003:

- AusBiotech
- Australasian College of Physical Scientists and Engineers in Medicine
- Australian & New Zealand College of Anaesthetists
- Australian College of Perioperative Nurses
- Australian Dental Association
- Australian Radiation Protection and Nuclear Safety Agency
- Australian Society of Anaesthetists
- Certification Body Australia (Certification Interests Australia)
- College of Biomedical Engineering Engineers Australia
- Department of Defence (Australian Government)
- Engineers Australia
- Medical Technology Association of Australia
- Medical Technology Association of New Zealand
- MidCentral District Health Board
- National Clinical Engineering Managers Forum (for District Health Boards)
- Queensland Health
- Therapeutic Goods Administration
- WorkSafe New Zealand

This Standard was issued in draft form for comment as DR AS/NZS IEC 60601.2.35:2022.

Keeping Standards up-to-date

Ensure you have the latest versions of our publications and keep up-to-date about Amendments, Rulings, Withdrawals, and new projects by visiting:

www.standards.org.au

www.standards.govt.nz

ISBN 978 1 76113 755 6

Australian/New Zealand Standard™

Medical electrical equipment

Part 2.35: Particular requirements for the basic safety and essential performance of heating devices using blankets, pads and mattresses and intended for heating in medical use

Originated as AS/NZS 3200.2.35:1999.
Jointly revised and redesignated as AS/NZS IEC 60601.2.35:2022.



© IEC Geneva Switzerland 2022 — All rights reserved

© Standards Australia Limited/the Crown in right of New Zealand, administered by the New Zealand Standards Executive 2022

All rights are reserved. No part of this work may be reproduced or copied in any form or by any means, electronic or mechanical, including photocopying, without the written permission of either the IEC or the publisher, unless otherwise permitted under the Copyright Act 1968 (Cth) or the Copyright Act 1994 (New Zealand). If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please see the contact details on the back cover or the contact us page of the website for further information.

Preface

This Standard was prepared by the Joint Standards Australia/Standards New Zealand Committee HE-003, Medical Electrical Equipment, to supersede AS/NZS 3200.2.35:1999, *Medical electrical equipment, Part 2.35: Particular requirements for safety—Blankets, pads and mattresses intended for heating in medical use*.

The objective of this document is to specify requirements for the basic safety and essential performance of heating devices using blankets, pads or mattresses in medical use, also referred to as ME equipment.

Requirements for heating devices intended to pre-warm a bed are also specified in this document.

This document does not apply to the following:

- (a) Heating devices intended for physiotherapy.
- (b) Infant radiant warmers (for more information, refer to IEC 60601-2-21).
- (c) Infant incubators (for more information, refer to IEC 60601-2-19).
- (d) Infant transport incubators (for more information, refer to IEC 60601-2-20).
- (e) Cooling devices.

This document is identical with, and has been reproduced from, IEC 60601-2-35:2020, *Medical electrical equipment – Part 2-35: Particular requirements for the basic safety and essential performance of heating devices using blankets, pads and mattresses and intended for heating in medical use*.

As this document has been reproduced from an International document, the following applies:

- (i) In the source text “this part of 60601 International Standard” should read “this document”.
- (ii) A full point substitutes for a comma when referring to a decimal marker.

Australian or Australian/New Zealand Standards that are identical adoptions of international normative references may be used interchangeably. Refer to the online catalogue for information on specific Standards.

The terms “normative” and “informative” are used in Standards to define the application of the appendices or annexes to which they apply. A “normative” appendix or annex is an integral part of a Standard, whereas an “informative” appendix or annex is only for information and guidance.

NOTES

CONTENTS

FOREWORD	4
INTRODUCTION	7
201.1 Scope, object and related standards	8
201.2 Normative references	10
201.3 Terms and definitions	10
201.4 General requirements	13
201.5 General requirements for testing ME EQUIPMENT	14
201.6 Classification of ME EQUIPMENT and ME SYSTEMS	14
201.7 ME EQUIPMENT identification, marking and documents	14
201.8 Protection against electrical HAZARDS from ME EQUIPMENT	18
201.9 Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS	24
201.10 Protection against unwanted and excessive radiation HAZARDS	26
201.11 Protection against excessive temperatures and other HAZARDS	26
201.12 Accuracy of controls and instruments and protection against hazardous outputs	28
201.13 HAZARDOUS SITUATIONS and fault conditions for ME EQUIPMENT	33
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS)	38
201.15 Construction of ME EQUIPMENT	38
201.16 ME SYSTEMS	42
201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS	42
202 ELECTROMAGNETIC DISTURBANCES – Requirements and tests	43
208 General requirements, tests and guidance for ALARM SYSTEMS IN MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS	43
210 * Requirements for the development of PHYSIOLOGIC CLOSED-LOOP CONTROLLERS	43
Annexes	44
Annex D (informative) Symbols on marking	44
Annex AA (informative) Particular guidance and rationale	45
Annex BB (normative) Determination of the LAGGING MATERIAL	57
Annex CC (normative) * Determination of heat transfer towards the PATIENT	58
Annex DD (normative) * Determination of heat transfer away from the PATIENT	60
Annex EE (normative) CONDITIONS OF ADEQUATE HEAT DISCHARGE	61
Annex FF (normative) Test procedure for maximum CONTACT SURFACE TEMPERATURE for FORCED AIR DEVICES	62
Annex GG (normative) Test procedure for maximum CONTACT SURFACE TEMPERATURE for FORCED AIR DEVICES under SINGLE FAULT CONDITION	64
Annex HH (normative) Safety test procedure for average CONTACT SURFACE TEMPERATURE for FORCED AIR DEVICES	65
Bibliography	67
Index of defined terms used in this document	69

Figure 201.101 – Positioning of temperature sensors on the contact surface of the
heated area of a HEATING DEVICE (see 201.12.4.101 and 201.12.4.105)..... 11

Figure 201.102 – Example of the positioning of temperature sensors on the contact surface of the heated areas of a HEATING DEVICE having more than one separately heated area	12
Figure 201.103 – Apparatus for the spark ignition test	22
Figure 201.104 – Ramp for the impact test on PADS	24
Figure 201.105 – Partial covering conditions.....	26
Figure 201.106 – Method of folding BLANKETS	35
Figure 201.107 – Examples of folds	37
Figure 201.108 – Positions of a BLANKET for the RUCK-RESISTANCE test.....	42
Figure AA.1 – Illustration of the main requirements of this document	45
Figure HH.1 – Sensor locations – Average CONTACT SURFACE TEMPERATURE.....	66
Table 201.101 – * Additional ESSENTIAL PERFORMANCE requirements	14
Table 201.102 – Temperature limits in dependency to time.....	39

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-35: Particular requirements for the basic safety and essential performance of heating devices using blankets, pads or mattresses and intended for heating in medical use

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 60601-2-35 has been prepared by IEC technical committee 62D: Electromedical equipment, of IEC technical committee 62: Electrical equipment in medical practice.

This second edition cancels and replaces IEC 80601-2-35 published in 2009 and Amendment 1:2016.

The text of this particular standard is based on the following documents:

FDIS	Report on voting
62D/1765/FDIS	62D/1777/RVD

Full information on the voting for the approval of this International Standard can be found in the report on voting indicated in the above table.

This document has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this document, the following print types are used:

- requirements and definitions: roman type;
- *test specifications: italic type;*
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type;
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this document, the term

- "clause" means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- "subclause" means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this document are preceded by the term "Clause" followed by the clause number. References to subclauses within this particular standard are by number only.

In this document, the conjunctive "or" is used as an "inclusive or" so a statement is true if any combination of the conditions is true.

The verbal forms used in this document conform to usage described in Clause 7 of the ISO/IEC Directives, Part 2. For the purposes of this document, the auxiliary verb:

- "shall" means that compliance with a requirement or a test is mandatory for compliance with this document;
- "should" means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this document;
- "may" is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

A list of all parts of the 60601 International Standard, published under the general title *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under "<http://webstore.iec.ch>" in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

NOTE The attention of users of this document is drawn to the fact that equipment manufacturers and testing organizations may need a transitional period following publication of a new, amended or revised IEC publication in which to make products in accordance with the new requirements and to equip themselves for conducting new or revised tests. It is the recommendation of the committees that the content of this publication be adopted for implementation nationally not earlier than 3 years from the date of publication.

IMPORTANT – The "colour inside" logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this publication using a colour printer.

INTRODUCTION

The minimum safety requirements specified in this particular standard are considered to provide for a practical degree of safety in the operation for HEATING DEVICES using BLANKETS, PADS or MATTRESSES and intended for heating in medical use.

While K (degree Kelvin) is the recognized unit and symbol for absolute temperature and temperature difference, °C has been used throughout this document because all measurements are commonly made using equipment marked with the Celsius temperature scale.

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-35: Particular requirements for the basic safety and essential performance of heating devices using blankets, pads or mattresses and intended for heating in medical use

201.1 Scope, object and related standards

Clause 1 of the general standard¹ applies, except as follows:

201.1.1 * Scope

Replacement:

This part of 60601 International Standard applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of HEATING DEVICES using BLANKETS, PADS or MATTRESSES in medical use, also referred to as ME EQUIPMENT. HEATING DEVICES intended to prewarm a bed are included in the scope of this document.

If a clause or subclause is specifically intended to be applicable to ME EQUIPMENT only, or to ME SYSTEMS only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to ME EQUIPMENT and to ME SYSTEMS, as relevant.

If a clause or subclause is specifically intended to be applicable to a specifically defined type of ME EQUIPMENT, as is the case with FORCED AIR DEVICES, then the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to ME EQUIPMENT and to ME SYSTEMS, as relevant.

HAZARDS inherent in the intended physiological function of ME EQUIPMENT or ME SYSTEMS within the scope of this document are not covered by specific requirements in this document, except in 7.2.13 and 8.4.1 of the general standard.

NOTE See also 4.2 of the general standard.

This document does not apply to:

- HEATING DEVICES intended for physiotherapy;
- INFANT RADIANT WARMERS; for information, see IEC 60601-2-21 [1]²;
- INFANT INCUBATORS; for information, see IEC 60601-2-19 [2];
- INFANT TRANSPORT INCUBATORS, for information, see IEC 60601-2-20 [3];
- cooling devices.

¹ The general standard is IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012, Medical electrical equipment – Part 1: General requirements for basic safety and essential performance.

² Figures in square brackets refer to the Bibliography.

201.1.2 Object

Replacement:

The object of this particular standard is to establish particular BASIC SAFETY and ESSENTIAL PERFORMANCE requirements, which minimize HAZARDS to PATIENTS, and OPERATORS for HEATING DEVICES using BLANKETS, PADS or MATTRESSES and intended for heating in medical use and to specify tests for demonstrating compliance with these requirements.

201.1.3 Collateral standards

Addition:

This particular standard refers to those applicable collateral standards that are listed in Clause 2 of the general standard and Clause 201.2 of this particular standard.

IEC 60601-1-2:2014, IEC 60601-1-8:2006 and IEC 60601-1-8:2006/AMD1:2012, and IEC 60601-1-10:2007 and IEC 60601-1-10:2007/AMD1:2013 apply as modified in Articles 202, 208 and 210 respectively. IEC 60601-1-3 does not apply. All other published collateral standards in the IEC 60601-1 series apply as published.

201.1.4 Particular standards

Replacement:

In the IEC 60601 series, particular standards may modify, replace or delete requirements contained in the general standard and collateral standards as appropriate for the particular ME EQUIPMENT under consideration, and may add other BASIC SAFETY and ESSENTIAL PERFORMANCE requirements.

A requirement of a particular standard takes priority over the general standard.

For brevity, IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012 are referred to in this particular standard as the general standard. Collateral standards are referred to by their document number.

The numbering of clauses and subclauses of this document corresponds to that of the general standard with the prefix "201" (e.g. 201.1 in this document addresses the content of Clause 1 of the general standard) or applicable collateral standard with the prefix "20x" where x is the final digit(s) of the collateral standard document number (e.g. 202.4 in this document addresses the content of Clause 4 of the IEC 60601-1-2 collateral standard, 203.4 in this document addresses the content of Clause 4 of the IEC 60601-1-3 collateral standard, etc.). The changes to the text of the general standard and applicable collateral standards are specified by the use of the following words:

"*Replacement*" means that the clause or subclause of the general standard or applicable collateral standard is replaced completely by the text of this particular standard.

"*Addition*" means that the text of this particular standard is additional to the requirements of the general standard or applicable collateral standard.

"*Amendment*" means that the clause or subclause of the general standard or applicable collateral standard is amended as indicated by the text of this particular standard.

Subclauses, figures or tables which are additional to those of the general standard are numbered starting from 201.101. However due to the fact that definitions in the general standard are numbered 3.1 through 3.147, additional definitions in this document are numbered beginning from 201.3.201. Additional annexes are lettered AA, BB, etc., and additional items aa), bb), etc.

Subclauses, figures or tables which are additional to those of a collateral standard are numbered starting from 20x, where "x" is the number of the collateral standard, e.g. 202 for IEC 60601-1-2, 203 for IEC 60601-1-3, etc.

The term "this document" is used to make reference to the general standard, any applicable collateral standards and this particular standard taken together.

Where there is no corresponding clause or subclause in this particular standard, the clause or subclause of the general standard or applicable collateral standard, although possibly not relevant, applies without modification; where it is intended that any part of the general standard or applicable collateral standard, although possibly relevant, is not to be applied, a statement to that effect is given in this particular standard.

201.2 Normative references

NOTE Informative references are listed in the Bibliography.

Clause 2 of the general standard applies, except as follows:

Addition:

IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*
IEC 60601-1:2005/AMD1:2012

IEC 60601-1-10:2007, *Medical electrical equipment – Part 1-10: General requirements for basic safety and essential performance – Collateral Standard: Requirements for the development of physiologic closed-loop controllers*
IEC 60601-1-10:2007/AMD1:2013

Replacement:

IEC 60384-14:2013, *Fixed capacitors for use in electronic equipment – Part 14: Sectional specification – Fixed capacitors for electromagnetic interference suppression and connection to the supply mains*
IEC 60384-14:2013/AMD1:2016

IEC 60601-1-2:2014, *Medical electrical equipment – Part 1-2: General requirements for basic safety and essential performance – Collateral standard: Electromagnetic disturbances – Requirements and tests*

201.3 Terms and definitions

For the purposes of this document, the terms and definitions specified in IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012, and the following apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <http://www.iso.org/obp>

NOTE An index of defined terms used in this document is found beginning on page 69.

Addition:

201.3.201.1

BLANKET

<other than FORCED AIR DEVICES> APPLIED PART of HEATING DEVICE, which can be folded, for use under or over a PATIENT

201.3.201.2

BLANKET

<FORCED AIR DEVICES> APPLIED PART of HEATING DEVICE intended to be used with a CONTROLLER to transfer thermal energy to all or part of the body of a PATIENT

201.3.202

CONDITIONS OF ADEQUATE HEAT DISCHARGE

conditions achieved when a HEATING DEVICE is supported and covered as specified in Annex EE


201.3.203.1

CONTACT SURFACE TEMPERATURE

<other than FORCED AIR DEVICES> temperature T_R at the reference point of the heated APPLIED PART

Note 1 to entry: See Figures 201.101 and 201.102.

Note 2 to entry: The CONTACT SURFACE TEMPERATURE for FORCED AIR DEVICES is measured by the test methods described in Annexes FF, GG and HH.



IEC


Key

T_R CONTACT SURFACE TEMPERATURE reference point on the contact surface

Some HEATING DEVICES may have unheated areas shown in the following figures as:



Figure 201.101 – Positioning of temperature sensors on the contact surface of the heated area of a HEATING DEVICE
(see 201.12.4.101 and 201.12.4.105)



IEC

The temperature at the centre point of any one of the heated areas closest to the centre of the HEATING DEVICE (in the example shown above 2, 3, 6, or 7) is treated as T_R .

Figure 201.102 – Example of the positioning of temperature sensors on the contact surface of the heated areas of a HEATING DEVICE having more than one separately heated area

201.3.203.2

CONTACT SURFACE TEMPERATURE

< FORCED AIR DEVICES> temperature resulting from the heat transferred to a target surface by the APPLIED PART

201.3.204

CONTROLLER

part of a HEATING DEVICE intended to supply and control thermal energy to a BLANKET, PAD or MATTRESS

Note 1 to entry: This includes the HOSE, if present.

201.3.205

FORCED AIR DEVICE

HEATING DEVICE that uses air as the heat transfer medium to warm a PATIENT and is comprised of a CONTROLLER and a BLANKET

201.3.206

FREE HOSING

hazardous practice or condition of using the CONTROLLER without a BLANKET

201.3.207

HEATING DEVICE

ME EQUIPMENT intended to supply heat to the whole or part of the body of a PATIENT by means of heated BLANKETS, PADS, or MATTRESSES

201.3.208

HIGH HEAT TRANSFER

thermal characteristic of a HEATING DEVICE as determined according to Annex CC or Annex DD

201.3.209

HOSE

component of the CONTROLLER that is the conduit for the heat transfer medium to and/or from the BLANKET, PAD or MATTRESS

201.3.210**INFANT**

PATIENT up to the age of three months and with a weight less than 10 kg

201.3.211**LAGGING MATERIAL**

polyurethane or polystyrene insulation material used in the test methods of this specification to assist in the determination of temperature

Note 1 to entry: Specifications for LAGGING MATERIAL are given in Annexes BB and FF.

201.3.212**LOW HEAT TRANSFER**

thermal characteristic of a HEATING DEVICE as determined according to Annex CC or Annex DD

201.3.213**MATTRESS**

APPLIED PART of a HEATING DEVICE, which provides resilient support to the whole body of a PATIENT

201.3.214**NOZZLE**

end of the HOSE that connects to the BLANKET, PAD or MATTRESS

201.3.215**OVER-BLANKET**

BLANKET designed to be used over a PATIENT

201.3.216**PAD**

APPLIED PART of HEATING DEVICE, which can be bent but not folded

201.3.217**RUCK**

unintended fold in a normally even surface

201.3.218**RUCK-RESISTANT BLANKET**

BLANKET having a construction such that RUCKING of the flexible part is unlikely

201.3.219**UNDER-BLANKET**

BLANKET designed to be used under a PATIENT

201.4 General requirements

Clause 4 of the general standard applies except as follows:

201.4.3 ESSENTIAL PERFORMANCE

Addition:

201.4.3.101 Additional requirements for ESSENTIAL PERFORMANCE

Additional ESSENTIAL PERFORMANCE requirements are found in the subclauses listed in Table 201.101.

Table 201.101 – * Additional ESSENTIAL PERFORMANCE requirements

Requirement	Subclause
ESSENTIAL PERFORMANCE requirement 1	201.12.4.104 or generation of a TECHNICAL ALARM CONDITION in compliance with 201.12.3.102

201.4.5 Alternative RISK CONTROL measures or test methods for me equipment or me systems

Addition:

This particular standard specifies safety requirements for HEATING DEVICES using BLANKETS, PADS or MATTRESSES, but alternate methods of compliance with a specific clause or subclause by demonstrating equivalent safety will not be judged non-compliant if the MANUFACTURER has demonstrated in his RISK MANAGEMENT FILE that the RISKS presented by the HAZARDS are of an acceptable level when weighed against the benefits of treatment using the device.

Additional subclause:

201.4.101 Combination of equipment

For equipment which combines several heat sources, the safety requirements of other relevant particular standards shall be considered. Further, the safety requirements of this particular standard shall be fulfilled with the combination of the other equipment, which is approved by the MANUFACTURER as stated in the instructions for use according to Clause 16 of the general standard (ME SYSTEMS).

201.5 General requirements for testing ME EQUIPMENT

Clause 5 of the general standard applies.

201.6 Classification of ME EQUIPMENT and ME SYSTEMS

Clause 6 of the general standard applies, except as follows:

201.6.2 Protection against electric shock

Replacement of the last paragraph with the following new paragraph:

HEATING DEVICES shall have TYPE BF APPLIED PARTS or TYPE CF APPLIED PARTS.

201.7 ME EQUIPMENT identification, marking and documents

Clause 7 of the general standard applies, except as follows:

201.7.2.1 Minimum requirements for marking on ME EQUIPMENT and on interchangeable parts

Addition:

201.7.2.1.101 * Additional minimum requirements for marking on ME EQUIPMENT and on interchangeable parts

201.7.2.1.101.1 HEATING DEVICES (other than for FORCED AIR DEVICES)

A HEATING DEVICE shall be marked as follows:

- a) to indicate how it is intended to be positioned in NORMAL USE, whether:
 - over the PATIENT;
 - under the PATIENT;
 - directly in contact with the PATIENT;
 - separated from the PATIENT by an intermediate layer, or layers (for example by a water-bed, other type of MATTRESS, or bedclothes);
 - it has to be used flat (that is without creases), or whether it can be wrapped around the PATIENT;
- b) to warn against possible HAZARDOUS SITUATIONS from penetration by sharp objects;
- c) to warn against possible HAZARDOUS SITUATIONS from folding or methods of storage, other than those specified by the MANUFACTURER;
- d) to warn against the possibility of a HAZARDOUS SITUATION if a partial covering is caused by pillows or other items having good thermal insulation being laid over part of the HEATING DEVICE;
- e) to specify, in the case of liquid-filled MATTRESSES intended to be used above a PAD, the minimum amount of liquid to which the MATTRESS should be filled, and a warning that if this minimum amount is not present a PATIENT could suffer a burn;
- f) to specify, in the case of a HEATING DEVICE supplied or controlled by an external unit, that the HEATING DEVICE shall only be used with the external unit specified by the MANUFACTURER of the HEATING DEVICE.

201.7.2.1.101.2 CONTROLLERS

A CONTROLLER for FORCED AIR DEVICES shall be marked as follows.

- a) The HOSE shall be marked within 15 cm of the NOZZLE to caution that the NOZZLE needs to be connected to a BLANKET. The following statement and the safety sign ISO 7010-M002 (see IEC 60601-1:2005, Table D.2, safety sign 10) shall accompany the "no FREE HOSE" safety sign shown in Annex D of this particular standard:
 "CAUTION! HOSE NOZZLE SHALL be connected to a compatible forced air BLANKET or thermal injury may occur."
- b) A caution that allowing the HOSE to contact the PATIENT can lead to thermal injury, if appropriate.
- c) A warning against using the device distal to arterial cross clamping and that non-observance can lead to thermal injury.

201.7.2.1.101.3 * Temperature sensors

A temperature sensor which is designed to be attached to or inserted into a PATIENT shall have its intended use identified clearly and unambiguously on or adjacent to the sensor.

201.7.2.1.101.4 APPLIED PARTS (for other than FORCED AIR DEVICES) with a large unheated border

For an APPLIED PART having an unheated border wider than 30 mm around the heated area, and where the requirements of 201.12.4.101 are not satisfied, the outer boundary of the heated area shall be marked on both sides of the APPLIED PART.

201.7.4.2 Control devices

Addition:

201.7.4.2.101 Additional requirements for control devices

NOTE See also 201.12.

Where provided, on ME EQUIPMENT other than FORCED AIR DEVICES, a control for setting the CONTACT SURFACE TEMPERATURE of a HEATING DEVICE shall indicate the temperature in intervals not greater than 1 °C (see also 201.12.1.101).

For FORCED AIR DEVICES, each heated temperature control position shall be marked in degrees C. Such marking shall be CLEARLY LEGIBLE.

201.7.9.2.2 Warning and safety notices

Addition:

201.7.9.2.2.101 Additional requirements for warning and safety notices

The instructions for use shall additionally contain the following:

- a) a strong recommendation that the surface of the HEATING DEVICE should be checked for freedom from mechanical damage prior to each application;
- b) an indication for the use of parts of HEATING DEVICES which are intended to be used together;
- c) statements, details and warnings on the use of the HEATING DEVICE in combination with other heat sources, if applicable;
- d) a warning statement that the use of materials of good thermal conductivity, such as water, gel and similar substances, with the HEATING DEVICE not switched on can decrease the temperature of the body of a PATIENT;
- e) a warning statement regarding the RISK of electrical shock, burns or electromagnetic interference with use of high frequency (HF) surgical instruments or endocardial catheters while a HEATING DEVICE is in use, if applicable;
- f) a statement that the OPERATOR should monitor the temperature of the PATIENT at regular intervals;
- g) a statement that means for drainage of liquid from a liquid-filled HEATING DEVICE is required (see also 201.11.6.5.101);
- h) a warning that a means for retaining a PATIENT either on or under a HEATING DEVICE may need to be used and that the means for retaining a PATIENT should not block the fluid pathways of the HEATING DEVICE;
- i) a warning that warming transdermal medications (patches) can increase drug delivery, resulting in possible HARM to the PATIENT;
- j) a statement that the HEATING DEVICE contains an ALARM SYSTEM with an interruption of power supply/SUPPLY MAINS ALARM CONDITION;
- k) for CONTROLLERS for FORCED AIR DEVICES, the following warning statements:
 - a caution that the HOSE, if allowed to contact the PATIENT, can lead to thermal injury, if appropriate;

- a description of system operating modes and PATIENT conditions where the HEATING DEVICE can be safely used;
- a statement that the OPERATOR should monitor the temperature of the PATIENT at regular intervals;
- * a statement that the HOSE NOZZLE needs to be connected to a BLANKET. The following statement shall accompany the "no FREE HOSING" safety sign shown in Annex D:
"CAUTION! HOSE NOZZLE SHALL be connected to a compatible forced air BLANKET or thermal injury may occur."

201.7.9.2.9 Operating instructions

Addition:

201.7.9.2.9.101 Additional requirements for start-up PROCEDURE

The instructions for use shall include a method for testing the function of the ALARM SYSTEM for each of the ALARM CONDITIONS specified in this document, if not performed automatically during start up.

201.7.9.2.9.102 Additional requirements for operating instructions

The following shall appear in the instructions for use:

- a) the approximate time required for the CONTACT SURFACE TEMPERATURE to heat up from $23\text{ °C} \pm 2\text{ °C}$ to 37 °C , when operated under CONDITIONS OF ADEQUATE HEAT DISCHARGE as specified in Annex EE;
- b) a description of how and when to verify the functionality of the ALARM SYSTEM.

201.7.9.2.9.103 Additional requirements for operating instructions for BLANKETS

The instructions for use for BLANKETS shall contain the following:

- a) a description of system operating modes and PATIENT conditions where the system can be safely used;
- b) a statement that the OPERATOR should monitor the temperature of the PATIENT at regular intervals;
- c) an indication of how the BLANKET is to be positioned during NORMAL USE, for example:
 - i) over the PATIENT,
 - ii) under the PATIENT,
 - iii) directly in contact with the PATIENT,
 - iv) separated from the PATIENT by an intermediate layer, or layers, of material,
 - v) whether it has to be used flat (without creases),
 - vi) whether it can be wrapped around the PATIENT;
- d) *a statement identifying which CONTROLLER(S) can be used safely with the BLANKET. The description shall include, as applicable, the CONTROLLER model number, revision level, product version, options or any other element that could affect the safety of the combination;
- e) *a warning against using the device distal to arterial cross clamping;

NOTE BLANKETS can be used proximal to the clamped artery without any additional RISK.

- f) a warning describing the possible HAZARD of using the device with ischemic limbs.

201.7.9.2.9.104 Additional requirements for operating instructions for temperature sensors

The instructions for use for temperature sensors required for temperature control, which are designed to be attached to, or inserted into, a PATIENT shall have their intended use described.

201.7.9.2.13 Maintenance

Addition:

201.7.9.2.13.101 Additional requirements for maintenance

The following shall appear in the documentation:

- particulars of any necessary calibration procedure(s);
- information as to how to confirm that the independent THERMAL CUT-OUT is operational.

201.8 Protection against electrical HAZARDS from ME EQUIPMENT

Clause 8 of the general standard applies, except as follows:

201.8.1 * Fundamental rule of protection against electric shock

Addition to item b):

- contact between a PATIENT and the heat-transfer fluid caused by a leak from a fluid-filled MATTRESS;
- * perforation of a liquid-filled MATTRESS heated by circulation of liquid from a remote unit;
- perforation of the ENCLOSURE of a HEATING DEVICE separated from the SUPPLY MAINS by a transformer (see 15.5 of the general standard).

201.8.5.1.2 MEANS OF PATIENT PROTECTION (MOPP)

Addition:

201.8.5.1.2.101 * Additional requirements for MEANS OF PATIENT PROTECTION (MOPP)

The electrical circuit within the APPLIED PART shall be isolated from earth by at least one MOPP and from MAINS by at least two MOPP. Where a transformer is used to achieve this isolation, it need not meet 15.5.3.

Compliance is checked by inspection of the HEATING DEVICES and examination of the circuit diagram.

201.8.7.4.7 Measurement of the PATIENT LEAKAGE CURRENT

Addition:

201.8.7.4.7.101 Additional requirements for measurement of the PATIENT LEAKAGE CURRENT

An APPLIED PART consisting of a surface of insulating material is tested using metal foil as mentioned in 8.7.4.7 of the general standard:

- a) for an APPLIED PART large enough for a PATIENT not to be in contact with the whole surface of the APPLIED PART, foil of area 100 cm × 30 cm shall be used; or
- b) for an APPLIED PART whose area is less than 100 cm × 30 cm, the whole APPLIED PART shall be covered by foil.

For a liquid-filled MATTRESS heated by circulation of liquid from a remote unit, the liquid is replaced by isotonic saline (9 g sodium chloride per litre of water). The PATIENT LEAKAGE CURRENT is measured using an electrode 5 mm × 5 mm positioned in the saline in contact with the surface which provides heat to the liquid;

201.8.8 Insulation

Amendment:

This subclause of the general standard does not apply to BLANKETS for FORCED AIR DEVICES that do not contain active electrically conductive wires.

201.8.8.4 Insulation other than wire insulation

Additional subclauses:

201.8.8.4.101 Mechanical and thermal resistance to damage of the ENCLOSURE of the flexible part of HEATING DEVICES

For BLANKETS and PADS consisting of woven material surrounding an electrical heating element, the woven material is tested by the ball-pressure test according to 8.8.4.1 a) of the general standard and additionally by the following ignition test.

Six samples having dimensions of 100 mm × 200 mm are cut from the ENCLOSURE. They are selected from parts of the ENCLOSURE so that no two samples contain the same warp thread or the same weft thread or; if this is not possible, the samples are selected so that the same threads do not appear in more than two samples. Any pieces of heating elements and trimming are removed from the samples.

The test apparatus, as shown in Figure 201.103, has two brass electrodes 3 mm in diameter, which are supported by brass pillars, mounted on a base plate of insulating material so that their axes are aligned. The base plate also supports a platform of insulating material having dimensions of 100 mm × 100 mm, which is located centrally between the brass pillars. Provision shall be made for the height of the platform to be adjusted.