



Designation: D 6378 – 99^{ε1}

An American National Standard

Standard Test Method for Determination of Vapor Pressure (VP_x) of Petroleum Products, Hydrocarbons, and Hydrocarbon-Oxygenate Mixtures (Triple Expansion Method)¹

This standard is issued under the fixed designation D 6378; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

^{ε1} NOTE—Note 17 was corrected editorially in September 1999.

1. Scope

1.1 This test method covers the use of automated vapor pressure instruments to determine the vapor pressure exerted in vacuum by volatile, liquid petroleum products, hydrocarbons, and hydrocarbon-oxygenate mixtures. This test method is suitable for testing samples with boiling points above 0°C (32°F) that exert a vapor pressure between 7 and 150 kPa (1.0 and 21psi) at 37.8°C (100°F) at a vapor-to-liquid ratio of 4:1. Measurements are made on liquid sample sizes in the range from 1 to 10 mL.

NOTE 1—The test method is suitable for the determination of the vapor pressure of volatile, liquid petroleum products at temperatures from 0 to 100 °C at vapor to liquid ratios of 4:1 to 1.1 ($X = 4$ to 1) and pressures up to 500 kPa (70 psi), but the precision statement (see Section 15.) may not be applicable.

1.2 The VP_x determined by this test method at a vapor-liquid ratio of 4:1 ($X = 4$) of gasoline and gasoline-oxygenate blends at 37.8°C can be correlated to the dry vapor pressure equivalent (DVPE) value determined by Test Method D 5191 (see Section 15.2).

1.3 The values stated in SI units are regarded as standard. The inch-pound units given in parentheses are provided for information only.

1.4 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.* For specific hazard statements, see Note 6.

2. Referenced Documents

2.1 ASTM Standards:

D 323 Test Method for Vapor Pressure of Petroleum Prod-

ucts (REID Method)²

D 1160 Test Method for Distillation of Petroleum Products at Reduced Pressure²

D 4057 Practice for Manual Sampling of Petroleum and Petroleum Products³

D 4177 Practice for Automatic Sampling of Petroleum and Petroleum Products³

D 4953 Test Method for Vapor Pressure of Gasoline and Gasoline-Oxygenate Blends (Dry Method)³

D 5191 Test Method for Vapor Pressure of Petroleum Products (Mini Method)³

D 5842 Practice for Sampling and Handling of Fuels for Volatility Measurements⁴

D 5854 Practice for Mixing and Handling of Liquid Samples of Petroleum and Petroleum Products⁴

3. Terminology

3.1 Definitions of Terms Specific to This Standard:

3.1.1 *dry vapor pressure equivalent (DVPE)*—a value calculated by a correlation equation from the total pressure (Test Method D 5191), which is equivalent to the value obtained on the sample by Test Method D 4953, Procedure A.

3.1.2 *partial pressure from dissolved air (PPA)*, n —the pressure exerted in vacuum from dissolved air that escapes from the liquid phase into the vapor phase.

3.1.3 *Reid vapor pressure equivalent (RVPE)*—a value calculated by a correlation equation from the TP_x , which is equivalent to the value obtained on the same by Test Method D 323.

3.1.4 *total pressure (TP_x)*, n —the pressure exerted in vacuum by air- and gas-containing petroleum products, components and feedstocks, and other liquids, in the absence of undissolved water at a vapor-liquid ration of $X:1$.

3.1.5 *vapor pressure (VP_x)*, n —the total pressure minus the PPA in the liquid at a vapor-liquid ratio of $X:1$.

¹ This test method is under the jurisdiction of ASTM Committee D-2 on Petroleum Products and Lubricants and is the direct responsibility of Subcommittee D02.08 on Volatility.

Current edition approved Jan. 10, 1999. Published April 1999.

² Annual Book of ASTM Standards, Vol 05.01.

³ Annual Book of ASTM Standards, Vol 05.02.

⁴ Annual Book of ASTM Standards, Vol 05.03.

$$VP_X = TP_X - PPA \quad (1)$$

4. Summary of Test Method

4.1 Employing a measuring chamber with a built-in piston, a sample of known volume is drawn into the temperature controlled chamber at 20°C or higher. After sealing the chamber, an expansion is performed in three steps to a final volume of (X+1) times that of the test specimen. After each expansion, the TP_X is determined. The PPA and the solubility of air in the specimen are calculated from the three resulting pressures. The temperature of the chamber is then increased to a specified value, and the TP_X is determined. The (VP_X) is calculated by subtracting the PPA in the liquid, which has been gas corrected for the temperature, from TP_X . For DUPE results, in accordance with Test Method D 5191, the volume of the measuring chamber shall be five times that of the total test specimen and the measuring temperature shall be 37.8°C.

NOTE 2—**Caution:** For liquids containing very low levels of high vapor pressure contaminants, which behave like a gas, this test method of determination of the PPA and gases may lead to wrong results since the partial pressure of the contaminants will be included in the PPA. This effect is shown when the value of the PPA and gases exceeds the average maximum limit of 7 kPa (1 psi).

5. Significance and Use

5.1 Vapor pressure is a very important physical property of volatile liquids for shipping and storage.

5.2 The vapor pressure of gasoline and gasoline-oxygenate blends is regulated by various government agencies.

5.3 Specifications for volatile petroleum products generally include vapor pressure limits to ensure products of suitable volatility performance.

5.4 In this test method, an air saturation procedure prior to the measurement is not required, thus eliminating losses of high volatile compounds during this step. This test method is faster and minimizes potential errors from improper air saturation. This test method permits VP_X determinations in the field.

5.5 This test method can be applied in on-line applications in which an air saturation procedure prior to the measurement cannot be performed.

6. Apparatus

6.1 The apparatus suitable for this test method employs a small volume, cylindrically shaped measuring chamber with associated equipment to control the chamber temperature within the range from 0 to 100 °C. The measuring chamber shall contain a movable piston with a minimum dead volume of less than 1 % of the total volume at the lowest position to allow sample introduction into the measuring chamber and expansion to the desired vapor-liquid ratio. A static pressure transducer shall be incorporated in the piston. The measuring chamber shall contain an inlet/outlet valve combination for sample introduction and expulsion. The piston and the valve combi-

nation shall be at the same temperature as the measuring chamber to avoid any condensation or excessive evaporation.⁵

6.1.1 The measuring chamber shall be designed to contain between 5 and 15 mL of liquid and vapor and be capable of maintaining a vapor-liquid ratio of 4:1 to 1:1. The accuracy of the adjusted vapor-liquid ratio shall be within 0.05.

NOTE 3—The measuring chamber employed by the instruments used in generating the precision and bias statements were constructed of nickel plated aluminum and stainless steel with a total volume of 5 mL. Measuring chambers exceeding a 5-mL capacity can be used, but the precision and bias statements (see Section 15) are not known to apply.

6.1.2 The pressure transducer shall have a minimum operational range from 0 to 200 kPa (0 to 71 psi) with a minimum resolution of 0.1 kPa (0.01 psi) and a minimum accuracy of ± 0.2 kPa (± 0.002 psi). The pressure measurement system shall include associated electronics and readout devices to display the resulting pressure reading.

6.1.3 Electronic temperature control shall be used to maintain the measuring chamber at the prescribed temperature within $\pm 0.1^\circ\text{C}$ for the duration of the test.

6.1.4 A platinum resistance thermometer shall be used for measuring the temperature of the measuring chamber. The minimum temperature range of the measuring device shall be from 20 to 100°C with a resolution of 0.1°C and an accuracy of $\pm 0.1^\circ\text{C}$.

6.1.5 The vapor pressure apparatus shall have provisions from rinsing the measuring chamber with a solvent of low vapor pressure or with the next sample to be tested.

6.2 *Vacuum Pump for Calibration*, capable of reducing the pressure in the measuring chamber to less than 0.01 kPa (0.001 psi) absolute.

6.3 *McLeod Vacuum Gage or Calibrated Electronic Vacuum Measuring Device for Calibration*, to cover at least the range from 0.01 to 0.67 kPa (0 to 5 mm Hg). The calibration of the electronic measuring device shall be regularly verified in accordance with Annex A1 of Test Method D 1160.

6.4 *Pressure Measuring Device for Calibration*, capable of measuring local station pressure with an accuracy and a resolution of 0.1 kPa (1 mm Hg), or better, at the same elevation relative to sea level as the apparatus in the laboratory.

NOTE 4—This test method does not give full details of instruments suitable for carrying out this test. Details on the installation, operation, and maintenance of each instrument may be found in the manufacturer's manual.

7. Reagents and Materials

7.1 *Purity of Reagents*—Use chemicals of at least 99 % purity for quality control checks (see Section 10). Unless otherwise indicated, it is intended that all reagents conform to the specifications of the Committee of Analytical Reagents of

⁵ Vapor pressure apparatus meeting these requirements are available from Grabner Instruments, A-1220 Vienna, Dr. Neurathgasse 1, Austria.