



Designation: D7740 – 11

Standard Practice for Optimization, Calibration, and Validation of Atomic Absorption Spectrometry for Metal Analysis of Petroleum Products and Lubricants¹

This standard is issued under the fixed designation D7740; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ε) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This practice covers information on the calibration and operational guidance for elemental measurements using atomic absorption spectrometry (AAS).

1.1.1 AAS *Related Standards*—Test Methods [D1318](#), [D3237](#), [D3340](#), [D3605](#), [D3831](#), [D4628](#), [D5056](#), [D5184](#), [D5863](#), [D6732](#); Practices [D7260](#) and [D7455](#); and Test Methods [D7622](#) and [D7623](#).

1.2 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.3 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

2. Referenced Documents

2.1 ASTM Standards:²

- [D1318](#) Test Method for Sodium in Residual Fuel Oil (Flame Photometric Method)
- [D3237](#) Test Method for Lead in Gasoline by Atomic Absorption Spectroscopy
- [D3340](#) Test Method for Lithium and Sodium in Lubricating Greases by Flame Photometer (Withdrawn 2013)³
- [D3605](#) Test Method for Trace Metals in Gas Turbine Fuels by Atomic Absorption and Flame Emission Spectroscopy
- [D3831](#) Test Method for Manganese in Gasoline By Atomic Absorption Spectroscopy

¹ This practice is under the jurisdiction of ASTM Committee [D02](#) on Petroleum Products, Liquid Fuels, and Lubricants and is the direct responsibility of Subcommittee [D02.03](#) on Elemental Analysis.

Current edition approved July 1, 2011. Published August 2011. DOI: 10.1520/D7740-11.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

³ The last approved version of this historical standard is referenced on www.astm.org.

- [D4057](#) Practice for Manual Sampling of Petroleum and Petroleum Products
- [D4177](#) Practice for Automatic Sampling of Petroleum and Petroleum Products
- [D4307](#) Practice for Preparation of Liquid Blends for Use as Analytical Standards
- [D4628](#) Test Method for Analysis of Barium, Calcium, Magnesium, and Zinc in Unused Lubricating Oils by Atomic Absorption Spectrometry
- [D5056](#) Test Method for Trace Metals in Petroleum Coke by Atomic Absorption
- [D5184](#) Test Methods for Determination of Aluminum and Silicon in Fuel Oils by Ashing, Fusion, Inductively Coupled Plasma Atomic Emission Spectrometry, and Atomic Absorption Spectrometry
- [D5863](#) Test Methods for Determination of Nickel, Vanadium, Iron, and Sodium in Crude Oils and Residual Fuels by Flame Atomic Absorption Spectrometry
- [D6299](#) Practice for Applying Statistical Quality Assurance and Control Charting Techniques to Evaluate Analytical Measurement System Performance
- [D6732](#) Test Method for Determination of Copper in Jet Fuels by Graphite Furnace Atomic Absorption Spectrometry
- [D6792](#) Practice for Quality System in Petroleum Products and Lubricants Testing Laboratories
- [D7260](#) Practice for Optimization, Calibration, and Validation of Inductively Coupled Plasma-Atomic Emission Spectrometry (ICP-AES) for Elemental Analysis of Petroleum Products and Lubricants
- [D7455](#) Practice for Sample Preparation of Petroleum and Lubricant Products for Elemental Analysis
- [D7622](#) Test Method for Total Mercury in Crude Oil Using Combustion and Direct Cold Vapor Atomic Absorption Method with Zeeman Background Correction
- [D7623](#) Test Method for Total Mercury in Crude Oil Using Combustion-Gold Amalgamation and Cold Vapor Atomic Absorption Method

3. Terminology

3.1 Definitions:

3.1.1 *absorbance*, *n*—logarithm to the base 10 of the ratio of the reciprocal of the transmittance.

3.1.2 *atomic absorption spectrometry*, *n*—analytical technique for measuring metal content of solutions, based on a combination of flame source, hollow cathode lamp, photomultiplier, and a readout device.

3.1.3 *atomizer*, *n*—usually a flame source used to decompose the chemical constituents in a solution to its elemental components.

3.1.4 *blank*, *n*—solution which is similar in composition and contents to the sample solution but does not contain the analyte being measured.

3.1.5 *burner*, *n*—flame device used to atomize the analyte by burning in a high temperature flame mixed of a fuel and an oxidant.

3.1.6 *calibration*, *n*—process by which the relationship between signal intensity and elemental concentration is determined for a specific element analysis.

3.1.7 *calibration curve*, *n*—plot of signal intensity versus elemental concentration using data obtained by making measurements with standards.

3.1.8 *calibration standard*, *n*—material with a certified value for a relevant property, issued by or traceable to a national organization such as NIST, and whose properties are known with sufficient accuracy to permit its use to evaluate the same property of another sample.

3.1.9 *certified reference material*, *n*—reference material one or more of whose property values are certified by a technically valid procedure, accompanied by a traceable certificate or other documentation which is issued by a certifying body.

3.1.10 *check standard*, *n*—material having an assigned (known) value (reference value) used to determine the accuracy of the measurement system or instrument.

3.1.10.1 *Discussion*—This practice is not used to calibrate the measurement instrument or system.

3.1.11 *detection limit*, *n*—concentration of an analyte that results in a signal intensity that is some multiple (typically two) times the standard deviation of the background intensity at the measurement wavelength.

3.1.12 *dilution factor*, *n*—ratio of sample weight of the aliquot taken to the final diluted volume of its solution.

3.1.12.1 *Discussion*—The dilution factor is used to multiply the observed reading and obtain the actual concentration of the analyte in the original sample.

3.1.13 *graphite furnace*, *n*—electrothermal device for atomizing the metal constituents.

3.1.14 *hollow cathode lamp*, *n*—device consisting of a quartz envelope containing a cathode of the metal to be determined and a suitable anode.

3.1.15 *hydride generation*, *n*—device to atomize some metals which form gaseous hydrides.

3.1.16 *monochromator*, *n*—device that isolates a single atomic resonance line from the line spectrum emitted by the hollow cathode lamp, excluding all other wavelengths.

3.1.17 *nebulizer*, *n*—device that generates an aerosol by flowing a liquid over a surface that contains an orifice from which gas flows at a high velocity.

3.1.18 *NIST*, *n*—National Institute of Standards and Technology, Gaithersburg, MD. Formerly known as National Bureau of Standards.

3.1.19 *precision*, *n*—closeness of agreement between test results obtained under prescribed conditions.

3.1.20 *quality assurance*, *n*—system of activities, the purpose of which is to provide to the producer and user of a product, measurement, or service the assurance that it meets the defined standards of quality with a stated level of confidence.

3.1.21 *quality control*, *n*—planned system of activities whose purpose is to provide a level of quality that meets the needs of users; also the uses of such a system.

3.1.22 *quality control sample*, *n*—for use in quality assurance program to determine and monitor the precision and stability of a measurement system; a stable and homogenous material having physical or chemical properties, or both, similar to those of typical samples tested by the analytical measurement system.

3.1.22.1 *Discussion*—This material should be properly stored to ensure sample integrity, and is available in sufficient quantity for repeated long term testing.

3.1.23 *reference material*, *n*—material with accepted reference value(s), accompanied by an uncertainty at a stated level of confidence for desired properties, which may be used for calibration or quality control purposes in the laboratory.

3.1.24 *refractory elements*, *n*—elements forming difficult-to-dissociate oxides during combustion.

3.1.25 *repeatability*, *n*—difference between two test results, obtained by the same operator with the same apparatus under constant operating conditions on identical test material would, in the long term and correct operation of the test method, exceed the values given only in one case in twenty.

3.1.26 *reproducibility*, *n*—difference between two single and independent results, obtained by different operators working in different laboratories on identical test materials, would in the long run, in the normal and correct operation of the test method, exceed the values given only one case in twenty.

3.1.27 *spectrometer*, *n*—instrument used to measure the emission or absorption spectrum emitted by a species in the vaporized sample.

3.1.28 *spectrum*, *n*—array of the components of an emission or absorption arranged in the order of some varying characteristics such as wavelength, mass, or energy.

3.1.29 *standard reference material*, *n*—trademark for reference materials certified by NIST.

4. Summary of Practice

4.1 An Atomic Absorption Spectrometer (AAS) is used to determine the metal composition of various materials.

Although usually AAS is done using a flame to atomize the metals, graphite furnace (GF-AAS) or cold vapor (CV-AAS) may also be used for metals at very low levels of concentration or some elements not amenable to flame atomization. This practice summarizes the protocols to be followed during calibration and verification of the instrument performance.

5. Significance and Use

5.1 Accurate elemental analysis of petroleum products and lubricants is necessary for the determination of chemical properties, which are used to establish compliance with commercial and regulatory specifications.

5.2 Atomic Absorption Spectrometry (AAS) is one of the most widely used analytical techniques in the oil industry for elemental analysis. There are at least twelve Standard Test Methods published by ASTM D02 Committee on Petroleum Products and Lubricants for such analysis. See [Table 1](#).

5.3 The advantage of using an AAS analysis include good sensitivity for most metals, relative freedom from interferences, and ability to calibrate the instrument based on elemental standards irrespective of their elemental chemical forms. Thus, the technique has been a method of choice in most of the oil industry laboratories. In many laboratories, AAS has been superseded by a superior ICP-AES technique (see [Practice D7260](#)).

5.4 Some of the ASTM AAS Standard Test Methods have also been issued by other standard writing bodies as technically equivalent standards. See [Table 2](#).

6. Interferences

6.1 Although over 70 elements can be determined by AAS usually with a precision of 1-3 % and with detection limits of the order of sub-mg/kg levels, and with little or no atomic spectral interference. However, there are several types of interferences possible: chemical, ionization, matrix, emission, spectral, and background absorption interferences. Since these interferences are well-defined, it is easy to eliminate or compensate for them. See [Table 3](#).

6.1.1 *Chemical Interferences*—If the sample for analysis contains a thermally stable compound with the analyte that is not totally decomposed by the energy of the flame, a chemical interference exists. They can normally be overcome or controlled by using a higher temperature flame or addition of a releasing agent to the sample and standard solutions.

6.1.2 *Ionization Interferences*—When the flame has enough energy to cause the removal of an electron from the atom, creating an ion, ionization interference can occur. They can be controlled by addition of an excess of an easily ionized element to both samples and standards. Normally alkali metals which have very low ionization potentials are used.

6.1.3 *Matrix Interferences*—These can cause either a suppression or enhancement of the analyte signal. Matrix interferences occur when the physical characteristics – viscosity, burning characteristics, surface tension – of the sample and standard differ considerably. To compensate for the matrix interferences, the matrix components in the sample and standard should be matched as closely as possible. Matrix interferences can also be controlled by diluting the sample solution until the effect of dissolved salts or acids is negligible. Sometimes, the method of standard addition is used to overcome this interference. See [6.2](#).

6.1.4 *Emission Interferences*—At high analyte concentrations, the atomic absorption analysis for highly emissive elements sometimes exhibits poor analytical precision, if the emission signal falls within the spectral bandpass being used. This interference can be compensated for by decreasing the slit width, increasing the lamp current, diluting the sample, and / or using a cooler flame.

6.1.5 *Spectral Interferences*—When an absorbing wavelength of an element present in the sample but not being determined falls within the bandwidth of the absorption line of the element of interest a spectral interference can occur. An interference by other atoms can occur when there is a sufficient overlapping between radiation and emitted by the excited atoms and other absorbing atoms. Usually the bandwidth is much wider than the width of the emission and absorption lines. Thus, interferences by other atoms are fortunately quite limited in AAS. The interference can result in erroneously high results. This can be overcome by using a smaller slit or selecting an alternate wavelength.

6.1.6 *Background Absorption Interferences*—There are two causes of background absorption: light scattering by particles in the flame and molecular absorption of light from the lamp by molecules in the flame. This interference cannot be corrected with standard addition method. The most common way to compensate for background absorption is to use a background corrector which utilizes a continuum source.

6.2 *Standard Addition Method*—One way of dealing with some of the interferences in the AAS methods is to use a

TABLE 1 Applications of AAS for Metal Analysis of Petroleum Products and Lubricants

ASTM Test Method	Matrix	Elements Determined
D1318	Residual Fuel Oil	Sodium
D3237	Gasoline	Lead
D3340	Greases	Lithium and Sodium
D3605	Gas Turbine Fuels	Calcium, Lead, Sodium, and Vanadium
D3831	Gasoline	Manganese
D4628	Automotive Lubricants	Barium, Calcium, Magnesium, and Zinc
D5056	Petroleum Coke	Aluminum, Calcium, Iron, Nickel, Silicon, Sodium, and Vanadium
D5184	Fuel Oils	Aluminum and Silicon
D5863	Crude and Fuel Oils	Iron, Nickel, Sodium, and Vanadium
D6732	Jet Fuels	Copper
D7622	Crude Oils	Mercury
D7623	Crude Oils	Mercury