



Designation: E502 – 21a

Standard Test Method for Selection and Use of ASTM Standards for the Determination of Flash Point of Chemicals by Closed Cup Methods¹

This standard is issued under the fixed designation E502; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This test method covers the determination of the flash point of liquid and solid chemical compounds flashing from below –10 to 370°C (16 to 700°F). The procedures and apparatus in Test Methods **D56**, **D93**, **D3278**, **D3828**, and **D3941** are to be used. Modification to these procedures are specified for tests on solids and viscous liquids. The significance of the results obtained is discussed along with possible sources of error and factors that might cause interference.

1.2 Suggestions for adapting this procedure to mixtures of chemicals are included (see **Appendix X2**).

1.3 This test method should be used to measure and describe the properties of materials, products, or assemblies in response to heat and flame under controlled laboratory conditions and should not be used to describe or appraise the fire hazard or fire risk of materials or assemblies under actual fire conditions. However, results of this test method may be used as elements of a fire risk assessment that take into account all of the factors that are pertinent to an assessment of the fire hazard of a particular end use.

1.4 The values stated in SI units are to be regarded as standard. The values given in parentheses after SI units are provided for information only and are not considered standard.

1.5 **Warning**—Mercury has been designated by the United States Environmental Protection Agency (EPA) and many state agencies as a hazardous material that can cause central nervous system, kidney, and liver damage. Mercury, or its vapor, may be hazardous to health and corrosive to materials. Caution should be taken when handling mercury and mercury containing products. See the applicable product Material Safety Data Sheet (MSDS) for details and EPA's website — <http://www.epa.gov/mercury/faq.htm> — for additional information. Users should be aware that selling mercury or mercury-containing products, or both, into your state may be prohibited by state law.

¹ This test method is under jurisdiction of ASTM Committee **E27** on Hazard Potential of Chemicals and is the direct responsibility of Subcommittee **E27.04** on Flammability and Ignitability of Chemicals.

Current edition approved June 1, 2021. Published July 2021. Originally approved in 1984. Last previous edition approved in 2021 as E502 – 21. DOI: 10.1520/E0502-21A.

1.6 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety, health, and environmental practices and determine the applicability of regulatory limitations prior to use.* See also Section 8.

1.7 *This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.*

2. Referenced Documents

2.1 ASTM Standards:²

- D56** Test Method for Flash Point by Tag Closed Cup Tester
- D92** Test Method for Flash and Fire Points by Cleveland Open Cup Tester
- D93** Test Methods for Flash Point by Pensky-Martens Closed Cup Tester
- D270** Methods of Sampling Petroleum and Petroleum Products (Withdrawn 1984)³
- D1310** Test Method for Flash Point and Fire Point of Liquids by Tag Open-Cup Apparatus
- D3278** Test Methods for Flash Point of Liquids by Small Scale Closed-Cup Apparatus
- D3828** Test Methods for Flash Point by Small Scale Closed Cup Tester
- D3934** Test Method for Flash/No Flash Test—Equilibrium Method by a Closed-Cup Apparatus
- D3941** Test Method for Flash Point by the Equilibrium Method With a Closed-Cup Apparatus
- E681** Test Method for Concentration Limits of Flammability of Chemicals (Vapors and Gases)
- E1232** Test Method for Temperature Limit of Flammability of Chemicals

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

³ The last approved version of this historical standard is referenced on www.astm.org.

3. Terminology

3.1 Definitions:

3.1.1 *flash point, n—in flash point test methods*, the lowest temperature of the test specimen, adjusted to account for variations in atmospheric pressure from 101.3 kPa, at which application of an ignition source causes the vapors of the test sample to ignite under specified conditions of test.

4. Summary of Test Method

4.1 The specimen is placed in a closed cup and, in the small scale method, equilibrated at a test temperature; in the Pensky-Martens method, heated at a controlled rate with stirring; and in the tag method, also heated at a controlled rate but without stirring. A small flame is directed into the vapor space of each cup at specified intervals, with simultaneous interruption of stirring in the Pensky-Martens method, to determine whether a flash occurs or not. In Test Method [D3941](#), the specimen is heated at a slower rate than in the other controlled heating methods, maintaining a small temperature differential between bath and specimen.

5. Significance and Use

5.1 The flash point measures the response of the sample to heat and flame under controlled laboratory conditions. It is only one of a number of properties that must be considered in assessing the overall flammability hazard of a material.

5.2 As a result of physical factors inherent in the apparatus and procedure, the closed cup flash point does not necessarily represent the minimum temperature at which a material can evolve flammable vapors, and the absence of a flash point does not guarantee nonflammability (see [Appendix X1](#) and [Appendix X2](#)).

5.3 Flash point is used in shipping and safety regulations to define flammable and combustible materials. Test Methods [D56](#), [D93](#), and [D3278](#) are specified as test methods for determining the flash point of these materials.

5.4 If the process or handling conditions dictate the usage of a flammable material at temperatures ranging upward from 5 to 10°C below the closed-cup flash point, then a flammable vapor might be present above the liquid. In such cases, it may be more appropriate to use the temperature limit of flammability (as determined by Test Method [E1232](#)) instead of flash point.

5.5 For single component samples, small-scale methods involving equilibrium procedures and only one flame pass per specimen are preferred.

5.6 For mixtures containing small concentrations of volatile components, special procedures are needed to minimize the loss of volatiles, with consequent elevation of the flash point, while the sample is being heated. (See [X2.5](#).)

5.7 In cases where errors caused by loss of volatiles, downwards flame direction and quenching are unacceptable, the “lower temperature limit of flammability” can be determined instead using Test Method [E1232](#). The temperature limit of flammability test chamber is sufficiently large to overcome flame quenching effects in most cases of practical importance, thus, usually indicating the presence of vapor-phase flammability if it does exist.

6. Interferences

6.1 Incorrect flash points can be obtained when testing chemicals corrosive to the materials of construction of the cup. (For example, certain amines and acid chlorides react with the standard aluminum small scale cup causing erroneously low flash points, perhaps due to hydrogen formation.) Cups employing alternative materials of construction, electroplating or plastic coating can provide corrosion resistance. Results in non-standard cups, particularly in non-equilibrium tests, may differ slightly from those obtained in this test method.

7. Apparatus

7.1 *Tag Closed-Cup Tester*, including thermometers, shall be as shown in Test Methods [D56](#) and [D3941](#).

7.2 *Pensky-Martens Closed-Cup Tester*, including thermometers, shall be as shown in Test Methods [D93](#).

7.3 *Small Scale Closed Tester*, including thermometers, shall be as shown in Test Methods [D3278](#) or [D3828](#).

NOTE 1—Some automatic flash point testers may save testing time and permit the use of small samples. If automatic testers are used, the user must be certain that all instructions for calibration and operation are followed to ensure that the results are equivalent to those obtained on the ASTM standard equipment. For regulation purposes or in cases of dispute, the flash point as determined on the manual tester shall be the accepted value.

NOTE 2—ASTM thermometers 33C or 33F may be used in the tag tester instead of those specified in Test Method [D56](#) when conducting tests at temperatures below −10°C (14°F). Slight stem corrections may be necessary and care should be taken to avoid freezing the mercury in the thermometer by cooling below −40°C (−40°F).

7.4 *Shield*, as described in Test Method [D3941](#) or [D1310](#).

8. Hazards

8.1 Toxicity of Chemical and Combustion Products:

8.1.1 Isolate or control operations on toxic or corrosive materials to prevent exposure to any personnel.

8.1.2 Since flash point tests are conducted in still air, the use of forced circulation for removal of toxic or nuisance fumes or combustion products is restricted. However, a laboratory fume hood equipped with an exhaust damper that can be completely closed provides an ideal location for maintaining draft-free conditions and provides the ability to readily exhaust dangerous vapors and combustion products when necessary.

8.1.3 Use respiratory and splash protective devices as appropriate with toxic or corrosive materials. In most cases, approved cartridge respirators are adequate respiratory protection for the concentrations normally encountered in flash-point testing. Certain toxic or unusual materials, however, may require an air-supplied respirator and extreme cases may require complete protective coverage such as an air-supplied plastic suit. (Two examples of the latter type of material are dimethyl sulfate and pure mercaptans.) Tests on these highly toxic or obnoxious materials may also be conducted in completely isolated, closed systems, such as glove boxes. In this case, procedures should ensure an uncontaminated air system in the box, and should prevent a buildup of vapors from the material under test.