

Standard Terminology Relating to Anesthesia and Respiratory Equipment¹

This standard is issued under the fixed designation F 2002; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

INTRODUCTION

Consistent use of terminology contributes to improved safety and performance of anesthetic and respiratory care equipment by helping manufacturers, operators, and users communicate more effectively. It is expected that this terminology will be used on labels and markings on equipment and documents that accompany equipment. If other terminology is used on labels or markings on equipment or documents that accompany equipment, or both, that terminology should be defined in the accompanying documents.

1. Scope

1.1 This terminology covers terms, symbols, and abbreviations associated with anesthesia and respiratory care equipment.

2. Referenced Documents

2.1 ASTM Standards:

- F 896 Specification for Flexible Fiberoptic Bronchoscopes²
- F 920 Specification for Minimum Performance and Safety Requirements for Resuscitators Intended for Use with Humans²
- F 927 Specification for Pediatric Tracheostomy Tubes²
- F 960 Specification for Medical and Surgical Suction and Drainage Systems²
- F 965 Specification for Rigid Laryngoscopes for Tracheal Intubation—Hook-on Fittings for Laryngoscope Handles and Blades with Lamps²
- F 984 Specification for Cutaneous Gas Monitoring Devices for Oxygen and Carbon Dioxide²
- F 1054 Specification for Conical Fittings of 15-mm and 22-mm Sizes²
- F 1100 Specification for Ventilators Intended for Use in Critical Care²

- F 1101 Specification for Ventilators Intended for Use During Anesthesia²
- F 1161 Specification for Minimum Performance and Safety Requirements for Components and Systems of Anesthesia Gas Machines²
- F 1195 Specification for Rigid Laryngoscopes for Tracheal Intubation—Hook-on Fittings for Fiberilluminated Blades and Handles²
- F 1204 Specification for Anesthesia Reservoir Bags²
- F 1205 Specification for Anesthesia Breathing Tubes²
- F 1208 Specification for Minimum Performance and Safety Requirements for Anesthesia Breathing Systems²
- F 1218 Specification for Bronchoscopes (Rigid)²
- F 1242 Specification for Cuffed and Uncuffed Tracheal Tubes²
- F 1243 Specification for Tracheal Tube Connectors²
- F 1343 Specification for Anesthetic Equipment—Scavenging Systems for Anesthetic Gases²
- F 1415 Specification for Pulse Oximeters²
- F 1452 Specification for Minimum Performance and Safety Requirements for Components and Systems of Anesthetic Gas Monitors²
- F 1456 Specification for Capnometers²
- F 1462 Specification for Oxygen Analyzers²
- F 1463 Specification for Alarm Signals in Medical Equipment Used in Anesthesia and Respiratory Care²
- F 1464 Specification for Oxygen Concentrators for Domiciliary Use²
- F 1573 Specification for Anesthetic Equipment—Oropharyngeal and Nasopharyngeal Airways²

¹ This terminology is under the jurisdiction of ASTM Committee F29 on Anesthetic and Respiratory Equipment and is the direct responsibility of Subcommittee F29.16 on Terminology.

Current edition approved Nov. 10, 2001. Published February 2002.

Originally published as F2002–00. Last previous edition F2002–01b.

² *Annual Book of ASTM Standards*, Vol 13.01.



F 1590 Specification for Tracheostomy Tube Connectors²
 F 1627 Specification for Tracheostomy Tubes—Pediatric Tracheostomy Tubes²

F 1628 Specification for Labeling and Marking of Cuffed and Uncuffed Tracheal Tubes and Related Treatments Intended for Use During Laser Surgery²

F 1666 Specification for Adult Tracheostomy Tubes²

2.2 Other Documents:

ISO 8835-2:1993 Inhalational anaesthesia systems—Part 2: Anaesthetic circle breathing systems³

ISO 9703-1:1992 Anaesthetic and respiratory alarm signals—Part 1: Visual alarm signals³

ISO 9360:1992 Anaesthetic and respiratory equipment—Heat and moisture exchangers for use in humidifying respired gases in humans³

ISO/FDIS 9170-2 Terminal units for medical gas pipeline systems—Part 2: Terminal units for anaesthetic gas scavenging systems³

ISO Standards from TC 210 Quality Management and Corresponding General Aspects for Medical Devices³

ISO 4135:1995 Anaesthesiology—Vocabulary³

3. Terminology

3.1 Definitions:

adjustable pressure limiting valve (APL), *n*—operator-adjustable device intended to limit the maximum pressure in an anesthesia breathing system when it is part of the flow pathway of the breathing system, by controlling the rate of escape of gases from the system.

anesthesia breathing system, *n*—those inspiratory and expiratory pathways through which gas flows at respiratory pressures between the fresh gas inlet, the patient connection port, and the exhaust valve or port.

anesthesia gas supply device, *n*—assembly of components that controls and delivers all gas flows and agent concentrations in the fresh gas into the anesthesia breathing system. Note: it may include a flow adjustment control(s), a gas mixing system, and an anesthetic vapor delivery device(s).

anesthesia gas supply piping, *n*—all pipe work, including unions, from the unidirectional valve(s) in the pipeline inlet(s), the piping connecting the high pressure supply to the gas pressure-reducing device/system, and from the gas pressure-reducing device/system outlet to the flow adjustment control(s) and auxiliary gas outlet. It includes piping leading to and from pneumatic alarm device(s), pressure indicator(s), oxygen flush, anesthetic vapor delivery device(s), and piping leading to the common gas outlet.

anesthesia ventilator, *n*—ventilator designed to be used with or integral to an anesthesia breathing system. **F 1101**

anesthesia workstation, *n*—system for administration of anesthesia to patients. It consists of the anesthesia gas supply device, anesthesia ventilator, and associated monitoring, alarm, and protection device(s).

anesthetic-gas scavenging systems, *n*—complete systems that collect and remove the excess gases released from equip-

ment used in administering anesthesia or exhaled by the patient during anesthesia for the purpose of conveying these gases to an appropriate place of discharge.

anesthetic vapor delivery device, *n*—device in which anesthetic agent is transformed from the liquid to the gaseous phase of controllable concentration.

applied part, *n*—entirety of all parts of equipment including the patient leads that come intentionally into contact with the patient to be examined or treated.

auxiliary oxygen flowmeter, *n*—self-contained oxygen flowmeter with its own flow adjustment control, flow indicator, and outlet that is not an integral part of the anesthesia gas supply device.

DISCUSSION—The auxiliary oxygen flowmeter is sometimes called the courtesy oxygen flowmeter.

backup, *n*—retrievable copy of information.

circle breathing system, *n*—breathing system in which the direction of gas flow is through separate inspiratory and expiratory pathways that form a circle.

common gas outlet (fresh gas outlet), *n*—that port through which the dispensed mixture from the anesthesia gas supply device is delivered to the anesthesia breathing system.

control center, *n*—information system component of an anesthesia workstation mediating the centralized alarm prioritization and management strategy(ies).

decision support system, *n*—information system that provides output information intended to facilitate decision making by the operator.

editing, *v*—changing a recorded value, comment, annotation, event, or other information at any time after it has been made part of the record.

flow adjustment control, *n*—device or assembly that controls the flow of gas(es) or gas mixtures. Note: the flow adjustment control is often called the flow control valve.

flow indicator, *n*—device that indicates the volume of a specific gas or gas mixture passing through it in a unit of time. Note: the flow indicator is sometimes called a flowmeter.

gas mixing system, *n*—device or assembly that receives oxygen and other medical gas(es) and delivers a mixture of these gases in concentrations controllable by the operator.

gas pressure reducing device/system, *n*—single device or group of devices designed to provide a range of controlled delivered pressures over a specific range of inlet pressures.

information system, *n*—digital system collecting, displaying, or processing information from primary or secondary data sources or both. Note: examples includes AARKs and decision support systems.

infusion device, *n*—mechanical or electromechanical device regulating the flow of liquids to the patient under positive pressure. Note: the positive pressure may be generated either by a pump or by gravitational force.

input device, *n*—device used to enter data or control the anesthesia information system manually.

DISCUSSION—Examples include keywords, mice, trackballs, light pens, barcode scanners, and microphones.

legible, *adj*—discernible or identifiable to the operator with 6 ft

³ Available from American National Standards Institute, 25 W. 43rd St., 4th Floor, New York, NY 10036.