

Microbiological safety cabinets —

Part 3: Specification for performance after installation

Committees responsible for this British Standard

The preparation of this British Standard was entrusted by the Health Care Standards Policy Committee (HCC/-) to Technical Committee HCC/28, upon which the following bodies were represented:

- Association of the British Pharmaceutical Industry
- Association of Clinical Pathologists
- Association of Medical Microbiologists
- British Health and Safety Society
- British Laboratory Ware Association
- British Occupational Hygiene Society
- Chartered Institution of Building Services Engineers
- Department of Health
- Guild of Hospital Pharmacists
- Health and Safety Executive
- Institute of Animal Health
- Institute of Biology
- Institute of Hospital Engineering
- Institute of Medical Laboratory Sciences
- Institution of Mechanical Engineers
- Manufacturing Science Finance
- Medical Research Council
- Ministry of Agriculture, Fisheries and Food
- Public Health Laboratory Service
- Royal College of Pathologists
- Coopted members

This British Standard, having been prepared under the direction of the Health Care Standards Policy Committee, was published under the authority of the Standards Board and comes into effect on 15 June 1992

© BSI 12-1998

The following BSI references relate to the work on this standard:
Committee reference HCC/28
Draft for comment 91/52391 DC

ISBN 0 580 20750 1

Amendments issued since publication

Amd. No.	Date	Comments

Contents

	Page
Committees responsible	Inside front cover
Foreword	ii
1 Scope	1
2 Definitions	1
3 Pressure integrity	1
4 Airflow	1
5 Operator protection factor for class I and class II cabinets	1
Appendix A Methods of air velocity measurement	2
Appendix B Recommendations for external contamination and cross contamination tests for class II cabinets	2
Publication(s) referred to	Inside back cover