

Agent specific filling systems for anaesthetic vaporizers —

Part 1: Rectangular keyed filling systems

The European Standard EN 1280-1:1997, with the incorporation of amendment A1:2000 has the status of a British Standard

ICS 11.040.10

Committees responsible for this British Standard

The preparation of this British Standard was entrusted to Technical Committee CH/44, Anaesthetic machines, breathing attachments, medical gas pipeline systems and hose assemblies, upon which the following bodies were represented:

- Association of Anaesthetists of Great Britain and Ireland
- Association of British Health-care Industries
- British Anaesthetic and Respiratory Equipment Manufacturers' Association
- Department of Health (NHS Estates)
- Department of Health (Medical Devices Agency)

The following body was also represented in the drafting of the standard, through a subcommittee/panel:

- Guild of Hospital Pharmacists

This British Standard, having been prepared under the direction of the Health and Environment Sector Board, was published under the authority of the Standards Board and comes into effect on 15 November 1997

© BSI 01-2001

The following BSI references relate to the work on this standard:
Committee reference CH/44
Draft for comment 93/509496 DC

ISBN 0 580 28641 X

Amendments issued since publication

Amd. No.	Date	Comments
11039	January 2001	Indicated by a sideline

Contents

	Page
Committees responsible	Inside front cover
National foreword	ii
Foreword	2
Text of EN 1280-1	5

National foreword

This British Standard has been prepared by Technical Committee CH/44 and is the English language version of EN 1280-1:1997 *Agent specific filling systems for anaesthetic vaporizers — Part 1: Rectangular keyed filling systems*, including amendment A1:2000 published by the European Committee for Standardization (CEN).

Cross-references

The British Standards which implement international or European publications referred to in this document may be found in the BSI Standards Catalogue under the section entitled “International Standards Correspondence Index”, or by using the “Find” facility of the BSI Standards Electronic Catalogue.

A British Standard does not purport to include all the necessary provisions of a contract. Users of British Standards are responsible for their correct application.

Compliance with a British Standard does not of itself confer immunity from legal obligations.

Summary of pages

This document comprises a front cover, an inside front cover, pages i and ii, the EN title page, pages 2 to 20, an inside back cover and a back cover.

The BSI copyright notice displayed in this document indicates when the document was last issued.

Sidelining in this document indicates the most recent changes by amendment.

ICS 11.040.10

Descriptors: Medical equipment, anaesthetic equipment, evaporators, filling devices, specifications

English version

Agent specific filling systems for anaesthetic vaporizers — Part 1: Rectangular keyed filling systems

(includes amendment A1:2000)

Systèmes de remplissage spécifiques à l'agent
pour évaporateurs d'anesthésie —
Partie 1: Systèmes de remplissage à clavettes
rectangulaires
(inclut l'amendement A1:2000)

Substanzspezifische Füllsysteme für
Anästhesiemittelverdampfer —
Teil 1: Rechteckige Kodierte Füllsysteme
(enthält Änderung A1:2000)

This European Standard was approved by CEN on 1997-02-15. Amendment A1 was approved by CEN on 2000-04-10. CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the Central Secretariat has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and United Kingdom.

CEN

European Committee for Standardization
Comité Européen de Normalisation
Europäisches Komitee für Normung

Central Secretariat: rue de Stassart 36, B-1050 Brussels

© 1997 All rights of exploitation in any form and by any means reserved worldwide for CEN national Members.

Ref. No. EN 1280-1:1997 + A1:2000 E
 Documents
Document Sharing Hub

Foreword

This European Standard has been prepared by Technical Committee CEN/TC 215, Respiratory and anaesthetic equipment, the Secretariat of which is held by BSI.

This European Standard is based on ISO 5360, prepared by Technical Committee ISO/TC 121 of the International Organization for Standardization (ISO). ISO/TC 121 participated in the preparation of this European Standard.

Annex B is normative and forms part of this European Standard. Annex A, Annex C and Annex ZA are for information only.

This European Standard has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

For relationship with EU Directives, see informative Annex ZA, which is an integral part of this standard.

Attention is drawn to the cautions given in clauses 13.2.2 and 13.3 b). See also informative references 11 and 12 in Annex C.

EN 1280 consists of the following Parts, under the general title *Agent specific filling systems for anaesthetic vaporizers*:

- *Part 1: Rectangular keyed filling systems;*
- *Part 2: Cylindrical keyed filling systems.*

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by September 1997, and conflicting national standards shall be withdrawn at the latest by the 13th of June 1998.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and the United Kingdom.

Foreword to amendment A1

This amendment EN 1280-1:1997/A1:2000 to the EN 1280-1:1997 has been prepared by Technical Committee CEN/TC 215, Respiratory and anaesthetic equipment, the Secretariat of which is held by BSI.

This amendment to the European Standard EN 1280-1:1997 shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by November 2000, and conflicting national standards shall be withdrawn at the latest by November 2000.

This amendment to the European Standard EN 1280-1:1997 has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Czech Republic, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and the United Kingdom

Contents

		Page
Foreword	2	
1 Scope	5	
2 Normative references	5	
3 Definitions	5	
4 Bottle	5	
5 Bottle collar	5	
6 Bottle adaptor	6	
7 Filler receptacle	6	
8 Filling rate	6	
9 Leakage	6	
10 Overfilling protection	6	
11 Colour coding	7	
12 Constructional requirements	7	
13 Information provided by the manufacturer	7	
Annex A (informative) Configurations of agent-specific filling systems	18	
Annex B (normative) Determination of the total leakage into the atmosphere of anaesthetic agent during the filling procedure	18	
Annex C (informative) Bibliography	19	
Annex ZA (informative) Clauses of this European Standard addressing essential requirements or other provisions of EU Directives	20	
Figure 1 — Two examples of threaded bottle necks for anaesthetic agent bottles	9	
Figure 2 — Configuration of agent-specific bottle collars	10	
Figure 3 — Alternative positions of agent-specific bottle collar	11	
Figure 4 — Configuration and dimensions of agent-specific male adaptors for use with enflurane, methoxyflurane and halothane	12	
Figure 5 — Configuration and dimensions of agent-specific male adaptors for use with isoflurane and sevoflurane	13	
Figure 6 — Configuration and dimensions of agent-specific bottle connectors	14	
Figure 7 — Configuration and dimensions of agent-specific filler receptacles for use with enflurane, methoxyflurane and halothane	15	
Figure 8 — Configuration and dimensions of agent-specific filler receptacles for use with isoflurane and sevoflurane	17	
		Figure A.1 — Different configurations of agent-specific filling system 18
		Table 1 — Dimensions of threaded bottle necks of anaesthetic agent bottles 8
		Table 2 — Dimensions and colour of agent-specific bottle collars and connectors 8
		Table 3 — Details of male adaptors for use with enflurane, methoxyflurane and halothane 12
		Table 4 — Details of male adaptors for use with isoflurane and sevoflurane 13
		Table 5 — Details of filler receptacles for use with enflurane, methoxyflurane and halothane 16
		Table 6 — Details of filler receptacles for use with isoflurane and sevoflurane 17
		Table ZA.1 — Correspondence between this European Standard and EU Directives 20

