



BSI Standards Publication

Medical electrical equipment

Part 1-12: General requirements for basic safety and essential performance — Collateral Standard: Requirements for medical electrical equipment and medical electrical systems intended for use in the emergency medical services environment

National foreword

This British Standard is the UK implementation of EN 60601-1-12:2015+A1:2020. It is identical to IEC 60601-1-12:2014, incorporating amendment 1:2020. It supersedes [BS EN 60601-1-12:2015](#), which is withdrawn.

The start and finish of text introduced or altered by amendment is indicated in the text by tags. Tags indicating changes to IEC text carry the number of the IEC amendment. For example, text altered by IEC amendment 1 is indicated by **A1** **A1**.

The UK participation in its preparation was entrusted to Technical Committee CH/62/1, Common aspects of Electrical Equipment used in Medical Practice.

A list of organizations represented on this committee can be obtained on request to its committee manager.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2020
Published by BSI Standards Limited 2020

ISBN 978 0 539 04587 1

ICS 11.040.50; 11.040.01

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 30 June 2015.

Amendments/corrigenda issued since publication

Date	Text affected
30 September 2020	Implementation of IEC amendment 1:2020 with CENELEC endorsement A1:2020

English Version

**Medical electrical equipment - Part 1-12: General requirements
for basic safety and essential performance - Collateral Standard:
Requirements for medical electrical equipment and medical
electrical systems intended for use in the emergency medical
services environment
(IEC 60601-1-12:2014)**

Appareils électromédicaux - Partie 1-12: Exigences
générales pour la sécurité de base et les performances
essentielle - Norme collatérale: Exigences pour les
appareils électromédicaux et les systèmes électromédicaux
destinés à être utilisés dans l'environnement des services
médicaux d'urgence
(IEC 60601-1-12:2014)

Medizinische elektrische Geräte - Teil 1-12: Allgemeine
Festlegungen für die Sicherheit einschließlich der
wesentlichen Leistungsmerkmale - Ergänzungsnorm:
Anforderungen an medizinische elektrische Geräte und
medizinische elektrische Systeme in der Umgebung für den
Notfalleinsatz
(IEC 60601-1-12:2014)

This European Standard was approved by CENELEC on 2014-07-24. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.



European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

European foreword

The text of document 62A/932/FDIS, future edition 1 of IEC 60601-1-12, prepared by SC 62A "Common aspects of electrical equipment used in medical practice", of IEC/TC 62 "Electrical equipment in medical practice", was submitted to the IEC-CENELEC parallel vote and approved by CENELEC as EN 60601-1-12:2015.

The following dates are fixed:

- latest date by which the document has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2015-11-22
- latest date by which the national standards conflicting with the document have to be withdrawn (dow) 2018-12-31

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC [and/or CEN] shall not be held responsible for identifying any or all such patent rights.

This document has been prepared under a mandate given to CENELEC by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

For the relationship with EU Directives 93/42/EEC and 90/385/EEC, see informative Annexes ZZA and ZZB, which are integral parts of this document.

Endorsement notice

The text of the International Standard IEC 60601-1-12:2014 was approved by CENELEC as a European Standard without any modification.

In the official version, for Bibliography, the following notes have to be added for the standards indicated:

IEC 60038:2009	NOTE	Harmonized as EN 60038:2011 (modified).
IEC 60065	NOTE	Harmonized as EN 60065.
IEC 60335-1:2010	NOTE	Harmonized as EN 60335-1:2012 (modified).
IEC 60364	NOTE	Harmonized in HD 384 / HD 60364 series (partly modified).
IEC 60721-3-7:1995 + A1:1996	NOTE	Harmonized as EN 60721-3-7:1995 (not modified) + A1:1997 (not modified).
IEC 60950-1:2005	NOTE	Harmonized as EN 60950-1:2006 (modified).
IEC 61032:1997	NOTE	Harmonized as EN 61032:1998 (not modified).
ISO 10651-2:2004	NOTE	Harmonized as EN ISO 10651-2:2009 (not modified).

Foreword to amendment A1

The text of document 62A/1396/FDIS, future IEC 60601-1-12/A1, prepared by SC 62A "Common aspects of electrical equipment used in medical practice" of IEC/TC 62 "Electrical equipment in medical practice" was submitted to the IEC-CENELEC parallel vote and approved by CENELEC as EN 60601-1-12:2015/A1:2020.

The following dates are fixed:

- latest date by which the document has to be implemented at national (dop) 2021-05-26 level by publication of an identical national standard or by endorsement
- latest date by which the national standards conflicting with the (dow) 2023-08-26 document have to be withdrawn

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC shall not be held responsible for identifying any or all such patent rights.

This document has been prepared under a mandate given to CENELEC by the European Commission and the European Free Trade Association.

Endorsement notice

The text of the International Standard IEC 60601-1-12:2014/A1:2020 was approved by CENELEC as a European Standard without any modification.

Annex ZA

(normative)

Normative references to international publications with their corresponding European publications

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE 1 Where an International Publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

NOTE 2 Up-to-date information on the latest versions of the European Standards listed in this annex is available here: www.cenelec.eu.

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
IEC 60601-1	2005	Medical electrical equipment - Part 1: General requirements for basic safety and essential performance	EN 60601-1	2006
-	-		+ corrigendum Mar.	2010
+ A1	2012		+ A1	2013
-	-		+ A12	2014
+ A2	2020		+A2	— ¹
IEC 60601-1-2	2014	Medical electrical equipment - Part 1-2: General requirements for basic safety and essential performance - Collateral Standard: Electromagnetic disturbances - Requirements and tests	EN 60601-1-2	2015
+ A1	2020		+A1	— ²
IEC 60601-1-6	2010	Medical electrical equipment - Part 1-6: General requirements for basic safety and essential performance - Collateral standard: Usability	EN 60601-1-6	2010
+ A1	2013		+ A1	2015
+ A2	2020		+A2	— ³

¹ Under preparation. Stage at time of publication: EN 60601-1:2006/FprA2:2020.

² Under preparation. Stage at time of publication: EN 60601-1-2:2015/FprA1:2020.

³ Under preparation. Stage at time of publication: EN 60601-1-6:2010/FprA2:2020.

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
IEC 60601-1-8	2006	Medical electrical equipment - Part 1-8: General requirements for basic safety and essential performance - Collateral Standard: General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems	EN 60601-1-8	2007
-	-		+ corrigendum Mar.	2010
+ A1	2012		+ A1	2013
-	-		+ AC	2014
-	-		+ A11	2017
+ A2	2020		+A2	— ⁴
IEC 60601-1-11	2015	Medical electrical equipment - Part 1-11: General requirements for basic safety and essential performance - Collateral Standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment	EN 60601-1-11	2015
+ A1	2020		+A1	— ⁵
ISO 7010	2019	Graphical symbols - Safety colours and safety signs - Registered safety signs	-	-

⁴ Under preparation. Stage at time of publication: EN 60601-1-8:2007/FprA2:2020.

⁵ Under preparation. Stage at time of publication: EN 60601-1-11:2015/FprA2:2020.

Annex ZZA
(informative)

Coverage of Essential Requirements of EU Directives

This European Standard has been prepared under a mandate given to CENELEC by the European Commission and the European Free Trade Association, and within its scope the Standard covers all relevant essential requirements given in Annex I of EU Directive 93/42/EEC of 14 June 1993 concerning medical devices.

Compliance with this standard provides one means of conformity with the specified essential requirements of the Directive concerned.

WARNING: Other requirements and other EU Directives can be applied to the products falling within the scope of this standard.

Annex ZZB
(informative)

Coverage of Essential Requirements of EU Directives

This European Standard has been prepared under a mandate given to CENELEC by the European Commission and the European Free Trade Association, and within its scope the Standard covers all relevant essential requirements given in Annex 1 of EU Directive 90/385/EEC of 20 June 1990 relating to active implantable medical devices.

Compliance with this standard provides one means of conformity with the specified essential requirements of the Directive concerned.

WARNING: Other requirements and other EU Directives can be applied to the products falling within the scope of this standard.

This page deliberately left blank

CONTENTS

FOREWORD	4
INTRODUCTION	7
INTRODUCTION to Amendment 1	7
1 Scope, object and related standards	8
1.1 * Scope	8
1.2 * Object	8
1.3 Related standards	9
1.3.1 IEC 60601-1	9
1.3.2 Particular standards	9
2 Normative references	9
3 Terms and definitions	10
4 General requirements	11
4.1 * Additional requirements for SUPPLY MAINS for ME EQUIPMENT and ME SYSTEMS	11
4.2 * Environmental conditions for ME EQUIPMENT	11
4.2.1 * Environmental conditions of transport and storage between uses	11
4.2.2 * Environmental operating conditions	13
5 * Classification of ME EQUIPMENT and ME SYSTEMS	15
6 ME EQUIPMENT identification, marking and documents	15
6.1 * Additional requirements for legibility of markings	15
6.2 * Additional requirements for marking of IP classification	16
6.3 * Instructions for use	16
6.3.1 Additional general requirements	16
6.3.2 * Additional requirements for an electrical power source	17
6.3.3 Additional requirements for ME EQUIPMENT start-up PROCEDURE	17
6.3.4 * Additional requirements for operating instructions	17
6.3.5 Additional requirements for ME EQUIPMENT messages	18
6.4 Technical description – FIXED or PERMANENTLY INSTALLED CLASS I ME EQUIPMENT	18
7 * Protection against electrical HAZARDS from ME EQUIPMENT	18
8 Protection against excessive temperatures and other HAZARDS	18
8.1 Additional requirements for ingress of water or particulate matter into ME EQUIPMENT and ME SYSTEMS	18
8.1.1 * Ingress of water or particulate matter into ME EQUIPMENT	18
8.1.2 * Ingress of water or particulate matter into ME SYSTEMS	19
8.2 Additional requirements for interruption of the power supply to ME EQUIPMENT and ME SYSTEM	19
8.3 * Additional requirements for INTERNAL ELECTRICAL POWER SOURCE for ME EQUIPMENT	20
9 * Accuracy of controls and instruments and protection against hazardous outputs	20
10 Construction of ME EQUIPMENT	21
10.1 * Additional requirements for mechanical strength of ME EQUIPMENT intended for the EMS ENVIRONMENT	21
10.1.1 General requirements for mechanical strength	21
10.1.2 * Requirements for mechanical strength for FIXED or PERMANENTLY INSTALLED ME EQUIPMENT intended for use in a road ambulance	21

10.1.3	* Requirements for mechanical strength for TRANSPORTABLE ME EQUIPMENT	22
10.1.4	* Requirements for mechanical strength for ME EQUIPMENT intended for airborne use	23
10.2	Requirements for mounting of ME EQUIPMENT.....	24
11	Additional requirements for electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS	24
Annex A	(informative) General guidance and rationale.....	25
A.1	General guidance.....	25
A.2	Rationale for particular clauses and subclauses.....	27
Annex B	(informative) Guide to marking and labelling requirements for ME EQUIPMENT and ME SYSTEMS	41
B.1	Marking on the outside of ME EQUIPMENT, ME SYSTEMS or their parts	41
B.2	ACCOMPANYING DOCUMENTS, instructions for use.....	41
B.3	ACCOMPANYING DOCUMENTS, technical description.....	42
Annex C	(informative) Symbols on marking.....	43
Bibliography		45
Index of defined terms used in this collateral standard		47
Figure A.1	– Saturation water vapour pressure as function of temperature.....	30
Table 1	– Mechanical strength test applicability	21
Table A.1	– Saturation water vapour pressure as function of temperature	31
Table B.1	– Marking on the outside of ME EQUIPMENT, ME SYSTEMS or their parts	41
Table B.2	– ACCOMPANYING DOCUMENTS, instructions for use	41
Table B.3	– ACCOMPANYING DOCUMENTS, technical description.....	42
Table C.1	– General symbols	43

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –

Part 1-12: General requirements for basic safety and essential performance – Collateral Standard: Requirements for medical electrical equipment and medical electrical systems intended for use in the emergency medical services environment

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International standard IEC 60601-1-12 has been prepared by a joint working group of IEC subcommittee 62A: Common aspects of electrical equipment used in medical practice, of IEC technical committee 62: Electrical equipment in medical practice, and ISO subcommittee SC3: Lung ventilators and related devices, of ISO technical committee 121: Anaesthetic and respiratory equipment.

This first edition constitutes a collateral standard to IEC 60601-1 (third edition): *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance* hereafter referred to as the general standard.

The text of this collateral standard is based on the following documents:

FDIS	Report on voting
62A/932/FDIS	62A/938/RVD

Full information on the voting for the approval of this collateral standard can be found in the report on voting indicated in the above table. In ISO, this International Standard has been approved by 18 P-members out of 19 having cast a vote.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In the 60601 series of publications, collateral standards specify general requirements for safety applicable to:

- a subgroup of MEDICAL ELECTRICAL EQUIPMENT (e.g. radiological equipment); or
- a specific characteristic of all MEDICAL ELECTRICAL EQUIPMENT, not fully addressed in the general standard (e.g. ALARM SYSTEMS).

In this collateral standard, the following print types are used:

- requirements and definitions: roman type.
- *test specifications: italic type.*
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS COLLATERAL STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this standard, the term

- “clause” means one of the numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.3.1 are all subclauses of Clause 7).

References to clauses within this standard are preceded by the term “Clause” followed by the clause number. References to subclauses within this collateral standard are by number only.

In this standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

Clauses, subclauses and definitions for which a rationale is provided in informative Annex A are marked with an asterisk (*).

A list of all parts of the IEC 60601 series, published under the general title: *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of this publication will remain unchanged until the stability date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

NOTE The attention of Member Bodies and National Committees is drawn to the fact that equipment manufacturers and testing organizations may need a transitional period following publication of a new, amended or revised ISO or IEC publication in which to make products in accordance with the new requirements and to equip themselves for conducting new or revised tests. It is the recommendation of the committee that the content of this publication be adopted for implementation nationally not earlier than 3 years from the date of publication.

IMPORTANT – The 'colour inside' logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.

INTRODUCTION

Medical practice is increasingly using MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS for monitoring, treatment or diagnosis of PATIENTS in the EMERGENCY MEDICAL SERVICES ENVIRONMENT (see 3.1). The safety of MEDICAL ELECTRICAL EQUIPMENT in this uncontrolled, rough environment is a cause for concern.

This collateral standard was developed with contributions from clinicians, engineers and regulators. The terminology, requirements, general recommendations and guidance of this collateral standard are intended to be useful for MANUFACTURERS of MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS and for technical committees responsible for the development of particular standards.

INTRODUCTION to Amendment 1

The first edition of IEC 60601-1-12 was published in 2014. Since the publication of IEC 60601-1-12:2014, the IEC Subcommittee (SC) 62A Secretariat has been collecting issues from a variety of sources including comments from National Committees. At the November 2015 meeting of IEC/SC 62A in Kobe, Japan, the subcommittee initiated a process to identify high-priority issues that need to be considered in an amendment and should not wait until the second edition of IEC 60601-1-12, which is presently targeted for publication sometime after 2024.

As directed in item 1 of Kobe Resolution 1, the IEC/SC 62A Chairman Advisory Group (CAG) considered the 27 issues collected by the SC/62A Secretariat for IEC 60601-1-12:2014 and determined that none met the selection criteria stated in Kobe Resolution 1.

However, an amendment is needed to update the references to IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020. In London in 2018, SC 62A approved the development of an administrative amendment to IEC 60601-1-12:2014.

Because this is an amendment to IEC 60601-1-12:2014, the style in force at the time of publication of IEC 60601-1-12 has been applied to this amendment. The style specified in ISO/IEC Directives Part 2:2018 has only been applied when implementing the new style guidance would not result in additional editorial changes.

Users of this document should note that when constructing the dated references to specific elements in a standard, such as definitions, amendments are only referenced if they modified the text being cited. For example, if a reference is made to a definition that has not been modified by an amendment, then the reference to the amendment is not included in the dated reference.

MEDICAL ELECTRICAL EQUIPMENT –

Part 1-12: General requirements for basic safety and essential performance – Collateral Standard: Requirements for medical electrical equipment and medical electrical systems intended for use in the emergency medical services environment

1 Scope, object and related standards

1.1 * Scope

This International Standard applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS, hereafter referred to as ME EQUIPMENT and ME SYSTEMS, which are intended, as indicated in the instructions for use by their MANUFACTURER, for use in the EMS ENVIRONMENT (EMERGENCY MEDICAL SERVICES ENVIRONMENT), as defined in 3.1.

NOTE 1 For the purposes of this standard, the intent of the MANUFACTURER is indicated in the instructions for use. The RESPONSIBLE ORGANIZATION and the OPERATOR need to be aware that any other use outside the MANUFACTURER'S INTENDED USE can result in a HAZARDOUS SITUATION for the PATIENT.

The EMS ENVIRONMENT includes

- responding to and providing life support at the scene of an emergency to a PATIENT reported as experiencing injury or illness in a pre-hospital setting, and transporting the PATIENT, while continuing such life support care, to an appropriate professional healthcare facility for further care.
- providing monitoring, treatment or diagnosis during transport between professional healthcare facilities.

This International Standard does not apply to ME EQUIPMENT and ME SYSTEMS intended solely for use in the HOME HEALTHCARE ENVIRONMENT covered by IEC 60601-1-11 or solely for use in professional healthcare facilities covered by IEC 60601-1 without the additions of IEC 60601-1-11 or this collateral standard. ME EQUIPMENT and ME SYSTEMS are often not solely intended for one environment. Such ME EQUIPMENT or ME SYSTEM can be intended for multiple use environments, and as such, if also intended for use in the EMS ENVIRONMENT, are within the scope of this standard.

EXAMPLE ME EQUIPMENT or ME SYSTEM intended for both the EMS ENVIRONMENT and the professional healthcare facility environment.

NOTE 2 EMS ENVIRONMENT ME EQUIPMENT and ME SYSTEMS can be used in locations with unreliable electrical sources and outdoor environmental conditions.

1.2 * Object

The object of this collateral standard is to provide general requirements for ME EQUIPMENT and ME SYSTEMS carried to the scene of an emergency and used there, as well as in transport, in situations where the ambient conditions differ from indoor conditions.

The object of this collateral standard is to specify general requirements that are in addition to those of the general standard and to serve as the basis for particular standards.

1.3 Related standards

1.3.1 IEC 60601-1

A1 For ME EQUIPMENT and ME SYSTEMS, this collateral standard complements IEC 60601-1. **A1**

When referring to IEC 60601-1 or to this collateral standard, either individually or in combination, the following conventions are used:

- "the general standard" designates IEC 60601-1 alone, **A1** including any amendments **A1**;
- "this collateral standard" designates IEC 60601-1-12 alone, **A1** including any amendments **A1**;
- "this standard" designates the combination of the general standard and this collateral standard.

1.3.2 Particular standards

A requirement in a particular standard takes priority over the corresponding requirement in this collateral standard.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE 1 The way in which these referenced documents are cited in normative requirements determines the extent (in whole or in part) to which they apply.

NOTE 2 Informative references are listed in the bibliography on page 45.

IEC 60068-2-27:2008, *Environmental testing – Part 2-27: Tests – Test Ea and guidance: Shock*

IEC 60068-2-31:2008, *Environmental testing – Part 2-31: Tests – Test Ec: Rough handling shocks, primarily for equipment-type specimens*

IEC 60068-2-64:2008, *Environmental testing – Part 2-64: Tests – Test Fh: Vibration, broadband random and guidance*

IEC 60529:1989, *Degrees of protection provided by enclosures (IP Code)*

IEC 60529:1989/AMD1:1999¹

A1 IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*

IEC 60601-1:2005/AMD1:2012

IEC 60601-1:2005/AMD2:2020 **A1**

A1 IEC 60601-1-2:2014, *Medical electrical equipment – Part 1-2: General requirements for basic safety and essential performance – Collateral Standard: Electromagnetic disturbances – Requirements and tests*

IEC 60601-1-2:2014/AMD1:2020 **A1**

A1 IEC 60601-1-6:2010, *Medical electrical equipment – Part 1-6: General requirements for basic safety and essential performance – Collateral standard: Usability* **A1**

¹ There exists a consolidated edition 2.1(2001) including IEC 60529:1989 and its Amendment 1:1999.