BS EN 61094-3:2016

Incorporating corrigendum December 2016

BSI Standards Publication

Electroacoustics — Measurement microphones

Part 3: Primary method for free-field calibration of laboratory standard microphones by the reciprocity technique (IEC 61094-3:2016)

Documents
Document Sharing Hub

Full document details: BS EN 61094-3:2016

National foreword

This British Standard is the UK implementation of EN 61094-3:2016. It is identical to IEC 61094-3:2016, incorporating corrigendum December 2016. It supersedes BS EN 61094-3:1996 which is withdrawn.

The UK participation in its preparation was entrusted to Technical Committee EPL/29, Electroacoustics.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2017.
Published by BSI Standards Limited 2017

ISBN 978 0 580 97575 2
ICS 17.140.50; 33.160.50

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 30 September 2016.

Amendments/corrigenda issued since publication

Date	Text affected
28 February 2017	Implementation of IEC corrigendum December 2016: subclause 5.7.2 corrected

Documents
Document Sharing Hub

Full document details: BS EN 61094-3:2016

EUROPEAN STANDARD

NORME EUROPÉENNE

EUROPÄISCHE NORM

EN 61094-3

September 2016

ICS 17.140.50; 33.160.50

Supersedes EN 61094-3:1995

English Version

Electroacoustics - Measurement microphones - Part 3: Primary method for free-field calibration of laboratory standard microphones by the reciprocity technique (IEC 61094-3:2016)

Électroacoustique - Microphones de mesure - Partie 3: Méthode primaire pour l'étalonnage en champ libre des microphones étalons de laboratoire par la méthode de réciprocité (IEC 61094-3:2016)

Messmikrofone - Teil 3: Primärverfahren zur Freifeld-Kalibrierung von Laboratoriums-Normalmikrofonen nach der Reziprozitätsmethode (IEC 61094-3:2016)

This European Standard was approved by CENELEC on 2016-07-19. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

© 2016 CENELEC All rights of exploitation in any form and by any means reserved worldwide for CENELEC Members.

Ref. No. EN 61094-3:2016 E

AVA Documents Document Sharing Hub

Full document details: BS EN 61094-3:2016

European foreword

The text of document 29/873/CDV, future edition 2 of IEC 61094-3, prepared by IEC TC 29, Electroacoustics, was submitted to the IEC-CENELEC parallel vote and was approved by CENELEC as EN 61094-3:2016.

The following dates are fixed:

- latest date by which the document has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2017-04-19
- latest date by which the national standards conflicting with the document have to be withdrawn (dow) 2019-07-19

This document supersedes EN 61094-3:1995.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC [and/or CEN] shall not be held responsible for identifying any or all such patent rights.

Endorsement notice

The text of the International Standard IEC 61094-3:2016 was approved by CENELEC as a European Standard without any modification.

In the official version, for Bibliography, the following note has to be added for the standard indicated:

IEC 61094-8:2012 NOTE Harmonized as EN 61094-8:2012.

Documents
Document Sharing Hub
Full document details: BS EN 61094-3:2016

Annex ZA
(normative)
Normative references to international publications with their corresponding European publications

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE 1 When an International Publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

NOTE 2 Up-to-date information on the latest versions of the European Standards listed in this annex is available here: www.cenelec.eu.

Publication	Year	Title	EN/HD	Year
IEC 61094-1	2000	Measurement microphones -- Part 1: Specifications for laboratory standard microphones	EN 61094-1	2000
IEC 61094-2	2009	Electroacoustics - Measurement microphones -- Part 2: Primary method for the pressure calibration of laboratory standard microphones by the reciprocity technique	EN 61094-2	2009
ISO 9613-1	-	Acoustics; attenuation of sound during propagation outdoors; part_1: calculation of the absorption of sound by the atmosphere	-	-
IEC/TS 61094-7	-	Measurement microphones -- Part 7: Values for the difference between free-field and pressure sensitivity levels of laboratory standard microphones	-	-
ISO/IEC Guide 98-3	-	Uncertainty of measurement - Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)	-	-

Documents
Document Sharing Hub
Full document details: BS EN 61094-3:2016

This page deliberately left blank

Documents
Document Sharing Hub

CONTENTS

FOREWORD....4
1 Scope....6
2 Normative references....6
3 Terms and definitions....6
4 Reference environmental conditions....7
5 Principles of free-field calibration by reciprocity....7
5.1 General principles....7
5.1.1 General....7
5.1.2 General principles using three microphones....7
5.1.3 General principles using two microphones and an auxiliary sound source....8
5.2 Basic expressions....8
5.3 Insert voltage technique....9
5.4 Free-field receiving characteristics of a microphone....9
5.5 Free-field transmitting characteristics of a microphone....10
5.6 Reciprocity procedure....11
5.7 Final expressions for the free-field sensitivity....11
5.7.1 Method using three microphones....11
5.7.2 Method using two microphones and an auxiliary sound source....12
6 Factors influencing the free-field sensitivity....12
6.1 General....12
6.2 Polarizing voltage....12
6.3 Shield configuration....12
6.4 Acoustic conditions....13
6.5 Position of the acoustic centre of a microphone....13
6.6 Dependence on environmental conditions....14
6.6.1 General....14
6.6.2 Static pressure....14
6.6.3 Temperature....14
6.6.4 Humidity....14
6.6.5 Transformation to reference environmental conditions....14
6.7 Considerations concerning measurement space....15
7 Calibration uncertainty components....15
7.1 General....15
7.2 Electrical transfer impedance....15
7.3 Deviations from ideal free-field conditions....15
7.4 Attenuation of sound in air....16
7.5 Polarizing voltage....16
7.6 Physical properties of air....16
7.7 Imperfection of theory....16
7.8 Uncertainty on free-field sensitivity level....17
Annex A (informative) Values for the position of the acoustic centre....19
Annex B (normative) Values of the air attenuation coefficient....20
B.1 General....20
B.2 Calculation procedure....20

AVA Documents
Document Sharing Hub
Full document details: BS EN 61094-3:2016

Annex C (informative) Environmental influence on the sensitivity of microphones 23
C.1 General.....23
C.2 Dependence on static pressure 23
C.3 Dependence on temperature 23
Annex D (informative) Application of time selective techniques for removal of unwanted reflections and acoustic interference between microphones 25
D.1 General.....25
D.2 Practical considerations 25
D.2.1 Signal-to-noise ratio 25
D.2.2 Reflections from walls and measurement rig 25
D.3 Frequency limitations 26
D.3.1 General 26
D.3.2 Measurements based on frequency sweeps 26
D.3.3 Measurements based on pure tones 26
D.4 Generating missing portions of the frequency response previous to transforming to the time-domain. 27
D.4.1 General 27
D.4.2 Missing frequencies below the minimum measurement frequency.....27
D.4.3 Missing frequencies above the maximum measured frequency 27
D.4.4 Filtering the extended frequency response.....28
Bibliography 29

Figure 1 – Equivalent circuit for a receiving microphone under free-field conditions 9
Figure 2 – Equivalent circuit for a transmitting microphone under free-field conditions.....10
Figure A.1 – Example of the estimated values of the acoustic centres of LS1P and LS2aP microphones given in the bibliographical references for Annex A 19

Table 1 – Uncertainty components 17
Table B.1 – Values for attenuation of sound pressure in air (in dB/m).....22

Documents
Document Sharing Hub
Full document details: BS EN 61094-3:2016

INTERNATIONAL ELECTROTECHNICAL COMMISSION

ELECTROACOUSTICS – MEASUREMENT MICROPHONES –

Part 3: Primary method for free-field calibration of laboratory standard microphones by the reciprocity technique

FOREWORD

1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
6) All users should ensure that they have the latest edition of this publication.
7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 61094-3 has been prepared by IEC technical committee 29: Electroacoustics.

This second edition cancels and replaces the first edition published in 1995. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

a) a new informative annex describing the use of time-selective techniques to minimize the influence of acoustic reflections from the measurement setup;
b) provision for the calibration of microphones in driven shield configuration.

Documents
Document Sharing Hub
Full document details: BS EN 61094-3:2016

The text of this standard is based on the following documents:

CDV	Report on voting
29/873/CDV	29/892A/RVC

Full information on the voting for the approval of this standard can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

A list of all parts in the IEC 61094 series, published under the general title *Electroacoustics – Measurement microphones*, can be found on the IEC website.

The committee has decided that the contents of this publication will remain unchanged until the stability date indicated on the IEC website under "http://webstore.iec.ch" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

IMPORTANT – The 'colour inside' logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.

Documents
Document Sharing Hub

Full document details: BS EN 61094-3:2016

ELECTROACOUSTICS – MEASUREMENT MICROPHONES –

Part 3: Primary method for free-field calibration of laboratory standard microphones by the reciprocity technique

1 Scope

This part of IEC 61094

- specifies a primary method of determining the complex free-field sensitivity of laboratory standard microphones so as to establish a reproducible and accurate basis for the measurement of sound pressure under free-field conditions,
- is applicable to laboratory standard microphones meeting the requirements of IEC 61094-1,
- is intended for use by laboratories with highly experienced staff and specialized equipment.

NOTE The calibration principle described in this part of IEC 61094 is also applicable to working standard microphones, preferably used without their protection grid.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 61094-1:2000, *Measurement microphones – Part 1: Specifications for laboratory standard microphones*

IEC 61094-2:2009, *Electroacoustics – Measurement microphones – Part 2: Primary method for pressure calibration of laboratory standard microphones by the reciprocity technique*

IEC TS 61094-7:2006, *Measurement microphones – Part 7: Values for the difference between free-field and pressure sensitivity levels of laboratory standard microphones*

ISO 9613-1, *Acoustics – Attenuation of sound during propagation outdoors – Part 1: Calculation of the absorption of sound by the atmosphere*

ISO/IEC Guide 98-3, *Uncertainty of measurement – Part 3: Guide to the expression of uncertainty in measurement* (GUM:1995)

3 Terms and definitions

For the purposes of this document, the terms and definitions given in IEC 61094-1, IEC 61094-2, ISO/IEC Guide 98-3 and the following apply.

3.1
phase
<free-field sensitivity of a microphone> phase angle between the open-circuit voltage and the sound pressure that would exist at the position of the acoustic centre of the microphone in the absence of the microphone, for a sinusoidal plane progressive wave of given frequency and direction of sound incidence, and for given environmental conditions

AVA Documents
Document Sharing Hub