

BS EN 61672-1:2013



BSI Standards Publication

Electroacoustics — Sound level meters

Part 1: Specifications

bsi.

...making excellence a habit.
AVA Documents™
Document Sharing Hub

National foreword

This British Standard is the UK implementation of EN 61672-1:2013. It is identical to IEC 61672-1:2013. It supersedes BS EN 61672-1:2003 which is withdrawn.

The UK participation in its preparation was entrusted to Technical Committee EPL/29, Electroacoustics.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2013

Published by BSI Standards Limited 2013

ISBN 978 0 580 68846 1

ICS 17.140.50

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 31 December 2013.

Amendments issued since publication

Date	Text affected
------	---------------

English version

**Electroacoustics -
Sound level meters -
Part 1: Specifications
(IEC 61672-1:2013)**

Electroacoustique -
Sonomètres -
Partie 1: Spécifications
(CEI 61672-1:2013)

Elektroakustik -
Schallpegelmesser -
Teil 1: Anforderungen
(IEC 61672-1:2013)

This European Standard was approved by CENELEC on 2013-11-04. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

CENELEC

European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

CEN-CENELEC Management Centre: Avenue Marnix 17, B - 1000 Brussels

© 2013 CENELEC - All rights of exploitation in any form and by any means reserved worldwide for CENELEC members.

Ref. No. EN 61672-1:2013 E

AVA Documents
Document Sharing Hub

Foreword

The text of document 29/812/FDIS, future edition 2 of IEC 61672-1, prepared by IEC/TC 29 "Electroacoustics" in cooperation with the International Organization of Legal Metrology (OIML), was submitted to the IEC-CENELEC parallel vote and approved by CENELEC as EN 61672-1:2013.

The following dates are fixed:

- latest date by which the document has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2013-08-04
- latest date by which the national standards conflicting with the document have to be withdrawn (dow) 2016-11-04

This document supersedes EN 61672-1:2003.

EN 61672-1:2013 includes the following significant technical changes with respect to EN 61672-1:2003.

In this second edition, conformance to specifications is demonstrated when

- a) measured deviations from design goals do not exceed the applicable acceptance limits, and
- b) the uncertainty of measurement does not exceed the corresponding maximum-permitted uncertainty, with both uncertainties determined for a coverage probability of 95 %.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC [and/or CEN] shall not be held responsible for identifying any or all such patent rights.

Endorsement notice

The text of the International Standard IEC 61672-1:2013 was approved by CENELEC as a European Standard without any modification.

Annex ZA

(normative)

Normative references to international publications with their corresponding European publications

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE When an international publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
IEC 60942	-	Electroacoustics - Sound calibrators	EN 60942	-
IEC 61000-4-2	2008	Electromagnetic compatibility (EMC) Part 4-2: Testing and measurement techniques - Electrostatic discharge immunity test	EN 61000-4-2	2009
IEC 61000-6-2	2005	Electromagnetic compatibility (EMC) Part 6-2: Generic standards - Immunity for industrial environments	EN 61000-6-2 + corr. September	2005 2005
IEC 61094-6	-	Measurement microphones Part 6: Electrostatic actuators for determination of frequency response	EN 61094-6	-
IEC 61183	-	Electroacoustics - Random-incidence and diffuse-field calibration of sound level meters	EN 61183	-
IEC 62585	-	Electroacoustics - Methods to determine corrections to obtain the free-field response of a sound level meter	EN 62585	-
ISO/IEC Guide 98-4	2012	Uncertainty of measurement Part 4: Role of measurement uncertainty in conformity assessment	-	-
ISO/IEC Guide 99		International vocabulary of metrology - Basic and general concepts and associated terms (VIM)	-	-
CISPR 16-1-1	2010	Specification for radio disturbance and	EN 55016-1-1	2010
+ corr. October	2010	immunity measuring apparatus and	-	-
+ corr. October	2011	methods	-	-
+ A1	2010	Part 1-1: Radio disturbance and immunity measuring apparatus - Measuring apparatus	+ A1	2010

CONTENTS

INTRODUCTION.....	6
1 Scope.....	7
2 Normative references	8
3 Terms and definitions	8
4 Reference environmental conditions	14
5 Performance specifications.....	14
5.1 General.....	14
5.2 Adjustments at the calibration check frequency	17
5.3 Corrections to indicated levels.....	17
5.3.1 General	17
5.3.2 Reflections and diffraction	17
5.3.3 Windscreens.....	18
5.3.4 Format for correction data	18
5.3.5 Corrections for use during periodic testing.....	19
5.4 Directional response.....	19
5.5 Frequency weightings.....	20
5.6 Level linearity	23
5.7 Self-generated noise	24
5.8 Time-weightings F and S	24
5.9 Toneburst response.....	24
5.10 Response to repeated tonebursts	26
5.11 Overload indication	27
5.12 Under-range indication	27
5.13 C-weighted peak sound level.....	27
5.14 Stability during continuous operation	28
5.15 High-level stability	28
5.16 Reset	29
5.17 Thresholds	29
5.18 Display	29
5.19 Analogue or digital output.....	29
5.20 Timing facilities	30
5.21 Radio frequency emissions and disturbances to a public power supply.....	30
5.22 Crosstalk	31
5.23 Power supply.....	31
6 Environmental, electrostatic, and radio-frequency requirements	32
6.1 General.....	32
6.2 Static pressure	32
6.3 Air temperature	32
6.4 Humidity.....	33
6.5 Electrostatic discharge	33
6.6 A.C. power-frequency and radio-frequency fields.....	33
6.7 Mechanical vibration.....	34
7 Provision for use with auxiliary devices	35
8 Marking	35
9 Instruction Manual	35

9.1	General	35
9.2	Information for operation	36
9.2.1	General	36
9.2.2	Design features	36
9.2.3	Power supply	37
9.2.4	Adjustments at the calibration check frequency	37
9.2.5	Corrections to indicated levels	37
9.2.6	Operating the sound level meter	37
9.2.7	Accessories	38
9.2.8	Influence of variations in environmental conditions	38
9.3	Information for testing	39
Annex A (informative)	Relationship between tolerance interval, corresponding acceptance interval and the maximum-permitted uncertainty of measurement	41
Annex B (normative)	Maximum-permitted uncertainties of measurement	42
Annex C (informative)	Example assessments of conformance to specifications of this standard	44
Annex D (normative)	Frequencies at fractional-octave intervals	47
Annex E (normative)	Analytical expressions for frequency-weightings C, A, and Z	49
Figure 1	– Principal steps involved in forming a time-weighted sound level	10
Figure A.1	– Relationship between tolerance interval, corresponding acceptance interval and the maximum-permitted uncertainty of measurement	41
Figure C.1	– Examples of assessment of conformance	46
Table 1	– Acceptance limits for the difference between a measured windscreen correction and the corresponding correction given in the Instruction Manual	18
Table 2	– Acceptance limits for deviations of directional response from the design goal	20
Table 3	– Frequency weightings and acceptance limits	22
Table 4	– Reference 4 kHz toneburst responses and acceptance limits	25
Table 5	– Reference differences for C-weighted peak sound levels and acceptance limits	28
Table 6	– Limits for conducted disturbance to the voltage of a public supply of electric power	31
Table B.1	– Maximum-permitted uncertainties of measurement for a coverage probability of 95 %	42
Table C.1	– Examples of assessment of conformance	45
Table D.1	– Frequencies at one-third-octave intervals	47
Table D.2	– Frequencies at one-sixth-octave intervals	48
Table D.3	– Frequencies at one-twelfth-octave intervals	48

INTRODUCTION

For assessments of conformance to performance specifications, this second edition of IEC 61672-1 uses different criteria than were used for the 2002 first edition.

In the period from 1961 to 1985, International Standards for sound level meters did not provide any requirements or recommendations to account for the uncertainty of measurement in assessments of conformance to specifications.

This absence of requirements or recommendations to account for uncertainty of measurement created ambiguity in determinations of conformance to specifications for situations where a measured deviation from a design goal was close to a limit of the allowed deviation. If conformance was determined based on whether a measured deviation did or did not exceed the limits, the end-user of the sound level meter incurred the risk that the true deviation from a design goal exceeded the limits.

To remove this ambiguity, IEC Technical Committee 29, at its meeting in 1996, adopted a policy to account for measurement uncertainty in assessments of conformance in International Standards that it prepares.

The first edition (2002) of IEC 61672-1 accounted for measurement uncertainty by giving two explicit criteria for determining conformance to the specifications. The two criteria were (a) that measured deviations from design goals, extended by the expanded uncertainty of measurement, do not exceed the applicable tolerance limits and (b) that the expanded uncertainty of measurement does not exceed agreed-upon maximum values. For most performance specifications, the tolerance limits were calculated essentially by extending the allowances for design and manufacturing from the 1979 and 1985 International Standards for sound level meters by the applicable maximum-permitted expanded uncertainties of measurement. Tolerance limits were intended to represent the limits for true deviations from design goals with a coverage probability of 95 %.

This second edition of IEC 61672-1 uses an amended criterion for assessing conformance to a specification. Conformance is demonstrated when (a) measured deviations from design goals do not exceed the applicable *acceptance limits* and (b) the uncertainty of measurement does not exceed the corresponding maximum-permitted uncertainty. Acceptance limits are analogous to the allowances for design and manufacturing implied in the first edition (2002) of IEC 61672-1. Actual and maximum-permitted uncertainties are determined for a coverage probability of 95 %. The amended criterion for assessing conformance does not necessitate any change to the design of a sound level meter in order to conform to the specifications of this International Standard.

The maximum-permitted uncertainties of measurement are not equivalent to the uncertainties associated with the measurement of a sound level. The uncertainty of a measured sound level is evaluated from the anticipated deviations of the electroacoustical performance of the sound level meter from the relevant design goals as well as estimates of the uncertainties associated with the specific measurement situation. Unless more-specific information is available, the evaluation of the contribution of a specific sound level meter to a total measurement uncertainty can be based on the acceptance limits and maximum-permitted uncertainties specified in this standard.

ELECTROACOUSTICS – SOUND LEVEL METERS –

Part 1: Specifications

1 Scope

This part of IEC 61672 gives electroacoustical performance specifications for three kinds of sound measuring instruments:

- a time-weighting sound level meter that measures exponential-time-weighted, frequency-weighted sound levels;
- an integrating-averaging sound level meter that measures time-averaged, frequency-weighted sound levels; and
- an integrating sound level meter that measures frequency-weighted sound exposure levels.

Sound level meters conforming to the requirements of this standard have a specified frequency response for sound incident on the microphone from one principal direction in an acoustic free field or successively from random directions.

Sound level meters specified in this standard are intended to measure sounds generally in the range of human hearing.

NOTE The AU frequency weighting specified in IEC 61012 can be applied for measurements of A-weighted sound levels of audible sound in the presence of a source that contains spectral components at frequencies greater than 20 kHz.¹

Two performance categories, class 1 and class 2, are specified in this standard. In general, specifications for class 1 and class 2 sound level meters have the same design goals and differ mainly in the acceptance limits and the range of operational temperature. Acceptance limits for class 2 are greater than, or equal to, those for class 1.

This standard is applicable to a range of designs for sound level meters. A sound level meter may be a self-contained hand-held instrument with an attached microphone and a built-in display device. A sound level meter may be comprised of separate components in one or more enclosures and may be capable of displaying a variety of acoustical signal levels. Sound level meters may include extensive analogue or digital signal processing, separately or in combination, with multiple analogue and digital outputs. Sound level meters may include general-purpose computers, recorders, printers, and other devices that form a necessary part of the complete instrument.

Sound level meters may be designed for use with an operator present or for automatic and continuous measurements of sound level without an operator present. Specifications in this standard for the response to sound waves apply without an operator present in the sound field.

¹ IEC 61012, *Filters for the measurement of audible sound in the presence of ultrasound*.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60942, *Electroacoustics – Sound calibrators*

IEC 61000-4-2:2008, *Electromagnetic compatibility (EMC) – Part 4-2: Testing and measurement techniques – Electrostatic discharge immunity test*

IEC 61000-6-2:2005, *Electromagnetic compatibility (EMC) – Part 6-2: Generic standards – Immunity for industrial environments*

IEC 61094-6, *Measurement microphones – Part 6: Electrostatic actuators for determination of frequency response*

IEC 61183, *Electroacoustics – Random-incidence and diffuse-field calibration of sound level meters*

IEC 62585, *Electroacoustics – Methods to determine corrections to obtain the free-field response of a sound level meter*

ISO/IEC Guide 98-4:2012, *Evaluation of measurement data – The role of measurement uncertainty in conformance assessment*

ISO/IEC Guide 99, *International vocabulary of metrology – Basic and general concepts and associated terms (VIM)*

CISPR 16-1-1:2010, *Specification for radio disturbance and immunity measuring apparatus and methods – Part 1-1: Radio disturbance and immunity measuring apparatus – Measuring apparatus²*

Amendment 1:2010

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO/IEC Guide 98-4, ISO/IEC Guide 99, and IEC 61000-6-2, as well as the following apply.

NOTE All quantities are expressed in SI units.

3.1 sound pressure

difference between an instantaneous total pressure and the corresponding static pressure

Note 1 to entry: Sound pressure is expressed in pascals (Pa).

3.2 sound pressure level

ten times the logarithm to the base 10 of the ratio of the time-mean-square of a sound-pressure signal to the square of the reference value

² CISPR = *International Special Committee on Radio Interference*.

Note 1 to entry: Sound pressure level is expressed in decibels (dB).

Note 2 to entry: The reference value is 20 µPa.

3.3 frequency weighting

difference, as a specified function of frequency, between the level of the frequency-weighted signal indicated on the display device and the corresponding level of a constant-amplitude sinusoidal input signal

Note 1 to entry: Level difference is expressed in decibels (dB).

3.4 time weighting

exponential function of time, of a specified time constant, that weights the square of a sound-pressure signal

3.5 sound level frequency-weighted sound pressure level

level with time weighting or time averaging of the square of a frequency-weighted sound-pressure signal

Note 1 to entry: Sound level is expressed in decibels (dB).

3.6 time-weighted sound level

ten times the logarithm to the base 10 of the ratio of the running time average of the time-weighted square of a frequency-weighted sound-pressure signal to the square of the reference value

Note 1 to entry: Time-weighted sound level is expressed in decibels (dB).

Note 2 to entry: For time-weighted sound level, example letter symbols are L_{AF} , L_{AS} , L_{CF} , and L_{CS} for frequency weightings A and C and time weightings F and S.


Note 3 to entry: In symbols and as an example, A-weighted and F-time-weighted sound level $L_{AF}(t)$ at observation time t can be represented by

$$L_{AF}(t) = 10 \lg \left[\frac{(1/\tau_F) \int_{-\infty}^t p_A^2(\xi) e^{-(t-\xi)/\tau_F} d\xi}{p_0^2} \right] \text{ dB} \quad (1)$$

where

- τ_F is the exponential time constant in seconds for the F time weighting;
- ξ is a dummy variable of time integration from some time in the past, as indicated by $-\infty$ for the lower limit of the integral, to the time of observation t ;
- $p_A(\xi)$ is the A-weighted instantaneous sound-pressure signal; and
- p_0 is the reference value of 20 µPa.

Note 4 to entry: The sketch in Figure 1 illustrates the process indicated by Equation (1).



IEC 2243/13

Key

- a Start with a frequency-weighted electrical input signal
- b Square the input signal
- c Apply a low-pass filter with one real pole at $-1/\tau$ (exponential time weighting)
- d Take the base-10 logarithm
- e Display the result in decibels with the square of a reference value of 20 µPa

Figure 1 – Principal steps involved in forming a time-weighted sound level

3.7

maximum time-weighted sound level

greatest time-weighted sound level within a stated time interval

Note 1 to entry: Maximum time-weighted sound level is expressed in decibels (dB).

Note 2 to entry: Example letter symbols for maximum time-weighted sound level are L_{AFmax} , L_{ASmax} , L_{CFmax} , and L_{CSmax} for frequency weightings A and C and time weightings F and S.

3.8

peak sound pressure

greatest sound pressure (positive or negative) during a stated time interval

Note 1 to entry: Peak sound pressure is expressed in pascals (Pa).

Note 2 to entry: A peak sound pressure can arise from a positive or negative instantaneous sound pressure.

3.9

peak sound level

ten times the logarithm to the base 10 of the ratio of the square of a frequency-weighted peak sound-pressure signal to the square of the reference value

Note 1 to entry: Peak sound level is expressed in decibels (dB).

Note 2 to entry: The reference value is 20 µPa.

3.10

time-averaged sound level

equivalent continuous sound level

ten times the logarithm to the base 10 of the ratio of the time average of the square of a frequency-weighted sound-pressure signal during a stated time interval to the square of the reference value

Note 1 to entry: Time-averaged or equivalent continuous sound level is expressed in decibels (dB).

Note 2 to entry: In symbols and as an example, time-averaged, A-weighted sound level $L_{Aeq,T}$ is given by

$$L_{Aeq,T} = 10 \lg \left[\frac{(1/T) \int_{t-T}^t p_A^2(\xi) d\xi}{p_0^2} \right] \text{ dB} \quad (2)$$

where

- ξ is a dummy variable of time integration over the averaging time interval ending at the time of observation t ;
- T is the averaging time interval;
- $p_A(\xi)$ is the A-weighted sound-pressure signal; and
- p_0 is the reference value of 20 µPa.

Note 3 to entry: In principle, time weighting is not involved in a determination of time-averaged sound level.

3.11

sound exposure

time integral of the square of a frequency-weighted sound-pressure signal over a stated time interval or event of stated duration

Note 1 to entry: Duration of integration is included implicitly in the time integral and is not always reported explicitly, although it is useful to state the nature of the event. For measurements of sound exposure over a specified time interval, duration of integration is usually reported and indicated by a suitable subscript to the letter symbol, for example as $E_{A,1h}$.

Note 2 to entry: In symbols and as an example, A-weighted sound exposure $E_{A,T}$ is represented by

$$E_{A,T} = \int_{t_1}^{t_2} p_A^2(t) dt \quad (3)$$

where $p_A^2(t)$ is the square of the A-weighted sound-pressure signal during integration time T starting at t_1 and ending at t_2 .

Note 3 to entry: The unit of sound exposure is pascal-squared seconds (Pa²s) if sound pressure is in pascals and running time is in seconds.

Note 4 to entry: For applications such as measurement of exposure to noise in the workplace, sound exposure in pascal-squared hours is more convenient than pascal-squared seconds.

3.12 sound exposure level

ten times the logarithm to the base 10 of the ratio of a sound exposure to the reference value

Note 1 to entry: Sound exposure level is expressed in decibels (dB).

Note 2 to entry: In symbols and as an example, A-weighted sound exposure level $L_{AE,T}$ is related to the corresponding time-averaged, A-weighted sound level $L_{Aeq,T}$ by

$$L_{AE,T} = 10 \lg \left[\frac{\int_{t_1}^{t_2} p_A^2(t) dt}{p_0^2 T_0} \right] \text{dB} = 10 \lg \left(\frac{E_{A,T}}{E_0} \right) \text{dB} = L_{Aeq,T} + 10 \lg \left(\frac{T}{T_0} \right) \text{dB} \quad (4)$$

where

- $E_{A,T}$ is the A-weighted sound exposure in pascal-squared seconds over time interval T (see Equation (3));
- E_0 is the reference value given by $p_0^2 T_0 = (20 \mu\text{Pa})^2 \times (1 \text{ s}) = 400 \times 10^{-12} \text{ Pa}^2\text{s}$;
- T is the measurement time interval, in seconds, starting at t_1 and ending at t_2 , and
- T_0 is the reference value of 1 s for sound exposure level.

Note 3 to entry: Time-averaged, A-weighted sound level $L_{Aeq,T}$ during averaging time interval T is related to the corresponding A-weighted sound exposure $E_{A,T}$, or the A-weighted sound exposure level $L_{AE,T}$, occurring within that interval by

$$E_{A,T} = p_0^2 T \left(10^{0,1 L_{Aeq,T}} \right) \quad (5)$$

or

$$L_{Aeq,T} = 10 \lg \left(\frac{E_{A,T}}{p_0^2 T} \right) \text{dB} = L_{AE,T} - 10 \lg \left(\frac{T}{T_0} \right) \text{dB} \quad (6)$$

3.13 microphone

electroacoustic transducer by which electrical signals are obtained from acoustic oscillations

[SOURCE: IEC 60050-801:1994, definition 801-26-01]

3.14 microphone reference point

point specified on, or close to, the microphone to describe the position of the microphone

Note 1 to entry: The microphone reference point can be at the centre of the diaphragm of the microphone.

3.15 reference direction

inward direction toward the microphone reference point and specified for determining the directional response and frequency weighting of a sound level meter

Note 1 to entry: The reference direction can be specified with respect to an axis of symmetry.

3.16**sound-incidence angle**

angle between the reference direction and a line between the acoustic centre of a sound source and the microphone reference point

Note 1 to entry: Sound-incidence angle is expressed in degrees.

3.17**relative directional response**

for any frequency weighting and at any frequency of incident sinusoidal sounds, and in a specified plane containing the principal axis of the microphone, sound level indicated at a given sound-incidence angle minus the sound level indicated for sound at the same frequency from the same source and incident from the reference direction

Note 1 to entry: Relative directional response is expressed in decibels.

3.18**directivity factor**

for a sound level meter, a measure of the deviation from an ideal directional response with equal sensitivity at all possible angles of sound incidence on the microphone

Note 1 to entry: Directivity factor is non-dimensional.

3.19**directivity index**

ten times the base-ten logarithm of the directivity factor

Note 1 to entry: Directivity index is expressed in decibels.

3.20**relative frequency-weighted free-field response**

for a given frequency, time-weighted or time-averaged, frequency-weighted sound level indicated by a sound level meter in response to plane progressive sinusoidal sound incident on the microphone from the reference direction minus the corresponding time-weighted or time-averaged sound level present at the position of the microphone reference point for the sound level meter and from the same sound source but in the absence of the sound level meter

Note 1 to entry: Relative frequency-weighted free-field response is expressed in decibels (dB).

Note 2 to entry: Relative frequency-weighted free-field response is called free-field sensitivity level in IEC 61183.

3.21**relative frequency-weighted random-incidence response**

for a given frequency, time-averaged, frequency-weighted sound level indicated by a sound level meter in response to random-incidence sound minus the time-averaged sound pressure level present at the position of the microphone reference point for the sound level meter and from the same sound source but in the absence of the sound level meter

Note 1 to entry: Relative frequency-weighted random-incidence response is expressed in decibels (dB).

Note 2 to entry: Relative frequency-weighted random-incidence response is called random-incidence sensitivity level in IEC 61183.

3.22**level range**

range of nominal sound levels measured for a particular setting of the controls of a sound level meter

Note 1 to entry: Level range is expressed in decibels (dB), for example, the 50 dB to 110 dB range.

3.23**reference sound pressure level**

sound pressure level specified for testing the electroacoustic performance of a sound level meter

Note 1 to entry: Reference sound pressure level is expressed in decibels (dB).

3.24**reference level range**

level range specified for testing the electroacoustic characteristics of a sound level meter and containing the reference sound pressure level

Note 1 to entry: Reference level range is expressed in decibels (dB), for example, the 50 dB to 110 dB range.

3.25**calibration check frequency**

nominal frequency of the sinusoidal sound pressure produced by a sound calibrator

3.26**level linearity deviation**

at a stated frequency, an indicated signal level minus the anticipated signal level

Note 1 to entry: Level linearity deviation is expressed in decibels (dB).

3.27**linear operating range**

on any level range and at a stated frequency, the range of sound levels over which level linearity deviations do not exceed the applicable acceptance limits specified in this standard

Note 1 to entry: Linear operating range is expressed in decibels (dB).

3.28**total range**

range of A-weighted sound levels, in response to sinusoidal signals, from the smallest sound level, on the most-sensitive level range, to the greatest sound level, on the least-sensitive level range, that can be measured without indication of overload or under-range and without exceeding the acceptance limits specified in this standard for level linearity deviation

Note 1 to entry: Total range is expressed in decibels (dB).

3.29**toneburst**

one or more complete cycles of a sinusoidal electrical signal starting and stopping at a zero crossing of the waveform

3.30**toneburst response**

maximum time-weighted sound level, or sound exposure level, measured in response to a toneburst minus the corresponding measured sound level of the steady input signal from which the toneburst was extracted

Note 1 to entry: Toneburst response is expressed in decibels (dB).

3.31**reference orientation**

orientation of a sound level meter for tests to demonstrate conformance to the specifications of this standard for emissions of, and immunity to the effects of exposure to, radio-frequency fields

3.32**coverage probability**

probability that the set of true quantity values of a measurand is contained within a specified coverage interval

[SOURCE: ISO/IEC Guide 98-4:2012, definition 3.2.8]

3.33**acceptance limit**

specified upper or lower bound of permissible measured quantity values

[SOURCE: ISO/IEC Guide 98-4:2012, definition 3.3.8]

4 Reference environmental conditions

Reference environmental conditions for specifying the electroacoustic performance of a sound level meter are:

- air temperature 23 °C;
- static pressure 101,325 kPa;
- relative humidity 50 %.

5 Performance specifications**5.1 General**

5.1.1 Generally, a sound level meter is a combination of a microphone, a preamplifier, a signal processor, and a display device. Performance specifications of this standard apply to any design for microphone and preamplifier that is appropriate for a sound level meter.

The signal processor includes the combined functions of an amplifier with a specified and controlled frequency response, a device to form the square of the frequency-weighted, time-varying sound-pressure signal, and a time integrator or time averager. Signal processing that is necessary to conform to the specifications of this standard is an integral part of a sound level meter.

In this standard, a display device provides a physical and visible display, or storage, of measurement results. Any stored measurement result shall be available for display by means of a manufacturer-specified device, for instance a computer with associated software.

5.1.2 Performance specifications of this Clause apply under the reference environmental conditions of Clause 4.

5.1.3 For specifying the maximum-permitted emission of, and immunity to the effects of exposure to, radio-frequency fields, sound level meters are classified into three groups as follows:

- group X sound level meters: self-contained instruments that include sound level measurement facilities according to this standard and which specify internal battery power for the normal mode of operation, requiring no external connections to other apparatus to measure sound levels;
- group Y sound level meters: self-contained instruments that include sound level measurement facilities according to this standard and which specify connection to a public supply of electric power for the normal mode of operation, requiring no external connections to other apparatus to measure sound levels; and

- group Z sound level meters: instruments that include sound level measurement facilities according to this standard and which require two or more items of equipment, which are essential constituent parts of the sound level meter, to be connected together by some means for the normal mode of operation. The separate items may be operated from internal batteries or from a public supply of electric power.

5.1.4 The configuration of the complete sound level meter and its normal mode of operation shall be stated in the Instruction Manual. If appropriate, the configuration of the complete sound level meter includes a windscreen and other devices that are installed around the microphone as integral components for the normal mode of operation.

5.1.5 A sound level meter that is stated in the Instruction Manual to be a class 1 or class 2 sound level meter shall conform to all relevant class 1 or class 2 specifications, respectively, that are provided in this standard. A class 2 sound level meter may provide some class 1 capabilities, but if any capability conforms only to the class 2 specifications, the instrument is a class 2 sound level meter. A sound level meter may be specified as a class 1 instrument in one configuration and a class 2 instrument in another configuration (for example, with a different microphone or preamplifier).

5.1.6 The Instruction Manual shall state the models of microphones with which the complete sound level meter conforms to the specifications for class 1 or class 2 performance for sound waves incident on the microphone from the reference direction in a free field or with random incidence, as applicable. The Instruction Manual shall describe appropriate procedures for use of the sound level meter.

5.1.7 The Instruction Manual shall state how the microphone and preamplifier are to be mounted, if applicable, to conform to the specifications for directional response and frequency weightings. An extension device or cable may be required to conform to the specifications. In this event, the sound level meter shall be stated in the Instruction Manual as conforming to the applicable specifications for directional response and frequency weighting only when the specified devices are installed.

5.1.8 Computer software may be an integral part of the sound level meter. The Instruction Manual shall describe the means by which a user can identify the version of the software that is installed to operate the functions of a sound level meter.

5.1.9 A sound level meter shall have frequency-weighting A. As a minimum, a time-weighting sound level meter shall provide a means to indicate A-frequency-weighted and F-time-weighted sound level. As a minimum, an integrating-averaging sound level meter shall provide a means to indicate A-weighted, time-averaged sound level. As a minimum, an integrating sound level meter shall provide a means to indicate A-weighted sound exposure level. Sound level meters may contain any or all of the design features for which performance specifications are given in this standard. A sound level meter shall conform to the applicable performance specifications for those design features that are provided.

If the sound level meter only indicates sound exposure level, time-averaged sound level shall be determined by application of Equation (6) for the averaging time.

5.1.10 Sound level meters conforming to class 1 acceptance limits also shall provide frequency-weighting C. Sound level meters that measure C-weighted peak sound levels shall also be able to measure C-weighted time-averaged sound levels. Frequency-weighting Z is optional. The Instruction Manual shall describe all frequency weightings that are provided.

5.1.11 A sound level meter may have more than one display device.

NOTE An analogue or digital output connection alone is not a display device.

5.1.12 A sound level meter may have more than one level range with a suitable level range control. The Instruction Manual shall (a) identify the level range(s) by the lower and upper