



BSI Standards Publication

Radiation protection instrumentation — Backpack-type radiation detector (BRD) for the detection of illicit trafficking of radioactive material

National foreword

This British Standard is the UK implementation of EN 62694:2016. It is identical to IEC 62694:2014.

The UK participation in its preparation was entrusted to Technical Committee NCE/2, Radiation protection and measurement.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2016.

Published by BSI Standards Limited 2016

ISBN 978 0 580 89966 9

ICS 13.280

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 30 November 2016.

Amendments/corrigenda issued since publication

Date	Text affected
------	---------------

EUROPEAN STANDARD
NORME EUROPÉENNE
EUROPÄISCHE NORM

EN 62694

October 2016

ICS 13.280

English Version

**Radiation protection instrumentation - Backpack-type radiation detector (BRD) for the detection of illicit trafficking of radioactive material
(IEC 62694:2014)**

Instrumentation pour la radioprotection - Détecteur de rayonnement de type sac-à-dos (BRD) pour la détection du trafic illicite des matières radioactives
(IEC 62694:2014)

Strahlenschutz-Messgeräte - Rucksack-Strahlenschutzdetektor zum Nachweis von unerlaubtem Transport radioaktiver Materialien
(IEC 62694:2014)

This European Standard was approved by CENELEC on 2016-09-05. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.



European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

© 2016 CENELEC All rights of exploitation in any form and by any means reserved worldwide for CENELEC Members.

European foreword

This document (EN 62694:2016) consists of the text of IEC 62694:2014 prepared by SC 45B “Radiation protection instrumentation” of IEC/TC 45 “Nuclear instrumentation”.

The following dates are fixed:

- latest date by which this document has to be implemented (dop) 2017-09-05
at national level by publication of an identical national standard or by endorsement
- latest date by which the national standards conflicting with this document have to be withdrawn (dow) 2019-09-05

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC [and/or CEN] shall not be held responsible for identifying any or all such patent rights.

Annex ZA (normative)

Normative references to international publications with their corresponding European publications

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE 1 When an International Publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

NOTE 2 Up-to-date information on the latest versions of the European Standards listed in this annex is available here: www.cenelec.eu

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
IEC 60050	Series	International Electrotechnical Vocabulary -	-	-
IEC 60050-393	2003	International Electrotechnical Vocabulary - Part 393: Nuclear instrumentation - Physical phenomena and basic concepts	-	-
IEC 60050-394	2007	International Electrotechnical Vocabulary - Part 394: Nuclear instrumentation - Instruments, systems, equipment and detectors	-	-
IEC 62706	-	Radiation protection instrumentation - Environmental, electromagnetic and mechanical performance requirements	-	-
IEC 62755	-	Radiation protection instrumentation - Data - format for radiation instruments used in the detection of illicit trafficking of radioactive materials	-	-

CONTENTS

FOREWORD.....	6
1 Scope.....	8
2 Normative references	8
3 Terms and definitions, abbreviations, quantities and units	8
3.1 Terms and definitions.....	8
3.2 Abbreviations.....	10
3.3 Quantities and units	11
4 General test procedure	11
4.1 Nature of test.....	11
4.2 Standard test conditions	11
4.3 Tests performed under standard test conditions	11
4.4 Test performed with variation of influence quantities	11
4.5 Statistical fluctuations	11
4.6 Uncertainties in the measurements	12
4.7 Background radiation during testing	12
4.8 BRD set up	12
4.9 Speed of moving sources and integration time for radionuclide identification	13
4.10 Radiation sources	13
4.11 Functionality tests	14
5 General requirements	15
5.1 Mass.....	15
5.1.1 Requirements	15
5.1.2 Method of test.....	16
5.2 Design requirements	16
5.2.1 Requirements	16
5.2.2 Method of test.....	16
5.3 Marking.....	16
5.3.1 Requirements	16
5.3.2 Method of test.....	16
5.4 Switches	16
5.4.1 Requirements	16
5.4.2 Method of test.....	16
5.5 Effective range of measurement – Energy.....	17
5.5.1 Requirements	17
5.5.2 Method of test.....	17
5.6 Effective range of measurement – Count rate	17
5.6.1 Requirements	17
5.6.2 Method of test.....	17
5.7 Operating parameters	17
5.7.1 Requirements	17
5.7.2 Method of test.....	17
5.8 Explosive atmospheres	17
5.8.1 Requirements	17
5.8.2 Method of test.....	18
5.9 Diagnostics	18
5.9.1 Requirements	18

	5.9.2	Method of test.....	18
5.10		Power supply	18
	5.10.1	Requirements	18
	5.10.2	Method of test.....	18
5.11		Data format.....	19
	5.11.1	Requirements	19
	5.11.2	Method of test.....	20
5.12		Data storage	21
	5.12.1	Requirements	21
	5.12.2	Method of test.....	21
5.13		Communication interface.....	21
	5.13.1	Requirements	21
	5.13.2	Method of test.....	21
5.14		User interface	21
	5.14.1	Display	21
	5.14.2	Basic indications.....	22
	5.14.3	Additional indications.....	22
	5.14.4	Indications for BRDs with radionuclide identification capabilities	23
	5.14.5	Indications for BRDs with radionuclide directionality capabilities	23
	5.14.6	Basic functions and controls	23
	5.14.7	Restricted functions and controls	24
6		Radiation detection requirements	24
6.1		False alarm test.....	24
	6.1.1	Requirements	24
	6.1.2	Method of test.....	24
6.2		Alarm response to photon radiation.....	25
	6.2.1	Requirements	25
	6.2.2	Method of test.....	25
6.3		Alarm response to neutron radiation	26
	6.3.1	Requirements	26
	6.3.2	Method of test.....	26
6.4		Personal radiation protection alarm and response time	27
	6.4.1	Requirements	27
	6.4.2	Method of test.....	27
6.5		Gamma-ray ambient dose equivalent rate indication	28
	6.5.1	Requirements	28
	6.5.2	Method of test.....	28
6.6		Angular dependence and verification of directional indication.....	28
	6.6.1	Requirements	28
	6.6.2	Method of test.....	28
6.7		Over range test.....	29
	6.7.1	Requirements	29
	6.7.2	Method of test.....	29
6.8		Neutron indication in the presence of photons.....	30
	6.8.1	Requirements	30
	6.8.2	Method of test.....	30
6.9		Detection of gradually increasing radiation levels.....	31

6.9.1	Requirements	31
6.9.2	Method of test.....	31
6.10	Networked area monitors	31
6.10.1	Requirements	31
6.10.2	Method of test.....	32
6.11	Radionuclide identification, when provided.....	32
6.11.1	General Requirements	32
6.11.2	Radionuclide identification library	33
6.11.3	Single radionuclide identification	33
6.11.4	Identification of shielded radionuclides	35
6.11.5	Simultaneous and masked radionuclide identification	35
6.11.6	Radionuclide not in library	36
6.11.7	Low-exposure rate identification	37
6.11.8	Over range characteristics for identification	37
6.11.9	Rejection of natural background variations	38
7	Environmental requirements	39
8	Mechanical requirements.....	39
9	Electromagnetic requirements	40
10	Documentation	40
10.1	Type test report	40
10.2	Certificate	40
10.3	Operation and maintenance manual.....	40
Annex A (informative)	Statistical considerations.....	46
A.1	Poisson distribution.....	46
A.2	Confidence intervals for Poisson distribution.....	46
A.3	False alarm testing	46
A.4	Binomial distribution.....	48
Annex B (informative)	List of expected progeny and expected impurities	50
Annex C (informative)	Summary of fluence rate calculations	52
Annex D (normative)	Calculation ambient dose equivalent rate	54
Bibliography.....		59
Figure 1	Diagram of testing angles when source passes at an angle of 0° in the horizontal plane (top view). The displayed source movement represents the test configuration at an angle of 0°.	44
Figure 2	Diagram of the two orthogonal planes (horizontal and vertical planes), the BRD reference point and testing angles	45
Figure 3	BRD setup and testing source positions for network area monitoring.....	45
Table 1	Standard test conditions.....	42
Table 2	Occurrence of functionality tests for environmental testing	42
Table 3	Occurrence of functionality tests for mechanical testing	43
Table 4	Emission frequency range	43
Table 5	Occurrence of functionality tests for electromagnetic testing	44
Table A.1	One-sided 95 % upper confidence bounds for the false alarm rate for a given number of false alarms observed over a given time period	47

Table A.2 – Necessary sample sizes (n) for different levels (p_o) and number of failures (k).....	49
Table B.1 – List of expected progeny and expected impurities	51
Table C.1 – Examples of fluence rate calculations	53
Table D.1 – Conversion coefficient $h^*_K(10)$ from air kerma, K, to ambient dose equivalent, $H^*(10)$, for mono-energetic and parallel photon beams.....	55
Table D.2 – Probabilities per disintegration for ^{232}Th and ^{226}Ra (in equilibrium) as a function of photon energy	56
Table D.3 – Values of the mass energy-transfer, mass energy-absorption, and mass attenuation coefficients for air	58

INTERNATIONAL ELECTROTECHNICAL COMMISSION

RADIATION PROTECTION INSTRUMENTATION – BACKPACK-TYPE RADIATION DETECTOR (BRD) FOR THE DETECTION OF ILLICIT TRAFFICKING OF RADIOACTIVE MATERIAL

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard 62694 has been prepared by subcommittee 45B: Radiation protection instrumentation, of IEC technical committee 45: Nuclear instrumentation.

The text of this standard is based on the following documents:

FDIS	Report on voting
45B/781/FDIS	45B/790/RVD

Full information on the voting for the approval of this standard can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

The committee has decided that the contents of this publication will remain unchanged until the stability date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

IMPORTANT – The 'colour inside' logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.

RADIATION PROTECTION INSTRUMENTATION – BACKPACK-TYPE RADIATION DETECTOR (BRD) FOR THE DETECTION OF ILLICIT TRAFFICKING OF RADIOACTIVE MATERIAL

1 Scope

This International Standard applies to backpack-type radiation detectors (BRDs) that are used for the detection of illicit trafficking of radioactive material. This standard establishes the operational and performance requirements for BRDs. BRDs are portable instruments designed to be worn during use. They may also be used as temporary area monitors in a stand-alone mode.

BRDs detect gamma radiation and may include neutron detection and/or the identification of gamma-ray emitting radionuclides. This standard establishes performance and testing requirements associated with radiation measurements and the expected electrical, mechanical, and environmental conditions while in use.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60050 (all parts): *International Electrotechnical Vocabulary* (available at <http://www.electropedia.org>)

IEC 60050-393:2003, *International Electrotechnical Vocabulary – Part 393: Nuclear instrumentation – Physical phenomena and basic concepts*

IEC 60050-394:2007, *International Electrotechnical Vocabulary – Part 394: Nuclear instrumentation – Instruments, systems, equipment and detectors*

IEC 62706, *Radiation protection instrumentation – Environmental, electromagnetic and mechanical performance requirements*

IEC 62755, *Radiation protection instrumentation – Data format for radiation instruments used in the detection of illicit trafficking of radioactive materials*

3 Terms and definitions, abbreviations, quantities and units

3.1 Terms and definitions

For the purposes of this document, the terms and definitions given in IEC 60050-393 and IEC 60050-394 apply, as well as the following.

3.1.1

accuracy

closeness of the agreement between the result of a measurement and the conventionally true value of the measurand

3.1.2**alarm**

audible, visual, or other signal activated when the instrument reading exceeds a preset value, falls outside of a preset range, or when the instrument detects the presence of the source of radiation according to a preset condition

[SOURCE: IEC 60050-393:2003, 393-18-03, modified]

3.1.3**background level**

radiation field in which there are no external sources present other than those in the natural background at the location of the measurements

3.1.4**backpack-type radiation detector**

instrument composed of several radiation detection components that are placed inside a backpack or other similar enclosure with an external user interface or control device

3.1.5**centre line**

horizontal or vertical line that describes the geometrical centre of an object

3.1.6**coefficient of variation**

V

ratio of the standard deviation s to the arithmetic mean $\bar{x}$ of a set of n measurements x_i given by the following formula:

$$V = \frac{s}{\bar{x}} = \frac{1}{\bar{x}} \sqrt{\frac{1}{n-1} \sum_{i=1}^n (x_i - \bar{x})^2}$$

[SOURCE: IEC 60050-394:2007, 394-40-14]

3.1.7**energy window**

part of the energy spectrum within an upper and lower energy limit

[SOURCE: IEC 60050-394:2007, 394-38-70]

3.1.8**keyhole markup language****KML**

is a file format used to display geographic data

Note 1 to entry: For example, see <http://www.opengeospatial.org/standards/kml/>.

3.1.9**fluence**

Φ

the quotient of dN by da , where dN is the number of particles incident on a sphere of cross-sectional area da : $\Phi = dN/da$

[SOURCE: IEC 60050-881:1983, 881-04-18]

3.1.10**fluence rate**

the *fluence rate*, $\dot{\phi}$, is the quotient of $d\Phi$ by dt , where $d\Phi$ is the increment of the fluence in the time interval dt , thus $\dot{\phi} = \frac{d\phi}{dt}$. The unit of fluence rate is $\text{m}^{-2}\text{s}^{-1}$

[SOURCE: ICRU Report 60:1998]

3.1.11**type test**

conformity test made on one or more items representative of the production

[SOURCE: IEC 60050-394:2007, 394-40-02]

3.1.12**user interface**

software and/or hardware that manages interactions between a user and equipment

3.1.13**variance**

σ^2

measure of dispersion, which is the sum of the squared deviation of observations from their mean divided by one less than the number of observations

$$\sigma^2 = \frac{1}{n-1} \sum_{i=1}^n (x_i - \bar{x})^2$$

3.2 Abbreviations

AC	alternating current
BRD	backpack-type radiation detector
cps	counts per second
DC	direct current
DU	depleted uranium
ESD	electrostatic discharge
FIFO	first in first out
GPS	global positioning system
HDPE	high density polyethylene
HEU	highly enriched uranium
HPGe	high purity germanium
KML	keyhole markup language
NORM	naturally occurring radioactive material
PMMA	polymethyl methacrylate
RGPu	reactor grade plutonium
WGPu	weapons grade plutonium
XML	eXtensible Markup Language

3.3 Quantities and units

In the present standard, units of the International System (SI) are used¹. The definitions of radiation quantities are given in IEC 60050-393 and IEC 60050-394.

Nevertheless, the following units may also be used:

- for energy: electron-volt (symbol: eV), $1 \text{ eV} = 1,602 \times 10^{-19} \text{ J}$;
- for time: years (symbol: y), days (symbol: d), hours (symbol: h), minutes (symbol: min);
- for distance: centimetre (symbol: cm), millimetre (symbol: mm), kilometre (symbol: km);

Multiples and submultiples of SI units will be used, when practicable, according to the SI system.

4 General test procedure

4.1 Nature of test

The tests in this standard are to be considered type tests, unless otherwise stated.

4.2 Standard test conditions

Except where otherwise specified, the tests in this standard shall be performed under the standard test conditions given in Table 1.

4.3 Tests performed under standard test conditions

For these tests, the value of temperature, pressure, relative humidity and gamma and neutron background at the time of the test shall be recorded. Values should be within the standard test conditions given in Table 1.

4.4 Test performed with variation of influence quantities

For those tests intended to determine the effects of variations in an influence quantity (e.g., temperature, humidity), all other influence quantities should be maintained at the standard test conditions given in Table 1 unless otherwise specified in the applicable test method.

4.5 Statistical fluctuations

For tests involving the use of radioactive sources to verify susceptibility to an environmental, electromagnetic, or mechanical condition the ambient dose equivalent rate produced by the sources to verify the BRD response shall be adjusted to reduce the magnitude of the statistical fluctuations.

If the magnitude of the statistical fluctuations of the BRD indication arising from the random nature of radiation alone is a significant fraction of the variation of the indication permitted in the test, then the ambient dose equivalent rate should be increased to ensure that the mean value of such readings may be estimated with sufficient accuracy to demonstrate compliance with the test in question.

It is recommended that the coefficient of variation (V , expressed in percentage) for each nominal mean reading be less than or equal to 12 %. For neutron or photon background measurements, attaining a coefficient of variation to meet this requirement may not be possible. Therefore, testing with neutrons or photons at background levels (i.e., testing

¹ International Bureau of Weights and Measures: The International System of Units, 8th edition, 2006

without radioactive source present) can be performed even when the coefficient of variation is larger than 12 %.

12 % is from statistical analysis techniques for dosimeter testing and has proven to be a simple way of determining when a group of readings are acceptable for compliance testing. The time interval between readings needs to be sufficiently long (i.e., larger than the integration time of the instrument) to ensure that the readings are statistically independent.

4.6 Uncertainties in the measurements

Unless otherwise stated for a specific quantity, the uncertainties for any measurable quantity (e.g., radiation field) should not exceed 15 % with a coverage factor of $k = 1$.

4.7 Background radiation during testing

Testing shall be performed in an area with a nominal natural radiation background that has only natural variation as defined in Table 1.

The gamma-ray background intensity shall be measured using a pressurized ion chamber or similar environmental radiation measurement device that is calibrated to provide the gamma-ray ambient dose equivalent rate, $\dot{H}^*(10)$. When testing spectrometric BRDs the gamma-ray background shall be characterized using a high resolution gamma-ray spectrometer (e.g., high purity germanium (HPGe) detector). The measured spectra shall be recorded. If the BRD is equipped with neutron detectors, the neutron background should be the natural background and should not be artificially modified during testing. The neutron background at the test location shall be measured and recorded.

The evaluation of the BRD shall be performed without the benefit of any radiation shielding against the natural background, except for that shielding that is part of the instrument.

4.8 BRD set up

The BRD shall be set up based on the manufacturer's specifications including background update mode, if applicable. Once set up for testing, no changes shall be made that could affect the overall response of the BRD. If more than one background update mode is available, testing should be performed in all modes when indicated in the specific clauses under the radiological tests.

When performing the radiological tests in Clause 6, the BRD shall be configured and oriented as it would be used. This may be achieved by using a phantom that would represent the human upper torso. The phantom shall be made of polymethyl methacrylate (PMMA). The phantom dimensions shall be 40 cm wide, 60 cm high and 15 cm thick.

The BRD shall be mounted on a stand or fixture made out of a material that does not have a large hydrogen content (e.g., foam, plastic). It is recommended to use materials such as aluminium for mounting the BRD to prevent possible additional moderation of the neutron source.

The reference point of the BRD should be marked by the manufacturer. If marking is not provided by the manufacturer, the reference point is defined as the imaginary point where the three mutually orthogonal lines that go through the center of the length, width and thickness of the BRD intersect (see Figure 2).

For static and dynamic tests described in Clause 6, the reference point of the BRD shall be positioned 1,5 m from the floor or ground surface. The centreline of the source shall be at the same height as the reference point of the BRD, 1,5 m from the floor or ground surface.

For static tests, the distance between the source and the centreline of the BRD shall be between 1 m and 3 m unless otherwise stated.

For dynamic tests, the line of source movement and detector centreline shall be kept parallel, the distance of closest approach between the source and the reference point of the BRD shall be between 1 m and 3 m unless otherwise stated, see Figure 1.

The phantom is not used when the BRD is evaluated for use as a stand-alone area radiation monitor. Testing as an area radiation monitor is performed if such claim is made by the manufacturer.

When performing the tests in Clauses 7, 8, and 9, the BRD shall not be mounted on a phantom. The BRD-to-source distance and the relative orientation and position between the BRD and the radiation source shall be adjusted to reduce the statistical fluctuations as discussed in 4.5. The testing distance, orientation and position of the BRD with respect to source shall be recorded for these tests. Due to the nature of the tests, there is no need for this standard to specify the BRD-to-source distance, and relative orientation and position between the BRD and the source.

4.9 Speed of moving sources and integration time for radionuclide identification

For static tests, the integration time required to perform a radionuclide identification shall be as specified by the manufacturer or a maximum of 1 min (whichever is the shortest).

During the static tests, the source shall be removed and placed back in the same location between trials. There shall be a 10 s minimum delay between each trial with the source either positioned at a distance where it does not affect the background surrounding the BRD or shielded during the delay.

For dynamic tests, the source or BRD shall be moved in a configuration that provides no shielding around the source other than that required for the specific test. The source speed shall be $1,2 \text{ m} \cdot \text{s}^{-1}$ (average walking speed) when tested at a distance of closest approach of 1,5 m, unless otherwise required in a test. If the distance of closest approach, d (expressed in m), is adjusted within 1 m and 3 m then the passage speed, v (expressed in $\text{m} \cdot \text{s}^{-1}$), shall be adjusted to $v = v_0 \times d/d_0$,

where $v_0 = 1,2 \text{ m} \cdot \text{s}^{-1}$ and $d_0 = 1,5 \text{ m}$.

During the dynamic tests, there shall be a 10 s minimum delay between each trial with the source either positioned at a distance where it does not affect the background surrounding the BRD or shielded during the delay.

NOTE For all dynamic tests, the source or the BRD can be moved relative to each other.

4.10 Radiation sources

Unless otherwise stated, tests involving the use of gamma radiation shall be carried out using ^{137}Cs for gross count measurements and ^{241}Am together with ^{60}Co for radionuclide identification (see Table 1).

The reference source for neutron radiation is ^{252}Cf . The neutron emission rate of the ^{252}Cf source shall be $20\,000 \text{ s}^{-1}$ ($\pm 20\%$) (see Table 1). The unmoderated reference neutron source shall be encapsulated in 1 cm of steel and shielded with 0,5 cm-thick lead in order to attenuate the possible gamma-ray emission from the ^{252}Cf source. The lead shall be placed outside the steel encapsulation. The moderation of the ^{252}Cf is achieved by surrounding the source in the presence of the 1 cm of steel encapsulation and 0,5 cm-thick lead shielding with 4 cm-thick high density polyethylene (HDPE) container (e.g., sphere, cylinder, box).

The sources shall be mounted on a stand or fixture made out of a material that does not have a large hydrogen content (e.g., foam, plastic). It is recommended to use materials such as aluminium for mounting the sources to prevent possible additional scattering and moderation of the neutron source.

The isotopic composition and activity of different naturally occurring radioactive materials (NORM), such as zircon, monazite and allanite, vary widely from sample to sample. Therefore, point sources are used to ensure greater consistency and traceability in performing measurements at different locations and at different times. The approximation of bulk NORM sources is done by surrounding ^{226}Ra and ^{232}Th sources (in equilibrium with their progeny) with 9 cm of polymethyl methacrylate (PMMA) each producing the same ambient dose equivalent rate, see Annex D.

As the ^{241}Am content in plutonium sources varies widely, when testing with WGPu the emission rate of the 60 keV line from ^{241}Am shall be no more than 10 times the emission rate of the 414 keV line for ^{239}Pu (e.g., if the emission rate for ^{239}Pu is 100 s^{-1} then the emission rate for ^{241}Am shall not exceed $1\,000\text{ s}^{-1}$), see Annex C. When needed, copper or cadmium should be used to reduce the contribution from ^{241}Am . The copper or cadmium shielded WGPu source shall be considered the bare WGPu source to be used in all radiological tests that require the use of WGPu.

All radioactive sources used for testing shall be checked for gamma-ray emitting impurities and isotopic composition, using a high purity germanium detector (HPGe), in order to determine the expected correct radionuclide identification by the BRD.

The fluence rate and ambient dose equivalent rate shall be determined using a photon cut-off energy of 40 keV.

When testing with point sources producing an ambient dose equivalent rate close to or below background levels (i.e., $0,05\text{ }\mu\text{Sv}\cdot\text{h}^{-1}$), the ambient dose equivalent rate shall be calculated as described in Annex D. If parallel beams (i.e., irradiator) are used, the testing laboratory shall describe and record the method used to calculate the ambient dose equivalent rate.

4.11 Functionality tests

For most tests in Clauses 7, 8, and 9 the BRD functionality is evaluated prior to and after the test and in some cases during the test as well. The BRD response after the test (post-test measurement) is compared with that measured prior to the test (pre-test measurement). Therefore, the source location relative to the BRD shall be the same for both the pre-test and post-test.

For BRDs without gamma-ray spectrometry capabilities, the gamma-ray measurements are performed using a ^{137}Cs source, producing an ambient dose equivalent rate of at least $0,2\text{ }\mu\text{Sv}\cdot\text{h}^{-1} \pm 20\%$ ($k = 1$) at the reference point of the BRD, unless otherwise stated in a test.

For BRDs with gamma-ray spectrometry capabilities, the gamma-ray measurements are performed using ^{241}Am and ^{60}Co sources placed simultaneously in front of the BRD, each source producing an ambient dose equivalent rate of at least $0,1\text{ }\mu\text{Sv}\cdot\text{h}^{-1} \pm 20\%$ ($k = 1$) at the reference point of the BRD, unless otherwise stated in a test.

The neutron measurements (if applicable) are performed using the moderated ^{252}Cf source with an emission rate of $20\,000\text{ s}^{-1}$ listed in 4.10 unless otherwise stated in a test.

Statistical fluctuations should be considered during these tests, see 4.5.

The pre-test and post-test measurements shall be carried out as follows:

Pre-test:

- Record 10 independent gamma count rate or ambient dose equivalent rate readings (if the BRD has a unit-less display record the displayed level).
- Verify that the gamma alarm is activated.

- c) If the BRD is equipped with a neutron sensor, verify that the system indicates the presence of a neutron source and record 10 independent neutron count rate readings.
- d) Verify that the neutron alarm is activated.
- e) Calculate the mean and standard deviation of the count rate or ambient dose equivalent rate readings.
- f) If the BRD has radionuclide identification capabilities, collect three spectra using an integration time of 1 min or that stated by the manufacturer (whichever is the shortest, see 4.9) and record the radionuclides identified.

Post-test:

- g) Place the sources in the same location relative to the BRD as for the pre-test.
- h) Record the same information as in the pre-test steps a) through f).
- i) In addition calculate the difference (expressed in percentage) between the post-test measured mean count rate or ambient dose equivalent rate readings and the corresponding values measured in the pre-test.

Acceptance criteria:

The results are acceptable if changes in response between the pre-test and post-test do not exceed:

- ± 15 % relative to the gamma readings pre-test values.
- ± 15 % relative to the neutron readings pre-test values.

For instruments with a unit-less display, the manufacturer shall provide a table to convert the unit-less reading into a gamma and/or neutron response (i.e., count rate, ambient dose equivalent rate). For instruments with a unit-less display, the post-test gamma and neutron response shall not have changed from the pre-test response by more than ± 1 unit.

The complete and correct results for identification at each post-test shall be the same or better than the identification results obtained at the pre-test. For example, if during a pre-test there are complete and correct results in two out of three trials, then the complete and correct results at each post-test point shall be two or more.

Depending on the environmental, mechanical or electromagnetic test, the BRD response during the test can be verified in two different ways, with or without radioactive sources present. When tested using radioactive sources, the BRD response during the test is performed in the same manner as the post-test. When tested without radioactive sources, the BRD response during testing is observed as not to produce any alarm, radionuclide identifications or spurious indication as a result of the environmental, mechanical or electromagnetic test.

NOTE When performing sequential tests (e.g., temperature followed by humidity followed by moisture), trends in the pre-test measurements may indicate degradation in the BRD performance (if source-to-detector distances and geometries are kept the same throughout the tests).

5 General requirements

5.1 Mass

5.1.1 Requirements

The BRD without gamma-ray spectrometry capabilities should weigh less than 10 kg including batteries and external user interface.

The manufacturer shall provide the weight for gamma-ray spectrometry type BRDs.

5.1.2 Method of test

Weigh the BRD together with the external user interface and record the result of the measurement. Review the manufacturer documentation and verify that the weight of the BRD and the external user interface are provided. The results of the verification shall be recorded.

5.2 Design requirements

5.2.1 Requirements

The BRD should be designed to:

- a) Distribute weight, as much as practical, with the heavier side being closer to the wearer's back and the weight distributed as evenly as possible from left to right for good balance,
- b) Be easy to wear, remove, and have adjustable openings around arms, and
- c) Look like a normal backpack to the maximum extent practical.

5.2.2 Method of test

The BRD shall be inspected and the results of the inspection shall be recorded.

5.3 Marking

5.3.1 Requirements

All external controls, displays, and adjustments shall be identifiable according to their functions. External markings on the controls, displays and adjustments shall be easily readable and remain fixed after normal decontamination procedures (e.g., water and mild, non-abrasive detergent). Internal controls needed for operation shall be identified through markings and identification in technical manuals.

5.3.2 Method of test

The BRD as used shall be inspected and the results of the inspection shall be recorded. The decontamination procedure is performed by cleaning the instrument with mild soap and a damp towel and verifying that no marking is removed. Review the technical manuals to identify the internal controls markings.

5.4 Switches

5.4.1 Requirements

Controls and adjustments that affect calibration and alarm settings shall be designed so that access to them is limited to authorized personnel.

Switches and other controls should be designed to ensure that the BRD can be operated properly when the user is wearing gloves while minimizing accidental switch operation.

5.4.2 Method of test

Verify by inspection of the BRD or manufacturer provided documentation that the controls and adjustments that affect calibration and alarm settings are designed so that access to them is limited to authorized personnel.

Three users shall individually operate the BRD, including switching modes and/or performing functions while wearing gloves. Each user shall record whether multiple switches are activated while operating the BRD. For test purposes, insulated gloves (i.e., cold weather protection type gloves) should be used to verify the requirement.