

BS EN 80601-2-30:2010+A1:2015

Incorporating corrigendum November 2010



BSI Standards Publication

Medical electrical equipment

Part 2-30: Particular requirements
for the basic safety and essential
performance of automated non-invasive
sphygmomanometers

bsi.

...making excellence a habit.
AVA **Documents**
Document Sharing Hub

National foreword

This British Standard is the UK implementation of EN 80601-2-30:2010+A1:2015. It is identical to IEC 80601-2-30:2009, incorporating corrigendum January 2010 and amendment 1:2013. It supersedes BS EN 80601-2-30:2010 which will be withdrawn on 14 April 2018.

The start and finish of text introduced or altered by corrigendum is indicated in the text by tags. Text altered by IEC corrigendum January 2010 is indicated in the text by AC AC.

The start and finish of text introduced or altered by amendment is indicated in the text by tags. Tags indicating changes to IEC text carry the number of the IEC amendment. For example, text altered by IEC amendment 1 is indicated by A1 A1.

The UK participation in its preparation was entrusted by Technical Committee CH/62, Electrical Equipment in Medical Practice, to Subcommittee CH/62/4, Electromedical equipment.

A list of organizations represented on this subcommittee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2015.

Published by BSI Standards Limited 2015

ISBN 978 0 580 73897 5

ICS 11.040.55

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 30 November 2010.

Amendments/corrigenda issued since publication

Date	Text affected
30 November 2010	Correction to supersession details
31 July 2015	Implementation of IEC amendment 1:2013 with CENELEC endorsement A1:2015. Annex ZA updated

**Medical electrical equipment -
Part 2-30: Particular requirements for the basic safety and essential
performance of automated non-invasive sphygmomanometers
(IEC 80601-2-30:2009 + corrigendum Jan. 2010)**

Appareils électromédicaux -
Partie 2-30: Exigences particulières
pour la sécurité de base
et les performances essentielles
de sphygmomanomètres non invasifs
automatiques
(CEI 80601-2-30:2009 + corrigendum Jan.
2010)

Medizinische elektrische Geräte -
Teil 2-30: Besondere Festlegungen
für die Sicherheit einschließlich
der wesentlichen Leistungsmerkmale
von nicht-invasiven Sphygmomanometern
von automatisierten Typ
(IEC 80601-2-30:2009 + corrigendum Jan.
2010)

This European Standard was approved by CENELEC on 2010-09-01. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the Central Secretariat has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and the United Kingdom.

CENELEC

European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

Management Centre: Avenue Marnix 17, B - 1000 Brussels

Foreword

The text of document 62D/721/FDIS, future edition 1 of IEC 80601-2-30, prepared by SC 62D, Electromedical equipment, of IEC TC 62, Electrical equipment in medical practice, and SC 3, Lung ventilators and related equipment, of ISO TC 121, Anaesthetic and respiratory equipment, was submitted to the IEC-CENELEC parallel vote and was approved by CENELEC as EN 80601-2-30 on 2010-09-01.

This European Standard supersedes EN 60601-2-30:2000.

EN 80601-2-30:2010 constitutes a major technical revision as well as an alignment with EN 60601-1:2006. Specific technical changes include: expansion of the scope to include all AUTOMATED SPHYGMOMANOMETERS including those where the PATIENT is the OPERATOR, identification of ESSENTIAL PERFORMANCE, new clinical accuracy requirements, additional mechanical strength requirements and prohibition of OPERATOR accessible 'Luer' connectors in the PNEUMATIC SYSTEM.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN and CENELEC shall not be held responsible for identifying any or all such patent rights.

The following dates were fixed:

- latest date by which the EN has to be implemented
at national level by publication of an identical
national standard or by endorsement (dop) 2011-06-01
- latest date by which the national standards conflicting
with the EN have to be withdrawn (dow) 2013-09-01

This European Standard has been prepared under a mandate given to CENELEC by the European Commission and the European Free Trade Association and covers essential requirements of EC Directive MDD (93/42/EEC). See Annex ZZ.

In this standard, the following print types are used:

- Requirements and definitions: roman type.
- *Test specifications: italic type.*
- Informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this standard, the term

- “clause” means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this standard are preceded by the term “Clause” followed by the clause number. References to subclauses within this particular standard are by number only.

In this standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

Annexes ZA and ZZ have been added by CENELEC.

Endorsement notice

The text of the International Standard IEC 80601-2-30:2009 was approved by CENELEC as a European Standard without any modification.

In the official version, for Bibliography, the following notes have to be added for the standards indicated:

- | | | |
|--------------------|------|---|
| [1] ISO 9919:2005 | NOTE | Harmonized as EN ISO 9919:2005 (not modified). |
| [3] ISO 21647:2004 | NOTE | Harmonized as EN ISO 21647:2004 (not modified). |

Foreword to amendment A1

The text of document 62D/1072/FDIS, future IEC 80601-2-30:2009/A1, prepared by SC 62D "Electrical equipment" of IEC/TC 62 "Electrical equipment in medical practice" was submitted to the IEC-CENELEC parallel vote and approved by CENELEC as EN 80601-2-30:2010/A1:2015.

The following dates are fixed:

- | | | |
|--|-------|------------|
| • latest date by which the document has to be implemented at national level by publication of an identical national standard or by endorsement | (dop) | 2016-01-14 |
| • latest date by which the national standards conflicting with the document have to be withdrawn | (dow) | 2018-04-14 |

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC [and/or CEN] shall not be held responsible for identifying any or all such patent rights.

This document has been prepared under a mandate given to CENELEC by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive.

For the relationship with EU Directive 93/42/EEC, see informative Annex ZZ, included in EN 80601-2-30:2010.

Endorsement notice

The text of the International Standard IEC 80601-2-30:2009/A1:2013 was approved by CENELEC as a European Standard without any modification.

Annex ZA (normative)

Normative references to international publications with their corresponding European publications

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE 1 When an International Publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

NOTE 2 Up-to-date information on the latest versions of the European Standards listed in this annex is available here: www.cenelec.eu

Annex ZA of EN 60601-1:2006 applies, except as follows:

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
Replace the reference to IEC 60601-1-2 by:				
IEC 60601-1-2 (mod)	2007	Medical electrical equipment - Part 1-2: General requirements for basic safety and essential performance - Collateral standard: Electromagnetic compatibility - Requirements and tests	EN 60601-1-2	2007
IEC 60601-1-6	2010	Medical electrical equipment - Part 1-6: General requirements for basic safety and essential performance - Collateral standard: Usability	EN 60601-1-6	2010
+A1	2013		+A1	2015
IEC 60601-1-8	2006	Medical electrical equipment - Part 1-8: General requirements for basic safety and essential performance - Collateral Standard: General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems	EN 60601-1-8 + corr. March	2007 2010
+A1	2012		+A1 +A1/AC	2013 2014
Addition:				
IEC 60068-2-27	2008	Environmental testing - Part 2-27: Tests - Test Ea and guidance: Shock	EN 60068-2-27	2009
IEC 60068-2-31	2008	Environmental testing - Part 2-31: Tests - Test Ec: Rough handling shocks, primarily for equipment-type specimens	EN 60068-2-31	2008
IEC 60068-2-64	2008	Environmental testing - Part 2-64: Tests - Test Fh: Vibration, broadband random and guidance	EN 60068-2-64	2008
IEC 60601-1-11	2010	Medical electrical equipment - Part 1-11: General requirements for basic safety and essential performance - Collateral standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment	EN 60601-1-11	2010

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
IEC 60601-2-2	2009	Medical electrical equipment - Part 2-2: Particular requirements for basic safety and essential performance of high frequency surgical equipment and high frequency surgical accessories	EN 60601-2-2	2009
IEC 62366	2007	Medical devices - Application of usability engineering to medical devices	EN 62366	2008
ISO 594-1	1986	Conical fittings with a 6 % (Luer) taper for syringes, needles and certain other medical equipment - Part 1: General requirements	EN 20594-1	1993
ISO 594-2	1991	Conical fittings with a 6 % (Luer) taper for syringes, needles and certain other medical equipment - Part 2: Lock fittings	-	-
ISO 81060-2	2013	Non-invasive sphygmomanometers - Part 2: Clinical investigation of automated measurement type	-	-

Annex ZZ (informative)

Coverage of Essential Requirements of EC Directives

This European Standard has been prepared under a mandate given to CENELEC by the European Commission and the European Free Trade Association and within its scope the standard covers all relevant essential requirements as given in Annex I of the EC Directive 93/42/EEC with the exception of ERs 3, 4, 7.1 and 12.1.

Compliance with this standard provides one means of conformity with the specified essential requirements of the Directive concerned.

WARNING: Other requirements and other EC Directives may be applicable to the products falling within the scope of this standard.

CONTENTS

FOREWORD.....	2
INTRODUCTION.....	10
201.1 Scope, object and related standards	11
201.2 Normative references	13
201.3 Terms and definitions.....	14
201.4 General requirements	16
201.5 General requirements for testing ME EQUIPMENT	16
201.6 Classification of ME EQUIPMENT and ME SYSTEMS.....	17
201.7 ME EQUIPMENT identification, marking and documents	17
201.8 Protection against electrical HAZARDS from ME EQUIPMENT.....	20
201.9 Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS	21
201.10 Protection against unwanted and excessive radiation HAZARDS	21
201.11 Protection against excessive temperatures and other HAZARDS	21
201.12 Accuracy of controls and instruments and protection against hazardous outputs.....	22
201.13 HAZARDOUS SITUATIONS and fault conditions	26
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS)	27
201.15 Construction of ME EQUIPMENT	27
201.16 ME SYSTEMS	29
201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS	29
201.101 Requirements for CUFFS	29
201.102 Connection tubing and CUFF connectors	29
201.103 Unauthorized access	29
201.104 * Maximum inflating time.....	30
201.105 * Automatic cycling modes	31
201.106 * Clinical accuracy	35
202 Electromagnetic compatibility – Requirements and tests	35
A1 206 Usability	38
211 Requirements for MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS used in the HOME HEALTHCARE ENVIRONMENT	38 A1
Annexes	39
Annex C (informative) Guide to marking and labelling requirements for ME EQUIPMENT and ME SYSTEMS	39
Annex AA (informative) Particular guidance and rationale	42
Annex BB (informative) Environmental aspects.....	50
Annex CC (informative) Reference to the essential principles	51
Bibliography.....	53
Index of defined terms	55
Figure 201.101 – CUFF pressure PROTECTION DEVICE, triggered by overpressure in SINGLE FAULT CONDITION.....	24
Figure 201.102 – CUFF pressure PROTECTION DEVICE, triggered by prolonged overpressure in SINGLE FAULT CONDITION	25
Figure 201.103 – CUFF pressure and maximum inflation time, NORMAL CONDITION and SINGLE FAULT CONDITION	31
Figure 201.104 – LONG-TERM AUTOMATIC MODE CUFF pressure in NORMAL CONDITION	32
Figure 201.105 – LONG-TERM AUTOMATIC MODE CUFF pressure in SINGLE FAULT CONDITION	32

Figure 201.106 – SHORT-TERM AUTOMATIC MODE CUFF pressure	33
Figure 201.107 – SELF-MEASUREMENT AUTOMATIC MODE CUFF pressure.....	34
Figure 202.101 – HF SURGICAL EQUIPMENT test layout	37
Figure 202.102 – Simulated PATIENT test set-up for HF SURGICAL EQUIPMENT	38
Table 201.101 – Distributed ESSENTIAL PERFORMANCE requirements	16
Table 201.102 – CUFF deflation pressure	21
Table 201.103 – CUFF inflation pressure	30
Table 201.C.101 – Marking on the outside of AUTOMATED SPHYGMOMANOMETERS or their parts	39
Table 201.C.102 – Marking of controls and instruments of AUTOMATED SPHYGMOMANOMETERS or their parts	40
Table 201.C.103 – ACCOMPANYING DOCUMENTS, general information for AUTOMATED SPHYGMOMANOMETERS	40
Table 201.C.104 – ACCOMPANYING DOCUMENTS, instructions for use of AUTOMATED SPHYGMOMANOMETERS	40
Table 201.C.105 – ACCOMPANYING DOCUMENTS, technical description of AUTOMATED SPHYGMOMANOMETERS	41
Table AA.1 – Summary of requirements by mode	48
Table BB.1 – Environmental aspects addressed by clauses of this standard.....	50
Table CC.1 – Correspondence between this particular standard and the essential principles.....	51

The committee has decided that the contents of this particular standard will remain unchanged until the maintenance result date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed;
- withdrawn;
- replaced by a revised edition, or
- amended.

NOTE The attention of National Committees is drawn to the fact that equipment manufacturers and testing organizations may need a transitional period following publication of a new, amended or revised IEC publication in which to make products in accordance with the new requirements and to equip themselves for conducting new or revised tests. It is the recommendation of the committee that the content of this publication be adopted for implementation nationally not earlier than 3 years from the date of publication for equipment newly designed and not earlier than 5 years from the date of publication for equipment already in production.

The contents of the corrigendum of January 2010 have been included in this copy.

INTRODUCTION

The minimum safety requirements specified in this particular standard are considered to provide for a practical degree of safety in the operation of an AUTOMATED SPHYGMOMANOMETER.

The requirements are followed by specifications for the relevant tests.

Following the decision taken by subcommittee 62D at the meeting in Washington in 1979, a "General guidance and rationale" section giving some explanatory notes, where appropriate, about the more important requirements is included in Annex AA.

It is considered that knowledge of the reasons for these requirements will not only facilitate the proper application of the standard but will, in due course, expedite any revision necessitated by changes in clinical practice or as a result of developments in technology. However, this annex does not form part of the requirements of this standard.

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-30: Particular requirements for the basic safety and essential performance of automated non-invasive sphygmomanometers

201.1 Scope, object and related standards

Clause 1 of the general standard¹⁾ applies, except as follows:

201.1.1 Scope

Replacement:

This International Standard applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of AUTOMATED SPHYGMOMANOMETERS, hereafter referred to as ME EQUIPMENT, which by means of an inflatable CUFF, are used for **A1** non-continuous **A1** indirect measurement of the BLOOD PRESSURE without arterial puncture.

NOTE 1 Equipment that performs indirect measurement of the BLOOD PRESSURE without arterial puncture does not directly measure the BLOOD PRESSURE. It only estimates the BLOOD PRESSURE.

This standard specifies requirements for the BASIC SAFETY and ESSENTIAL PERFORMANCE for this ME EQUIPMENT and its ACCESSORIES, including the requirements for the accuracy of a DETERMINATION.

This standard covers electrically-powered intermittent, indirect measurement of the BLOOD PRESSURE without arterial puncture, ME EQUIPMENT with automatic methods for estimating BLOOD PRESSURE, including BLOOD PRESSURE monitors for the HOME HEALTHCARE ENVIRONMENT.

Requirements for indirect measurement of the BLOOD PRESSURE without arterial puncture ME EQUIPMENT with an electrically-powered PRESSURE TRANSDUCER and/or displays used in conjunction with a stethoscope or other manual methods for determining BLOOD PRESSURE (NON-AUTOMATED SPHYGMOMANOMETERS) are specified in document ISO 81060-1.

If a clause or subclause is specifically intended to be applicable to ME EQUIPMENT only, or to ME SYSTEMS only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to ME EQUIPMENT and to ME SYSTEMS, as relevant.

HAZARDS inherent in the intended physiological function of ME EQUIPMENT or ME SYSTEMS within the scope of this standard are not covered by specific requirements in this standard except in 201.11 and 201.105.3.3, as well as 7.2.13 and 8.4.1 of IEC 60601-1.

NOTE 2 See also 4.2 of the general standard.

201.1.2 Object

Replacement:

The object of this particular standard is to establish particular BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for an AUTOMATED SPHYGMOMANOMETER as defined in 201.3.201.

¹⁾ The general standard is IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance* **A1** including Amendment 1:2012 **A1**.

201.1.3 Collateral standards

Addition:

This particular standard refers to those applicable collateral standards that are listed in Clause 2 of IEC 60601-1 and Clause 2 of this particular standard.

IEC 60601-1-2 is amended by this particular standard. IEC 60601-1-3 does not apply. All other published collateral standards in the IEC 60601-1 series apply as published.

201.1.4 Particular standards

Replacement:

In the IEC 60601 series, particular standards may modify, replace or delete requirements contained in the general standard and its collateral standards as appropriate for the particular ME EQUIPMENT under consideration, and may add other BASIC SAFETY and ESSENTIAL PERFORMANCE requirements.

A requirement of a particular standard takes priority over the general standard.

For brevity, IEC 60601-1 is referred to in this particular standard as the general standard. Collateral standards are referred to by their document number.

The numbering of clauses and subclauses of this particular standard corresponds to that of the general standard with the prefix "201" (e.g. 201.1 in this standard addresses the content of Clause 1 of the general standard) or applicable collateral standard with the prefix "20x" where x is the final digit(s) of the collateral standard document number (e.g. 202.4 in this particular standard addresses the content of Clause 4 of the 60601-1-2 collateral standard, 203.4 in this particular standard addresses the content of Clause 4 of the 60601-1-3 collateral standard, etc.). The changes to the text of the general standard are specified by the use of the following words:

"Replacement" means that the clause or subclause of the general standard or applicable collateral standard is replaced completely by the text of this particular standard.

"Addition" means that the text of this particular standard is additional to the requirements of the general standard or applicable collateral standard.

"Amendment" means that the clause or subclause of the general standard or applicable collateral standard is amended as indicated by the text of this particular standard.

Subclauses, figures or tables which are additional to those of the general standard are numbered starting from 201.101. However due to the fact that definitions in the general standard are numbered 3.1 through 3.139, additional definitions in this standard are numbered beginning from 201.3.201. Additional annexes are lettered AA, BB, etc., and additional items aa), bb), etc.

Subclauses or figures which are additional to those of a collateral standard are numbered starting from 20x, where "x" is the number of the collateral standard, e.g. 202 for IEC 60601-1-2, 203 for IEC 60601-1-3, etc.

The term "this standard" is used to make reference to the general standard, any applicable collateral standards and this particular standard taken together.

Where there is no corresponding clause or subclause in this particular standard, the clause or subclause of the general standard or applicable collateral standard, although possibly not relevant, applies without modification; where it is intended that any part of the general

standard or applicable collateral standard, although possibly relevant, is not to be applied, a statement to that effect is given in this particular standard.

201.2 Normative references

NOTE Informative references are listed in the bibliography beginning on page 49.

Clause 2 of the general standard applies, except as follows:

A1 Amendment of the following references: **A1**

IEC 60601-1-2:2007, *Medical electrical equipment – Part 1-2: General requirements for basic safety and essential performance – Collateral standard: Electromagnetic compatibility – Requirements and tests*

A1 IEC 60601-1-6:2010, *Medical electrical equipment – Part 1-6: General requirements for basic safety and essential performance – Collateral standard: Usability*
Amendment 1:2013

IEC 60601-1-8:2006, *Medical electrical equipment – Part 1-8: General requirements for basic safety and essential performance – Collateral standard: General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems*
Amendment 1:2012 **A1**

Addition:

IEC 60068-2-27:2008, *Environmental testing – Part 2-27: Tests – Test Ea and guidance: Shock*

IEC 60068-2-31:2008, *Environmental testing – Part 2-31: Tests – Test Ec: Rough handling shocks, primarily for equipment-type specimens*

IEC 60068-2-64:2008, *Environmental testing – Part 2-64: Tests – Test Fh: Vibration, broadband random and guidance*

A1 IEC 60601-1-11:2010, *Medical electrical equipment – Part 1-11: General requirements for basic safety and essential performance – Collateral standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment* **A1**

IEC 60601-2-2:2009, *Medical electrical equipment – Part 2-2: Particular requirements for the basic safety and essential performance of high frequency surgical equipment and high frequency surgical accessories*

A1 IEC 62366:2007, *Medical devices – Application of usability engineering to medical devices* **A1**

ISO 594-1:1986, *Conical fittings with a 6 % (Luer) taper for syringes, needles and certain other medical equipment – Part 1: General requirements*

ISO 594-2:1991, *Conical fittings with 6 % (Luer) taper for syringes, needles and certain other medical equipment – Part 2: Lock fittings*

A1 ISO 81060-2:2013, *Non-invasive sphygmomanometers – Part 2: Clinical investigation of automated measurement type* **A1**

201.3 Terms and definitions

A1 For the purposes of this document, the terms and definitions given in IEC 60601-1:2005+A1:2012, IEC 60601-1-2:2007, IEC 60601-1-8:2006+A1:2012, and IEC 60601-2-2:2009 apply, except as follows: **A1**

NOTE An index of defined terms is found beginning on page 51.

Addition:

201.3.201

AUTOMATED SPHYGMOMANOMETER

ME EQUIPMENT used for the non-invasive estimation of the BLOOD PRESSURE by utilizing an inflatable CUFF, a PRESSURE TRANSDUCER, a valve for deflation, and/or displays used in conjunction with automatic methods for determining BLOOD PRESSURE

NOTE Components of an AUTOMATED SPHYGMOMANOMETER include manometer, CUFF, valve for deflation (often in combination with the valve for rapidly exhausting the PNEUMATIC SYSTEM), pump for inflation of the BLADDER, and connection tubing.

201.3.202

BLADDER

part of the CUFF that is inflatable

[ISO 81060-1:2007, definition 3.2]

201.3.203

BLOOD PRESSURE

pressure in the systemic arterial system of the body

[ISO 81060-1:2007, definition 3.3]

201.3.204

CUFF

part of the AUTOMATED SPHYGMOMANOMETER that is wrapped around the limb of the PATIENT

NOTE A CUFF usually comprises a BLADDER and an inelastic part that encloses the BLADDER, or has an integral BLADDER (i.e., the CUFF, including the BLADDER, is one piece).

[ISO 81060-1:2007, definition 3.5, modified]

201.3.205

DETERMINATION (value)

result of the process of estimating BLOOD PRESSURE by the AUTOMATED SPHYGMOMANOMETER

201.3.206

DIASTOLIC BLOOD PRESSURE (value)

minimum value of the BLOOD PRESSURE as a result of relaxation of the systemic ventricle

NOTE Because of hydrostatic effects, this value should be determined with the CUFF at the level of the heart.

201.3.207

HOME HEALTHCARE ENVIRONMENT

A1 dwelling place in which a PATIENT lives or other places where PATIENTS are present, excluding professional healthcare facility environments where OPERATORS with medical training are continually available when PATIENTS are present

NOTE 1 Professional healthcare facilities include hospitals, physician offices, freestanding surgical centres, dental offices, freestanding birthing centres, limited care facilities, multiple treatment facilities and emergency medical services. **A1**

A1 NOTE 2 For the purpose of this particular standard, nursing homes are considered the HOME HEALTHCARE ENVIRONMENT.

NOTE 3 Other places where PATIENTS are present include the outdoor environment and in vehicles.

EXAMPLES In a car, bus, train, boat or plane, in a wheelchair or walking outdoors.

[SOURCE: IEC 60601-1-11:2010, definition 3.2] **A1**

201.3.208

LONG-TERM AUTOMATIC MODE

mode in which a timer, set by the OPERATOR, initiates multiple DETERMINATIONS

201.3.209

MEAN ARTERIAL PRESSURE (value)

value of the integral of one heartbeat cycle of the BLOOD PRESSURE curve divided by the time of that cycle

NOTE Because of hydrostatic effects, this value should be determined with the CUFF at the level of the heart.

201.3.210

NEONATAL MODE

mode of AUTOMATED SPHYGMOMANOMETER for use with neonates or infants

NOTE 1 The approximate age range for a newborn (neonate) is from birth to 1 month. [10]²⁾

NOTE 2 The approximate age range for an infant is from 1 month to 2 years. [10] For the purposes of this standard, up to 3 years of age are considered infants (see ISO 81060-2, 6.1.3).

NOTE 3 The NEONATAL MODE is used to limit the maximum pressure to 150 mmHg and frequently has a different algorithm from other modes intended for older PATIENTS.

201.3.211

NON-AUTOMATED SPHYGMOMANOMETER

ME EQUIPMENT used for the non-invasive measurement of the BLOOD PRESSURE by utilizing an inflatable CUFF with a pressure-sensing element, a valve for deflation, and display used in conjunction with a stethoscope or other manual methods for estimating BLOOD PRESSURE

NOTE Components of these instruments include manometer, CUFF, valve for deflation (often in combination with the valve for rapidly exhausting the PNEUMATIC SYSTEM), hand pump or electro-mechanical pump for inflation of the BLADDER, and connection tubing. A NON-AUTOMATED SPHYGMOMANOMETER can also contain electro-mechanical components for pressure control.

[ISO 81060-1:2007, definition 3.11, modified]

201.3.212

PATIENT SIMULATOR

equipment for simulating the oscillometric CUFF pulses and/or auscultatory signals during inflation and deflation

NOTE This equipment is not used for testing accuracy but is used in assessing stability of performance.

201.3.213

PNEUMATIC SYSTEM

part of the AUTOMATED SPHYGMOMANOMETER that includes all pressurized and pressure-controlling components

EXAMPLES CUFF, tubing, connectors, valves, PRESSURE TRANSDUCER and pump.

[ISO 81060-1:2007, definition 3.16, modified]

²⁾ Figures in square brackets refer to the Bibliography.

201.3.214

PRESSURE TRANSDUCER

component that transforms sensed pressure into an electrical signal

201.3.215

PROTECTION DEVICE

part of ME EQUIPMENT that, without intervention by the OPERATOR, protects the PATIENT from hazardous output due to incorrect delivery of energy or substances

201.3.216

SELF-MEASUREMENT AUTOMATIC MODE

A1 mode of AUTOMATED SPHYGMOMANOMETER that is manually initiated and overseen by the OPERATOR and in which a limited number of repeated DETERMINATIONS are made over a limited period **A1**

201.3.217

* SHORT-TERM AUTOMATIC MODE

A1 mode of AUTOMATED SPHYGMOMANOMETER that is manually initiated by the OPERATOR and in which rapid repetitive automatic DETERMINATIONS are made within a specified time period **A1**

201.3.218

SYSTOLIC BLOOD PRESSURE (value)

maximum value of the BLOOD PRESSURE as a result of the contraction of the systemic ventricle

NOTE Because of hydrostatic effects, this value should be determined with the CUFF at the level of the heart.

201.4 General requirements

Clause 4 of the general standard applies, except as follows:

201.4.3 Essential performance

Additional subclause:

201.4.3.101 Additional essential performance requirements

Additional ESSENTIAL PERFORMANCE requirements for an AUTOMATED SPHYGMOMANOMETER are found in the subclauses listed in Table 201.101.

Table 201.101 – Distributed ESSENTIAL PERFORMANCE requirements

Requirement	Subclause
Electrosurgery interference recovery	202.6.2.101
Limits of the error of the manometer, or generation of a TECHNICAL ALARM CONDITION	201.12.1.102 201.11.8.102 201.12.1.101
A1 Reproducibility A1 of the BLOOD PRESSURE DETERMINATION and low and high BLOOD PRESSURE PHYSIOLOGICAL ALARM CONDITIONS (if provided), or generation of a TECHNICAL ALARM CONDITION	201.12.1.107 201.12.3.101 201.11.8.102 201.12.1.101

201.5 General requirements for testing ME EQUIPMENT

Clause 5 of the general standard applies.