



BSI Standards Publication

Medical electrical equipment

Part 2-28: Particular requirements for the basic safety and essential performance of X-ray tube assemblies for medical diagnosis

National foreword

This British Standard is the UK implementation of EN IEC 60601-2-28:2019. It is identical to IEC 60601-2-28:2017. It supersedes BS EN 60601-2-28:2010, which is withdrawn.

The UK participation in its preparation was entrusted to Technical Committee CH/62/2, Diagnostic imaging equipment.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2019
Published by BSI Standards Limited 2019

ISBN 978 0 580 91560 4

ICS 11.040.55

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 30 September 2019.

Amendments/corrigenda issued since publication

Date	Text affected
------	---------------

English Version

**Medical electrical equipment - Part 2-28: Particular requirements
for the basic safety and essential performance of X-ray tube
assemblies for medical diagnosis
(IEC 60601-2-28:2017)**

Appareils électromédicaux - Partie 2-28 : Exigences
particulières pour la sécurité de base et les performances
essentielles des gaines équipées pour diagnostic médical
(IEC 60601-2-28:2017)

Medizinische elektrische Geräte - Teil 2-28: Besondere
Festlegungen für die Sicherheit einschließlich der
wesentlichen Leistungsmerkmale von Röntgenstrahlern für
die medizinische Diagnostik
(IEC 60601-2-28:2017)

This European Standard was approved by CENELEC on 2019-08-07. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.



European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

European foreword

The text of document 62B/1040/FDIS, future edition 3 of IEC 60601-2-28, prepared by SC 62B "Diagnostic imaging equipment" of IEC/TC 62 "Electrical equipment in medical practice" was submitted to the IEC-CENELEC parallel vote and approved by CENELEC as EN IEC 60601-2-28:2019.

The following dates are fixed:

- latest date by which the document has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2020-05-07
- latest date by which the national standards conflicting with the document have to be withdrawn (dow) 2022-08-07

This document supersedes EN 60601-2-28:2010.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC shall not be held responsible for identifying any or all such patent rights.

Endorsement notice

The text of the International Standard IEC 60601-2-28:2017 was approved by CENELEC as a European Standard without any modification.

In the official version, for Bibliography, the following notes have to be added for the standards indicated:

IEC 60601-1-9	NOTE	Harmonized as EN 60601-1-9
IEC 60601-1-10	NOTE	Harmonized as EN 60601-1-10
IEC 61010-1:2010	NOTE	Harmonized as EN 61010-1:2010 (not modified)
ISO 13732-1	NOTE	Harmonized as EN ISO 13732-1

Annex ZA

(normative)

Normative references to international publications with their corresponding European publications

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE 1 Where an International Publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

NOTE 2 Up-to-date information on the latest versions of the European Standards listed in this annex is available here: www.cenelec.eu.

The annex ZA of EN 60601-1:2006 applies, except as follows:

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
<i>Addition</i>				
IEC 60336	-	Medical electrical equipment - X-ray tube assemblies for medical diagnosis - Characteristics of focal spots	EN 60336	-
IEC 60522	-	Determination of the permanent filtration of X-ray tube assemblies	EN 60522	-
IEC TR 60788	2004	Medical electrical equipment - Glossary of-defined terms		-
<i>Replacement</i>				
IEC 60601-1-3	2008	Medical electrical equipment - Part 1-3: General requirements for basic safety and essential performance - Collateral Standard: Radiation protection in diagnostic X-ray equipment	EN 60601-1-3	2008
			+EN 60601-1-2010	
			3:2008/corrigendum	
			Mar. 2010	
			+A11	2016

CONTENTS

FOREWORD	3
201. 1 Scope, object and related standards	5
201. 2 Normative references	6
201. 3 Terms and definitions	7
201. 4 General requirements	7
201. 5 General requirements for testing ME EQUIPMENT	8
201. 6 Classification of ME EQUIPMENT and ME SYSTEMS	8
201. 7 ME EQUIPMENT identification, marking and documents	9
201. 8 Protection against electrical HAZARDS from ME EQUIPMENT	13
201. 9 Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS	14
201. 10 Protection against unwanted and excessive RADIATION HAZARDS	15
201. 11 Protection against excessive temperatures and other HAZARDS	15
201. 12 Accuracy of controls and instruments and protection against hazardous outputs	16
201. 13 HAZARDOUS SITUATIONS and fault conditions for ME EQUIPMENT	16
201. 14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS)	17
201. 15 Construction of ME EQUIPMENT	17
201. 16 ME SYSTEMS	17
201. 17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS	17
203 RADIATION protection in diagnostic X-RAY EQUIPMENT	18
Annexes	20
Annex AA (informative) Test of X-RAY TUBE ASSEMBLIES for expelled parts-related and/or tube implosion-related RISKS	20
Bibliography	22
Index of defined terms used in this particular standard	23

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –

**Part 2-28: Particular requirements for the basic safety
and essential performance of X-ray tube assemblies for medical diagnosis**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as “IEC Publication(s)”). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 60601-2-28 has been prepared by subcommittee 62B: Diagnostic imaging equipment, of IEC technical committee 62: Electrical equipment in medical practice.

This third edition cancels and replaces the second edition published in 2010. This edition constitutes a technical revision.

The third edition of this particular standard has been prepared to fit IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012 (the amended third edition of IEC 60601-1), which is referred to as the general standard. Apart from the changes related to the amendment of IEC 60601-1, changes related to technical improvements are also included.

The text of this standard is based on the following documents:

FDIS	Report on voting
62B/1040/FDIS	62B/1051/RVD

Full information on the voting for the approval of this particular standard can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this standard, the following print types are used:

- requirements and definitions: roman type;
- *test specifications: italic type;*
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type;
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this standard, the term

- “clause” means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this standard are preceded by the term “Clause” followed by the clause number. References to subclauses within this particular standard are by number only.

In this standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Clause 7 of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

A list of all parts of the IEC 60601 series, published under the general title *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of this publication will remain unchanged until the stability date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-28: Particular requirements for the basic safety and essential performance of X-ray tube assemblies for medical diagnosis

201.1 Scope, object and related standards

Clause 1 of the general standard¹ applies, except as follows:

201.1.1 Scope

Replacement:

This part of IEC 60601 applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of X-RAY TUBE ASSEMBLIES and to components thereof, intended for medical diagnosis and imaging.

Where the general standard IEC 60601-1 and the collateral standard IEC 60601-1-3 refer to ME EQUIPMENT, this is interpreted as X-RAY TUBE ASSEMBLIES in this particular standard. If a clause or subclause is specifically intended to be applicable to ME EQUIPMENT only, or to ME SYSTEMS only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to ME EQUIPMENT and to ME SYSTEMS, as relevant.

NOTE This document is also applicable to the X-RAY TUBE ASSEMBLY aspects of X-RAY SOURCE ASSEMBLIES and X-RAY TUBE HEADS.

201.1.2 Object

Replacement:

The object of this particular standard is to establish particular BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for X-RAY TUBE ASSEMBLIES for medical diagnosis.

201.1.3 Collateral standards

Addition:

This particular standard refers to those applicable collateral standards that are listed in Clause 2 of the general standard and Clause 201.2 of this particular standard.

IEC 60601-1-3:2008 and IEC 60601-1-3:2008/AMD1:2013 applies as modified in Clause 203. IEC 60601-1-2, IEC 60601-1-6, IEC 60601-1-8, IEC 60601-1-9, IEC 60601-1-10, IEC 60601-1-11 and IEC 60601-1-12 do not apply. All other published collateral standards in the IEC 60601-1 series apply as published.

NOTE 101 IEC 60601-1-2 does not apply because RISKS for the X-RAY TUBE ASSEMBLY outside the system may only be indicative of RISKS for the system due to the difference in electromagnetic environment.

NOTE 102 IEC 60601-1-6 and IEC 60601-1-8 do not apply because X-RAY TUBE ASSEMBLIES are not operated as a stand-alone device.

NOTE 103 X-RAY TUBE ASSEMBLIES are not in the scope of IEC 60601-1-10, IEC 60601-1-11 and IEC 60601-1-12.

¹ The general standard is IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*.