



BSI Standards Publication

Medical electrical equipment

Part 2-46: Particular requirements for the basic safety
and essential performance of operating tables

National foreword

This British Standard is the UK implementation of EN IEC 60601-2-46:2024. It is identical to IEC 60601-2-46:2023. It supersedes BS EN IEC 60601-2-46:2019, which will be withdrawn on 31 July 2027.

The UK participation in its preparation was entrusted to Technical Committee CH/62/4, Electromedical equipment.

A list of organizations represented on this committee can be obtained on request to its committee manager.

Contractual and legal considerations

This publication has been prepared in good faith, however no representation, warranty, assurance or undertaking (express or implied) is or will be made, and no responsibility or liability is or will be accepted by BSI in relation to the adequacy, accuracy, completeness or reasonableness of this publication. All and any such responsibility and liability is expressly disclaimed to the full extent permitted by the law.

This publication is provided as is, and is to be used at the recipient's own risk.

The recipient is advised to consider seeking professional guidance with respect to its use of this publication.

This publication is not intended to constitute a contract. Users are responsible for its correct application.

© The British Standards Institution 2024
Published by BSI Standards Limited 2024

ISBN 978 0 539 16781 8

ICS 11.140

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 30 September 2024.

Amendments/corrigenda issued since publication

Date	Text affected
------	---------------

English Version

Medical electrical equipment - Part 2-46: Particular requirements
for the basic safety and essential performance of operating
tables
(IEC 60601-2-46:2023)

Appareils électromédicaux - Partie 2-46: Exigences
particulières pour la sécurité de base et les performances
essentiels des tables d'opération
(IEC 60601-2-46:2023)

Medizinische elektrische Geräte - Teil 2-46: Besondere
Festlegungen für die Sicherheit einschließlich der
wesentlichen Leistungsmerkmale von Operationstischen
(IEC 60601-2-46:2023)

This European Standard was approved by CENELEC on 2024-07-31. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.



European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

European foreword

The text of document 62D/1939/CDV, future edition 4 of IEC 60601-2-46, prepared by SC 62D "Particular medical equipment, software, and systems" of IEC/TC 62 "Medical equipment, software, and systems" was submitted to the IEC-CENELEC parallel vote and approved by CENELEC as EN IEC 60601-2-46:2024.

The following dates are fixed:

- latest date by which the document has to be implemented at national (dop) 2025-05-01 level by publication of an identical national standard or by endorsement
- latest date by which the national standards conflicting with the (dow) 2027-07-31 document have to be withdrawn

This document supersedes EN IEC 60601-2-46:2019 and all of its amendments and corrigenda (if any).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC shall not be held responsible for identifying any or all such patent rights.

Any feedback and questions on this document should be directed to the users' national committee. A complete listing of these bodies can be found on the CENELEC website.

Endorsement notice

The text of the International Standard IEC 60601-2-46:2023 was approved by CENELEC as a European Standard without any modification.

In the official version, for Bibliography, the following notes have to be added for the standard indicated:

IEC 60601-1-8:2006	NOTE	Approved as EN 60601-1-8:2007 (not modified) +A11:2017
IEC 60601-1-8:2006/A1:2012	NOTE	Approved as EN 60601-1-8:2007/A1:2013 (not modified)
IEC 60601-1-8:2006/A2:2020	NOTE	Approved as EN 60601-1-8:2007/A2:2021 (not modified)
IEC 60601-1-9:2007	NOTE	Approved as EN 60601-1-9:2008 (not modified)
IEC 60601-1-9:2007/A1:2013	NOTE	Approved as EN 60601-1-9:2008/A1:2013 (not modified)
IEC 60601-1-9:2007/A2:2020	NOTE	Approved as EN 60601-1-9:2008/A2:2020 (not modified)
IEC 60601-1-10:2007	NOTE	Approved as EN 60601-1-10:2008 (not modified)
IEC 60601-1-10:2007/A1:2013	NOTE	Approved as EN 60601-1-10:2008/A1:2015 (not modified)
IEC 60601-1-10:2007/A2:2020	NOTE	Approved as EN 60601-1-10:2008/A2:2021 (not modified)
IEC 60601-1-11:2015	NOTE	Approved as EN 60601-1-11:2015 (not modified)
IEC 60601-1-11:2015/A1:2020	NOTE	Approved as EN 60601-1-11:2015/A1:2021 (not modified)
IEC 60601-1-12:2014	NOTE	Approved as EN 60601-1-12:2015 (not modified)

IEC 60601-1-12:2014/A1:2020	NOTE	Approved as EN 60601-1-12:2015/A1:2020 (not modified)
IEC 60601-2-35:2020	NOTE	Approved as EN IEC 60601-2-35:2021 (not modified)
IEC 60601-2-52:2009/A1:2015	NOTE	Approved as EN 60601-2-52:2010/A1:2015 (not modified)
IEC 62366-1:2015	NOTE	Approved as EN 62366-1:2015 (not modified)
IEC 62366-1:2015/A1:2020	NOTE	Approved as EN 62366-1:2015/A1:2020 (not modified)
ISO 7494-1:2018	NOTE	Approved as EN ISO 7494-1:2018 (not modified)
ISO 20342-1:2022	NOTE	Approved as EN ISO 20342-1:2022 (not modified)

Annex ZA
(normative)

Normative references to international publications
with their corresponding European publications

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE 1 Where an International Publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

NOTE 2 Up-to-date information on the latest versions of the European Standards listed in this annex is available here: www.cenelec.eu.

Annex ZA of EN 60601-1:2006¹, applies, except as follows:

Replace:

Publication	Year	Title	EN/HD	Year
IEC 60601-1-3	2008	Medical electrical equipment - Part 1-3: General requirements for basic safety and essential performance - Collateral Standard: Radiation protection in diagnostic X-ray equipment	EN 60601-1-3	2008
-	-		+ AC	2010
+ A1	2013		+ A1	2013
-	-		+ AC	2014
-	-		+ A11	2016
+ A2	2021		+ A2	2021
ISO 2878	2017	Rubber, vulcanized or thermoplastic - Antistatic and conductive products - Determination of electrical resistance	-	-

¹ As impacted by EN 60601-1:2006/AC:2010, EN 60601-1:2006/A1:2013, EN 60601-1:2006/A1:2013/AC:2014, EN 60601-1:2006/A12:2014, EN 60601-1:2006/A2:2021, EN 60601-1:2006/AC:2022-12 and EN 60601-1:2006/A13:2024.

Add:

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
IEC 60601-2-2	2017	Medical electrical equipment - Part 2-2: Particular requirements for the basic safety and essential performance of high frequency surgical equipment and high frequency surgical accessories	EN IEC 60601-2-2	2018
IEC 60601-2-43	2022	Medical electrical equipment - Part 2-43: Particular requirements for the basic safety and essential performance of X-ray equipment for interventional procedures	EN IEC 60601-2-43	2023
IEC 60601-2-54	2022	Medical electrical equipment - Part 2-54: Particular requirements for the basic safety and essential performance of X-ray equipment for radiography and radioscopy	EN IEC 60601-2-54	2024

CONTENTS

FOREWORD.....	3
INTRODUCTION.....	6
201.1 Scope, object and related standards	7
201.2 Normative references.....	9
201.3 Terms and definitions.....	9
201.4 General requirements	10
201.5 General requirements for testing ME EQUIPMENT.....	10
201.6 Classification of ME EQUIPMENT and ME SYSTEMS	10
201.7 ME EQUIPMENT identification, marking and documents	11
201.8 Protection against electrical HAZARDS from ME EQUIPMENT.....	11
201.9 Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS.....	11
201.10 Protection against unwanted and excessive radiation HAZARDS	14
201.11 Protection against excessive temperatures and other HAZARDS	14
201.12 Accuracy of controls and instruments and protection against hazardous outputs	15
201.13 HAZARDOUS SITUATIONS and fault conditions for ME EQUIPMENT	15
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS).....	15
201.15 Construction of ME EQUIPMENT	15
201.16 ME SYSTEMS.....	15
201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS	15
202 Electromagnetic disturbances – Requirements and tests	16
203 *Radiation protection in diagnostic X-ray equipment	19
Annexes	20
Annex G (normative) Protection against HAZARDS of ignition of flammable anaesthetic mixtures.....	21
Annex AA (informative) Particular guidance and rationale.....	22
Bibliography.....	26
Index of defined terms used in this particular standard.....	27
Figure 202.101 – ENCLOSURE ad hoc test.....	17
Figure 202.102 – POWER SUPPLY CORD ad hoc test.....	18
Figure 202.103 – ACCESSORY cable ad hoc test	18
Figure AA.1 – Recommended distribution of mass in excess of 135 kg and examples of application	22
Figure AA.2 – Typical stress-strain curve.....	24
Figure AA.3 – Typical bending line along the length L_0 of a beam	24
Table 201.101 – Determination of TENSILE SAFETY FACTOR.....	13
Table AA.1 – Recommended distribution of mass in excess of 135 kg and examples of application	23

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –**Part 2-46: Particular requirements for the basic safety
and essential performance of operating tables**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

IEC 60601-2-46 has been prepared by subcommittee 62D: Particular medical equipment, software, and systems, of IEC technical committee 62: Medical equipment, software, and systems. It is an International Standard.

This fourth edition cancels and replaces the third edition published in 2016. This edition constitutes a technical revision.

This edition includes the following significant technical change with respect to the previous edition: structural alignment with IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020.

The text of this International Standard is based on the following documents:

Draft	Report on voting
62D/1939/CDV	62D/1989/RVC

Full information on the voting for its approval can be found in the report on voting indicated in the above table.

The language used for the development of this International Standard is English.

This document was drafted in accordance with ISO/IEC Directives, Part 2, and developed in accordance with ISO/IEC Directives, Part 1 and ISO/IEC Directives, IEC Supplement, available at www.iec.ch/members_experts/refdocs. The main document types developed by IEC are described in greater detail at www.iec.ch/publications.

In this document, the following print types are used:

- requirements and definitions: roman type;
- *test specifications: italic type*;
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type;
- TERMS DEFINED IN CLAUSE 3 OF IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 AND IEC 60601-1:2005/AMD2:2020, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this document, the term

- "clause" means one of the numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes Subclauses 7.1, 7.2, etc.);
- "subclause" means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this document are preceded by the term "Clause" followed by the clause number. References to subclauses within this particular standard are by number only.

In this document, the conjunctive "or" is used as an "inclusive or" so a statement is true if any combination of the conditions is true.

The verbal forms used in this document conform to usage described in Clause 7 of the ISO/IEC Directives, Part 2. For the purposes of this document, the auxiliary verb:

- "shall" means that compliance with a requirement or a test is mandatory for compliance with this document;
- "should" means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this document;
- "may" is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

A list of all parts of the IEC 60601 series, published under the general title *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under webstore.iec.ch in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.