



BSI Standards Publication

Evaluation and routine testing in medical imaging departments

Part 3-5: Acceptance and constancy tests — Imaging
performance of computed tomography X-ray equipment

National foreword

This British Standard is the UK implementation of EN IEC 61223-3-5:2019. It is identical to IEC 61223-3-5:2019 incorporating corrigendum June 2022. It supersedes BS EN 61223-3-5:2004, which is withdrawn.

The start and finish of text introduced or altered by corrigendum is indicated in the text by tags. Text altered by IEC corrigendum June 2022 is indicated in the text by AC1 AC1.

The UK participation in its preparation was entrusted to Technical Committee CH/62/2, Diagnostic imaging equipment.

A list of organizations represented on this committee can be obtained on request to its committee manager.

Contractual and legal considerations

This publication has been prepared in good faith, however no representation, warranty, assurance or undertaking (express or implied) is or will be made, and no responsibility or liability is or will be accepted by BSI in relation to the adequacy, accuracy, completeness or reasonableness of this publication. All and any such responsibility and liability is expressly disclaimed to the full extent permitted by the law.

This publication is provided as is, and is to be used at the recipient's own risk.

The recipient is advised to consider seeking professional guidance with respect to its use of this publication.

This publication is not intended to constitute a contract. Users are responsible for its correct application.

© The British Standards Institution 2022
Published by BSI Standards Limited 2022

ISBN 978 0 539 22972 1

ICS 11.040.50

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 30 November 2019.

Amendments/corrigenda issued since publication

Date	Text affected
31 August 2022	Implementation of IEC corrigendum June 2022

English Version

Evaluation and routine testing in medical imaging departments -
Part 3-5: Acceptance tests and constancy tests - Imaging
performance of computed tomography X-ray equipment
(IEC 61223-3-5:2019)

Essais d'évaluation et de routine dans les services
d'imagerie médicale - Partie 3-5: Essais d'acceptation et de
constance - Performance d'imagerie des équipements de
tomodensitométrie à rayonnement X
(IEC 61223-3-5:2019)

Bewertung und routinemäßige Prüfung in Abteilungen für
medizinische Bildgebung - Teil 3-5: Abnahmeprüfungen -
Leistungsmerkmale zur Bildgebung von
Röntgeneinrichtungen für Computertomographie
(IEC 61223-3-5:2019)

This European Standard was approved by CENELEC on 2019-10-21. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.



European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

European foreword

The text of document 62B/1134/FDIS, future edition 2 of IEC 61223-3-5, prepared by SC 62B "Diagnostic imaging equipment" of IEC/TC 62 "Electrical equipment in medical practice" was submitted to the IEC-CENELEC parallel vote and approved by CENELEC as EN IEC 61223-3-5:2019.

The following dates are fixed:

- latest date by which the document has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2020-07-21
- latest date by which the national standards conflicting with the document have to be withdrawn (dow) 2022-10-21

This document supersedes EN 61223-3-5:2004 and all of its amendments and corrigenda (if any).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC shall not be held responsible for identifying any or all such patent rights.

Endorsement notice

The text of the International Standard IEC 61223-3-5:2019 was approved by CENELEC as a European Standard without any modification.

In the official version, for Bibliography, the following note has to be added for the standard indicated:

IEC 61674:2012 NOTE Harmonized as EN 61674:2013 (not modified)

Endorsement notice to corrigendum AC:2022-07

The text of the corrigendum IEC 61223-3-5:2019/COR1:2022 was approved by CENELEC as EN IEC 61223-3-5:2019/AC:2022-07 without any modification.

Annex ZA

(normative)

Normative references to international publications with their corresponding European publications

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE 1 Where an International Publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

NOTE 2 Up-to-date information on the latest versions of the European Standards listed in this annex is available here: www.cenelec.eu.

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
IEC 60601-1	2005	Medical electrical equipment - Part 1: General requirements for basic safety and essential performance	EN 60601-1	2006
+ A1	2012		+ A1	2013
-	-		+ A12	2014
IEC 60601-2-44	2009	Medical electrical equipment - Part 2-44: Particular requirements for the basic safety and essential performance of X-ray equipment for computed tomography	EN 60601-2-44	2009
-	-		+ A11	2011
+ A1	2012		+ A1	2012
+ A2	2016		+ A2	2016
IEC/TR 60788	2004	Medical electrical equipment - Glossary of defined terms		-

CONTENTS

FOREWORD.....	5
INTRODUCTION.....	7
1 Scope and object.....	8
2 Normative references	9
3 Terms and definitions	9
4 General aspects of ACCEPTANCE TESTS and CONSTANCY TESTS	17
4.1 General conditions to be considered in test procedures	17
4.1.1 General	17
4.1.2 Preconditions.....	18
4.1.3 BASELINE VALUES	18
4.1.4 Identification and recording of equipment, instrumentation, and test conditions	18
4.1.5 TEST DEVICES	18
4.2 Documents and data for the tests in the ACCOMPANYING DOCUMENTS	19
4.3 Scope of tests.....	20
4.4 Considerations for selection of ACCEPTANCE and CONSTANCY TESTS	20
4.5 Measuring equipment.....	21
4.6 Actions to be taken after a MAJOR SERVICE ACTION	21
4.7 Establishment of BASELINE VALUES.....	21
4.8 Frequency of CONSTANCY TESTS.....	22
5 Test methods for CT SCANNERS	22
5.1 Positioning of the PATIENT SUPPORT	22
5.1.1 Summary	22
5.1.2 Test equipment.....	22
5.1.3 Test procedure	22
5.1.4 Data evaluation	23
5.1.5 Criteria to be applied	23
5.1.6 Constancy testing	23
5.2 PATIENT positioning accuracy	24
5.2.1 Axial PATIENT positioning accuracy	24
5.2.2 Sagittal and coronal PATIENT positioning light accuracy (if available).....	25
5.2.3 Constancy testing – Axial, sagittal, and coronal positioning light accuracy	25
5.3 RECONSTRUCTED SECTION THICKNESS.....	26
5.3.1 General	26
5.3.2 RECONSTRUCTED SECTION THICKNESS for axial scanning	26
5.3.3 RECONSTRUCTED SECTION THICKNESS for helical scanning	28
5.4 Dose	28
5.4.1 Summary	28
5.4.2 Test equipment.....	28
5.4.3 Test procedure	28
5.4.4 Data evaluation	29
5.4.5 Criteria to be applied	30
5.4.6 Constancy testing	30
5.5 MEAN CT NUMBER, magnitude of NOISE, and UNIFORMITY	31
5.5.1 Summary	31
5.5.2 Test equipment.....	32

5.5.3	Test procedure	32
5.5.4	Scan conditions	32
5.5.5	Criteria to be applied for ACCEPTANCE TEST	35
5.5.6	Criteria to be applied for CONSTANCY TESTS	36
5.6	SPATIAL RESOLUTION (high contrast)	38
5.6.1	Summary	38
5.6.2	Information to be supplied in the ACCOMPANYING DOCUMENTS	38
5.6.3	Test equipment	38
5.6.4	Test procedure	38
5.6.5	Data evaluation	39
5.6.6	Criteria to be applied	39
5.6.7	Constancy testing	39
5.7	Automatic exposure control (AEC)	39
AC1 5.8	LOW CONTRAST RESOLUTION	40 AC1
Annex A (informative) Visual method for LOW CONTRAST RESOLUTION		41
Annex B (informative) DOSE PROFILE		42
B.1	Summary	42
B.2	Methods	42
B.2.1	Point dosimeter method	42
B.2.2	Film method	42
B.2.3	Criteria to be applied	42
Annex C (informative) Accuracy of the gantry tilt		43
C.1	Summary	43
C.2	Method A	43
C.2.1	Test equipment	43
C.2.2	Test procedure	43
C.2.3	Data evaluation	43
C.2.4	Criteria to be applied	43
C.3	Method B	44
C.3.1	Test equipment	44
C.3.2	Test procedure	44
C.3.3	Data evaluation	44
C.4	Criteria to be applied	44
Annex D (informative) Characterization of z-axis SPATIAL RESOLUTION		45
Annex E (informative) Helical RECONSTRUCTED SECTION THICKNESS		46
E.1	Summary	46
E.2	Test equipment	46
E.3	Test procedure	46
E.4	Data evaluation	46
Annex F (informative) Guidance on action to be taken		47
F.1	Failing the ESTABLISHED CRITERIA at first measurement	47
F.2	Failing the ESTABLISHED CRITERIA after repeated measurement	47
F.3	Marginally failing the ESTABLISHED CRITERIA	47
F.4	Substantially failing the ESTABLISHED CRITERIA	47
F.5	History of repeatedly failing the ESTABLISHED CRITERIA	48
F.6	Failing the established CONSTANCY CRITERIA but passing the established ACCEPTANCE CRITERIA	48
F.7	Cases not covered by Clauses F.1 to F.5	48

Annex G (informative) Automated exposure control (AEC)	49
G.1 Overview	49
G.2 Test equipment	49
G.3 Test procedure	49
G.4 Size-dependent modulation evaluation	49
G.4.1 Size-dependent modulation evaluation for Adult Body PROTOCOL ELEMENTS	49
G.4.2 Size-dependent modulation evaluation for Paediatric Body PROTOCOL ELEMENTS	50
G.5 Longitudinal modulation evaluation	50
G.6 Data evaluation	51
G.6.1 Size-dependent modulation evaluation	51
G.6.2 Longitudinal modulation evaluation	51
G.7 Criteria to be applied	51
G.7.1 Size-dependent modulation evaluation	51
G.7.2 Longitudinal modulation evaluation	51
Annex H (informative) Mapping of IEC requirements to regulations	52
Annex I (informative) Overview of criteria for acceptance and constancy testing for 5.5	54
Annex J (informative) Overview of criteria and frequency for all acceptance and constancy testing	55
Bibliography	59
Index of defined terms	60
 Figure 1 – Coordinate system	 14
Figure 2 – Illustration of $N \times T$, R and $(N \times T) + R$	16
Figure G.1 – TEST DEVICE aligned	50
 Table 1 – Test pattern for $CTDI_{free\ air}$ for Adult Body PROTOCOL ELEMENTS	 29
Table 2 – Combination of PROTOCOL ELEMENTS and PHANTOMS used for ACCEPTANCE TEST scans	33
Table 3 – Combination of PROTOCOL ELEMENTS and PHANTOMS used for CONSTANCY TEST scans	33
Table H.1 – Mapping of IEC requirements to regulations	52
Table I.1 – Overview of criteria for ACCEPTANCE and CONSTANCY TESTING for 5.5	54
Table J.1 – Overview of criteria and frequency	55

INTERNATIONAL ELECTROTECHNICAL COMMISSION

**EVALUATION AND ROUTINE TESTING IN
MEDICAL IMAGING DEPARTMENTS –****Part 3-5: Acceptance and constancy tests – Imaging
performance of computed tomography X-ray equipment**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 61223-3-5 has been prepared by subcommittee 62B: Diagnostic imaging equipment, of IEC technical committee 62: Electrical equipment in medical practice.

This second edition cancels and replaces the first edition published in 2004, and the second edition of IEC 61223-2-6 published in 2006. This edition constitutes a technical revision.

This edition includes the following significant technical change with respect to the previous edition and to IEC 61223-2-6:

- a) modification of the RADIATION protection and control;
- b) modification of the acceptance testing;
- c) introduction of constancy testing.

The text of this International Standard is based on the following documents:

FDIS	Report on voting
62B/1134/FDIS	62B/1145/RVD

Full information on the voting for the approval of this International Standard can be found in the report on voting indicated in the above table.

This document has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this document, the following print types are used:

- requirements, compliance with which can be tested, and definitions: in roman type;
- explanations, advice, notes, general statements, exceptions and references: in smaller type;
- *test specifications: in italic type;*
- TERMS USED THROUGHOUT THIS DOCUMENT THAT HAVE BEEN LISTED IN THE INDEX OF DEFINED TERMS AND DEFINED IN CLAUSE 3, OR IN OTHER STANDARDS: SMALL CAPITALS.

A list of all parts in the IEC 61223 series, published under the general title *Evaluation and routine testing in medical imaging departments*, can be found on the IEC website.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under "<http://webstore.iec.ch>" in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

NOTE The attention of the users of this document is drawn to the fact that equipment MANUFACTURERS and testing organizations may need a transitional period following publication of a new, amended or revised IEC publication in which to make products in accordance with the new requirements and to equip themselves for conducting new or revised tests. It is the recommendation of the committee that the content of this publication be adopted for implementation nationally not earlier than 3 years from the date of publication.

INTRODUCTION

This part of IEC 61223 gives methods for acceptance testing and constancy testing for medical diagnostic CT equipment.

The complete set of ACCEPTANCE TESTS is to be carried out after new equipment has been installed, or a subset of the tests is to be carried out after each MAJOR SERVICE ACTION that is made to existing equipment. This is done in order to facilitate verification of applicable safety and performance standards, regulations, and published and/or contractual specifications that influence the image quality, RADIATION OUTPUT and PATIENT positioning.

To maintain the homogeneity of this document with the other IEC standards addressing CT SCANNERS, the measuring methods and the terminology are taken as applicable from the CT safety standard IEC 60601-2-44:2009, IEC 60601-2-44:2009/AMD1:2012 and IEC 60601-2-44:2009/AMD2:2016.

Some provisions or statements in this document require additional information, which is presented in the annexes.

IEC 61223-3-5 is referenced by IEC 60601-2-44:2009 and IEC 60601-2-44:2009/AMD1:2012 using an undated reference. This can suggest the reference to change from IEC 61223-3-5:2004 to IEC 61223-3-5:2019 with the date of its publication. However, the IEC technical subcommittee 62B who prepared both standards does not intend this immediate change of reference. The IEC technical subcommittee 62B clearly recommends in the foreword of both standards the necessity for MANUFACTURERS and testing organizations for a transitional period to make products in accordance with the new requirements and to equip themselves for conducting new or revised tests. Therefore, the reference in IEC 60601-2-44 has to be seen as a dated reference towards IEC 61223-3-5:2004, for a transitional period of not less than 3 years from the date of publication of this document. The IEC technical subcommittee 62B intends to clarify this undated reference with the preparation of a new version 4 of IEC 60601-2-44.

EVALUATION AND ROUTINE TESTING IN MEDICAL IMAGING DEPARTMENTS –

Part 3-5: Acceptance and Constancy tests – Imaging performance of computed tomography X-ray equipment

1 Scope and object

This part of IEC 61223 applies to CT SCANNERS that conform to IEC 60601-2-44:2009, IEC 60601-2-44:2009/AMD1:2012 and IEC 60601-2-44:2009/AMD2:2016.

IEC 60601-2-44 and this document

- defines the essential parameters which describe the performance of CT SCANNERS with regard to image quality, RADIATION OUTPUT and PATIENT positioning; the list of parameters to be tested can be found in 4.3,
- defines the methods of testing the essential parameters, and
- evaluates compliance with the tolerances of the parameters SPECIFIED by the ACCOMPANYING DOCUMENTS.

The methods defined in IEC 60601-2-44 and this document rely on non-invasive measurements, using appropriate test equipment, performed during or after installation. Signed statements covering steps in the installation procedure can be used as part of the ACCEPTANCE TEST report.

This document applies to ACCEPTANCE TESTS and CONSTANCY TESTS on a CT SCANNER. The aim of the ACCEPTANCE TESTS is to verify compliance of the installation or MAJOR SERVICE ACTION with specifications affecting the image quality, RADIATION OUTPUT and PATIENT positioning. The CONSTANCY TESTS are performed to ensure that the functional performance of EQUIPMENT meets ESTABLISHED CRITERIA and to enable the early recognition of changes in the properties of components of the EQUIPMENT, and to verify compliance with specifications affecting the image quality, RADIATION OUTPUT and PATIENT positioning.

This document also contains requirements associated with ACCEPTANCE TEST and CONSTANCY TEST for the ACCOMPANYING DOCUMENTS of the CT SCANNER.

This document does not apply to

- aspects of mechanical and electrical safety, and
- aspects of mechanical, electrical and software performance, unless they are essential for performing the ACCEPTANCE TESTS and CONSTANCY TESTS, and are directly affecting image quality, RADIATION OUTPUT and PATIENT positioning.

NOTE 1 If a user of this document wishes to apply this document to CT SCANNERS that were designed to comply with editions of IEC 60601-2-44:2009 and earlier, understanding and adjustment for the different definitions that have been used for $CTDI_{vol}$ is critical. Additionally, the ACCOMPANYING DOCUMENTS for CT scanners that were designed and manufactured to these older editions can be referenced to obtain applicable specifications.

NOTE 2 It is possible the accompanying documents that were compiled in accordance with IEC 60601-2-44:2009 and IEC 60601-2-44:2009/AMD1:2012 or IEC 60601-2-44:2009, IEC 60601-2-44:2009/AMD1:2012 and IEC 60601-2-44:2009/AMD2:2016 do not include all the needed content and specifications identified in this document prior to the completion of the transition period to this document.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*
IEC 60601-1:2005/AMD1:2012

IEC 60601-2-44:2009, *Medical electrical equipment – Part 2-44: Particular requirements for the basic safety and essential performance of X-ray equipment for computed tomography*
IEC 60601-2-44:2009/AMD1:2012
IEC 60601-2-44:2009/AMD2:2016

IEC TR 60788:2004, *Medical electrical equipment – Glossary of defined terms*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in IEC TR 60788, IEC 60601-2-44:2009, IEC 60601-2-44:2009/AMD1:2012 and IEC 60601-2-44:2009/AMD2:2016 and the following apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <http://www.iso.org/obp>

NOTE 1 Terms printed in SMALL CAPITALS are used in accordance with their definitions in the documents referred to in the Index of defined terms at the end of this document.

NOTE 2 Attention is drawn to the fact that in cases where the concept addressed is not strongly confined to the definition given in one of the publications listed above, a corresponding term is printed in lower case letters.

NOTE 3 Associated conditions qualifying the usage of certain terms are given below.

3.1

ACCEPTANCE TEST

test performed after new equipment has been installed, or MAJOR SERVICE ACTIONS have been made to existing equipment, in order to verify that the functional performance of equipment meets ESTABLISHED CRITERIA from the MANUFACTURER, contractual specifications, and/or requirements of this document

Note 1 to entry: The ESTABLISHED CRITERIA verified are specifications affecting the image quality, RADIATION OUTPUT and PATIENT positioning. Additionally, during or immediately after the ACCEPTANCE TEST, the BASELINE VALUES for CONSTANCY TEST are established.

3.2

CONSTANCY TEST

test performed to verify that the functional performance of EQUIPMENT meets ESTABLISHED CRITERIA and to enable the early recognition of changes in the properties of components of the EQUIPMENT

Note 1 to entry: The test verifies conformance with specifications affecting the image quality, radiation output and PATIENT positioning.

3.3

CT CONDITIONS OF OPERATION

selectable parameters governing the operation of a CT SCANNER

Note 1 to entry: Examples of such conditions include NOMINAL TOMOGRAPHIC SECTION THICKNESS, CT PITCH FACTOR, FILTRATION, peak X-RAY TUBE VOLTAGE and either X-RAY TUBE CURRENT and LOADING TIME or CURRENT TIME PRODUCT.

Note 2 to entry: Some CT CONDITIONS OF OPERATION may vary during the exposure.

Note 3 to entry: CT CONDITIONS OF OPERATION include parameters that are derived by the system from the user-selectable parameters.

[SOURCE: IEC 60601-2-44:2009 and IEC 60601-2-44:2009/AMD1:2012, 201.3.202]

3.4

CT DOSE INDEX 100

$CTDI_{100}$

integral of the DOSE PROFILE representative of a single axial scan along a line perpendicular to the TOMOGRAPHIC PLANE divided by $N \times T$ according to the following:

for $N \times T$ less than or equal to 40 mm

$$CTDI_{100} = \int_{-50 \text{ mm}}^{+50 \text{ mm}} \frac{D(z)}{N \times T} dz$$

for $N \times T$ greater than 40 mm (all CT CONDITIONS OF OPERATION except collimation are kept the same for these measurements)

$$CTDI_{100} = \int_{-50 \text{ mm}}^{+50 \text{ mm}} \frac{D_{\text{Ref}}(z)}{(N \times T)_{\text{Ref}}} dz \times \frac{CTDI_{\text{free air}, N \times T}}{CTDI_{\text{free air}, \text{Ref}}}$$

where

$D(z)$ is the DOSE PROFILE representative of a single axial scan along a line z perpendicular to the TOMOGRAPHIC PLANE, where dose is reported as ABSORBED DOSE in air and is evaluated within a polymethylmethacrylate (PMMA) dosimetry PHANTOM (see IEC 60601-2-44:2009 and IEC 60601-2-44:2009/AMD1:2012, 203.108);

$(N \times T)_{\text{Ref}}$ is a SPECIFIC $N \times T$ of 20 mm or the largest $N \times T$ available not greater than 20 mm;

$D_{\text{Ref}}(z)$ is the DOSE PROFILE representative of a single axial scan along a line z perpendicular to the TOMOGRAPHIC PLANE, where dose is reported as ABSORBED DOSE in air and is evaluated within a polymethylmethacrylate (PMMA) dosimetry PHANTOM (see IEC 60601-2-44:2009 and IEC 60601-2-44:2009/AMD1:2012, 203.108) for $(N \times T)_{\text{Ref}}$;

$CTDI_{\text{free air}, N \times T}$ is the $CTDI_{\text{free air}}$ (IEC 60601-2-44:2009 and IEC 60601-2-44:2009/AMD1:2012, 201.3.215) for a SPECIFIC value of $N \times T$;

$CTDI_{\text{free air}, \text{Ref}}$ is the $CTDI_{\text{free air}}$ (IEC 60601-2-44:2009 and IEC 60601-2-44:2009/AMD1:2012, 201.3.215) for $(N \times T)_{\text{Ref}}$;

N is the number of TOMOGRAPHIC SECTIONS produced in a single axial scan of the X-ray source;

T is the NOMINAL TOMOGRAPHIC SECTION THICKNESS

Note 1 to entry: The dose is reported as ABSORBED DOSE to air, but for practical purposes the evaluation of ABSORBED DOSE to air within a PMMA dosimetry PHANTOM is well approximated by measurement of the AIR KERMA.

Note 2 to entry: This definition assumes that the DOSE PROFILE is centred on $z = 0$.

Note 3 to entry: A single axial scan is typically a 360° rotation of the X-ray source.

Note 4 to entry: When the TOMOGRAPHIC SECTIONS overlap, for example in CT SCANNERS with a "sliding FOCAL SPOT", the denominator of the integral needs to be replaced by the total nominal width along z of overlapping

TOMOGRAPHIC SECTIONS. For example, if the percentage of overlap is 50 %, then the denominator would be replaced by $0,5 \times N \times T$.

Note 5 to entry: Typically the z-axis is the axis of rotation.

Note 6 to entry: See Annex CC of IEC 60601-2-44:2009 and IEC 60601-2-44:2009/AMD1:2012 for explanation.

[SOURCE: IEC 60601-2-44:2009 and IEC 60601-2-44:2009/AMD1:2012, 201.3.203, modified – In the term, "computed tomography" has been replaced by "CT". Note 6 has been deleted, and Note 7 renumbered Note 6.]

3.5

CT DOSE INDEX MEASURED FREE-IN-AIR

$CTDI_{\text{free air}}$

integral of the DOSE PROFILE representative of a single axial scan along a line through the ISOCENTRE and perpendicular to the TOMOGRAPHIC PLANE divided by $N \times T$ according to the following

$$CTDI_{\text{free air}} = \int_{-L/2}^{+L/2} \frac{D(z)}{N \times T} dz$$

where

$D(z)$ is the DOSE PROFILE representative of a single axial scan along a line z through ISOCENTRE and perpendicular to the TOMOGRAPHIC PLANE, where dose is reported as ABSORBED DOSE in air and is evaluated free-in-air in the absence of a PHANTOM and the PATIENT SUPPORT;

N is the number of TOMOGRAPHIC SECTIONS produced in a single axial scan of the X-ray source;

T is the NOMINAL TOMOGRAPHIC SECTION THICKNESS;

L is at least $(N \times T) + 40$ mm, but not less than 100 mm

Note 1 to entry: This definition assumes that the DOSE PROFILE is centred on $z = 0$.

Note 2 to entry: When the TOMOGRAPHIC SECTIONS overlap, for example in CT SCANNERS with a "z-flying FOCAL SPOT", the denominator of the integral needs to be replaced by the total nominal width along z of overlapping TOMOGRAPHIC SECTIONS. For example, if the percentage of overlap is 50 %, then the denominator would be replaced by $0,5 \times N \times T$.

Note 3 to entry: Typically, a RADIATION DETECTOR of length L or longer is used. Annex DD of IEC 60601-2-44:2009, IEC 60601-2-44:2009/AMD1:2016 and IEC 60601-2-44:2009/AMD2:2016 provides an example for alternate measurements.

[SOURCE: IEC 60601-2-44:2009 and IEC 60601-2-44:2009/AMD1:2012, 201.3.215, modified – In the term, "computed tomography" has been replaced by "CT".]

3.6

CT NUMBER

number used to represent the mean X-ray ATTENUATION associated with each elemental area of the COMPUTED TOMOGRAPHY image

Note 1 to entry: The CT NUMBER is normally expressed in Hounsfield units. MEASURED VALUES of the linear ATTENUATION coefficients are transformed into CT NUMBERS using the international Hounsfield scale, using the expression:

$$CT\ NUMBER = \frac{\mu_{\text{material}} - \mu_{\text{water}}}{\mu_{\text{water}}} \times 1\ 000$$

where

μ is the linear ATTENUATION coefficient.

Note 2 to entry: The CT NUMBER scale is defined so that water has a value of 0 and air a value of -1 000 (μ_{air} is assumed to be 0).

3.7

CT PITCH FACTOR

in helical scanning the ratio of the PATIENT SUPPORT travel Δd along the z direction per rotation of the X-ray source divided by the product of the NOMINAL TOMOGRAPHIC SECTION THICKNESS T and the number of TOMOGRAPHIC SECTIONS N :

$$CT \text{ pitch factor} = \frac{\Delta d}{N \times T}$$

where

Δd is the PATIENT SUPPORT travel along the z -direction per rotation of the X-RAY SOURCE;

T is the NOMINAL TOMOGRAPHIC SECTION THICKNESS;

N is the number of TOMOGRAPHIC SECTIONS produced in a single axial scan of the X-RAY SOURCE.

Note 1 to entry: Although the CT PITCH FACTOR is associated with helical scanning, its definition refers to parameters T and N that are defined only for axial scanning. This definition of CT PITCH FACTOR presumes that these axial-scanning parameters T and N correspond to the same collimation and active-detector configuration as that of the helical scanning for which the CT PITCH FACTOR is being evaluated.

Note 2 to entry: When the TOMOGRAPHIC SECTIONS overlap, for example in CT SCANNERS with a "z-flying FOCAL SPOT", the denominator of the integral needs to be replaced by the total nominal width along z of overlapping TOMOGRAPHIC SECTIONS. For example, if the percentage of overlap is 50 %, then the denominator would be replaced by $0,5 \times N \times T$.

Note 3 to entry: CT PITCH FACTOR will be a function of time when Δd is variable during the exposure.

Note 4 to entry: The terms "helical" is used in this document as a synonym for the term "spiral".

[SOURCE: IEC 60601-2-44:2009 and IEC 60601-2-44:2009/AMD1:2012, 201.3.204, modified – Note 2 has been replaced by a new note.]

3.8

CT SCANNER

X-RAY EQUIPMENT intended to generate cross-sectional images of the body by computer reconstruction of X-ray transmission data obtained at different angles, which may include signal analysis and display equipment, PATIENT SUPPORT, support parts and ACCESSORIES

Note 1 to entry: The scope of IEC 60601-2-44:2009 is limited to CT SCANNERS intended to be used for both head and body imaging, characterized by an ENCLOSURE of the X-ray source(s) and imaging detector(s) in a common protective cover in the shape of a toroid.

Note 2 to entry: Secondary imaging processing is not included in the scope of IEC 60601-2-44:2009.

[SOURCE: IEC 60601-2-44:2009, 201.3.201]

3.9

DOSE PROFILE

representation of the dose as a function of position along a line

[SOURCE: IEC 60601-2-44:2009, 201.3.205]

3.10

FULL WIDTH AT HALF-MAXIMUM

FWHM

interval parallel to the abscissa between the points on a curve with the value of one-half of the maximum of the curve

3.11

IMAGE DISPLAY DEVICE

device capable of displaying images from an input signal provided by an imaging system

3.12**LOW CONTRAST RESOLUTION**

smallest size that can be individually resolved at a given level of contrast and for a SPECIFIED shape from a uniform background

3.13**MAJOR SERVICE ACTION**

service action that may significantly affect RADIATION OUTPUT, image quality, or PATIENT positioning, and requires an ACCEPTANCE TEST as described in the ACCOMPANYING DOCUMENTS

3.14**MEAN CT NUMBER**

mean value of the CT NUMBERS of all pixels within a certain defined REGION OF INTEREST

3.15**NOISE**

variation of CT NUMBERS from a mean value in a defined area in the image of a uniform substance

Note 1 to entry: The magnitude of NOISE is indicated by the standard deviation of the CT NUMBERS of a uniform substance in the REGION OF INTEREST.

3.16**NOMINAL TOMOGRAPHIC SECTION THICKNESS**

in CT SCANNERS, the TOMOGRAPHIC SECTION THICKNESS which is selected and indicated on the CONTROL PANEL

Note 1 to entry: The RECONSTRUCTED SECTION THICKNESS might or might not be equal to the NOMINAL TOMOGRAPHIC SECTION THICKNESS.

[SOURCE: IEC 60601-2-44:2009, 201.3.206, modified – Note 1 to entry has been rephrased.]

3.17**RECONSTRUCTED SECTION THICKNESS**

FULL WIDTH AT HALF MAXIMUM of the SENSITIVITY PROFILE of a reconstructed image

3.18**REGION OF INTEREST****ROI**

localized part of an image, of defined shape and dimension, which is of particular interest at a given time

Note 1 to entry: This note applies to the French language only.

3.19**SENSITIVITY PROFILE**

relative response of a system for COMPUTED TOMOGRAPHY as a function of position along a line perpendicular to the TOMOGRAPHIC PLANE

[SOURCE: IEC 60601-2-44:2009, 201.3.207]

3.20**SPATIAL RESOLUTION**

<CT SCANNERS> ability to resolve spatially adjacent objects in the image, when the difference in ATTENUATION between the objects and the background is large compared to NOISE


Note 1 to entry: Normally, a difference in ATTENUATION coefficient between the object and the background resulting in a difference of the respective CT NUMBERS of several hundred Hounsfield units is regarded as large enough.

NOTE 2 to entry: SPATIAL RESOLUTION is also known as "high-contrast resolution" and is best described by the MODULAR TRANSFER FUNCTION.

3.21

TOMOGRAPHIC PLANE

geometric plane perpendicular to the axis of rotation at the centre of the X-RAY FIELD in z (see Figure 1)



Key

- 1 TOMOGRAPHIC PLANE
- 2 PHANTOM

Figure 1 – Coordinate system

[SOURCE: IEC 60601-2-44:2009, 201.3.208, modified – Figure 1 has been added, and the parenthesis replaced in the definition.]

3.22

TOMOGRAPHIC SECTION

for CT SCANNERS with a single detector row, the volume over which TRANSMISSION data of X-RADIATION are collected in a single axial scan; for CT SCANNERS with multiple detector rows along the z-axis, the volume over which data are collected by a single acquisition channel representing a single detector row or a selected grouping of detector rows

[SOURCE: IEC 60601-2-44:2009, 201.3.209]

3.23

TOMOGRAPHIC SECTION THICKNESS

FULL WIDTH AT HALF MAXIMUM of the SENSITIVITY PROFILE taken at the ISOCENTRE of a TOMOGRAPHIC SECTION

[SOURCE: IEC 60601-2-44:2009, 2.3.210]

3.24

UNIFORMITY

consistency of the CT NUMBERS of the image of a homogeneous material across the scan field

3.25**VOLUME $CTDI_w$** $CTDI_{vol}$

a) for axial scanning

$$CTDI_{vol} = \frac{N \times T}{\Delta d} CTDI_w$$

where

N is the number of TOMOGRAPHIC SECTIONS produced in a single axial scan of the X-ray source;

T is the NOMINAL TOMOGRAPHIC SECTION THICKNESS;

Δd is the PATIENT SUPPORT travel in z-direction between consecutive scans

Note 1 to entry: For the selected CT CONDITIONS OF OPERATION, but irrespective of any scanning length that may be used clinically, the VOLUME $CTDI_w$ ($CTDI_{vol}$) is an index of dose based on a convention of a 100 mm range of integration along the z-axis. For axial scanning, $CTDI_{vol}$ corresponds to the average dose that would accrue in the PHANTOM central section of volume equal to the cross-sectional area $\times \Delta d$.

Note 2 to entry: For axial scanning with a total table travel much less than $N \times T$, $CTDI_{vol}$ as defined overestimates the average dose that would accrue in the PHANTOM central section of volume equal to the cross-sectional area $\times \Delta d$.

b) for helical scanning

$$CTDI_{vol} = \frac{CTDI_w}{CT \text{ pitch factor}}$$

Note 3 to entry: CT PITCH FACTOR will be a function of time when Δd is variable during the exposure.

Note 4 to entry: For the selected CT CONDITIONS OF OPERATION, but irrespective of any scanning length that may be used clinically, the VOLUME $CTDI_w$ ($CTDI_{vol}$) is an index of dose based on a convention of 100 mm range of integration along the z-axis. For helical scanning, $CTDI_{vol}$ corresponds to the average dose that would accrue in the centre of a 100 mm scan length.

Note 5 to entry: For helical scanning, when the product of a small number of rotations times the table travel per rotation is much less than $N \times T$, $CTDI_{vol}$ as defined overestimates the average dose that would accrue in the centre of a 100 mm scan length within a PHANTOM

c) for scanning without movement of the PATIENT SUPPORT

$$CTDI_{vol} = n \times CTDI_w$$

where n is equal to the number of rotations

Note 6 to entry: c) includes situations where the PATIENT SUPPORT may be moved manually, for example, during an interventional procedure.

Note 7 to entry: For scanning without movement of the PATIENT SUPPORT and for situations where the PATIENT SUPPORT may be moved manually, this definition overestimates the dose as it includes assumed scatter contribution from adjacent slices.

Note 8 to entry: For scanning without movement of the PATIENT SUPPORT, $CTDI_{vol}$ corresponds to the dose that would accrue in the PHANTOM central section of volume equal to the cross-sectional area $\times N \times T$ were there n congruent sequences of contiguous scanning, each sequence of length 100 mm.

d) for axial scanning without gaps and helical scanning, both involving back-and-forth PATIENT SUPPORT movement between two positions (shuttle mode)

$$CTDI_{vol} = n \frac{N \times T}{(N \times T) + R} CTDI_w$$

where

N is the number of TOMOGRAPHIC SECTIONS produced in a single axial scan of the X-ray source;

T is the NOMINAL TOMOGRAPHIC SECTION THICKNESS;


n is equal to the total number of rotations for the entire scan series;

R is the distance between the two positions;

$CTDI_W$ is the WEIGHTED $CTDI_{100}$.

Note 9 to entry: See Figure 2.

Note 10 to entry: $CTDI_W$ is evaluated as the time weighed $CTDI_W$ reflecting the varying CT CONDITIONS OF OPERATION.



F FOCAL SPOT

I ISOCENTRE

Z z-direction

Figure 2 – Illustration of $N \times T$, R and $(N \times T) + R$

[SOURCE: IEC 60601-2-44:2009 and IEC 60601-2-44:2009/AMD1:2012, 201.3.212, modified – The notes have been renumbered.]

3.26

WEIGHTED $CTDI_{100}$

$CTDI_W$

value defined as

$$CTDI_W = \frac{1}{3}CTDI_{100(\text{centre})} + \frac{2}{3}CTDI_{100(\text{peripheral})}$$

where

$CTDI_{100(\text{centre})}$ is the value of $CTDI_{100}$ measured in the centre of the CT dosimetry PHANTOM (IEC 60601-2-44:2009 203.108), and

$CTDI_{100(\text{peripheral})}$ is the average of the four values of $CTDI_{100}$ measured around the dosimetry PHANTOM periphery, according to IEC 60601-2-44:2009 203.109.1 a) 4)

[SOURCE: IEC 60601-2-44:2009, definition 201.3.211, modified – In the key of $CTDI_{100(\text{centre})}$, the parenthesis has been added. In the key of $CTDI_{100(\text{peripheral})}$, the reference "203.109.1 a) 2) and 3)" has been replaced by "203.109.1 a) 4)".]