



BSI Standards Publication

Ultrasonics — Hydrophones

Part 3: Properties of hydrophones for ultrasonic fields

National foreword

This British Standard is the UK implementation of EN IEC 62127-3:2023. It is identical to IEC 62127-3:2022. It supersedes BS EN 62127-3:2007+A1:2013, which is withdrawn.

The UK participation in its preparation was entrusted to Technical Committee EPL/87, Ultrasonics.

A list of organizations represented on this committee can be obtained on request to its committee manager.

Contractual and legal considerations

This publication has been prepared in good faith, however no representation, warranty, assurance or undertaking (express or implied) is or will be made, and no responsibility or liability is or will be accepted by BSI in relation to the adequacy, accuracy, completeness or reasonableness of this publication. All and any such responsibility and liability is expressly disclaimed to the full extent permitted by the law.

This publication is provided as is, and is to be used at the recipient's own risk.

The recipient is advised to consider seeking professional guidance with respect to its use of this publication.

This publication is not intended to constitute a contract. Users are responsible for its correct application.

© The British Standards Institution 2023
Published by BSI Standards Limited 2023

ISBN 978 0 539 03740 1
ICS 17.140.50

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 28 February 2023.

Amendments/corrigenda issued since publication

Date	Text affected
------	---------------

English Version

Ultrasonics - Hydrophones - Part 3: Properties of hydrophones
for ultrasonic fields
(IEC 62127-3:2022)

Ultrasons - Hydrophones - Partie 3: Propriétés des
hydrophones pour les champs ultrasoniques
(IEC 62127-3:2022)

Ultraschall - Hydrophone - Teil 3: Eigenschaften von
Hydrophonen zur Verwendung in Ultraschallfeldern
(IEC 62127-3:2022)

This European Standard was approved by CENELEC on 2023-01-23. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.



European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

European foreword

The text of document 87/818/FDIS, future edition 2 of IEC 62127-3, prepared by IEC/TC 87 "Ultrasonics" was submitted to the IEC-CENELEC parallel vote and approved by CENELEC as EN IEC 62127-3:2023.

The following dates are fixed:

- latest date by which the document has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2023-10-23
- latest date by which the national standards conflicting with the document have to be withdrawn (dow) 2026-01-23

This document supersedes EN 62127-3:2007 and all of its amendments and corrigenda (if any).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC shall not be held responsible for identifying any or all such patent rights.

Any feedback and questions on this document should be directed to the users' national committee. A complete listing of these bodies can be found on the CENELEC website.

Endorsement notice

The text of the International Standard IEC 62127-3:2022 was approved by CENELEC as a European Standard without any modification.

In the official version, for Bibliography, the following notes have to be added for the standard indicated:

IEC 60565-1 NOTE Approved as EN IEC 60565-1

IEC 60565-2 NOTE Approved as EN IEC 60565-2

Annex A (normative)

Normative references to international publications with their corresponding European publications

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE 1 Where an International Publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

NOTE 2 Up-to-date information on the latest versions of the European Standards listed in this annex is available here: www.cenelec.eu.

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
IEC 62127-1	-	Ultrasonics - Hydrophones - Part 1: Measurement and characterization of medical ultrasonic fields	EN IEC 62127-1	-
IEC 62127-2	-	Ultrasonics - Hydrophones - Part 2: Calibration for ultrasonic fields up to 40 MHz	EN 62127-2	-

INTERNATIONAL STANDARD

NORME INTERNATIONALE



**Ultrasonics – Hydrophones –
Part 3: Properties of hydrophones for ultrasonic fields**

**Ultrasons – Hydrophones –
Partie 3: Propriétés des hydrophones pour les champs ultrasoniques**

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

COMMISSION
ELECTROTECHNIQUE
INTERNATIONALE

ICS 17.140.50

ISBN 978-2-8322-6210-8

**Warning! Make sure that you obtained this publication from an authorized distributor.
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.**

CONTENTS

FOREWORD	4
INTRODUCTION	6
1 Scope	7
2 Normative references	7
3 Terms, definitions and symbols	7
4 List of symbols	11
5 Hydrophone characteristics	11
5.1 General	11
5.2 Basic information	11
5.3 Sensitivity	12
5.4 Frequency response	13
5.4.1 Stated frequency band	13
5.4.2 Frequency dependence	13
5.5 Directional response	13
5.5.1 General	13
5.5.2 Determination of the directional response	13
5.5.3 Symmetry of directional response	14
5.6 Effective hydrophone size	15
5.6.1 General	15
5.6.2 Model of directional response	15
5.6.3 Fitting of experimental response to theoretical predictions	16
5.7 Dynamic range, linearity and electromagnetic interference	16
5.8 Electric output characteristics	17
5.8.1 Hydrophone without pre-amplifier	17
5.8.2 Hydrophone assembly	18
5.8.3 Output lead configuration	18
5.9 Environmental aspects	18
5.9.1 Temperature range	18
5.9.2 Water tightness	18
5.9.3 Water properties and incompatible materials	18
5.9.4 Exposed material	18
5.10 Guidance manual	19
5.11 List of hydrophone characteristics	19
Annex A (informative) Examples of information on hydrophone properties	20
A.1 General	20
A.2 Basic information	20
A.3 Sensitivity and frequency response	20
A.4 Directional response	23
A.5 Effective dimension	23
A.6 Dynamic range, linearity and electromagnetic interference	25
A.6.1 Lower dynamic limit	25
A.6.2 Upper dynamic limit	25
A.7 Electric output characteristics	25
A.8 Environmental aspects	25
Annex B (informative) Rationale	27

B.1	General.....	27
B.2	Changes to the determination of directional response	27
B.3	Changes to the determination of effective radius.....	27
Annex C (informative)	Membrane hydrophone directivity model	29
C.1	General.....	29
C.2	Details of model.....	29
Bibliography		30
Figure A.1	– Frequency response of 0,2 mm needle hydrophone in the range 1 MHz to 40 MHz	21
Figure A.2	– Frequency response of 0,2 mm needle hydrophone in the range 100 kHz to 1 MHz	22
Figure A.3	– Directional response of 0,2 mm needle hydrophone	23
Figure A.4	– Effective radius of 0,2 mm needle hydrophone.....	24
Figure A.5	– Comparison of modelled and experimentally derived directional response	24
Table A.1	– Example of basic information for 0,2 mm needle hydrophone assembly	20

INTERNATIONAL ELECTROTECHNICAL COMMISSION

ULTRASONICS – HYDROPHONES –

Part 3: Properties of hydrophones for ultrasonic fields

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as “IEC Publication(s)”). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

IEC 62127-3 has been prepared by IEC technical committee 87: Ultrasonics. It is an International Standard.

This second edition cancels and replaces the first edition published in 2007 and Amendment 1:2013. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition.

- a) The upper frequency limit of 40 MHz has been removed.
- b) Hydrophone sensitivity definitions have been changed to recognize sensitivities as complex-valued quantities.
- c) Procedures to determine the effective hydrophone size have been changed according to the rationale outlined in Annex B.
- d) Requirements on the frequencies for which the effective hydrophone size shall be provided have been changed to achieve practicality for increased frequency bands.
- e) The new Annex B and Annex C have been added.

f) Annex A has been updated to reflect the changes of the normative parts.

The text of this International Standard is based on the following documents:

Draft	Report on voting
87/818/FDIS	87/824/RVD

Full information on the voting for its approval can be found in the report on voting indicated in the above table.

The language used for the development of this International Standard is English.

This document was drafted in accordance with ISO/IEC Directives, Part 2, and developed in accordance with ISO/IEC Directives, Part 1 and ISO/IEC Directives, IEC Supplement, available at www.iec.ch/members_experts/refdocs. The main document types developed by IEC are described in greater detail at www.iec.ch/publications.

A list of all parts of IEC 62127 series, published under the general title *Ultrasonics – Hydrophones*, can be found on the IEC website.

NOTE Words in **bold** in the text are defined in Clause 3.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under webstore.iec.ch in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

IMPORTANT – The "colour inside" logo on the cover page of this document indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.

INTRODUCTION

The spatial and temporal distribution of acoustic pressure in an ultrasonic field in a liquid medium is commonly determined using miniature ultrasonic **hydrophones**. The properties of these **hydrophones** have been dealt with in a number of IEC standards in various aspects. The purpose of this document is to bring together all these specifications and to establish a common standard on the properties of ultrasonic **hydrophones**. The main **hydrophone** application in this context is the measurement of ultrasonic fields emitted by medical diagnostic equipment in water. Other medical applications are field measurements for therapy equipment such as that used in lithotripsy, high-intensity focused ultrasound (HIFU) and physiotherapy. **Hydrophones** are also used extensively in non-medical applications for both product development and quality control including:

- mapping of the ultrasound field within ultrasonic cleaning baths;
- characterization of acoustic fields used in transmission measurement systems (e.g. ultrasonic spectrometers, ultrasonic attenuation meters and velocimeters);
- characterization of acoustic fields used in reflection measurement systems (e.g. Doppler flowmeters).

While the term **hydrophone** can be used in a wider sense, it is understood here as referring to miniature piezoelectric **hydrophones**. It is this instrument type that is used today in various areas of ultrasonics and, in particular, to quantitatively characterize the field structure of medical diagnostic instruments. With regard to other pressure sensor types, such as those based on fibre optics, some of the requirements of this document are applicable to these as well but others are not. If in the future these other **hydrophone** types gain more importance in field measurement practice, their properties will have to be dealt with in a revised version of this document or in a separate one.

Underwater **hydrophones** as covered by IEC 60500, IEC 60565-1, and IEC 60565-2 are not included in this document, although there is an overlap in the frequency ranges. Underwater **hydrophones** are used in natural waters, even in the ocean, and this leads to different technical concepts and requirements. In addition, the main direction of acoustic incidence in underwater applications is at various angles and often at right angles to the **hydrophone axis**, whereas in this document it is assumed that the main direction of acoustic incidence is in the direction of the **hydrophone axis**.

Historically, ultrasonic **hydrophones** were used almost exclusively as amplitude sensors. However, the complex-valued nature of a **hydrophone's** system response function is well understood and IEC 62127-1:2022 makes use of this within the deconvolution procedures it contains. In this document, requirements are specified for the amplitude aspect of the **hydrophone** sensitivity and recommendations are provided for the phase aspect which can be derived either via calibration, or via calculation methods that are discussed in IEC 62127-1:2022.

ULTRASONICS – HYDROPHONES –

Part 3: Properties of hydrophones for ultrasonic fields

1 Scope

This part of IEC 62127 specifies relevant **hydrophone** characteristics.

This document is applicable to:

- **hydrophones** employing piezoelectric sensor elements, designed to measure the pulsed and continuous wave ultrasonic fields generated by ultrasonic equipment;
- **hydrophones** used for measurements made in water;
- **hydrophones** with or without an associated pre-amplifier.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 62127-1, *Ultrasonics – Hydrophones – Part 1: Measurement and characterization of medical ultrasonic fields*

IEC 62127-2, *Ultrasonics – Hydrophones – Part 2: Calibration for ultrasonic fields up to 40 MHz*

3 Terms, definitions and symbols

For the purposes of this document, the terms and definitions given in IEC 62127-1, IEC 62127-2 and the following apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <http://www.iso.org/obp>

3.1

acoustic pulse waveform

temporal waveform of the **instantaneous acoustic pressure** at a specified position in an acoustic field and displayed over a period sufficiently long to include all significant acoustic information in a single pulse or tone-burst, or one or more cycles in a continuous wave

Note 1 to entry: Temporal waveform is a representation (e.g. oscilloscope presentation or formula) of the **instantaneous acoustic pressure**.

[SOURCE: IEC 62127-1:2022, 3.1]