



BSI Standards Publication

Medical electrical equipment

Part 2-49: Particular requirements for the basic safety and essential performance of multifunction patient monitors

National foreword

This British Standard is the UK implementation of EN IEC 80601-2-49:2019+A1:2024. It is identical to IEC 80601-2-49:2018, incorporating amendment 1:2024. It supersedes BS EN IEC 80601-2-49:2019, which is withdrawn.

The start and finish of text introduced or altered by amendment is indicated in the text by tags. Tags indicating changes to IEC text carry the number of the IEC amendment. For example, text altered by IEC amendment 1 is indicated by A1 A1.

The UK participation in its preparation was entrusted to Technical Committee CH/62/4, Electromedical equipment.

A list of organizations represented on this committee can be obtained on request to its committee manager.

Contractual and legal considerations

This publication has been prepared in good faith, however no representation, warranty, assurance or undertaking (express or implied) is or will be made, and no responsibility or liability is or will be accepted by BSI in relation to the adequacy, accuracy, completeness or reasonableness of this publication. All and any such responsibility and liability is expressly disclaimed to the full extent permitted by the law.

This publication is provided as is, and is to be used at the recipient's own risk.

The recipient is advised to consider seeking professional guidance with respect to its use of this publication.

This publication is not intended to constitute a contract. Users are responsible for its correct application.

© The British Standards Institution 2024
Published by BSI Standards Limited 2024

ISBN 978 0 539 17216 4
ICS 11.040.55

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 31 October 2019.

Amendments/corrigenda issued since publication

Date	Text affected
31 October 2024	Implementation of IEC amendment 1:2024 with CENELEC endorsement A1:2024

English Version

**Medical electrical equipment - Part 2-49: Particular requirements
for the basic safety and essential performance of multifunction
patient monitoring equipment
(IEC 80601-2-49:2018)**

Appareils électromédicaux - Partie 2-49: Exigences
particulières pour la sécurité de base et les performances
essentielles des appareils de surveillance multifonction des
patients
(IEC 80601-2-49:2018)

Medizinische elektrische Geräte - Teil 2-49: Besondere
Festlegungen für die Sicherheit einschließlich der
wesentlichen Leistungsmerkmale von multifunktionalen
Patientenüberwachungsgeräten
(IEC 80601-2-49:2018)

This European Standard was approved by CENELEC on 2019-08-07. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.



European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

European foreword

The text of document 62D/1547/FDIS, future edition 1.0 of IEC 80601-2-49, prepared by SC 62D "Electromedical equipment" of IEC/TC 62 "Electrical equipment in medical practice" was submitted to the IEC-CENELEC parallel vote and approved by CENELEC as EN IEC 80601-2-49:2019.

The following dates are fixed:

- latest date by which the document has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2020-05-07
- latest date by which the national standards conflicting with the document have to be withdrawn (dow) 2022-08-07

This document supersedes EN 60601-2-49:2015.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC shall not be held responsible for identifying any or all such patent rights.

Endorsement notice

The text of the International Standard IEC 80601-2-49:2018 was approved by CENELEC as a European Standard without any modification.

In the official version, for Bibliography, the following notes have to be added for the standards indicated:

IEC 60601-2-16	NOTE	Harmonized as EN IEC 60601-2-16
ISO 80601-2-13	NOTE	Harmonized as EN ISO 80601-2-13
ISO 80601-2-56	NOTE	Harmonized as EN ISO 80601-2-56
ISO 80601-2-72	NOTE	Harmonized as EN ISO 80601-2-72
IEC 62366-1:2015	NOTE	Harmonized as EN 62366-1:2015 (not modified)

European foreword to amendment A1

The text of document 62D/2146/FDIS, future edition 1 of IEC 80601-2-49/AMD1, prepared by SC 62D "Particular medical equipment, software, and systems" of IEC/TC 62 "Medical equipment, software, and systems" was submitted to the IEC-CENELEC parallel vote and approved by CENELEC as EN IEC 80601-2-49:2019/A1:2024.

The following dates are fixed:

- latest date by which the document has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2025-10-31
- latest date by which the national standards conflicting with the document have to be withdrawn (dow) 2027-10-31

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC shall not be held responsible for identifying any or all such patent rights.

Any feedback and questions on this document should be directed to the users' national committee. A complete listing of these bodies can be found on the CENELEC website.

Endorsement notice

The text of the International Standard IEC 80601-2-49:2018/AMD1:2024 was approved by CENELEC as a European Standard without any modification.

Annex ZA

(normative)

Normative references to international publications with their corresponding European publications

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE 1 Where an International Publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

NOTE 2 Up-to-date information on the latest versions of the European Standards listed in this annex is available here: www.cenelec.eu.

The Annex ZA of EN 60601-1:2006 applies, except as follows:

<u>Publication</u> <i>Replacement</i>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
IEC 60601-1-2	2014	Medical electrical equipment - Part 1-2: General requirements for basic safety and essential performance - Collateral Standard: Electromagnetic disturbances - Requirements and tests	EN 60601-1-2	2015
+ A1	2020		+ A1	2021
IEC 60601-1-6	2010	Medical electrical equipment - Part 1-6: General requirements for basic safety and essential performance - Collateral standard: Usability	EN 60601-1-6	2010
+ A1	2013		+ A1	2015
+ A2	2020		+ A2	2021
IEC 60601-1-8	2006	Medical electrical equipment - Part 1-8: General requirements for basic safety and essential performance - Collateral standard: General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems	EN 60601-1-8	2007
+ A1	2012		+ A1	2013
-	-		+ AC	2014
+ A2	2020		+ A2	2021
IEC 60529	1989	Degrees of protection provided by enclosures (IP Code)	EN 60529	1991
			+EN 60529:1991/corrigendum May 1993	1993

EN IEC 80601-2-49:2019+A1:2024 (E)

<u>Publication Addition</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
IEC 60601-1	2005	Medical electrical equipment - Part 1: General requirements for basic safety and essential performance	EN 60601-1	2006
-	-		+ corrigendum Mar. 2010	
+ A1	2012		+ A1	2013
-	-		+ A12	2014
+ A2	2020		+ A2	2021
-	-		+ A13	2024
IEC 60601-1-11	2015	Medical electrical equipment - Part 1-11: General requirements for basic safety and essential performance - Collateral standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment	EN 60601-1-11	2015
+ 1	2020		+ 1	2021
IEC 60601-1-12	2014	Medical Electrical Equipment - Part 1-12: General requirements for basic safety and essential performance - Collateral Standard: Requirements for medical electrical equipment and medical electrical systems used in the emergency medical services environment	EN 60601-1-12	2015
+ 1	2020		+ 1	2020
IEC 60601-2-2	2017	Medical electrical equipment - Part 2-2: Particular requirements for the basic safety and essential performance of high frequency surgical equipment and high frequency surgical accessories	EN IEC 60601-2-2	2018
IEC 60601-2-27	2011	Medical electrical equipment - Part 2-27: Particular requirements for the basic safety and essential performance of electrocardiographic monitoring equipment	EN 60601-2-27	2014
IEC 60601-2-34	2011	Medical electrical equipment - Part 2-34: Particular requirements for the basic safety and essential performance of invasive blood pressure monitoring equipment	EN 60601-2-34	2014

CONTENTS

FOREWORD.....	4
INTRODUCTION.....	7
A1 INTRODUCTION to Amendment 1	7 A1
201.1 Scope, object and related standards	8
201.2 Normative references.....	10
201.3 Terms and definitions.....	11
201.4 General requirements	11
201.5 General requirements for testing ME EQUIPMENT	12
201.6 Classification of ME EQUIPMENT and ME SYSTEMS.....	12
201.7 ME EQUIPMENT identification, marking and documents	13
201.8 Protection against electrical HAZARDS from ME EQUIPMENT	14
201.9 Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS	16
201.10 Protection against unwanted and excessive radiation HAZARDS	16
201.11 Protection against excessive temperatures and other HAZARDS	16
201.12 Accuracy of controls and instruments and protection against hazardous outputs	17
201.13 HAZARDOUS SITUATIONS and fault conditions for ME EQUIPMENT	18
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS).....	18
201.15 Construction of ME EQUIPMENT	18
201.16 ME SYSTEMS	19
201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS	19
202 Electromagnetic disturbances – Requirements and tests.....	19
206 USABILITY	24
208 General requirements, tests and guidance for ALARM SYSTEMS IN MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS.....	24
Annexes	27
Annex AA (informative) Particular guidance and rationale	28
Bibliography.....	36
Index of defined terms used in this particular standard.....	37
Figure 201.101 – MULTIFUNCTION PATIENT MONITOR with single PATIENT circuit (6) with multiple PHYSIOLOGICAL MONITORING UNITS and multiple PATIENT circuits (7) each with a single PHYSIOLOGICAL MONITORING UNIT	15
Figure 202.101 – Test layout for conducted and radiated EMISSIONS and IMMUNITY test	20
Figure 202.102 – Test circuit for HF SURGICAL EQUIPMENT protection measurement according to 202.8.102 with PATIENT CONNECTIONS	22
Figure 202.103 – Test setup for HF SURGICAL EQUIPMENT protection measurement according to 202.8.102	23
Figure 202.104 – Test circuit for HF SURGICAL EQUIPMENT protection measurement according to 202.8.102 with non-conductive APPLIED PART.....	24
Figure AA.1 – Example of a pre-configured MULTIFUNCTION PATIENT MONITOR	29
Figure AA.2 – Example of a modular MULTIFUNCTION PATIENT MONITOR	29
Figure AA.3 – Example of a MULTIFUNCTION PATIENT MONITOR connected to a central station	29

Figure AA.4 – Example of a MULTIFUNCTION PATIENT MONITOR integrated into a ventilator	30
Figure AA.5 – Single PATIENT circuit with multiple PHYSIOLOGICAL MONITORING UNITS and PATIENT cables	32
Table 201.101 – ESSENTIAL PERFORMANCE requirements.....	12

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-49: Particular requirements for the basic safety and essential performance of multifunction patient monitors

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International standard IEC 80601-2-49 has been prepared by a Joint Working Group of IEC subcommittee 62D: Electromedical equipment, of IEC technical committee 62: Electrical equipment in medical practice, and of ISO subcommittee SC3: Lung ventilators and related equipment, of ISO technical committee 121: Anaesthetic and respiratory equipment.

This first edition cancels and replaces the second edition of IEC 60601-2-49, published in 2011. This edition constitutes a technical revision to align with the current edition and Amendment to IEC 60601-1, new versions of collateral standards and amendments thereto. Major changes are in Clause 208 because many of the former requirements are now addressed by IEC 60601-1-8.

It is published as a double logo standard.

The text of this standard is based on the following documents:

FDIS	Report on voting
62D/1547/FDIS	62D/1559/RVD

Full information on the voting for the approval of this document can be found in the report on voting indicated in the above table. In ISO, the standard has been approved by XXX P members out of YYY having cast a vote.

In this document, the following print types are used:

- requirements and definitions: roman type;
- *test specifications: italic type;*
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type;
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this document, the term

- "clause" means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- "subclause" means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this document are preceded by the term "Clause" followed by the clause number. References to subclauses within this particular standard are by number only.

In this document, the conjunctive "or" is used as an "inclusive or" so a statement is true if any combination of the conditions is true.

The verbal forms used in this document conform to usage described in Clause 7 of the ISO/IEC Directives, Part 2. For the purposes of this document, the auxiliary verb:

- "shall" means that compliance with a requirement or a test is mandatory for compliance with this document;
- "should" means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this document;
- "may" is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

A list of all parts of the 80601 International Standard, published under the general title *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under "<http://webstore.iec.ch>" in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

NOTE The attention of users of this document is drawn to the fact that equipment manufacturers and testing organizations may need a transitional period following publication of a new, amended or revised IEC publication in which to make products in accordance with the new requirements and to equip themselves for conducting new or revised tests. It is the recommendation of the committees that the content of this publication be adopted for implementation nationally not earlier than 3 years from the date of publication.

INTRODUCTION

This particular standard concerns the BASIC SAFETY and ESSENTIAL PERFORMANCE of MULTIFUNCTION PATIENT MONITORS. It amends and supplements IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*, hereinafter referred to as the general standard.

The aim of this edition is to bring this particular standard up to date with reference to the edition 3.1 of the general standard and new versions of collateral standards and amendments thereto through technical changes.

The requirements of this particular standard take priority over those of the general standard.

A "Particular guidance and rationale" for the requirements of this particular standard is included in Annex AA. It is considered that knowledge of the reasons for these requirements will not only facilitate the proper application of the standard but will, in due course, expedite any revision necessitated by changes in clinical practice or as a result of developments in technology. However, this Annex AA does not form part of the requirements of this document.

INTRODUCTION to Amendment 1

At the October 2019 meeting of IEC SC 62D in Shanghai, China, the subcommittee discussed the need for administrative/technical changes to most 62D standards after completion of the amendment projects within the IEC 60601-1 series. Those projects were all completed and the amendments published in 2020.

The full list of IEC SC 62D documents that will be amended or revised may be found within the IEC document 62D/1792/DC. The results and comments on the DC may be found within 62D/1808/INF. The review report for this amendment is 62D/1835A/RR. 

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-49: Particular requirements for the basic safety and essential performance of multifunction patient monitors

201.1 Scope, object and related standards

Clause 1 of the general standard¹ applies, except as follows:

201.1.1 * Scope

Replacement:

This part of the 80601 International Standard applies to BASIC SAFETY and ESSENTIAL PERFORMANCE requirements of MULTIFUNCTION PATIENT MONITORS as defined in 201.3.201, hereafter referred to as ME EQUIPMENT or MEDICAL ELECTRICAL SYSTEMS. This particular standard applies to MULTIFUNCTION PATIENT MONITORS intended for use in professional healthcare facilities as well as in the EMERGENCY MEDICAL SERVICE ENVIRONMENT or the HOME HEALTHCARE ENVIRONMENT.

The scope of this document is restricted to ME EQUIPMENT or MEDICAL ELECTRICAL SYSTEMS intended for connection to a single PATIENT that has two or more PHYSIOLOGICAL MONITORING UNITS.

NOTE For purposes of this document, a pregnant mother and her fetus(es) are considered a single PATIENT.

This document does not specify requirements for individual PHYSIOLOGICAL MONITORING UNITS such as ECG, invasive pressure and pulse oximetry. The particular standards related to these PHYSIOLOGICAL MONITORING UNITS specify requirements from the perspective of stand-alone ME EQUIPMENT. This particular standard addresses the additional requirements related to MULTIFUNCTION PATIENT MONITORS. MULTIFUNCTION PATIENT MONITORS can be integrated into other ME EQUIPMENT or MEDICAL ELECTRICAL SYSTEMS. When this is the case, other relevant standards also apply.

EXAMPLE 1 MULTIFUNCTION PATIENT MONITOR incorporated into a critical care ventilator where ISO 80601-2-12 also applies.

EXAMPLE 2 MULTIFUNCTION PATIENT MONITOR incorporated into a homecare ventilator for dependent PATIENT where ISO 80601-2-72 also applies.

EXAMPLE 3 MULTIFUNCTION PATIENT MONITOR incorporated into anesthetic workstation where ISO 80601-2-13 also applies.

EXAMPLE 4 MULTIFUNCTION PATIENT MONITOR incorporated into haemodialysis equipment, IEC 60601-2-16 also applies.

This document does not apply to implantable parts of MULTIFUNCTION PATIENT MONITORS.

201.1.2 Object

Replacement:

The object of this particular standard is to establish BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for MULTIFUNCTION PATIENT MONITORS as defined in 201.3.201.

¹ [A1] The general standard is IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020. *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance.* [A1]