



BSI Standards Publication

Pressure regulators for use with medical gases

Part 1: Pressure regulators and pressure regulators with flow-metering devices

National foreword

This British Standard is the UK implementation of EN ISO 10524-1:2019+A1:2023. It is identical to ISO 10524-1:2018, incorporating amendment 1:2023. It supersedes BS EN ISO 10524-1:2019, which is withdrawn.

The start and finish of text introduced or altered by amendment is indicated in the text by tags. Tags indicating changes to ISO text carry the number of the ISO amendment. For example, text altered by ISO amendment 1 is indicated by **A1** **A1**.

The UK participation in its preparation was entrusted to Technical Committee CH/121/6, Medical gas supply systems.

A list of organizations represented on this committee can be obtained on request to its committee manager.

Contractual and legal considerations

This publication has been prepared in good faith, however no representation, warranty, assurance or undertaking (express or implied) is or will be made, and no responsibility or liability is or will be accepted by BSI in relation to the adequacy, accuracy, completeness or reasonableness of this publication. All and any such responsibility and liability is expressly disclaimed to the full extent permitted by the law.

This publication is provided as is, and is to be used at the recipient's own risk.

The recipient is advised to consider seeking professional guidance with respect to its use of this publication.

This publication is not intended to constitute a contract. Users are responsible for its correct application.

© The British Standards Institution 2023
Published by BSI Standards Limited 2023

ISBN 978 0 539 25069 5

ICS 11.040.10

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 28 February 2019.

Amendments/corrigenda issued since publication

Date	Text affected
30 November 2023	Implementation of ISO amendment 1:2023 with CEN endorsement A1:2023

English Version

Pressure regulators for use with medical gases - Part 1:
Pressure regulators and pressure regulators with flow-
metering devices (ISO 10524-1:2018)

Détendeurs pour l'utilisation avec les gaz
médicaux- Partie 1: Détendeurs et détendeurs-
débitmètres (ISO 10524-1:2018)

Druckminderer zur Verwendung mit medizinischen
Gasen - Teil 1: Druckminderer und Druckminderer
mit Durchflussmessgeräten (ISO 10524-1:2018)

This European Standard was approved by CEN on 13 December 2018.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

European foreword

This document (EN ISO 10524-1:2019) has been prepared by Technical Committee ISO/TC 121 "Anaesthetic and respiratory equipment" in collaboration with Technical Committee CEN/TC 215 "Respiratory and anaesthetic equipment" the secretariat of which is held by BSI.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by July 2019, and conflicting national standards shall be withdrawn at the latest by July 2019.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 10524-1:2006.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 10524-1:2018 has been approved by CEN as EN ISO 10524-1:2019 without any modification.

European foreword to amendment 1

This document (EN ISO 10524-1:2019/A1:2023) has been prepared by Technical Committee ISO/TC 121 "Anaesthetic and respiratory equipment" in collaboration with Technical Committee CEN/TC 215 "Respiratory and anaesthetic equipment" the secretariat of which is held by BSI.

This Amendment to the European Standard EN ISO 10524-1:2019 shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by May 2024, and conflicting national standards shall be withdrawn at the latest by May 2024.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

Any feedback and questions on this document should be directed to the users' national standards body/national committee. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

Endorsement notice

The text of ISO 10524-1:2018/Amd 1:2023 has been approved by CEN as EN ISO 10524-1:2019/A1:2023 without any modification.

Contents

Page

Foreword	v
Introduction	vi
1 Scope	1
2 Normative references	1
3 Terms and definitions	2
4 Nomenclature	4
5 General requirements	4
5.1 Safety	4
5.2 Usability	4
5.3 Alternative construction	4
5.4 Materials	4
6 Design requirements	5
6.1 General	5
6.2 Indicator for cylinder pressure or cylinder content	6
6.2.1 General	6
6.2.2 PRESSURE GAUGES, pressure indicators, and FLOWGAUGES	6
6.3 Integrated electronic device	6
6.4 Connections	6
6.4.1 Inlet connector	6
6.4.2 Outlet connectors	7
6.5 * Requirements for outlet pressure	7
6.5.1 PRESSURE OUTLET	7
6.5.2 FLOW OUTLET	8
6.6 Flow-metering device	8
6.7 Flow control and indication	8
6.8 Pressure-adjusting device	8
6.9 * Filtration	9
6.10 * PRESSURE-RELIEF DEVICE	9
6.11 Leakage	9
6.12 Mechanical strength	10
6.12.1 Resistance of the high-pressure side	10
6.12.2 Resistance of the low-pressure side to pneumatic pressure	10
6.12.3 Resistance of the low-pressure side to inlet pressure, P_1	10
6.13 * Resistance to ignition	10
6.14 Requirements for PRESSURE REGULATORS with FLOWMETERS	10
6.14.1 Calibration	10
6.14.2 ACCURACY OF FLOW	10
6.14.3 Stability of flow	11
6.14.4 Legibility	11
6.14.5 Flow control device	11
6.15 Requirements for PRESSURE REGULATORS fitted with FLOWGAUGES	11
6.15.1 Calibration	11
6.15.2 ACCURACY OF FLOW	11
6.15.3 Stability of flow	11
6.16 Requirements for PRESSURE REGULATORS fitted with fixed ORIFICES	12
6.16.1 Stability and ACCURACY OF FLOW	12
6.16.2 * Flow setting torque	12
6.16.3 Removal of a fixed ORIFICE	12
6.16.4 Legibility	12
6.17 Endurance	12
6.17.1 FLOW SELECTOR	12
6.17.2 PRESSURE REGULATOR	13

7	Construction requirements	13
7.1	* Cleanliness	13
7.2	Lubricants	13
7.3	Loosening torques	13
8	Test methods for type tests	14
8.1	General conditions	14
8.1.1	General	14
8.1.2	Ambient conditions	14
8.1.3	Test gas	14
8.1.4	Reference conditions	14
8.2	Test schedule	14
8.3	Test methods for outlet pressure	16
8.3.1	Test equipment	16
8.3.2	Test methods for determining outlet pressure limits for a PRESSURE REGULATOR fitted with a PRESSURE OUTLET	17
8.3.3	Test method for determining the outlet pressure limit for a PRESSURE REGULATOR fitted with a FLOW OUTLET	18
8.4	Test method for a PRESSURE-RELIEF DEVICE	18
8.5	Test methods for leakage	18
8.5.1	External leakage	18
8.5.2	Internal leakage	19
8.6	Test method for mechanical strength	19
8.6.1	Test method for the high-pressure side	19
8.6.2	Test method for resistance of the low-pressure side to pneumatic pressure	19
8.6.3	Test method for the resistance of the low pressure side to P_1	19
8.7	Test method for resistance to ignition	20
8.7.1	General	20
8.7.2	Test procedure for ADJUSTABLE and PRE-SET PRESSURE REGULATORS	20
8.8	Test method for ACCURACY OF FLOW of PRESSURE REGULATORS fitted with FLOWMETERS or FLOWGAUGES	20
8.9	Test method for the stability of flow of PRESSURE REGULATOR fitted with FLOWMETERS or FLOWGAUGES	20
8.10	Test method for stability and ACCURACY OF FLOW of PRESSURE REGULATORS fitted with fixed ORIFICES	21
8.11	Test method for flow setting and loosening torques	21
8.11.1	General	21
8.11.2	Test method for verifying no stable position between two settings	21
8.12	Test method for durability of markings and colour coding	21
8.13	*FLOW SELECTOR endurance test	21
8.14	PRESSURE REGULATOR endurance test	22
9	Marking, colour coding, and packaging	22
9.1	Marking	22
9.2	Colour coding	24
9.3	Packaging	24
10	Information to be supplied by the manufacturer	24
	Annex A (informative) Rationale	26
	Annex B (informative) Reported regional and national deviations of colour coding and nomenclature for medical gases	31
	Bibliography	33

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 121, *Anaesthetic and respiratory equipment*, Subcommittee SC 6, *Medical gas systems*.

This second edition cancels and replaces the first edition (ISO 10524-1:2006), which has been technically revised.

The main changes compared to the previous edition are as follows:

- the common requirements have been aligned with ISO 10524-2 and ISO 10524-3;
- this document has been restructured according to the new ISO template and associated renumbering;
- a complete schedule has been introduced;
- all type tests have been reviewed.

A list of all parts in the ISO 10524 series can be found on the ISO website.

Introduction

PRESSURE REGULATORS are used to reduce high cylinder pressure to a lower pressure suitable for use with medical equipment or for delivery of gas directly to a patient.

These functions cover a wide range of inlet and outlet pressures and flows which require specific design characteristics. It is important that the operating characteristics of PRESSURE REGULATORS are specified and tested in a defined manner.

A PRESSURE REGULATOR normally has coupled to it a device which controls the flow, such as a flow control device or a fixed ORIFICE. The flow can be indicated by a FLOWMETER or by a FLOWGAUGE.

It is essential that regular inspection and maintenance be undertaken to ensure that the PRESSURE REGULATOR continues to meet the requirements of this document.

This document pays particular attention to

- use of suitable materials,
- safety (mechanical strength, leakage, safe relief of excess pressure and resistance to ignition),
- GAS SPECIFICity,
- cleanliness,
- type testing,
- marking, and
- information supplied by the manufacturer.

[Annex A](#) contains rationale statements for some of the requirements of this document. The clauses and subclauses marked with an asterisk (*) after their number have corresponding rationale included to provide additional insight into the reasoning that led to the requirements and recommendations that have been incorporated into this document. It is considered that knowledge of the reasons for the requirements will not only facilitate the proper application of this document, but will expedite any subsequent revisions.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in [Annex A](#).

In this document, the following print types are used:

- requirements and definitions: roman type;
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type;
- *test specifications: italic type;*
- TERMS DEFINED IN CLAUSE 3 OF THIS DOCUMENT OR AS NOTED: SMALL CAPITALS TYPE.

Pressure regulators for use with medical gases —

Part 1:

Pressure regulators and pressure regulators with flow-metering devices

1 Scope

This document specifies the design, construction, type testing, and marking requirements for PRESSURE REGULATORS (as defined in [3.18](#)) intended for the administration of medical gases and their mixtures in the treatment, management, diagnostic evaluation and care of patients or for gases used for driving surgical tools.

Examples of gases include oxygen, medical air and oxygen/nitrous oxide mixtures.

This document applies to PRESSURE REGULATORS:

- a) intended to be connected to cylinders by the operator;
- b) with integral flow-metering devices intended to be connected to cylinders by the operator;
- c) that are an integral part of medical equipment (e.g. anaesthetic workstations, lung ventilators, resuscitators).

A PRESSURE REGULATOR can be provided with PRESSURE OUTLET or FLOW OUTLET, and can be adjustable or pre-set.

PRESSURE REGULATORS are intended to be fitted to refillable cylinders with a WORKING PRESSURE up to 30 000 kPa (300 bar) and can be provided with devices which control and measure the flow of the medical gas delivered.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 32, *Gas cylinders for medical use — Marking for identification of content*

ISO 407, *Small medical gas cylinders — Pin-index yoke-type valve connections*

ISO 5145, *Cylinder valve outlets for gases and gas mixtures — Selection and dimensioning*

ISO 7000, *Graphical symbols for use on equipment — Registered symbols*

ISO 9170-1, *Terminal units for medical gas pipeline systems — Part 1: Terminal units for use with compressed medical gases and vacuum*

ISO 10297:2014, *Gas cylinders — Cylinder valves — Specification and type testing*

ISO 14971, *Medical devices — Application of risk management to medical devices*

ISO 15001:2010, *Anaesthetic and respiratory equipment — Compatibility with oxygen*

EN 837-1, *PRESSURE GAUGES — Part 1: Bourdon tube PRESSURE GAUGES — Dimensions, metrology, requirements and testing*

EN 13544-2, *Respiratory therapy equipment — Part 2: Tubing and connectors*

IEC 60601-1, *Medical electrical equipment — Part 1: General requirements for basic safety and essential performance*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <https://www.electropedia.org/>

3.1

ACCURACY OF FLOW

difference between the indicated value and the actual value of the flow

Note 1 to entry: It is expressed in per cent.

3.2

ADJUSTABLE PRESSURE REGULATOR

PRESSURE REGULATOR (3.18) that is provided with a means of operator adjustment of the outlet pressure

3.3

CONTENT INDICATOR

device that displays the amount of gas remaining in the cylinder

Note 1 to entry: The content can be expressed either in percentage of content or cylinder pressure.

3.4

FLOWGAUGE

device that measures pressure and which is calibrated in units of flow

Note 1 to entry: The FLOWGAUGE does not measure flow. It indicates flow by measuring the pressure upstream of a fixed *ORIFICE* (3.13).

3.5

FLOWMETER

device that measures and indicates the flow of a specific gas or gas mixture

3.6

FLOW SELECTOR

means for selecting the flow and indicating the flow selected

3.7

FLOW OUTLET

outlet intended to deliver a controlled flow of gas

3.8

GAS-SPECIFIC

quality of having characteristics that prevent connection between different gas services

3.9

GAS-SPECIFIC CONNECTION POINT

part of the terminal unit which is the receptor for a *GAS-SPECIFIC* (3.8) probe

3.10**NIPPLE**

portion of a connector which is pushed into and secured within the bore (lumen) of a hose

3.11**NOMINAL INLET PRESSURE**

P_1

upstream *WORKING PRESSURE* (3.21) specified by the manufacturer for which the *PRESSURE REGULATOR* (3.18) is intended to be used

3.12**NOMINAL OUTLET PRESSURE**

P_2

nominal downstream pressure under flow conditions specified by the manufacturer

3.13**ORIFICE**

restriction of known cross-section that delivers a constant flow of gas when supplied with gas at a constant upstream pressure

3.14**OXIDIZING GAS**

any gas or gas mixture more oxidizing than air, i.e. any gas or gas mixture that is able, at atmospheric pressure, to support the combustion more than a reference oxidizer consisting of 23,5 % oxygen in nitrogen

[SOURCE: ISO 10156:2017, 3.1.5, modified]

3.15**PRE-SET PRESSURE REGULATOR**

PRESSURE REGULATOR (3.18) that is not provided with a means of operator adjustment of the outlet pressure

3.16**PRESSURE GAUGE**

device that measures and indicates pressure

3.17**PRESSURE OUTLET**

outlet intended to deliver gas at a controlled pressure

3.18**PRESSURE REGULATOR**

device that reduces the inlet pressure and maintains the set outlet pressure within specified limits

3.19**PRESSURE-RELIEF DEVICE**

device intended to relieve excess pressure at a pre-set value

3.20**SINGLE FAULT CONDITION**

condition in which a single means for protection against a safety hazard in equipment is defective or a single external abnormal condition is present

3.21**WORKING PRESSURE**

settled pressure of a compressed gas at a uniform reference temperature of 15 °C in a full gas cylinder

Note 1 to entry: This definition does not apply to liquefied gases (e.g. carbon dioxide) or dissolved gases (e.g. acetylene).

4 Nomenclature

Examples of PRESSURE REGULATORS (see figures) with terminology are given in [Annex A](#).

5 General requirements

5.1 Safety

PRESSURE REGULATORS shall, when transported, stored, installed, operated in normal use and maintained according to the instructions of the manufacturer, present no risks with an unacceptable level, under normal condition or SINGLE FAULT CONDITION, identified using risk management procedures in accordance with ISO 14971.

The risks associated with the ignition of metallic and non-metallic materials, including the potential release of toxic products in an oxygen-enriched environment, shall be assessed according to the principles defined in ISO 15001.

The design of the PRESSURE REGULATOR should be such that in the event of internal ignition, the consequences of the ignition are contained and the gas vented safely.

Check compliance by inspection of the risk management file.

NOTE A situation in which a fault is not detected is considered a normal condition. Fault conditions/hazardous situations can remain undetected over a period of time and as a consequence can lead to an unacceptable risk. In that case, a fault condition subsequently detected needs to be considered as a SINGLE FAULT CONDITION. Specific risk control measures to deal with such situations need to be determined within the risk management process.

5.2 Usability

The manufacturer shall address, in a usability engineering process, any risks resulting from poor usability.

Check compliance by inspection of the usability engineering file.

NOTE For information related to usability, see other documents; for example, IEC 62366-1[6] and IEC/TR 62366-2[7].

5.3 Alternative construction

PRESSURE REGULATORS and components, or parts thereof, using materials or having forms of construction different from those detailed in this document, shall be presumed to be in compliance with the safety objectives of this document if it can be demonstrated that at least an equivalent degree of safety is obtained (i.e. compliance with requirements presumes that risks have been mitigated to acceptable levels) unless objective evidence to the contrary becomes available.

Objective evidence may be obtained by post-market surveillance.

Evidence of at least an equivalent degree of safety shall be provided by the manufacturer.

NOTE Regional or national regulations can require the provision of evidence to a competent authority or a conformity assessment body, e.g. to a notified body in the European Economic Area (EEA) upon request.

5.4 Materials

5.4.1 * The materials which come in contact with the medical gas in normal condition shall be resistant to corrosion and compatible with oxygen, the other medical gases and their mixtures in the temperature range specified in [6.1](#).

NOTE 1 Corrosion resistance includes resistance against moisture and surrounding materials.

NOTE 2 Oxygen compatibility is usually defined as the ability of a material to coexist with oxygen and a moderate ignition source. The aim of using oxygen-compatible materials is to develop a system design which has a low probability of ignition and minimizes consequences based on the use of materials exhibiting good compatibility, low energy release if ignited or by minimizing the quantities of non-metallic components.

NOTE 3 Many materials which do not burn in air will do so in an oxygen-enriched atmosphere, particularly under pressure. Similarly, materials which can be ignited in air require lower ignition energies to ignite in an oxygen-enriched atmosphere. Many such materials can be ignited by friction at a valve seat or by adiabatic compression when an oxygen-enriched gas at high pressure is rapidly introduced into a system initially at low pressure.

NOTE 4 Halogenated polymers such as polytetrafluoroethylene (PTFE), polychlorotrifluoroethylene (PTCFE) and fluoroelastomers (FKM) can release highly toxic products during thermal decomposition.

NOTE 5 Design considerations and criteria for the selection of metallic and non-metallic materials are given in ISO 15001.

5.4.2 Materials that are liable to shed particles which can come in contact with the medical gas in normal condition or SINGLE FAULT CONDITION shall not be used for highly strained components and parts liable to wear.

EXAMPLE Springs.

NOTE See ISO 15001:2010, Annex C.

5.4.3 * Aluminium, aluminium alloys or alloys with aluminium content greater than 2,5 % shall not be used for components whose surfaces come into contact with OXIDIZING GASES or gas mixtures at cylinder pressure in normal or SINGLE FAULT CONDITION.

5.4.4 Consideration should be given to the avoidance of stainless steel and other ferrous alloys for components whose surfaces come into contact with OXIDIZING GASES or gas mixtures at cylinder pressure in normal or SINGLE FAULT CONDITION.

5.4.5 The materials shall permit the PRESSURE REGULATOR and its components to meet the requirements of [Clause 5](#) in the temperature range of -20 °C to +60 °C.

NOTE Regional or national environmental conditions can require deviation from this range of temperatures.

5.4.6 PRESSURE REGULATORS shall meet the requirements of this document after being packed for transport and storage and being exposed to environmental conditions, as stated by the manufacturer.

Evidence of conformity with the requirements of [Clause 5](#) shall be provided by the manufacturer upon request.

6 Design requirements

6.1 General

The operation of the PRESSURE REGULATOR shall comply with the requirements of this document between -20 °C and +60 °C.

NOTE Regional or national regulations can specify additional design requirements.