



BSI Standards Publication

Biological evaluation of medical devices

Part 18: Chemical characterization of medical device
materials within a risk management process

National foreword

This British Standard is the UK implementation of EN ISO 10993-18:2020+A1:2023. It is identical to ISO 10993-18:2020 incorporating amendment 1:2022. It supersedes BS EN ISO 10993-18:2020, which is withdrawn.

The start and finish of text introduced or altered by amendment is indicated in the text by tags. Tags indicating changes to ISO text carry the number of the ISO amendment. For example, text altered by ISO amendment 1 is indicated by A1 A1.

The UK participation in its preparation was entrusted to Technical Committee CH/194, Biological evaluation of medical devices.

A list of organizations represented on this committee can be obtained on request to its committee manager.

Contractual and legal considerations

This publication has been prepared in good faith, however no representation, warranty, assurance or undertaking (express or implied) is or will be made, and no responsibility or liability is or will be accepted by BSI in relation to the adequacy, accuracy, completeness or reasonableness of this publication. All and any such responsibility and liability is expressly disclaimed to the full extent permitted by the law.

This publication is provided as is, and is to be used at the recipient's own risk.

The recipient is advised to consider seeking professional guidance with respect to its use of this publication.

This publication is not intended to constitute a contract. Users are responsible for its correct application.

This publication has been prepared under a mandate given to the European Standards Organizations by the European Commission and the European Free Trade Association. It is intended to support requirements of the EU legislation detailed in the European Foreword. A European Annex, usually Annex ZA or ZZ, describes how this publication relates to that EU legislation.

For the Great Britain market (England, Scotland and Wales), if UK Government has designated this publication for conformity with UKCA marking (or similar) legislation, it may contain an additional National Annex. Where such a National Annex exists, it shows the correlation between this publication and the relevant UK legislation. If there is no National Annex of this kind, the relevant Annex ZA or ZZ in the body of the European text will indicate the relationship to UK regulation applicable in Great Britain. References to EU legislation may need to be read in accordance with the UK designation and the applicable UK law. Further information on designated standards can be found at www.bsigroup.com/standardsandregulation.

For the Northern Ireland market, UK law will continue to implement relevant EU law subject to periodic confirmation. Therefore Annex ZA/ZZ in the European text, and references to EU legislation, are still valid for this market.

UK Government is responsible for legislation. For information on legislation and policies relating to that legislation, consult the relevant pages of www.gov.uk.

© The British Standards Institution 2023
Published by BSI Standards Limited 2023

ISBN 978 0 539 18132 6

ICS 11.100.20

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 30 June 2020.

Amendments/corrigenda issued since publication

Date	Text affected
31 December 2020	Implementation of CEN correction notice 2 December 2020: Annex ZA replaced
31 July 2023	Implementation of ISO amendment 1:2022 with CEN endorsement A1:2023

English Version

**Biological evaluation of medical devices - Part 18:
Chemical characterization of medical device materials
within a risk management process (ISO 10993-18:2020)**

Évaluation biologique des dispositifs médicaux - Partie
18 : Caractérisation chimique des matériaux des
dispositifs médicaux au sein d'un processus de gestion
du risque (ISO 10993-18:2020)

Biologische Beurteilung von Medizinprodukten - Teil
18: Chemische Charakterisierung von Werkstoffen für
Medizinprodukte im Rahmen eines
Risikomanagementsystems (ISO 10993-18:2020)

This European Standard was approved by CEN on 21 July 2019.

This European Standard was corrected and reissued by the CEN-CENELEC Management Centre on 02 December 2020.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

European foreword

This document (EN ISO 10993-18:2020) has been prepared by Technical Committee ISO/TC 194 "Biological and clinical evaluation of medical devices" in collaboration with Technical Committee CEN/TC 206 "Biological and clinical evaluation of medical devices" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by November 2020, and conflicting national standards shall be withdrawn at the latest by November 2020.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 10993-18:2009.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

For the relationship with EU Directive(s) see informative Annex ZA, ZB and ZC, which are an integral part of this document.

The following referenced documents are indispensable for the application of this document. For undated references, the latest edition of the referenced document (including any amendments) applies. For dated references, only the edition cited applies. However, for any use of this standard 'within the meaning of Annex ZA', the user should always check that any referenced document has not been superseded and that its relevant contents can still be considered the generally acknowledged state-of-art.

When an IEC or ISO standard is referred to in the ISO standard text, this shall be understood as a normative reference to the corresponding EN standard, if available, and otherwise to the dated version of the ISO or IEC standard, as listed below.

NOTE The way in which these referenced documents are cited in normative requirements determines the extent (in whole or in part) to which they apply.

Table — Correlations between undated normative references and dated EN and ISO standards

Normative references as listed in Clause 2 of the ISO standard	Equivalent dated standard	
	EN	ISO or IEC
ISO 10993-1	EN ISO 10993-1:2020	ISO 10993-1:2018
ISO 10993-17	EN ISO 10993-17:2009	ISO 10993-17:2002
ISO 14971	EN ISO 14971:2020	ISO 14971:2020

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 10993-18:2020 has been approved by CEN as EN ISO 10993-18:2020 without any modification.

European foreword to amendment A1

This document (EN ISO 10993-18:2020/A1:2023) has been prepared by Technical Committee ISO/TC 194 "Biological and clinical evaluation of medical devices" in collaboration with Technical Committee CEN/TC 206 "Biological and clinical evaluation of medical devices" the secretariat of which is held by DIN.

This Amendment to the European Standard EN ISO 10993-18:2020 shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by January 2024, and conflicting national standards shall be withdrawn at the latest by January 2024.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document has been prepared under a Standardization Request given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s) / Regulation(s).

For the relationship with EU Directive(s) / Regulation(s), see informative Annex ZA, which is an integral part of this document.

Any feedback and questions on this document should be directed to the users' national standards body/national committee. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

Endorsement notice

The text of ISO 10993-18:2020/Amd 1:2022 has been approved by CEN as EN ISO 10993-18:2020/A1:2023 without any modification.

Annex ZA (informative)

Relationship between this European Standard the General Safety and Performance Requirements of Regulation (EU) 2017/745 aimed to be covered

This European standard has been prepared under a Commission's standardization request M/575 to provide one voluntary means of conforming to the General Safety and Performance Requirements of Regulation (EU) 2017/745 of 5 April 2017 concerning medical devices [OJ L 117] and to system or process requirements including those relating to quality management systems, risk management, post-market surveillance systems, clinical investigations, clinical evaluation or post-market clinical follow-up.

Once this standard is cited in the Official Journal of the European Union under that Regulation, compliance with the normative clauses of this standard given in Table ZA.1 and application of the edition of the normatively referenced standards as given in Table ZA.2 confers, within the limits of the scope of this standard, a presumption of conformity with the corresponding General Safety and Performance Requirements of that Regulation, and associated EFTA Regulations.

Where a definition in this standard differs from a definition of the same term set out in Regulation (EU) 2017/745, the differences shall be indicated in the Annex Z. For the purpose of using this standard in support of the requirements set out in Regulation (EU) 2017/745, the definitions set out in this Regulation prevail.

Where the European standard is an adoption of an International Standard, the scope of this document can differ from the scope of the European Regulation that it supports. As the scope of the applicable regulatory requirements differ from nation to nation and region to region the standard can only support European regulatory requirements to the extent of the scope of the European Regulation for medical devices ((EU) 2017/745).

NOTE 1 Where a reference from a clause of this standard to the risk management process is made, the risk management process needs to be in compliance with Regulation (EU) 2017/745. This means that risks have to be 'reduced as far as possible', 'reduced to the lowest possible level', 'reduced as far as possible and appropriate', 'removed or reduced as far as possible', 'eliminated or reduced as far as possible', 'removed or minimized as far as possible', or 'minimized', according to the wording of the corresponding General Safety and Performance Requirement.

NOTE 2 The manufacturer's policy for determining **acceptable risk** must be in compliance with General Safety and Performance Requirements 1, 2, 3, 4, 5, 8, 9, 10, 11, 14, 16, 17, 18, 19, 20, 21 and 22 of the Regulation.

NOTE 3 When a General Safety and Performance Requirement does not appear in Table ZA.1, it means that it is not addressed by this European Standard.

Table ZA.1 — Correspondence between this European Standard and Annex I of Regulation (EU) 2017/745 [OJ L 117] and to system or process requirements including those relating to quality management systems, risk management, post-market surveillance systems, clinical investigations, clinical evaluation or post-market clinical follow-up

General Safety and Performance Requirements of Regulation (EU) 2017/745	Clause(s) / sub-clause(s) of this EN	Remarks / Notes
10.1 a), b) and h)	5 and 6	<p>This standard provides requirements and recommendations for the chemical characterization of medical devices as part of a risk management process, including a qualitative characterization of the chemicals and materials used, a quantitative characterization of the amounts of chemicals and materials used and an evaluation of chemical release (leachable and extractable profile) in both the design and manufacturing processes. This chemical characterization can be used to define or confirm chemical specifications [10.1 h)] and evaluate the risk of toxicity [10.1 a) and b)] and biocompatibility [10.1 b)]. Flammability [10.1 a)] is not covered.</p> <p>For 10.1 b), ADME (absorption, distribution, metabolism, and excretion) is not covered. .</p> <p>For 10.1 h), physical specifications are not covered.</p>
10.2	5 and 6	<p>This standard provides requirements and recommendations for the chemical characterization of medical devices as part of a risk management process, including an evaluation of the release of contaminants and residues (composition, leachable and extractable profile) in both the design and manufacturing processes. Packaging is not covered. Aspects of contaminants and residues during transport and storage are not covered.</p>
10.4.1 (First paragraph, first sentence)	5 and 6	<p>This standard provides requirements and recommendations for the chemical characterization of medical devices as part of a risk management process, including an evaluation of chemical substances that may be released from the medical device (composition, leachable and extractable profile) in both the design and manufacturing processes.</p> <p>Particles and wear debris are not covered.</p>

Table ZA.2 — Applicable Standards to confer presumption of conformity as described in this Annex ZA

Column 1 Reference in Clause 2	Column 2 International Standard Edition	Column 3 Title	Column 4 Corresponding European Standard Edition
ISO 10993-1	ISO 10993-1:2018	<i>Biological evaluation of medical devices — Part 1: Evaluation and testing within a risk management process</i>	EN ISO 10993-1:2020
ISO 10993-17	ISO 10993-17:2002	<i>Biological evaluation of medical devices — Part 17: Establishment of allowable limits for leachable substances</i>	EN ISO 10993-17:2009
ISO 14971	ISO 14971:2019	<i>Medical devices — Application of risk management to medical devices</i>	EN ISO 14971:2019 EN ISO 14971:2019 + A11:2021

The documents listed in the Column 1 of table ZA.2, in whole or in part, are normatively referenced in this document, i.e. are indispensable for its application. The achievement of the presumption of conformity is subject to the application of the edition of Standards as listed in Column 4 or, if no European Standard Edition exists, the International Standard Edition given in Column 2 of table ZA.2.

WARNING 1 — Presumption of conformity stays valid only as long as a reference to this European standard is maintained in the list published in the Official Journal of the European Union. Users of this standard should consult frequently the latest list published in the Official Journal of the European Union.

WARNING 2 — Other Union legislation may be applicable to the product(s) falling within the scope of this standard.

Annex ZB (informative)

Relationship between this European Standard and the essential requirements of Directive 90/385/EEC [OJ L 189] aimed to be covered

This European Standard has been prepared under a Commission's joint standardization request M/BC/CEN/89/9 concerning harmonized standards relating to horizontal aspects in the field of medical devices to provide one voluntary means of conforming to essential requirements of Council Directive 90/385/EEC of 20 June 1990 on the approximation of the laws of the Member States relating to active implantable medical devices [OJ L 189].

Once this standard is cited in the Official Journal of the European Union under that Directive, compliance with the normative clauses of this standard given in Table ZB.1 confers, within the limits of the scope of this standard, a presumption of conformity with the corresponding essential requirements of that Directive and associated EFTA regulations.

NOTE 1 Where a reference from a clause of this standard to the risk management process is made, the risk management process needs to be in compliance with Directive 90/385/EEC as amended by 2007/47/EC. This means that risks have to be reduced 'as far as possible', 'to a minimum', 'to the lowest possible level', 'minimized' or 'removed', according to the wording of the corresponding essential requirement.

NOTE 2 The manufacturer's policy for determining acceptable risk must be in compliance with Essential Requirements 1, 4, 5, 8, 9 and 10 of the Directive.

NOTE 3 This Annex ZB is based on normative references according to the table of references in the European foreword, replacing the references in the core text.

NOTE 4 When an Essential Requirement does not appear in Table ZB.1, it means that it is not addressed by this European Standard.

Table ZB.1 — Correspondence between this European Standard and Annex I of Directive 90/385/EEC [OJ L 189]

Essential Requirements of Directive 90/385/EEC	Clause(s)/sub-clause(s) of this EN	Remarks/Notes
9 (only first and second indent)	5 and 6	<i>This standard provides requirements and recommendations for evaluating the chemical characterization of medical devices as part of a risk management process, including a qualitative characterization of the chemicals and materials used, a quantitative characterization of the amounts of chemicals and materials used and an evaluation of chemical release (leachable and extractable profile) in both the design and manufacturing processes. This chemical characterization can be used to evaluate the risk of toxicity (first indent) and biocompatibility (second indent).</i>

General Note: Presumption of conformity depends on also complying with the relevant parts of the ISO 10993 series.

WARNING 1 — Presumption of conformity stays valid only as long as a reference to this European Standard is maintained in the list published in the Official Journal of the European Union. Users of this standard should consult frequently the latest list published in the Official Journal of the European Union.

WARNING 2 — Other Union legislation may be applicable to the products falling within the scope of this standard.

Annex ZC (informative)

Relationship between this European standard and the General Safety and Performance Requirements of Regulation (EU) 2017/745 aimed to be covered

This European standard has been prepared under a Commission's standardisation request [Full reference to the request "M/xxx"] to provide one voluntary means of conforming to the General Safety and Performance Requirements of Regulation (EU) 2017/745 of 5 April 2017 concerning medical devices [OJ L 117].

Once this standard is cited in the Official Journal of the European Union under that Regulation, compliance with the normative clauses of this standard given in Table ZC.1 confers, within the limits of the scope of this standard, a presumption of conformity with the corresponding General Safety and Performance Requirements of that Regulation, and associated EFTA regulations.

NOTE 1 Where a reference from a clause of this standard to the risk management process is made, the risk management process needs to be in compliance with Regulation (EU) 2017/745. This means that risks have to be 'reduced as far as possible', 'reduced to the lowest possible level', 'reduced as far as possible and appropriate', 'removed or reduced as far as possible', 'eliminated or reduced as far as possible', 'removed or minimized as far as possible', or 'minimized', according to the wording of the corresponding General Safety and Performance Requirement.

NOTE 2 The manufacturer's policy for determining acceptable risk must be in compliance with General Safety and Performance Requirements 1, 2, 3, 4, 5, 8, 9, 10, 11, 14, 16, 17, 18, 19, 20, 21 and 22 of the Regulation.

NOTE 3 This Annex ZC is based on normative references according to the table of references in the European Foreword, replacing the references in the core text.

NOTE 4 When a General Safety and Performance Requirement does not appear in Table ZC.1, it means that it is not addressed by this European Standard.

**Table ZC.1 — Correspondence between this European Standard and
Annex I of Regulation (EU) 2017/745 [OJ L 117]**

General Safety and Performance Requirements of Regulation (EU) 2017/745	Clause(s)/sub-clause(s) of this EN	Remarks/Notes
--	---------------------------------------	---------------

10.1 a), b) and h)	5 and 6	<p><i>This standard provides requirements and recommendations for evaluating the chemical characterization of medical devices as part of a risk management process, including a qualitative characterization of the chemicals and materials used, a quantitative characterization of the amounts of chemicals and materials used and an evaluation of chemical release (leachable and extractable profile) in both the design and manufacturing processes. This chemical characterization can be used to evaluate the risk of toxicity [10.1 a) and b)] and biocompatibility [10.1 b)]. Flammability [10.1 a)] is not covered.</i></p> <p><i>For 10.1 b), ADME (absorption, distribution, metabolism, and excretion) is not covered. .</i></p> <p><i>For 10.1 h), physical specifications are not covered.</i></p>
10.2	5 and 6	<p><i>This standard provides requirements and recommendations for evaluating the chemical characterization of medical devices as part of a risk management process, including an evaluation of chemical release (leachable and extractable profile) in both the design and manufacturing processes. Packaging is not covered. Aspects of contaminants and residues during transport and storage are not covered.</i></p>
10.4.1 (First paragraph, first sentence)	5 and 6	<p><i>This standard provides requirements and recommendations for evaluating the chemical characterization of medical devices as part of a risk management process, including an evaluation of chemical release (leachable and extractable profile) in both the design and manufacturing processes.</i></p> <p><i>Particles and debris are not covered.</i></p>

WARNING 1: Presumption of conformity stays valid only as long as a reference to this European standard is maintained in the list published in the Official Journal of the European Union. Users of this standard should consult frequently the latest list published in the Official Journal of the European Union.

WARNING 2: Other Union legislation may be applicable to the product(s) falling within the scope of this standard.

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	2
4 Symbols and abbreviated terms	6
5 Characterization procedure	7
5.1 General	7
5.2 Establish medical device configuration and material composition	10
5.2.1 General	10
5.2.2 Information gathering	11
5.2.3 Information generation	11
5.3 Assess material/chemical equivalence to a clinically established material or medical device	12
5.4 Assess the hypothetical worst-case chemical release based on total exposure to the medical device's chemical constituents	13
5.4.1 Establish the hypothetical worst-case chemical release	13
5.4.2 Assess the hypothetical worst-case chemical release	13
5.5 Establish an analytical evaluation threshold	14
5.6 Estimate the chemical release; perform extraction study	14
5.7 Assess the estimated chemical release (extractables profile)	17
5.8 Determine the actual chemical release; perform leachables study	17
5.9 Assess the actual chemical release (leachables profile)	19
5.10 Exiting the chemical characterization process	19
6 Chemical characterization parameters and methods	19
6.1 General	19
6.2 Material composition	20
6.3 Extractables and leachables	22
6.4 Structural composition or configuration	24
6.5 Analytical methods	25
7 Reporting of the chemical characterization data	26
Annex A (informative) General principles of chemical characterization	27
Annex B (informative) Information sources for chemical characterization	31
Annex C (informative) Principles for establishing biological equivalence	35
Annex D (informative) Principles of sample extraction	38
Annex E (informative) Calculation and application of the analytical evaluation threshold (AET)	50
Annex F (informative) Qualification of analytical methods used for extractables/leachables	58
Annex G (informative) Reporting details for analytical methods and chemical data	61
Bibliography	64

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 194, *Biological and clinical evaluation of medical devices*.

This second edition cancels and replaces the first edition (ISO 10993-18:2005), which has been technically revised. The main changes compared to the previous edition are as follows:

- greater integration and harmonization with ISO 10993-1, ISO 10993-12, and ISO 10993-17;
- a revised and expanded chemical characterization process flowchart;
- a strengthened explanation that analytical testing is not necessarily required;
- added a number of definitions (e.g. medical device configuration, materials of construction, and material composition);
- clarified testing approaches unique to chemical characterization (i.e. digestion and dissolution for hazard identification);
- added discussion of considerations related to analytical method qualification;
- added informative annexes on general principles, vehicle extraction considerations, and the analytical evaluation threshold (AET; concentration threshold below which extractables or leachables identification is unneeded).

A list of all parts in the ISO 10993 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

ISO 10993-1 serves as a framework in which to plan a biological evaluation which, as scientific knowledge advances our understanding of the basic mechanisms of tissue responses, minimizes the number and exposure of test animals. Preference is given to the assessment of chemical/physical properties and testing with *in vitro* models in situations within a risk assessment process. These methods are used when the results yield equally relevant information to that obtained from *in vivo* models.

The characterization procedure and its associated flowchart is based on the principles in ISO 10993-1; specifically, that the biological evaluation and risk assessment process is most efficient and effective if it is based on the minimum amount of acceptable and necessary chemical information that can establish that a medical device presents an acceptable health risk.

ISO 10993-1:2018, 4.2 states that in the selection of materials to be used in medical device manufacture, the first consideration shall be fitness for purpose with regard to characteristics and properties of the material, which can include chemical, toxicological, physical, electrical, morphological and mechanical properties. Furthermore, ISO 10993-1:2018, 6.1 states that gathering physical and chemical information on the medical device or component is a crucial first step in the biological evaluation process and its associated process of material characterization.

Lastly, ISO 10993-1:2018, and by reference ISO 14971, points out that a biological risk analysis depends on what is known about the material formulation, what nonclinical and clinical safety and toxicological data exist, and on the nature and duration of body contact with the medical device.

The requirements specified in this document are intended to yield the following information, which will be of value in assessing the biological response to the materials as represented in the final product.

- The identities and quantities, as appropriate, of the materials of construction of the medical device (device configuration).
- The identities and quantities, as appropriate, of the chemical constituents in each material of construction (material composition).
- The identities and quantities, as appropriate, of chemical substances used in the medical device's manufacturing process, including processing aids and residues.
- The potential of the medical device and/or its materials of construction to release chemical substances to which a potentially affected individual could be exposed to during clinical conditions of use.

The composition of the materials of construction is mainly established by the suppliers of these materials. The composition can change during manufacture of a medical device. Other medical device characteristics are chiefly established by component suppliers or device manufacturers to address the performance and quality requirements to be met by the finished medical device as well as the production, storage and distribution processes experienced by the medical device.

Biological evaluation of medical devices —

Part 18:

Chemical characterization of medical device materials within a risk management process

1 Scope

This document specifies a framework for the identification, and if necessary, quantification of constituents of a medical device, allowing the identification of biological hazards and the estimation and control of biological risks from material constituents, using a generally stepwise approach to the chemical characterization which can include one or more of the following:

- the identification of its materials of construction (medical device configuration);
- the characterization of the materials of construction via the identification and quantification of their chemical constituents (material composition);
- the characterization of the medical device for chemical substances that were introduced during manufacturing (e.g. mould release agents, process contaminants, sterilization residues);
- the estimation (using laboratory extraction conditions) of the potential of the medical device, or its materials of construction, to release chemical substances under clinical use conditions (extractables);
- the measurement of chemical substances released from a medical device under its clinical conditions of use (leachables).

This document can also be used for chemical characterization (e.g. the identification and/or quantification) of degradation products. Information on other aspects of degradation assessment are covered in ISO 10993-9, ISO 10993-13, ISO 10993-14 and ISO 10993-15.

The ISO 10993 series is applicable when the material or medical device has direct or indirect body contact (see ISO 10993-1 for categorization by nature of body contact).

This document is intended for suppliers of materials and manufacturers of medical devices, to support a biological evaluation.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 10993-1, *Biological evaluation of medical devices — Part 1: Evaluation and testing within a risk management process*

ISO 10993-17, *Biological evaluation of medical devices — Part 17: Establishment of allowable limits for leachable substances*

ISO 14971, *Medical devices — Application of risk management to medical devices*

3 Terms and definitions

For the purposes of this document, the definitions in ISO 10993-1 and the following apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available from <https://www.iso.org/obp>
- IEC Electropedia: available from <http://www.electropedia.org/>

3.1 **accelerated extraction**

extraction whose duration is shorter than the duration of clinical use but whose conditions do not result in a chemical change to the substances being extracted

Note 1 to entry: See also [Annex D](#).

3.2 **analytical evaluation threshold** **AET**

threshold below which the analyst need not identify or quantify leachables or extractables or report them for potential toxicological assessment

Note 1 to entry: See [Annex E](#).

3.3 **analytically expedient**

situation where an extraction vehicle can be directly evaluated with generally available analytical methods with the sensitivity and selectivity necessary to achieve a designated reporting threshold such as the AET

3.4 **analytical screening method**

method whose purpose is to discover, identify and semi-quantitatively estimate the concentration of all relevant analytes in a test sample above an established reporting threshold (such as the AET)

3.5 **analytical targeting method**

method whose purpose is to quantify, with an appropriately high degree of accuracy and precision, specified analytes in a specified test sample over a specified concentration range

3.6 **chemical characterization**

process of obtaining chemical information, accomplished either by information gathering or by information generation, for example, by literature review or chemical testing

3.7 **chemical information**

qualitative and quantitative, if applicable, knowledge related to the configuration, composition and production of the medical device and/or its materials of construction, thereby establishing the identities and amounts of constituents present in the materials and device

Note 1 to entry: See also [5.2.1](#), [5.2.2](#), [5.2.3](#), and [Annex B](#).

Note 2 to entry: Chemical information can be used to establish the hypothetical worst-case release of chemicals from a medical device, predicated on the circumstance that all chemicals present in the device are released from the device under its clinical conditions of use.

3.8 **clinically established**

medical device, component, or material of construction which has been used extensively for specified and established clinical uses for which biocompatibility has been established

3.9**component**

item which forms one part of a medical device, but is not itself a medical device

3.10**constituent**

chemical that is present in a finished medical device or its materials of construction

Note 1 to entry: Constituents may be intentionally present (e.g. an additive such as an antioxidant) or unintentionally present (e.g. an impurity or degradant).

3.11**convertor**

person or company who converts or fabricates a basic raw material into a semi-finished product (e.g. a former of lengths of rod, tubing, or plastic components)

3.12**digestion**

process of completely solubilizing a medical device, one or more of its components or one or more of its materials of construction by breaking it down into its fundamental structural units, including its elemental constituents or monomeric units

3.13**dissolution**

process of completely solubilizing a medical device, one or more of its components or one or more of its materials of construction, generally preserving the molecular structures of its constituents

3.14**exaggerated extraction**

extraction that is intended to result in a greater number or amount of chemical constituents being released as compared to the amount generated under the clinical conditions of use

Note 1 to entry: It is important to ensure that the exaggerated extraction does not result in a chemical change of the material or the substances being extracted.

3.15**exhaustive extraction**

multi-step extraction conducted until the amount of material extracted in a subsequent extraction step is less than 10 % by gravimetric analysis (or achieved by other means) of that determined in the initial extraction step

3.16**extractable**

substance that is released from a medical device or material of construction when the medical device or material is extracted using laboratory extraction conditions and vehicles

3.17**extraction power**

ability of an extraction vehicle to extract (or leach) substances from a medical device, component or material of construction

Note 1 to entry: The extraction power of an extraction vehicle is impacted by its physicochemical properties, including, but not limited to, its polarity, pH and dielectric constant.

3.18**extraction vehicle**

medium (solution or solvent) which is used to extract (or leach) a test article for the purpose of establishing the test article's extractables or leachables profile

Note 1 to entry: It is preferred that extraction vehicles be analytically expedient.

Note 2 to entry: For some medical devices (e.g. infusion systems) that are labelled for use with a drug, the most appropriate extraction medium may be the drug product or drug product vehicle.

3.19

identification

process of assigning a molecular structure and chemical name to an organic compound or assigning constituent elements or molecular structure as appropriate, and a chemical name to an inorganic compound

3.20

information gathering

process of collecting existing chemical information, including available test results, that is relevant to chemical characterization

3.21

information generation

process of producing chemical information via laboratory testing

3.22

leachable

substance that is released from a medical device or material during its clinical use

Note 1 to entry: For many medical devices, a leachables study is not practical due to challenges with reproducing actual clinical conditions, so simulated-use extraction studies are often performed instead. See definition for simulated-use extraction.

3.23

manufacturer

natural or legal person who manufactures or fully refurbishes a medical device, or has a device designed, manufactured, or fully refurbished, and markets that medical device under its name or trademark

3.24

material composition

listing of the constituents that are contained in a material (qualitative) and the amount of each substance in the material (quantitative)

Note 1 to entry: A material's composition establishes the hypothetical situation in which the total amount of all substances present in a medical device are released during clinical use. These amounts can be derived directly from known composition; experimentally, they can be derived from digestion, dissolution, and, in many cases, exhaustive extraction studies.

3.25

material of construction

individual raw material that is used to produce a component

EXAMPLE Polymer resins.

3.26

medical device configuration

listing of a medical device's components (qualitative), including a listing of the component's materials of construction (qualitative) and the proportion of each material in each component (quantitative)

Note 1 to entry: Device configuration should also take into account the shape and relative arrangement of the parts in the medical device and surface properties (topography and chemistry).

3.27

potentially affected individual

person having direct or indirect body contact with the medical device

Note 1 to entry: See ISO 10993-1 for categorization by nature of body contact.

3.28**qualification**

process of establishing that an analytical method is suitable for its intended use

3.29**qualitative analysis**

analytical approach which estimates an analyte's concentration by using the response from a surrogate substance (or substances) chosen without specifically addressing or considering the relative responses of the analyte and the surrogate(s)

3.30**quantification**

process of assigning a concentration to an analyte present in a sample

Note 1 to entry: There are several possible levels as shown in [3.31](#), [3.32](#) and [3.33](#).

3.31**estimated quantitative analysis**

analytical approach which estimates an analyte's concentration by using the response from a surrogate substance chosen without specifically addressing or considering the relative responses of the analyte and the surrogate

3.32**semi-quantitative analysis**

analytical approach which provides an analyte's concentration by using the response from a surrogate substance (or substances), specifically accounting for the relative responses of the analyte and the surrogate

3.33**quantitative analysis**

analytical approach which establishes the most accurate estimate of an analyte's concentration by using a response function (calibration curve) generated specifically for the analyte via the use of a reference standard

Note 1 to entry: Estimated quantitative analysis is generally less accurate than semi-quantitative analysis, which is generally less accurate than quantitative analysis.

3.34**safety concern threshold****SCT**

threshold below which a leachable (or an extractable as a probable leachable) has a dose so low that it presents a negligible safety concern from carcinogenic and non-carcinogenic toxic effects

Note 1 to entry: See Reference [\[27\]](#).

3.35**simulated-use extraction**

extraction using a method that simulates clinical use

Note 1 to entry: A simulated-use extraction is performed to estimate the type and amount of substances that are expected to be released from a medical device during its clinical use. A simulated-use extraction is designed to produce an extractables profile that represents the worst-case leachables profile, meaning that all leachables are also extractables and the levels of all individual extractables are at least equal to the level of all individual leachables.

3.36**solubilisation**

action or process of using a vehicle to dissolve part or all of a test article

Note 1 to entry: Leaching, extraction, dissolution, and digestion are (progressively more complete) sub-categories of solubilisation.

3.37

sponsor

individual or organization that plans, commissions, and takes responsibility for testing of a medical device

3.38

supplier

person or company who manufactures or provides the materials of construction or components to be used in the manufacture of a medical device

3.39

threshold of toxicological concern

TTC

level of exposure for constituents, below which there would be no appreciable risk to human health

Note 1 to entry: See ISO/TS 21726 for full context.

3.40

toxicological risk assessment

act of determining the potential of a chemical to elicit an adverse effect based on a specified level of exposure

4 Symbols and abbreviated terms

The abbreviated terms given in [Table 1](#) are used in this document.

Table 1 — Methodology abbreviations

Abbreviated term	Analytical method
2D PAGE	Two-dimensional polyacrylamide gel electrophoresis
AES	Atomic emission spectroscopy
AET	Analytical evaluation threshold
DMTA	Dynamic mechanical thermal analysis
DSC	Differential scanning calorimetry
FID	Flame ionization detection
FTIR	Fourier transform infrared spectroscopy
GC	Gas chromatography
GPC/SEC	Gel permeation chromatography/size exclusion chromatography
HPLC (or LC)	High performance liquid chromatography (or liquid chromatography)
HS	Headspace sampling
IC	Ion chromatography
ICP	Inductively coupled plasma
IR	Infrared spectroscopy
MS ^a	Mass spectrometry
NMR	Nuclear magnetic resonance spectroscopy
NVOC	Non-volatile organic compound
NVR	Non-volatile residue
SEM-EDS (or SEM-EDX)	Scanning electron microscopy-energy dispersive X-ray spectroscopy
SVOC	Semi-volatile organic compound
TOC	Total organic carbon
UV	Ultraviolet spectroscopy
^a Mass spectrometry is frequently combined with other techniques (especially chromatographic) in coupled methods such as GC-MS, LC-MS and MS-MS.	

Table 1 (continued)

Abbreviated term	Analytical method
VOC	Volatile organic compound
XPS	X-ray photoelectron spectroscopy
XRF	X-ray fluorescence
^a Mass spectrometry is frequently combined with other techniques (especially chromatographic) in coupled methods such as GC-MS, LC-MS and MS-MS.	

5 Characterization procedure

5.1 General

The chemical characterization information, either collected or generated, and augmented with additional supporting information as appropriate, can be used for a range of important applications, for example:

- supporting the overall biological safety of a medical device (ISO 10993-1 and ISO 14971);
- supporting the biological safety of a reprocessed medical device;
- determining the amount of chemical substances that might be leached from a medical device under the conditions of its clinical use, to support performing a toxicological risk assessment (ISO 10993-17);
- supporting equivalence of a proposed medical device to a clinically established device, used for the same type of clinical exposure, with regards to either the device's configuration or its extractables/leachables profiles and any subsequent relevant evaluations;
- supporting equivalence of a clinically established medical device, used for the same type of clinical exposure, after changes in the manufacturing process, (including, but not limited, to changes in the sterilization process), manufacturing sites, suppliers of materials or components, etc.;
- supporting equivalence of a proposed material of construction to a clinically established material of construction with regards to either the material's composition or its extractables profiles and any subsequent relevant evaluations;
- supporting equivalence of a final medical device to a prototype device with regards to the use of data secured on the prototype to support the assessment of the final device, specifically considering relevant information such as composition, device configuration and extractable profile obtained for either the device or its materials of construction; or
- screening of potential new materials for chemical suitability in a medical device for a proposed clinical application.

These important applications notwithstanding, chemical characterization alone can be insufficient to establish the equivalence or biocompatibility of materials and medical devices, and cannot unilaterally substitute for biological testing. However, chemical characterization in combination with risk assessment can be a necessary part of judging chemical equivalence and assessing biocompatibility, and if appropriately conducted can be used in lieu of certain biological tests.

Chemical characterization of a medical device provides the necessary input into the device's biological evaluation and toxicological risk assessment (see ISO 10993-1 and ISO 10993-17). A flowchart describing the general chemical characterization process is given in [Figure 1](#). This flowchart represents the chemical characterization portion of the overall biological evaluation flow as discussed in ISO 10993-1 and is meant to illustrate the characterization process that is described in this clause. This general flowchart is supplemented with additional flowcharts (see [Figures 2 to 4](#)) that provide greater detail to specific steps in the general process.

The requirements and guidance for each step of the chemical characterization process are specified in [5.2](#) to [5.10](#). When specified in the applicable flowchart, knowledgeable and experienced individuals shall compile existing information relevant to the chemical characterization (information gathering) and assess its adequacy as the basis for a toxicological risk assessment of the material/medical device. If the existing information is insufficient to complete the assessment, additional information shall be gathered or produced by testing (information generation) to enable the toxicological risk assessment.

This procedure should consider each of the direct and indirect contact materials of construction used in a medical device in addition to the requirement for chemical characterization of the finished medical device. Since the chemical nature of a medical device can be affected by its processing during its construction (e.g. sterilization), the effect of this processing on the device shall be taken into account in the design and interpretation of the chemical characterization.

At each step of the characterization procedure, the adequacy of the available data as the basis for performing the risk assessment shall be established. The available data can be considered adequate if it reflects or exceeds the conditions of clinical use and a risk assessment based on the available data can be completed. Inadequacies in the data can be addressed by filling gaps in such data (e.g. literature review) and/or supplementing the data via analytical testing.

The flowcharts have the following types of process steps; start/stop, decision points, information gathering and evaluation, and analytical testing. Each type of step is represented by a geometric shape. Start/stop steps are identified as ovals, a decision step is identified as a diamond, an information gathering/evaluation step is represented as a parallelogram, and a step that involves analytical testing is represented as a rectangle.

The steps and actions defined in [5.4.2](#), [5.7](#) and [5.9](#) are part of the risk assessment process and represent the points at which chemical information is provided for assessment. As such, they are for the most part, outside the scope of chemical characterization, which is the focus of this document. These steps are included to indicate the important link between chemical characterization and risk assessment (see ISO 10993-1, ISO 10993-17, and ISO 14971).

The characterization procedure and its associated flowchart system is based on the principles in ISO 10993-1; specifically, that the biological evaluation and toxicological risk assessment process is most efficient and effective if it is based on the appropriate (minimum) amount of acceptable and necessary chemical information that can establish that a medical device presents an acceptable health risk. Thus, the first step of the procedure is to establish the configuration of the medical device and the composition of the device's materials of construction so that it can be compared to a clinically established device or assessed based on hypothetical worst-case chemical release (i.e. "it all comes out"). This assessment should include potential contaminants, degradants, processing aids and additives which could be introduced by the manufacturing process. If an assessment based on the hypothetical worst-case chemical release leads to the conclusion that there is an acceptable risk, then the process can be completed with the collection or generation of a minimum amount of information. On the other hand, if the conclusion of acceptable health risk cannot be supported, then additional data shall be collected, following a step-wise process from determining and evaluating the medical device's hypothetical worst-case chemical release to the actual chemical release under clinical conditions of use. In any and all cases, the information collected shall reflect (or exceed) and be assessed according to the clinical conditions of use.

In using the flowcharts, it is not always necessary to complete all steps in the entire sequence; thus, the flowchart system has multiple points of exit. For example, if one can demonstrate that a hypothetical exposure to all of the chemical constituents of a medical device presents an acceptable health risk, additional chemical testing is not necessary, the characterization is complete and the flowcharts are exited and biological evaluation continued according to ISO 10993-1.