BS EN ISO 15002:2024

BSI Standards Publication

Flow control devices for connection to a medical gas supply system

Documents
Document Sharing Hub

Full document details: BS EN ISO 15002:2024

National foreword

This British Standard is the UK implementation of EN ISO 15002:2024. It is identical to ISO 15002:2023. It supersedes BS EN ISO 15002:2008+A1:2019, which is withdrawn.

The UK participation in its preparation was entrusted to Technical Committee CH/121/6, Medical gas supply systems.

A list of organizations represented on this committee can be obtained on request to its committee manager.

Contractual and legal considerations

This publication has been prepared in good faith, however no representation, warranty, assurance or undertaking (express or implied) is or will be made, and no responsibility or liability is or will be accepted by BSI in relation to the adequacy, accuracy, completeness or reasonableness of this publication. All and any such responsibility and liability is expressly disclaimed to the full extent permitted by the law.

This publication is provided as is, and is to be used at the recipient's own risk.

The recipient is advised to consider seeking professional guidance with respect to its use of this publication.

This publication is not intended to constitute a contract. Users are responsible for its correct application.

© The British Standards Institution 2024
Published by BSI Standards Limited 2024

ISBN 978 0 539 01204 0

ICS 11.040.10

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 31 March 2024.

Amendments/corrigenda issued since publication

Date	Text affected

Documents
Document Sharing Hub

Full document details: BS EN ISO 15002:2024

EUROPEAN STANDARD

NORME EUROPÉENNE

EUROPÄISCHE NORM

EN ISO 15002

March 2024

ICS 11.040.10

Supersedes EN ISO 15002:2008

English Version

Flow control devices for connection to a medical gas supply system (ISO 15002:2023)

Dispositifs de contrôle du débit pour raccordement à un système d'alimentation en gaz médicaux (ISO 15002:2023)

Durchflussregeleinrichtungen zum Anschluss an ein Versorgungssystem für medizinische Gase (ISO 15002:2023)

This European Standard was approved by CEN on 23 March 2024.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and United Kingdom.

EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

© 2024 CEN All rights of exploitation in any form and by any means reserved worldwide for CEN national Members.

Ref. No. EN ISO 15002:2024 E

Documents
Document Sharing Hub

Full document details: BS EN ISO 15002:2024

European foreword

The text of ISO 15002:2023 has been prepared by Technical Committee ISO/TC 121 "Anaesthetic and respiratory equipment" of the International Organization for Standardization (ISO) and has been taken over as EN ISO 15002:2024 by Technical Committee CEN/TC 215 "Respiratory and anaesthetic equipment" the secretariat of which is held by BSI.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by September 2024, and conflicting national standards shall be withdrawn at the latest by September 2024.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 15002:2008.

Any feedback and questions on this document should be directed to the users' national standards body. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

Endorsement notice

The text of ISO 15002:2023 has been approved by CEN as EN ISO 15002:2024 without any modification.

Documents
Document Sharing Hub

Full document details: BS EN ISO 15002:2024

INTERNATIONAL STANDARD

ISO 15002

Third edition
2023-08

Flow control devices for connection to a medical gas supply system

Dispositifs de contrôle du débit pour raccordement à un système d'alimentation en gaz médicaux

Reference number
ISO 15002:2023(E)

Documents
Document Sharing Hub
© ISO 2023

Full document details: BS EN ISO 15002:2024

COPYRIGHT PROTECTED DOCUMENT

© ISO 2023

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

© ISO 2023 – All rights reserved

Documents
Document Sharing Hub

Full document details: BS EN ISO 15002:2024

Contents

Page

Foreword ... **iv**
Introduction ... **vi**
1 Scope ... **1**
2 Normative references ... **2**
3 Terms and definitions ... **2**
4 General requirements ... **3**
4.1 Risk management ... 3
4.2 Usability ... 3
4.3 Materials ... 4
4.4 Oxygen compatibility ... 4
4.5 Environmental conditions ... 4
5 Design requirements ... **5**
5.1 General ... 5
5.2 Gas inlets ... 6
5.3 Outlet connectors ... 6
5.4 Mechanical strength ... 6
5.5 Leakage ... 6
5.6 Flow indication ... 7
5.7 Accuracy ... 7
5.8 Stability of indicated flow ... 8
5.9 Continuous increase in flow ... 8
5.10 Security of components ... 8
5.11 Flow controls and flow selectors ... 8
6 Packaging ... **9**
7 Information supplied by the manufacturer ... **9**
7.1 General ... 9
7.2 Marking ... 9
7.3 Instructions for use ... 10
Annex A (informative) **Rationale** ... **11**
Annex B (informative) **Examples of *flow control devices* and gas supply inlets** ... **13**
Annex C (normative) **Test methods** ... **15**
Annex D (informative) **Hazard identification for risk assessment** ... **20**
Bibliography ... **22**

© ISO 2023 – All rights reserved

Documents
Document Sharing Hub
Full document details: BS EN ISO 15002:2024

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

ISO draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at www.iso.org/patents. ISO shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 121, *Anaesthetic and respiratory equipment*, Subcommittee SC 6, *Medical gas supply systems*.

This third edition cancels and replaces the second edition (ISO 15002:2008), which has been technically revised. It also incorporates the Amendment ISO 15002:2008/Amd.1:2018.

The main changes are as follows:

— title changed as the requirements for *flow control devices* are the same regardless of the gas supply and they control the flow, they do not measure the flow;

— layout changed from requirements for each type of *flow control device* to the common requirements as they are the same for each *flow control device*;

— test methods have been rationalised and put into a new Annex C;

— hazard identification list added as a new Annex D;

— the maximum flow that can be achieved when the flow control is opened fully has been included as a marking requirement on the device so that the user will know what could be delivered to the patient. A rationale has also been added to cover this marking requirement;

— a new requirement has been added for stability of setting;

— the environmental conditions have been aligned with IEC 60601-1-12, emergency equipment, as *flow control devices* are used in such environments; and

— the requirement for accuracy has been rationalised for clarity.

© ISO 2023 – All rights reserved
Documents
Document Sharing Hub
Full document details: BS EN ISO 15002:2024

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html

© ISO 2023 – All rights reserved

Documents
Document Sharing Hub
Full document details: BS EN ISO 15002:2024

Introduction

Flow control devices are used to administer a prescribed flow of gas to a patient interface device (e.g. nasal cannula, facemask) from a pressure gas source, such as a medical gas supply system. These devices need to deliver accurate flows under varying conditions of temperature and inlet pressures. Therefore, it is important that the performance characteristics be specified and tested in a defined manner.

Annex A provides additional insight into the reasoning that led to the requirements and recommendations that have been incorporated in this document. It is considered that knowledge of the reasons for the requirements will not only facilitate the proper application of this document but will expedite any subsequent revisions.

© ISO 2023 – All rights reserved
Documents
Document Sharing Hub
Full document details: BS EN ISO 15002:2024

Flow control devices for connection to a medical gas supply system

1 Scope

1.1 This document specifies requirements for *flow control devices* that can be connected by the user either directly, by means of a probe or a *gas-specific* connector, or indirectly by means of a low-pressure hose assembly conforming with ISO 5359 to:

a) a terminal unit conforming with ISO 9170-1 of a medical gas pipeline system conforming with ISO 7396-1:2016;

b) the pressure outlet of a regulator conforming with ISO 10524-1:2018; or

c) to the pressure outlet of a valve integrated pressure regulator (VIPR) conforming with ISO 10524-3 (see 5.2 gas inlets).

1.2 This document applies to the following types of *flow control devices* (FCDs):

a) *flowmeters*;

b) *flowgauge FCDs*; and

c) *fixed orifice FCDs*.

NOTE *Flow control devices* that are classed as medical electrical equipment can be subject to additional requirements of IEC 60601-1.

1.3 This document applies to *flow control devices* for the following gases:

— oxygen;

— oxygen 93 %;

— nitrous oxide;

— medical air;

— carbon dioxide;

— oxygen/nitrous oxide mixture 50/50 (% volume fraction);

— oxygen-enriched air;

— helium;

— xenon; and

— specified mixtures of the gases listed above.

NOTE *Flow control devices* can be available for other gases.

1.4 This document does not apply to *flow control devices* that are:

a) for use with gases for driving surgical tools;

b) an integral part of a regulator (see ISO 10524-1:2018); or

© ISO 2023 – All rights reserved

Documents
Document Sharing Hub
Full document details: BS EN ISO 15002:2024