



BSI Standards Publication

## Medical devices — Sleep apnoea breathing therapy — Masks and application accessories

---

## National foreword

This British Standard is the UK implementation of EN ISO 17510:2025. It is identical to ISO 17510:2025. It supersedes PD CEN/TS 17510:2020 and BS EN ISO 17510:2020, which are withdrawn.

The UK participation in its preparation was entrusted to Technical Committee CH/121/9, Lung Ventilators & Related Equipment.

A list of organizations represented on this committee can be obtained on request to its committee manager.

### Contractual and legal considerations

This publication has been prepared in good faith, however no representation, warranty, assurance or undertaking (express or implied) is or will be made, and no responsibility or liability is or will be accepted by BSI in relation to the adequacy, accuracy, completeness or reasonableness of this publication. All and any such responsibility and liability is expressly disclaimed to the full extent permitted by the law.

This publication is provided as is, and is to be used at the recipient's own risk.

The recipient is advised to consider seeking professional guidance with respect to its use of this publication.

This publication is not intended to constitute a contract. Users are responsible for its correct application.

© The British Standards Institution 2025  
Published by BSI Standards Limited 2025

ISBN 978 0 539 30559 3

ICS 11.040.10

### Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 31 December 2025.

### Amendments/corrigenda issued since publication

| Date | Text affected |
|------|---------------|
|------|---------------|

English Version

## Medical devices - Sleep apnoea breathing therapy - Masks and application accessories (ISO 17510:2025)

Dispositifs médicaux - Thérapie respiratoire de l'apnée  
du sommeil - Masques et accessoires d'application (ISO  
17510:2025)

Medizinische Geräte - Schlafapnoe-Atemtherapie -  
Masken und Anwendungszubehör (ISO 17510:2025)

This European Standard was approved by CEN on 4 December 2025.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

**CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels**

## **European foreword**

This document (EN ISO 17510:2025) has been prepared by Technical Committee ISO/TC 121 "Anaesthetic and respiratory equipment" in collaboration with Technical Committee CEN/TC 215 "Respiratory and anaesthetic equipment" the secretariat of which is held by BSI.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by June 2026, and conflicting national standards shall be withdrawn at the latest by June 2026.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 17510:2020.

Any feedback and questions on this document should be directed to the users' national standards body/national committee. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

## **Endorsement notice**

The text of ISO 17510:2025 has been approved by CEN as EN ISO 17510:2025 without any modification.

# Contents

Page

|                                                                                                                       |           |
|-----------------------------------------------------------------------------------------------------------------------|-----------|
| <b>Foreword</b>                                                                                                       | <b>iv</b> |
| <b>Introduction</b>                                                                                                   | <b>v</b>  |
| <b>1 Scope</b>                                                                                                        | <b>1</b>  |
| <b>2 Normative references</b>                                                                                         | <b>1</b>  |
| <b>3 Terms and definitions</b>                                                                                        | <b>2</b>  |
| <b>4 Information supplied by the manufacturer</b>                                                                     | <b>8</b>  |
| 4.1 General                                                                                                           | 8         |
| 4.2 Accompanying information                                                                                          | 8         |
| <b>5 Construction requirements</b>                                                                                    | <b>10</b> |
| 5.1 Mask connectors                                                                                                   | 10        |
| 5.2 Biocompatibility                                                                                                  | 10        |
| 5.2.1 Patient contacting                                                                                              | 10        |
| 5.2.2 Gas pathway contacting                                                                                          | 10        |
| 5.3 Protection against rebreathing                                                                                    | 10        |
| 5.3.1 Normal condition protection                                                                                     | 11        |
| 5.3.2 Single fault condition protection                                                                               | 11        |
| 5.4 Cleaning, disinfection, and sterilization                                                                         | 11        |
| 5.4.1 Single patient multiple use                                                                                     | 11        |
| 5.4.2 Multiple patient multiple use                                                                                   | 12        |
| 5.5 Breathing during single fault condition                                                                           | 12        |
| 5.6 Breathing system filter                                                                                           | 13        |
| <b>6 Audible acoustic energy</b>                                                                                      | <b>13</b> |
| <b>7 Measurement uncertainty</b>                                                                                      | <b>13</b> |
| <b>Annex A (informative) Particular guidance and rationale</b>                                                        | <b>14</b> |
| <b>Annex B (normative) Exhaust flow test procedure</b>                                                                | <b>18</b> |
| <b>Annex C (normative) Resistance to flow (pressure drop)</b>                                                         | <b>20</b> |
| <b>Annex D (normative) Anti-asphyxia valve pressure testing</b>                                                       | <b>22</b> |
| <b>Annex E (normative) Determination of the inspiratory and expiratory pressure drop under single fault condition</b> | <b>24</b> |
| <b>Annex F (normative) Carbon dioxide rebreathing</b>                                                                 | <b>26</b> |
| <b>Annex G (normative) Audible acoustic energy</b>                                                                    | <b>29</b> |
| <b>Annex H (informative) Guide to information supplied by the manufacturer</b>                                        | <b>31</b> |
| <b>Annex I (informative) Reference to the IMDRF essential principles and labelling principles</b>                     | <b>32</b> |
| <b>Annex J (informative) Terminology — Alphabetized index of defined terms</b>                                        | <b>34</b> |
| <b>Bibliography</b>                                                                                                   | <b>36</b> |

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

ISO draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at [www.iso.org/patents](http://www.iso.org/patents). ISO shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html).

This document was prepared jointly by Technical Committee ISO/TC 121, *Anaesthetic and respiratory equipment*, Subcommittee SC 3, *Respiratory devices and related equipment used for patient care*, and Technical Committee IEC/TC 62, *Medical equipment, software, and systems*, Subcommittee SC D, *Particular medical equipment, software, and systems*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 215, *Respiratory and anaesthetic equipment*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This second edition cancels and replaces the first edition (ISO 17510:2015), which has been technically revised.

The main changes are as follows:

- harmonization with IEC 60050-880 sources, where appropriate;
- adding disclosure requirements for magnets in *headgear*;
- updated *processing* requirements;
- updated noise requirements;
- referencing ISO 18562-1, for *biocompatibility of gas pathways*;
- harmonization with ISO 20417, where appropriate.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at [www.iso.org/members.html](http://www.iso.org/members.html).

## Introduction

Sleep apnoea is the clinically significant intermittent absences of normal respiration occurring during sleep. The awareness of the *risks* associated with sleep apnoea has grown significantly in recent years. As a result, the use of *sleep apnoea breathing therapy equipment* has become common. This document covers basic safety and essential performance requirements for *masks* and other application *accessories* needed to protect *patients* during use of this equipment.

*Sleep apnoea breathing therapy equipment* is covered by ISO 80601-2-70. [Figure A.1](#) shows the typical elements of this document together with the *sleep apnoea breathing therapy equipment* of ISO 80601-2-70 that form a sleep apnoea breathing system.

In this document, the following print types are used:

- Requirements and definitions: roman type.
- Informative material appearing outside of tables, such as notes, examples, and references: in smaller type. Normative text of tables is also in a smaller type.
- *Terms defined in [Clause 3](#) in this document or as noted: italics.*

In referring to the structure of this document, the term:

- “clause” means one of the numbered divisions within the table of contents, inclusive of all subdivisions (e.g. [Clause 5](#) includes [5.1](#), [5.2](#), etc.);
- “subclause” means a numbered subdivision of a clause (e.g. [5.1](#), [5.2](#), and [5.3.1](#) are all subclauses of [Clause 5](#)).

References to clauses within this document are preceded by the term “Clause” followed by the clause number. References to subclauses within this particular document are by number only.

In this document, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

In this document, the following verbal forms are used:

- “shall” indicates a requirement;
- “should” indicates a recommendation;
- “may” indicates a permission; and
- “can” is used to describe a possibility or capability.



# Medical devices — Sleep apnoea breathing therapy — Masks and application accessories

## 1 Scope

This document specifies requirements for *masks* and *accessories*, including any connecting element, that are required to connect the *patient-connection port* of *sleep apnoea breathing therapy equipment* to a *patient* for the application of sleep apnoea breathing therapy (e.g. nasal masks, exhaust ports and headgear).

This document applies to *masks* and their *accessories* used to connect *sleep apnoea breathing therapy equipment* to the *patient*.

The requirements in this document take priority over the requirements in ISO 18190.

This document does not cover *oral appliances*.

NOTE This document has been prepared to address the relevant *essential principles*<sup>[14]</sup> and labelling principles<sup>[15]</sup> of the International Medical Devices Regulators Forum (IMDRF) as indicated in [Annex I](#).

## 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 3744:2010, *Acoustics — Determination of sound power levels and sound energy levels of noise sources using sound pressure — Engineering methods for an essentially free field over a reflecting plane*

ISO 4871:1996, *Acoustics — Declaration and verification of noise emission values of machinery and equipment*

ISO 5356-1:2015, *Anaesthetic and respiratory equipment — Conical connectors — Part 1: Cones and sockets*

ISO 5356-2:2012, *Anaesthetic and respiratory equipment — Conical connectors — Part 2: Screw-threaded weight-bearing connectors*

ISO 5356-2:2012/Amd 1:2019, *Anaesthetic and respiratory equipment — Conical connectors — Part 2: Screw-threaded weight-bearing connectors — Amendment 1*

ISO 10993-1:2018, *Biological evaluation of medical devices — Part 1: Evaluation and testing within a risk management process*

ISO 14937:2009, *Sterilization of health care products — General requirements for characterization of a sterilizing agent and the development, validation and routine control of a sterilization process for medical devices*

ISO 17664-1:2021, *Processing of health care products — Information to be provided by the medical device manufacturer for the processing of medical devices — Part 1: Critical and semi-critical medical devices*

ISO 17664-2:2021, *Processing of health care products — Information to be provided by the medical device manufacturer for the processing of medical devices — Part 2: Non-critical medical devices*

ISO 18562-1:2024, *Biocompatibility evaluation of breathing gas pathways in healthcare applications — Part 1: Evaluation and testing within a risk management process*

ISO 20417:—<sup>1)</sup>, *Medical devices — Information to be supplied by the manufacturer*

1) Under preparation. Stage at the time of publication: ISO/DIS 20417:2025.

ISO 23328-1:2003, *Breathing system filters for anaesthetic and respiratory use — Part 1: Salt test method to assess filtration performance*

ISO 23328-2:2002, *Breathing system filters for anaesthetic and respiratory use — Part 2: Non-filtration aspects*

IEC 61672-1:2013, *Electroacoustics — Sound level meters — Part 1: Specifications*

IEC Guide 115:2023, *Application of uncertainty of measurement to conformity assessment activities in the electrotechnical sector*

### 3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 20417:— and the following apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <https://www.electropedia.org/>

NOTE An alphabetical index of defined terms is found in [Annex J](#).

#### 3.1

##### **anti-asphyxia valve**

valve used on a breathing *mask* intended to allow spontaneous breathing when the lung ventilator or breathing therapy equipment is not providing adequate pressure or flow

[SOURCE: ISO 4135:2022, 3.6.3.7]

#### 3.2

##### **biocompatibility**

ability of a *medical device*, *accessory* or material to perform with an appropriate host response in a specific application

Note 1 to entry: A *medical device* or *accessory* can produce some level of adverse effect, but that level can be determined to be acceptable when considering the *benefit* provided.

[SOURCE: ISO 18562-1:2024, 3.6]

#### 3.3

##### **breathing system filter**

##### **BSF**

device intended to reduce transmission of particulates, including microorganisms, in a breathing system

[SOURCE: ISO 4135:2022, 3.6.1.5]

#### 3.4

##### **breathing tube**

non-rigid tube used to convey gases or vapours within the *user-detachable* section of a breathing system

[SOURCE: ISO 4135:2022, 3.1.4.4, modified — Deleted Note 1 to entry.]

#### 3.5

##### **cleaning**

*process* to remove contaminants to the extent necessary for further *processing* or for *intended use*

Note 1 to entry: *Cleaning* of a used *product* consists of the removal of adherent soil (e.g. blood, protein substances and other debris) from the surfaces, crevices, serrations, joints and lumens of a *medical device* by a manual or automated *process* that prepares the items for safe handling or further *processing*.

Note 2 to entry: *Cleaning* of a new *product* can occur before initial use or during production before release for distribution.

[SOURCE: ISO 17664-2:2021, 3.1, modified — Replaced “removal of” with “*process to remove*”. In Note 1 to entry, replaced “and/or” with “or” and added “of a used *product*”. Added Note 2 to entry.]

### 3.6

#### conformity

fulfilment of a requirement

Note 1 to entry: In English the word “conformance” is synonymous but deprecated. In French the word “compliance” is synonymous but deprecated.

[SOURCE: ISO 9000:2015, 3.6.11, modified — Deleted Note 2 to entry.]

### 3.7

#### disinfection

*process to inactivate viable microorganisms to a level previously specified as being appropriate for a defined purpose*

[SOURCE: ISO 17664-2:2021, 3.5]

### 3.8

#### exhaust flow

<airway device> flow from the *mask* or application *accessory* to atmosphere other than the leak due to improper seal to the face

Note 1 to entry: The *exhaust flow* can pass through openings in the *mask*, the connecting element and the *mask* or through the *anti-asphyxia valve*.

Note 2 to entry: The *exhaust flow* discharges exhaled gases to atmosphere to reduce *rebreathing* of CO<sub>2</sub>.

[SOURCE: ISO 4135:2022, 3.8.1.3]

### 3.9

#### exhaust port

port through which gas is discharged to the atmosphere or to an anaesthetic gas scavenging system

[SOURCE: ISO 4135:2022, 3.1.4.11, modified — Deleted notes to entry.]

### 3.10

#### gas output port

port of the device through which gas is delivered at respiratory pressures to a *user-detachable* part of a breathing system

[SOURCE: ISO 4135:2022, 3.1.4.22, modified — Deleted Note 1 to entry and deprecated term.]

### 3.11

#### gas pathway

interior surfaces over which respiratory gases or liquids pass that can be inspired

EXAMPLE 1 The ventilator breathing *system*, inlet filter, gas mixer, blower and internal piping.

EXAMPLE 2 Enclosed chamber of an incubator including the mattress or the inner surface of an oxygen hood.

EXAMPLE 3 The inner surfaces of *breathing tubes*, tracheal tubes or *masks* and mouthpieces.

Note 1 to entry: The *gas pathway* is bounded by the fluid ports through which gases or liquids enter the *medical device* or *accessory*. This can include the *patient* interface or the interior surfaces of enclosures that are in contact with gases or liquids that can be inspired.

Note 2 to entry: The *gas pathway* can include some surfaces in the expiratory pathway.

Note 3 to entry: The fluids from a *gas pathway* can be ingested.

[SOURCE: ISO 18562-1:2024, 3.11, modified — In the definition, added “respiratory”. In Note 1 to entry, replaced “ports” with “fluid ports”. Replaced Note 3 to entry.]

### 3.12

#### **harm**

injury or damage to the health of people or animals, or damage to property or the environment

[SOURCE: ISO/IEC Guide 63:2019, 3.1, modified — Added “or animals”.]

### 3.13

#### **headgear**

part that is used to fix the *mask* to the *patient*

Note 1 to entry: The *headgear* may be an integral part of the *mask*.

### 3.14

#### **healthcare professional**, adj

appropriately trained, knowledgeable and skilled, providing systematic preventive, curative, promotional or rehabilitative healthcare services

EXAMPLE *Healthcare professional user, healthcare professional responsible organization.*

[SOURCE: ISO 80601-2-12:2023, 201.3.247, modified — Deleted “to families or communities”. Added example.]

### 3.15

#### **intended use**

use for which a product or *process* is intended according to the specifications, instructions for use and other *information supplied by the manufacturer*

Note 1 to entry: The intended medical indication, *patient* population, part of the body or type of tissue interacted with, *user* profile, use environment, and operating principle are typical elements of the *intended use*.

Note 2 to entry: *Intended use* is different from *normal use*. Both include the concept of use as intended by the *manufacturer*. However, *intended use* focuses on the medical purpose whereas *normal use* incorporates not only the medical purpose, but maintenance, transport, *processing*, etc. as well.

Note 3 to entry: Product includes service.

[SOURCE: ISO/IEC Guide 63:2019, 3.4, modified — Deleted “or service” and replaced “instructions” with “instructions for use”, “provided” with “supplied” and added “other”. Added Notes 2 and 3 to entry.]

### 3.16

#### **inspection**

*conformity* evaluation by observation and judgement accompanied as appropriate by measurement, testing or gauging

### 3.17

#### **manufacturer**

organization with responsibility for design or manufacture of a *medical device* or *accessory* with the intention of making the *medical device* or *accessory* available for use, under its name; whether such a *medical device* or *accessory* is designed or manufactured by that organization itself or on its behalf by another organization

Note 1 to entry: This organization has ultimate legal responsibility for ensuring compliance with all applicable policies for the *medical device* or *accessory* in the countries or jurisdictions where it is intended to be made available or sold, unless this responsibility is specifically imposed on another organization by the authority having jurisdiction.

Note 2 to entry: The *manufacturer's* responsibilities include meeting both pre-market requirements and post-market requirements, such as adverse event reporting and notification of corrective actions.

Note 3 to entry: “Design or manufacture” can include specification development, production, fabrication, assembly, *processing*, packaging, repackaging, labelling, relabelling, sterilization, installation or remanufacturing of a *medical device* or *accessory*; or putting a collection of *medical devices* or *accessories*, and possibly other products, together for a medical purpose.

Note 4 to entry: Any organization which assembles or adapts a *medical device* or *accessory* that has already been supplied by another organization for an individual *patient*, in accordance with the *instructions for use*, is not the *manufacturer*, provided the assembly or adaptation does not change the *intended use* of the *medical device* or *accessory*.

Note 5 to entry: Any organization which changes the *intended use* of or modifies a *medical device* or *accessory* without acting on behalf of the original *manufacturer* and which makes it available for use under its own name, is considered the *manufacturer* of the modified *medical device* or *accessory*.

Note 6 to entry: An authorized representative, distributor or importer who only adds their own address and contact details to the *medical device*, *accessory* or the packaging, without covering or changing the existing *information supplied by the manufacturer*, is not considered a *manufacturer*.

Note 7 to entry: To the extent that an *accessory* is subject to the applicable policies of a *medical device*, the organization responsible for the design or manufacture of that *accessory* is considered to be a *manufacturer*.

[SOURCE: ISO/IEC Guide 63:2019, 3.6, modified — Added “or *accessory*”, replaced “whether or not” with “whether”, “and/or” with “or”, “natural or legal person” and “person” with “organization”, “regulatory requirements” with “applicable policies”, “Regulatory Authority within that jurisdiction” with “authority having jurisdiction”, inserted “IMDRF” and replaced “labelling” with “*information supplied by the manufacturer*”.]

### 3.18

#### **mask**

device which provides a non-invasive interface between the *patient's* airway and a *patient-connection port* or other connection to a source of respirable gas

[SOURCE: ISO 4135:2022, 3.8.6.4]

### 3.19

#### **ME equipment**

##### **medical electrical equipment**

electrical *medical device* having an applied part or transferring energy or substances to or from the *patient* or detecting such energy or substance transfer to or from the *patient*

Note 1 to entry: If *ME equipment* is provided with more than one connection to a particular mains, then the combination of the *ME equipment* and its mains connections is a medical electrical system.

Note 2 to entry: The *ME equipment* includes those *accessories* and detachable parts as defined by the *manufacturer*.

Note 3 to entry: Not all electrical equipment used in medical practice falls within this definition.

Note 4 to entry: The definition of *medical device* is broader than that of *ME equipment*.

Note 5 to entry: The transfer or detecting of energy includes optical radiation, ionising radiation, ultrasound and magnetic resonance.

[SOURCE: IEC 60601-1:2005, 3.63, modified — Replaced “electrical equipment” with “electrical *medical device*” and “supply mains” with “mains”, added “or substances”, deleted item b), deleted “and which is provided with not more than one connection to a particular mains”. Added Notes 1, 2, 4 and 5 to entry.]

### 3.20

#### **normal condition**

condition of a product in which all *risk control measures* are intact

[SOURCE: IEC 60601-1:2005, 3.70, modified — Added “of a product” and replaced “means provided for protection against *hazards*” with “*risk control measures*”.]

### 3.21

#### **oral appliance**

device intended to maintain the oral airway by mechanical means and which achieves its purpose independently of *sleep apnoea breathing therapy equipment*

### 3.22

#### **patient-connection port**

port of a breathing system intended for connection to an airway device

Note 1 to entry: The *patient-connection port* is the end of the breathing system proximal to the *patient*.

Note 2 to entry: The *patient-connection port* is typically a connector suitable for connection to an airway device such as a tracheal tube, tracheostomy tube, face *mask* or supralaryngeal airway.

Note 3 to entry: Current product standards typically specify that the *patient-connection port* is required to be in the form of specific standardized connectors, for example, a connector conforming to ISO 5356-1.

Note 4 to entry: In anaesthetic and respiratory equipment where its function is dependent upon a design feature of the component that connects the equipment to the *patient's* airway, then there is no *patient-connection-port*. Examples include non-invasive ventilation using a face *mask* with an integrated expiratory port, or oxygen therapy using an open *mask*.

[SOURCE: ISO 4135:2022, 3.1.4.41]

### 3.23

#### **procedure**

specified way to carry out an activity or a *process*

Note 1 to entry: *Procedures* can be documented or not.

[SOURCE: ISO 9000:2015, 3.4.5]

### 3.24

#### **process**

set of interrelated or interacting activities that use inputs to deliver an intended result

Note 1 to entry: Whether the “intended result” of a *process* is called output, product or service depends on the context of the reference.

Note 2 to entry: Inputs to a *process* are generally the outputs of other *processes* and outputs of a *process* are generally the inputs to other *processes*.

Note 3 to entry: Two or more interrelated and interacting *processes* in series can also be referred to as a *process*.

[SOURCE: ISO 9000:2015, 3.4.1, modified — Deleted Notes 4, 5 and 6 to entry.]

### 3.25

#### **rated**

assigned by the *manufacturer* for a specified operating condition

Note 1 to entry: A *rated* value can have a tolerance.

[SOURCE: IEC 60601-1:2005, 3.97, modified — Deleted “term referring to a value”.]

### 3.26

#### **rebreathing**

inhalation of expired gas mixture from which carbon dioxide may or may not have been removed

[SOURCE: ISO 4135:2022, 3.1.5.12]

### 3.27

#### **risk**

combination of the probability of occurrence of *harm* and the severity of that *harm*

Note 1 to entry: The probability of occurrence includes the exposure to a hazardous situation and the possibility to avoid or limit the *harm*.

Note 2 to entry: In French, the term “risqué” also denotes the potential source of *harm*, in English “hazard”.

[SOURCE: IEC 60050-903:2013, 903-01-07, modified — Added Note 1 to entry.]