BS EN ISO 18778:2022

BSI Standards Publication

Respiratory equipment — Particular requirements for basic safety and essential performance of infant cardiorespiratory monitors

Documents
Document Sharing Hub

Full document details: BS EN ISO 18778:2022

National foreword

This British Standard is the UK implementation of EN ISO 18778:2022. It is identical to ISO 18778:2022. It supersedes BS EN ISO 18778:2009, which is withdrawn.

The UK participation in its preparation was entrusted to Technical Committee CH/121/9, Lung Ventilators & Related Equipment.

A list of organizations represented on this committee can be obtained on request to its committee manager.

Contractual and legal considerations

This publication has been prepared in good faith, however no representation, warranty, assurance or undertaking (express or implied) is or will be made, and no responsibility or liability is or will be accepted by BSI in relation to the adequacy, accuracy, completeness or reasonableness of this publication. All and any such responsibility and liability is expressly disclaimed to the full extent permitted by the law.

This publication is provided as is, and is to be used at the recipient's own risk.

The recipient is advised to consider seeking professional guidance with respect to its use of this publication.

This publication is not intended to constitute a contract. Users are responsible for its correct application.

© The British Standards Institution 2022
Published by BSI Standards Limited 2022

ISBN 978 0 539 12988 5

ICS 11.040.10; 11.040.55

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 31 October 2022.

Amendments/corrigenda issued since publication

Date	Text affected

Documents
Document Sharing Hub

Full document details: BS EN ISO 18778:2022

EUROPEAN STANDARD

NORME EUROPÉENNE

EUROPÄISCHE NORM

EN ISO 18778

October 2022

ICS 11.040.10; 11.040.55

Supersedes EN ISO 18778:2009

English Version

Respiratory equipment - Particular requirements for basic safety and essential performance of infant cardiorespiratory monitors (ISO 18778:2022)

Matériel respiratoire - Exigences particulières relatives à la sécurité de base et aux performances essentielles des moniteurs cardiorespiratoires pour nourrissons (ISO 18778:2022)

Medizinische elektrische Geräte - Besondere Festlegungen für die Sicherheit einschließlich der wesentlichen Leistungsmerkmale von kardiorespiratorischen Überwachungsgeräten für Kleinkinder (ISO 18778:2022)

This European Standard was approved by CEN on 17 June 2022.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and United Kingdom.

EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

© 2022 CEN All rights of exploitation in any form and by any means reserved worldwide for CEN national Members.

Ref. No. EN ISO 18778:2022 E

Documents
Document Sharing Hub

Full document details: BS EN ISO 18778:2022

European foreword

This document (EN ISO 18778:2022) has been prepared by Technical Committee ISO/TC 121 "Anaesthetic and respiratory equipment" in collaboration with Technical Committee CEN/TC 215 “Respiratory and anaesthetic equipment” the secretariat of which is held by BSI.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by April 2023, and conflicting national standards shall be withdrawn at the latest by April 2023.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 18778:2009.

Any feedback and questions on this document should be directed to the users’ national standards body/national committee. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

Endorsement notice

The text of ISO 18778:2022 has been approved by CEN as EN ISO 18778:2022 without any modification.

Documents
Document Sharing Hub

Full document details: BS EN ISO 18778:2022

Contents

Page

Foreword ... **vi**

Introduction ... **viii**

1 Scope ... **1**

2 Normative references ... **1**

3 Terms and definitions ... **2**

4 General requirements ... **6**
4.1 General ... 6
4.2 *Essential performance* ... 6
4.3 *ME equipment* or *ME system* parts that contact the *patient* ... 6
4.4 *Single fault condition* for *ME equipment* ... 6

5 General requirements for testing of *ME equipment* ... **6**
5.1 General ... 6
5.2 *Infant cardiorespiratory monitor* testing errors ... 6

6 Classification of *ME equipment* and *ME systems* ... **7**
6.1 General ... 7
6.2 Additional requirements for classification of *ME equipment* and *ME systems* ... 7

7 *ME equipment* identification, *marking* and documents ... **7**
7.1 General ... 7
7.2 *Information to be supplied by the manufacturer* ... 7
7.3 Additional requirements for *accessories* ... 7
7.4 Additional requirements for *marking* on the outside of *ME equipment* or *ME equipment* parts ... 8
7.5 General *instructions for use* ... 8
7.6 Additional requirements for warnings and safety notices ... 8
7.7 Additional requirements for start-up *procedure* ... 9
7.8 Additional requirements for operating instructions ... 9
7.8.1 General ... 9
7.8.2 *Lay operator* operating instructions ... 9
7.8.3 *Healthcare professional operator* operating instructions ... 10
7.9 *Cleaning*, *disinfection*, and *sterilization* ... 10
7.10 Additional requirements for maintenance ... 10
7.11 Additional requirements for *accessories*, supplementary equipment, used material ... 10
7.12 Additional requirements for the *technical description* ... 10

8 Protection against electrical *hazards* from *ME equipment* ... **11**

9 Protection against *mechanical hazards* of *ME equipment* and *ME systems* ... **11**
9.1 General ... 11
9.2 Additional requirements for instability from unwanted lateral movement ... 11
9.3 Grips and other handling devices ... 11

10 Protection against unwanted and excessive radiation *hazards* ... **11**

11 Protection against excessive temperatures and other *hazards* ... **11**
11.1 General ... 11
11.2 *Cleaning* and *disinfection* of *ME equipment* or *ME system* ... 12
11.3 *Sterilization* of *ME equipment* or *ME system* ... 12
11.4 *Biocompatibility* of *ME equipment* and *ME systems* ... 12
11.5 Interruption of the power supply / *supply mains* to *ME equipment* ... 13
11.5.1 General ... 13
11.5.2 Power sources ... 13
11.5.3 Alternative power supply/*supply mains* ... 14

12 Accuracy of controls and instruments and protection against hazardous outputs ... **14**

AVA Documents
Document Sharing Hub

© ISO 2022 – All rights reserved

Full document details: BS EN ISO 18778:2022

12.1 General ... 14
12.2 Accuracy of controls and instruments ... 14
12.3 Accuracy of controls and instruments ... 15
12.3.1 General ... 15
12.3.2 Cardiorespiratory monitoring ... 15
12.3.3 Direct monitoring - respiration ... 15
12.3.4 Indirect monitoring – heart rate ... 15
12.3.5 Indirect monitoring from pulse oximetry ... 15
12.3.6 Apnoeic *patient alarm condition* ... 16
12.3.7 Sensor fault ... 17
12.3.8 Clinical performance evaluation ... 17
12.4 *Usability* of *ME equipment* ... 17

13 *Hazardous situations* and fault conditions for *ME equipment* ... 18

14 *Programmable electrical medical systems (PEMS)* ... 18

15 Construction of *ME equipment* ... 18
15.1 General ... 18
15.2 Mode of operation ... 18
15.3 Pre-use check ... 18

16 *ME systems* ... 19

17 Electromagnetic compatibility of *ME equipment* and *ME systems* ... 19

18 Requirements for the *accessories* ... 19
18.1 General ... 19
18.2 Labelling ... 19

19 Training ... 19

20 *Functional connection* ... 19
20.1 General ... 19
20.2 Connection to an electronic health record ... 19
20.3 Connection to a *distributed alarm system* ... 19

21 Electromagnetic disturbances – Requirements and tests ... 20
21.1 General ... 20
21.2 Compliance criteria ... 20
21.3 Requirements applicable to all *ME equipment* and *ME systems* ... 20
21.4 Additional general requirements ... 20

22 *Usability* ... 20
22.1 General ... 20
22.2 *Primary operating functions* ... 21

23 General requirements, tests and guidance for *alarm systems* in *medical electrical equipment* and *medical electrical systems* ... 21
23.1 General ... 21
23.2 Volume and characteristics of auditory *alarm signals* and *information signals* ... 21
23.3 Additional requirements for termination of *alarm signal* inactivation ... 21
23.4 Additional requirements for *alarm system* logging ... 22

24 Requirements for *medical electrical equipment* and *medical electrical systems* used in the *home healthcare environment* ... 22

Annex A (informative) **General guidance and rationale ... 23**

Annex B (informative) **Sequence of testing ... 28**

Annex C (informative) **Guide to *marking* and labelling requirements for *ME equipment* and *ME systems* ... 29**

Annex D (informative) ***Symbols* on *marking* ... 32**

© ISO 2022 – All rights reserved
Documents
Document Sharing Hub
Full document details: BS EN ISO 18778:2022

Annex E (informative) **Examples of the connection of the measuring device (MD) for measurement of the *patient leakage current* and *patient auxiliary current*** 33
Annex F (informative) **Suitable measuring supply circuits** 34
Annex G (informative) **Protection against *hazards* of ignition of flammable anaesthetic mixtures** 35
Annex H (informative) ***PEMS* structure, *PEMS development life-cycle* and documentation** 36
Annex I (informative) ***ME systems* aspects** 37
Annex J (informative) **Survey of insulation paths** 38
Annex K (informative) **Simplified *patient leakage current* diagrams** 39
Annex L (informative) **Insulated winding wires for use without interleaved insulation** 40
Annex M (informative) **Reduction of pollution degrees** 41
Annex N (informative) **Data interface requirements** 42
Annex O (informative) **Considerations for a clinical performance study** 45
Annex P (informative) **Reference to the IMDRF *essential principles* and labelling guidances** 47
Annex Q (informative) **Reference to the *essential principles*** 49
Annex R (informative) **Reference to the general safety and performance requirements** 51
Bibliography 54
Terminology — Alphabetized index of defined terms 55

© ISO 2022 – All rights reserved

Documents
Document Sharing Hub
Full document details: BS EN ISO 18778:2022

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 121, *Anaesthetic and respiratory equipment*, Subcommittee SC 3, *Lung ventilators and related equipment*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 215, *Respiratory and anaesthetic equipment*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This second edition cancels and replaces the first (ISO 18778:2005), which has been technically revised.

The main changes are as follows:

— extending the scope to include the *infant cardiorespiratory monitor* and its *accessories,* where the characteristics of those *accessories* can affect the *basic safety* or *essential performance* of the *infant cardiorespiratory monitor,* and thus not only the *infant cardiorespiratory monitor* itself;

— identification of *essential performance* of an *infant cardiorespiratory monitor* and its *accessories*;

— harmonization with the third edition of IEC 60601-1;

and the following additions:

— tests for *infant cardiorespiratory monitor* performance;

— tests for mechanical strength (via IEC 60601-1-11);

— requirements for *transit-operable* use;

— new *symbols*;

— requirements for an *infant cardiorespiratory monitor* as a component of an *ME system*;

— requirement for both a direct measurement of respiration, and an indirect measurement of apnoeic activity;

— tests for *enclosure* integrity (water ingress via IEC 60601-1-11);

© ISO 2022 – All rights reserved
Documents
Document Sharing Hub
Full document details: BS EN ISO 18778:2022

— tests for *cleaning* and *disinfection procedures* (via IEC 60601-1-11); and

— harmonization with ISO 20417.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

© ISO 2022 – All rights reserved

Documents
Document Sharing Hub
Full document details: BS EN ISO 18778:2022

Introduction

This document specifies requirements for *infant cardiorespiratory monitors* called in previous working documents "infant apnoea monitors or infant monitors". *Infant cardiorespiratory monitors* are intended to be used primarily to monitor cardiorespiratory parameters for *patients* less than 3 years of age. *Infant cardiorespiratory monitors* are required to include at least one direct measurement of respiration and one indirect measurement of apnoeic activity such as heart rate or oxygen saturation. *Infant cardiorespiratory monitors* are intended for use in the *home healthcare environment. Infant cardiorespiratory monitors* are frequently used in locations where *supply mains* is not reliable. *Infant cardiorespiratory monitors* are often supervised by non-healthcare personnel (*lay operators*) with varying levels of training. An *infant cardiorespiratory monitor* conforming with this document can be used elsewhere (i.e., in healthcare facilities).

Annex A contains guidance or rationale to indicated clauses and subclauses.

Annex C contains a guide to the *marking* and labelling requirements in this document.

Annex D contains a summary of the *symbols* referenced in this document.

If a clause or subclause is specifically intended to be applicable to *ME equipment* only, or to *ME systems* only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to *ME equipment* and to *ME systems*, as relevant.

Hazards inherent in the intended physiological function of *ME equipment* or *ME systems* within the scope of this document are not covered by specific requirements in this document except in IEC 60601-1:2005+AMD1:2012+AMD2:2020, 7.2.13 and 8.4.1.

NOTE 1 Additional information can be found in IEC 60601-1:2005+AMD1:2012+AMD2:2020, 4.2.

The object of this document is to establish particular *basic safety* and *essential performance* requirements for an *infant cardiorespiratory monitor,* as defined in 3.10, and its *accessories*.

Accessories are included because the combination of the *infant cardiorespiratory monitor* and the *accessories* needs to be adequately safe. *Accessories* can have a significant impact on the *basic safety* or *essential performance* of the *infant cardiorespiratory monitor*.

NOTE 2 This document has been prepared to address the relevant *essential principles*[6] and labelling[7] guidances of the International Medical Devices Regulators Forum (IMDRF) as indicated in Annex P.

NOTE 3 This document has been prepared to address the relevant *essential principles of safety and performance* of ISO 16142-1:2016 as indicated in Annex Q.

NOTE 4 This document has been prepared to address the relevant general safety and performance requirements of European regulation (EU) 2017/745[8] as indicated in Annex R.

© ISO 2022 – All rights reserved
Documents
Document Sharing Hub
Full document details: BS EN ISO 18778:2022

Respiratory equipment — Particular requirements for basic safety and essential performance of infant cardiorespiratory monitors

1 Scope

This document applies to the *basic safety* and *essential performance* of an *infant cardiorespiratory monitor,* as defined in 3.10, hereafter also referred to as *ME equipment,* in combination with its *accessories*:

— intended for use in the *home healthcare environment*;

— intended for use by a *lay operator*;

— intended to monitor cardiorespiratory parameters in sleeping or resting children under three years of age; and

— intended for *transit-operable* use.

NOTE An *infant cardiorespiratory monitor* can also be used in professional health care facilities.

This document is also applicable to those *accessories* intended by their *manufacturer* to be connected to the *infant cardiorespiratory monitor*, where the characteristics of those *accessories* can affect the *basic safety* or *essential performance* of the *infant cardiorespiratory monitor*.

EXAMPLE probes, cables *distributed alarm system*

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 10993-1:2018, *Biological evaluation of medical devices — Part 1: Evaluation and testing within a risk management process*

ISO 14155:2020, *Clinical investigation of medical devices for human subjects — Good clinical practice*

ISO 16142-1:2016, *Medical devices — Recognized essential principles of safety and performance of medical devices — Part 1: General essential principles and additional specific essential principles for all non-IVD medical devices and guidance on the selection of standards*

ISO 17664-2:2021, *Processing of health care products — Information to be provided by the medical device manufacturer for the processing of medical devices — Part 2: Non-critical medical devices*

ISO 18562-1:2017, *Biocompatibility evaluation of breathing gas pathways in healthcare applications — Part 1: Evaluation and testing within a risk management process*

ISO 20417:2021, *Medical devices — Information to be supplied by the manufacturer*

ISO 80601-2-61:2017, *Medical electrical equipment — Part 2-61: Particular requirements for basic safety and essential performance of pulse oximeter equipment*

IEC 60601-1:2005+AMD1:2012+AMD2:2020, *Medical electrical equipment — Part 1: General requirements for basic safety and essential performance*

© ISO 2022 – All rights reserved

Documents
Document Sharing Hub
Full document details: BS EN ISO 18778:2022

IEC 60601-1-2:2014+AMD1:2020, *Medical electrical equipment — Part 1-2: General requirements for basic safety and essential performance — Collateral Standard: Electromagnetic disturbances - Requirements and tests*

IEC 60601-1-6:2010+AMD1:2013+AMD2:2020, *Medical electrical equipment — Part 1-6: General requirements for basic safety and essential performance — Collateral standard: Usability*

IEC 60601-1-8:2006+AMD1:2012+AMD2:2020, *Medical electrical equipment — Part 1-8: General requirements for basic safety and essential performance — Collateral Standard: General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems*

IEC 60601-1-11:2015+AMD1:2020, *Medical electrical equipment — Part 1-11: General requirements for basic safety and essential performance — Collateral Standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment*

IEC 60601-2-27:2011, *Medical electrical equipment – Part 2-27: Particular requirements for the basic safety and essential performance of electrocardiographic monitoring equipment*

IEC 62366-1:2015+AMD1:2020, *Medical devices — Application of usability engineering to medical devices*

IEC 62570:2014, *Standard practice for marking medical devices and other items for safety in the magnetic resonance environment*

IEC Guide 115:2021, *Application of uncertainty of measurement to conformity assessment activities in the electrotechnical sector*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 16142-1, ISO 17664-2, ISO 18562-1, IEC 60601-1, IEC 60601-1-2, IEC 60601-1-6, IEC 60601-1-8, IEC 60601-1-11, IEC 62366-1, and the following apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

— ISO Online browsing platform: available at https://www.iso.org/obp

— IEC Electropedia: available at https://www.electropedia.org/

3.1
accompanying information
information accompanying or *marked* on a *medical device* or *accessory* for the *user* or those accountable for the installation, use, *processing*, maintenance, decommissioning and disposal of the *medical device* or *accessory*, particularly regarding safe use

Note 1 to entry: The *accompanying information* shall be regarded as part of the *medical device* or *accessory.*

Note 2 to entry: The *accompanying information* can consist of the *label*, *marking*, *instructions for use*, *technical description*, installation manual, quick reference guide, etc.

Note 3 to entry: *Accompanying information* is not necessarily a written or printed document but could involve auditory, visual, or tactile materials and multiple media types (e.g., CD/DVD-ROM, USB stick, website).

[SOURCE: ISO 20417:2021, 3.2, modified — deleted note 4.]

3.2
apnoea
cessation of breathing lasting 10 s or more

© ISO 2022 – All rights reserved
Documents
Document Sharing Hub
Full document details: BS EN ISO 18778:2022

3.3
biocompatibility
ability to be in contact with a living system without producing an unacceptable adverse effect

Note 1 to entry: Medical devices may produce some level of adverse effect, but that level may be determined to be acceptable when considering the benefits provided by the medical device.

[SOURCE: ISO 18562-1:2017, 3.2]

3.4
central apnoea
apnoea where there is a cessation of output from the central respiratory centres, and no respiratory effort

3.5
cleaning
removal of contaminants to the extent necessary for further *processing* or for *intended use*

Note 1 to entry: *Cleaning* consists of the removal, usually with detergent and water, of adherent soil (e.g. blood, protein substances, and other debris) from the surfaces, crevices, serrations, joints, and lumens of a medical device by a manual or automated *process* that prepares the items for safe handling or further *processing*.

[SOURCE: ISO 17664-2:2017, 3.1, modified — replaced "and/or" with "or" in note 1.]

3.6
clinical investigation
systematic investigation in one or more human subjects, undertaken to assess the clinical performance, effectiveness or safety of a medical device

Note 1 to entry: For the purpose of this document, “clinical trial” or “clinical study” are synonymous with “clinical investigation”.

[SOURCE: ISO 14155:2020, 3.8]

3.7
clinical investigation plan
CIP
document that states the rationale, objectives, design and pre-specified analysis, methodology, organization, monitoring, conduct and record-keeping of the *clinical investigation* (3.6)

Note 1 to entry: For the purpose of this document “protocol” is synonymous with “*CIP*”. However, protocol has many different meanings, some not related to *clinical investigation*, and these can differ from country to country. Therefore, the term *CIP* is used in this document.

[SOURCE: ISO 14155:2020, 3.9]

3.8
disinfection
process to reduce the number of viable microorganisms to a level previously specified as being appropriate for a defined purpose

[SOURCE: ISO 17664-2:2021, 3.3]

3.9
healthcare professional
individual with appropriate training, knowledge and skills who provides preventive, curative, promotional or rehabilitative healthcare services in a systematic way to people, families or communities

[SOURCE: ISO 4135:2022, 3.1.6.2]

3.10
infant cardiorespiratory monitor
ME equipment intended to monitor cardiorespiratory parameters for *patients* less than 3 years of age

© ISO 2022 – All rights reserved

AVA Documents
Document Sharing Hub
Full document details: BS EN ISO 18778:2022

3.11
information supplied by the manufacturer
information related to the identification and use of a *medical device* or *accessory*, in whatever form provided, intended to ensure the safe and effective use of the *medical device* or *accessory*

Note 1 to entry: For the purposes of this document, e-documentation is included in *information supplied by the manufacturer*.

Note 2 to entry: For the purposes of this document, shipping documents and promotional material are excluded from *information supplied by the manufacturer*. However, some authorities having jurisdiction can consider such supplemental information as *information supplied by the manufacturer*.

Note 3 to entry: The primary purpose of *information supplied by the manufacturer* is to identify the *medical device* and its *manufacturer*, and provide essential information about its safety, performance, and appropriate use to the *user* or other relevant persons.

[SOURCE: ISO 20417:2021, 3.10, modified — deleted note 4.]

3.12
instructions for use
IFU
portion of the *accompanying information* that is essential for the safe and effective use of a *medical device* or *accessory* directed to the *user* of the *medical device*

Note 1 to entry: For the purposes of this document, a *user* can be either a *lay user* or professional *user* with relevant specialized training.

Note 2 to entry: For the purposes of this document, instructions for the professional *processing* between uses of a *medical device* or *accessory* can be included in the *instructions for use*.

Note 3 to entry: The *instructions for use*, or portions thereof, can be located on the display of a *medical device* or *accessory*.

Note 4 to entry: *Medical devices* or *accessories* that can be used safely and effectively without *instructions for use* are exempted from having *instructions for use* by some *authorities having jurisdiction*.

[SOURCE: ISO 20417:2021, 3.11, modified — deleted note 5.]

3.13
marking
information, in text or graphical format, durably affixed, printed, etched (or equivalent) to a *medical device* or *accessory*

Note 1 to entry: For the purposes of this document, the term *marked* is used to designate the corresponding act.

Note 2 to entry: For the purposes of this document, *marking* is different from 'direct marking' as commonly described in unique device identification (UDI) standards and regulations. A UDI 'direct marking' is a type of *marking*.

[SOURCE: ISO 20417:2021, 3.16, modified — deleted note 3.]

3.14
obstructive apnoea
apnoea due to airway obstruction

3.15
processing
<preparation of medical device, *accessory*> activity to prepare a new or used *medical device* or *accessory* for its *intended use*

[SOURCE: ISO 20417:2021, 3.20]

© ISO 2022 – All rights reserved
Documents
Document Sharing Hub
Full document details: BS EN ISO 18778:2022

3.16
single use
<medical device, *accessory*> intended by the *manufacturer* to be used on an individual *patient* or specimen during a single *procedure* and then disposed of

Note 1 to entry: A *single use* medical device or *accessory* is not intended by its *manufacturer* to be further processed and used again.

[SOURCE: ISO 20417:2021, 3.26]

3.17
sterile
free from viable microorganisms

[SOURCE: ISO 20417:2021, 3.28]

3.18
sterilization
process used to render product free from viable microorganisms

Note 1 to entry: In a *sterilization process*, the nature of microbial inactivation is exponential and thus the survival of a microorganism on an individual item can be expressed in terms of probability. While this probability can be reduced to a very low number, it can never be reduced to zero.

[SOURCE: ISO 17664-2:2021, 3.17]

3.19
symbol
graphical representation appearing on the label and/or associated documentation of a medical device that communicates characteristic information without the need for the supplier or receiver of the information to have knowledge of the language of a particular nation or people

Note 1 to entry: The *symbol* can be an abstract pictorial or a graphical representation, or one that uses familiar objects, including alphanumeric characters (with sufficient justification).

[SOURCE: ISO 20417:2021, 3.29]

3.20
technical description
portion of the *accompanying information* directed to the *responsible organization* and *service personnel* that is essential for preparation for the first use and safe use, maintenance or repair as well as *processing*, transport or storage for the *expected service life* of a *medical device*

Note 1 to entry: The *technical description* may be included in the *instructions for use*.

[SOURCE: ISO 20417:2021, 3.30, modified — deleted note 2.]

3.21
validation
confirmation, through the provision of *objective evidence*, that the requirements for a specific *intended use* or application have been fulfilled

Note 1 to entry: The *objective evidence* needed for a *validation* is the result of a test or other form of determination such as performing alternative calculations or reviewing documents.

Note 2 to entry: The term "*validated*" is used to designate the corresponding status.

Note 3 to entry: The use conditions for *validation* can be real or simulated.

[SOURCE: ISO 9000:2015, 3.8.13]

© ISO 2022 – All rights reserved

Documents
Document Sharing Hub
Full document details: BS EN ISO 18778:2022

4 General requirements

4.1 General

Clause 4 of IEC 60601-1:2005+AMD1:2012+AMD2:2020 applies with the following modifications:

4.2 *Essential performance*

NOTE There is guidance or rationale for this subclause contained in Clause A.3.

In addition to subclause 4.3 of IEC 60601-1:2005+AMD1:2012+AMD2:2020, additional *essential performance* requirements are given in the subclauses listed in Table 1.

Table 1 — Distributed *essential performance* requirements

Requirement	Subclause
Generating the apnoeic *patient alarm condition*	12.3.6 [a]
or	
Generation of a *technical alarm condition*	
Sensor fault	12.3.7
Internal electrical power source 5 min remaining	11.5.2 f) 1) iii)
[a] Subclause 21.4 b) 5) indicates the method of evaluating generating the apnoeic *alarm condition* as acceptance criteria following specific tests required by this document.	

4.3 *ME equipment* or *ME system* parts that contact the *patient*

In addition to subclause 4.6 of IEC 60601-1:2005+AMD1:2012+AMD2:2020, the following applies prior to the compliance check:

The parts or *accessories* that can come into contact with the *patient* shall be subject to the requirements for *applied parts* according to this subclause.

4.4 *Single fault condition* for *ME equipment*

In addition to subclause 4.7 of IEC 60601-1:2005+AMD1:2012+AMD2:2020, the following applies:

An *infant cardiorespiratory monitor* shall include at least one direct measurement of respiration and one indirect measurement of apnoeic activity such as pulse rate or oxygen saturation.

5 General requirements for testing of *ME equipment*

5.1 General

Clause 5 of IEC 60601-1:2005+AMD1:2012+AMD2:2020 applies.

5.2 *Infant cardiorespiratory monitor* testing errors

NOTE There is guidance or rationale for this subclause contained in Clause A.3.

a) For the purposes of this document, acceptance criteria for testing declared tolerances shall use the type A evaluation method (statistical uncertainty) *procedure* from IEC Guide 115:2021, 4.4.2.

b) Test equipment and methods shall be selected and controlled to ensure that the uncertainty (with coverage factor k = 2, for confidence of ~95 %) is no more than 30 % of the disclosed tolerance for the parameter being tested.

© ISO 2022 – All rights reserved
Documents Document Sharing Hub
Full document details: BS EN ISO 18778:2022

c) The *manufacturer* shall disclose the measurement uncertainty of each disclosed tolerance in the *technical description*.

Check conformance by inspection of the *instructions for use* and the *technical description*.

6 Classification of *ME equipment* and *ME systems*

6.1 General

Clause 6 of IEC 60601-1:2005+AMD1:2012+AMD2:2020 applies.

6.2 Additional requirements for classification of *ME equipment* and *ME systems*

An *infant cardiorespiratory monitor* shall be *transit-operable*.

7 *ME equipment* identification, *marking* and documents

7.1 General

Clause 7 of IEC 60601-1:2005+AMD1:2012+AMD2:2020 applies.

7.2 *Information to be supplied by the manufacturer*

a) The *information supplied by the manufacturer* of an *infant cardiorespiratory monitor* and its *accessories* shall conform with ISO 20417.

b) In applying ISO 20417:2021, the terms in this document and those in IEC 60601-1:2005+AMD1:2012+AMD2:2020 shall be used as follows.

   1) The term "*accompanying information*" shall assume the same meaning as *accompanying documents*.

   2) The term "*medical device*" shall assume the same meaning as *ME equipment*.

   3) The term "*user*" shall assume the same meaning as *operator*.

   4) The term "*patient*" shall include animals.

Check conformance by application of ISO 20417:2021.

7.3 Additional requirements for *accessories*

In addition to subclause 7.2.4 of IEC 60601-1:2005+AMD1:2012+AMD2:2020, the following applies:

Accessories supplied separately shall:

a) fulfil the requirements of 7.4; and

b) be *marked* with an indication of any limitations or adverse effects of the *accessory* on the *basic safety* or *essential performance* of the *infant cardiorespiratory monitor*, if applicable.

   1) If *marking* the *accessory* is not practicable, this information may be placed in the *instructions for use*.

NOTE The *manufacturer* of the *accessory* can be the *infant cardiorespiratory monitor manufacturer* or another entity ("third-party manufacturer", healthcare provider or durable medical equipment provider) and all these entities are expected to verify conformance with this requirement. Additional requirements are found in 18.1.

© ISO 2022 – All rights reserved

Documents
Document Sharing Hub
Full document details: BS EN ISO 18778:2022

Check conformance by inspection and inspection of the *risk management file* for any limitations or adverse effects of the *accessory*.

7.4 Additional requirements for *marking* on the outside of *ME equipment* or *ME equipment* parts

In addition to subclause 7.2 of IEC 60601-1:2005+AMD1:2012+AMD2:2020, the following applies:

a) For an *infant cardiorespiratory monitor* intended to be used in the magnetic resonance (MR) environment, the *infant cardiorespiratory monitor*, its parts and *accessories* shall have *clearly legible markings* conforming with:

   1) *symbol* 7.3.1-1 of IEC 62570:2014 (see Table D.1, *symbol* 1) if 'MR Safe';

   2) *symbol* 7.3.1-2 of IEC 62570:2014 (see Table D.1, *symbol* 2) if 'MR Safe'; or

   3) *symbol* 7.3.2 of IEC 62570:2014 (see Table D.1, *symbol* 3) if 'MR Conditional'.

b) For an *infant cardiorespiratory monitor* not intended to be used in the magnetic resonance (MR) environment, the *infant cardiorespiratory monitor*, its parts and *accessories* shall have *clearly legible markings* conforming with IEC 62570:2014 *symbol* 7.3.3 (see Table D.1, *symbol* 4) if 'MR Unsafe'.

Check conformance by inspection.

7.5 General *instructions for use*

In addition to subclause 7.9.2.1 of IEC 60601-1:2005+AMD1:2012+AMD2:2020, the following applies.

Add as third bullet following the first paragraph:

— the intended position of the *operator*;

Add after note 5:

a) Separate *instructions for use* shall be provided for:

   1) the *lay operator*; and

   2) the *healthcare professional operator*.

b) The *manufacturer* may choose in which *instructions for use* to place the information required by this document unless otherwise indicated in this document based on *risk management* and *usability* considerations.

c) The *healthcare professional operator instructions for use* shall include the information contained in the *lay operator instructions for use*.

Check conformance by inspection of the *instructions for use*, the *risk management file* and the *usability engineering file*.

7.6 Additional requirements for warnings and safety notices

In addition to subclause 7.9.2.2 of IEC 60601-1:2005+AMD1:2012+AMD2:2020, the following applies.

The *instructions for use* shall include:

a) advice of other *hazards* and *risks* associated with the *infant cardiorespiratory monitor*, including applicable examples.

   EXAMPLE WARNING: Do not use the monitor with another impedance monitor as interference can cause missed apnoea events.

© ISO 2022 – All rights reserved
Documents
Document Sharing Hub
Full document details: BS EN ISO 18778:2022

b) if the *infant cardiorespiratory monitor* only detects *central apnoea*, a warning to the effect that: Warning: Do not attempt to use this monitor to detect obstructive sleep apnoea. Obstructive sleep apnoea cannot be detected by this monitor.

c) if the *infant cardiorespiratory monitor* is not 'MR safe' or 'MR Conditional', a warning to the effect that: Warning: It is unsafe to bring this monitor into an area of magnetic resonance (MR) equipment.

Check conformance by inspection of the *instructions for use*.

7.7 Additional requirements for start-up *procedure*

NOTE 1 There is guidance or rationale for this subclause contained in Clause A.3.

In addition to subclause 7.9.2.8 of IEC 60601-1:2005+AMD1:2012+AMD2:2020, the following applies.

NOTE 2 For the purposes of this document, a start-up *procedure* is a pre-use functional test that is used for the initial setup for a *patient* to determine whether the *infant cardiorespiratory monitor* is ready for use.

a) The *instructions for use* for the *lay operator* shall disclose a method by which the following can be functionally tested to determine if operating correctly:

   1) the assembled sensors and related *accessories*; and

   2) the switchover to and operation from the *internal electrical power supply.*

   NOTE 3 Additional requirements are also found in 15.3.

b) The *instructions for use* for the *healthcare professional operator* shall disclose a test method:

   1) by which functions necessary for *normal use* can be tested to determine if they are operating correctly; and

   2) by which one can determine whether or not the sensors and related *accessories* are suitable for use.

c) These test methods, or portions thereof, may:

   1) be performed automatically by the *infant cardiorespiratory monitor;* or

   2) require *operator* action.

d) These test methods should be as automated as practicable.

Check conformance by inspection of the *instructions for use*.

7.8 Additional requirements for operating instructions

7.8.1 General

In addition to subclause 7.9.2.8 of IEC 60601-1:2005+AMD1:2012+AMD2:2020, the following applies.

7.8.2 *Lay operator* operating instructions

The *instructions for use* intended for the *lay operator* shall include:

a) a description of a means to determine the operation time of the *internal electrical power source*.

b) a description of how to connect and test the connection of a *distributed alarm system,* if provided.

© ISO 2022 – All rights reserved

Documents
Document Sharing Hub
Full document details: BS EN ISO 18778:2022

7.8.3 *Healthcare professional operator* operating instructions

The *instructions for use* intended for the *healthcare professional operator* shall include

a) a description of how at least the following *alarm conditions* can be functionally tested:

   1) *apnoea*; and

   2) sensor fault.

Check conformance by inspection of the *instructions for use*.

7.9 *Cleaning*, *disinfection*, and *sterilization*

Subclause 7.9.2.12 of IEC 60601-1:2005+AMD1:2012+AMD2:2020 applies with the following modification:

Add after normal use:

and *single fault condition*

Add to the bulleted list:

a) any pre-use *cleaning* and *disinfection* or *cleaning* and *sterilization procedures* for the *infant cardiorespiratory monitor* and any *accessories* including any specific *procedures* necessary before the *infant cardiorespiratory monitor* is transferred to another *patient*;

b) *procedures* for *cleaning*, *disinfection* or *sterilization* and the recommended frequencies;

c) any limitations on the number of *cleaning*, *disinfection* or *sterilization* cycles.

7.10 Additional requirements for maintenance

In addition to subclause 7.9.2.13 of IEC 60601-1:2005+AMD1:2012+AMD2:2020, the following applies.

The *instructions for use* shall disclose

a) a description of periodic safety inspections that should be performed by the *operator*.

b) the care and maintenance *procedures* for the *internal electrical power source*, including instructions for recharging and, if applicable, replacement.

Check conformance by inspection of the *instructions for use*.

7.11 Additional requirements for *accessories*, supplementary equipment, used material

In addition to subclause 7.9.2.14 of IEC 60601-1:2005+AMD1:2012+AMD2:2020, the following applies.

If applicable, the *instructions for use* shall disclose the positioning of sensors.

7.12 Additional requirements for the *technical description*

In addition to subclause 7.9.3 of IEC 60601-1:2005+AMD1:2012+AMD2:2020, the following applies.

a) The *technical description* shall include a description of a method for checking the proper functioning of the *alarm system* for each of the *alarm conditions* specified in this document, if not performed automatically during the start-up *procedure*.

b) The *technical description* shall disclose which checks are performed automatically.

Check conformance by inspection of the *technical description*.

© ISO 2022 – All rights reserved
Documents
Document Sharing Hub
Full document details: BS EN ISO 18778:2022

8 Protection against electrical *hazards* from *ME equipment*

Clause 8 of IEC 60601-1:2005+AMD1:2012+AMD2:2020 applies.

9 Protection against *mechanical hazards* of *ME equipment* and *ME systems*

9.1 General

Clause 9 of IEC 60601-1:2005+AMD1:2012+AMD2:2020 applies including the following additional subclauses.

9.2 Additional requirements for instability from unwanted lateral movement

In addition to subclause 9.4.3 of IEC 60601-1:2005+AMD1:2012+AMD2:2020, the following applies:

a) An *infant cardiorespiratory monitor* shall include a means by which the *infant cardiorespiratory monitor* can be secured without the use of a *tool* to prevent unwanted movement during transport while in use.

EXAMPLE 1 Means to restrain physically the *infant cardiorespiratory monitor* during transport in a personal vehicle or in a baby carriage.

b) The means shall secure the *infant cardiorespiratory monitor* to withstand accelerations or decelerations of 1,0 g longitudinal (forward, backward) and 1,0 g transverse (left, right) for at least 3 s in each orientation.

EXAMPLE 2 Attach the *infant cardiorespiratory monitor* to an armature at a 1 m radius from an axis of horizontal rotation. When rotating through a circle every 2 s at constant speed, the lateral (centripetal) acceleration is approximately 1,0 g[9].

Check conformance by functional testing.

9.3 Grips and other handling devices

Subclause 9.4.4 of IEC 60601-1:2005+AMD1:2012+AMD2:2020 applies with the following modifications.

Replace list item b) with the following:

b) An *infant cardiorespiratory monitor* shall require only one hand to be carried.

EXAMPLE 1 Small enough that it does not require a handle.

EXAMPLE 2 Equipped with a handle that does not require more than one hand.

EXAMPLE 3 Equipped with a carrying strap.

Check conformance by carrying with one hand.

10 Protection against unwanted and excessive radiation *hazards*

Clause 10 of IEC 60601-1:2005+AMD1:2012+AMD2:2020 applies.

11 Protection against excessive temperatures and other *hazards*

11.1 General

Clause 11 of IEC 60601-1:2005+AMD1:2012+AMD2:2020 applies including the following additions.

© ISO 2022 – All rights reserved

Documents
Document Sharing Hub
Full document details: BS EN ISO 18778:2022