



BSI Standards Publication

Microbiology of the food chain — Horizontal method for the determination of *Vibrio* spp.

Part 1: Detection of potentially enteropathogenic *Vibrio parahaemolyticus*, *Vibrio cholerae* and *Vibrio vulnificus*

National foreword

This British Standard is the UK implementation of EN ISO 21872-1:2017+A1:2023. It is identical to ISO 21872-1:2017, incorporating amendment 1:2023. It supersedes BS EN ISO 21872-1:2017, which is withdrawn.

The start and finish of text introduced or altered by amendment is indicated in the text by tags. Tags indicating changes to ISO text carry the number of the ISO amendment. For example, text altered by ISO amendment 1 is indicated by A1 A1.

The UK participation in its preparation was entrusted to Technical Committee AW/9, Microbiology.

A list of organizations represented on this committee can be obtained on request to its committee manager.

Contractual and legal considerations

This publication has been prepared in good faith, however no representation, warranty, assurance or undertaking (express or implied) is or will be made, and no responsibility or liability is or will be accepted by BSI in relation to the adequacy, accuracy, completeness or reasonableness of this publication. All and any such responsibility and liability is expressly disclaimed to the full extent permitted by the law.

This publication is provided as is, and is to be used at the recipient's own risk.

The recipient is advised to consider seeking professional guidance with respect to its use of this publication.

This publication is not intended to constitute a contract. Users are responsible for its correct application.

© The British Standards Institution 2023
Published by BSI Standards Limited 2023

ISBN 978 0 539 16203 5

ICS 07.100.30

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 30 September 2017.

Amendments/corrigenda issued since publication

Date	Text affected
31 March 2023	Implementation of ISO amendment 1:2023 with CEN endorsement A1:2023

English Version

Microbiology of the food chain - Horizontal method for the
determination of *Vibrio* spp. - Part 1: Detection of
potentially enteropathogenic *Vibrio parahaemolyticus*,
Vibrio cholerae and *Vibrio vulnificus* (ISO 21872-1:2017)

Microbiologie de la chaîne alimentaire - Méthode
horizontale pour la détermination des *Vibrio* spp. -
Partie 1: Recherche des espèces de *Vibrio*
parahaemolyticus, *Vibrio cholerae* et *Vibrio vulnificus*
potentiellement entéropathogènes (ISO 21872-1:2017)

Mikrobiologie von Lebensmitteln und Futtermitteln -
Horizontales Verfahren zum Nachweis von potentiell
enteropathogenen *Vibrio* spp. - Teil 1: Nachweis von
vibrio parahaemolyticus und *vibrio cholerae* (ISO
21872-1:2017)

This European Standard was approved by CEN on 14 May 2017.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

European foreword

This document (EN ISO 21872-1:2017) has been prepared by Technical Committee ISO/TC 34 "Food products" in collaboration with Technical Committee CEN/TC 275 "Food analysis - Horizontal methods" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by January 2018 and conflicting national standards shall be withdrawn at the latest by January 2018.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 21872-1:2017 has been approved by CEN as EN ISO 21872-1:2017 without any modification.

European foreword to amendment 1

This document (EN ISO 21872-1:2017/A1:2023) has been prepared by Technical Committee ISO/TC 34 "Food products" in collaboration with Technical Committee CEN/TC 463 "Microbiology of the food chain" the secretariat of which is held by AFNOR.

This Amendment to the European Standard EN ISO 21872-1:2017 shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by September 2023, and conflicting national standards shall be withdrawn at the latest by September 2023.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

Any feedback and questions on this document should be directed to the users' national standards body/national committee. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

Endorsement notice to amendment 1

The text of ISO 21872-1:2017/Amd 1:2023 has been approved by CEN as EN ISO 21872-1:2017/A1:2023 without any modification.

Contents

Page

Foreword	v
Introduction	vi
1 Scope	1
2 Normative references	1
3 Terms and definitions	2
4 Principle	2
4.1 General	2
4.2 Primary enrichment in a liquid selective medium	2
4.3 Secondary enrichment in a liquid selective medium	3
4.4 Isolation and identification	3
4.5 Confirmation	3
5 Culture media and reagents	3
5.1 Enrichment medium: alkaline saline peptone water (ASPW)	4
5.2 Solid selective isolation media	4
5.2.1 First medium: thiosulphate, citrate, bile and sucrose agar medium (TCBS)	4
5.2.2 Second medium	4
5.3 Saline nutrient agar (SNA)	4
5.4 Reagent for detection of oxidase	4
5.5 Biochemical tests	4
5.5.1 L-lysine decarboxylase saline medium (LDC)	4
5.5.2 Arginine dihydrolase saline medium (ADH)	4
5.5.3 Reagent for detection of β -galactosidase	5
5.5.4 Saline medium for detection of indole	5
5.5.5 Saline peptone waters	5
5.5.6 Sodium chloride solution	5
5.6 PCR	5
5.6.1 Tris acetate EDTA buffer (TAE) (or a buffer allowing similar performance for the purpose)	5
5.6.2 Mastermix	5
5.6.3 Primers and probes	5
5.6.4 Positive control material	5
5.6.5 Negative extraction control	6
6 Equipment and consumables	6
7 Sampling	6
8 Preparation of the test sample	6
9 Procedure (See Figure A.1)	7
9.1 Test portion and initial suspension	7
9.2 Primary selective enrichment	7
9.3 Secondary selective enrichment	7
9.4 Isolation and identification	8
9.5 Confirmation	8
9.5.1 General	8
9.5.2 Selection of colonies for confirmation and preparation of pure cultures	9
9.5.3 Tests for presumptive identification	9
9.5.4 Biochemical confirmation	10
9.5.5 PCR confirmation	12
9.5.6 DNA extraction	12
9.5.7 Conventional PCR	12
9.5.8 Real-time PCR	13
10 Expression of results	13

11	Performance characteristics of the method	13
11.1	Interlaboratory study	13
11.2	Sensitivity	14
11.3	Specificity	14
11.4	LOD ₅₀	14
12	Test report	14
Annex A (normative)	Diagram of procedure	15
Annex B (normative)	Composition and preparation of the culture media and reagents	17
Annex C (informative)	Conventional PCR for the detection of <i>Vibrio parahaemolyticus</i>, thermostable direct haemolysin (<i>tdh</i>) and thermostable direct related haemolysin (<i>trh</i>) genes, <i>Vibrio cholerae</i> and <i>Vibrio vulnificus</i>	23
Annex D (informative)	Real-time PCR for the detection of <i>Vibrio parahaemolyticus</i>, thermostable direct haemolysin gene (<i>tdh</i>) and <i>Vibrio vulnificus</i>	27
Annex E (informative)	Results of an interlaboratory study	29
Bibliography		32

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

This document was prepared by the European Committee for Standardization (CEN) Technical Committee CEN/TC 275, *Food analysis — Horizontal methods*, in collaboration with ISO Technical Committee TC 34, *Food products*, Subcommittee SC 9, *Microbiology*, in accordance with the agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This first edition cancels and replaces ISO/TS 21872-1:2007, which has been technically revised. It also incorporates ISO/TS 21872-1:2007/Cor.1:2008.

The main changes are as follows:

- introduction of optional molecular identification methods for major food borne *Vibrio* spp. (*V. parahaemolyticus*, including potentially enteropathogenic strains, *V. vulnificus* and *V. cholerae*);
- performance characteristics of the method have been added in [Annex E](#).

A list of all parts in the ISO 21872 series can be found on the ISO website.

Introduction

Because of the large variety of food and feed products, the horizontal method described in this document may not be appropriate in every detail for certain products. In this case, different methods, which are specific to these products may be used if absolutely necessary for justified technical reasons. Nevertheless, every attempt will be made to apply this horizontal method as far as possible.

The main changes, listed in the foreword, introduced in this document compared to ISO/TS 21872-1:2007 are considered as major (see ISO 17468).

When this document is next reviewed, account will be taken of all information then available regarding the extent to which this horizontal method has been followed and the reasons for deviations from this method in the case of particular products.

The harmonization of test methods cannot be immediate and, for certain groups of products, International Standards and/or national standards may already exist that do not comply with this horizontal method. It is hoped that when such standards are reviewed they will be changed to comply with this document so that eventually the only remaining departures from this horizontal method will be those necessary for well-established technical reasons.

Microbiology of the food chain — Horizontal method for the determination of *Vibrio* spp. —

Part 1:

Detection of potentially enteropathogenic *Vibrio parahaemolyticus*, *Vibrio cholerae* and *Vibrio vulnificus*

WARNING — In order to safeguard the health of laboratory personnel, it is essential that tests for detection of *Vibrio* spp., and particularly toxigenic *Vibrio cholerae*, be conducted only in laboratories equipped for this purpose and under the supervision of an experienced microbiologist, and that great care is exercised in the disposal of contaminated material.

1 Scope

This document specifies a horizontal method for the detection of enteropathogenic *Vibrio* spp., which causes human illness in or via the intestinal tract. The species detectable by the methods specified include *Vibrio parahaemolyticus*, *Vibrio cholerae* and *Vibrio vulnificus*.

It is applicable to the following:

- products intended for human consumption and the feeding of animals;
- environmental samples in the area of food production and food handling.

NOTE 1 This method may not be appropriate in every detail for certain products (see Introduction).

NOTE 2 The World Health Organization (WHO) has identified that *V. parahaemolyticus*, *V. cholerae* and *V. vulnificus* are the major food-borne *Vibrio* spp. However, the method in this document can also be appropriate for the identification of other *Vibrio* spp. causing illness in humans.^[1]

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 6887-1:2017, *Microbiology of the food chain — Preparation of test samples, initial suspension and decimal dilutions for microbiological examination — Part 1: General rules for the preparation of the initial suspension and decimal dilutions*

ISO 6887-3, *Microbiology of the food chain — Preparation of test samples, initial suspension and decimal dilutions for microbiological examination — Part 3: Specific rules for the preparation of fish and fishery products*

ISO 7218, *Microbiology of food and animal feeding stuffs — General requirements and guidance for microbiological examinations*

ISO 11133, *Microbiology of food, animal feed and water — Preparation, production, storage and performance testing of culture media*

ISO 22118, *Microbiology of food and animal feeding stuffs — Polymerase chain reaction (PCR) for the detection and quantification of food-borne pathogens — Performance characteristics*

ISO 22119, *Microbiology of food and animal feeding stuffs — Real-time polymerase chain reaction (PCR) for the detection of food-borne pathogens — General requirements and definitions*

ISO 22174, *Microbiology of food and animal feeding stuffs — Polymerase chain reaction (PCR) for the detection of food-borne pathogens — General requirements and definitions*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <http://www.iso.org/obp>

3.1

potentially enteropathogenic *Vibrio* spp.

microorganism which forms typical colonies on solid selective media and which possesses the described biochemical or molecular characteristics when the test is performed in accordance with this document

Note 1 to entry: This document describes specific procedures for *V. parahaemolyticus*, *V. cholerae* and *V. vulnificus*.

3.2

detection of potentially enteropathogenic *Vibrio* spp.

determination of the presence or absence of potentially enteropathogenic *Vibrio* spp. (3.1) (*V. parahaemolyticus*, *V. cholerae* and *V. vulnificus*) in a determined quantity of product, when the test is performed in accordance with this document

4 Principle

4.1 General

The detection of potentially enteropathogenic *Vibrio* spp. (*V. parahaemolyticus*, *V. cholerae* and *V. vulnificus*) requires four successive phases, as shown in the procedure diagram in [Annex A](#).

Recovery of certain *Vibrio* spp. from foodstuffs may be improved by the use of different incubation temperatures depending upon the target species or state of the food matrix. For example, recovery of *V. parahaemolyticus* and *V. cholerae* in fresh products is enhanced by enrichment at 41,5 °C whereas for *V. vulnificus*, and for *V. parahaemolyticus* and *V. cholerae* in deep frozen (<-18 °C),^[2] dried or salted products, recovery is enhanced by enrichment at 37 °C. If detection of *V. parahaemolyticus*, *V. cholerae* and *V. vulnificus* is required, all specified incubation temperatures should be used. If detection of *V. parahaemolyticus*, *V. cholerae* and *V. vulnificus* together is not required, the specific procedure(s) may be selected according to the species being sought. Such a selection should be clearly specified in the test report.

NOTE *V. parahaemolyticus*, *V. cholerae* and *V. vulnificus* may be present in small numbers and are often accompanied by a much larger number of other microorganisms belonging to the *Vibrionaceae* family or to other families.

4.2 Primary enrichment in a liquid selective medium

Inoculation of the test portion in the primary enrichment medium alkaline saline peptone water (ASPW) (5.1) at ambient temperature, followed by incubation at 41,5 °C for 6 h and/or 37 °C for 6 h.

The incubation conditions are determined by the target species and food product state.

For detection of all target species in deep frozen, dried or salted products, primary enrichment should be at 37 °C.

For detection of *V. vulnificus* in all products, primary enrichment should be at 37 °C.

For detection of *V. parahaemolyticus* and/or *V. cholerae* only, in fresh products, primary enrichment should be at 41,5 °C.

4.3 Secondary enrichment in a liquid selective medium

Inoculation of the second enrichment medium (ASPW) with the cultures obtained in [4.2](#).

Incubation of inoculated enrichment medium at 41,5 °C for 18 h and/or 37 °C for 18 h.

For detection of *V. vulnificus* in all products, secondary enrichment should be at 37 °C.

For detection of *V. parahaemolyticus* and/or *V. cholerae* only, in all products, secondary enrichment should be at 41,5 °C.

4.4 Isolation and identification

From the cultures obtained in [4.2](#) and in [4.3](#), inoculation of two solid selective media:

- thiosulfate citrate bile and sucrose agar (TCBS) medium ([5.2.1](#));
- another appropriate solid selective medium (left to the choice of the laboratory), such as chromogenic agar, complementary to the TCBS medium ([5.2.2](#)).

Incubation of the TCBS medium at 37 °C, then examination after 24 h. Incubation of the second selective medium according to the manufacturer's recommendations.

4.5 Confirmation

Presumptive colonies of *V. parahaemolyticus*, *V. cholerae* and *V. vulnificus* isolated in [4.4](#) are subcultured and confirmed by means of appropriate biochemical and/or polymerase chain reaction (PCR) tests.

Biochemical and/or PCR confirmation of isolates may be used for species identification, however, reliable detection of enteropathogenic *V. parahaemolyticus* as determined by presence of the direct thermostable haemolysin (*tdh*) and/or direct related haemolysin (*trh*) genes can only be carried out using PCR tests.

It has been demonstrated that by screening ASPW broths using conventional PCR, the absence of amplification of *Vp-toxR* can indicate no detection of *V. parahaemolyticus*.^[3] To reduce the amount of downstream testing and, if shown to be reliable by the user laboratory, screening of incubated enrichment broths may be used. This approach is only recommended after secondary enrichment and does not apply to molecular targets for other *Vibrio* spp.

NOTE 1 Validation data generated in the preparation of this document demonstrated that PCR based identification can be achieved by conventional PCR for *V. parahaemolyticus*, *V. vulnificus* and *V. cholerae* or real-time PCR for *V. parahaemolyticus* and *V. vulnificus*. The PCR methods used in the development of this document are given in [Annexes C](#) and [D](#).

NOTE 2 Validation data for screening of secondary enrichment broths for amplification of *Vp-toxR* were not generated in the preparation of this document.

5 Culture media and reagents

For general laboratory practice, refer to ISO 7218.

For clarity of the text, details of the composition of culture media and reagents and their preparation are described in [Annex B](#).

A1 For performance testing of culture media and reagents, follow the procedures in accordance with ISO 11133 and Clause B.12. **A1**

NOTE Primers, probes and PCR running conditions used in the development of this document are given in [Annexes C](#) and [D](#).

5.1 Enrichment medium: alkaline saline peptone water (ASPW)

As specified in [B.1](#).

5.2 Solid selective isolation media

5.2.1 First medium: thiosulphate, citrate, bile and sucrose agar medium (TCBS)

As specified in [B.2](#). **A1** Text deleted. **A1**

5.2.2 Second medium

The selection of the second medium is left to the choice of the test laboratory. Preparation of the medium should be strictly according to the manufacturer's instructions.

5.3 Saline nutrient agar (SNA)

As specified in [B.3](#).

5.4 Reagent for detection of oxidase

As specified in [B.4](#).

5.5 Biochemical tests

5.5.1 L-lysine decarboxylase saline medium (LDC)

As specified in [B.5](#).

5.5.2 Arginine dihydrolase saline medium (ADH)

As specified in [B.6](#).

5.5.3 Reagent for detection of β -galactosidase

As specified in [B.7](#).

5.5.4 Saline medium for detection of indole

As specified in [B.8](#).

5.5.5 Saline peptone waters

As specified in [B.9](#).

5.5.6 Sodium chloride solution

As specified in [B.10](#).

5.6 PCR

5.6.1 Tris acetate EDTA buffer (TAE) (or a buffer allowing similar performance for the purpose)

As specified in [B.11](#).

5.6.2 Mastermix

Reagents shall be added in quantities as specified by the manufacturer's instructions. See [Annexes C](#) and [D](#) for example details of master mixes used in the development of this document.

5.6.3 Primers and probes

Primer (and hydrolysis probe) sequences if required shall be published in a peer-reviewed journal and be verified for use against a broad range of target *Vibrio* spp. and non-target strains. With advances in whole genome sequencing of bacterial strains, more appropriate species-specific markers may be identified in the future.

For *V. parahaemolyticus* the target region should be *toxR*.

For determination of pathogenic strains of *V. parahaemolyticus* genes encoding the thermostable direct (TDH) and the thermostable direct related (TRH) haemolysins should be targeted.

For *V. cholerae* the target region for conventional PCR should be the 16S-23S rRNA intergenic spacer region *prVC*.

For *V. vulnificus* the target region should be the *V. vulnificus* haemolysin region (*vvha*).

Target regions other than those specified above for the identification of *V. parahaemolyticus*, *V. vulnificus* and *V. cholerae* can be used if they have been shown to demonstrate equivalent performance to those used in the development of this document and described in [Annexes C](#) or [D](#), are published in a peer-reviewed journal and are verified against a broad range of target *Vibrio* spp. and non-target strains.

See [Annex C](#) for example details of primers used for conventional PCR and [Annex D](#) for primers and hydrolysis probes for real-time PCR used in the development of this document.

5.6.4 Positive control material

Separate control material shall be used for each target *Vibrio* spp. See [Annexes C](#) and [D](#) for example details of control strains used in the development of this document.

5.6.5 Negative extraction control

Nuclease free water or sterile NaCl 0,85 % extracted according to [9.5.6](#).

6 Equipment and consumables

Disposable equipment is acceptable in the same way as reusable glassware, if the specifications are similar.

Ordinary microbiology laboratory equipment as specified in ISO 7218, and in particular the following.

6.1 Refrigerator, adjustable to $5\text{ }^{\circ}\text{C} \pm 3,0\text{ }^{\circ}\text{C}$.

6.2 Incubator, adjustable to $37\text{ }^{\circ}\text{C} \pm 1,0\text{ }^{\circ}\text{C}$.

6.3 Incubator, adjustable to $41,5\text{ }^{\circ}\text{C} \pm 1,0\text{ }^{\circ}\text{C}$.

6.4 Freezer, adjustable to $<-15\text{ }^{\circ}\text{C}$.

6.5 Micro-centrifuge tubes, with a capacity of 1,5 ml and 2,0 ml.

6.6 Micro-centrifuge, for reaction tubes with a capacity of 1,5 ml and 2,0 ml and capable of running at 10 000*g*.

6.7 Heating block capable of operating at $95\text{ }^{\circ}\text{C} \pm 2,0\text{ }^{\circ}\text{C}$ or equivalent.

6.8 Vortex.

6.9 Graduated pipettes and pipette filter tips, for volumes between 1 μl and 1 000 μl .

6.10 Associated consumables for conventional or real-time PCR, e.g. optical plates and caps, optical plate holder, suitable for use with the selected PCR machine.

6.11 Conventional or real-time PCR machine, gel electrophoresis and UV visualization equipment as appropriate.

7 Sampling

It is important that the laboratory receives a truly representative sample which has not been damaged or modified during transport and storage.

Sampling does not form part of the method specified in this document. See the International Standard specific to the relevant product. If a specific document does not exist, it is recommended that the relevant parties reach agreement on this subject.

8 Preparation of the test sample

Prepare the test sample in accordance with ISO 6887-1 and ISO 6887-3 and/or the document concerning the product to be examined. If a specific document does not exist, it is recommended that the relevant parties reach agreement on this subject.