

Breathing system filters for anaesthetic and respiratory use —

Part 2: Non-filtration aspects

ICS 11.040.10

National foreword

This British Standard is the UK implementation of EN ISO 23328-2:2008. It is identical to ISO 23328-2:2002. It supersedes BS EN 13328-2:2002 which is withdrawn.

The UK participation in its preparation was entrusted by Technical Committee CH/121, Anaesthetic and respiratory equipment, to Subcommittee CH/121/5, Lung ventilators, tracheal tubes and related equipment.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 29 August 2008

© BSI 2008

ISBN 978 0 580 59142 6

Amendments/corrigenda issued since publication

Date	Comments

English Version

Breathing system filters for anaesthetic and respiratory use -
Part 2: Non-filtration aspects (ISO 23328-2:2002)

Filtres pour matériel d'anesthésie et de réanimation
respiratoire - Partie 2: Aspects autres que la filtration (ISO
23328-2:2002)

Filter für Atemsysteme zur Anwendung bei Anästhesie und
Beatmung - Teil 2: Aspekte, die nicht die Filtration betreffen
(ISO 23328-2:2002)

This European Standard was approved by CEN on 24 February 2008.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: rue de Stassart, 36 B-1050 Brussels

Foreword

The text of ISO 23328-2:2002 has been prepared by Technical Committee ISO/TC 121 “Anaesthetic and respiratory equipment” of the International Organization for Standardization (ISO) and has been taken over as EN ISO 23328-2:2008 by Technical Committee CEN/TC 215 “Respiratory and anaesthetic equipment” the secretariat of which is held by BSI.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by September 2008, and conflicting national standards shall be withdrawn at the latest by September 2008.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN 13328-2:2002.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EC Directive(s).

For relationship with EC Directive(s), see informative Annex ZA, which is an integral part of this document.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and the United Kingdom.

Endorsement notice

The text of ISO 23328-2:2002 has been approved by CEN as a EN ISO 23328-2:2008 without any modification.

Contents

Page

Foreword	iv
Introduction.....	v
1 Scope	1
2 Normative references.....	1
3 Terms and definitions	1
4 BSF port connectors	2
4.1 BSF breathing system and patient connection ports.....	2
4.2 Accessory ports	2
5 Test methods	2
5.1 Ambient conditions of test	2
5.2 Measurement of pressure drop	2
5.3 Test for gas leakage.....	3
6 Packaging of sterile BSF	3
7 Marking.....	3
7.1 Use of symbols.....	3
7.2 Marking of BSF	3
7.3 Marking of package.....	3
7.4 BSF intended for single use.....	3
8 Information to be provided by the manufacturer	4
Annex ZA (informative) Relationship between this European Standard and the Essential Requirements of EU Directive 93/42/EEC Medical devices.....	6
Bibliography.....	5