BS EN ISO 8253-3:2022

BSI Standards Publication

Acoustics — Audiometric test methods

Part 3: Speech audiometry

Documents
Document Sharing Hub

Full document details: BS EN ISO 8253-3:2022

National foreword

This British Standard is the UK implementation of EN ISO 8253-3:2022. It is identical to ISO 8253-3:2022. It supersedes BS EN ISO 8253-3:2012, which is withdrawn.

The UK participation in its preparation was entrusted to Technical Committee EH/1/1, Hearing.

A list of organizations represented on this committee can be obtained on request to its committee manager.

Contractual and legal considerations

This publication has been prepared in good faith, however no representation, warranty, assurance or undertaking (express or implied) is or will be made, and no responsibility or liability is or will be accepted by BSI in relation to the adequacy, accuracy, completeness or reasonableness of this publication. All and any such responsibility and liability is expressly disclaimed to the full extent permitted by the law.

This publication is provided as is, and is to be used at the recipient's own risk.

The recipient is advised to consider seeking professional guidance with respect to its use of this publication.

This publication is not intended to constitute a contract. Users are responsible for its correct application.

© The British Standards Institution 2022
Published by BSI Standards Limited 2022

ISBN 978 0 539 06662 3

ICS 13.140

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 31 March 2022.

Amendments/corrigenda issued since publication

Date	Text affected

Documents
Document Sharing Hub

Full document details: BS EN ISO 8253-3:2022

EUROPEAN STANDARD

NORME EUROPÉENNE

EUROPÄISCHE NORM

EN ISO 8253-3

February 2022

ICS 13.140

Supersedes EN ISO 8253-3:2012

English Version

Acoustics - Audiometric test methods - Part 3: Speech audiometry (ISO 8253-3:2022)

Acoustique - Méthodes d'essais audiométriques - Partie 3: Audiométrie vocale (ISO 8253-3:2022)

Akustik - Audiometrische Prüfverfahren - Teil 3: Sprachaudiometrie (ISO 8253-3:2022)

This European Standard was approved by CEN on 18 January 2022.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.

EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

© 2022 CEN All rights of exploitation in any form and by any means reserved worldwide for CEN national Members.

Ref. No. EN ISO 8253-3:2022 E

Documents
Document Sharing Hub

Full document details: BS EN ISO 8253-3:2022

Contents

Page

European foreword **3**

Documents
Document Sharing Hub

Full document details: BS EN ISO 8253-3:2022

European foreword

This document (EN ISO 8253-3:2022) has been prepared by Technical Committee ISO/TC 43 "Acoustics" in collaboration with Technical Committee CEN/TC 211 “Acoustics” the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by August 2022, and conflicting national standards shall be withdrawn at the latest by August 2022.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 8253-3:2012.

Any feedback and questions on this document should be directed to the users’ national standards body/national committee. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 8253-3:2022 has been approved by CEN as EN ISO 8253-3:2022 without any modification.

AVA Documents
Document Sharing Hub

Full document details: BS EN ISO 8253-3:2022

This page deliberately left blank

Documents
Document Sharing Hub

Full document details: BS EN ISO 8253-3:2022

INTERNATIONAL STANDARD

ISO 8253-3

Third edition
2022-01

Acoustics — Audiometric test methods —

Part 3: Speech audiometry

Acoustique — Méthodes d'essais audiométriques —

Partie 3: Audiométrie vocale

Reference number
ISO 8253-3:2022(E)

Documents
Document Sharing Hub
© ISO 2022

Full document details: BS EN ISO 8253-3:2022

COPYRIGHT PROTECTED DOCUMENT

© ISO 2022

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

© ISO 2022 – All rights reserved

Documents
Document Sharing Hub

Full document details: BS EN ISO 8253-3:2022

Contents

Page

Foreword ... **v**
Introduction ... **vi**
1 Scope ... **1**
2 Normative references ... **1**
3 Terms and definitions ... **2**
4 Requirements for recording of speech material ... **5**
4.1 General requirements ... 5
4.2 Calibration signal ... 5
4.3 Reference recording ... 5
4.4 Recording environment ... 6
4.5 Recording setup ... 6
4.6 Signals and levels ... 6
4.6.1 Calibration signals ... 6
4.6.2 Speech signals ... 7
4.6.3 Announcing signals ... 7
4.6.4 Competing sound signals ... 7
4.7 Phonemic equivalence across test lists ... 7
4.8 Perceptual equivalence across test lists ... 7
5 Validation of speech material recordings ... **8**
5.1 General ... 8
5.2 Determination of reference speech recognition curve ... 8
5.3 Determination of perceptual equivalence of test lists ... 9
5.4 Determination of test-retest reliability ... 9
5.5 Documentation ... 9
6 Requirements for speech audiometry ... **10**
6.1 Audiometric equipment ... 10
6.2 Ambient sound pressure levels in test room for speech audiometry ... 10
6.3 Sound field speech audiometry ... 10
7 Preparation and instruction of test subject ... **11**
7.1 General ... 11
7.2 Preparation of test subject ... 11
7.3 Instruction of test subject ... 11
8 Subject's response mode ... **11**
9 Determination of speech detection threshold level ... **12**
10 Determination of speech recognition threshold level ... **13**
11 Determination of speech recognition scores ... **14**
12 Contralateral masking ... **14**
13 Speech audiometry with competing sound ... **15**
13.1 Type of competing sound ... 15
13.2 Presentation of competing sound ... 15
13.3 Speech and competing sound levels ... 15
13.4 Test procedure ... 16
13.4.1 General ... 16
13.4.2 Speech recognition score at fixed SNR ... 16
13.4.3 Speech recognition threshold level with a competing sound ... 16
14 Format of speech audiogram ... **16**
15 Measurement uncertainty ... **17**

Documents
Document Sharing Hub

© ISO 2022 – All rights reserved

Full document details: BS EN ISO 8253-3:2022

16 Maintenance and calibration of equipment ... **17**
16.1 General ... 17
16.2 Stage A: Routine checking and listening tests ... 17
16.3 Stage B: Periodic electroacoustic tests ... 18
16.4 Stage C: Basic calibration tests ... 18
16.5 Intervals between tests ... 18
Annex A (informative) **Example of speech materials** ... **20**
Annex B (informative) **Examples of competing sound** ... **21**
Annex C (informative) **Typical results** ... **22**
Annex D (informative) **Optimization of perceptual equivalence of test lists** ... **26**
Annex E (informative) **Measurement uncertainty** ... **27**
Annex F (informative) **Determination of minimum number of subjects** ... **32**
Annex G (informative) **Determination of test-retest reliability** ... **34**
Bibliography ... **39**

© ISO 2022 – All rights reserved
Documents
Document Sharing Hub
Full document details: BS EN ISO 8253-3:2022

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 43, *Acoustics*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 211, *Acoustics,* in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This third edition cancels and replaces the second edition (ISO 8253-3:2012), which has been technically revised. The main changes compared to the previous edition are as follows:

— The technical requirements for recording of speech material were adapted to recent equipment and technology.

— The determination of reference speech recognition curves was revised. An annex that gives advices on how to determine the minimum number of subjects was introduced.

— The determination of speech recognition threshold levels is described in a more general manner.

— Symbols for the graphical representation of speech audiometry results were introduced.

A list of all the parts in the ISO 8253 series can be found on the ISO website.

© ISO 2022 – All rights reserved

Documents
Document Sharing Hub
Full document details: BS EN ISO 8253-3:2022

Introduction

Speech audiometry is used for the assessment of hearing in connection with diagnostic evaluation and audiological rehabilitation.

The results of speech audiometry depend on the speech material and test method used. This document sets conditions for speech materials in order to ensure minimum requirements of precision and comparability between different tests using different speech materials including materials in different languages. It also specifies procedures to be used when testing speech recognition.

© ISO 2022 – All rights reserved
Documents
Document Sharing Hub
Full document details: BS EN ISO 8253-3:2022

Acoustics — Audiometric test methods —

Part 3:
Speech audiometry

1 Scope

This document specifies basic methods for speech recognition tests for audiological applications.

NOTE Examples of speech materials are given in Annex A.

In order to ensure minimum requirements of precision and comparability between different test procedures including speech recognition tests in different languages, this document specifies requirements for the composition, validation and evaluation of speech test materials, and the realization of speech recognition tests. This document does not specify the contents of the speech material because of the variety of languages.

Furthermore, this document also specifies the determination of reference values and requirements for the realization and manner of presentation. In addition, there are features of speech tests described which are important to be specified, but which are not understood as a requirement.

This document specifies procedures and requirements for speech audiometry with the recorded test material being presented by an audiometer through a transducer, e.g., an earphone, bone vibrator, or loudspeaker arrangement for sound field audiometry. Methods for using noise either for masking the non-test ear or as a competing sound are described.

Some test subjects, for example children, can require modified test procedures not specified in this document.

Specialized tests, such as those used for evaluating directional hearing and dichotic hearing, are outside the scope of this document.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 8253-1, *Acoustics — Audiometric test methods — Part 1: Pure-tone air and bone conduction audiometry*

ISO 8253-2, *Acoustics — Audiometric test methods — Part 2: Sound field audiometry with pure-tone and narrow-band test signals*

ISO/IEC Guide 98-3, *Uncertainty of measurement — Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)*

IEC 60645-1:2017, *Electroacoustics — Audiometric equipment — Part 1: Equipment for pure-tone and speech audiometry*

IEC 61672-1, *Electroacoustics — Sound level meters — Part 1: Specifications*

© ISO 2022 – All rights reserved

Documents
Document Sharing Hub
Full document details: BS EN ISO 8253-3:2022

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 8253-1 and ISO 8253-2 and the following apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

— ISO Online browsing platform: available at https://www.iso.org/obp

— IEC Electropedia: available at https://www.electropedia.org/

3.1
speech signal
acoustic signal which carries information in a given language

Note 1 to entry: A speech signal can be a voice signal or an acoustic signal simulating a voice signal.

3.2
test item
particular monosyllabic or polysyllabic word, or *logatom* (3.12), or spondee, or sentence, or time-limited segment of connected speech, used in accordance with defined rules of presentation and scoring in a speech audiometric procedure

Note 1 to entry: Scoring may be based on a complete test item or parts thereof being correctly recognized.

3.3
test list
number of selected *test items* (3.2), presented and scored as a single unit

3.4
set of test items
selected number of *test items* (3.2) from a test list

3.5
speech material
entire *set of test items* (3.4) which is used for speech recognition tests

Note 1 to entry: Usually the speech material is subdivided into several test lists.

3.6
open-set test
test in which the number of alternative responses to each *test item* (3.2) is unlimited

3.7
closed-set test
test in which the number of alternative responses to each *test item* (3.2) is limited

3.8
phoneme
the smallest phonetic unit that distinguishes one word from another in a particular language

3.9
phoneme class
subdivision of phonemes that show characteristic similarities in vocal production mode as well as in acoustical signal properties

3.10
phoneme distribution
relative distribution of the various phonemes within a given *speech material* (3.5)

© ISO 2022 – All rights reserved
Documents
Document Sharing Hub
Full document details: BS EN ISO 8253-3:2022

3.11
syllable
segment of speech consisting of an (optional) onset, a nucleus, and an (optional) coda, where the nucleus most often consists of a vowel sound and the onset and coda most often consist of one or more consonant sounds

3.12
logatom
syllabic unit that has no verbal meaning to the listener

Note 1 to entry: A logatom is sometimes called a "nonsense syllable".

3.13
carrier phrase
sentence or phrase in which a *test item* (3.2) is embedded such that the correct recognition of the *test item* (3.2) is not dependent upon the context or meaning of the sentence or phrase

3.14
connected speech
continuous speech with natural intonation, consisting of consecutive sentences with logical connections

3.15
reference recording
master recording
well-defined recording that represents the *speech material* (3.5) and that is used for validation and application of the *speech material* (3.5)

3.16
speech level
sound pressure level of a specified *speech material* (3.5) as measured in an appropriate acoustic coupler, ear simulator or in a sound field, with a specified frequency and time weighting

Note 1 to entry: The frequency and the time weighting should be according to IEC 61672-1.

3.17
speech detection threshold level
for a given test subject, for a specified *speech material* (3.5) and a specified manner of signal presentation, the speech level at which it is detected as sound (not necessarily understood) in 50 % of the trials

Note 1 to entry: Is synonymous with the expression "speech awareness threshold".

3.18
speech recognition score
for a given test subject, for a specified *speech material* (3.5), a specified manner of signal presentation and at a specified speech level, the percentage of correctly recognized *test items* (3.2) or scorable items if the scoring method is not based on whole *test items* (3.2)

3.19
speech audiogram
graphical representation of the results of speech audiometry, where the speech recognition score is plotted as percentage along the ordinate and speech level, in dB, or hearing level for speech, in dB, or speech-to-noise level difference, in dB, along the abscissa

3.20
maximum speech recognition score
for a given test subject, for a specified *speech material* (3.5) and a specified manner of signal presentation, the maximum speech recognition score obtained regardless of the speech level

© ISO 2022 – All rights reserved

Documents
Document Sharing Hub
Full document details: BS EN ISO 8253-3:2022

3.21
speech recognition threshold level
for a given test subject, for a specified *speech material* (3.5) and a specified manner of signal presentation, the lowest speech level at which the speech recognition score is equal to 50 % or another specified value

Note 1 to entry: Speech recognition threshold has been called "speech reception threshold".

3.22
reference speech recognition threshold level
for a specified *speech material* (3.5) and a specified manner of signal presentation, the median value of the speech recognition threshold levels of a sufficiently large number of otologically normal persons, of both sexes, between 18 years and 25 years inclusive for whom the *speech material* (3.5) is appropriate

3.23
hearing level for speech
for a specified *speech material* (3.5) and a specified manner of signal presentation, the speech level minus the appropriate reference speech recognition threshold level

3.24
optimum speech level
for a given test subject, for a specified speech material and a specified manner of signal presentation, the speech level or range of speech levels at which maximum speech recognition score is obtained

3.25
half-optimum speech level
for a given test subject, for a specified speech material and a specified manner of signal presentation, the speech level at which half of the maximum speech recognition score is obtained

3.26
most comfortable level for speech
for a given test subject, a specific *speech material* (3.5) and a specified manner of presentation, the speech level at which the loudness of the *speech signal* (3.1) is judged by the test subject to be most comfortable

3.27
speech recognition curve
for a given test subject, for a specified *speech material* (3.5) and a specified manner of signal presentation, a curve that describes the speech recognition score as a function of speech level

3.28
reference speech recognition curve
for a specified *speech material* (3.5) and a specified manner of presentation, a curve that describes the speech recognition score as a function of speech level for a sufficiently large number of otologically normal persons of both sexes, aged between 18 years and 25 years inclusive and for whom the *speech material* (3.5) is appropriate

3.29
competing sound
additional sound that is presented during speech recognition tests

3.30
competing sound level
sound pressure level of a competing sound as measured in an appropriate acoustic coupler, ear simulator or in a sound field, with a specified frequency and time weighting

Note 1 to entry: The frequency and the time weighting should be according to IEC 61672-1.

© ISO 2022 – All rights reserved
Documents
Document Sharing Hub
Full document details: BS EN ISO 8253-3:2022

3.31
SNR
speech-to-noise ratio
speech-to-noise level difference
difference between the speech level and the competing sound level

3.32
effective masking level for speech
level of a specified masking sound equivalent to that hearing level for speech to which the speech recognition threshold level for a specified *speech signal* (3.1) would be raised by the presence of that masking sound

4 Requirements for recording of speech material

4.1 General requirements

Each recording shall contain the following signals:

a) the speech material;

b) a signal for the calibration of the equipment for speech audiometry.

4.2 Calibration signal

The calibration signal shall meet the following requirements:

a) A minimum duration of 60 s.

b) The calibration signal for sound-field speech audiometry shall be a weighted random noise, e.g., as specified in IEC 60645-1:2017, 6.5.3, a band of noise centred at 1 kHz and having a bandwidth of one-third octave, or a frequency-modulated tone at 1 kHz having a bandwidth of at least one-third octave — the modulating signal shall be either sinusoidal or triangular with a repetition rate in the range from 4 Hz to 20 Hz.

c) The calibration signal for earphones or bone vibrators shall be as specified in b). Alternatively, a pure-tone signal may be used.

4.3 Reference recording

The voice characteristics of the speaker influence the test results. For general-use tests, a suitable speaker, either male or female, shall have a normal and clear articulation and should speak with a generally well understood dialect. The speaker shall be instructed to maintain clarity, natural pace, and vocal effort and avoid emphasis on keywords.

NOTE 1 Tests with representative speech facilitate the comparison of results.

NOTE 2 For some applications, e.g., localized speech tests, a specific dialect can be required.

A reference recording mainly consists of the elements specified in 4.1. The reference recording shall not be modified. All speech materials used for testing shall be equivalent copies of the reference recording. Any changes of the reference recording (e.g., a new recording or changes of levels or frequency content) require a new validation (see Clause 5).

For speech audiometry intended to be performed in noise, the reference recording shall also include the associated competing sound or sounds, either on the same or on separate channels.

© ISO 2022 – All rights reserved

Documents
Document Sharing Hub
Full document details: BS EN ISO 8253-3:2022