

BS EN ISO 8835-3:2009+A1:2010



BSI Standards Publication

Inhalational anaesthesia systems

Part 3: Transfer and receiving systems of active anaesthetic gas scavenging systems (ISO 8835-3:2007/AMD 1:2010)

bsi.

...making excellence a habit.
AVA Documents™
Document Sharing Hub

National foreword

This British Standard is the UK implementation of EN ISO 8835-3:2009+A1:2010. It is identical to ISO 8835-3:2007, incorporating amendment 1:2010.

The start and finish of text introduced or altered by amendment is indicated in the text by tags. Tags indicating changes to ISO text carry the number of the ISO amendment. For example, text altered by ISO amendment 1 is indicated by **A1** **A1**.

EN ISO 8835-3:2009 had not been previously implemented as a British Standard because of safety concerns over clauses 2, 5.2, 7.1 and 8.6, and it was subject to a formal appeal against ratification by BSI. The previous edition, EN ISO 8835-3:2007, was not implemented as a British standard for the same reason. The concern over safety has now been removed by the incorporation of amendment 1:2010.

The UK participation in its preparation was entrusted to Technical Committee CH/121/1, Breathing attachments and anaesthetic machines.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© BSI 2011

ISBN 978 0 580 71307 1

ICS 11.040.10

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 30 June 2011.

Amendments/corrigenda issued since publication

Date	Text affected
------	---------------

English Version

Inhalational anaesthesia systems - Part 3: Transfer and
receiving systems of active anaesthetic gas scavenging systems
(ISO 8835-3:2007/AMD 1:2010)

Systèmes d'anesthésie par inhalation - Partie 3: Systèmes
de transfert et de réception des systèmes d'évacuation des
gaz d'anesthésie (ISO 8835-3:2007/AMD 1:2010)

Systeme für die Inhalationsanästhesie - Teil 3:
Weiterleitungs- und Aufnahmesysteme von aktiven
Anästhesiegas-Fortleitungssystemen (ISO 8835-
3:2007/AMD 1:2010)

This amendment A1 modifies the European Standard EN ISO 8835-3:2009; it was approved by CEN on 14 October 2010.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for inclusion of this amendment into the relevant national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This amendment exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: Avenue Marnix 17, B-1000 Brussels

Foreword

The text of ISO 8835-3:2007 has been prepared by Technical Committee ISO/TC 121 “Anaesthetic and respiratory equipment” of the International Organization for Standardization (ISO) and has been taken over as EN ISO 8835-3:2009 by Technical Committee CEN/TC 215 “Respiratory and anaesthetic equipment” the secretariat of which is held by BSI.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by September 2009, and conflicting national standards shall be withdrawn at the latest by March 2010.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 8835-3:2007.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EC Directive.

For relationship with EC Directive, see informative Annex ZA, which is an integral part of this document.

Other European Standards relating to anaesthetic workstations and their components prepared by CEN/TC 215 which, together with EN 60601-2-13:2006, replace EN 740:1998 in total, are:

- EN ISO 8835-2:2007, Inhalational anaesthesia systems – Part 2: Anaesthetic breathing systems (ISO 8835-2:2007)
- EN ISO 8835-3:2007, Inhalational anaesthesia systems – Part 3: Transfer and receiving systems of active anaesthetic gas scavenging systems (ISO 8835-3:2007)
- EN ISO 8835-4:2004, Inhalational anaesthesia systems – Part 4: Anaesthetic vapour delivery devices (ISO 8835-4:2004)
- EN ISO 8835-5:2004, Inhalational anaesthesia systems – Part 5: Anaesthetic ventilators (ISO 8835-5:2004)

Attention is also drawn to ISO/TS 18835:2004, Inhalational anaesthesia systems — Draw-over vaporizers and associated equipment.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and the United Kingdom.

Endorsement notice

The text of ISO 8835-3:2007 has been approved by CEN as a EN ISO 8835-3:2009 without any modification.

Foreword to amendment A1

This document (EN ISO 8835-3:2009/A1:2010) has been prepared by Technical Committee ISO/TC 121 “Anaesthetic and respiratory equipment” of the International Organization for Standardization (ISO) in collaboration with Technical Committee CEN/TC 215 “Respiratory and anaesthetic equipment” the secretariat of which is held by BSI.

This Amendment to the European Standard EN ISO 8835:2009 shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by April 2011, and conflicting national standards shall be withdrawn at the latest by April 2011.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive.

For relationship with EU Directive, see informative Annex ZA, which is an integral part of this document.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and the United Kingdom.

Endorsement notice

The text of ISO 8835-3:2007/AMD 1:2010 has been approved by CEN as a EN ISO 8835-3:2009/A1:2010 without any modification.

Annex ZA (informative)

Relationship between this European Standard and the Essential Requirements of EU Directive 93/42/EEC

This European Standard has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association to provide a means of conforming to Essential Requirements of the New Approach Directive 93/42/EEC on medical devices.

Once this standard is cited in the Official Journal of the European Union under that Directive and has been implemented as a national standard in at least one Member State, compliance with the clauses of this standard given in Table ZA.1 confers, within the limits of the scope of this standard, a presumption of conformity with the corresponding Essential Requirements of that Directive and associated EFTA regulations.

Table ZA.1 - Relationship between this European Standard and the Essential Requirements of EU Directive 93/42/EEC

Clause(s)/sub-clause(s) of this EN	Essential Requirements (ERs) of Directive 93/42/EEC	Qualifying remarks/Notes
All	2, 12.7.4	
4.1	7.5 (1 st paragraph)	This relevant Essential Requirement is not completely addressed in this EN.
11, 12	1, 2 nd paragraph, 1 st and 2 nd dash	These relevant Essential Requirements are not fully addressed in this EN
11, 12	13.1	
11, 12	13.3 b)	
11 d)	13.6 h)	
	13.6 q)	This relevant Essential Requirement is not addressed in this EN
12 b)	13.3 a)	This relevant Essential Requirement is not completely addressed in this EN

Warning – Other requirements and other EU Directives may be applicable to the products falling within the scope of this International standard.

Contents

Page

Foreword.....	vi
Introduction	vii
* 1 Scope	1
2 Normative references	1
3 Terms and definitions.....	2
4 General requirements and alternative test methods.....	4
4.1 Materials	4
4.2 Means of pressure relief	4
4.3 Alternative test methods.....	4
* 5 Patient and environmental protection	4
5.1 Normal operating conditions.....	4
5.2 Single fault conditions	5
* 6 Connectors	5
* 7 Transfer systems	5
7.1 Inlet of transfer systems	5
7.2 Outlet of transfer systems	6
* 8 Receiving systems.....	6
8.1 General.....	6
8.2 Inlet of receiving systems.....	6
8.3 Outlet of receiving systems	7
8.4 Visual indicator	7
8.5 Particle filter	7
8.6 Receiving hoses.....	7
9 Extract flow resistance.....	7
9.1 Low-flow transfer and receiving systems	7
9.2 High-flow transfer and receiving systems	7
10 Electrical requirements	8
* 11 Information to be supplied by the manufacturer.....	8
12 Marking	8
* 13 Colour coding.....	8
Annex A (informative) Typical test arrangement and method for pressure measurement at inlet to AGSS	9
Annex B (informative) Typical test arrangement and method for sub-atmospheric pressure limitation	13
Annex C (informative) Typical test arrangement and method for testing of induced flow.....	14
Annex D (informative) Typical test arrangement and method for resistance to extract flow.....	15
Annex E (informative) Typical test arrangement and method for spillage.....	17
Annex F (informative) Rationale	20
Annex G (informative) Environmental aspects	22
Bibliography	23

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 8835-3 was prepared by Technical Committee ISO/TC 121, *Anaesthetic and respiratory equipment*, Subcommittee SC 1, *Breathing attachments and anaesthetic machines*.

This second edition cancels and replaces the first edition (ISO 8835-3:1997), which has been technically revised.

ISO 8835 consists of the following parts, under the general title *Inhalational anaesthesia systems*:

- *Part 2: Anaesthetic breathing systems*
- *Part 3: Transfer and receiving systems of active anaesthetic gas scavenging systems*
- *Part 4: Anaesthetic vapour delivery devices*
- *Part 5: Anaesthetic ventilators*

Introduction

This part of ISO 8835 is intended to ensure that, for all practical purposes, an active AGSS will remove essentially all gases delivered to it and thereby reduce atmospheric pollution to a small fraction of the uncontrolled level.

It is recognized that there are many factors affecting conditions within the operator's working environment, which are outside the control of manufacturers of active AGSSs. These include room ventilation, leakage from equipment and the choice of anaesthetic technique, all of which are variable. Furthermore, the amount of pollutant taken up by personnel will be affected by other factors, such as the duration of exposure, their position in relation to any source of pollution, etc.

Atmospheric pollution by anaesthetic gases is the subject of considerable discussion, and opinions differ as to the limits that should be allowed in the working environment. Recommendations on permissible levels are therefore not included in this part of ISO 8835 but can be specified in national standards.

The committee responsible for this part of ISO 8835 has been primarily concerned with limiting the risks to the patient, which the transfer and receiving systems of AGSS can introduce by altering the function of breathing systems. The wide range of anaesthetic machines, ventilators and related equipment in general use today has been taken into account.

Annex F contains rationale statements for some of the requirements of this part of ISO 8835. The clauses and subclauses marked with an asterix (*) before their number have corresponding rationale contained in Annex F, included to provide additional insight into the reasoning that led to the requirements and recommendations that have been incorporated in this International Standard.

This page deliberately set blank

Inhalational anaesthesia systems —

Part 3:

Transfer and receiving systems of active anaesthetic gas scavenging systems

* 1 Scope

This part of ISO 8835 specifies requirements for transfer and receiving systems of active anaesthetic gas scavenging systems (active AGSSs) intended to reduce exposure of healthcare personnel to anaesthetic gases and vapours while providing patient protection (e.g. against excessive flow and pressure). This part of ISO 8835 also specifies requirements for transfer and receiving systems of active anaesthetic gas scavenging systems in which the power device is integral with the transfer and receiving system.

This part of ISO 8835 does not specify requirements for

- disposal systems which are covered by ISO 7396-2,
- non-active AGSSs (passive AGSSs),
- proximity gas extraction systems (i.e. systems not directly connected to the breathing system or associated equipment),
- transfer and receiving systems intended for use with flammable anaesthetic as determined by Annex DD of IEC 60601-2-13:2003.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 594-2, *Conical fittings with a 6 % (Luer) taper for syringes, needles and certain other medical equipment — Part 2: Lock fittings*

ISO 4135, *Anaesthetic and respiratory equipment — Vocabulary*

ISO 5356-1, *Anaesthetic and respiratory equipment — Conical connectors: Part 1: Cones and sockets*

ISO 5356-2, *Anaesthetic and respiratory equipment — Conical connectors: Part 2: Screw-threaded weight-bearing connections*

ISO 5359:2008, *Low-pressure hose assemblies for use with medical gases*

ISO 7000:2004, *Graphical symbols for use on equipment — Index and synopsis*

ISO 7396-2, *Medical gas pipeline systems — Part 2: Anaesthetic gas scavenging disposal systems*