



BSI Standards Publication

Health informatics — Device interoperability

Part 10407: Personal health device communication —
Device specialization — Blood pressure monitor

National foreword

This British Standard is the UK implementation of EN ISO/IEEE 11073-10407:2022. It is identical to ISO/IEEE 11073-10407:2022. It supersedes BS EN ISO 11073-10407:2011, which is withdrawn.

The UK participation in its preparation was entrusted to Technical Committee IST/35, Health informatics.

A list of organizations represented on this committee can be obtained on request to its committee manager.

Contractual and legal considerations

This publication has been prepared in good faith, however no representation, warranty, assurance or undertaking (express or implied) is or will be made, and no responsibility or liability is or will be accepted by BSI in relation to the adequacy, accuracy, completeness or reasonableness of this publication. All and any such responsibility and liability is expressly disclaimed to the full extent permitted by the law.

This publication is provided as is, and is to be used at the recipient's own risk.

The recipient is advised to consider seeking professional guidance with respect to its use of this publication.

This publication is not intended to constitute a contract. Users are responsible for its correct application.

© The British Standards Institution 2023
Published by BSI Standards Limited 2023

ISBN 978 0 539 22600 3

ICS 35.240.80

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 31 January 2023.

Amendments/corrigenda issued since publication

Date	Text affected
------	---------------

English Version

Health informatics - Device interoperability - Part 10407:
Personal health device communication - Device
specialization - Blood pressure monitor (ISO/IEEE 11073-
10407:2022)

Informatique de santé - Interopérabilité des dispositifs
- Partie 10407: Communication entre dispositifs de
santé personnels - Spécialisation des dispositifs -
Moniteur de pression sanguine (ISO/IEEE 11073-
10407:2022)

Medizinische Informatik - Kommunikation von Geräten
für die persönliche Gesundheit - Teil 10407:
Gerätespezifikation - Blutdruckmonitor (ISO/IEEE
11073-10407:2022)

This European Standard was approved by CEN on 4 December 2022.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

European foreword

This document (EN ISO/IEEE 11073-10407:2022) has been prepared by Technical Committee ISO/TC 215 "Health informatics" in collaboration with Technical Committee CEN/TC 251 "Health informatics" the secretariat of which is held by NEN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by June 2023, and conflicting national standards shall be withdrawn at the latest by June 2023.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 11073-10407:2011.

Any feedback and questions on this document should be directed to the users' national standards body/national committee. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

Endorsement notice

The text of ISO/IEEE 11073-10407:2022 has been approved by CEN as EN ISO/IEEE 11073-10407:2022 without any modification.

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted (see www.iso.org/directives).

IEEE Standards documents are developed within the IEEE Societies and the Standards Coordinating Committees of the IEEE Standards Association (IEEE-SA) Standards Board. The IEEE develops its standards through a consensus development process, approved by the American National Standards Institute, which brings together volunteers representing varied viewpoints and interests to achieve the final product. Volunteers are not necessarily members of the Institute and serve without compensation. While the IEEE administers the process and establishes rules to promote fairness in the consensus development process, the IEEE does not independently evaluate, test, or verify the accuracy of any of the information contained in its standards.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

ISO/IEEE 11073-10407 was prepared by the *IEEE 11073 Standards Committee of the IEEE Engineering in Medicine and Biology Society* (as IEEE Std 11073-10407-2020) and drafted in accordance with its editorial rules. It was adopted, under the “fast-track procedure” defined in the Partner Standards Development Organization cooperation agreement between ISO and IEEE, by Technical Committee ISO/TC 215, *Health informatics*.

This second edition cancels and replaces the first edition (ISO/IEEE 11073-10407:2010), which has been technically revised.

A list of all parts in the ISO/IEEE 11073 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Health informatics—Personal health device communication

Part 10407: Device specialization— Blood pressure monitor

Developed by the

IEEE 11073™ Standards Committee
of the
IEEE Engineering in Medicine and Biology Society

Approved 30 January 2020

IEEE SA Standards Board

Abstract: Within the context of the ISO/IEEE 11073 family of standards for device communication, this standard establishes a normative definition of communication between personal telehealth blood pressure monitor devices and compute engines (e.g., cell phones, personal computers, personal health appliances, and set top boxes) in a manner that enables plug-and-play interoperability. It leverages appropriate portions of existing standards including ISO/IEEE 11073 terminology, information models, application profile standards, and transport standards. It specifies the use of specific term codes, formats, and behaviors in telehealth environments restricting optionality in base frameworks in favor of interoperability. This standard defines a common core of communication functionality for personal telehealth blood pressure monitors.

Keywords: blood pressure monitor, IEEE 11073-10407™, medical device communication, personal health devices

The Institute of Electrical and Electronics Engineers, Inc.
3 Park Avenue, New York, NY 10016-5997, USA

Copyright © 2020 by The Institute of Electrical and Electronics Engineers, Inc.
All rights reserved. Published 30 April 2020. Printed in the United States of America.

IEEE is a registered trademark in the U.S. Patent & Trademark Office, owned by The Institute of Electrical and Electronics Engineers, Incorporated.

PDF: ISBN 978-1-5044-6460-4 STD24062
Print: ISBN 978-1-5044-6461-1 STDPD24062

IEEE prohibits discrimination, harassment, and bullying.

For more information, visit <https://www.ieee.org/about/corporate/governance/p9-26.html>.

No part of this publication may be reproduced in any form, in an electronic retrieval system or otherwise, without the prior written permission of the publisher.

Important Notices and Disclaimers Concerning IEEE Standards Documents

IEEE documents are made available for use subject to important notices and legal disclaimers. These notices and disclaimers, or a reference to this page, appear in all standards and may be found under the heading “Important Notices and Disclaimers Concerning IEEE Standards Documents.” They can also be obtained on request from IEEE or viewed at <https://standards.ieee.org/ipr/disclaimers.html>.

Notice and Disclaimer of Liability Concerning the Use of IEEE Standards Documents

IEEE Standards documents (standards, recommended practices, and guides), both full-use and trial-use, are developed within IEEE Societies and the Standards Coordinating Committees of the IEEE Standards Association (“IEEE SA”) Standards Board. IEEE (“the Institute”) develops its standards through a consensus development process, approved by the American National Standards Institute (“ANSI”), which brings together volunteers representing varied viewpoints and interests to achieve the final product. IEEE Standards are documents developed through scientific, academic, and industry-based technical working groups. Volunteers in IEEE working groups are not necessarily members of the Institute and participate without compensation from IEEE. While IEEE administers the process and establishes rules to promote fairness in the consensus development process, IEEE does not independently evaluate, test, or verify the accuracy of any of the information or the soundness of any judgments contained in its standards.

IEEE Standards do not guarantee or ensure safety, security, health, or environmental protection, or ensure against interference with or from other devices or networks. Implementers and users of IEEE Standards documents are responsible for determining and complying with all appropriate safety, security, environmental, health, and interference protection practices and all applicable laws and regulations.

IEEE does not warrant or represent the accuracy or content of the material contained in its standards, and expressly disclaims all warranties (express, implied and statutory) not included in this or any other document relating to the standard, including, but not limited to, the warranties of: merchantability; fitness for a particular purpose; non-infringement; and quality, accuracy, effectiveness, currency, or completeness of material. In addition, IEEE disclaims any and all conditions relating to: results; and workmanlike effort. IEEE standards documents are supplied “AS IS” and “WITH ALL FAULTS.”

Use of an IEEE standard is wholly voluntary. The existence of an IEEE standard does not imply that there are no other ways to produce, test, measure, purchase, market, or provide other goods and services related to the scope of the IEEE standard. Furthermore, the viewpoint expressed at the time a standard is approved and issued is subject to change brought about through developments in the state of the art and comments received from users of the standard.

In publishing and making its standards available, IEEE is not suggesting or rendering professional or other services for, or on behalf of, any person or entity nor is IEEE undertaking to perform any duty owed by any other person or entity to another. Any person utilizing any IEEE Standards document, should rely upon his or her own independent judgment in the exercise of reasonable care in any given circumstances or, as appropriate, seek the advice of a competent professional in determining the appropriateness of a given IEEE standard.

IN NO EVENT SHALL IEEE BE LIABLE FOR ANY DIRECT, INDIRECT, INCIDENTAL, SPECIAL, EXEMPLARY, OR CONSEQUENTIAL DAMAGES (INCLUDING, BUT NOT LIMITED TO: PROCUREMENT OF SUBSTITUTE GOODS OR SERVICES; LOSS OF USE, DATA, OR PROFITS; OR BUSINESS INTERRUPTION) HOWEVER CAUSED AND ON ANY THEORY OF LIABILITY, WHETHER IN CONTRACT, STRICT LIABILITY, OR TORT (INCLUDING NEGLIGENCE OR OTHERWISE) ARISING IN ANY WAY OUT OF THE PUBLICATION, USE OF, OR RELIANCE UPON ANY STANDARD, EVEN IF ADVISED OF THE POSSIBILITY OF SUCH DAMAGE AND REGARDLESS OF WHETHER SUCH DAMAGE WAS FORESEEABLE.

Translations

The IEEE consensus development process involves the review of documents in English only. In the event that an IEEE standard is translated, only the English version published by IEEE should be considered the approved IEEE standard.

Official statements

A statement, written or oral, that is not processed in accordance with the IEEE SA Standards Board Operations Manual shall not be considered or inferred to be the official position of IEEE or any of its committees and shall not be considered to be, or be relied upon as, a formal position of IEEE. At lectures, symposia, seminars, or educational courses, an individual presenting information on IEEE standards shall make it clear that his or her views should be considered the personal views of that individual rather than the formal position of IEEE.

Comments on standards

Comments for revision of IEEE Standards documents are welcome from any interested party, regardless of membership affiliation with IEEE. However, IEEE does not provide consulting information or advice pertaining to IEEE Standards documents. Suggestions for changes in documents should be in the form of a proposed change of text, together with appropriate supporting comments. Since IEEE standards represent a consensus of concerned interests, it is important that any responses to comments and questions also receive the concurrence of a balance of interests. For this reason, IEEE and the members of its societies and Standards Coordinating Committees are not able to provide an instant response to comments or questions except in those cases where the matter has previously been addressed. For the same reason, IEEE does not respond to interpretation requests. Any person who would like to participate in revisions to an IEEE standard is welcome to join the relevant IEEE working group.

Comments on standards should be submitted to the following address:

Secretary, IEEE SA Standards Board
445 Hoes Lane
Piscataway, NJ 08854 USA

Laws and regulations

Users of IEEE Standards documents should consult all applicable laws and regulations. Compliance with the provisions of any IEEE Standards document does not imply compliance to any applicable regulatory requirements. Implementers of the standard are responsible for observing or referring to the applicable regulatory requirements. IEEE does not, by the publication of its standards, intend to urge action that is not in compliance with applicable laws, and these documents may not be construed as doing so.

Copyrights

IEEE draft and approved standards are copyrighted by IEEE under US and international copyright laws. They are made available by IEEE and are adopted for a wide variety of both public and private uses. These include both use, by reference, in laws and regulations, and use in private self-regulation, standardization, and the promotion of engineering practices and methods. By making these documents available for use and adoption by public authorities and private users, IEEE does not waive any rights in copyright to the documents.

Photocopies

Subject to payment of the appropriate fee, IEEE will grant users a limited, non-exclusive license to photocopy portions of any individual standard for company or organizational internal use or individual, non-commercial use only. To arrange for payment of licensing fees, please contact Copyright Clearance Center, Customer Service, 222 Rosewood Drive, Danvers, MA 01923 USA; +1 978 750 8400. Permission to photocopy portions of any individual standard for educational classroom use can also be obtained through the Copyright Clearance Center.

Updating of IEEE Standards documents

Users of IEEE Standards documents should be aware that these documents may be superseded at any time by the issuance of new editions or may be amended from time to time through the issuance of amendments, corrigenda, or errata. An official IEEE document at any point in time consists of the current edition of the document together with any amendments, corrigenda, or errata then in effect.

Every IEEE standard is subjected to review at least every 10 years. When a document is more than 10 years old and has not undergone a revision process, it is reasonable to conclude that its contents, although still of some value, do not wholly reflect the present state of the art. Users are cautioned to check to determine that they have the latest edition of any IEEE standard.

In order to determine whether a given document is the current edition and whether it has been amended through the issuance of amendments, corrigenda, or errata, visit IEEE Xplore at <https://ieeexplore.ieee.org/> or contact IEEE at the address listed previously. For more information about the IEEE SA or IEEE's standards development process, visit the IEEE SA Website at <https://standards.ieee.org>.

Errata

Errata, if any, for IEEE standards can be accessed via <https://standards.ieee.org/standard/index.html>. Search for standard number and year of approval to access the web page of the published standard. Errata links are located under the Additional Resources Details section. Errata are also available in IEEE Xplore: <https://ieeexplore.ieee.org/browse/standards/collection/ieee/>. Users are encouraged to periodically check for errata.

Patents

Attention is called to the possibility that implementation of this standard may require use of subject matter covered by patent rights. By publication of this standard, no position is taken by the IEEE with respect to the existence or validity of any patent rights in connection therewith. If a patent holder or patent applicant has filed a statement of assurance via an Accepted Letter of Assurance, then the statement is listed on the IEEE SA Website at <https://standards.ieee.org/about/sasb/patcom/patents.html>. Letters of Assurance may indicate whether the Submitter is willing or unwilling to grant licenses under patent rights without compensation or under reasonable rates, with reasonable terms and conditions that are demonstrably free of any unfair discrimination to applicants desiring to obtain such licenses.

Essential Patent Claims may exist for which a Letter of Assurance has not been received. The IEEE is not responsible for identifying Essential Patent Claims for which a license may be required, for conducting inquiries into the legal validity or scope of Patent Claims, or determining whether any licensing terms or conditions provided in connection with submission of a Letter of Assurance, if any, or in any licensing agreements are reasonable or non-discriminatory. Users of this standard are expressly advised that determination of the validity of any patent rights, and the risk of infringement of such rights, is entirely their own responsibility. Further information may be obtained from the IEEE Standards Association.

Participants

At the time this IEEE standard was completed, the Personal Health Device Working Group had the following membership:

Daidi Zhong, Co-Chair
Michael J. Kirwan, Co-Chair

Karsten Aalders	Cory Condek	Robert Hall
Charles R. Abbruscato	Todd H. Cooper	Shu Han
Nabil Abujbara	David Cornejo	Nathaniel Hamming
Maher Abuzaid	Douglas Coup	Rickey L. Hampton
James Agnew	Nigel Cox	Sten Hanke
Manfred Aigner	Hans Crommenacker	Aki Harma
Jorge Alberola	Tomio Crosley	Jordan Hartmann
David Aparisi	Allen Curtis	Kai Hassing
Lawrence Arne	Jesús Daniel Trigo	Wolfgang Heck
Diego B. Arquillo	David Davenport	Nathaniel Heintzman
Serafin Arroyo	Russell Davis	Charles Henderson
Muhammad Asim	Sushil K. Deka	Jun-Ho Her
Kit August	Ciro de la Vega	Helen B. Hernandez
Doug Baird	Pedro de-las-Heras-Quiros	Timothy L. Hirou
David Baker	Jim Dello Stritto	Allen Hobbs
Anindya Bakshi	Kent Dicks	Alex Holland
Ananth Balasubramanian	Hyoungdo Do	Arto Holopainen
Sunlee Bang	Jonathan Dougherty	Kris Holtzclaw
M. Jonathan Barkley	Xiaolian Duan	Xinyi Hong
Gilberto Barrón	Sourav Dutta	Robert Hoy
David Bean	Jakob Ehrensvar	Di Hu
John Bell	Fredrik Einberg	Anne Huang
Olivia Bellamou-Huet	Javier Escayola Calvo	Ron Huby
Rudy Belliardi	Mark Estes	David Hughes
Daniel Bernstein	Leonardo Estevez	Robert D. Hughes
George A. Bertos	Bosco T. Fernandes	Jiyoung Huh
Chris Biernacki	Christoph Fischer	Hugh Hunter
Ola Björnsne	Morten Flintrup	Philip O. Isaacson
Thomas Blackadar	Joseph W. Forler	Atsushi Ito
Thomas Bluethner	Russell Foster	Michael Jaffe
Douglas P. Bogia	Eric Freudenthal	Praduman Jain
Xavier Boniface	Matthias Frohner	Hu Jin
Shannon Boucousis	Ken Fuchs	Danny Jochelson
Julius Broma	Jing Gao	Akiyoshi Kabe
Lyle G. Bullock, Jr.	Marcus Garbe	Steve Kahle
Bernard Burg	John Garguilo	Tomio Kamioka
Chris Burns	Rick Geimer	James J. Kang
Jeremy Byford-Rew	Igor Gejdos	Kei Kariya
Satya Calloji	Ferenc Gerbovics	Andy Kaschl
Carole C. Carey	Alan Godfrey	Junzo Kashiara
Craig Carlson	Nicolae Goga	Colin Kennedy
Santiago Carot-Nemesio	Julian Goldman	Ralph Kent
Randy W. Carroll	Raul Gonzalez Gomez	Laurie M. Kermes
Seungchul Chae	Chris Gough	Ahmad Kheirandish
Peggy Chien	Channa Gowda	Junhyung Kim
David Chiu	Charles M. Gropper	Minho Kim
Jinyong Choi	Amit Gupta	Min-Joon Kim
Chia-Chin Chong	Jeff Guttmacher	Taekon Kim
Saeed A. Choudhary	Rasmus Haahr	Tetsuya Kimura
Jinhan Chung	Christian Habermann	Alfred Kloos
John A. Cogan	Michael Hagerty	Jeongmee Koh
John T. Collins	Jerry Hahn	Jean-Marc Koller

John Koon	BegonyaOtal	Nicholas Steblay
Patty Krantz	Marco Paleari	Lars Steubesand
Raymond Krasinski	Bud Panjwani	John (Ivo) Stivorice
Alexander Kraus	Carl Pantiskas	Raymond A. Strickland
Ramesh Krishna	Harry P. Pappas	Chandrasekaran Subramaniam
Geoffrey Kruse	Hanna Park	Hermann Suominen
Falko Kuester	Jong-Tae Park	Lee Surprenant
Rafael Lajara	Myungeun Park	Ravi Swami
Pierre Landau	Soojun Park	Ray Sweidan
Jaechul Lee	Phillip E. Pash	Na Tang
Jong-Muk Lee	Tong-Bi Pei	Haruyuyki Tatsumi
Kyong Ho Lee	Soren Petersen	Isabel Tejero
Rami Lee	James Petisce	Tom Thompson
Sungkee Lee	Peter Piction	Jonas Tirén
Woojae Lee	Michael Pliskin	Janet Traub
Qiong Li	Varshney Prabodh	Gary Tschautscher
Xiangchen Li	Jeff Price	Masato Tsuchid
Yingsong Li	Harald Prinzhorn	Ken Tubman
Zhuofang Li	Harry Qiu	Akib Uddin
Patrick Lichter	Tanzilur Rahman	Sunil Unadkat
Lin Lin	Phillip Raymond	Fabio Urbani
Jisoon Lim	Terrie Reed	Philipp Urbauer
Joon-Ho Lim	Barry Reinhold	Laura Vanzago
Liang Liu	Brian Reinhold	Alpo Värri
Xiaoming Liu	Melvin I. Reynolds	Andrei Vasilateanu
Wei-Jung Lo	John G. Rhoads	Dalimar Velez
Charles Lowe	Jeffrey S. Robbins	Martha Velez
Don Ludolph	Chris Roberts	Rudi Voon
Christian Lusztick	Moskowitz Robert	Barry Vornbrock
Bob MacWilliams	Stefan Robert	Isobel Walker
Srikkanth Madhurbootheswaran	Scott M. Robertson	David Wang
Miriam L. Makhoul	Timothy Robertson	Linling Wang
Romain Marmot	David Rosales	Jerry P. Wang
Sandra Martinez	Bill Saltzstein	Yao Wang
Miguel Martínez de Espronceda	Giovanna Sannino	Yi Wang
Cámara	Jose A. Santos-Cadenas	Steve Warren
Peter Mayhew	Stefan Sauermann	Fujio Watanabe
Jim McCain	John Sawyer	Toru Watsuji
László Meleg	Alois Schloegl	Kathleen Wible
Alexander Mense	Paul S. Schluter	Paul Williamson
Behnaz Minaei	Mark G. Schnell	Jan Wittenber
Jinsei Miyazaki	Richard A. Schrenker	Jia-Rong Wu
Erik Moll	Antonio Scorpiniti	Will Wykeham
Darr Moore	KwangSeok Seo	Ariton Xhafa
Chris Morel	Riccardo Serafin	Ricky Yang
Carsten Mueglitz	Sid Shaw	Melanie S. Yeung
Soundharya Nagasubramanian	Frank Shen	Qiang Yin
Alex Neefus	Min Shih	Done-Sik Yoo
Trong-Nghia Nguyen-Dobinsky	Mazen Shihabi	Zhi Yu
Michael E. Nidd	Redmond Shouldice	Jianchao Zeng
Jim Niswander	Sternly K. Simon	Jason Zhang
Hiroaki Niwamoto	Marjorie Skubic	Jie Zhao
Thomas Norgall	Robert Smith	Thomas Zhao
Yoshiteru Nozoe	Ivan Soh	Yuanhong Zhong
Abraham Ofek	Motoki Sone	Miha Zoubek
Brett Olive	Emily Sopensky	Szymon Zyskoter
	Rajagopalan Srinivasan	

The following members of the individual Standards Association balloting group voted on this standard. Balloters may have voted for approval, disapproval, or abstention.

Robert Aiello
Bjoern Andersen
Lyle G. Bullock, Jr.
Keith Chow
Malcolm Clarke
David Fuschi
Randall Groves
Nathaniel Hamming

Werner Hoelzl
Noriyuki Ikeuchi
Atsushi Ito
Raj Jain
Piotr Karocki
Raymond Krasinski
H. Moll

Iulian Profir
Beth Pumo
Stefan Schlichting
Janek Schumann
Walter Struppler
Oren Yuen
Janusz Zalewski
Daidi Zhong

When the IEEE SA Standards Board approved this standard on 30 January 2020, it had the following membership:

Gary Hoffman, *Chair*
Vacant Position, *Vice Chair*
Jean-Philippe Faure, *Past Chair*
Konstantinos Karachalios, *Secretary*

Ted Burse
J. Travis Griffith
Grace Gu
Guido R. Hiertz
Joseph L. Koepfinger*
John D. Kulick
David J. Law

Howard Li
Dong Liu
Kevin Lu
Paul Nikolich
Damir Novosel
Jon Walter Rosdahl

Dorothy Stanley
Mehmet Ulema
Lei Wang
Sha Wei
Philip B. Winston
Daidi Zhong
Jingyi Zhou

*Member Emeritus

Introduction

This introduction is not part of IEEE Std 11073-10407-2020, Health informatics—Personal health device communication—Part 10407: Device specialization—Blood pressure monitor.

ISO/IEEE 11073 standards enable communication between medical devices and external computer systems. This document uses the optimized framework created in IEEE Std 11073-20601™-2019 and describes a specific, interoperable communication approach for blood pressure monitors.^a These standards align with and draw on the existing clinically focused standards to provide support for communication of data from personal health devices.

^a Information on references can be found in Clause 2.

Contents

1. Overview	12
1.1 Scope	12
1.2 Purpose	12
1.3 Word usage	12
1.4 Context	13
2. Normative references.....	13
3. Definitions, acronyms, and abbreviations	14
3.1 Definitions	14
3.2 Acronyms and abbreviations	15
4. Introduction to ISO/IEEE 11073 personal health devices	15
4.1 General	15
4.2 Introduction to IEEE 11073-20601 modeling constructs	16
4.3 Compliance with other standards.....	17
5. Blood pressure monitor device concepts and modalities.....	17
5.1 General	17
5.2 Systolic and diastolic pressure	18
5.3 Mean arterial pressure.....	18
5.4 Pulse rate	18
5.5 Blood pressure measurement status.....	18
6. Blood pressure monitor domain information model.....	18
6.1 Overview	18
6.2 Class extensions.....	18
6.3 Object instance diagram	18
6.4 Types of configuration.....	20
6.5 Medical device system object.....	20
6.6 Numeric objects.....	25
6.7 Real-time sample array objects.....	30
6.8 Enumeration objects	34
6.9 PM-store objects.....	35
6.10 Scanner objects.....	35
6.11 Class extension objects.....	36
6.12 Blood pressure monitor information model extensibility rules.....	36
7. Blood pressure monitor service model	36
7.1 General	36
7.2 Object access services.....	36
7.3 Object access event report services	38
8. Blood pressure monitor communication model.....	38
8.1 Overview	38
8.2 Communication characteristics.....	38
8.3 Association procedure	39
8.4 Configuring procedure.....	41
8.5 Operating procedure	43
8.6 Time synchronization	43

9. Test associations	44
9.1 General	44
9.2 Behavior with standard configuration	44
9.3 Behavior with extended configurations	44
10. Conformance	44
10.1 Applicability	44
10.2 Conformance specification	45
10.3 Levels of conformance	45
10.4 Implementation conformance statements	45
Annex A (informative) Bibliography	51
Annex B (normative) Additional ASN.1 definitions	52
B.1 Device and sensor status bit mapping	52
Annex C (normative) Allocation of identifiers	53
Annex D (informative) Message sequence examples	55
Annex E (informative) Protocol data unit examples	57
E.1 General	57
E.2 Association information exchange	57
E.3 Configuration information exchange	60
E.4 GET MDS attributes service	63
E.5 Data reporting	65
E.6 Disassociation	66
Annex F (informative) Revision history	67

Health informatics—Personal health device communication

Part 10407: Device specialization— Blood pressure monitor

1. Overview

1.1 Scope

Within the context of the ISO/IEEE 11073 family of standards for device communication, this standard establishes a normative definition of communication between personal telehealth blood pressure monitor devices and compute engines (e.g., cell phones, personal computers, personal health appliances, and set top boxes) in a manner that enables plug-and-play interoperability. It leverages appropriate portions of existing standards including ISO/IEEE 11073 terminology, information models, application profile standards, and transport standards. It specifies the use of specific term codes, formats, and behaviors in telehealth environments restricting optionality in base frameworks in favor of interoperability. This standard defines a common core of communication functionality for personal telehealth blood pressure monitors.

1.2 Purpose

This standard addresses a need for an openly defined, independent standard for controlling information exchange to and from personal health devices and compute engines (e.g., cell phones, personal computers, personal health appliances, and set top boxes). Interoperability is the key to growing the potential market for these devices and to enabling people to be better informed participants in the management of their health.

1.3 Word usage

The word *shall* indicates mandatory requirements strictly to be followed in order to conform to the standard and from which no deviation is permitted (*shall* equals *is required to*).^{1,2}

¹ The use of the word *must* is deprecated and cannot be used when stating mandatory requirements; *must* is used only to describe unavoidable situations.

² The use of *will* is deprecated and cannot be used when stating mandatory requirements; *will* is used only in statements of fact.

IEEE P11073-10407/D12
Health informatics—Personal health device communication
Part 10407: Device specialization—Blood pressure monitor

The word *should* indicates that among several possibilities one is recommended as particularly suitable, without mentioning or excluding others; or that a certain course of action is preferred but not necessarily required (*should* equals *is recommended that*).

The word *may* is used to indicate a course of action permissible within the limits of the standard (*may* equals *is permitted to*).

The word *can* is used for statements of possibility and capability, whether material, physical, or causal (*can* equals *is able to*).

1.4 Context

See IEEE Std 11073-20601-2019TM for an overview of the environment within which this standard is written.³

This document, IEEE Std 11073-10407, defines the device specialization for the blood pressure monitor, being a specific agent type, and provides a description of the device concepts, its capabilities, and its implementation according to this standard.

This standard is based on IEEE Std 11073-20601-2019, which in turn draws information from both ISO/IEEE 11073-10201:2004 [B6] and ISO/IEEE 11073-20101:2004 [B7].⁴ The medical device encoding rules (MDERs) used within this standard are fully described in IEEE Std 11073-20601-2019.

This standard defines specialized nomenclature codes that will be collected in future revisions of IEEE Std 11073-10101. Between this standard, IEEE Std 11073-10101-2019, IEEE Std 11073-20601-2019, and other IEEE Std 11073-104xx, all required nomenclature codes for implementation are documented. New codes may be defined in newer versions / revisions of each of these documents. In the case of a conflict, where one term code has been assigned to two separate semantic concepts with different RefIDs, in general the oldest definition that is in actual use should take precedence. The same policy applies when one RefID has two different code values assigned in different specifications. The resolution of such conflicts will be determined through joint action by the responsible working groups and other stakeholders, and any corrective actions will be published as corrigenda.

NOTE—In this standard, IEEE Std 11073-104zz is used to refer to the collection of device specialization standards that utilize IEEE Std 11073-20601-2019, where zz can be any number from 01 to 99, inclusive.⁵

2. Normative references

The following referenced documents are indispensable for the application of this document (i.e., they must be understood and used; therefore, each referenced document is cited in text, and its relationship to this document is explained). For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments or corrigenda) applies.

IEEE Std 11073-10101TM-2019, Health informatics—Point-of-care medical device communication—Part 10101: Nomenclature.^{6,7}

³ Information on references can be found in Clause 2.

⁴ The numbers in brackets correspond to the numbers of the bibliography in Annex A.

⁵ Notes in text, tables, and figures are given for information only and do not contain requirements needed to implement the standard.

⁶ The IEEE standards or products referred to in this clause are trademarks of The Institute of Electrical and Electronics Engineers, Inc.

⁷ IEEE publications are available from The Institute of Electrical and Electronics Engineers (<https://standards.ieee.org/>).

IEEE Std 11073-20601™-2019, Health informatics—Personal health device communication—Part 20601: Application profile—Optimized Exchange Protocol.

See Annex A for all informative material referenced by this standard.

3. Definitions, acronyms, and abbreviations

3.1 Definitions

For the purposes of this document, the following terms and definitions apply. The *IEEE Standards Dictionary Online* should be consulted for terms not defined in this clause.⁹

agent: A node that collects and transmits personal health data to an associated manager.

blood pressure: The cyclic pressure (i.e., amount of force applied over a given area divided by the size of this area) exerted by blood against the walls of blood vessels. Noninvasive blood pressure measurement is typically performed at the brachial artery (arm) or radial artery (wrist). There are usually two numbers reported for blood pressure, and with the home monitors, a third number is typically available. The first, and higher, number is produced by the contraction of the heart (See: systolic pressure). The second, lower number is produced by relaxation of the heart (See: diastolic pressure). The third number is the mean arterial pressure.

class: In object-oriented modeling, a class describes the attributes, methods, and events that objects instantiated from the class utilize.

compute engine: *See:* **manager.**

device: A term used to refer to a physical apparatus implementing either an agent or a manager role.

diastolic pressure: This is minimum pressure achieved during the cardiac cycle. It is typically the second and the lower of the readings given as the blood pressure.

handle: An unsigned 16-bit number that is locally unique and identifies one of the object instances within an agent.

manager: A node receiving data from one or more agent systems. Some examples of managers include a cellular phone, health appliance, set top box, or a computer system.

mean arterial pressure: value of the integral of one cycle of the blood pressure curve divided by the period between successive heart beats.

object: In object-oriented modeling, a particular instantiation of a class. The instantiation realizes attributes, methods, and events from the class.

obj-handle: *See:* **handle.**

personal health device: A device used in personal health applications.

personal telehealth device: *See:* **personal health device.**

⁹ *IEEE Standards Dictionary Online* is available at <https://dictionary.ieee.org>.

pulse: The frequency of the cardiac cycle as reported by the blood pressure monitor.

pulse pressure: The systolic pressure minus the diastolic pressure.

systolic pressure: This maximum value of the arterial blood pressure as a result of the contraction of the left ventricle. It is typically the first and the higher of the readings given as the blood pressure.

3.2 Acronyms and abbreviations

APDU	application protocol data unit
ASN.1	Abstract Syntax Notation One
BPM	beats per minute
DIM	domain information model
EUI-64	extended unique identifier (64 bits)
ICS	implementation conformance statement
MAP	mean arterial pressure
MDC	medical device communication
MDER	medical device encoding rules
MDS	medical device system
MOC	managed object class
PDU	protocol data unit
RT-SA	real-time sample array
PHD	personal health device
VMO	virtual medical object
VMS	virtual medical system

4. Introduction to ISO/IEEE 11073 personal health devices

4.1 General

This standard and the remainder of the series of ISO/IEEE 11073 personal health device (PHD) standards fit in the larger context of the ISO/IEEE 11073 series of standards. The full suite of standards enables agents to interconnect and interoperate with managers and with computerized health-care information systems. See IEEE Std 11073-20601-2019 for a description of the guiding principles for this series of ISO/IEEE 11073 Personal Health Device standards.

IEEE Std 11073-20601-2019 supports the modeling and implementation of an extensive set of personal health devices. This standard defines aspects of the blood pressure monitor device. It describes all aspects necessary to implement the application layer services and data exchange protocol between an ISO/IEEE 11073 PHD blood pressure monitor agent and a manager. This standard defines a subset of the objects and functionality contained in IEEE Std 11073-20601-2019 and extends and adds definitions where appropriate. All new definitions are given in Annex B in Abstract Syntax Notation One (ASN.1) [B8].

All nomenclature codes referenced in this standard are collected in Annex C. Annex C may contain definitions of codes that are used by this standard and that are not yet present in IEEE Std 11073-10101-2019 and IEEE Std 11073-20601-2019.

4.2 Introduction to IEEE 11073-20601 modeling constructs

4.2.1 General

The ISO/IEEE 11073 series of standards, and in particular the IEEE Std 11073-20601-2019, is based on an object-oriented systems management paradigm. The overall system model is divided into three principal components: the domain information model (DIM), the service model, and the communication model. See IEEE Std 11073-20601-2019 for a detailed description of the modeling constructs.

4.2.2 Domain information model

The DIM is a hierarchical model that describes an agent as a set of objects. These objects and their attributes represent the elements that control behavior and report on the status of the agent and data that an agent can communicate to a manager. Communication between the agent and the manager is defined by the application protocol in IEEE Std 11073-20601-2019.

4.2.3 Service model

The service model defines the conceptual mechanisms for the data exchange services. Such services are mapped to messages that are exchanged between the agent and the manager. Protocol messages within the ISO/IEEE 11073 series of standards are defined in ASN.1. The messages defined in IEEE Std 11073-20601-2019 can coexist with messages defined in other standard application profiles defined in the ISO/IEEE 11073 series of standards.

4.2.4 Communication model

In general, the communication model supports the topology of one or more agents communicating over logical point-to-point connections to a single manager. For each logical point-to-point connection, the dynamic system behavior is defined by a connection state machine as specified in IEEE Std 11073-20601-2019.

4.2.5 Implementing the models

An agent implementing this standard shall implement all mandatory elements of the information, service, and communication models as well as all conditional elements where the condition is met. The agent should implement the recommended elements, and it may implement any combination of the optional elements. A manager implementing this standard shall utilize at least one of the mandatory, conditional, recommended, or optional elements. In this context, “utilize” means use the element as part of the primary function of the manager device. For example, a manager whose primary function is to display data would need to display a piece of data in the element in order to utilize it.

4.3 Compliance with other standards

Devices that comply with this standard may also be required to comply with other domain- and device-specific standards that supersede the requirements of this standard with respect to issues including safety, reliability, and risk management. A user of this standard is expected to be familiar with all other such standards that apply and to comply with any higher specifications thus imposed.

Typically, medical devices should comply with the IEC 60601-1 [B1] base standards with respect to electrical and mechanical safety and any device-specific standard as might be defined in the IEC 60601-2 [B2] series of standards. Software aspects may apply through standards such as IEC 62304 [B3]. Devices that comply with this standard implement higher layers of network software and utilize lower layers as appropriate to the application. The requirements on performance of such applications and conformance are defined elsewhere and are outside the scope of this standard. Moreover, the use of any medical equipment is subject to risk assessment and risk management appropriate to the application. Some relevant examples are ISO 14971 [B5] and IEC 80001-1 [B4]. The requirements of such risk assessment, risk management, and conformance are outside the scope of this standard. The applicable versions of the referenced safety-related standards may differ per country.

5. Blood pressure monitor device concepts and modalities

5.1 General

This clause presents the general concepts of blood pressure monitor devices. In the context of personal health devices in this family of standards, a blood pressure monitor is a device that measures blood pressure [i.e., systolic, diastolic, and mean arterial pressure (MAP)] and pulse noninvasively. Blood pressure monitor devices considered in this standard typically inflate a cuff to occlude an artery and then to measure the reaction of the artery while the pressure is released with the results being converted into systolic, diastolic, and MAP values. The pulse rate is determined at the same time.

Blood pressure monitor devices may use a variety of techniques for measuring blood pressure and pulse rate. One typical method is the oscillometric method where oscillations in cuff pressure are analyzed to obtain blood pressure values. Another technique is the automated auscultatory method where the device uses a microphone to detect Korotkoff sounds during cuff deflation. Auscultatory devices measure the systolic and diastolic values and estimates the MAP.

In home monitors, the oscillometric method is typically used, allowing the measurement to be done electronically. On the oscillometric method, small pressure changes (oscillations) occur in the cuff as a result of blood pressure pulses during the inflation or deflation of the cuff and are detected. These oscillations, which first increase and then decrease, are stored together with the corresponding cuff pressure values in the automated sphygmomanometer. With these stored values, the systolic, diastolic, and mean blood pressure values can be mathematically derived using an appropriate algorithm.

Blood pressure is historically most frequently measured using the units of millimeters of mercury (mmHg). Kilopascals (kPa) may also be used. This standard supports both mmHg and kPa.