BS ISO 17973:2024

BSI Standards Publication

Surface chemical analysis — Medium-resolution Auger electron spectrometers — Calibration of energy scales for elemental analysis

Documents
Document Sharing Hub

Full document details: BS ISO 17973:2024

National foreword

This British Standard is the UK implementation of ISO 17973:2024. It supersedes BS ISO 17973:2016, which is withdrawn.

The UK participation in its preparation was entrusted to Technical Committee CII/60, Surface chemical analysis.

A list of organizations represented on this committee can be obtained on request to its committee manager.

Contractual and legal considerations

This publication has been prepared in good faith, however no representation, warranty, assurance or undertaking (express or implied) is or will be made, and no responsibility or liability is or will be accepted by BSI in relation to the adequacy, accuracy, completeness or reasonableness of this publication. All and any such responsibility and liability is expressly disclaimed to the full extent permitted by the law.

This publication is provided as is, and is to be used at the recipient's own risk.

The recipient is advised to consider seeking professional guidance with respect to its use of this publication.

This publication is not intended to constitute a contract. Users are responsible for its correct application.

© The British Standards Institution 2024
Published by BSI Standards Limited 2024

ISBN 978 0 539 25859 2

ICS 71.040.40

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 31 July 2024.

Amendments/corrigenda issued since publication

Date	Text affected

Documents
Document Sharing Hub

Full document details: BS ISO 17973:2024

International Standard

ISO 17973

Third edition
2024-07

Surface chemical analysis — Medium-resolution Auger electron spectrometers — Calibration of energy scales for elemental analysis

Analyse chimique des surfaces — Spectromètres d'électrons Auger à résolution moyenne — Étalonnage des échelles d'énergie pour l'analyse élémentaire

Reference number
ISO 17973:2024(en)

© ISO 2024

Documents
Document Sharing Hub

Full document details: BS ISO 17973:2024

COPYRIGHT PROTECTED DOCUMENT

© ISO 2024

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Documents
Document Sharing Hub
Full document details: BS ISO 17973:2024

Contents

Page

Foreword ..... **iv**
Introduction ..... **v**
1 Scope **1**
2 Normative references **1**
3 Terms and definitions **1**
4 Symbols and abbreviated terms **1**
5 Outline of method **2**
6 Energy scale calibration procedures **3**
6.1 Obtaining reference samples 3
6.2 Mounting samples 3
6.3 Cleaning samples 3
6.4 Choosing spectrometer settings for energy calibration 5
6.5 Operating the instrument 5
6.6 Measurement of reference peaks 6
6.7 Determining measured kinetic energies of reference peaks 7
6.8 Determining correction of instrument kinetic energy scale 8
6.9 Next calibration 9
Bibliography **11**

© ISO 2024 – All rights reserved

Documents
Document Sharing Hub
Full document details: BS ISO 17973:2024

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

ISO draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at www.iso.org/patents. ISO shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

The committee responsible for this document is Technical Committee ISO/TC 201, *Surface chemical analysis*, Subcommittee SC 7, *Electron spectroscopies.*

This third edition cancels and replaces the second edition (ISO 17973:2016), which has been technically revised.

The main changes are as follows:

— NOTE added to 6.3 to provide example sputtering conditions for cleaning sample surface;

— text added to 6.5 regarding need to check detector is operating within its linear region.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

© ISO 2024 – All rights reserved

Documents
Document Sharing Hub
Full document details: BS ISO 17973:2024